

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENTA05
 Lab Code: ACE SDG NO.: Q2579
 Lab File ID: VX046961.D Date Analyzed: 07/14/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:16
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	302863	5.56	503750	6.76	451308	10.05
UPPER LIMIT	605726	6.056	1007500	7.263	902616	10.549
LOWER LIMIT	151432	5.056	251875	6.263	225654	9.549
EPA SAMPLE NO.						
WC-A2-13-G	354879	5.56	603011	6.77	549695	10.06
WC-A2-14-G	353132	5.56	601642	6.77	548369	10.06
VX0714WBL01	330854	5.56	558126	6.77	513014	10.06
VX0714WBS01	311322	5.57	518025	6.77	465228	10.06
VX0714WBSD01	295675	5.56	497201	6.77	449456	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.