

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENTA05
 Lab Code: ACE SDG NO.: Q2579
 Lab File ID: VX046988.D Date Analyzed: 07/15/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:18
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	298555	5.56	499158	6.76	443564	10.06
UPPER LIMIT	597110	6.056	998316	7.263	887128	10.555
LOWER LIMIT	149278	5.056	249579	6.263	221782	9.555
EPA SAMPLE NO.						
WC-A2-13-GRE	352031	5.56	597406	6.77	550618	10.06
WC-A2-14-GRE	339200	5.56	586472	6.77	536571	10.06
VX0715WBL01	334020	5.56	564806	6.77	510782	10.06
VX0715WBS01	318386	5.56	538385	6.77	482579	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.