

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENTA05
 Lab Code: ACE SDG NO.: Q2579
 Lab File ID: VX046988.D Date Analyzed: 07/15/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:18
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	217822	12.018				
UPPER LIMIT	435644	12.518				
LOWER LIMIT	108911	11.518				
EPA SAMPLE NO.						
WC-A2-13-GRE	287789	12.02				
WC-A2-14-GRE	277997	12.02				
VX0715WBL01	265478	12.02				
VX0715WBS01	241598	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.