

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT

SDG No.: Q2580

Contract: ENTA05

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	16.7	20.0	83	65 - 135	P	07/14/2025	14:26	LB136469
	Barium	99.1	100	99	65 - 135	P	07/14/2025	14:26	LB136469
	Cadmium	5.57	6.0	93	65 - 135	P	07/14/2025	14:26	LB136469
	Chromium	10.6	10.0	106	65 - 135	P	07/14/2025	14:26	LB136469
	Lead	9.89	12.0	82	65 - 135	P	07/14/2025	14:26	LB136469
	Selenium	20.4	20.0	102	65 - 135	P	07/14/2025	14:26	LB136469
	Silver	11.9	10.0	119	65 - 135	P	07/14/2025	14:26	LB136469
CRI01	Arsenic	15.4	20.0	77	65 - 135	P	07/17/2025	14:39	LB136534
	Barium	101	100	102	65 - 135	P	07/17/2025	14:39	LB136534
	Cadmium	5.72	6.0	95	65 - 135	P	07/17/2025	14:39	LB136534
	Chromium	10.2	10.0	102	65 - 135	P	07/17/2025	14:39	LB136534
	Lead	11.9	12.0	99	65 - 135	P	07/17/2025	14:39	LB136534
	Selenium	19.6	20.0	98	65 - 135	P	07/17/2025	14:39	LB136534
	Silver	9.57	10.0	96	65 - 135	P	07/17/2025	14:39	LB136534
CRA	Mercury	0.23	0.2	116	70 - 130	CV	07/25/2025	11:46	LB136613