

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q2580

Contract: ENTA05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168846BS							
Arsenic	mg/Kg	40.0	35.5		89	80 - 120	P
Barium	mg/Kg	10.0	10.2		102	80 - 120	P
Cadmium	mg/Kg	10.0	9.35		94	80 - 120	P
Chromium	mg/Kg	20.0	19.6		98	80 - 120	P
Lead	mg/Kg	50.0	46.8		94	80 - 120	P
Selenium	mg/Kg	100	97.2		97	80 - 120	P
Silver	mg/Kg	3.8	3.68		97	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q2580

Contract: ENTA05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169010BS Mercury	mg/Kg	0.25	0.26		106	80 - 124	CV