

rteres

Instrument :

FID_F

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 07/15/2025

Supervised By :mohammad ahmed 07/16/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF07142
Data File : FF016153.D
Signal(s) : FID2B.ch
Acq On : 14 Jul 2025 15:19
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF071025.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.072	2.032	2.169	PB	356895	6385774	97.55%	6.192%
2	4.616	4.579	4.672	BB	608849	6546460	100.00%	6.348%
3	6.795	6.759	6.844	BB	647257	6452663	98.57%	6.257%
4	8.629	8.589	8.687	BB	625154	6307081	96.34%	6.116%
5	10.239	10.190	10.294	BB	573302	6213377	94.91%	6.025%
6	11.686	11.599	11.732	BB	540141	6240388	95.32%	6.051%
7	13.000	12.940	13.039	BB	536296	6193496	94.61%	6.006%
8	14.202	14.167	14.244	VB	481563	5985522	91.43%	5.804%
9	15.104	14.952	15.136	PV	384342	5302187	80.99%	5.141%
10	15.310	15.215	15.365	BB	465156	5953814	90.95%	5.773%
11	16.333	15.869	16.372	BV	441282	5847943	89.33%	5.671%
12	17.287	17.200	17.339	BV	446366	5814980	88.83%	5.639%
13	18.178	18.121	18.217	VV	422129	5948351	90.86%	5.768%
14	19.015	18.950	19.049	BV	396316	5721233	87.39%	5.548%
15	19.801	19.750	19.854	BV	366924	5280271	80.66%	5.120%
16	20.542	20.494	20.580	VV	318379	5033719	76.89%	4.881%
17	21.310	21.264	21.349	VV	215157	4084600	62.39%	3.961%
18	22.276	22.160	22.324	BV	153274	3814034	58.26%	3.698%
Sum of corrected areas:						103125893		

FF071025.M Tue Jul 15 01:14:11 2025