

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: ENTA05
ProjectID: 540 Degraw St, Brooklyn, NY - E9309
Lab Code: ACE SDG No.: Q2580
DataFile: FF016160.D Analyst Name: YP\AJ Analyst Date: 07-14-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	87782848	103274	106641	3.157