

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2580

Client ID : SSTDCCC0.4

Date Analyzed: 07/14/2025

Lab File ID: BN037483.D

Time Analyzed: 09:10

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1935	7.732	4597	10.51	2166	14.36
UPPER LIMIT	3870	8.232	9194	11.009	4332	14.855
LOWER LIMIT	967.5	7.232	2298.5	10.009	1083	13.855
EPA SAMPLE NO.						
01 PB168831BL	2220	7.73	5460	10.51	2504	14.36
02 PB168831BS	2484	7.73	6222	10.51	2912	14.36
03 WC-URBAN-FILL-B2	2050	7.73	5267	10.51	2641	14.36
04 WC-URBAN-FILL-B2DL	2104	7.73	5369	10.51	2602	14.36
05 WC-URBAN-FILL-B2MS	2103	7.73	5429	10.51	2752	14.36
06 WC-URBAN-FILL-B2MSD	2101	7.73	5450	10.52	2896	14.36

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.