



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: ENTA05
Lab Code: ACE SDG No.: Q2580
Instrument ID: MSVOA_W Calibration Date/Time: 07/11/2025 11:50
Lab File ID: VW031817.D Init. Calib. Date(s): 06/30/2025 06/30/2025
Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:54 12:55
GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	62.53	-25.062	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.