

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENTA05
Lab Code: ACE SDG NO.: Q2580
Lab File ID: VW031817.D Date Analyzed: 07/11/2025
Instrument ID: MSVOA_W Time Analyzed: 11:50
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	198972	7.96	375996	8.85	344776	11.64
UPPER LIMIT	397944	8.459	751992	9.349	689552	12.135
LOWER LIMIT	99486	7.459	187998	8.349	172388	11.135
EPA SAMPLE NO.						
WC-URBAN-FILL-B1	194384	7.97	385504	8.86	362157	11.64
VW0711SBL01	167609	7.96	357632	8.86	345815	11.63
VW0711SBS01	205659	7.97	389380	8.85	347490	11.63
VW0711SBSD01	203412	7.97	392643	8.86	357866	11.63

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.