

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENTA05
 Lab Code: ACE SDG NO.: Q2580
 Lab File ID: VW031817.D Date Analyzed: 07/11/2025
 Instrument ID: MSVOA_W Time Analyzed: 11:50
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	160985	13.556				
UPPER LIMIT	321970	14.056				
LOWER LIMIT	80492.5	13.056				
EPA SAMPLE NO.						
WC-URBAN-FILL-B1	168181	13.56				
VW0711SBL01	162450	13.56				
VW0711SBS01	173515	13.56				
VW0711SBSD01	171573	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.