

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2581

Client ID : SSTD020233

Date Analyzed: 07/18/2025

Lab File ID: BN037522.D

Time Analyzed: 13:57

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	83966	7.722	344876	10.50	217939	14.36
UPPER LIMIT	167932	8.222	689752	11.004	435878	14.857
LOWER LIMIT	41983	7.222	172438	10.004	108970	13.857
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	79770	7.72	316123	10.50	194877	14.36

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.