

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX070225

Review By	John Carlone	Review On	7/3/2025 8:44:30 AM
Supervise By	Maresh Dadoda	Supervise On	7/3/2025 3:09:59 PM
SubDirectory	VX070225	HP Acquire Method	HP Processing Method 82X070225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134642 VP134633,VP134634,VP134635,VP134638,VP134639,VP134640 VP134641

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046859.D	02 Jul 2025 11:12		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX046860.D	02 Jul 2025 12:11		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX046861.D	02 Jul 2025 12:37	%D Failed for com.#13 in 5 ppb	JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX046862.D	02 Jul 2025 13:18		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX046863.D	02 Jul 2025 13:39	LR-13	JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX046864.D	02 Jul 2025 14:10		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX046865.D	02 Jul 2025 14:31		JC/MD	Ok,M
8	VIBLK	VIBLK	VX046866.D	02 Jul 2025 14:53		JC/MD	Ok
9	VSTDICV050	ICVVX070225	VX046867.D	02 Jul 2025 15:14		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071825

Review By	John Carlone	Review On	7/21/2025 8:40:45 AM
Supervise By	Maresh Dadoda	Supervise On	7/21/2025 1:56:12 PM
SubDirectory	VX071825	HP Acquire Method	MSVOA_X HP Processing Method 82x070225w.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134820
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134822,VP134823 VP133935

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047037.D	18 Jul 2025 07:32		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047038.D	18 Jul 2025 10:20	CCC fail High com. # 25	JC/MD	Ok,M
3	VX0718MBL01	VX0718MBL01	VX047039.D	18 Jul 2025 10:58		JC/MD	Ok
4	VX0718WBL01	VX0718WBL01	VX047040.D	18 Jul 2025 11:19		JC/MD	Ok
5	VX0718WBS01	VX0718WBS01	VX047041.D	18 Jul 2025 11:43	BS Failed High for com.#25	JC/MD	Ok,M
6	VX0718WBSD01	VX0718WBSD01	VX047042.D	18 Jul 2025 12:05	BSD Failed High for com.#25	JC/MD	Ok,M
7	Q2581-01	STORAGE-BLANK-SO	VX047043.D	18 Jul 2025 12:53	vial A pH<2	JC/MD	Ok
8	Q2581-02	STORAGE-BLANK-WA	VX047044.D	18 Jul 2025 13:14	vial A pH<2	JC/MD	Ok
9	Q2581-03	STORAGE-BLANK-WA	VX047045.D	18 Jul 2025 13:36	vial A pH<2	JC/MD	Ok
10	Q2581-04	STORAGE-BLANK-SAI	VX047046.D	18 Jul 2025 13:57	vial A pH<2	JC/MD	Ok
11	Q2637-01	STORAGE-BLANK-SO	VX047047.D	18 Jul 2025 14:18	vial A pH<2	JC/MD	Ok
12	Q2637-02	STORAGE-BLANK-WA	VX047048.D	18 Jul 2025 14:39	vial A pH<2	JC/MD	Ok
13	Q2637-03	STORAGE-BLANK-WA	VX047049.D	18 Jul 2025 15:00	vial A pH<2	JC/MD	Ok
14	Q2637-04	STORAGE-BLANK-SAI	VX047050.D	18 Jul 2025 15:21	vial A pH<2	JC/MD	Ok
15	Q2646-01	FRAC TANK	VX047051.D	18 Jul 2025 15:42	CCC Failed High,BS,BSD Failed High,Run @ lower dilution	JC/MD	Not Ok
16	VSTDCCC050	VSTDCCC050EC	VX047052.D	18 Jul 2025 16:03		JC/MD	Ok,M

M : Manual Integration