

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Ardmore Chemical

SDG No.: Q2585

Contract: ARDM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168881BS							
Lead	ug/L	500	481		96	85 - 115	P
Zinc	ug/L	100	107		107	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Ardmore Chemical

SDG No.: Q2585

Contract: ARDM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168942BS Mercury	ug/L	4.0	4.15		104	85 - 115	CV