

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071725\
 Data File : BF143148.D
 Acq On : 17 Jul 2025 15:06
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF071725

Quant Time: Jul 17 16:24:49 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 17 15:14:05 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.598	0.576	3.7	83	-0.03
3	Pyridine	1.552	1.576	-1.5	89	-0.02
4	n-Nitrosodimethylamine	0.801	0.817	-2.0	87	-0.02
5 S	2-Fluorophenol	1.264	1.219	3.6	85	0.00
6	Aniline	2.217	2.277	-2.7	89	0.00
7 S	Phenol-d6	1.591	1.554	2.3	86	0.00
8	2-Chlorophenol	1.325	1.375	-3.8	90	0.00
9	Benzaldehyde	1.193	1.117	6.4	67	0.00
10 C	Phenol	1.733	1.769	-2.1	89	0.00
11	bis(2-Chloroethyl)ether	1.327	1.348	-1.6	89	0.00
12	1,3-Dichlorobenzene	1.439	1.454	-1.0	89	0.00
13 C	1,4-Dichlorobenzene	1.449	1.476	-1.9	90	0.00
14	1,2-Dichlorobenzene	1.378	1.430	-3.8	92	0.00
15	Benzyl Alcohol	1.215	1.274	-4.9	90	0.00
16	2,2'-oxybis(1-Chloropropane	2.461	2.545	-3.4	91	0.00
17	2-Methylphenol	1.114	1.153	-3.5	91	0.00
18	Hexachloroethane	0.502	0.526	-4.8	92	0.00
19 P	n-Nitroso-di-n-propylamine	1.021	1.029	-0.8	90	0.00
20	3+4-Methylphenols	1.351	1.432	-6.0	92	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
22	Acetophenone	0.474	0.477	-0.6	91	0.00
23 S	Nitrobenzene-d5	0.401	0.362	9.7	79	0.00
24	Nitrobenzene	0.374	0.390	-4.3	91	0.00
25	Isophorone	0.708	0.718	-1.4	91	0.00
26 C	2-Nitrophenol	0.162	0.182	-12.3	93	0.00
27	2,4-Dimethylphenol	0.331	0.337	-1.8	91	0.00
28	bis(2-Chloroethoxy)methane	0.425	0.424	0.2	90	0.00
29 C	2,4-Dichlorophenol	0.279	0.285	-2.2	90	0.00
30	1,2,4-Trichlorobenzene	0.302	0.306	-1.3	91	0.00
31	Naphthalene	0.985	0.989	-0.4	90	0.00
32	Benzoic acid	0.169	0.181	-7.1	88	0.00
33	4-Chloroaniline	0.402	0.409	-1.7	91	0.00
34 C	Hexachlorobutadiene	0.184	0.190	-3.3	92	0.00
35	Caprolactam	0.084	0.086	-2.4	87	0.00
36 C	4-Chloro-3-methylphenol	0.303	0.308	-1.7	89	0.00
37	2-Methylnaphthalene	0.599	0.604	-0.8	91	0.00
38	1-Methylnaphthalene	0.617	0.628	-1.8	91	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
40	1,2,4,5-Tetrachlorobenzene	0.577	0.591	-2.4	93	0.00
41 P	Hexachlorocyclopentadiene	0.376	0.403	-7.2	93	0.00
42 S	2,4,6-Tribromophenol	0.207	0.207	0.0	85	0.00
43 C	2,4,6-Trichlorophenol	0.400	0.402	-0.5	91	0.00
44	2,4,5-Trichlorophenol	0.400	0.430	-7.5	92	0.00
45 S	2-Fluorobiphenyl	1.462	1.261	13.7	80	0.00
46	1,1'-Biphenyl	1.560	1.601	-2.6	93	0.00
47	2-Chloronaphthalene	1.155	1.160	-0.4	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071725\
 Data File : BF143148.D
 Acq On : 17 Jul 2025 15:06
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF071725

Quant Time: Jul 17 16:24:49 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 17 15:14:05 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.362	0.396	-9.4	91	0.00
49	Acenaphthylene	1.935	1.936	-0.1	90	0.00
50	Dimethylphthalate	1.304	1.311	-0.5	89	0.00
51	2,6-Dinitrotoluene	0.265	0.291	-9.8	91	0.00
52 C	Acenaphthene	1.144	1.149	-0.4	91	0.00
53	3-Nitroaniline	0.310	0.333	-7.4	91	0.00
54 P	2,4-Dinitrophenol	0.098	0.113	-15.3	93	0.00
55	Dibenzofuran	1.698	1.697	0.1	90	0.00
56 P	4-Nitrophenol	0.232	0.245	-5.6	86	0.00
57	2,4-Dinitrotoluene	0.333	0.370	-11.1	90	0.00
58	Fluorene	1.274	1.249	2.0	89	0.00
59	2,3,4,6-Tetrachlorophenol	0.331	0.346	-4.5	90	0.00
60	Diethylphthalate	1.289	1.312	-1.8	90	0.00
61	4-Chlorophenyl-phenylether	0.635	0.641	-0.9	91	0.00
62	4-Nitroaniline	0.266	0.279	-4.9	86	0.00
63	Azobenzene	1.343	1.357	-1.0	89	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.00
65	4,6-Dinitro-2-methylphenol	0.094	0.108	-14.9	93	0.00
66 c	n-Nitrosodiphenylamine	0.704	0.709	-0.7	89	0.00
67	4-Bromophenyl-phenylether	0.238	0.243	-2.1	90	0.00
68	Hexachlorobenzene	0.249	0.251	-0.8	88	0.00
69	Atrazine	0.194	0.207	-6.7	89	0.00
70 C	Pentachlorophenol	0.147	0.155	-5.4	88	0.00
71	Phenanthrene	1.062	1.063	-0.1	88	0.00
72	Anthracene	1.083	1.098	-1.4	89	0.00
73	Carbazole	0.946	0.957	-1.2	87	0.00
74	Di-n-butylphthalate	1.070	1.132	-5.8	89	0.00
75 C	Fluoranthene	0.975	0.995	-2.1	88	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
77	Benzydine	0.771	0.833	-8.0	78	0.00
78	Pyrene	1.707	1.703	0.2	89	0.00
79 S	Terphenyl-d14	1.344	1.165	13.3	79	0.00
80	Butylbenzylphthalate	0.523	0.573	-9.6	94	0.00
81	Benzo(a)anthracene	1.339	1.337	0.1	93	0.00
82	3,3'-Dichlorobenzidine	0.446	0.488	-9.4	102	0.00
83	Chrysene	1.204	1.271	-5.6	94	0.00
84	Bis(2-ethylhexyl)phthalate	0.791	0.862	-9.0	98	0.00
85 c	Di-n-octyl phthalate	1.434	1.544	-7.7	100	0.00
86 I	Perylene-d12	1.000	1.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	1.410	1.445	-2.5	98	0.00
88	Benzo(b)fluoranthene	1.222	1.322	-8.2	110	0.00
89	Benzo(k)fluoranthene	1.090	1.023	6.1	87	0.00
90 C	Benzo(a)pyrene	1.126	1.142	-1.4	98	0.00
91	Dibenzo(a,h)anthracene	1.148	1.168	-1.7	98	0.00
92	Benzo(g,h,i)perylene	1.110	1.149	-3.5	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071725\
Data File : BF143148.D
Acq On : 17 Jul 2025 15:06
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF071725

Quant Time: Jul 17 16:24:49 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 17 15:14:05 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0