

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2585

Client ID : SSTDCCC040

Date Analyzed: 07/18/2025

Lab File ID: BF143159.D

Time Analyzed: 11:24

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	158986	6.963	609943	8.25	308568	10.00
UPPER LIMIT	317972	7.463	1219890	8.745	617136	10.498
LOWER LIMIT	79493	6.463	304972	7.745	154284	9.498
EPA SAMPLE NO.						
01 PB168904BS	140840	6.96	541427	8.25	282216	10.00
02 PB168904BSD	134233	6.96	524933	8.25	272308	10.00
03 PB168904BL	144227	6.96	555488	8.24	293300	10.00
04 EFF-WW	91787	6.96	319077	8.24	149160 *	10.00

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.