

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ARDM01
 Lab Code: ACE SDG NO.: Q2585
 Lab File ID: VN087353.D Date Analyzed: 07/18/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:15
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	46358	7.80	246415	9.08	220107	11.85
UPPER LIMIT	92716	8.3	492830	9.583	440214	12.347
LOWER LIMIT	23179	7.3	123208	8.583	110054	11.347
EPA SAMPLE NO.						
EFF-WW	44521	7.80	204474	9.08	181978	11.85
VN0718WBL01	41937	7.79	219664	9.08	193250	11.85
VN0718WBS01	47522	7.80	244435	9.08	221298	11.85
VN0718WBSD01	42433	7.80	214950	9.08	202696	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.