

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136454

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/11/2025	16:55
2	CAL2	1	Water	NA	NA	20.2	7.01	07/11/2025	16:56
3	CAL3	1	Water	NA	NA	20.3	10.02	07/11/2025	16:59
4	ICV	1	Water	NA	NA	20.2	7.01	07/11/2025	17:00
5	CCV1	1	Water	NA	NA	20.2	2.01	07/11/2025	17:02
6	Q2588-01	1	Water	NA	NA	20.9	4.74	07/11/2025	17:15
7	Q2591-04	1	Water	NA	NA	21.1	4.98	07/11/2025	17:22
8	Q2591-04DUP	1	Water	NA	NA	21.4	4.99	07/11/2025	17:25
9	CCV2	1	Water	NA	NA	20.3	12.02	07/11/2025	17:30

WORKLIST(Hardcopy Internal Chain)

136454

WorkList Name : PH W Q2591 WorkList ID : 190684 Department : Wet-Chemistry Date : 07-11-2025 16:31:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2588-01	MH-7112025	Water	pH	Cool 4 deg C	EURO03	D41	07/11/2025	9040C
Q2591-04	NWB-2187	Water	pH	Cool 4 deg C	PSEG03	D41	07/11/2025	9040C

Date/Time 07-11-25 16:45
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 07-11-25 18:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]