

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB136475

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/15/2025	08:40
2	CAL2	1	Water	NA	NA	20.2	7.00	07/15/2025	08:41
3	CAL3	1	Water	NA	NA	20.2	10.02	07/15/2025	08:42
4	ICV	1	Water	NA	NA	20.2	7.00	07/15/2025	08:41
5	CCV1	1	Water	NA	NA	20.2	2.01	07/15/2025	08:45
6	Q2592-02	1	Solid	20.02	20	22.8	6.94	07/15/2025	09:00
7	Q2592-02DUP	1	Solid	20.03	20	22.9	6.96	07/15/2025	09:02
8	CCV2	1	Water	NA	NA	20.3	12.02	07/15/2025	09:33

WORKLIST(Hardcopy Internal Chain)

W 136475

WorkList Name : corrsivity q2600 WorkList ID : 190719 Department : Wet-Chemistry Date : 07-15-2025 08:30:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2592-02	WC-SOIL-20250711	Solid	Corrosivity	Cool 4 deg C	PARS02	D51	07/11/2025	9045D

Date/Time 07/15/25 08:33
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 07/15/25 13:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]