

### Initial and Continuing Calibration Blank Summary

**Client:** PARSONS Engineering of New York, Inc.

**SDG No.:** Q2592

**Project:** Con Edison - East River Site 2

**RunNo.:** LB136502

| Analyte                                    | Units | Result   | Acceptance Limits | Conc Qual | MDL     | RDL   | Analysis Date |
|--------------------------------------------|-------|----------|-------------------|-----------|---------|-------|---------------|
| Sample ID: <b>ICB1</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 07/16/2025    |
| Sample ID: <b>CCB1</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 07/16/2025    |
| Sample ID: <b>CCB2</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 07/16/2025    |
| Sample ID: <b>CCB3</b><br>Reactive Cyanide | mg/L  | < 0.0025 | 0.0025            | U         | 0.00096 | 0.005 | 07/16/2025    |