

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2592

Contract: PARS02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	243000	250000	97	216000	294000	07/17/2025	14:46	LB136534
	Antimony	20.2			-50	50	07/17/2025	14:46	LB136534
	Arsenic	-5.59			-20	20	07/17/2025	14:46	LB136534
	Barium	5.91	6.0	98	-94	106	07/17/2025	14:46	LB136534
	Beryllium	1.12			-6	6	07/17/2025	14:46	LB136534
	Cadmium	-0.95	1.0	95	-5	7	07/17/2025	14:46	LB136534
	Calcium	232000	240000	97	208000	282000	07/17/2025	14:46	LB136534
	Chromium	48.4	52.0	93	42	62	07/17/2025	14:46	LB136534
	Cobalt	21.5			-30	30	07/17/2025	14:46	LB136534
	Copper	-6.18	2.0	309	-18	22	07/17/2025	14:46	LB136534
	Iron	89400	100000	89	85600	116500	07/17/2025	14:46	LB136534
	Lead	1.20			-12	12	07/17/2025	14:46	LB136534
	Magnesium	242000	260000	93	216000	294000	07/17/2025	14:46	LB136534
	Manganese	3.52	7.0	50	-13	27	07/17/2025	14:46	LB136534
	Nickel	1.33	2.0	66	-38	42	07/17/2025	14:46	LB136534
	Potassium	3.55			0	0	07/17/2025	14:46	LB136534
	Selenium	-9.33			-20	20	07/17/2025	14:46	LB136534
	Silver	-1.63			-10	10	07/17/2025	14:46	LB136534
	Sodium	127			0	0	07/17/2025	14:46	LB136534
	Thallium	2.74			-40	40	07/17/2025	14:46	LB136534
	Vanadium	9.54			-40	40	07/17/2025	14:46	LB136534
	Zinc	11.6			-40	40	07/17/2025	14:46	LB136534
ICSAB01	Aluminum	243000	250000	97	209000	285000	07/17/2025	14:51	LB136534
	Antimony	586	620	94	525	711	07/17/2025	14:51	LB136534
	Arsenic	89.5	100	90	88.4	120	07/17/2025	14:51	LB136534
	Barium	490	540	91	437	637	07/17/2025	14:51	LB136534
	Beryllium	480	500	96	420	570	07/17/2025	14:51	LB136534
	Cadmium	954	970	98	826	1120	07/17/2025	14:51	LB136534
	Calcium	233000	230000	101	199000	271000	07/17/2025	14:51	LB136534
	Chromium	528	540	98	460	624	07/17/2025	14:51	LB136534
	Cobalt	493	480	103	404	548	07/17/2025	14:51	LB136534
	Copper	455	510	89	434	588	07/17/2025	14:51	LB136534
	Iron	90300	99000	91	84400	114500	07/17/2025	14:51	LB136534
	Lead	46.7	49.0	95	37	61	07/17/2025	14:51	LB136534
	Magnesium	242000	250000	97	210000	286000	07/17/2025	14:51	LB136534
	Manganese	475	510	93	430	584	07/17/2025	14:51	LB136534
	Nickel	941	950	99	810	1100	07/17/2025	14:51	LB136534
	Potassium	3.18			0	0	07/17/2025	14:51	LB136534
	Selenium	42.8	46.0	93	26	66	07/17/2025	14:51	LB136534
	Silver	204	200	102	170	232	07/17/2025	14:51	LB136534
	Sodium	145			0	0	07/17/2025	14:51	LB136534
	Thallium	91.9	110	84	68	148	07/17/2025	14:51	LB136534

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2592

Contract: PARS02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	470	490	96	417	565	07/17/2025	14:51	LB136534
	Zinc	999	950	105	809	1095	07/17/2025	14:51	LB136534