



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2592

Contract: PARS02

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.23	0.2	116	70 - 130	CV	07/17/2025	10:26	LB136516
CRI01	Arsenic	18.9	20.0	95	65 - 135	P	07/17/2025	13:46	LB136533
	Barium	103	100	103	65 - 135	P	07/17/2025	13:46	LB136533
	Cadmium	5.94	6.0	99	65 - 135	P	07/17/2025	13:46	LB136533
	Chromium	9.06	10.0	91	65 - 135	P	07/17/2025	13:46	LB136533
	Lead	11.1	12.0	92	65 - 135	P	07/17/2025	13:46	LB136533
	Selenium	18.7	20.0	94	65 - 135	P	07/17/2025	13:46	LB136533
	Silver	10.8	10.0	108	65 - 135	P	07/17/2025	13:46	LB136533