



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2592

Contract: PARS02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB31	Mercury	0.076	+/-0.2	U	0.20	CV	07/17/2025	10:12	LB136516

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2592

Contract: PARS02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB18	Mercury	0.076	+/-0.2	U	0.20	CV	07/17/2025	10:21	LB136516
CCB19	Mercury	0.076	+/-0.2	U	0.20	CV	07/17/2025	11:05	LB136516
CCB20	Mercury	0.076	+/-0.2	U	0.20	CV	07/17/2025	11:35	LB136516
CCB21	Mercury	0.076	+/-0.2	U	0.20	CV	07/17/2025	11:55	LB136516

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2592

Contract: PARS02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	13:41	LB136533
	Barium	14.6	+/-50	U	100	P	07/17/2025	13:41	LB136533
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	13:41	LB136533
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	13:41	LB136533
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	13:41	LB136533
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	13:41	LB136533
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	13:41	LB136533

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2592

Contract: PARS02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	14:25	LB136533
	Barium	14.6	+/-50	U	100	P	07/17/2025	14:25	LB136533
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	14:25	LB136533
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	14:25	LB136533
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	14:25	LB136533
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	14:25	LB136533
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	14:25	LB136533
CCB02	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	15:02	LB136533
	Barium	14.6	+/-50	U	100	P	07/17/2025	15:02	LB136533
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	15:02	LB136533
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	15:02	LB136533
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	15:02	LB136533
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	15:02	LB136533
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	15:02	LB136533
CCB03	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	15:56	LB136533
	Barium	14.6	+/-50	U	100	P	07/17/2025	15:56	LB136533
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	15:56	LB136533
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	15:56	LB136533
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	15:56	LB136533
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	15:56	LB136533
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	15:56	LB136533
CCB04	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	16:46	LB136533
	Barium	14.6	+/-50	U	100	P	07/17/2025	16:46	LB136533
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	16:46	LB136533
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	16:46	LB136533
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	16:46	LB136533
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	16:46	LB136533
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	16:46	LB136533
CCB05	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	18:02	LB136533
	Barium	14.6	+/-50	U	100	P	07/17/2025	18:02	LB136533
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	18:02	LB136533
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	18:02	LB136533
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	18:02	LB136533
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	18:02	LB136533
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	18:02	LB136533
CCB06	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	19:28	LB136533
	Barium	14.6	+/-50	U	100	P	07/17/2025	19:28	LB136533
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	19:28	LB136533
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	19:28	LB136533

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2592

Contract: PARS02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	07/17/2025	19:28	LB136533
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	19:28	LB136533
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	19:28	LB136533
CCB07	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	20:20	LB136533
	Barium	14.6	+/-50	U	100	P	07/17/2025	20:20	LB136533
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	20:20	LB136533
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	20:20	LB136533
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	20:20	LB136533
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	20:20	LB136533
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	20:20	LB136533