

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2592

Contract: PARS02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	9.77			-20	20	07/17/2025	13:51	LB136533
	Barium	1.77	6.0	30	-94	106	07/17/2025	13:51	LB136533
	Cadmium	-1.32	1.0	132	-5	7	07/17/2025	13:51	LB136533
	Chromium	51.4	52.0	99	42	62	07/17/2025	13:51	LB136533
	Lead	-4.86			-12	12	07/17/2025	13:51	LB136533
	Selenium	10.3			-20	20	07/17/2025	13:51	LB136533
	Silver	2.71			-10	10	07/17/2025	13:51	LB136533
ICSAB01	Arsenic	109	104	105	88.4	120	07/17/2025	13:56	LB136533
	Barium	520	537	97	437	637	07/17/2025	13:56	LB136533
	Cadmium	976	972	100	826	1120	07/17/2025	13:56	LB136533
	Chromium	512	542	94	460	624	07/17/2025	13:56	LB136533
	Lead	41.1	49.0	84	37	61	07/17/2025	13:56	LB136533
	Selenium	57.8	46.0	126	26	66	07/17/2025	13:56	LB136533
	Silver	218	201	108	170	232	07/17/2025	13:56	LB136533