

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PARS02
Lab Code: ACE **SDG NO.:** Q2592
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073804.D **Time Analyzed:** 16:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.637	5.539	5.739	455.310	500.000	-8.9
Aroclor-1016-2	5.659	5.560	5.760	494.680	500.000	-1.1
Aroclor-1016-3	5.720	5.622	5.822	488.830	500.000	-2.2
Aroclor-1016-4	5.818	5.720	5.920	493.640	500.000	-1.3
Aroclor-1016-5	6.110	6.012	6.212	507.350	500.000	1.5
Aroclor-1260-1	7.226	7.129	7.329	489.480	500.000	-2.1
Aroclor-1260-2	7.480	7.383	7.583	415.640	500.000	-16.9
Aroclor-1260-3	7.838	7.740	7.940	474.930	500.000	-5.0
Aroclor-1260-4	8.061	7.964	8.164	486.570	500.000	-2.7
Aroclor-1260-5	8.379	8.283	8.483	492.390	500.000	-1.5
Decachlorobiphenyl	10.171	10.076	10.276	52.010	50.000	4.0
Tetrachloro-m-xylene	4.487	4.388	4.588	51.230	50.000	2.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PARS02
Lab Code: ACE **SDG NO.:** Q2592
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073804.D **Time Analyzed:** 16:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.856	4.758	4.958	472.940	500.000	-5.4
Aroclor-1016-2	4.873	4.775	4.975	467.040	500.000	-6.6
Aroclor-1016-3	5.050	4.951	5.151	476.820	500.000	-4.6
Aroclor-1016-4	5.092	4.993	5.193	476.550	500.000	-4.7
Aroclor-1016-5	5.305	5.207	5.407	509.550	500.000	1.9
Aroclor-1260-1	6.334	6.237	6.437	472.180	500.000	-5.6
Aroclor-1260-2	6.523	6.425	6.625	473.960	500.000	-5.2
Aroclor-1260-3	6.674	6.577	6.777	460.750	500.000	-7.9
Aroclor-1260-4	7.143	7.046	7.246	486.760	500.000	-2.6
Aroclor-1260-5	7.386	7.289	7.489	485.010	500.000	-3.0
Decachlorobiphenyl	8.777	8.680	8.880	53.400	50.000	6.8
Tetrachloro-m-xylene	3.779	3.678	3.878	47.220	50.000	-5.6