

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PARS02
Lab Code: ACE **SDG NO.:** Q2592
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073819.D **Time Analyzed:** 21:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.634	5.539	5.739	406.560	500.000	-18.7
Aroclor-1016-2	5.655	5.560	5.760	433.980	500.000	-13.2
Aroclor-1016-3	5.718	5.622	5.822	433.080	500.000	-13.4
Aroclor-1016-4	5.815	5.720	5.920	439.190	500.000	-12.2
Aroclor-1016-5	6.107	6.012	6.212	444.730	500.000	-11.1
Aroclor-1260-1	7.223	7.129	7.329	426.830	500.000	-14.6
Aroclor-1260-2	7.477	7.383	7.583	361.680	500.000	-27.7
Aroclor-1260-3	7.835	7.740	7.940	411.850	500.000	-17.6
Aroclor-1260-4	8.058	7.964	8.164	422.130	500.000	-15.6
Aroclor-1260-5	8.376	8.283	8.483	430.280	500.000	-13.9
Decachlorobiphenyl	10.166	10.076	10.276	45.300	50.000	-9.4
Tetrachloro-m-xylene	4.484	4.388	4.588	45.830	50.000	-8.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PARS02
Lab Code: ACE **SDG NO.:** Q2592
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073819.D **Time Analyzed:** 21:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.854	4.758	4.958	469.990	500.000	-6.0
Aroclor-1016-2	4.871	4.775	4.975	464.330	500.000	-7.1
Aroclor-1016-3	5.048	4.951	5.151	474.880	500.000	-5.0
Aroclor-1016-4	5.090	4.993	5.193	470.170	500.000	-6.0
Aroclor-1016-5	5.303	5.207	5.407	502.970	500.000	0.6
Aroclor-1260-1	6.332	6.237	6.437	439.520	500.000	-12.1
Aroclor-1260-2	6.521	6.425	6.625	445.700	500.000	-10.9
Aroclor-1260-3	6.672	6.577	6.777	430.730	500.000	-13.9
Aroclor-1260-4	7.142	7.046	7.246	447.390	500.000	-10.5
Aroclor-1260-5	7.384	7.289	7.489	445.990	500.000	-10.8
Decachlorobiphenyl	8.774	8.680	8.880	50.560	50.000	1.1
Tetrachloro-m-xylene	3.777	3.678	3.878	46.870	50.000	-6.3