

Surrogate Summary

SDG No.: **Q2592**

Client: **PARSONS Engineering of New York, Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP073553.D	PIBLK-PP073553.D	Tetrachloro-m-xyl	1	20	16.2	81		60	140
		Decachlorobiphen	1	20	17.3	87		60	140
		Tetrachloro-m-xyl	2	20	16.2	81		60	140
		Decachlorobiphen	2	20	17.1	85		60	140
I.BLK-PP073808.D	PIBLK-PP073808.D	Tetrachloro-m-xyl	1	20	19.6	98		60	140
		Decachlorobiphen	1	20	19.7	99		60	140
		Tetrachloro-m-xyl	2	20	18.3	91		60	140
		Decachlorobiphen	2	20	20.7	104		60	140
PB168852BL	PB168852BL	Tetrachloro-m-xyl	1	20	19.6	98		32	144
		Decachlorobiphen	1	20	19.9	100		32	175
		Tetrachloro-m-xyl	2	20	18.5	92		32	144
		Decachlorobiphen	2	20	21.3	106		32	175
PB168852BS	PB168852BS	Tetrachloro-m-xyl	1	20	20.4	102		32	144
		Decachlorobiphen	1	20	20.5	103		32	175
		Tetrachloro-m-xyl	2	20	18.1	91		32	144
		Decachlorobiphen	2	20	21.9	110		32	175
Q2592-01	WC-SOIL-20250711	Tetrachloro-m-xyl	1	20	18.3	92		32	144
		Decachlorobiphen	1	20	14.7	74		32	175
		Tetrachloro-m-xyl	2	20	18.0	90		32	144
		Decachlorobiphen	2	20	16.6	83		32	175
I.BLK-PP073823.D	PIBLK-PP073823.D	Tetrachloro-m-xyl	1	20	18.1	91		60	140
		Decachlorobiphen	1	20	17.7	89		60	140
		Tetrachloro-m-xyl	2	20	18.1	90		60	140
		Decachlorobiphen	2	20	20.4	102		60	140
Q2598-01MS	OK-01-071425MS	Tetrachloro-m-xyl	1	20	16.5	83		32	144
		Decachlorobiphen	1	20	19.0	95		32	175
		Tetrachloro-m-xyl	2	20	18.9	94		32	144
		Decachlorobiphen	2	20	20.1	100		32	175
Q2598-01MSD	OK-01-071425MSD	Tetrachloro-m-xyl	1	20	17.9	90		32	144
		Decachlorobiphen	1	20	20.0	100		32	175
		Tetrachloro-m-xyl	2	20	21.2	106		32	144
		Decachlorobiphen	2	20	21.9	110		32	175
I.BLK-PP073838.D	PIBLK-PP073838.D	Tetrachloro-m-xyl	1	20	17.8	89		60	140
		Decachlorobiphen	1	20	18.1	90		60	140
		Tetrachloro-m-xyl	2	20	18.3	91		60	140
		Decachlorobiphen	2	20	21.6	108		60	140