

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2592</u>	Analytical Method:	<u>8082A</u>
Client:	<u>PARSONS Engineering of New York, Inc</u>	DataFile :	<u>PP073825.D</u>

		Sample				Rec		RPD		Limits	
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2598-01MS	Client Sample ID:		OK-01-071425MS							
(Column 1)											
AR1016	171.8	0	154	ug/kg	90				55	146	
AR1260	171.8	0	143	ug/kg	83				54	119	
Lab Sample ID:	Q2598-01MS	Client Sample ID:		OK-01-071425MS							
(Column 2)											
AR1016	171.8	0	147	ug/kg	86				55	146	
AR1260	171.8	0	136	ug/kg	79				54	119	