

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: PARS02

Lab Code: ACE

SDG NO.: Q2592

Instrument ID: ECD_S

Calibration Date(s): 07/21/2025 07/21/2025

Calibration Times: 15:02 16:39

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS031157.D</u>	CF 500 =	<u>PS031158.D</u>		
CF 750 =		<u>PS031159.D</u>	CF 1000 =	<u>PS031160.D</u>	CF 1500 =	<u>PS031161.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	24592600000	22780000000	21638700000	20857200000	19899400000	21953600000	8
2,4-D	4193860000	3820730000	3654460000	3552490000	3453030000	3734920000	8
2,4-DCAA	5091100000	4403340000	4248720000	4081100000	3917010000	4348250000	10

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: PARS02

Lab Code: ACE

SDG NO.: Q2592

Instrument ID: ECD_S

Calibration Date(s): 07/21/2025 07/21/2025

Calibration Times: 15:02 16:39

GC Column: RTX-CLP2

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS031157.D</u>	CF 500 =	<u>PS031158.D</u>		
CF 750 =		<u>PS031159.D</u>	CF 1000 =	<u>PS031160.D</u>	CF 1500 =	<u>PS031161.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	16137400000	15454900000	14840900000	14348100000	13689700000	14894200000	6
2,4-D	1930260000	1742600000	1656200000	1604210000	1558320000	1698320000	9
2,4-DCAA	1147310000	1039810000	988394000	963101000	936229000	1014970000	8