

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance

**Contract:** PARS02

**Lab Code:** ACE

**SDG NO.:** Q2592

**Continuing Calib Date:** 07/18/2025

**Initial Calibration Date(s):** 07/11/2025 07/11/2025

**Continuing Calib Time:** 11:55

**Initial Calibration Time(s):** 16:00 17:36

**GC Column:** RTX-CLP **ID:** 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.33       | 7.33      | 7.23      | 7.43 | 0.00       |
| 2,4-D            | 8.46       | 8.47      | 8.37      | 8.57 | 0.01       |
| 2,4,5-TP(Silvex) | 9.35       | 9.35      | 9.25      | 9.45 | 0.00       |

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance

**Contract:** PARS02

**Lab Code:** ACE

**SDG NO.:** Q2592

**Continuing Calib Date:** 07/18/2025

**Initial Calibration Date(s):** 07/11/2025 07/11/2025

**Continuing Calib Time:** 11:55

**Initial Calibration Time(s):** 16:00 17:36

**GC Column:** RTX-CLP2 **ID:** 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| 2,4-DCAA         | 7.76       | 7.77      | 7.67      | 7.87  | 0.01       |
| 2,4-D            | 9.02       | 9.03      | 8.93      | 9.13  | 0.01       |
| 2,4,5-TP(Silvex) | 9.93       | 9.93      | 9.83      | 10.03 | 0.00       |