

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PARS02
Lab Code: ACE **SDG NO.:** Q2592
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031123.D **Time Analyzed:** 02:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.345	9.253	9.453	734.610	712.500	3.1
2,4-D	8.460	8.367	8.567	770.900	705.000	9.3
2,4-DCAA	7.327	7.233	7.433	713.140	750.000	-4.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PARS02
Lab Code: ACE **SDG NO.:** Q2592
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031123.D **Time Analyzed:** 02:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.931	9.832	10.032	608.020	712.500	-14.7
2,4-D	9.026	8.927	9.127	591.050	705.000	-16.2
2,4-DCAA	7.768	7.669	7.869	616.160	750.000	-17.8