

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PARS02
Lab Code: ACE **SDG NO.:** Q2592
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031132.D **Time Analyzed:** 15:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.345	9.253	9.453	770.640	712.500	8.2
2,4-D	8.459	8.367	8.567	824.550	705.000	17.0
2,4-DCAA	7.326	7.233	7.433	705.130	750.000	-6.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PARS02
Lab Code: ACE **SDG NO.:** Q2592
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031132.D **Time Analyzed:** 15:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.928	9.832	10.032	642.510	712.500	-9.8
2,4-D	9.023	8.927	9.127	665.480	705.000	-5.6
2,4-DCAA	7.764	7.669	7.869	642.990	750.000	-14.3