

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PARS02
Lab Code: ACE **SDG NO.:** Q2592
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/21/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031170.D **Time Analyzed:** 20:16

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.342	9.242	9.442	717.590	712.500	0.7
2,4-D	8.456	8.356	8.556	715.850	705.000	1.5
2,4-DCAA	7.324	7.225	7.425	745.520	750.000	-0.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PARS02
Lab Code: ACE **SDG NO.:** Q2592
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/21/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031170.D **Time Analyzed:** 20:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.929	9.829	10.029	712.950	712.500	0.1
2,4-D	9.024	8.924	9.124	695.090	705.000	-1.4
2,4-DCAA	7.766	7.666	7.866	737.580	750.000	-1.7