

Surrogate Summary

SDG No.: Q2592

Client: PARSONS Engineering of New York, Inc.

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PS031005.D	PIBLK-PS031005.D	2,4-DCAA	1	500	356	71		61	136
		2,4-DCAA	2	500	496	99		61	136
I.BLK-PS031109.D	PIBLK-PS031109.D	2,4-DCAA	1	500	468	94		61	136
		2,4-DCAA	2	500	494	99		61	136
PB168886BL	PB168886BL	2,4-DCAA	1	500	479	96		61	136
		2,4-DCAA	2	500	537	107		61	136
Q2592-02	WC-SOIL-20250711	2,4-DCAA	1	500	340	68		61	136
		2,4-DCAA	2	500	365	73		61	136
Q2592-02MS	WC-SOIL-20250711MS	2,4-DCAA	1	500	457	91		61	136
		2,4-DCAA	2	500	453	91		61	136
Q2592-02MSD	WC-SOIL-20250711MSD	2,4-DCAA	1	500	448	90		61	136
		2,4-DCAA	2	500	439	88		61	136
I.BLK-PS031122.D	PIBLK-PS031122.D	2,4-DCAA	1	500	469	94		61	136
		2,4-DCAA	2	500	495	99		61	136
I.BLK-PS031125.D	PIBLK-PS031125.D	2,4-DCAA	1	500	462	92		61	136
		2,4-DCAA	2	500	501	100		61	136
PB168847TB	PB168847TB	2,4-DCAA	1	500	510	102		61	136
		2,4-DCAA	2	500	548	110		61	136
I.BLK-PS031131.D	PIBLK-PS031131.D	2,4-DCAA	1	500	472	94		61	136
		2,4-DCAA	2	500	505	101		61	136
I.BLK-PS031156.D	PIBLK-PS031156.D	2,4-DCAA	1	500	397	79		61	136
		2,4-DCAA	2	500	504	101		61	136
I.BLK-PS031163.D	PIBLK-PS031163.D	2,4-DCAA	1	500	397	79		61	136
		2,4-DCAA	2	500	496	99		61	136
PB168886BS	PB168886BS	2,4-DCAA	1	500	511	102		61	136
		2,4-DCAA	2	500	506	101		61	136
I.BLK-PS031169.D	PIBLK-PS031169.D	2,4-DCAA	1	500	406	81		61	136
		2,4-DCAA	2	500	500	100		61	136