

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2592</u>	Analytical Method:	<u>8151A</u>
Client:	<u>PARSONS Engineering of New York, Inc</u>	DataFile :	<u>PS031116.D</u>

	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q2592-02MSD (Column 1)		Client Sample ID:	WC-SOIL-20250711MSD								
	2,4-D	50	0	70.1	ug/L	140	*	1		65	135	20
	2,4,5-TP(Silvex)	50	0	63.5	ug/L	127		1		62	139	20
Lab Sample ID:	Q2592-02MSD (Column 2)		Client Sample ID:	WC-SOIL-20250711MSD								
	2,4-D	50	0	55.8	ug/L	112		1		65	135	20
	2,4,5-TP(Silvex)	50	0	52.6	ug/L	105		3		62	139	20