

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PARS02
Lab Code: ACE **SDG NO.:** Q2592
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/18/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096477.D **Time Analyzed:** 17:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.018	8.920	9.120	47.270	50.000	-5.5
Endrin	6.544	6.445	6.645	53.550	50.000	7.1
gamma-BHC (Lindane)	4.312	4.213	4.413	53.590	50.000	7.2
Heptachlor	4.905	4.806	5.006	52.990	50.000	6.0
Heptachlor epoxide	5.662	5.565	5.765	54.650	50.000	9.3
Methoxychlor	7.463	7.364	7.564	48.960	50.000	-2.1
Tetrachloro-m-xylene	3.536	3.438	3.638	52.750	50.000	5.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PARS02
Lab Code: ACE **SDG NO.:** Q2592
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/18/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096477.D **Time Analyzed:** 17:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.993	7.896	8.096	49.570	50.000	-0.9
Endrin	5.710	5.612	5.812	54.060	50.000	8.1
gamma-BHC (Lindane)	3.666	3.567	3.767	55.800	50.000	11.6
Heptachlor	4.015	3.917	4.117	55.050	50.000	10.1
Heptachlor epoxide	4.800	4.702	4.902	54.060	50.000	8.1
Methoxychlor	6.682	6.584	6.784	49.540	50.000	-0.9
Tetrachloro-m-xylene	2.828	2.729	2.929	54.780	50.000	9.6