

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:		
Project:	Con Edison - East River Site 2		Date Received:	07/16/25	
Client Sample ID:	PB168847TB		SDG No.:	Q2592	
Lab Sample ID:	PB168847TB		Matrix:	TCLP	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL096470.D	1	07/16/25 11:02	07/18/25 16:01	PB168887

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.50	U	0.037	0.50	ug/L
76-44-8	Heptachlor	0.50	U	0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	0.50	U	0.096	0.50	ug/L
72-20-8	Endrin	0.50	U	0.032	0.50	ug/L
72-43-5	Methoxychlor	0.50	U	0.11	0.50	ug/L
8001-35-2	Toxaphene	10.0	U	1.70	10.0	ug/L
57-74-9	Chlordane	5.00	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.8		57 - 171	109%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.3		61 - 148	111%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit