

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2592</u>	Analytical Method:	<u>8081B</u>
Client:	<u>PARSONS Engineering of New York, Inc</u>	DataFile :	<u>PL096463.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result				Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2592-02MSD (Column 1)	Client Sample ID:		WC-SOIL-20250711MSD								
	gamma-BHC (Lindane)	5	0	5.10	ug/L	102		2		60	152	20
	Heptachlor	5	0	5.00	ug/L	100		2		56	147	20
	Heptachlor epoxide	5	0	5.30	ug/L	106		0		77	143	20
	Endrin	5	0	5.30	ug/L	106		2		76	144	20
	Methoxychlor	5	0	4.90	ug/L	98		4		70	142	20
Lab Sample ID:	Q2592-02MSD (Column 2)	Client Sample ID:		WC-SOIL-20250711MSD								
	gamma-BHC (Lindane)	5	0	5.40	ug/L	108		2		60	152	20
	Heptachlor	5	0	5.10	ug/L	102		2		56	147	20
	Heptachlor epoxide	5	0	5.20	ug/L	104		2		77	143	20
	Endrin	5	0	5.20	ug/L	104		4		76	144	20
	Methoxychlor	5	0	4.70	ug/L	94		2		70	142	20