

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2592

Analytical Method: 8081B

Client: PARSONS Engineering of New York, Inc

Datafile : PL096475.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168887BS (Column 1)	gamma-BHC (Lindane)	0.5	0.50	ug/L	100				82	129		
	Heptachlor	0.5	0.50	ug/L	100				79	127		
	Heptachlor epoxide	0.5	0.53	ug/L	105				81	124		
	Endrin	0.5	0.51	ug/L	102				81	128		
	Methoxychlor	0.5	0.45	ug/L	91				78	108		
PB168887BS (Column 2)	gamma-BHC (Lindane)	0.5	0.54	ug/L	108				82	129		
	Heptachlor	0.5	0.53	ug/L	105				79	127		
	Heptachlor epoxide	0.5	0.53	ug/L	106				81	124		
	Endrin	0.5	0.50	ug/L	100				81	128		
	Methoxychlor	0.5	0.46	ug/L	93				78	108		