

rteres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG07182
Data File : FG016329.D
Signal(s) : FID1A.ch
Acq On : 18 Jul 2025 09:24
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 07/21/2025
Supervised By :mohammad ahmed 07/23/2025

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070925.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.120	2.076	2.224	VB	342479	4993123	84.49%	5.309%
2	4.670	4.627	4.729	BB	517464	5345593	90.45%	5.684%
3	6.854	6.810	6.911	BB	545618	5466474	92.50%	5.813%
4	8.696	8.647	8.768	BB	532355	5570268	94.26%	5.923%
5	10.316	10.270	10.358	BV	528350	5648246	95.57%	6.006%
6	11.770	11.733	11.820	BB	506287	5812904	98.36%	6.181%
7	13.092	13.055	13.146	VV	493936	5909763	100.00%	6.284%
8	14.300	14.264	14.355	VB	467722	5700225	96.45%	6.061%
9	15.207	15.143	15.237	BV	359004	5008719	84.75%	5.326%
10	15.412	15.334	15.469	BB	459568	5721499	96.81%	6.084%
11	16.442	16.289	16.486	BB	427962	5569725	94.25%	5.923%
12	17.399	17.303	17.440	BV	399447	5452560	92.26%	5.798%
13	18.293	18.209	18.344	PV	394345	5530703	93.59%	5.881%
14	19.130	19.002	19.167	PV	376866	5383602	91.10%	5.725%
15	19.920	19.797	19.953	VV	339680	4886621	82.69%	5.196%
16	20.663	20.600	20.701	BV	279687	4277354	72.38%	4.548%
17	21.457	21.397	21.654	VV	191737	4233937	71.64%	4.502%
18	22.466	22.405	22.550	VV	129925	3531325	59.75%	3.755%
Sum of corrected areas:						94042640		

FG070925.M Fri Jul 18 11:43:06 2025