

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: PARS02
ProjectID: Con Edison - East River Site 2
Lab Code: ACE SDG No.: Q2592
DataFile: FG016342.D Analyst Name: YP\AJ Analyst Date: 07-18-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	86883822	102216	103531	1.27