284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# DATA PACKAGE

## SEMI-VOLATILE ORGANICS

**PROJECT NAME : CON EDISON - EAST RIVER SITE 2**

**PARSONS ENGINEERING OF NEW YORK, INC.**

**301 Plainfield Road**

**Suite 350**

**Syracuse, NY - 13212**

**Phone No: 315-451-9560**

**ORDER ID : Q2592**

**ATTENTION : Zohar Lavy**

**Laboratory Certification ID # 20012**

Table Of Contents for Q2592


1) SEMI-VOLATILE DATA 2
2) Signature Page 4
3) Case Narrative 5
4) Qualifier Page 7
5) Conformance/Non Conformance 8
6) QA Checklist 10
7) Chronicle 11
8) Hit Summary 12
9) QC Data Summary For SVOCMS Group1 13
- 9.1) Deuterated Monitoring Compound Summary 14
- 9.2) MS/MSD Summary 15
- 9.3) LCS/LCSD Summary 19
- 9.4) Method Blank Summary 21
- 9.5) GS/MS Tune Summary 22
- 9.6) Internal Standard Area and RT Summary 27

10) Sample Data 33
- 10.1) WC-SOIL-20250711 34

11) Calibration Data Summary 62
- 11.1) Initial Calibration Data 3
  - 11.1.1) BF071525 63
  - 11.1.2) BP070325 475
- 11.2) Continued Calibration Data 885
  - 11.2.1) BF143117.D 885
  - 11.2.2) BP025155.D 942
  - 11.2.3) BP025187.D 999

12) QC Sample Data 1056
- 12.1) Tune Raw Data 1057
- 12.2) Method Blank Data 1077
- 12.3) LCS Data 1094
- 12.4) MS Data 1146
- 12.5) MSD Data 1198

13) Manual Integration 1250
14) Analytical Runlogs 1257
15) Percent Solid 1269
16) Extraction Logs 1273

16.1) PB168854.pdf 1273
16.2) PB168854IC.pdf 1275
17) Standard Prep Logs 1276
18) Shipping Document 1439
18.1) Chain Of Custody 1440
18.2) Lab Certificate 1441
18.3) Internal COC 1442

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Cover Page

**Order ID :** Q2592

**Project ID :** Con Edison - East River Site 2

**Client :** PARSONS Engineering of New York, Inc.

| Lab Sample Number | Client Sample Number |
|---|---|
| Q2592-01 | WC-SOIL-20250711 |
| Q2592-02 | WC-SOIL-20250711 |

I certify that the data package is in compliance with the terms and conditions of the contract,both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 7/22/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

# CASE NARRATIVE

**PARSONS Engineering of New York, Inc.**
**Project Name: Con Edison - East River Site 2**
**Project # N/A**
**Order ID # Q2592**
**Test Name: SVOCMS Group1**

**A. Number of Samples and Date of Receipt:**
1 Solid sample was received on 07/11/2025.

**B. Parameters**
According to the Chain of Custody document, the following analyses were requested: SVOCMS Group1. This data package contains results for SVOCMS Group1.

**C. Analytical Techniques:**
The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um dfThe samples were analyzed on instrument BNA_P using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOCMS Group1 was based on method 8270E and extraction was done based on method 3541.

**D. QA/ QC Samples:**
The Holding Times were met for all analysis.
The Surrogate recoveries were met for all analysis.
The Internal Standards Areas met the acceptable requirements.
The Retention Times were acceptable for all samples.
The MS recoveries met the requirements for all compounds.
The MSD recoveries met the requirements for all compounds.
The RPD met criteria .

The Blank Spike for {PB168854BS} with File ID: BF143119.D met requirements for all samples except for 4-Nitroaniline[106%],is failing high but no positive hit in associate samples therefore no corrective action taken.

The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 20% in the Initial Calibration (8270-BF071525.M) for 2-Nitrophenol, 2-Nitroaniline, 2,6-Dinitrotoluene, 2,4-Dinitrotoluene, Butylbenzylphthalate, Bis(2-ethylhexyl)phthalate, and Di-n-octyl phthalate these compounds are passing on Linear Regression and 2,4-Dinitrophenol,4,6-Dinitro-2-methylphenol, these Compounds are passing on Qudratic regression.

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The % RSD is greater than 20% in the Initial Calibration (8270-BP070325.M) for 2-Nitrophenol,2-Nitroaniline,2,4-Dinitrophenol,2,4-Dinitrotoluene,4,6-Dinitro-2-methylphenol,Butylbenzylphthalate,Di-n-octyl phthalate these compounds are passing on Linear Regression.

The Continuous Calibration File ID BP025155.D met the requirements except for 2,4-Dinitrophenol,2-Nitrophenol,3,3-Dichlorobenzidine,4,6-Dinitro-2-methylphenol,Bis(2-ethylhexyl)phthalate,Hexachlorocyclopentadiene and 2,4,6-Tribromophenol are biased high therefore no corrective action was taken.

The Continuous Calibration File ID BP025187.D met the requirements except for 2,4-Dinitrophenol and 4,6-Dinitro-2-methylphenol, are biased high,and no positive hit in associated sampls, therefore no corrective action was taken.

The Tuning criteria met requirements.

Sample WC-SOIL-20250711 analyzed with direct 5X dilution due to dirty, viscous and concentrated matrix.

**E. Additional Comments:**
The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The soil samples results are based on a dry weight basis.
Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**
Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

284 Sheffield Street, Mountainside NJ 07092 Tel. 908-789-8900 Fax: 908-789-8922

# DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following " Results Qualifiers" are used:

| | |
|---|---|
| Value | If the result is a value greater than or equal to the detection limit, report the value |
| **U** | Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required. |
| **ND** | Indicates the analyte was analyzed for, but not detected |
| **J** | Indicates an estimated value**.** This flag is used:<br>(1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.)<br>(2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others. |
| **B** | Indicates the analyte was found in the blank as well as the sample report as "12 B". |
| **E** | Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis. |
| **D** | This flag identifies all compounds identified in an analysis at a secondary dilution factor. |
| **P** | This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P". |
| **N** | This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used. |
| **A** | This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product. |
| **Q** | Indicates the LCS did not meet the control limits requirements |

## GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2592 MATRIX: Solid

METHOD: 8270E/3541

| | | NA | NO | YES |
|---|---|---|---|---|
| 1. | Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks) | | | ✓ |
| 2. | GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ) | | | ✓ |
| 3. | GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series. | | | ✓ |
| 4. | GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series. | | | ✓ |
| 5. | GC/MS Calibration Requirements. | | ✓ | |

The % RSD is greater than 20% in the Initial Calibration (8270-BF071525.M) for 2-Nitrophenol, 2-Nitroaniline, 2,6-Dinitrotoluene, 2,4-Dinitrotoluene, Butylbenzylphthalate, Bis(2-ethylhexyl)phthalate, and Di-n-octyl phthalate these compounds are passing on Linear Regression and 2,4-Dinitrophenol,4,6-Dinitro-2-methylphenol, these Compounds are passing on Qudratic regression.

The % RSD is greater than 20% in the Initial Calibration (8270-BP070325.M) for 2-Nitrophenol,2-Nitroaniline,2,4-Dinitrophenol,2,4-Dinitrotoluene,4,6-Dinitro-2-methylphenol,Butylbenzylphthalate,Di-n-octyl phthalate these compounds are passing on Linear Regression.

The Continuous Calibration File ID BP025155.D met the requirements except for 2,4-Dinitrophenol,2-Nitrophenol,3,3-Dichlorobenzidine,4,6-Dinitro-2-methylphenol,Bis(2-ethylhexyl)phthalate,Hexachlorocyclopentadiene and 2,4,6-Tribromophenol are biased high therefore no corrective action was taken.

The Continuous Calibration File ID BP025187.D met the requirements except for 2,4-Dinitrophenol and 4,6-Dinitro-2-methylphenol, are biased high,and no positive hit in associated sampls, therefore no corrective action was taken.

| | | | | |
|---|---|---|---|---|
| 6. | Blank Contamination - If yes, list compounds and concentrations in each blank: | | ✓ | |
| 7. | Surrogate Recoveries Meet Criteria | | | ✓ |

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

## GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

| | | NA | NO | YES |
|---|---|---|---|---|
| 8. | Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria | | ✓ | |
| | If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| | The Blank Spike for {PB168854BS} with File ID: BF143119.D met requirements for all samples except for 4-Nitroaniline[106%],is failing high but no positive hit in associate samples therefore no corrective action taken. | | | |
| 9. | Internal Standard Area/Retention Time Shift Meet Criteria | | | ✓ |
| | Comments: | | | |
| 10. | Extraction Holding Time Met | | | ✓ |
| | If not met, list number of days exceeded for each sample: | | | |
| 11. | Analysis Holding Time Met | | | ✓ |
| | If not met, list number of days exceeded for each sample: | | | |

ADDITIONAL COMMENTS:

Sample WC-SOIL-20250711 analyzed with direct 5X dilution due to dirty, viscous and concentrated matrix.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

_____ _____

QA REVIEW Date

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

## APPENDIX A

## QA REVIEW GENERAL DOCUMENTATION

**Project #:** Q2592

| | Completed |
|---|---|
| **For thorough review, the report must have the following:** | |
| **GENERAL:** | |
| **Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)** | ✓ |
| **Check chain-of-custody for proper relinquish/return of samples** | ✓ |
| **Is the chain of custody signed and complete** | ✓ |
| **Check internal chain-of-custody for proper relinquish/return of samples /sample extracts** | ✓ |
| **Collect information for each project id from server. Were all requirements followed** | ✓ |
| **COVER PAGE:** | |
| **Do numbers of samples correspond to the number of samples in the Chain of Custody on login page** | ✓ |
| **Do lab numbers and client Ids on cover page agree with the Chain of Custody** | ✓ |
| **CHAIN OF CUSTODY:** | |
| **Do requested analyses on Chain of Custody agree with form I results** | ✓ |
| **Do requested analyses on Chain of Custody agree with the log-in page** | ✓ |
| **Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody** | ✓ |
| **Were the samples received within hold time** | ✓ |
| **Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle** | ✓ |
| **ANALYTICAL:** | |
| **Was method requirement followed?** | ✓ |
| **Was client requirement followed?** | ✓ |
| **Does the case narrative summarize all QC failure?** | ✓ |
| **All runlogs and manual integration are reviewed for requirements** | ✓ |
| **All manual calculations and /or hand notations verified** | ✓ |

**QA Review Signature:** SOHIL JODHANI **Date:** 07/22/2025

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

# LAB CHRONICLE

| | | | |
|---|---|---|---|
| **OrderID:** | Q2592 | **OrderDate:** | 7/11/2025 3:05:25 PM |
| **Client:** | PARSONS Engineering of New York, Inc. | **Project:** | Con Edison - East River Site 2 |
| **Contact:** | Zohar Lavy | **Location:** | D51,VOA Ref. #2 Soil |

| LabID | ClientID | Matrix | Test | Method | Sample Date | Prep Date | Anal Date | Received |
|---|---|---|---|---|---|---|---|---|
| **Q2592-01** | **WC-SOIL-20250711** | **SOIL** | | | **07/11/25** | | | **07/11/25** |
| | | | SVOCMS Group1 | 8270E | | 07/15/25 | 07/16/25 | |
| **Q2592-02** | **WC-SOIL-20250711** | **TCLP** | | | **07/11/25** | | | **07/11/25** |
| | | | TCLP BNA | 8270E | | 07/16/25 | 07/16/25 | |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Hit Summary Sheet
**SW-846**

**SDG No.:** Q2592

**Client:** PARSONS Engineering of New York, Inc.

| Sample ID | Client ID | Matrix | Parameter | | Concentration | C | MDL | RDL | Units |
|---|---|---|---|---|---|---|---|---|---|
| **Client ID :** | **WC-SOIL-20250711** | | | | | | | | |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Fluoranthene | | 760.000 | J | 200 | 1100 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Pyrene | | 690.000 | J | 230 | 1100 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Butylbenzylphthalate | | 860.000 | J | 460 | 1100 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Benzo(a)anthracene | | 490.000 | J | 150 | 1100 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Chrysene | | 570.000 | J | 130 | 1100 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Bis(2-ethylhexyl)phthalate | | 2,000.000 | | 390 | 1100 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Benzo(b)fluoranthene | | 810.000 | J | 120 | 1100 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Benzo(a)pyrene | | 620.000 | J | 190 | 1100 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Indeno(1,2,3-cd)pyrene | | 470.000 | J | 190 | 1100 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Benzo(g,h,i)perylene | | 640.000 | J | 170 | 1100 | ug/Kg |
| | | | **Total Svoc :** | | **7,910.00** | | | | |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Benzenamine, 2-methoxy- | * | 680.000 | J | 0 | 0 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Benzene, 1-isocyanato-2-methoxy | * | 470.000 | J | 0 | 0 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Benzophenone | * | 510.000 | J | 0 | 0 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | n-Hexadecanoic acid | * | 980.000 | J | 0 | 0 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Octadecanoic acid | * | 770.000 | J | 0 | 0 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Phenol, 4,4-(1-methylethylidene)b | * | 1,500.000 | J | 0 | 0 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | unknown16.113 | * | 460.000 | J | 0 | 0 | ug/Kg |
| Q2592-01 | WC-SOIL-20250711 | SOIL | Benzoic acid | * | 1,400.000 | J | 1200 | 2100 | ug/Kg |
| | | | **Total Tics :** | | **6,770.00** | | | | |
| | | | **Total Concentration:** | | **14,680.00** | | | | |

# QC SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Surrogate Summary

**SW-846**

**SDG No.:** Q2592

**Client:** PARSONS Engineering of New York, Inc.

**Analytical Method:** 8270E

| Lab Sample ID | Client ID | Parameter | Spike (PPM) | Result (PPM) | Recovery (%) | Qual | Limits (%) Low | Limits (%) High |
|---|---|---|---|---|---|---|---|---|
| PB168854BL | PB168854BL | 2-Fluorophenol | 150 | 110 | 73 | | 18 | 112 |
| | | Phenol-d6 | 150 | 110 | 73 | | 15 | 107 |
| | | Nitrobenzene-d5 | 100 | 82.0 | 82 | | 18 | 107 |
| | | 2-Fluorobiphenyl | 100 | 70.8 | 71 | | 20 | 109 |
| | | 2,4,6-Tribromophenol | 150 | 131 | 87 | | 10 | 116 |
| | | Terphenyl-d14 | 100 | 77.2 | 77 | | 10 | 105 |
| PB168854BS | PB168854BS | 2-Fluorophenol | 150 | 112 | 75 | | 18 | 112 |
| | | Phenol-d6 | 150 | 114 | 76 | | 15 | 107 |
| | | Nitrobenzene-d5 | 100 | 86.0 | 86 | | 18 | 107 |
| | | 2-Fluorobiphenyl | 100 | 72.4 | 72 | | 20 | 109 |
| | | 2,4,6-Tribromophenol | 150 | 140 | 94 | | 10 | 116 |
| | | Terphenyl-d14 | 100 | 79.7 | 80 | | 10 | 105 |
| Q2592-01 | WC-SOIL-20250711 | 2-Fluorophenol | 150 | 70.8 | 47 | | 18 | 112 |
| | | Phenol-d6 | 150 | 71.5 | 48 | | 15 | 107 |
| | | Nitrobenzene-d5 | 100 | 53.2 | 53 | | 18 | 107 |
| | | 2-Fluorobiphenyl | 100 | 55.4 | 55 | | 20 | 109 |
| | | 2,4,6-Tribromophenol | 150 | 97.1 | 65 | | 10 | 116 |
| | | Terphenyl-d14 | 100 | 58.2 | 58 | | 10 | 105 |
| Q2597-01MS | CHRT-26899MS | 2-Fluorophenol | 150 | 81.7 | 54 | | 18 | 112 |
| | | Phenol-d6 | 150 | 79.9 | 53 | | 15 | 107 |
| | | Nitrobenzene-d5 | 100 | 60.7 | 61 | | 18 | 107 |
| | | 2-Fluorobiphenyl | 100 | 62.3 | 62 | | 20 | 109 |
| | | 2,4,6-Tribromophenol | 150 | 106 | 71 | | 10 | 116 |
| | | Terphenyl-d14 | 100 | 63.9 | 64 | | 10 | 105 |
| Q2597-01MSD | CHRT-26899MSD | 2-Fluorophenol | 150 | 82.8 | 55 | | 18 | 112 |
| | | Phenol-d6 | 150 | 80.9 | 54 | | 15 | 107 |
| | | Nitrobenzene-d5 | 100 | 59.8 | 60 | | 18 | 107 |
| | | 2-Fluorobiphenyl | 100 | 60.6 | 61 | | 20 | 109 |
| | | 2,4,6-Tribromophenol | 150 | 105 | 70 | | 10 | 116 |
| | | Terphenyl-d14 | 100 | 62.7 | 63 | | 10 | 105 |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Matrix Spike/Matrix Spike Duplicate Summary**

**SW-846**

**SDG No.:** Q2592 **Analytical Method:** SW8270E

**Client:** PARSONS Engineering of New York, Inc. **DataFile:** BP025191.D

| Parameter | Spike | Sample Result | Result | Units | Rec | Rec Qual | RPD | RPD Qual | Limits Low | Limits High | RPD |
|---|---|---|---|---|---|---|---|---|---|---|---|
| **Lab Sample ID: Q2597-01MS** | | **Client Sample ID: CHRT-26899MS** | | | | | | | | | |
| Benzaldehyde | 1000 | 0 | 880 | ug/Kg | 88 | | | | 10 | 171 | |
| Phenol | 1000 | 0 | 860 | ug/Kg | 86 | | | | 51 | 122 | |
| bis(2-Chloroethyl)ether | 1000 | 0 | 840 | ug/Kg | 84 | | | | 54 | 125 | |
| 2-Chlorophenol | 1000 | 0 | 910 | ug/Kg | 91 | | | | 51 | 121 | |
| 2-Methylphenol | 1000 | 0 | 930 | ug/Kg | 93 | | | | 47 | 125 | |
| 2,2-oxybis(1-Chloropropane) | 1000 | 0 | 800 | ug/Kg | 80 | | | | 46 | 119 | |
| Acetophenone | 1000 | 0 | 800 | ug/Kg | 80 | | | | 55 | 128 | |
| 3+4-Methylphenols | 1000 | 0 | 900 | ug/Kg | 90 | | | | 49 | 125 | |
| N-Nitroso-di-n-propylamine | 1000 | 0 | 790 | ug/Kg | 79 | | | | 59 | 119 | |
| Hexachloroethane | 1000 | 0 | 920 | ug/Kg | 92 | | | | 51 | 116 | |
| Nitrobenzene | 1000 | 0 | 880 | ug/Kg | 88 | | | | 47 | 124 | |
| Isophorone | 1000 | 0 | 900 | ug/Kg | 90 | | | | 49 | 127 | |
| 2-Nitrophenol | 1000 | 0 | 1000 | ug/Kg | 100 | | | | 43 | 131 | |
| 2,4-Dimethylphenol | 1000 | 0 | 1200 | ug/Kg | 120 | | | | 63 | 151 | |
| bis(2-Chloroethoxy)methane | 1000 | 0 | 850 | ug/Kg | 85 | | | | 51 | 119 | |
| 2,4-Dichlorophenol | 1000 | 0 | 950 | ug/Kg | 95 | | | | 50 | 122 | |
| Naphthalene | 1000 | 0 | 840 | ug/Kg | 84 | | | | 51 | 121 | |
| 4-Chloroaniline | 1000 | 0 | 440 | ug/Kg | 44 | | | | 10 | 100 | |
| Hexachlorobutadiene | 1000 | 0 | 980 | ug/Kg | 98 | | | | 44 | 126 | |
| Caprolactam | 1000 | 0 | 860 | ug/Kg | 86 | | | | 51 | 134 | |
| 4-Chloro-3-methylphenol | 1000 | 0 | 880 | ug/Kg | 88 | | | | 57 | 132 | |
| 2-Methylnaphthalene | 1000 | 0 | 770 | ug/Kg | 77 | | | | 59 | 123 | |
| Hexachlorocyclopentadiene | 2000 | 0 | 2500 | ug/Kg | 125 | | | | 10 | 175 | |
| 2,4,6-Trichlorophenol | 1000 | 0 | 1100 | ug/Kg | 110 | | | | 33 | 141 | |
| 2,4,5-Trichlorophenol | 1000 | 0 | 1000 | ug/Kg | 100 | | | | 38 | 135 | |
| 1,1-Biphenyl | 1000 | 0 | 880 | ug/Kg | 88 | | | | 55 | 131 | |
| 2-Chloronaphthalene | 1000 | 0 | 910 | ug/Kg | 91 | | | | 48 | 124 | |
| 2-Nitroaniline | 1000 | 0 | 940 | ug/Kg | 94 | | | | 47 | 134 | |
| Dimethylphthalate | 1000 | 0 | 850 | ug/Kg | 85 | | | | 54 | 120 | |
| Acenaphthylene | 1000 | 0 | 950 | ug/Kg | 95 | | | | 57 | 125 | |
| 2,6-Dinitrotoluene | 1000 | 0 | 1000 | ug/Kg | 100 | | | | 48 | 127 | |
| 3-Nitroaniline | 1000 | 0 | 810 | ug/Kg | 81 | | | | 10 | 112 | |
| Acenaphthene | 1000 | 0 | 850 | ug/Kg | 85 | | | | 70 | 121 | |
| 2,4-Dinitrophenol | 2000 | 0 | 1100 | ug/Kg | 55 | | | | 10 | 155 | |
| 4-Nitrophenol | 2000 | 0 | 1800 | ug/Kg | 90 | | | | 10 | 175 | |
| Dibenzofuran | 1000 | 0 | 810 | ug/Kg | 81 | | | | 52 | 114 | |
| 2,4-Dinitrotoluene | 1000 | 0 | 900 | ug/Kg | 90 | | | | 41 | 140 | |
| Diethylphthalate | 1000 | 0 | 810 | ug/Kg | 81 | | | | 51 | 119 | |
| 4-Chlorophenyl-phenylether | 1000 | 0 | 820 | ug/Kg | 82 | | | | 48 | 122 | |
| Fluorene | 1000 | 0 | 790 | ug/Kg | 79 | | | | 53 | 118 | |
| 4-Nitroaniline | 1000 | 0 | 830 | ug/Kg | 83 | | | | 29 | 140 | |
| 4,6-Dinitro-2-methylphenol | 1000 | 0 | 690 | ug/Kg | 69 | | | | 10 | 160 | |
| N-Nitrosodiphenylamine | 1000 | 0 | 970 | ug/Kg | 97 | | | | 73 | 118 | |
| 4-Bromophenyl-phenylether | 1000 | 0 | 1100 | ug/Kg | 110 | | | | 65 | 121 | |
| Hexachlorobenzene | 1000 | 0 | 1000 | ug/Kg | 100 | | | | 67 | 118 | |
| Atrazine | 1000 | 0 | 1200 | ug/Kg | 120 | | | | 45 | 175 | |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Matrix Spike/Matrix Spike Duplicate Summary**

**SW-846**

**SDG No.:** Q2592 **Analytical Method:** SW8270E

**Client:** PARSONS Engineering of New York, Inc. **DataFile:** BP025191.D

| Parameter | Spike | Sample Result | Result | Units | Rec | Rec Qual | RPD | RPD Qual | Limits Low | Limits High | RPD |
|---|---|---|---|---|---|---|---|---|---|---|---|
| Pentachlorophenol | 2000 | 0 | 2000 | ug/Kg | 100 | | | | 13 | 153 | |
| Phenanthrene | 1000 | 0 | 880 | ug/Kg | 88 | | | | 52 | 128 | |
| Anthracene | 1000 | 0 | 920 | ug/Kg | 92 | | | | 62 | 124 | |
| Carbazole | 1000 | 0 | 870 | ug/Kg | 87 | | | | 59 | 119 | |
| Di-n-butylphthalate | 1000 | 0 | 920 | ug/Kg | 92 | | | | 55 | 125 | |
| Fluoranthene | 1000 | 100 | 910 | ug/Kg | 81 | | | | 44 | 125 | |
| Pyrene | 1000 | 110 | 850 | ug/Kg | 74 | | | | 37 | 122 | |
| Butylbenzylphthalate | 1000 | 0 | 930 | ug/Kg | 93 | | | | 44 | 135 | |
| 3,3-Dichlorobenzidine | 1000 | 0 | 560 | ug/Kg | 56 | | | | 15 | 112 | |
| Benzo(a)anthracene | 1000 | 62.5 | 930 | ug/Kg | 87 | | | | 53 | 119 | |
| Chrysene | 1000 | 92.0 | 930 | ug/Kg | 84 | | | | 57 | 121 | |
| bis(2-Ethylhexyl)phthalate | 1000 | 120 | 1100 | ug/Kg | 98 | | | | 42 | 169 | |
| Di-n-octyl phthalate | 1000 | 0 | 1000 | ug/Kg | 100 | | | | 51 | 156 | |
| Benzo(b)fluoranthene | 1000 | 140 | 930 | ug/Kg | 79 | | | | 52 | 117 | |
| Benzo(k)fluoranthene | 1000 | 46.9 | 890 | ug/Kg | 84 | | | | 57 | 134 | |
| Benzo(a)pyrene | 1000 | 63.6 | 1000 | ug/Kg | 94 | | | | 70 | 142 | |
| Indeno(1,2,3-cd)pyrene | 1000 | 62.9 | 1000 | ug/Kg | 94 | | | | 40 | 129 | |
| Dibenz(a,h)anthracene | 1000 | 0 | 980 | ug/Kg | 98 | | | | 43 | 123 | |
| Benzo(g,h,i)perylene | 1000 | 76.3 | 990 | ug/Kg | 91 | | | | 24 | 125 | |
| 1,2,4,5-Tetrachlorobenzene | 1000 | 0 | 950 | ug/Kg | 95 | | | | 52 | 134 | |
| 1,4-Dioxane | 1000 | 0 | 680 | ug/Kg | 68 | | | | 46 | 112 | |
| 2,3,4,6-Tetrachlorophenol | 1000 | 0 | 950 | ug/Kg | 95 | | | | 24 | 146 | |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Matrix Spike/Matrix Spike Duplicate Summary**

**SW-846**

**SDG No.:** Q2592 **Analytical Method:** SW8270E

**Client:** PARSONS Engineering of New York, Inc. **DataFile:** BP025192.D

| Parameter | Spike | Sample Result | Result | Units | Rec | Rec Qual | RPD | RPD Qual | Limits Low | Limits High | RPD |
|---|---|---|---|---|---|---|---|---|---|---|---|
| **Lab Sample ID: Q2597-01MSD** | | **Client Sample ID: CHRT-26899MSD** | | | | | | | | | |
| Benzaldehyde | 1000 | 0 | 870 | ug/Kg | 87 | 1 | | | 10 | 171 | 20 |
| Phenol | 1000 | 0 | 870 | ug/Kg | 87 | 1 | | | 51 | 122 | 20 |
| bis(2-Chloroethyl)ether | 1000 | 0 | 840 | ug/Kg | 84 | 0 | | | 54 | 125 | 20 |
| 2-Chlorophenol | 1000 | 0 | 930 | ug/Kg | 93 | 2 | | | 51 | 121 | 20 |
| 2-Methylphenol | 1000 | 0 | 930 | ug/Kg | 93 | 0 | | | 47 | 125 | 20 |
| 2,2-oxybis(1-Chloropropane) | 1000 | 0 | 800 | ug/Kg | 80 | 0 | | | 46 | 119 | 20 |
| Acetophenone | 1000 | 0 | 800 | ug/Kg | 80 | 0 | | | 55 | 128 | 20 |
| 3+4-Methylphenols | 1000 | 0 | 900 | ug/Kg | 90 | 0 | | | 49 | 125 | 20 |
| N-Nitroso-di-n-propylamine | 1000 | 0 | 790 | ug/Kg | 79 | 0 | | | 59 | 119 | 20 |
| Hexachloroethane | 1000 | 0 | 920 | ug/Kg | 92 | 0 | | | 51 | 116 | 20 |
| Nitrobenzene | 1000 | 0 | 860 | ug/Kg | 86 | 2 | | | 47 | 124 | 20 |
| Isophorone | 1000 | 0 | 890 | ug/Kg | 89 | 1 | | | 49 | 127 | 20 |
| 2-Nitrophenol | 1000 | 0 | 1000 | ug/Kg | 100 | 0 | | | 43 | 131 | 20 |
| 2,4-Dimethylphenol | 1000 | 0 | 1200 | ug/Kg | 120 | 0 | | | 63 | 151 | 20 |
| bis(2-Chloroethoxy)methane | 1000 | 0 | 840 | ug/Kg | 84 | 1 | | | 51 | 119 | 20 |
| 2,4-Dichlorophenol | 1000 | 0 | 960 | ug/Kg | 96 | 1 | | | 50 | 122 | 20 |
| Naphthalene | 1000 | 0 | 840 | ug/Kg | 84 | 0 | | | 51 | 121 | 20 |
| 4-Chloroaniline | 1000 | 0 | 460 | ug/Kg | 46 | 4 | | | 10 | 100 | 20 |
| Hexachlorobutadiene | 1000 | 0 | 980 | ug/Kg | 98 | 0 | | | 44 | 126 | 20 |
| Caprolactam | 1000 | 0 | 850 | ug/Kg | 85 | 1 | | | 51 | 134 | 20 |
| 4-Chloro-3-methylphenol | 1000 | 0 | 880 | ug/Kg | 88 | 0 | | | 57 | 132 | 20 |
| 2-Methylnaphthalene | 1000 | 0 | 760 | ug/Kg | 76 | 1 | | | 59 | 123 | 20 |
| Hexachlorocyclopentadiene | 2000 | 0 | 2300 | ug/Kg | 115 | 8 | | | 10 | 175 | 20 |
| 2,4,6-Trichlorophenol | 1000 | 0 | 1100 | ug/Kg | 110 | 0 | | | 33 | 141 | 20 |
| 2,4,5-Trichlorophenol | 1000 | 0 | 1000 | ug/Kg | 100 | 0 | | | 38 | 135 | 20 |
| 1,1-Biphenyl | 1000 | 0 | 860 | ug/Kg | 86 | 2 | | | 55 | 131 | 20 |
| 2-Chloronaphthalene | 1000 | 0 | 900 | ug/Kg | 90 | 1 | | | 48 | 124 | 20 |
| 2-Nitroaniline | 1000 | 0 | 940 | ug/Kg | 94 | 0 | | | 47 | 134 | 20 |
| Dimethylphthalate | 1000 | 0 | 830 | ug/Kg | 83 | 2 | | | 54 | 120 | 20 |
| Acenaphthylene | 1000 | 0 | 940 | ug/Kg | 94 | 1 | | | 57 | 125 | 20 |
| 2,6-Dinitrotoluene | 1000 | 0 | 1000 | ug/Kg | 100 | 0 | | | 48 | 127 | 20 |
| 3-Nitroaniline | 1000 | 0 | 800 | ug/Kg | 80 | 1 | | | 10 | 112 | 20 |
| Acenaphthene | 1000 | 0 | 820 | ug/Kg | 82 | 4 | | | 70 | 121 | 20 |
| 2,4-Dinitrophenol | 2000 | 0 | 1100 | ug/Kg | 55 | 0 | | | 10 | 155 | 20 |
| 4-Nitrophenol | 2000 | 0 | 1800 | ug/Kg | 90 | 0 | | | 10 | 175 | 20 |
| Dibenzofuran | 1000 | 0 | 790 | ug/Kg | 79 | 3 | | | 52 | 114 | 20 |
| 2,4-Dinitrotoluene | 1000 | 0 | 900 | ug/Kg | 90 | 0 | | | 41 | 140 | 20 |
| Diethylphthalate | 1000 | 0 | 800 | ug/Kg | 80 | 1 | | | 51 | 119 | 20 |
| 4-Chlorophenyl-phenylether | 1000 | 0 | 810 | ug/Kg | 81 | 1 | | | 48 | 122 | 20 |
| Fluorene | 1000 | 0 | 790 | ug/Kg | 79 | 0 | | | 53 | 118 | 20 |
| 4-Nitroaniline | 1000 | 0 | 820 | ug/Kg | 82 | 1 | | | 29 | 140 | 20 |
| 4,6-Dinitro-2-methylphenol | 1000 | 0 | 670 | ug/Kg | 67 | 3 | | | 10 | 160 | 20 |
| N-Nitrosodiphenylamine | 1000 | 0 | 950 | ug/Kg | 95 | 2 | | | 73 | 118 | 20 |
| 4-Bromophenyl-phenylether | 1000 | 0 | 1100 | ug/Kg | 110 | 0 | | | 65 | 121 | 20 |
| Hexachlorobenzene | 1000 | 0 | 1000 | ug/Kg | 100 | 0 | | | 67 | 118 | 20 |
| Atrazine | 1000 | 0 | 1200 | ug/Kg | 120 | 0 | | | 45 | 175 | 20 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Matrix Spike/Matrix Spike Duplicate Summary**

**SW-846**

**SDG No.:** Q2592 **Analytical Method:** SW8270E

**Client:** PARSONS Engineering of New York, Inc. **DataFile:** BP025192.D

| Parameter | Spike | Sample Result | Result | Units | Rec | Rec Qual | RPD | RPD Qual | Limits Low | Limits High | RPD |
|---|---|---|---|---|---|---|---|---|---|---|---|
| Pentachlorophenol | 2000 | 0 | 2000 | ug/Kg | 100 | 0 | | | 13 | 153 | 20 |
| Phenanthrene | 1000 | 0 | 850 | ug/Kg | 85 | 3 | | | 52 | 128 | 20 |
| Anthracene | 1000 | 0 | 910 | ug/Kg | 91 | 1 | | | 62 | 124 | 20 |
| Carbazole | 1000 | 0 | 860 | ug/Kg | 86 | 1 | | | 59 | 119 | 20 |
| Di-n-butylphthalate | 1000 | 0 | 890 | ug/Kg | 89 | 3 | | | 55 | 125 | 20 |
| Fluoranthene | 1000 | 100 | 890 | ug/Kg | 79 | 3 | | | 44 | 125 | 20 |
| Pyrene | 1000 | 110 | 860 | ug/Kg | 75 | 1 | | | 37 | 122 | 20 |
| Butylbenzylphthalate | 1000 | 0 | 920 | ug/Kg | 92 | 1 | | | 44 | 135 | 20 |
| 3,3-Dichlorobenzidine | 1000 | 0 | 600 | ug/Kg | 60 | 7 | | | 15 | 112 | 20 |
| Benzo(a)anthracene | 1000 | 62.5 | 930 | ug/Kg | 87 | 0 | | | 53 | 119 | 20 |
| Chrysene | 1000 | 92.0 | 930 | ug/Kg | 84 | 0 | | | 57 | 121 | 20 |
| bis(2-Ethylhexyl)phthalate | 1000 | 120 | 1100 | ug/Kg | 98 | 0 | | | 42 | 169 | 20 |
| Di-n-octyl phthalate | 1000 | 0 | 1000 | ug/Kg | 100 | 0 | | | 51 | 156 | 20 |
| Benzo(b)fluoranthene | 1000 | 140 | 960 | ug/Kg | 82 | 4 | | | 52 | 117 | 20 |
| Benzo(k)fluoranthene | 1000 | 46.9 | 850 | ug/Kg | 80 | 5 | | | 57 | 134 | 20 |
| Benzo(a)pyrene | 1000 | 63.6 | 1000 | ug/Kg | 94 | 0 | | | 70 | 142 | 20 |
| Indeno(1,2,3-cd)pyrene | 1000 | 62.9 | 1000 | ug/Kg | 94 | 0 | | | 40 | 129 | 20 |
| Dibenz(a,h)anthracene | 1000 | 0 | 970 | ug/Kg | 97 | 1 | | | 43 | 123 | 20 |
| Benzo(g,h,i)perylene | 1000 | 76.3 | 980 | ug/Kg | 90 | 1 | | | 24 | 125 | 20 |
| 1,2,4,5-Tetrachlorobenzene | 1000 | 0 | 940 | ug/Kg | 94 | 1 | | | 52 | 134 | 20 |
| 1,4-Dioxane | 1000 | 0 | 670 | ug/Kg | 67 | 1 | | | 46 | 112 | 20 |
| 2,3,4,6-Tetrachlorophenol | 1000 | 0 | 940 | ug/Kg | 94 | 1 | | | 24 | 146 | 20 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

**SW-846**

**SDG No.:** Q2592 **Analytical Method:** 8270E

**Client:** PARSONS Engineering of New York, Inc. **DataFile:** BF143119.D

| Lab Sample ID | Parameter | Spike | Result | Unit | Rec | RPD | Qual | RPD Qual | Limits Low | Limits High | Limits RPD |
|---|---|---|---|---|---|---|---|---|---|---|---|
| PB168854BS | Benzaldehyde | 1700 | 1200 | ug/Kg | 71 | | | | 10 | 133 | |
| | Phenol | 1700 | 1500 | ug/Kg | 88 | | | | 62 | 112 | |
| | bis(2-Chloroethyl)ether | 1700 | 1400 | ug/Kg | 82 | | | | 60 | 101 | |
| | 2-Chlorophenol | 1700 | 1500 | ug/Kg | 88 | | | | 65 | 112 | |
| | 2-Methylphenol | 1700 | 1500 | ug/Kg | 88 | | | | 61 | 108 | |
| | 2,2-oxybis(1-Chloropropane) | 1700 | 1400 | ug/Kg | 82 | | | | 51 | 100 | |
| | Acetophenone | 1700 | 1400 | ug/Kg | 82 | | | | 66 | 98 | |
| | 3+4-Methylphenols | 1700 | 1500 | ug/Kg | 88 | | | | 58 | 111 | |
| | N-Nitroso-di-n-propylamine | 1700 | 1500 | ug/Kg | 88 | | | | 63 | 95 | |
| | Hexachloroethane | 1700 | 1500 | ug/Kg | 88 | | | | 72 | 108 | |
| | Nitrobenzene | 1700 | 1600 | ug/Kg | 94 | | | | 57 | 101 | |
| | Isophorone | 1700 | 1500 | ug/Kg | 88 | | | | 59 | 99 | |
| | 2-Nitrophenol | 1700 | 1600 | ug/Kg | 94 | | | | 61 | 111 | |
| | 2,4-Dimethylphenol | 1700 | 1500 | ug/Kg | 88 | | | | 46 | 141 | |
| | bis(2-Chloroethoxy)methane | 1700 | 1500 | ug/Kg | 88 | | | | 66 | 97 | |
| | 2,4-Dichlorophenol | 1700 | 1500 | ug/Kg | 88 | | | | 62 | 107 | |
| | Naphthalene | 1700 | 1400 | ug/Kg | 82 | | | | 62 | 100 | |
| | 4-Chloroaniline | 1700 | 620 | ug/Kg | 36 | | | | 16 | 100 | |
| | Hexachlorobutadiene | 1700 | 1500 | ug/Kg | 88 | | | | 53 | 98 | |
| | Caprolactam | 1700 | 1700 | ug/Kg | 100 | | | | 67 | 110 | |
| | 4-Chloro-3-methylphenol | 1700 | 1500 | ug/Kg | 88 | | | | 58 | 112 | |
| | 2-Methylnaphthalene | 1700 | 1500 | ug/Kg | 88 | | | | 60 | 104 | |
| | Hexachlorocyclopentadiene | 3300 | 3100 | ug/Kg | 94 | | | | 45 | 165 | |
| | 2,4,6-Trichlorophenol | 1700 | 1600 | ug/Kg | 94 | | | | 59 | 102 | |
| | 2,4,5-Trichlorophenol | 1700 | 1600 | ug/Kg | 94 | | | | 61 | 98 | |
| | 1,1-Biphenyl | 1700 | 1400 | ug/Kg | 82 | | | | 57 | 103 | |
| | 2-Chloronaphthalene | 1700 | 1400 | ug/Kg | 82 | | | | 58 | 99 | |
| | 2-Nitroaniline | 1700 | 1600 | ug/Kg | 94 | | | | 66 | 101 | |
| | Dimethylphthalate | 1700 | 1500 | ug/Kg | 88 | | | | 61 | 99 | |
| | Acenaphthylene | 1700 | 1500 | ug/Kg | 88 | | | | 63 | 101 | |
| | 2,6-Dinitrotoluene | 1700 | 1600 | ug/Kg | 94 | | | | 61 | 104 | |
| | 3-Nitroaniline | 1700 | 1100 | ug/Kg | 65 | | | | 28 | 100 | |
| | Acenaphthene | 1700 | 1500 | ug/Kg | 88 | | | | 57 | 104 | |
| | 2,4-Dinitrophenol | 3300 | 3200 | ug/Kg | 97 | | | | 37 | 128 | |
| | 4-Nitrophenol | 3300 | 3400 | ug/Kg | 103 | | | | 48 | 119 | |
| | Dibenzofuran | 1700 | 1500 | ug/Kg | 88 | | | | 63 | 99 | |
| | 2,4-Dinitrotoluene | 1700 | 1700 | ug/Kg | 100 | | | | 60 | 106 | |
| | Diethylphthalate | 1700 | 1600 | ug/Kg | 94 | | | | 60 | 101 | |
| | 4-Chlorophenyl-phenylether | 1700 | 1500 | ug/Kg | 88 | | | | 58 | 98 | |
| | Fluorene | 1700 | 1400 | ug/Kg | 82 | | | | 61 | 101 | |
| | 4-Nitroaniline | 1700 | 1800 | ug/Kg | 106 | | * | | 64 | 103 | |
| | 4,6-Dinitro-2-methylphenol | 1700 | 1700 | ug/Kg | 100 | | | | 76 | 113 | |
| | N-Nitrosodiphenylamine | 1700 | 1500 | ug/Kg | 88 | | | | 71 | 99 | |
| | 4-Bromophenyl-phenylether | 1700 | 1500 | ug/Kg | 88 | | | | 66 | 102 | |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

**SW-846**

**SDG No.:** Q2592 **Analytical Method:** 8270E

**Client:** PARSONS Engineering of New York, Inc. **DataFile:** BF143119.D

| Lab Sample ID | Parameter | Spike | Result | Unit | Rec | RPD | Qual | RPD Qual | Limits Low | Limits High | RPD |
|---|---|---|---|---|---|---|---|---|---|---|---|
| PB168854BS | Hexachlorobenzene | 1700 | 1500 | ug/Kg | 88 | | | | 64 | 98 | |
| | Atrazine | 1700 | 1700 | ug/Kg | 100 | | | | 47 | 152 | |
| | Pentachlorophenol | 3300 | 3300 | ug/Kg | 100 | | | | 67 | 105 | |
| | Phenanthrene | 1700 | 1400 | ug/Kg | 82 | | | | 59 | 103 | |
| | Anthracene | 1700 | 1500 | ug/Kg | 88 | | | | 61 | 105 | |
| | Carbazole | 1700 | 1500 | ug/Kg | 88 | | | | 61 | 99 | |
| | Di-n-butylphthalate | 1700 | 1700 | ug/Kg | 100 | | | | 58 | 104 | |
| | Fluoranthene | 1700 | 1500 | ug/Kg | 88 | | | | 57 | 107 | |
| | Pyrene | 1700 | 1500 | ug/Kg | 88 | | | | 59 | 103 | |
| | Butylbenzylphthalate | 1700 | 1700 | ug/Kg | 100 | | | | 55 | 103 | |
| | 3,3-Dichlorobenzidine | 1700 | 780 | ug/Kg | 46 | | | | 42 | 91 | |
| | Benzo(a)anthracene | 1700 | 1600 | ug/Kg | 94 | | | | 60 | 102 | |
| | Chrysene | 1700 | 1500 | ug/Kg | 88 | | | | 59 | 101 | |
| | bis(2-Ethylhexyl)phthalate | 1700 | 1600 | ug/Kg | 94 | | | | 54 | 135 | |
| | Di-n-octyl phthalate | 1700 | 1500 | ug/Kg | 88 | | | | 52 | 137 | |
| | Benzo(b)fluoranthene | 1700 | 1500 | ug/Kg | 88 | | | | 62 | 109 | |
| | Benzo(k)fluoranthene | 1700 | 1700 | ug/Kg | 100 | | | | 62 | 109 | |
| | Benzo(a)pyrene | 1700 | 1600 | ug/Kg | 94 | | | | 63 | 103 | |
| | Indeno(1,2,3-cd)pyrene | 1700 | 1500 | ug/Kg | 88 | | | | 63 | 101 | |
| | Dibenz(a,h)anthracene | 1700 | 1500 | ug/Kg | 88 | | | | 61 | 112 | |
| | Benzo(g,h,i)perylene | 1700 | 1500 | ug/Kg | 88 | | | | 70 | 108 | |
| | 1,2,4,5-Tetrachlorobenzene | 1700 | 1400 | ug/Kg | 82 | | | | 53 | 101 | |
| | 1,4-Dioxane | 1700 | 1200 | ug/Kg | 71 | | | | 50 | 96 | |
| | 2,3,4,6-Tetrachlorophenol | 1700 | 1700 | ug/Kg | 100 | | | | 59 | 108 | |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID: PB168854BL

Lab Name: Alliance | Contract: PARS02

Lab Code: ACE | SDG NO.: Q2592

Lab File ID: BF143118.D | Lab Sample ID: PB168854BL

Instrument ID: BNA_F | Date Extracted: 07/15/2025

Matrix: (soil/water) SOIL | Date Analyzed: 07/16/2025

Level: (low/med) LOW | Time Analyzed: 20:09

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA SAMPLE NO. | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED |
|---|---|---|---|
| PB168854BS | PB168854BS | BF143119.D | 07/16/2025 |
| WC-SOIL-20250711 | Q2592-01 | BP025166.D | 07/16/2025 |
| CHRT-26899MS | Q2597-01MS | BP025191.D | 07/18/2025 |
| CHRT-26899MSD | Q2597-01MSD | BP025192.D | 07/18/2025 |

COMMENTS:

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance Contract: PARS02

Lab Code: ACE SDG NO.: Q2592

Lab File ID: BF143092.D DFTPP Injection Date: 07/15/2025

Instrument ID: BNA_F DFTPP Injection Time: 12:35

| m/e | ION ABUNDANCE CRITERIA | % RELATIVE ABUNDANCE |
|---|---|---|
| 68 | Less than 2.0% of mass 69 | 0.6 ( 2 ) 1 |
| 69 | Mass 69 relative abundance | 100 |
| 70 | Less than 2.0% of mass 69 | 0.2 ( 0.5 ) 1 |
| 197 | Less than 2.0% of mass 198 | 0.3 |
| 198 | Base Peak, 100% relative abundance | 100 |
| 199 | 5.0 to 9.0% of mass 198 | 6.5 |
| 365 | Greater than 1% of mass 198 | 4.0 |
| 441 | Present, but less than mass 443 | 77.7 |
| 442 | Greater than 50% of mass 198 | 100 |
| 443 | 15.0 - 24.0% of mass 442 | 19.4 (19.4) 2 |

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA SAMPLE NO. | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|---|---|---|---|---|
| SSTDICC2.5 | SSTDICC2.5 | BF143093.D | 07/15/2025 | 13:04 |
| SSTDICC005 | SSTDICC005 | BF143094.D | 07/15/2025 | 13:35 |
| SSTDICC010 | SSTDICC010 | BF143095.D | 07/15/2025 | 14:05 |
| SSTDICC020 | SSTDICC020 | BF143096.D | 07/15/2025 | 14:36 |
| SSTDICCC040 | SSTDICCC040 | BF143097.D | 07/15/2025 | 15:05 |
| SSTDICC050 | SSTDICC050 | BF143098.D | 07/15/2025 | 15:35 |
| SSTDICC060 | SSTDICC060 | BF143099.D | 07/15/2025 | 16:05 |
| SSTDICC080 | SSTDICC080 | BF143100.D | 07/15/2025 | 16:35 |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance Contract: PARS02

Lab Code: ACE SDG NO.: Q2592

Lab File ID: BF143116.D DFTPP Injection Date: 07/16/2025

Instrument ID: BNA_F DFTPP Injection Time: 19:09

| m/e | ION ABUNDANCE CRITERIA | % RELATIVE ABUNDANCE |
|---|---|---|
| 68 | Less than 2.0% of mass 69 | 0.6 ( 1.9 ) 1 |
| 69 | Mass 69 relative abundance | 100 |
| 70 | Less than 2.0% of mass 69 | 0.2 ( 0.6 ) 1 |
| 197 | Less than 2.0% of mass 198 | 0.7 |
| 198 | Base Peak, 100% relative abundance | 100 |
| 199 | 5.0 to 9.0% of mass 198 | 6.8 |
| 365 | Greater than 1% of mass 198 | 3.6 |
| 441 | Present, but less than mass 443 | 79.1 |
| 442 | Greater than 50% of mass 198 | 100 |
| 443 | 15.0 - 24.0% of mass 442 | 19.1 (19.1) 2 |

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA SAMPLE NO. | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|---|---|---|---|---|
| SSTDCCC040 | SSTDCCC040 | BF143117.D | 07/16/2025 | 19:39 |
| PB168854BL | PB168854BL | BF143118.D | 07/16/2025 | 20:09 |
| PB168854BS | PB168854BS | BF143119.D | 07/16/2025 | 20:39 |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance | Contract: PARS02

Lab Code: ACE | SDG NO.: Q2592

Lab File ID: BP025070.D | DFTPP Injection Date: 07/03/2025

Instrument ID: BNA_P | DFTPP Injection Time: 08:58

| m/e | ION ABUNDANCE CRITERIA | % RELATIVE ABUNDANCE |
|---|---|---|
| 68 | Less than 2.0% of mass 69 | 0.5 ( 1.7 ) 1 |
| 69 | Mass 69 relative abundance | 100 |
| 70 | Less than 2.0% of mass 69 | 0.1 ( 0.4 ) 1 |
| 197 | Less than 2.0% of mass 198 | 0.3 |
| 198 | Base Peak, 100% relative abundance | 100 |
| 199 | 5.0 to 9.0% of mass 198 | 6.8 |
| 365 | Greater than 1% of mass 198 | 3.7 |
| 441 | Present, but less than mass 443 | 79.7 |
| 442 | Greater than 50% of mass 198 | 100.0 |
| 443 | 15.0 - 24.0% of mass 442 | 18 (19.4) 2 |

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA SAMPLE NO. | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|---|---|---|---|---|
| SSTDICC2.5 | SSTDICC2.5 | BP025071.D | 07/03/2025 | 09:39 |
| SSTDICC005 | SSTDICC005 | BP025072.D | 07/03/2025 | 10:21 |
| SSTDICC010 | SSTDICC010 | BP025073.D | 07/03/2025 | 11:02 |
| SSTDICC020 | SSTDICC020 | BP025074.D | 07/03/2025 | 11:44 |
| SSTDICCC040 | SSTDICCC040 | BP025075.D | 07/03/2025 | 12:25 |
| SSTDICC050 | SSTDICC050 | BP025076.D | 07/03/2025 | 13:07 |
| SSTDICC060 | SSTDICC060 | BP025077.D | 07/03/2025 | 13:49 |
| SSTDICC080 | SSTDICC080 | BP025078.D | 07/03/2025 | 14:31 |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance Contract: PARS02

Lab Code: ACE SDG NO.: Q2592

Lab File ID: BP025154.D DFTPP Injection Date: 07/16/2025

Instrument ID: BNA_P DFTPP Injection Time: 11:19

| m/e | ION ABUNDANCE CRITERIA | % RELATIVE ABUNDANCE |
|---|---|---|
| 68 | Less than 2.0% of mass 69 | 0.5 ( 1.7 ) 1 |
| 69 | Mass 69 relative abundance | 100 |
| 70 | Less than 2.0% of mass 69 | 0.2 ( 0.6 ) 1 |
| 197 | Less than 2.0% of mass 198 | 0.2 |
| 198 | Base Peak, 100% relative abundance | 100 |
| 199 | 5.0 to 9.0% of mass 198 | 6.8 |
| 365 | Greater than 1% of mass 198 | 3.6 |
| 441 | Present, but less than mass 443 | 78.5 |
| 442 | Greater than 50% of mass 198 | 100 |
| 443 | 15.0 - 24.0% of mass 442 | 18.5 (19.8) 2 |

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA SAMPLE NO. | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|---|---|---|---|---|
| SSTDCCC040 | SSTDCCC040 | BP025155.D | 07/16/2025 | 12:42 |
| WC-SOIL-20250711 | Q2592-01 | BP025166.D | 07/16/2025 | 20:22 |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance Contract: PARS02

Lab Code: ACE SDG NO.: Q2592

Lab File ID: BP025186.D DFTPP Injection Date: 07/18/2025

Instrument ID: BNA_P DFTPP Injection Time: 09:41

| m/e | ION ABUNDANCE CRITERIA | % RELATIVE ABUNDANCE |
|---|---|---|
| 68 | Less than 2.0% of mass 69 | 0.4 ( 1.8 ) 1 |
| 69 | Mass 69 relative abundance | 100 |
| 70 | Less than 2.0% of mass 69 | 0.1 ( 0.5 ) 1 |
| 197 | Less than 2.0% of mass 198 | 0.0 |
| 198 | Base Peak, 100% relative abundance | 100 |
| 199 | 5.0 to 9.0% of mass 198 | 6.8 |
| 365 | Greater than 1% of mass 198 | 4.2 |
| 441 | Present, but less than mass 443 | 80.7 |
| 442 | Greater than 50% of mass 198 | 100 |
| 443 | 15.0 - 24.0% of mass 442 | 19.2 (19.2) 2 |

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA SAMPLE NO. | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|---|---|---|---|---|
| SSTDCCC040 | SSTDCCC040 | BP025187.D | 07/18/2025 | 11:04 |
| CHRT-26899MS | Q2597-01MS | BP025191.D | 07/18/2025 | 13:52 |
| CHRT-26899MSD | Q2597-01MSD | BP025192.D | 07/18/2025 | 14:33 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q2592

Client ID : SSTDCCC040 Date Analyzed: 07/16/2025

Lab File ID: BF143117.D Time Analyzed: 19:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

| | | IS1 (DCB) AREA # | RT # | IS2 (NPT) AREA # | RT # | IS3 (ANT) AREA # | RT # |
|---|---|---|---|---|---|---|---|
| | 12 HOUR STD | 135149 | 6.969 | 510164 | 8.25 | 254065 | 10.00 |
| | UPPER LIMIT | 270298 | 7.469 | 1020330 | 8.751 | 508130 | 10.504 |
| | LOWER LIMIT | 67574.5 | 6.469 | 255082 | 7.751 | 127033 | 9.504 |
| | EPA SAMPLE NO. | | | | | | |
| 01 | PB168854BS | 139596 | 6.97 | 538805 | 8.25 | 272708 | 10.00 |
| 02 | PB168854BL | 138731 | 6.96 | 528830 | 8.25 | 273109 | 10.00 |

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

## SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q2592

Client ID: SSTDCCC040 Date Analyzed: 07/16/2025

Lab File ID: BF143117.D Time Analyzed: 19:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

| | | IS4 (PHN) AREA # | RT # | IS5 (CRY) AREA # | RT # | IS6 (PRY) AREA # | RT # |
|---|---|---|---|---|---|---|---|
| | 12 HOUR STD | 397173 | 11.492 | 210714 | 14.127 | 255524 | 15.633 |
| | UPPER LIMIT | 794346 | 11.992 | 421428 | 14.627 | 511048 | 16.133 |
| | LOWER LIMIT | 198587 | 10.992 | 105357 | 13.627 | 127762 | 15.133 |
| | EPA SAMPLE NO. | | | | | | |
| 01 | PB168854BS | 455589 | 11.49 | 240950 | 14.13 | 262866 | 15.63 |
| 02 | PB168854BL | 466923 | 11.49 | 262735 | 14.13 | 225267 | 15.63 |

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q2592

Client ID : SSTDCCC040 Date Analyzed: 07/16/2025

Lab File ID: BP025155.D Time Analyzed: 12:42

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

| | | IS1 (DCB) AREA # | RT # | IS2 (NPT) AREA # | RT # | IS3 (ANT) AREA # | RT # |
|---|---|---|---|---|---|---|---|
| | 12 HOUR STD | 564997 | 7.443 | 2237580 | 10.20 | 1465070 | 14.10 |
| | UPPER LIMIT | 1129990 | 7.943 | 4475160 | 10.696 | 2930140 | 14.595 |
| | LOWER LIMIT | 282499 | 6.943 | 1118790 | 9.696 | 732535 | 13.595 |
| | EPA SAMPLE NO. | | | | | | |
| 01 | WC-SOIL-20250711 | 569178 | 7.44 | 2223520 | 10.19 | 1404760 | 14.10 |

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q2592

Client ID: SSTDCCC040 Date Analyzed: 07/16/2025

Lab File ID: BP025155.D Time Analyzed: 12:42

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

| | | IS4 (PHN) AREA # | RT # | IS5 (CRY) AREA # | RT # | IS6 (PRY) AREA # | RT # |
|---|---|---|---|---|---|---|---|
| | 12 HOUR STD | 2922940 | 16.93 | 2997870 | 21.36 | 3472370 | 24.501 |
| | UPPER LIMIT | 5845880 | 17.43 | 5995740 | 21.86 | 6944740 | 25.001 |
| | LOWER LIMIT | 1461470 | 16.43 | 1498940 | 20.86 | 1736190 | 24.001 |
| | EPA SAMPLE NO. | | | | | | |
| 01 | WC-SOIL-20250711 | 2849470 | 16.92 | 3032750 | 21.35 | 3540160 | 24.50 |

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q2592

Client ID : SSTDCCC040 Date Analyzed: 07/18/2025

Lab File ID: BP025187.D Time Analyzed: 11:04

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

| | | IS1 (DCB) AREA # | RT # | IS2 (NPT) AREA # | RT # | IS3 (ANT) AREA # | RT # |
|---|---|---|---|---|---|---|---|
| | 12 HOUR STD | 531610 | 7.437 | 2146080 | 10.19 | 1427550 | 14.10 |
| | UPPER LIMIT | 1063220 | 7.937 | 4292160 | 10.69 | 2855100 | 14.601 |
| | LOWER LIMIT | 265805 | 6.937 | 1073040 | 9.69 | 713775 | 13.601 |
| | EPA SAMPLE NO. | | | | | | |
| 01 | CHRT-26899MS | 725237 | 7.44 | 2872260 | 10.18 | 1756650 | 14.10 |
| 02 | CHRT-26899MSD | 718069 | 7.44 | 2875360 | 10.19 | 1762350 | 14.10 |

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

## SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE SDG NO.: Q2592

Client ID: SSTDCCC040 Date Analyzed: 07/18/2025

Lab File ID: BP025187.D Time Analyzed: 11:04

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

| | | IS4 (PHN) AREA # | RT # | IS5 (CRY) AREA # | RT # | IS6 (PRY) AREA # | RT # |
|---|---|---|---|---|---|---|---|
| | 12 HOUR STD | 2869300 | 16.925 | 3091730 | 21.348 | 3620680 | 24.495 |
| | UPPER LIMIT | 5738600 | 17.425 | 6183460 | 21.848 | 7241360 | 24.995 |
| | LOWER LIMIT | 1434650 | 16.425 | 1545870 | 20.848 | 1810340 | 23.995 |
| | EPA SAMPLE NO. | | | | | | |
| 01 | CHRT-26899MS | 3030590 | 16.92 | 3384850 | 21.36 | 4114550 | 24.51 |
| 02 | CHRT-26899MSD | 3047800 | 16.92 | 3398850 | 21.36 | 4139650 | 24.52 |

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

# SAMPLE DATA

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | 07/11/25 |
| Project: | Con Edison - East River Site 2 | Date Received: | 07/11/25 |
| Client Sample ID: | WC-SOIL-20250711 | SDG No.: | Q2592 |
| Lab Sample ID: | Q2592-01 | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 76.7 |
| Sample Wt/Vol: | 30.07 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BP025166.D | 5 | 07/15/25 09:05 | 07/16/25 20:22 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| **TARGETS** | | | | | | |
| 100-52-7 | Benzaldehyde | 2100 | U | 1000 | 2100 | ug/Kg |
| 108-95-2 | Phenol | 1100 | U | 140 | 1100 | ug/Kg |
| 111-44-4 | bis(2-Chloroethyl)ether | 1100 | U | 160 | 1100 | ug/Kg |
| 95-57-8 | 2-Chlorophenol | 1100 | U | 160 | 1100 | ug/Kg |
| 95-48-7 | 2-Methylphenol | 1100 | U | 190 | 1100 | ug/Kg |
| 108-60-1 | 2,2-oxybis(1-Chloropropane) | 1100 | U | 240 | 1100 | ug/Kg |
| 98-86-2 | Acetophenone | 1100 | U | 190 | 1100 | ug/Kg |
| 65794-96-9 | 3+4-Methylphenols | 2100 | U | 270 | 2100 | ug/Kg |
| 621-64-7 | n-Nitroso-di-n-propylamine | 520 | U | 310 | 520 | ug/Kg |
| 67-72-1 | Hexachloroethane | 1100 | U | 110 | 1100 | ug/Kg |
| 98-95-3 | Nitrobenzene | 1100 | U | 120 | 1100 | ug/Kg |
| 78-59-1 | Isophorone | 1100 | U | 210 | 1100 | ug/Kg |
| 88-75-5 | 2-Nitrophenol | 1100 | U | 380 | 1100 | ug/Kg |
| 105-67-9 | 2,4-Dimethylphenol | 1100 | U | 420 | 1100 | ug/Kg |
| 111-91-1 | bis(2-Chloroethoxy)methane | 1100 | U | 200 | 1100 | ug/Kg |
| 120-83-2 | 2,4-Dichlorophenol | 1100 | U | 180 | 1100 | ug/Kg |
| 91-20-3 | Naphthalene | 1100 | U | 150 | 1100 | ug/Kg |
| 106-47-8 | 4-Chloroaniline | 1100 | U | 230 | 1100 | ug/Kg |
| 87-68-3 | Hexachlorobutadiene | 1100 | U | 160 | 1100 | ug/Kg |
| 105-60-2 | Caprolactam | 2100 | U | 340 | 2100 | ug/Kg |
| 59-50-7 | 4-Chloro-3-methylphenol | 1100 | U | 190 | 1100 | ug/Kg |
| 91-57-6 | 2-Methylnaphthalene | 1100 | U | 170 | 1100 | ug/Kg |
| 77-47-4 | Hexachlorocyclopentadiene | 2100 | U | 750 | 2100 | ug/Kg |
| 88-06-2 | 2,4,6-Trichlorophenol | 1100 | U | 130 | 1100 | ug/Kg |
| 95-95-4 | 2,4,5-Trichlorophenol | 1100 | U | 190 | 1100 | ug/Kg |
| 92-52-4 | 1,1-Biphenyl | 1100 | U | 140 | 1100 | ug/Kg |
| 91-58-7 | 2-Chloronaphthalene | 1100 | U | 150 | 1100 | ug/Kg |
| 88-74-4 | 2-Nitroaniline | 1100 | U | 310 | 1100 | ug/Kg |
| 131-11-3 | Dimethylphthalate | 1100 | U | 180 | 1100 | ug/Kg |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | 07/11/25 |
| Project: | Con Edison - East River Site 2 | Date Received: | 07/11/25 |
| Client Sample ID: | WC-SOIL-20250711 | SDG No.: | Q2592 |
| Lab Sample ID: | Q2592-01 | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 76.7 |
| Sample Wt/Vol: | 30.07 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BP025166.D | 5 | 07/15/25 09:05 | 07/16/25 20:22 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| 208-96-8 | Acenaphthylene | 1100 | U | 190 | 1100 | ug/Kg |
| 606-20-2 | 2,6-Dinitrotoluene | 1100 | U | 220 | 1100 | ug/Kg |
| 99-09-2 | 3-Nitroaniline | 1100 | U | 300 | 1100 | ug/Kg |
| 83-32-9 | Acenaphthene | 1100 | U | 140 | 1100 | ug/Kg |
| 51-28-5 | 2,4-Dinitrophenol | 2100 | U | 1500 | 2100 | ug/Kg |
| 100-02-7 | 4-Nitrophenol | 2100 | U | 700 | 2100 | ug/Kg |
| 132-64-9 | Dibenzofuran | 1100 | U | 150 | 1100 | ug/Kg |
| 121-14-2 | 2,4-Dinitrotoluene | 1100 | U | 330 | 1100 | ug/Kg |
| 84-66-2 | Diethylphthalate | 1100 | U | 180 | 1100 | ug/Kg |
| 7005-72-3 | 4-Chlorophenyl-phenylether | 1100 | U | 170 | 1100 | ug/Kg |
| 86-73-7 | Fluorene | 1100 | U | 160 | 1100 | ug/Kg |
| 100-01-6 | 4-Nitroaniline | 1100 | UQ | 420 | 1100 | ug/Kg |
| 534-52-1 | 4,6-Dinitro-2-methylphenol | 2100 | U | 670 | 2100 | ug/Kg |
| 86-30-6 | n-Nitrosodiphenylamine | 1100 | U | 210 | 1100 | ug/Kg |
| 101-55-3 | 4-Bromophenyl-phenylether | 1100 | U | 180 | 1100 | ug/Kg |
| 118-74-1 | Hexachlorobenzene | 1100 | U | 160 | 1100 | ug/Kg |
| 1912-24-9 | Atrazine | 1100 | U | 220 | 1100 | ug/Kg |
| 87-86-5 | Pentachlorophenol | 2100 | U | 330 | 2100 | ug/Kg |
| 85-01-8 | Phenanthrene | 1100 | U | 140 | 1100 | ug/Kg |
| 120-12-7 | Anthracene | 1100 | U | 220 | 1100 | ug/Kg |
| 86-74-8 | Carbazole | 1100 | U | 200 | 1100 | ug/Kg |
| 84-74-2 | Di-n-butylphthalate | 1100 | U | 310 | 1100 | ug/Kg |
| 206-44-0 | Fluoranthene | 760 | J | 200 | 1100 | ug/Kg |
| 129-00-0 | Pyrene | 690 | J | 230 | 1100 | ug/Kg |
| 85-68-7 | Butylbenzylphthalate | 860 | J | 460 | 1100 | ug/Kg |
| 91-94-1 | 3,3-Dichlorobenzidine | 2100 | U | 240 | 2100 | ug/Kg |
| 56-55-3 | Benzo(a)anthracene | 490 | J | 150 | 1100 | ug/Kg |
| 218-01-9 | Chrysene | 570 | J | 130 | 1100 | ug/Kg |
| 117-81-7 | Bis(2-ethylhexyl)phthalate | 2000 | | 390 | 1100 | ug/Kg |
| 117-84-0 | Di-n-octyl phthalate | 2100 | U | 560 | 2100 | ug/Kg |
| 205-99-2 | Benzo(b)fluoranthene | 810 | J | 120 | 1100 | ug/Kg |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | 07/11/25 |
| Project: | Con Edison - East River Site 2 | Date Received: | 07/11/25 |
| Client Sample ID: | WC-SOIL-20250711 | SDG No.: | Q2592 |
| Lab Sample ID: | Q2592-01 | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 76.7 |
| Sample Wt/Vol: | 30.07 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BP025166.D | 5 | 07/15/25 09:05 | 07/16/25 20:22 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| 207-08-9 | Benzo(k)fluoranthene | 1100 | U | 150 | 1100 | ug/Kg |
| 50-32-8 | Benzo(a)pyrene | 620 | J | 190 | 1100 | ug/Kg |
| 193-39-5 | Indeno(1,2,3-cd)pyrene | 470 | J | 190 | 1100 | ug/Kg |
| 53-70-3 | Dibenzo(a,h)anthracene | 1100 | U | 180 | 1100 | ug/Kg |
| 191-24-2 | Benzo(g,h,i)perylene | 640 | J | 170 | 1100 | ug/Kg |
| 95-94-3 | 1,2,4,5-Tetrachlorobenzene | 1100 | U | 170 | 1100 | ug/Kg |
| 123-91-1 | 1,4-Dioxane | 1100 | U | 290 | 1100 | ug/Kg |
| 58-90-2 | 2,3,4,6-Tetrachlorophenol | 1100 | U | 180 | 1100 | ug/Kg |
| **SURROGATES** | | | | | | |
| 367-12-4 | 2-Fluorophenol | 70.8 | | 18 - 112 | 47% | SPK: 150 |
| 13127-88-3 | Phenol-d6 | 71.5 | | 15 - 107 | 48% | SPK: 150 |
| 4165-60-0 | Nitrobenzene-d5 | 53.2 | | 18 - 107 | 53% | SPK: 100 |
| 321-60-8 | 2-Fluorobiphenyl | 55.4 | | 20 - 109 | 55% | SPK: 100 |
| 118-79-6 | 2,4,6-Tribromophenol | 97.1 | | 10 - 116 | 65% | SPK: 150 |
| 1718-51-0 | Terphenyl-d14 | 58.2 | | 10 - 105 | 58% | SPK: 100 |
| **INTERNAL STANDARDS** | | | | | | |
| 3855-82-1 | 1,4-Dichlorobenzene-d4 | 569000 | 7.443 | | | |
| 1146-65-2 | Naphthalene-d8 | 2220000 | 10.19 | | | |
| 15067-26-2 | Acenaphthene-d10 | 1400000 | 14.101 | | | |
| 1517-22-2 | Phenanthrene-d10 | 2850000 | 16.919 | | | |
| 1719-03-5 | Chrysene-d12 | 3030000 | 21.348 | | | |
| 1520-96-3 | Perylene-d12 | 3540000 | 24.495 | | | |
| **TENTATIVE IDENTIFIED COMPOUNDS** | | | | | | |
| 65-85-0 | Benzoic acid | 1400 | J | | 9.46 | ug/Kg |
| 000090-04-0 | Benzenamine, 2-methoxy- | 680 | J | | 9.91 | ug/Kg |
| 000700-87-8 | Benzene, 1-isocyanato-2-methoxy- | 470 | J | | 10.4 | ug/Kg |
| 000119-61-9 | Benzophenone | 510 | J | | 15.5 | ug/Kg |
| | unknown16.113 | 460 | J | | 16.1 | ug/Kg |
| 000057-10-3 | n-Hexadecanoic acid | 980 | J | | 17.9 | ug/Kg |
| 000057-11-4 | Octadecanoic acid | 770 | J | | 19.3 | ug/Kg |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | 07/11/25 |
| Project: | Con Edison - East River Site 2 | Date Received: | 07/11/25 |
| Client Sample ID: | WC-SOIL-20250711 | SDG No.: | Q2592 |
| Lab Sample ID: | Q2592-01 | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 76.7 |
| Sample Wt/Vol: | 30.07 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BP025166.D | 5 | 07/15/25 09:05 | 07/16/25 20:22 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| 000080-05-7 | Phenol, 4,4-(1-methylethylidene)b | 1500 | J | | 19.5 | ug/Kg |

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025166.D
Acq On : 16 Jul 2025 20:22
Operator : CG/JU
Sample : Q2592-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025

Quant Time: Jul 17 01:10:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| Internal Standards | | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.443 | 152 | 569178 | 20.000 | ng | | 0.00 |
| 21) Naphthalene-d8 | 10.190 | 136 | 2223517 | 20.000 | ng | | 0.00 |
| 39) Acenaphthene-d10 | 14.101 | 164 | 1404763 | 20.000 | ng | | 0.01 |
| 64) Phenanthrene-d10 | 16.919 | 188 | 2849474 | 20.000 | ng | | 0.01 |
| 76) Chrysene-d12 | 21.348 | 240 | 3032748 | 20.000 | ng | | 0.00 |
| 86) Perylene-d12 | 24.495 | 264 | 3540161 | 20.000 | ng | | 0.00 |
| System Monitoring Compounds | | | | | | | |
| 5) 2-Fluorophenol | 5.102 | 112 | 454833 | 14.159 | ng | | 0.00 |
| 7) Phenol-d6 | 6.643 | 99 | 599229 | 14.294 | ng | | 0.01 |
| 23) Nitrobenzene-d5 | 8.572 | 82 | 356018 | 10.648 | ng | | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.625 | 330 | 322413 | 19.428 | ng | | 0.00 |
| 45) 2-Fluorobiphenyl | 12.689 | 172 | 1004357 | 11.076 | ng | | 0.00 |
| 79) Terphenyl-d14 | 19.683 | 244 | 1605339 | 11.634 | ng | | 0.01 |
| Target Compounds | | | | | | | Qvalue |
| 32) Benzoic acid | 9.460 | 122 | 7225 | 6.474 | ng | | 84 |
| 75) Fluoranthene | 19.066 | 202 | 627897 | 3.528 | ng | | 98 |
| 78) Pyrene | 19.436 | 202 | 594146 | 3.163 | ng | # | 95 |
| 80) Butylbenzylphthalate | 20.430 | 149 | 21902 | 3.961 | ng | # | 83 |
| 81) Benzo(a)anthracene | 21.330 | 228 | 415157 | 2.259 | ng | | 91 |
| 83) Chrysene | 21.389 | 228 | 461571m | 2.610 | ng | | |
| 84) Bis(2-ethylhexyl)phtha... | 21.324 | 149 | 891651 | 9.045 | ng | # | 98 |
| 87) Indeno(1,2,3-cd)pyrene | 28.030 | 276 | 491763 | 2.180 | ng | # | 93 |
| 88) Benzo(b)fluoranthene | 23.512 | 252 | 772310 | 3.748 | ng | # | 89 |
| 90) Benzo(a)pyrene | 24.348 | 252 | 504238 | 2.875 | ng | # | 82 |
| 92) Benzo(g,h,i)perylene | 29.100 | 276 | 571607 | 2.968 | ng | | 94 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025166.D
Acq On : 16 Jul 2025 20:22
Operator : CG/JU
Sample : Q2592-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Quant Time: Jul 17 01:10:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.443 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:152 Resp: 56917

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 153.6 | 126.1 | 189.1 |
| 115 | 54.7 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#5
2-Fluorophenol
Concen: 14.159 ng
RT: 5.102 min Scan# 368
Delta R.T. 0.006 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:112 Resp: 454833

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 51.5 | 42.0 | 63.0 |
| 63 | 26.6 | 22.2 | 33.2 |






#7
Phenol-d6
Concen: 14.294 ng
RT: 6.643 min Scan# 6
Delta R.T. 0.012 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion: 99 Resp: 59922

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 16.3 | 13.6 | 20.4 |
| 71 | 30.0 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.190 min Scan# 1233
Delta R.T. 0.006 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:136 Resp: 2223517

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 11.0 | 8.7 | 13.1 |
| 54 | 6.7 | 5.4 | 8.2 |
| 68 | 5.5 | 4.5 | 6.7 |






#23
Nitrobenzene-d5
Concen: 10.648 ng
RT: 8.572 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion: 82 Resp: 35601

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 50.4 | 38.4 | 57.6 |
| 54 | 45.2 | 37.8 | 56.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#32
Benzoic acid
Concen: 6.474 ng
RT: 9.460 min Scan# 1109
Delta R.T. -0.041 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:122 Resp: 7225

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 136.8 | 101.0 | 141.0 |
| 77 | 104.3 | 67.6 | 107.6 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.101 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:164 Resp: 140476

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 100.2 | 80.3 | 120.5 |
| 160 | 44.7 | 35.9 | 53.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#42
2,4,6-Tribromophenol
Concen: 19.428 ng
RT: 15.625 min Scan# 2157
Delta R.T. 0.006 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:330 Resp: 322413

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 95.9 | 77.7 | 116.5 |
| 141 | 27.6 | 26.0 | 39.0 |






#45
2-Fluorobiphenyl
Concen: 11.076 ng
RT: 12.689 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:172 Resp: 100435

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.5 | 28.2 | 42.2 |
| 170 | 23.2 | 18.7 | 28.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.919 min Scan# 2377
Delta R.T. 0.012 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:188 Resp: 2849474

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 8.7 | 7.3 | 10.9 |
| 80 | 8.9 | 7.8 | 11.6 |






#75
Fluoranthene
Concen: 3.528 ng
RT: 19.066 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:202 Resp: 62789

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 11.7 | 0.0 | 30.4 |
| 203 | 18.2 | 0.0 | 37.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.348 min Scan# 3130
Delta R.T. 0.006 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:240 Resp: 3032748

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.1 | 9.0 | 13.4 |
| 236 | 25.8 | 20.3 | 30.5 |






#78
Pyrene
Concen: 3.163 ng
RT: 19.436 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:202 Resp: 59414

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 22.1 | 17.1 | 25.7 |
| 203 | 21.9 | 14.1 | 21.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#79
Terphenyl-d14
Concen: 11.634 ng
RT: 19.683 min Scan# 2847
Delta R.T. 0.012 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:244 Resp: 1605339

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.6 | 5.9 | 8.9 |
| 122 | 10.1 | 9.4 | 14.2 |






#80
Butylbenzylphthalate
Concen: 3.961 ng
RT: 20.430 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:149 Resp: 2190

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 80.8 | 50.6 | 76.0 |
| 206 | 20.9 | 17.2 | 25.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#81
Benzo(a)anthracene
Concen: 2.259 ng
RT: 21.330 min Scan# 3127
Delta R.T. 0.006 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:228 Resp: 415157

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 33.1 | 22.1 | 33.1 |
| 229 | 23.1 | 16.0 | 24.0 |






#83
Chrysene
Concen: 2.610 ng m
RT: 21.389 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:228 Resp: 46157

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 31.1 | 24.4 | 36.6 |
| 229 | 25.2 | 16.0 | 24.0# |




**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 9.045 ng
RT: 21.324 min Scan# 3126
Delta R.T. -0.006 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:149 Resp: 891651

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 28.5 | 21.9 | 32.9 |
| 279 | 5.4 | 3.5 | 5.3# |






#86
Perylene-d12
Concen: 20.000 ng
RT: 24.495 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:264 Resp: 354016

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.8 | 18.6 | 28.0 |
| 265 | 22.3 | 17.7 | 26.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#87
Indeno(1,2,3-cd)pyrene
Concen: 2.180 ng
RT: 28.030 min Scan# 4266
Delta R.T. -0.006 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:276 Resp: 491763

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 17.5 | 13.0 | 19.6 |
| 277 | 24.4 | 15.9 | 23.9# |






#88
Benzo(b)fluoranthene
Concen: 3.748 ng
RT: 23.512 min Scan# 34
Delta R.T. -0.000 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:252 Resp: 77231(

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 24.0 | 17.7 | 26.5 |
| 125 | 19.7 | 8.1 | 12.1# |




**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#90
Benzo(a)pyrene
Concen: 2.875 ng
RT: 24.348 min Scan# 3640
Delta R.T. -0.006 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:252 Resp: 504238

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 26.1 | 17.6 | 26.4 |
| 125 | 25.4 | 9.0 | 13.4# |






#92
Benzo(g,h,i)perylene
Concen: 2.968 ng
RT: 29.100 min Scan# 4
Delta R.T. -0.024 min
Lab File: BP025166.D
Acq: 16 Jul 2025 20:22

Tgt Ion:276 Resp: 57160

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 26.4 | 19.3 | 28.9 |
| 138 | 24.7 | 16.9 | 25.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025166.D
Acq On : 16 Jul 2025 20:22
Operator : CG/JU
Sample : Q2592-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

Integration Parameters: rteint.p
Integrator: RTE
Smoothing : ON Filtering: 5
Sampling : 1 Min Area: 3 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC: BP025166.D\data.ms

| peak # | R.T. min | first scan | max scan | last scan | PK TY | peak height | corr. area | corr. % max. | % of total |
|---|---|---|---|---|---|---|---|---|---|
| 1 | 4.660 | 288 | 293 | 299 | rVB | 96309 | 137782 | 1.37% | 0.220% |
| 2 | 5.102 | 362 | 368 | 376 | rVB | 1020201 | 1427318 | 14.18% | 2.281% |
| 3 | 6.643 | 624 | 630 | 639 | rVB | 1026560 | 1579594 | 15.70% | 2.524% |
| 4 | 7.443 | 759 | 766 | 773 | rBV | 1999225 | 3039986 | 30.21% | 4.858% |
| 5 | 8.572 | 952 | 958 | 965 | rVB | 639101 | 1006265 | 10.00% | 1.608% |
| 6 | 9.907 | 1179 | 1185 | 1198 | rBV | 279463 | 667547 | 6.63% | 1.067% |
| 7 | 10.190 | 1225 | 1233 | 1240 | rBV | 2569141 | 4259804 | 42.33% | 6.807% |
| 8 | 10.384 | 1260 | 1266 | 1272 | rBV | 270350 | 463328 | 4.60% | 0.740% |
| 9 | 11.654 | 1477 | 1482 | 1488 | rBV | 228408 | 373689 | 3.71% | 0.597% |
| 10 | 12.689 | 1652 | 1658 | 1665 | rVB | 1947313 | 2746429 | 27.29% | 4.389% |
| 11 | 13.554 | 1798 | 1805 | 1813 | rBV4 | 36022 | 110114 | 1.09% | 0.176% |
| 12 | 14.101 | 1891 | 1898 | 1906 | rBV | 3967708 | 5789402 | 57.53% | 9.251% |
| 13 | 14.960 | 2040 | 2044 | 2050 | rBV | 231865 | 291783 | 2.90% | 0.466% |
| 14 | 15.501 | 2131 | 2136 | 2144 | rBV2 | 386081 | 678189 | 6.74% | 1.084% |
| 15 | 15.625 | 2152 | 2157 | 2163 | rBV | 1444315 | 2200682 | 21.87% | 3.517% |
| 16 | 15.725 | 2171 | 2174 | 2183 | rVB3 | 120691 | 215582 | 2.14% | 0.344% |
| 17 | 16.078 | 2229 | 2234 | 2236 | rBV3 | 169954 | 289770 | 2.88% | 0.463% |
| 18 | 16.113 | 2236 | 2240 | 2248 | rVB3 | 402352 | 712956 | 7.08% | 1.139% |
| 19 | 16.719 | 2339 | 2343 | 2348 | rBV4 | 228871 | 374303 | 3.72% | 0.598% |
| 20 | 16.919 | 2371 | 2377 | 2382 | rBV2 | 4146580 | 6659122 | 66.17% | 10.641% |
| 21 | 16.966 | 2382 | 2385 | 2390 | rVB | 314451 | 410628 | 4.08% | 0.656% |
| 22 | 17.901 | 2540 | 2544 | 2552 | rVB | 1124602 | 1500609 | 14.91% | 2.398% |
| 23 | 19.066 | 2738 | 2742 | 2746 | rBV | 1100318 | 1790737 | 17.79% | 2.862% |
| 24 | 19.254 | 2771 | 2774 | 2781 | rVB | 1328262 | 1781504 | 17.70% | 2.847% |
| 25 | 19.495 | 2811 | 2815 | 2821 | rBV | 2076482 | 3381618 | 33.60% | 5.404% |
| 26 | 19.683 | 2843 | 2847 | 2852 | rVB | 3401284 | 4312062 | 42.85% | 6.890% |
| 27 | 21.348 | 3123 | 3130 | 3135 | rVB2 | 5378542 | 10063901 | 100.00% | 16.082% |
| 28 | 24.495 | 3662 | 3665 | 3674 | rVB | 3269594 | 6315321 | 62.75% | 10.092% |

Sum of corrected areas: 62580025

Sample Results: **BP025166.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025166.D
Acq On : 16 Jul 2025 20:22
Operator : CG/JU
Sample : Q2592-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025166.D
Acq On : 16 Jul 2025 20:22
Operator : CG/JU
Sample : Q2592-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

**********
Peak Number 1 Benzenamine, 2-methoxy- Concentration Rank 4

| R.T. | EstConc | Area | Relative to ISTD | R.T. |
|---|---|---|---|---|
| 9.907 | 3.13 ng | 667547 | Naphthalene-d8 | 10.190 |

| Hit# of 5 | Tentative ID | MW | MolForm | CAS# | Qual |
|---|---|---|---|---|---|
| 1 | Benzenamine, 2-methoxy- | 123 | C7H9NO | 000090-04-0 | 96 |
| 2 | Pyrazinamide | 123 | C5H5N3O | 000098-96-4 | 74 |
| 3 | Bicyclo[2.2.2]oct-2-ene | 108 | C8H12 | 000931-64-6 | 64 |
| 4 | 1,3-Phenylenediamine | 108 | C6H8N2 | 000108-45-2 | 64 |
| 5 | Pyrazine | 80 | C4H4N2 | 000290-37-9 | 50 |






Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025166.D
Acq On : 16 Jul 2025 20:22
Operator : CG/JU
Sample : Q2592-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

**********
Peak Number 2 Benzene, 1-isocyanato-2-met... Concentration Rank 6

| R.T. | EstConc | Area | Relative to ISTD | R.T. |
|---|---|---|---|---|
| 10.384 | 2.18 ng | 463328 | Naphthalene-d8 | 10.190 |

| Hit# of 5 | Tentative ID | MW | MolForm | CAS# | Qual |
|---|---|---|---|---|---|
| 1 | Benzene, 1-isocyanato-2-methoxy- | 149 | C8H7NO2 | 000700-87-8 | 94 |
| 2 | Benzene, 1-isocyanato-4-methoxy- | 149 | C8H7NO2 | 005416-93-3 | 49 |
| 3 | 1H-Benzotriazole, 4-methoxy- | 149 | C7H7N3O | 027799-90-2 | 47 |
| 4 | 3',4'-Formoxylidide | 149 | C9H11NO | 006639-60-7 | 47 |
| 5 | 1H-Benzotriazole, 5-methoxy- | 149 | C7H7N3O | 027799-91-3 | 46 |






Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025166.D
Acq On : 16 Jul 2025 20:22
Operator : CG/JU
Sample : Q2592-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

**********

Peak Number 3 Benzophenone Concentration Rank 5

| R.T. | EstConc | Area | Relative to ISTD | R.T. |
|---|---|---|---|---|
| 15.501 | 2.34 ng | 678189 | Acenaphthene-d10 | 14.101 |

| Hit# of 5 | Tentative ID | MW | MolForm | CAS# | Qual |
|---|---|---|---|---|---|
| 1 | Benzophenone | 182 | C13H10O | 000119-61-9 | 86 |
| 2 | Acetamide, N-(2-chlorobenzofuran... | 209 | C10H8ClNO2 | 067382-11-0 | 27 |
| 3 | 2-N-Diethylmelamine | 182 | C7H14N6 | 002073-31-6 | 22 |
| 4 | 7-Phosphabicyclo[2.2.1]hept-2-en... | 182 | C11H19P | 1000151-83-5 | 22 |
| 5 | Sydnone, 4-chloro-3-phenyl- | 196 | C8H5ClN2O2 | 005226-93-7 | 14 |






Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025166.D
Acq On : 16 Jul 2025 20:22
Operator : CG/JU
Sample : Q2592-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

**********

Peak Number 4 unknown16.113 Concentration Rank 7

| R.T. | EstConc | Area | Relative to ISTD | R.T. |
|---|---|---|---|---|
| 16.113 | 2.14 ng | 712956 | Phenanthrene-d10 | 16.919 |

| Hit# of 5 | Tentative ID | MW | MolForm | CAS# | Qual |
|---|---|---|---|---|---|
| 1 | Oxyhydrostinine | 205 | C11H11NO3 | 000552-29-4 | 45 |
| 2 | 2-(Piperidin-1-yl)pyrimidine | 163 | C9H13N3 | 051957-36-9 | 38 |
| 3 | N-(4-Formyl-phenyl)-N-methyl-for... | 163 | C9H9NO2 | 079213-80-2 | 30 |
| 4 | 1-Amino-4,6-dimethyl-2-oxo-1,2-d... | 163 | C8H9N3O | 001562-12-5 | 30 |
| 5 | 2H-1-Benzopyran-2-one, 3,4-dihyd... | 162 | C10H10O2 | 000092-47-7 | 30 |






Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025166.D
Acq On : 16 Jul 2025 20:22
Operator : CG/JU
Sample : Q2592-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

**********

Peak Number 5 n-Hexadecanoic acid Concentration Rank 2

| R.T. | EstConc | Area | Relative to ISTD | R.T. |
|---|---|---|---|---|
| 17.901 | 4.51 ng | 1500610 | Phenanthrene-d10 | 16.919 |

| Hit# of 5 | Tentative ID | MW | MolForm | CAS# | Qual |
|---|---|---|---|---|---|
| 1 | n-Hexadecanoic acid | 256 | C16H32O2 | 000057-10-3 | 99 |
| 2 | Tetradecanoic acid | 228 | C14H28O2 | 000544-63-8 | 91 |
| 3 | Tridecanoic acid | 214 | C13H26O2 | 000638-53-9 | 89 |
| 4 | Pentadecanoic acid | 242 | C15H30O2 | 001002-84-2 | 87 |
| 5 | Octadecanoic acid | 284 | C18H36O2 | 000057-11-4 | 83 |






Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025166.D
Acq On : 16 Jul 2025 20:22
Operator : CG/JU
Sample : Q2592-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

**********

Peak Number 6 Octadecanoic acid Concentration Rank 3

| R.T. | EstConc | Area | Relative to ISTD | R.T. |
|---|---|---|---|---|
| 19.254 | 3.54 ng | 1781500 | Chrysene-d12 | 21.348 |

| Hit# of 5 | Tentative ID | MW | MolForm | CAS# | Qual |
|---|---|---|---|---|---|
| 1 | Octadecanoic acid | 284 | C18H36O2 | 000057-11-4 | 98 |
| 2 | Hexacosanoic acid | 396 | C26H52O2 | 000506-46-7 | 72 |
| 3 | Pentadecanoic acid | 242 | C15H30O2 | 001002-84-2 | 64 |
| 4 | Tetradecanoic acid | 228 | C14H28O2 | 000544-63-8 | 60 |
| 5 | Ethanol, 2-(tetradecyloxy)- | 258 | C16H34O2 | 002136-70-1 | 41 |






Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025166.D
Acq On : 16 Jul 2025 20:22
Operator : CG/JU
Sample : Q2592-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

**********
Peak Number 7 Phenol, 4,4'-(1-methylethyl... Concentration Rank 1

| R.T. | EstConc | Area | Relative to ISTD | R.T. |
|---|---|---|---|---|
| 19.495 | 6.72 ng | 3381620 | Chrysene-d12 | 21.348 |

| Hit# of 5 | Tentative ID | MW | MolForm | CAS# | Qual |
|---|---|---|---|---|---|
| 1 | Phenol, 4,4'-(1-methylethylidene... | 228 | C15H16O2 | 000080-05-7 | 97 |
| 2 | 4,4'-(1,3-Dimethylbutylidene)bis... | 270 | C18H22O2 | 006807-17-6 | 83 |
| 3 | 2,2-Bis-(4-hyxroxyphenyl)-butane | 242 | C16H18O2 | 000077-40-7 | 72 |
| 4 | 2,5-bis(Trimethylsilyl)thiophene | 228 | C10H20SSi2 | 017906-71-7 | 59 |
| 5 | 4-Hydroxy-.gamma.-(4-hydroxyphen... | 300 | C18H20O4 | 007297-85-0 | 59 |






Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025166.D
Acq On : 16 Jul 2025 20:22
Operator : CG/JU
Sample : Q2592-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
WC-SOIL-20250711

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

| | | | | | --Internal Standard--- | | | |
|---|---|---|---|---|---|---|---|---|
| TIC Top Hit name | RT | EstConc | Units | Response | # | RT | Resp | Conc |
| Benzenamine, 2-... | 9.907 | 3.1 | ng | 667547 | 2 | 10.190 | 4259800 | 20.0 |
| Benzene, 1-isoc... | 10.384 | 2.2 | ng | 463328 | 2 | 10.190 | 4259800 | 20.0 |
| Benzophenone | 15.501 | 2.3 | ng | 678189 | 3 | 14.101 | 5789400 | 20.0 |
| unknown16.113 | 16.113 | 2.1 | ng | 712956 | 4 | 16.919 | 6659120 | 20.0 |
| n-Hexadecanoic ... | 17.901 | 4.5 | ng | 1500610 | 4 | 16.919 | 6659120 | 20.0 |
| Octadecanoic acid | 19.254 | 3.5 | ng | 1781500 | 5 | 21.348 | 10063900 | 20.0 |
| Phenol, 4,4'-(1... | 19.495 | 6.7 | ng | 3381620 | 5 | 21.348 | 10063900 | 20.0 |

# CALIBRATION SUMMARY

Response Factor Report BNA_F

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF071525.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Jul 15 17:53:25 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BF143093.D 5 =BF143094.D 10 =BF143095.D 20 =BF143096.D 40 =BF143097.D 50 =BF143098.D 60 =BF143099.D 80 =BF143100.D

| | | Compound | 2.5 | 5 | 10 | 20 | 40 | 50 | 60 | 80 | Avg | %RSD |
|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1) | I | 1,4-Dichlorobenzen... | ----------ISTD---------- | | | | | | | | | |
| 2) | | 1,4-Dioxane | | 0.637 | 0.635 | 0.640 | 0.627 | 0.661 | 0.626 | 0.586 | 0.630 | 3.58 |
| 3) | | Pyridine | | 1.606 | 1.541 | 1.616 | 1.594 | 1.681 | 1.612 | 1.519 | 1.596 | 3.33 |
| 4) | | n-Nitrosodimet... | | | 0.788 | 0.825 | 0.816 | 0.879 | 0.840 | 0.812 | 0.827 | 3.70 |
| 5) | S | 2-Fluorophenol | | 1.334 | 1.272 | 1.315 | 1.255 | 1.305 | 1.232 | 1.148 | 1.266 | 4.98 |
| 6) | | Aniline | | 2.330 | 2.254 | 2.280 | 2.205 | 2.351 | 2.196 | 2.069 | 2.241 | 4.26 |
| 7) | S | Phenol-d6 | | 1.685 | 1.616 | 1.640 | 1.565 | 1.643 | 1.547 | 1.441 | 1.591 | 5.11 |
| 8) | | 2-Chlorophenol | | 1.239 | 1.208 | 1.321 | 1.273 | 1.363 | 1.286 | 1.212 | 1.272 | 4.51 |
| 9) | | Benzaldehyde | | | 1.196 | 1.149 | 1.451 | 1.452 | 1.163 | 0.864 | 1.212 | 18.15 |
| 10) | C | Phenol | | 1.766 | 1.742 | 1.782 | 1.709 | 1.780 | 1.694 | 1.584 | 1.723 | 4.07 |
| 11) | | bis(2-Chloroet... | | 1.425 | 1.363 | 1.366 | 1.309 | 1.371 | 1.291 | 1.209 | 1.334 | 5.26 |
| 12) | | 1,3-Dichlorobe... | | 1.547 | 1.482 | 1.470 | 1.414 | 1.471 | 1.388 | 1.286 | 1.437 | 5.83 |
| 13) | C | 1,4-Dichlorobe... | | 1.570 | 1.473 | 1.504 | 1.412 | 1.482 | 1.386 | 1.292 | 1.446 | 6.27 |
| 14) | | 1,2-Dichlorobe... | | 1.468 | 1.409 | 1.430 | 1.364 | 1.417 | 1.321 | 1.241 | 1.379 | 5.57 |
| 15) | | Benzyl Alcohol | | | 1.126 | 1.204 | 1.167 | 1.255 | 1.189 | 1.132 | 1.179 | 4.09 |
| 16) | | 2,2'-oxybis(1-... | | 2.713 | 2.532 | 2.528 | 2.405 | 2.517 | 2.342 | 2.171 | 2.458 | 6.99 |
| 17) | | 2-Methylphenol | | 1.132 | 1.081 | 1.124 | 1.083 | 1.150 | 1.077 | 1.035 | 1.098 | 3.63 |
| 18) | | Hexachloroethane | | 0.465 | 0.440 | 0.486 | 0.472 | 0.500 | 0.476 | 0.455 | 0.471 | 4.22 |
| 19) | P | n-Nitroso-di-n... | 0.956 | 1.029 | 0.988 | 1.002 | 0.964 | 1.017 | 0.945 | 0.897 | 0.975 | 4.44 |
| 20) | | 3+4-Methylphenols | | | 1.399 | 1.419 | 1.356 | 1.421 | 1.293 | 1.189 | 1.346 | 6.74 |
| 21) | I | Naphthalene-d8 | ----------ISTD---------- | | | | | | | | | |
| 22) | | Acetophenone | | 0.538 | 0.516 | 0.500 | 0.469 | 0.482 | 0.451 | 0.419 | 0.482 | 8.33 |
| 23) | S | Nitrobenzene-d5 | | 0.283 | 0.319 | 0.363 | 0.372 | 0.399 | 0.390 | 0.370 | 0.357 | 11.59 |
| 24) | | Nitrobenzene | | 0.282 | 0.317 | 0.357 | 0.356 | 0.385 | 0.366 | 0.347 | 0.344 | 9.94 |
| 25) | | Isophorone | | 0.705 | 0.688 | 0.697 | 0.678 | 0.718 | 0.687 | 0.651 | 0.689 | 3.11 |
| 26) | C | 2-Nitrophenol | | 0.063 | 0.075 | 0.106 | 0.120 | 0.141 | 0.146 | 0.147 | 0.114 | 29.95 |
| 27) | | 2,4-Dimethylph... | | 0.325 | 0.322 | 0.326 | 0.320 | 0.332 | 0.315 | 0.299 | 0.320 | 3.38 |
| 28) | | bis(2-Chloroet... | | 0.444 | 0.434 | 0.435 | 0.409 | 0.424 | 0.404 | 0.375 | 0.418 | 5.63 |
| 29) | C | 2,4-Dichloroph... | | 0.237 | 0.254 | 0.267 | 0.267 | 0.280 | 0.271 | 0.254 | 0.262 | 5.43 |
| 30) | | 1,2,4-Trichlor... | | 0.316 | 0.305 | 0.307 | 0.295 | 0.303 | 0.287 | 0.267 | 0.297 | 5.40 |
| 31) | | Naphthalene | | 1.099 | 1.048 | 1.037 | 0.966 | 0.987 | 0.937 | 0.878 | 0.993 | 7.52 |
| 32) | | Benzoic acid | | | 0.080 | 0.111 | 0.144 | 0.159 | 0.171 | 0.172 | 0.139 | 26.27 |
| 33) | | 4-Chloroaniline | | 0.420 | 0.420 | 0.414 | 0.398 | 0.409 | 0.381 | 0.353 | 0.399 | 6.12 |
| 34) | C | Hexachlorobuta... | | 0.185 | 0.176 | 0.183 | 0.177 | 0.182 | 0.176 | 0.163 | 0.178 | 4.10 |
| 35) | | Caprolactam | | | 0.073 | 0.078 | 0.083 | 0.087 | 0.086 | 0.083 | 0.082 | 6.36 |
| 36) | C | 4-Chloro-3-met... | | 0.285 | 0.290 | 0.296 | 0.293 | 0.307 | 0.293 | 0.277 | 0.291 | 3.26 |
| 37) | | 2-Methylnaphth... | | 0.644 | 0.626 | 0.615 | 0.580 | 0.597 | 0.562 | 0.524 | 0.592 | 6.91 |
| 38) | | 1-Methylnaphth... | | 0.678 | 0.660 | 0.631 | 0.603 | 0.615 | 0.578 | 0.537 | 0.615 | 7.81 |

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF071525.M

| | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|
| 39) | I | Acenaphthene-d10 | ----------ISTD---------- | | | | | | | | |
| 40) | | 1,2,4,5-Tetrac... | 0.616 | 0.600 | 0.602 | 0.578 | 0.601 | 0.561 | 0.522 | 0.583 | 5.56 |
| 41) | P | Hexachlorocycl... | | 0.254 | 0.306 | 0.323 | 0.356 | 0.356 | 0.344 | 0.323 | 12.16 |
| 42) | S | 2,4,6-Tribromo... | 0.136 | 0.157 | 0.181 | 0.188 | 0.204 | 0.194 | 0.188 | 0.178 | 13.23 |
| 43) | C | 2,4,6-Trichlor... | 0.300 | 0.312 | 0.377 | 0.376 | 0.400 | 0.393 | 0.377 | 0.362 | 10.94 |
| 44) | | 2,4,5-Trichlor... | 0.336 | 0.367 | 0.390 | 0.393 | 0.418 | 0.377 | 0.378 | 0.380 | 6.62 |
| 45) | S | 2-Fluorobiphenyl | 1.777 | 1.645 | 1.602 | 1.456 | 1.460 | 1.355 | 1.238 | 1.505 | 12.18 |
| 46) | | 1,1'-Biphenyl | 1.750 | 1.687 | 1.671 | 1.567 | 1.610 | 1.496 | 1.384 | 1.595 | 7.84 |
| 47) | | 2-Chloronaphth... | 1.271 | 1.203 | 1.206 | 1.159 | 1.195 | 1.120 | 1.048 | 1.172 | 6.10 |
| 48) | | 2-Nitroaniline | 0.180 | 0.224 | 0.296 | 0.328 | 0.372 | 0.363 | 0.355 | 0.303 | 24.51 |
| 49) | | Acenaphthylene | 2.060 | 2.011 | 2.034 | 1.959 | 1.989 | 1.867 | 1.746 | 1.952 | 5.65 |
| 50) | | Dimethylphthalate | 1.310 | 1.259 | 1.292 | 1.257 | 1.302 | 1.219 | 1.169 | 1.258 | 4.01 |
| 51) | | 2,6-Dinitrotol... | 0.133 | 0.176 | 0.222 | 0.234 | 0.262 | 0.262 | 0.247 | 0.219 | 22.07 |
| 52) | C | Acenaphthene | 1.273 | 1.204 | 1.173 | 1.127 | 1.168 | 1.088 | 1.032 | 1.152 | 6.85 |
| 53) | | 3-Nitroaniline | 0.180 | 0.227 | 0.283 | 0.288 | 0.318 | 0.313 | 0.295 | 0.272 | 18.56 |
| 54) | P | 2,4-Dinitrophenol | | 0.038 | 0.056 | 0.069 | 0.086 | 0.090 | 0.097 | 0.073 | 31.18 |
| 55) | | Dibenzofuran | 1.895 | 1.795 | 1.785 | 1.697 | 1.729 | 1.604 | 1.516 | 1.717 | 7.37 |
| 56) | P | 4-Nitrophenol | | 0.164 | 0.201 | 0.224 | 0.239 | 0.237 | 0.232 | 0.216 | 13.37 |
| 57) | | 2,4-Dinitrotol... | 0.147 | 0.197 | 0.253 | 0.295 | 0.326 | 0.319 | 0.320 | 0.265 | 26.25 |
| 58) | | Fluorene | 1.478 | 1.387 | 1.328 | 1.246 | 1.284 | 1.186 | 1.105 | 1.288 | 9.67 |
| 59) | | 2,3,4,6-Tetrac... | 0.253 | 0.274 | 0.302 | 0.312 | 0.339 | 0.323 | 0.309 | 0.302 | 9.71 |
| 60) | | Diethylphthalate | 1.221 | 1.217 | 1.257 | 1.225 | 1.260 | 1.187 | 1.130 | 1.214 | 3.68 |
| 61) | | 4-Chlorophenyl... | 0.666 | 0.642 | 0.631 | 0.614 | 0.625 | 0.586 | 0.555 | 0.617 | 5.92 |
| 62) | | 4-Nitroaniline | 0.169 | 0.208 | 0.237 | 0.249 | 0.270 | 0.263 | 0.259 | 0.236 | 15.25 |
| 63) | | Azobenzene | 1.425 | 1.399 | 1.375 | 1.328 | 1.381 | 1.282 | 1.208 | 1.343 | 5.65 |
| 64) | I | Phenanthrene-d10 | ----------ISTD---------- | | | | | | | | |
| 65) | | 4,6-Dinitro-2-... | | 0.035 | 0.053 | 0.064 | 0.079 | 0.084 | 0.086 | 0.067 | 30.46 |
| 66) | c | n-Nitrosodiphe... | 0.735 | 0.709 | 0.736 | 0.682 | 0.717 | 0.685 | 0.654 | 0.703 | 4.30 |
| 67) | | 4-Bromophenyl-... | 0.231 | 0.219 | 0.234 | 0.222 | 0.239 | 0.228 | 0.219 | 0.227 | 3.38 |
| 68) | | Hexachlorobenzene | 0.249 | 0.237 | 0.247 | 0.235 | 0.250 | 0.239 | 0.231 | 0.241 | 3.12 |
| 69) | | Atrazine | 0.158 | 0.171 | 0.184 | 0.181 | 0.195 | 0.185 | 0.184 | 0.180 | 6.72 |
| 70) | C | Pentachlorophenol | | 0.098 | 0.125 | 0.132 | 0.149 | 0.145 | 0.145 | 0.132 | 14.47 |
| 71) | | Phenanthrene | 1.190 | 1.136 | 1.111 | 1.042 | 1.089 | 1.021 | 0.968 | 1.079 | 6.95 |
| 72) | | Anthracene | 1.158 | 1.126 | 1.126 | 1.057 | 1.106 | 1.042 | 0.983 | 1.085 | 5.62 |
| 73) | | Carbazole | 1.038 | 1.026 | 1.008 | 0.938 | 0.998 | 0.927 | 0.873 | 0.972 | 6.27 |
| 74) | | Di-n-butylphth... | 0.746 | 0.846 | 0.913 | 0.921 | 0.983 | 0.942 | 0.897 | 0.893 | 8.62 |
| 75) | C | Fluoranthene | 1.044 | 1.053 | 1.009 | 0.960 | 0.998 | 0.938 | 0.869 | 0.981 | 6.59 |
| 76) | I | Chrysene-d12 | ----------ISTD---------- | | | | | | | | |
| 77) | | Benzidine | | 0.664 | 0.786 | 0.834 | 0.674 | 0.616 | 0.514 | 0.681 | 17.02 |
| 78) | | Pyrene | 2.018 | 1.956 | 1.955 | 1.842 | 1.918 | 1.716 | 1.585 | 1.856 | 8.33 |
| 79) | S | Terphenyl-d14 | 1.563 | 1.471 | 1.437 | 1.343 | 1.385 | 1.246 | 1.133 | 1.368 | 10.53 |
| 80) | | Butylbenzylpht... | 0.179 | 0.233 | 0.310 | 0.343 | 0.385 | 0.395 | 0.384 | 0.318 | 26.22 |
| 81) | | Benzo(a)anthra... | 1.296 | 1.304 | 1.406 | 1.347 | 1.403 | 1.326 | 1.247 | 1.333 | 4.34 |
| 82) | | 3,3'-Dichlorob... | | 0.314 | 0.406 | 0.385 | 0.428 | 0.411 | 0.377 | 0.387 | 10.35 |
| 83) | | Chrysene | 1.286 | 1.234 | 1.211 | 1.201 | 1.270 | 1.228 | 1.162 | 1.227 | 3.41 |
| 84) | | Bis(2-ethylhex... | 0.273 | 0.343 | 0.479 | 0.515 | 0.578 | 0.591 | 0.583 | 0.480 | 26.25 |
| 85) | c | Di-n-octyl pht... | | 0.474 | 0.811 | 0.922 | 1.073 | 1.115 | 1.167 | 0.927 | 27.83 |

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF071525.M

| | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|
| 86) | I | Perylene-d12 | ----------ISTD---------- | | | | | | | |
| 87) | | Indeno(1,2,3-c... | 1.345 | 1.411 | 1.526 | 1.521 | 1.620 | 1.519 | 1.470 | 1.488 | 5.98 |
| 88) | | Benzo(b)fluora... | 1.063 | 1.191 | 1.171 | 1.082 | 1.297 | 1.185 | 1.087 | 1.154 | 7.17 |
| 89) | | Benzo(k)fluora... | 1.175 | 1.066 | 1.111 | 1.148 | 1.088 | 1.060 | 1.056 | 1.101 | 4.23 |
| 90) | C | Benzo(a)pyrene | 0.989 | 1.019 | 1.108 | 1.124 | 1.196 | 1.124 | 1.078 | 1.091 | 6.41 |
| 91) | | Dibenzo(a,h)an... | 1.114 | 1.162 | 1.266 | 1.240 | 1.330 | 1.222 | 1.178 | 1.216 | 5.90 |
| 92) | | Benzo(g,h,i)pe... | 1.148 | 1.187 | 1.263 | 1.258 | 1.349 | 1.260 | 1.214 | 1.240 | 5.20 |

(#) = Out of Range




Response = 1.498e-001 * Amt - 2.944e-002
Coef of Det (r^2) = 0.992826  Curve Fit: wlr(1/a)
Method: \\nsvcsrv\HPCHEM1\BNA F\Methods\8270-BF071525.M
Calibration Table Last Updated: Tue Jul 15 17:53:25 2025




Response = 3.746e-001 * Amt - 5.948e-002
Coef of Det (r^2) = 0.996906  Curve Fit: wlr(1/a)
Method: Q2502-SVOCMS:Group1
Calibration Table Last Updated: Tue Jul 15 17:53:25 2025




R = 1.009e-002 A*A + 6.225e-002 A - 1.545e-002
Coef of Det (r^2) = 0.996906  Curve Fit: Quadratic w(1/a)
Method: \\nsvocms\chemsrv\HPCHEM1\BNA F\Methods\8270-BF071525.M
Calibration Table Last Updated: Tue Jul 15 17:53:25 2025




Response = 3.341e-001 * Amt - 5.679e-002
Coef of Det (r^2) = 0.997543  Curve Fit: wlr(1/a)
Method: Q2502-SVOCMS:Group1
Calibration Table Last Updated: Tue Jul 15 17:53:25 2025




Response = 2.640e-001 * Amt - 3.691e-002
Coef of Det (r^2) = 0.997212  Curve Fit: wlr(1/a)
Method: \\chemsrv\HPCHEM1\BNA F\Methods\8270-BF071525.M
Calibration Table Last Updated: Tue Jul 15 17:53:25 2025




R = 6.691e-003 A*A + 6.646e-002 A - 1.863e-002
Coef of Det (r^2) = 0.995577  Curve Fit: Quadratic w(1/a)
Method: \\nsvcsrv\HPCHEM1\BNA F\Methods\8270-BF071525.M
Calibration Table Last Updated: Tue Jul 15 17:53:25 2025




Response = 4.013e-001 * Amt - 6.845e-002
Coef of Det (r^2) = 0.996788  Curve Fit: wlr(1/a)
Method: \\nt-mscsrv\HPCHEM1\BNA F\Methods\8270-BF071525.M
Calibration Table Last Updated: Tue Jul 15 17:53:25 2025




Response = 6.053e-001 * Amt - 1.032e-001
Coef of Det (r^2) = 0.996945  Curve Fit: wlr(1/a)
Method: Q2502-SVOCMS:Group1
Calibration Table Last Updated: Tue Jul 15 17:53:25 2025




Response = 1.236e+000 * Amt - 4.136e-001
Coef of Det (r^2) = 0.995967  Curve Fit: wlr(1/a)
Method: \\nsvolsrv\HPCHEM1\BNA F\Methods\8270-BF071525.M
Calibration Table Last Updated: Tue Jul 15 17:53:25 2025

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143093.D
Acq On : 15 Jul 2025 13:04
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC2.5

Quant Time: Jul 15 17:36:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 6.969 | 152 | 130499 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 8.245 | 136 | 494118 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 10.004 | 164 | 253763 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 11.486 | 188 | 424376 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 14.127 | 240 | 244536 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 15.633 | 264 | 187520 | 20.000 | ng | 0.00 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 0.000 | 112 | 0d | 0.000 | ng | |
| 7) Phenol-d6 | 0.000 | 99 | 0d | 0.000 | ng | |
| 23) Nitrobenzene-d5 | 0.000 | 82 | 0d | 0.000 | ng | |
| 42) 2,4,6-Tribromophenol | 0.000 | 330 | 0d | 0.000 | ng | |
| 45) 2-Fluorobiphenyl | 0.000 | 172 | 0d | 0.000 | ng | |
| 79) Terphenyl-d14 | 0.000 | 244 | 0d | 0.000 | ng | |
| Target Compounds | | | | | | Qvalue |
| 19) n-Nitroso-di-n-propyla... | 7.363 | 70 | 15601 | 2.453 | ng | 95 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143093.D
Acq On : 15 Jul 2025 13:04
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC2.5

Quant Time: Jul 15 17:36:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.969 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC2.5

Tgt Ion:152 Resp: 130499

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 155.8 | 126.5 | 189.7 |
| 115 | 59.9 | 47.0 | 70.6 |







#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.59 min

Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

Tgt Ion: 112

| Sig | Exp Ratio |
|---|---|
| 112 | 100 |
| 64 | 68.3 |
| 63 | 34.1 |






#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.59 min

Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC2.5

Tgt Ion: 99

| Sig | Exp Ratio |
|---|---|
| 99 | 100 |
| 42 | 21.5 |
| 71 | 34.2 |




#19
n-Nitroso-di-n-propylamine
Concen: 2.453 ng
RT: 7.363 min Scan# 879
Delta R.T. -0.012 min
Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

Tgt Ion: 70 Resp: 15601

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 71.9 | 53.2 | 79.8 |
| 101 | 10.0 | 7.5 | 11.3 |
| 130 | 19.7 | 15.7 | 23.5 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC2.5

Tgt Ion:136 Resp: 494118

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 11.0 | 8.9 | 13.3 |
| 54 | 11.2 | 8.3 | 12.5 |
| 68 | 7.2 | 5.3 | 7.9 |







#23
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 7.53 min

Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

Tgt Ion: 82

| Sig | Exp Ratio |
|---|---|
| 82 | 100 |
| 128 | 41.6 |
| 54 | 59.3 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC2.5

Tgt Ion:164 Resp: 253763

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 99.1 | 79.0 | 118.6 |
| 160 | 44.7 | 35.8 | 53.6 |







#42
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 10.79 min

Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

Tgt Ion: 330

| Sig | Exp Ratio |
|---|---|
| 330 | 100 |
| 332 | 97.4 |
| 141 | 32.4 |






#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 9.33 min

Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC2.5

Tgt Ion: 172

| Sig | Exp Ratio |
|---|---|
| 172 | 100 |
| 171 | 35.3 |
| 170 | 23.3 |




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.486 min Scan# 1580
Delta R.T. -0.006 min
Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

Tgt Ion:188 Resp: 424376

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 11.2 | 8.6 | 13.0 |
| 80 | 12.7 | 9.3 | 13.9 |






#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.127 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC2.5

Tgt Ion:240 Resp: 244536

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 12.5 | 10.0 | 15.0 |
| 236 | 25.3 | 20.0 | 30.0 |







#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 13.07 min

Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

Tgt Ion: 244

| Sig | Exp Ratio |
|---|---|
| 244 | 100 |
| 212 | 6.5 |
| 122 | 11.0 |






#86
Perylene-d12
Concen: 20.000 ng
RT: 15.633 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF143093.D
Acq: 15 Jul 2025 13:04

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC2.5

Tgt Ion:264 Resp: 187520

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.0 | 18.5 | 27.7 |
| 265 | 21.3 | 17.5 | 26.3 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143094.D
Acq On : 15 Jul 2025 13:35
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jul 15 17:37:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| **Internal Standards** | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 6.969 | 152 | 119528 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 8.245 | 136 | 456442 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 10.004 | 164 | 230973 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 11.486 | 188 | 382578 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 14.127 | 240 | 212909 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 15.633 | 264 | 174154 | 20.000 | ng | 0.00 |
| **System Monitoring Compounds** | | | | | | |
| 5) 2-Fluorophenol | 5.581 | 112 | 79738 | 10.540 | ng | 0.00 |
| 7) Phenol-d6 | 6.575 | 99 | 100706 | 10.590 | ng | -0.01 |
| 23) Nitrobenzene-d5 | 7.516 | 82 | 64500m | 7.927 | ng | -0.01 |
| 42) 2,4,6-Tribromophenol | 10.786 | 330 | 15703 | 7.625 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 9.322 | 172 | 205273 | 11.814 | ng | 0.00 |
| 79) Terphenyl-d14 | 13.069 | 244 | 166389 | 11.424 | ng | 0.00 |
| **Target Compounds** | | | | | | Qvalue |
| 2) 1,4-Dioxane | 2.810 | 88 | 19027 | 5.052 | ng | 95 |
| 3) Pyridine | 3.587 | 79 | 48004 | 5.034 | ng | 97 |
| 6) Aniline | 6.622 | 93 | 69613 | 5.198 | ng | 99 |
| 8) 2-Chlorophenol | 6.745 | 128 | 37026 | 4.871 | ng | 98 |
| 10) Phenol | 6.587 | 94 | 52784 | 5.127 | ng | 98 |
| 11) bis(2-Chloroethyl)ether | 6.692 | 93 | 42574 | 5.342 | ng | 98 |
| 12) 1,3-Dichlorobenzene | 6.910 | 146 | 46218 | 5.382 | ng | 98 |
| 13) 1,4-Dichlorobenzene | 6.987 | 146 | 46918 | 5.431 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 7.140 | 146 | 43854 | 5.323 | ng | 97 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.240 | 45 | 81065 | 5.518 | ng | 99 |
| 17) 2-Methylphenol | 7.198 | 107 | 33834 | 5.158 | ng | 98 |
| 18) Hexachloroethane | 7.487 | 117 | 13889 | 4.938 | ng | 99 |
| 19) n-Nitroso-di-n-propyla... | 7.363 | 70 | 30747 | 5.277 | ng | 99 |
| 22) Acetophenone | 7.363 | 105 | 61393 | 5.580 | ng | 95 |
| 24) Nitrobenzene | 7.534 | 77 | 32206 | 4.100 | ng | 99 |
| 25) Isophorone | 7.775 | 82 | 80494 | 5.118 | ng | 99 |
| 26) 2-Nitrophenol | 7.857 | 139 | 7204 | 6.037 | ng | 96 |
| 27) 2,4-Dimethylphenol | 7.887 | 122 | 37098 | 5.082 | ng | 98 |
| 28) bis(2-Chloroethoxy)met... | 7.987 | 93 | 50637 | 5.310 | ng | 98 |
| 29) 2,4-Dichlorophenol | 8.092 | 162 | 27047 | 4.529 | ng | 96 |
| 30) 1,2,4-Trichlorobenzene | 8.187 | 180 | 36023 | 5.315 | ng | 99 |
| 31) Naphthalene | 8.269 | 128 | 125461 | 5.535 | ng | 100 |
| 33) 4-Chloroaniline | 8.310 | 127 | 47947 | 5.260 | ng | 98 |
| 34) Hexachlorobutadiene | 8.392 | 225 | 21105 | 5.207 | ng | 98 |
| 36) 4-Chloro-3-methylphenol | 8.775 | 107 | 32542 | 4.892 | ng | 96 |
| 37) 2-Methylnaphthalene | 8.957 | 142 | 73511 | 5.436 | ng | 99 |
| 38) 1-Methylnaphthalene | 9.057 | 142 | 77390 | 5.518 | ng | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 9.128 | 216 | 35579 | 5.286 | ng | 97 |
| 43) 2,4,6-Trichlorophenol | 9.228 | 196 | 17295 | 4.135 | ng | 97 |
| 44) 2,4,5-Trichlorophenol | 9.263 | 196 | 19416 | 4.427 | ng | 99 |
| 46) 1,1'-Biphenyl | 9.422 | 154 | 101056 | 5.486 | ng | 99 |
| 47) 2-Chloronaphthalene | 9.445 | 162 | 73385 | 5.423 | ng | 99 |
| 48) 2-Nitroaniline | 9.528 | 65 | 10398 | 5.579 | ng | 91 |
| 49) Acenaphthylene | 9.863 | 152 | 118965 | 5.276 | ng | 99 |
| 50) Dimethylphthalate | 9.710 | 163 | 75659 | 5.207 | ng | 99 |
| 51) 2,6-Dinitrotoluene | 9.769 | 165 | 7659 | 5.308 | ng | 93 |

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143094.D
Acq On : 15 Jul 2025 13:35
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jul 15 17:37:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 52) Acenaphthene | 10.033 | 154 | 73531 | 5.526 | ng | | 99 |
| 53) 3-Nitroaniline | 9.933 | 138 | 10382 | 3.307 | ng | # | 90 |
| 55) Dibenzofuran | 10.204 | 168 | 109432 | 5.518 | ng | | 99 |
| 57) 2,4-Dinitrotoluene | 10.169 | 165 | 8515 | 5.606 | ng | | 87 |
| 58) Fluorene | 10.551 | 166 | 85318 | 5.738 | ng | | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 10.322 | 232 | 14610 | 4.192 | ng | # | 98 |
| 60) Diethylphthalate | 10.416 | 149 | 70533 | 5.031 | ng | | 98 |
| 61) 4-Chlorophenyl-phenyle... | 10.539 | 204 | 38452 | 5.395 | ng | | 96 |
| 62) 4-Nitroaniline | 10.539 | 138 | 9779 | 3.581 | ng | | 83 |
| 63) Azobenzene | 10.698 | 77 | 82291 | 5.308 | ng | | 99 |
| 66) n-Nitrosodiphenylamine | 10.651 | 169 | 70293 | 5.231 | ng | | 99 |
| 67) 4-Bromophenyl-phenylether | 11.033 | 248 | 22127 | 5.086 | ng | | 95 |
| 68) Hexachlorobenzene | 11.098 | 284 | 23819 | 5.160 | ng | | 97 |
| 69) Atrazine | 11.175 | 200 | 15103 | 4.393 | ng | | 99 |
| 71) Phenanthrene | 11.510 | 178 | 113784 | 5.511 | ng | | 99 |
| 72) Anthracene | 11.563 | 178 | 110778 | 5.335 | ng | | 100 |
| 73) Carbazole | 11.710 | 167 | 99235 | 5.335 | ng | | 100 |
| 74) Di-n-butylphthalate | 12.051 | 149 | 71352 | 4.179 | ng | | 99 |
| 75) Fluoranthene | 12.704 | 202 | 99884 | 5.320 | ng | | 100 |
| 78) Pyrene | 12.927 | 202 | 107389 | 5.436 | ng | | 99 |
| 80) Butylbenzylphthalate | 13.545 | 149 | 9546 | 5.646 | ng | | 92 |
| 81) Benzo(a)anthracene | 14.116 | 228 | 68990 | 4.863 | ng | | 99 |
| 83) Chrysene | 14.151 | 228 | 68451 | 5.239 | ng | | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 14.110 | 149 | 14537 | 5.667 | ng | | 100 |
| 87) Indeno(1,2,3-cd)pyrene | 17.162 | 276 | 58565 | 4.521 | ng | | 98 |
| 88) Benzo(b)fluoranthene | 15.186 | 252 | 46276 | 4.607 | ng | | 98 |
| 89) Benzo(k)fluoranthene | 15.215 | 252 | 51177 | 5.340 | ng | | 99 |
| 90) Benzo(a)pyrene | 15.568 | 252 | 43046 | 4.531 | ng | | 97 |
| 91) Dibenzo(a,h)anthracene | 17.186 | 278 | 48506 | 4.581 | ng | | 97 |
| 92) Benzo(g,h,i)perylene | 17.627 | 276 | 49989 | 4.631 | ng | | 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143094.D
Acq On : 15 Jul 2025 13:35
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jul 15 17:37:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.969 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:152 Resp: 11952

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 156.2 | 126.5 | 189.7 |
| 115 | 59.5 | 47.0 | 70.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#2
1,4-Dioxane
Concen: 5.052 ng
RT: 2.810 min Scan# 105
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 88 Resp: 19027

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 94.1 | 70.5 | 105.7 |
| 43 | 34.9 | 27.3 | 40.9 |






#3
Pyridine
Concen: 5.034 ng
RT: 3.587 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 79 Resp: 4800

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 89.6 | 69.2 | 103.8 |
| 51 | 42.3 | 33.3 | 49.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#5
2-Fluorophenol
Concen: 10.540 ng
RT: 5.581 min Scan# 576
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:112 Resp: 79738

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 66.2 | 54.6 | 82.0 |
| 63 | 32.8 | 27.3 | 40.9 |






#6
Aniline
Concen: 5.198 ng
RT: 6.622 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 93 Resp: 6961

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 40.6 | 33.1 | 49.7 |
| 65 | 21.0 | 16.9 | 25.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#7
Phenol-d6
Concen: 10.590 ng
RT: 6.575 min Scan# 745
Delta R.T. -0.012 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 99 Resp: 100706

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 21.4 | 17.2 | 25.8 |
| 71 | 33.2 | 27.4 | 41.0 |






#8
2-Chlorophenol
Concen: 4.871 ng
RT: 6.745 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:128 Resp: 3702

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 30.6 | 12.8 | 52.8 |
| 64 | 55.9 | 36.1 | 76.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#10
Phenol
Concen: 5.127 ng
RT: 6.587 min Scan# 747
Delta R.T. -0.012 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 94 Resp: 52784

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 30.3 | 9.8 | 49.8 |
| 66 | 39.1 | 20.5 | 60.5 |






#11
bis(2-Chloroethyl)ether
Concen: 5.342 ng
RT: 6.692 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 93 Resp: 4257

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 84.4 | 61.9 | 101.9 |
| 95 | 32.9 | 12.2 | 52.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#12
1,3-Dichlorobenzene
Concen: 5.382 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:146 Resp: 46218

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 65.2 | 50.1 | 75.1 |
| 75 | 33.8 | 27.4 | 41.0 |






#13
1,4-Dichlorobenzene
Concen: 5.431 ng
RT: 6.987 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:146 Resp: 4691

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.7 | 50.6 | 76.0 |
| 111 | 43.3 | 35.0 | 52.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#14
1,2-Dichlorobenzene
Concen: 5.323 ng
RT: 7.140 min Scan# 841
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:146 Resp: 43854

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.4 | 49.8 | 74.6 |
| 111 | 46.3 | 35.4 | 53.0 |






#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.518 ng
RT: 7.240 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 45 Resp: 8106

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 10.7 | 0.0 | 31.1 |
| 79 | 8.8 | 0.0 | 28.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#17
2-Methylphenol
Concen: 5.158 ng
RT: 7.198 min Scan# 851
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:107 Resp: 33834

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 111.6 | 90.9 | 136.3 |
| 77 | 45.9 | 37.5 | 56.3 |
| 79 | 43.8 | 36.9 | 55.3 |






#18
Hexachloroethane
Concen: 4.938 ng
RT: 7.487 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:117 Resp: 1388

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 95.7 | 76.7 | 115.1 |
| 201 | 95.6 | 75.0 | 112.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#19
n-Nitroso-di-n-propylamine
Concen: 5.277 ng
RT: 7.363 min Scan# 879
Delta R.T. -0.012 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 70 Resp: 30747

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 67.2 | 53.2 | 79.8 |
| 101 | 9.1 | 7.5 | 11.3 |
| 130 | 18.2 | 15.7 | 23.5 |




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:136 Resp: 45644

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.9 | 8.9 | 13.3 |
| 54 | 10.9 | 8.3 | 12.5 |
| 68 | 7.1 | 5.3 | 7.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#22
Acetophenone
Concen: 5.580 ng
RT: 7.363 min Scan# 879
Delta R.T. -0.012 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:105 Resp: 61393

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 8.5 | 7.2 | 10.8 |
| 51 | 38.8 | 27.0 | 40.4 |
| 120 | 22.7 | 18.2 | 27.2 |






#23
Nitrobenzene-d5
Concen: 7.927 ng m
RT: 7.516 min Scan# 9
Delta R.T. -0.012 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 82 Resp: 6450

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 37.2 | 33.3 | 49.9 |
| 54 | 59.3 | 47.4 | 71.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#24
Nitrobenzene
Concen: 4.100 ng
RT: 7.534 min Scan# 908
Delta R.T. -0.012 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 77 Resp: 32206

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 42.1 | 34.5 | 51.7 |
| 65 | 14.5 | 11.3 | 16.9 |






#25
Isophorone
Concen: 5.118 ng
RT: 7.775 min Scan# 94
Delta R.T. -0.012 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 82 Resp: 8049

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 6.6 | 5.6 | 8.4 |
| 138 | 17.6 | 14.3 | 21.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#26
2-Nitrophenol
Concen: 6.037 ng
RT: 7.857 min Scan# 963
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:139 Resp: 7204

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 29.4 | 23.4 | 35.0 |
| 65 | 58.5 | 43.4 | 65.2 |






#27
2,4-Dimethylphenol
Concen: 5.082 ng
RT: 7.887 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:122 Resp: 3709

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 110.4 | 91.3 | 136.9 |
| 121 | 56.9 | 45.5 | 68.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#28
bis(2-Chloroethoxy)methane
Concen: 5.310 ng
RT: 7.987 min Scan# 985
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 93 Resp: 50637

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 31.9 | 26.2 | 39.4 |
| 123 | 11.2 | 9.6 | 14.4 |






#29
2,4-Dichlorophenol
Concen: 4.529 ng
RT: 8.092 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:162 Resp: 2704

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 65.7 | 43.1 | 83.1 |
| 98 | 41.6 | 19.2 | 59.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#30
1,2,4-Trichlorobenzene
Concen: 5.315 ng
RT: 8.187 min Scan# 1019
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:180 Resp: 36023

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 93.9 | 74.7 | 112.1 |
| 145 | 32.9 | 24.5 | 36.7 |






#31
Naphthalene
Concen: 5.535 ng
RT: 8.269 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:128 Resp: 12546

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.4 | 9.0 | 13.4 |
| 127 | 13.2 | 10.5 | 15.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#33
4-Chloroaniline
Concen: 5.260 ng
RT: 8.310 min Scan# 1040
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:127 Resp: 47947

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 31.4 | 26.2 | 39.4 |
| 65 | 35.2 | 28.6 | 42.8 |
| 92 | 19.5 | 17.1 | 25.7 |






#34
Hexachlorobutadiene
Concen: 5.207 ng
RT: 8.392 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:225 Resp: 2110

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 64.1 | 50.3 | 75.5 |
| 227 | 64.8 | 50.1 | 75.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#36
4-Chloro-3-methylphenol
Concen: 4.892 ng
RT: 8.775 min Scan# 1119
Delta R.T. -0.012 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:107 Resp: 32542

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 24.1 | 20.2 | 30.2 |
| 142 | 74.1 | 62.4 | 93.6 |






#37
2-Methylnaphthalene
Concen: 5.436 ng
RT: 8.957 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:142 Resp: 7351

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 88.7 | 70.2 | 105.2 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#38
1-Methylnaphthalene
Concen: 5.518 ng
RT: 9.057 min Scan# 1167
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:142 Resp: 77390

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 90.9 | 72.2 | 108.4 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:164 Resp: 23097

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 98.5 | 79.0 | 118.6 |
| 160 | 44.0 | 35.8 | 53.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.286 ng
RT: 9.128 min Scan# 1179
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:216 Resp: 35579

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 81.8 | 62.6 | 94.0 |
| 179 | 21.6 | 16.2 | 24.2 |
| 108 | 20.5 | 15.8 | 23.6 |






#42
2,4,6-Tribromophenol
Concen: 7.625 ng
RT: 10.786 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:330 Resp: 1570

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 98.4 | 77.9 | 116.9 |
| 141 | 35.3 | 25.9 | 38.9 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#43
2,4,6-Trichlorophenol
Concen: 4.135 ng
RT: 9.228 min Scan# 1196
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:196 Resp: 17295

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 92.8 | 76.9 | 115.3 |
| 200 | 29.5 | 23.5 | 35.3 |






#44
2,4,5-Trichlorophenol
Concen: 4.427 ng
RT: 9.263 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:196 Resp: 1941

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.9 | 76.4 | 114.6 |
| 97 | 53.7 | 42.4 | 63.6 |
| 132 | 31.2 | 23.9 | 35.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#45
2-Fluorobiphenyl
Concen: 11.814 ng
RT: 9.322 min Scan# 1212
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:172 Resp: 205273

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 34.1 | 28.2 | 42.4 |
| 170 | 22.5 | 18.6 | 28.0 |






#46
1,1'-Biphenyl
Concen: 5.486 ng
RT: 9.422 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:154 Resp: 10105

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 39.9 | 20.1 | 60.1 |
| 76 | 14.2 | 0.0 | 33.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#47
2-Chloronaphthalene
Concen: 5.423 ng
RT: 9.445 min Scan# 1233
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:162 Resp: 73385

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 40.6 | 33.0 | 49.6 |
| 164 | 31.6 | 26.1 | 39.1 |






#48
2-Nitroaniline
Concen: 5.579 ng
RT: 9.528 min Scan# 1
Delta R.T. -0.005 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 65 Resp: 1039

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 62.6 | 51.3 | 76.9 |
| 138 | 83.1 | 77.8 | 116.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#49
Acenaphthylene
Concen: 5.276 ng
RT: 9.863 min Scan# 1304
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:152 Resp: 118965

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.1 | 15.8 | 23.8 |
| 153 | 12.6 | 10.4 | 15.6 |






#50
Dimethylphthalate
Concen: 5.207 ng
RT: 9.710 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:163 Resp: 7565

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 3.9 | 3.4 | 5.2 |
| 164 | 10.1 | 8.2 | 12.2 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#51
2,6-Dinitrotoluene
Concen: 5.308 ng
RT: 9.769 min Scan# 1288
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:165 Resp: 7659

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 75.4 | 57.4 | 86.0 |
| 89 | 72.9 | 51.8 | 77.8 |






#52
Acenaphthene
Concen: 5.526 ng
RT: 10.033 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:154 Resp: 7353

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 109.4 | 86.7 | 130.1 |
| 152 | 52.2 | 41.4 | 62.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#53
3-Nitroaniline
Concen: 3.307 ng
RT: 9.933 min Scan# 1316
Delta R.T. -0.012 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:138 Resp: 10382

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 18.3 | 10.1 | 15.1# |
| 92 | 131.5 | 96.4 | 144.6 |






#55
Dibenzofuran
Concen: 5.518 ng
RT: 10.204 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:168 Resp: 10943

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 36.1 | 29.8 | 44.6 |
| 169 | 13.6 | 10.8 | 16.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#57
2,4-Dinitrotoluene
Concen: 5.606 ng
RT: 10.169 min Scan# 1356
Delta R.T. -0.011 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:165 Resp: 8515

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 59.3 | 40.6 | 61.0 |
| 89 | 90.5 | 62.3 | 93.5 |
| 182 | 2.8 | 2.6 | 3.8 |






#58
Fluorene
Concen: 5.738 ng
RT: 10.551 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:166 Resp: 8531

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 91.9 | 73.7 | 110.5 |
| 167 | 13.5 | 10.9 | 16.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#59
2,3,4,6-Tetrachlorophenol
Concen: 4.192 ng
RT: 10.322 min Scan# 1382
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:232 Resp: 14610

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 43.7 | 36.9 | 55.3 |
| 130 | 1.5 | 1.8 | 2.8# |
| 166 | 29.0 | 23.0 | 34.6 |






#60
Diethylphthalate
Concen: 5.031 ng
RT: 10.416 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:149 Resp: 7053

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 20.4 | 17.3 | 25.9 |
| 150 | 12.1 | 9.9 | 14.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#61
4-Chlorophenyl-phenylether
Concen: 5.395 ng
RT: 10.539 min Scan# 1419
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:204 Resp: 38452

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 32.5 | 25.8 | 38.6 |
| 141 | 62.4 | 46.8 | 70.2 |






#62
4-Nitroaniline
Concen: 3.581 ng
RT: 10.539 min Scan# 14
Delta R.T. -0.018 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:138 Resp: 977

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 64.5 | 30.4 | 70.4 |
| 108 | 81.6 | 49.5 | 89.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#63
Azobenzene
Concen: 5.308 ng
RT: 10.698 min Scan# 1446
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion: 77 Resp: 82291

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 23.1 | 4.0 | 44.0 |
| 105 | 21.8 | 1.4 | 41.4 |
| 51 | 33.4 | 14.3 | 54.3 |






#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.486 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:188 Resp: 38257

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 11.2 | 8.6 | 13.0 |
| 80 | 12.2 | 9.3 | 13.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#66
n-Nitrosodiphenylamine
Concen: 5.231 ng
RT: 10.651 min Scan# 1438
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:169 Resp: 70293

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 65.2 | 53.0 | 79.6 |
| 167 | 35.1 | 28.4 | 42.6 |






#67
4-Bromophenyl-phenylether
Concen: 5.086 ng
RT: 11.033 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

Tgt Ion:248 Resp: 2212

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 98.3 | 76.7 | 115.1 |
| 141 | 70.8 | 61.7 | 92.5 |

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#68
Hexachlorobenzene
Concen: 5.160 ng
RT: 11.098 min Scan# 1514
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:284 Resp: 23819

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 37.9 | 28.9 | 43.3 |
| 249 | 31.9 | 24.6 | 37.0 |






#69
Atrazine
Concen: 4.393 ng
RT: 11.175 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:200 Resp: 1510

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 27.9 | 9.1 | 49.1 |
| 215 | 50.4 | 31.0 | 71.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#71
Phenanthrene
Concen: 5.511 ng
RT: 11.510 min Scan# 1584
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:178 Resp: 113784

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.0 | 15.1 | 22.7 |
| 179 | 15.4 | 12.2 | 18.4 |






#72
Anthracene
Concen: 5.335 ng
RT: 11.563 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:178 Resp: 11077

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.3 | 14.7 | 22.1 |
| 179 | 15.2 | 12.4 | 18.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#73
Carbazole
Concen: 5.335 ng
RT: 11.710 min Scan# 1618
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:167 Resp: 99235

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.3 | 17.0 | 25.4 |
| 139 | 13.0 | 10.2 | 15.4 |






#74
Di-n-butylphthalate
Concen: 4.179 ng
RT: 12.051 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:149 Resp: 7135

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 8.7 | 7.3 | 10.9 |
| 104 | 5.1 | 3.8 | 5.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#75
Fluoranthene
Concen: 5.320 ng
RT: 12.704 min Scan# 1787
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:202 Resp: 99884

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 12.7 | 0.0 | 32.4 |
| 203 | 17.4 | 0.0 | 37.4 |






#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.127 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:240 Resp: 212909

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 12.4 | 10.0 | 15.0 |
| 236 | 24.8 | 20.0 | 30.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#78
Pyrene
Concen: 5.436 ng
RT: 12.927 min Scan# 1825
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:202 Resp: 107389

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 19.1 | 15.5 | 23.3 |
| 203 | 17.3 | 14.0 | 21.0 |






#79
Terphenyl-d14
Concen: 11.424 ng
RT: 13.069 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:244 Resp: 16638

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.3 | 5.2 | 7.8 |
| 122 | 11.9 | 8.8 | 13.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#80
Butylbenzylphthalate
Concen: 5.646 ng
RT: 13.545 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:149 Resp: 9546

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 78.4 | 57.0 | 85.4 |
| 206 | 19.7 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 4.863 ng
RT: 14.116 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:228 Resp: 6899

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 25.9 | 21.3 | 31.9 |
| 229 | 19.7 | 15.5 | 23.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#83
Chrysene
Concen: 5.239 ng
RT: 14.151 min Scan# 2033
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:228 Resp: 68451

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 28.2 | 23.4 | 35.2 |
| 229 | 18.9 | 15.3 | 22.9 |






#84
Bis(2-ethylhexyl)phthalate
Concen: 5.667 ng
RT: 14.110 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:149 Resp: 1453

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 27.4 | 21.9 | 32.9 |
| 279 | 3.6 | 3.4 | 5.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.633 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:264 Resp: 174154

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 22.5 | 18.5 | 27.7 |
| 265 | 21.5 | 17.5 | 26.3 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 4.521 ng
RT: 17.162 min Scan# 2
Delta R.T. -0.018 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:276 Resp: 5856

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 29.8 | 22.6 | 34.0 |
| 277 | 25.1 | 20.2 | 30.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#88
Benzo(b)fluoranthene
Concen: 4.607 ng
RT: 15.186 min Scan# 2209
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:252 Resp: 46276

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.0 | 17.5 | 26.3 |
| 125 | 10.6 | 8.6 | 13.0 |






#89
Benzo(k)fluoranthene
Concen: 5.340 ng
RT: 15.215 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:252 Resp: 5117

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.1 | 17.4 | 26.2 |
| 125 | 11.0 | 8.5 | 12.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#90
Benzo(a)pyrene
Concen: 4.531 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.006 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:252 Resp: 43046

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 23.1 | 17.5 | 26.3 |
| 125 | 12.9 | 9.4 | 14.0 |






#91
Dibenzo(a,h)anthracene
Concen: 4.581 ng
RT: 17.186 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:278 Resp: 4850

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 19.7 | 14.6 | 22.0 |
| 279 | 24.6 | 18.8 | 28.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#92
Benzo(g,h,i)perylene
Concen: 4.631 ng
RT: 17.627 min Scan# 2624
Delta R.T. -0.018 min
Lab File: BF143094.D
Acq: 15 Jul 2025 13:35

Tgt Ion:276 Resp: 49989

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.6 | 19.2 | 28.8 |
| 138 | 26.6 | 19.8 | 29.6 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143095.D
Acq On : 15 Jul 2025 14:05
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

Quant Time: Jul 15 17:38:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| Internal Standards | | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 6.969 | 152 | 129158 | 20.000 | ng | | 0.00 |
| 21) Naphthalene-d8 | 8.245 | 136 | 488635 | 20.000 | ng | | 0.00 |
| 39) Acenaphthene-d10 | 10.004 | 164 | 248013 | 20.000 | ng | | 0.00 |
| 64) Phenanthrene-d10 | 11.486 | 188 | 412977 | 20.000 | ng | | 0.00 |
| 76) Chrysene-d12 | 14.127 | 240 | 231826 | 20.000 | ng | | 0.00 |
| 86) Perylene-d12 | 15.633 | 264 | 195100 | 20.000 | ng | | 0.00 |
| System Monitoring Compounds | | | | | | | |
| 5) 2-Fluorophenol | 5.581 | 112 | 164234 | 20.090 | ng | | 0.00 |
| 7) Phenol-d6 | 6.575 | 99 | 208774 | 20.317 | ng | | -0.01 |
| 23) Nitrobenzene-d5 | 7.516 | 82 | 156007m | 17.910 | ng | | -0.01 |
| 42) 2,4,6-Tribromophenol | 10.786 | 330 | 39055 | 17.661 | ng | | 0.00 |
| 45) 2-Fluorobiphenyl | 9.322 | 172 | 407861 | 21.860 | ng | | 0.00 |
| 79) Terphenyl-d14 | 13.069 | 244 | 340992 | 21.502 | ng | | 0.00 |
| Target Compounds | | | | | | | Qvalue |
| 2) 1,4-Dioxane | 2.810 | 88 | 40980 | 10.070 | ng | | 98 |
| 3) Pyridine | 3.581 | 79 | 99499 | 9.656 | ng | | 98 |
| 4) n-Nitrosodimethylamine | 3.504 | 42 | 50912 | 9.536 | ng | # | 96 |
| 6) Aniline | 6.622 | 93 | 145577 | 10.060 | ng | | 99 |
| 8) 2-Chlorophenol | 6.745 | 128 | 78009 | 9.498 | ng | | 98 |
| 9) Benzaldehyde | 6.510 | 77 | 77218 | 9.862 | ng | | 97 |
| 10) Phenol | 6.587 | 94 | 112508 | 10.114 | ng | | 98 |
| 11) bis(2-Chloroethyl)ether | 6.693 | 93 | 88047 | 10.224 | ng | | 99 |
| 12) 1,3-Dichlorobenzene | 6.910 | 146 | 95695 | 10.313 | ng | | 99 |
| 13) 1,4-Dichlorobenzene | 6.987 | 146 | 95134 | 10.191 | ng | | 99 |
| 14) 1,2-Dichlorobenzene | 7.140 | 146 | 90960 | 10.218 | ng | | 100 |
| 15) Benzyl Alcohol | 7.092 | 79 | 72743 | 9.554 | ng | | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.240 | 45 | 163494 | 10.299 | ng | | 99 |
| 17) 2-Methylphenol | 7.204 | 107 | 69808 | 9.849 | ng | | 99 |
| 18) Hexachloroethane | 7.487 | 117 | 28400 | 9.345 | ng | | 98 |
| 19) n-Nitroso-di-n-propyla... | 7.363 | 70 | 63831 | 10.139 | ng | | 95 |
| 20) 3+4-Methylphenols | 7.351 | 107 | 90318 | 10.391 | ng | # | 83 |
| 22) Acetophenone | 7.363 | 105 | 126127 | 10.708 | ng | | 95 |
| 24) Nitrobenzene | 7.540 | 77 | 77369 | 9.201 | ng | | 97 |
| 25) Isophorone | 7.775 | 82 | 167986 | 9.978 | ng | | 99 |
| 26) 2-Nitrophenol | 7.857 | 139 | 18432 | 8.966 | ng | | 97 |
| 27) 2,4-Dimethylphenol | 7.887 | 122 | 78706 | 10.072 | ng | | 98 |
| 28) bis(2-Chloroethoxy)met... | 7.987 | 93 | 106045 | 10.387 | ng | | 98 |
| 29) 2,4-Dichlorophenol | 8.092 | 162 | 62164 | 9.724 | ng | | 99 |
| 30) 1,2,4-Trichlorobenzene | 8.187 | 180 | 74581 | 10.280 | ng | | 98 |
| 31) Naphthalene | 8.269 | 128 | 256031 | 10.551 | ng | | 99 |
| 32) Benzoic acid | 7.939 | 122 | 19661m | 10.880 | ng | | |
| 33) 4-Chloroaniline | 8.310 | 127 | 102550 | 10.510 | ng | | 96 |
| 34) Hexachlorobutadiene | 8.392 | 225 | 42959 | 9.901 | ng | | 97 |
| 35) Caprolactam | 8.634 | 113 | 17855 | 8.944 | ng | | 99 |
| 36) 4-Chloro-3-methylphenol | 8.775 | 107 | 70747 | 9.935 | ng | | 97 |
| 37) 2-Methylnaphthalene | 8.957 | 142 | 152822 | 10.557 | ng | | 99 |
| 38) 1-Methylnaphthalene | 9.057 | 142 | 161228 | 10.739 | ng | | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 9.128 | 216 | 74348 | 10.287 | ng | | 99 |
| 41) Hexachlorocyclopentadiene | 9.122 | 237 | 31472 | 7.854 | ng | | 100 |
| 43) 2,4,6-Trichlorophenol | 9.234 | 196 | 38748 | 8.628 | ng | | 98 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143095.D
Acq On : 15 Jul 2025 14:05
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

Quant Time: Jul 15 17:38:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 9.263 | 196 | 45455 | 9.653 | ng | | 98 |
| 46) 1,1'-Biphenyl | 9.422 | 154 | 209259 | 10.580 | ng | | 99 |
| 47) 2-Chloronaphthalene | 9.445 | 162 | 149175 | 10.267 | ng | | 99 |
| 48) 2-Nitroaniline | 9.534 | 65 | 27816 | 9.163 | ng | | 98 |
| 49) Acenaphthylene | 9.863 | 152 | 249435 | 10.302 | ng | | 100 |
| 50) Dimethylphthalate | 9.710 | 163 | 156072 | 10.003 | ng | | 100 |
| 51) 2,6-Dinitrotoluene | 9.769 | 165 | 21789 | 9.451 | ng | | 96 |
| 52) Acenaphthene | 10.033 | 154 | 149318 | 10.451 | ng | | 100 |
| 53) 3-Nitroaniline | 9.939 | 138 | 28105 | 8.336 | ng | | 98 |
| 54) 2,4-Dinitrophenol | 10.039 | 184 | 4729 | 10.239 | ng | # | 1 |
| 55) Dibenzofuran | 10.204 | 168 | 222555 | 10.451 | ng | | 98 |
| 56) 4-Nitrophenol | 10.081 | 139 | 20334 | 7.589 | ng | | 97 |
| 57) 2,4-Dinitrotoluene | 10.175 | 165 | 24420 | 9.293 | ng | # | 95 |
| 58) Fluorene | 10.551 | 166 | 171975 | 10.771 | ng | | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 10.322 | 232 | 33995 | 9.084 | ng | | 99 |
| 60) Diethylphthalate | 10.416 | 149 | 150934 | 10.027 | ng | | 100 |
| 61) 4-Chlorophenyl-phenyle... | 10.545 | 204 | 79583 | 10.400 | ng | | 99 |
| 62) 4-Nitroaniline | 10.545 | 138 | 25757 | 8.785 | ng | | 96 |
| 63) Azobenzene | 10.698 | 77 | 173437 | 10.418 | ng | | 98 |
| 65) 4,6-Dinitro-2-methylph... | 10.581 | 198 | 7158 | 10.290 | ng | | 96 |
| 66) n-Nitrosodiphenylamine | 10.651 | 169 | 146433 | 10.094 | ng | | 98 |
| 67) 4-Bromophenyl-phenylether | 11.033 | 248 | 45279 | 9.641 | ng | | 97 |
| 68) Hexachlorobenzene | 11.104 | 284 | 49035 | 9.840 | ng | | 96 |
| 69) Atrazine | 11.175 | 200 | 35230 | 9.492 | ng | | 99 |
| 70) Pentachlorophenol | 11.286 | 266 | 20245 | 7.405 | ng | | 97 |
| 71) Phenanthrene | 11.510 | 178 | 234601 | 10.526 | ng | | 100 |
| 72) Anthracene | 11.563 | 178 | 232459 | 10.372 | ng | | 99 |
| 73) Carbazole | 11.710 | 167 | 211853 | 10.551 | ng | | 99 |
| 74) Di-n-butylphthalate | 12.051 | 149 | 174639 | 9.475 | ng | | 100 |
| 75) Fluoranthene | 12.704 | 202 | 217346 | 10.725 | ng | | 99 |
| 77) Benzidine | 12.816 | 184 | 76979 | 9.747 | ng | | 99 |
| 78) Pyrene | 12.927 | 202 | 226752 | 10.542 | ng | | 99 |
| 80) Butylbenzylphthalate | 13.545 | 149 | 26977 | 9.210 | ng | | 94 |
| 81) Benzo(a)anthracene | 14.116 | 228 | 151099 | 9.781 | ng | | 99 |
| 82) 3,3'-Dichlorobenzidine | 14.074 | 252 | 36435 | 8.122 | ng | | 99 |
| 83) Chrysene | 14.151 | 228 | 142990 | 10.050 | ng | | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 14.110 | 149 | 39724 | 9.073 | ng | | 97 |
| 85) Di-n-octyl phthalate | 14.727 | 149 | 54952 | 10.526 | ng | | 96 |
| 87) Indeno(1,2,3-cd)pyrene | 17.168 | 276 | 137620 | 9.484 | ng | | 100 |
| 88) Benzo(b)fluoranthene | 15.186 | 252 | 116154 | 10.321 | ng | | 100 |
| 89) Benzo(k)fluoranthene | 15.216 | 252 | 103994 | 9.686 | ng | | 98 |
| 90) Benzo(a)pyrene | 15.568 | 252 | 99366 | 9.336 | ng | | 99 |
| 91) Dibenzo(a,h)anthracene | 17.186 | 278 | 113316 | 9.553 | ng | | 99 |
| 92) Benzo(g,h,i)perylene | 17.627 | 276 | 115816 | 9.577 | ng | | 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143095.D
Acq On : 15 Jul 2025 14:05
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jul 15 17:38:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.969 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:152 Resp: 12915

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 154.7 | 126.5 | 189.7 |
| 115 | 59.3 | 47.0 | 70.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#2
1,4-Dioxane
Concen: 10.070 ng
RT: 2.810 min Scan# 105
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 88 Resp: 40980

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 86.3 | 70.5 | 105.7 |
| 43 | 34.8 | 27.3 | 40.9 |






#3
Pyridine
Concen: 9.656 ng
RT: 3.581 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 79 Resp: 9949

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 88.9 | 69.2 | 103.8 |
| 51 | 41.5 | 33.3 | 49.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#4
n-Nitrosodimethylamine
Concen: 9.536 ng
RT: 3.504 min Scan# 223
Delta R.T. -0.018 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 42 Resp: 50912

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 111.6 | 92.8 | 139.2 |
| 44 | 9.1 | 5.8 | 8.6# |






#5
2-Fluorophenol
Concen: 20.090 ng
RT: 5.581 min Scan# 5
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:112 Resp: 16423

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 69.1 | 54.6 | 82.0 |
| 63 | 33.7 | 27.3 | 40.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#6
Aniline
Concen: 10.060 ng
RT: 6.622 min Scan# 753
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 93 Resp: 145577

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 41.0 | 33.1 | 49.7 |
| 65 | 20.1 | 16.9 | 25.3 |






#7
Phenol-d6
Concen: 20.317 ng
RT: 6.575 min Scan# 74
Delta R.T. -0.012 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 99 Resp: 20877

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 22.2 | 17.2 | 25.8 |
| 71 | 32.8 | 27.4 | 41.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#8
2-Chlorophenol
Concen: 9.498 ng
RT: 6.745 min Scan# 774
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:128 Resp: 78009

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.6 | 12.8 | 52.8 |
| 64 | 58.5 | 36.1 | 76.1 |






#9
Benzaldehyde
Concen: 9.862 ng
RT: 6.510 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 77 Resp: 7721

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 91.6 | 76.3 | 116.3 |
| 106 | 91.1 | 72.8 | 112.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#10
Phenol
Concen: 10.114 ng
RT: 6.587 min Scan# 747
Delta R.T. -0.012 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 94 Resp: 112508

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 29.2 | 9.8 | 49.8 |
| 66 | 39.1 | 20.5 | 60.5 |






#11
bis(2-Chloroethyl)ether
Concen: 10.224 ng
RT: 6.693 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 93 Resp: 8804

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 82.9 | 61.9 | 101.9 |
| 95 | 31.1 | 12.2 | 52.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#12
1,3-Dichlorobenzene
Concen: 10.313 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:146 Resp: 95695

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 61.6 | 50.1 | 75.1 |
| 75 | 33.3 | 27.4 | 41.0 |






#13
1,4-Dichlorobenzene
Concen: 10.191 ng
RT: 6.987 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:146 Resp: 9513

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.6 | 50.6 | 76.0 |
| 111 | 42.2 | 35.0 | 52.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#14
1,2-Dichlorobenzene
Concen: 10.218 ng
RT: 7.140 min Scan# 841
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:146 Resp: 90960

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.6 | 49.8 | 74.6 |
| 111 | 44.4 | 35.4 | 53.0 |






#15
Benzyl Alcohol
Concen: 9.554 ng
RT: 7.092 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 79 Resp: 7274

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 74.9 | 61.1 | 91.7 |
| 77 | 62.8 | 49.9 | 74.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.299 ng
RT: 7.240 min Scan# 858
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 45 Resp: 163494

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 10.8 | 0.0 | 31.1 |
| 79 | 8.6 | 0.0 | 28.4 |






#17
2-Methylphenol
Concen: 9.849 ng
RT: 7.204 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:107 Resp: 6980

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 113.3 | 90.9 | 136.3 |
| 77 | 45.1 | 37.5 | 56.3 |
| 79 | 45.0 | 36.9 | 55.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#18
Hexachloroethane
Concen: 9.345 ng
RT: 7.487 min Scan# 900
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:117 Resp: 28400

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 94.7 | 76.7 | 115.1 |
| 201 | 95.9 | 75.0 | 112.4 |




#19
n-Nitroso-di-n-propylamine
Concen: 10.139 ng
RT: 7.363 min Scan# 8
Delta R.T. -0.012 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 70 Resp: 6383

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 71.4 | 53.2 | 79.8 |
| 101 | 9.2 | 7.5 | 11.3 |
| 130 | 19.2 | 15.7 | 23.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#20
3+4-Methylphenols
Concen: 10.391 ng
RT: 7.351 min Scan# 877
Delta R.T. -0.012 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:107 Resp: 90318

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 91.6 | 71.3 | 111.3 |
| 77 | 61.4 | 77.2 | 117.2# |
| 79 | 29.1 | 10.7 | 50.7 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.245 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:136 Resp: 48863

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 11.1 | 8.9 | 13.3 |
| 54 | 11.2 | 8.3 | 12.5 |
| 68 | 7.0 | 5.3 | 7.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#22
Acetophenone
Concen: 10.708 ng
RT: 7.363 min Scan# 879
Delta R.T. -0.012 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:105 Resp: 126127

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 8.5 | 7.2 | 10.8 |
| 51 | 38.5 | 27.0 | 40.4 |
| 120 | 22.0 | 18.2 | 27.2 |






#23
Nitrobenzene-d5
Concen: 17.910 ng m
RT: 7.516 min Scan# 9
Delta R.T. -0.012 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 82 Resp: 15600

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 36.8 | 33.3 | 49.9 |
| 54 | 60.0 | 47.4 | 71.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#24
Nitrobenzene
Concen: 9.201 ng
RT: 7.540 min Scan# 909
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 77 Resp: 77369

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 41.2 | 34.5 | 51.7 |
| 65 | 13.4 | 11.3 | 16.9 |






#25
Isophorone
Concen: 9.978 ng
RT: 7.775 min Scan# 94
Delta R.T. -0.012 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 82 Resp: 16798

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 6.7 | 5.6 | 8.4 |
| 138 | 17.4 | 14.3 | 21.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#26
2-Nitrophenol
Concen: 8.966 ng
RT: 7.857 min Scan# 963
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:139 Resp: 18432

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 33.4 | 23.4 | 35.0 |
| 65 | 54.8 | 43.4 | 65.2 |






#27
2,4-Dimethylphenol
Concen: 10.072 ng
RT: 7.887 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:122 Resp: 7870

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 116.5 | 91.3 | 136.9 |
| 121 | 58.1 | 45.5 | 68.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#28
bis(2-Chloroethoxy)methane
Concen: 10.387 ng
RT: 7.987 min Scan# 985
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 93 Resp: 106045

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.1 | 26.2 | 39.4 |
| 123 | 11.0 | 9.6 | 14.4 |






#29
2,4-Dichlorophenol
Concen: 9.724 ng
RT: 8.092 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:162 Resp: 6216

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 62.3 | 43.1 | 83.1 |
| 98 | 39.8 | 19.2 | 59.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#30
1,2,4-Trichlorobenzene
Concen: 10.280 ng
RT: 8.187 min Scan# 1019
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:180 Resp: 74581

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 91.6 | 74.7 | 112.1 |
| 145 | 32.5 | 24.5 | 36.7 |






#31
Naphthalene
Concen: 10.551 ng
RT: 8.269 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:128 Resp: 25603

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 10.9 | 9.0 | 13.4 |
| 127 | 12.9 | 10.5 | 15.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#32
Benzoic acid
Concen: 10.880 ng m
RT: 7.939 min Scan# 977
Delta R.T. -0.048 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:122 Resp: 19661

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 117.5 | 105.4 | 145.4 |
| 77 | 99.5 | 78.6 | 118.6 |






#33
4-Chloroaniline
Concen: 10.510 ng
RT: 8.310 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:127 Resp: 10255

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 30.9 | 26.2 | 39.4 |
| 65 | 33.3 | 28.6 | 42.8 |
| 92 | 19.4 | 17.1 | 25.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#34
Hexachlorobutadiene
Concen: 9.901 ng
RT: 8.392 min Scan# 1054
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:225 Resp: 42959

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 65.7 | 50.3 | 75.5 |
| 227 | 63.9 | 50.1 | 75.1 |






#35
Caprolactam
Concen: 8.944 ng
RT: 8.634 min Scan# 1
Delta R.T. -0.035 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:113 Resp: 1785

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 206.2 | 186.2 | 226.2 |
| 56 | 150.0 | 134.1 | 174.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#36
4-Chloro-3-methylphenol
Concen: 9.935 ng
RT: 8.775 min Scan# 1119
Delta R.T. -0.012 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:107 Resp: 70747

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 24.1 | 20.2 | 30.2 |
| 142 | 75.2 | 62.4 | 93.6 |






#37
2-Methylnaphthalene
Concen: 10.557 ng
RT: 8.957 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:142 Resp: 15282

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 88.2 | 70.2 | 105.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#38
1-Methylnaphthalene
Concen: 10.739 ng
RT: 9.057 min Scan# 1167
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:142 Resp: 161228

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 89.0 | 72.2 | 108.4 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:164 Resp: 24801

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 98.3 | 79.0 | 118.6 |
| 160 | 44.3 | 35.8 | 53.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 10.287 ng
RT: 9.128 min Scan# 1179
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:216 Resp: 74348

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.7 | 62.6 | 94.0 |
| 179 | 20.7 | 16.2 | 24.2 |
| 108 | 19.9 | 15.8 | 23.6 |






#41
Hexachlorocyclopentadiene
Concen: 7.854 ng
RT: 9.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:237 Resp: 3147

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 61.8 | 41.6 | 81.6 |
| 272 | 12.6 | 0.0 | 32.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#42
2,4,6-Tribromophenol
Concen: 17.661 ng
RT: 10.786 min Scan# 1461
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:330 Resp: 39055

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 98.1 | 77.9 | 116.9 |
| 141 | 33.8 | 25.9 | 38.9 |






#43
2,4,6-Trichlorophenol
Concen: 8.628 ng
RT: 9.234 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:196 Resp: 3874

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.6 | 76.9 | 115.3 |
| 200 | 32.5 | 23.5 | 35.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#44
2,4,5-Trichlorophenol
Concen: 9.653 ng
RT: 9.263 min Scan# 1202
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:196 Resp: 45455

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 92.5 | 76.4 | 114.6 |
| 97 | 53.8 | 42.4 | 63.6 |
| 132 | 31.1 | 23.9 | 35.9 |






#45
2-Fluorobiphenyl
Concen: 21.860 ng
RT: 9.322 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:172 Resp: 40786

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 34.8 | 28.2 | 42.4 |
| 170 | 23.0 | 18.6 | 28.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#46
1,1'-Biphenyl
Concen: 10.580 ng
RT: 9.422 min Scan# 1229
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:154 Resp: 209259

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 39.5 | 20.1 | 60.1 |
| 76 | 14.0 | 0.0 | 33.7 |






#47
2-Chloronaphthalene
Concen: 10.267 ng
RT: 9.445 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:162 Resp: 14917

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 41.1 | 33.0 | 49.6 |
| 164 | 31.8 | 26.1 | 39.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#48
2-Nitroaniline
Concen: 9.163 ng
RT: 9.534 min Scan# 1248
Delta R.T. 0.001 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 65 Resp: 27816

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 63.7 | 51.3 | 76.9 |
| 138 | 99.7 | 77.8 | 116.6 |






#49
Acenaphthylene
Concen: 10.302 ng
RT: 9.863 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:152 Resp: 24943

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 19.8 | 15.8 | 23.8 |
| 153 | 12.8 | 10.4 | 15.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#50
Dimethylphthalate
Concen: 10.003 ng
RT: 9.710 min Scan# 1278
Delta R.T. -0.012 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:163 Resp: 156072

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.3 | 3.4 | 5.2 |
| 164 | 10.2 | 8.2 | 12.2 |






#51
2,6-Dinitrotoluene
Concen: 9.451 ng
RT: 9.769 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:165 Resp: 2178

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 74.4 | 57.4 | 86.0 |
| 89 | 69.0 | 51.8 | 77.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#52
Acenaphthene
Concen: 10.451 ng
RT: 10.033 min Scan# 1333
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:154 Resp: 149318

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 108.2 | 86.7 | 130.1 |
| 152 | 52.4 | 41.4 | 62.2 |






#53
3-Nitroaniline
Concen: 8.336 ng
RT: 9.939 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:138 Resp: 2810

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 14.5 | 10.1 | 15.1 |
| 92 | 122.1 | 96.4 | 144.6 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#54
2,4-Dinitrophenol
Concen: 10.239 ng
RT: 10.039 min Scan# 1334
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:184 Resp: 4729

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 321.1 | 149.3 | 223.9# |
| 154 | 2815.4 | 1048.0 | 1572.0# |






#55
Dibenzofuran
Concen: 10.451 ng
RT: 10.204 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:168 Resp: 22255

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 36.0 | 29.8 | 44.6 |
| 169 | 13.3 | 10.8 | 16.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#56
4-Nitrophenol
Concen: 7.589 ng
RT: 10.081 min Scan# 1341
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:139 Resp: 20334

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 73.8 | 49.4 | 89.4 |
| 65 | 106.1 | 88.2 | 128.2 |






#57
2,4-Dinitrotoluene
Concen: 9.293 ng
RT: 10.175 min Scan# 1
Delta R.T. -0.005 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:165 Resp: 2442

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 53.0 | 40.6 | 61.0 |
| 89 | 82.8 | 62.3 | 93.5 |
| 182 | 2.4 | 2.6 | 3.8# |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#58
Fluorene
Concen: 10.771 ng
RT: 10.551 min Scan# 1421
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:166 Resp: 171975

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 91.8 | 73.7 | 110.5 |
| 167 | 13.5 | 10.9 | 16.3 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 9.084 ng
RT: 10.322 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:232 Resp: 3399

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 46.3 | 36.9 | 55.3 |
| 130 | 2.0 | 1.8 | 2.8 |
| 166 | 30.0 | 23.0 | 34.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#60
Diethylphthalate
Concen: 10.027 ng
RT: 10.416 min Scan# 1398
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:149 Resp: 150934

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 21.3 | 17.3 | 25.9 |
| 150 | 12.3 | 9.9 | 14.9 |






#61
4-Chlorophenyl-phenylether
Concen: 10.400 ng
RT: 10.545 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:204 Resp: 7958

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 31.4 | 25.8 | 38.6 |
| 141 | 58.6 | 46.8 | 70.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#62
4-Nitroaniline
Concen: 8.785 ng
RT: 10.545 min Scan# 1420
Delta R.T. -0.012 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:138 Resp: 25757

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 53.5 | 30.4 | 70.4 |
| 108 | 72.3 | 49.5 | 89.5 |






#63
Azobenzene
Concen: 10.418 ng
RT: 10.698 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion: 77 Resp: 17343

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 22.5 | 4.0 | 44.0 |
| 105 | 20.6 | 1.4 | 41.4 |
| 51 | 33.5 | 14.3 | 54.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.486 min Scan# 1580
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:188 Resp: 412977

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 11.4 | 8.6 | 13.0 |
| 80 | 12.6 | 9.3 | 13.9 |






#65
4,6-Dinitro-2-methylphenol
Concen: 10.290 ng
RT: 10.581 min Scan# 14
Delta R.T. -0.005 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:198 Resp: 715

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 66.0 | 44.6 | 84.6 |
| 105 | 54.0 | 29.3 | 69.3 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#66
n-Nitrosodiphenylamine
Concen: 10.094 ng
RT: 10.651 min Scan# 1438
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:169 Resp: 146433

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 64.2 | 53.0 | 79.6 |
| 167 | 34.9 | 28.4 | 42.6 |






#67
4-Bromophenyl-phenylether
Concen: 9.641 ng
RT: 11.033 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:248 Resp: 4527

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 97.5 | 76.7 | 115.1 |
| 141 | 73.9 | 61.7 | 92.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#68
Hexachlorobenzene
Concen: 9.840 ng
RT: 11.104 min Scan# 1515
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:284 Resp: 49035

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 32.6 | 28.9 | 43.3 |
| 249 | 30.1 | 24.6 | 37.0 |






#69
Atrazine
Concen: 9.492 ng
RT: 11.175 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:200 Resp: 3523

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 28.9 | 9.1 | 49.1 |
| 215 | 50.0 | 31.0 | 71.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#70
Pentachlorophenol
Concen: 7.405 ng
RT: 11.286 min Scan# 1546
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:266 Resp: 20245

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 63.0 | 51.8 | 77.6 |
| 264 | 64.6 | 49.7 | 74.5 |






#71
Phenanthrene
Concen: 10.526 ng
RT: 11.510 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:178 Resp: 23460

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.0 | 15.1 | 22.7 |
| 179 | 15.4 | 12.2 | 18.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#72
Anthracene
Concen: 10.372 ng
RT: 11.563 min Scan# 1593
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:178 Resp: 232459

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 17.7 | 14.7 | 22.1 |
| 179 | 15.2 | 12.4 | 18.6 |






#73
Carbazole
Concen: 10.551 ng
RT: 11.710 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:167 Resp: 21185

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 20.5 | 17.0 | 25.4 |
| 139 | 12.2 | 10.2 | 15.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#74
Di-n-butylphthalate
Concen: 9.475 ng
RT: 12.051 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:149 Resp: 174639

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.1 | 7.3 | 10.9 |
| 104 | 4.8 | 3.8 | 5.8 |






#75
Fluoranthene
Concen: 10.725 ng
RT: 12.704 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:202 Resp: 21734

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 11.9 | 0.0 | 32.4 |
| 203 | 17.3 | 0.0 | 37.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.127 min Scan# 2029
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:240 Resp: 231826

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.7 | 10.0 | 15.0 |
| 236 | 25.4 | 20.0 | 30.0 |






#77
Benzidine
Concen: 9.747 ng
RT: 12.816 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:184 Resp: 7697

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 16.4 | 12.4 | 18.6 |
| 183 | 10.8 | 8.6 | 12.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#78
Pyrene
Concen: 10.542 ng
RT: 12.927 min Scan# 1825
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:202 Resp: 226752

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 18.8 | 15.5 | 23.3 |
| 203 | 16.9 | 14.0 | 21.0 |






#79
Terphenyl-d14
Concen: 21.502 ng
RT: 13.069 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:244 Resp: 34099

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.5 | 5.2 | 7.8 |
| 122 | 12.7 | 8.8 | 13.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#80
Butylbenzylphthalate
Concen: 9.210 ng
RT: 13.545 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:149 Resp: 26977

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 76.7 | 57.0 | 85.4 |
| 206 | 20.2 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 9.781 ng
RT: 14.116 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:228 Resp: 151099

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 26.0 | 21.3 | 31.9 |
| 229 | 19.0 | 15.5 | 23.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#82
3,3'-Dichlorobenzidine
Concen: 8.122 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:252 Resp: 36435

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 65.2 | 51.5 | 77.3 |
| 126 | 12.3 | 9.4 | 14.0 |






#83
Chrysene
Concen: 10.050 ng
RT: 14.151 min Scan# 2(
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:228 Resp: 14299(

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 29.5 | 23.4 | 35.2 |
| 229 | 19.6 | 15.3 | 22.9 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#84
Bis(2-ethylhexyl)phthalate
Concen: 9.073 ng
RT: 14.110 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:149 Resp: 39724

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 25.6 | 21.9 | 32.9 |
| 279 | 3.6 | 3.4 | 5.0 |






#85
Di-n-octyl phthalate
Concen: 10.526 ng
RT: 14.727 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:149 Resp: 5495

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.6 | 1.2 | 1.8 |
| 43 | 15.1 | 10.6 | 16.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.633 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:264 Resp: 195100

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 22.4 | 18.5 | 27.7 |
| 265 | 21.4 | 17.5 | 26.3 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 9.484 ng
RT: 17.168 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:276 Resp: 13762

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 28.5 | 22.6 | 34.0 |
| 277 | 25.1 | 20.2 | 30.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#88
Benzo(b)fluoranthene
Concen: 10.321 ng
RT: 15.186 min Scan# 2209
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:252 Resp: 116154

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.0 | 17.5 | 26.3 |
| 125 | 10.9 | 8.6 | 13.0 |






#89
Benzo(k)fluoranthene
Concen: 9.686 ng
RT: 15.216 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:252 Resp: 10399

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.2 | 17.4 | 26.2 |
| 125 | 11.6 | 8.5 | 12.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#90
Benzo(a)pyrene
Concen: 9.336 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.006 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:252 Resp: 99366

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.1 | 17.5 | 26.3 |
| 125 | 12.1 | 9.4 | 14.0 |






#91
Dibenzo(a,h)anthracene
Concen: 9.553 ng
RT: 17.186 min Scan# 2
Delta R.T. -0.012 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:278 Resp: 11331

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 18.7 | 14.6 | 22.0 |
| 279 | 23.0 | 18.8 | 28.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#92
Benzo(g,h,i)perylene
Concen: 9.577 ng
RT: 17.627 min Scan# 2624
Delta R.T. -0.018 min
Lab File: BF143095.D
Acq: 15 Jul 2025 14:05

Tgt Ion:276 Resp: 115816

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.1 | 19.2 | 28.8 |
| 138 | 25.9 | 19.8 | 29.6 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143096.D
Acq On : 15 Jul 2025 14:36
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jul 15 17:39:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| **Internal Standards** | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 6.969 | 152 | 153233 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 8.251 | 136 | 579905 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 10.004 | 164 | 283300 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 11.492 | 188 | 451524 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 14.127 | 240 | 236667 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 15.633 | 264 | 262419 | 20.000 | ng | 0.00 |
| **System Monitoring Compounds** | | | | | | |
| 5) 2-Fluorophenol | 5.587 | 112 | 402974 | 41.549 | ng | 0.00 |
| 7) Phenol-d6 | 6.581 | 99 | 502651 | 41.231 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 7.522 | 82 | 420674m | 40.693 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 10.792 | 330 | 102277 | 40.489 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 9.322 | 172 | 907497 | 42.581 | ng | 0.00 |
| 79) Terphenyl-d14 | 13.074 | 244 | 680179 | 42.012 | ng | 0.00 |
| **Target Compounds** | | | | | | Qvalue |
| 2) 1,4-Dioxane | 2.810 | 88 | 98010 | 20.301 | ng | 99 |
| 3) Pyridine | 3.581 | 79 | 247677 | 20.259 | ng | 99 |
| 4) n-Nitrosodimethylamine | 3.516 | 42 | 126355 | 19.948 | ng | 100 |
| 6) Aniline | 6.628 | 93 | 349337 | 20.349 | ng | 100 |
| 8) 2-Chlorophenol | 6.751 | 128 | 202437 | 20.775 | ng | 98 |
| 9) Benzaldehyde | 6.516 | 77 | 176058 | 18.952 | ng | 99 |
| 10) Phenol | 6.592 | 94 | 273102m | 20.693 | ng | |
| 11) bis(2-Chloroethyl)ether | 6.692 | 93 | 209360 | 20.491 | ng | 98 |
| 12) 1,3-Dichlorobenzene | 6.910 | 146 | 225218 | 20.458 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 6.987 | 146 | 230409 | 20.804 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 7.139 | 146 | 219102 | 20.745 | ng | 99 |
| 15) Benzyl Alcohol | 7.098 | 79 | 184535 | 20.428 | ng | 98 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.239 | 45 | 387425 | 20.571 | ng | 100 |
| 17) 2-Methylphenol | 7.204 | 107 | 172219 | 20.480 | ng | 99 |
| 18) Hexachloroethane | 7.486 | 117 | 74466 | 20.654 | ng | 100 |
| 19) n-Nitroso-di-n-propyla... | 7.369 | 70 | 153555 | 20.559 | ng | 98 |
| 20) 3+4-Methylphenols | 7.357 | 107 | 217393 | 21.080 | ng | 91 |
| 22) Acetophenone | 7.369 | 105 | 289716 | 20.725 | ng | 98 |
| 24) Nitrobenzene | 7.539 | 77 | 206781 | 20.721 | ng | 98 |
| 25) Isophorone | 7.781 | 82 | 404382 | 20.239 | ng | 100 |
| 26) 2-Nitrophenol | 7.857 | 139 | 61544 | 18.099 | ng | 98 |
| 27) 2,4-Dimethylphenol | 7.886 | 122 | 188840 | 20.363 | ng | 99 |
| 28) bis(2-Chloroethoxy)met... | 7.986 | 93 | 252183 | 20.814 | ng | 99 |
| 29) 2,4-Dichlorophenol | 8.098 | 162 | 154978 | 20.427 | ng | 98 |
| 30) 1,2,4-Trichlorobenzene | 8.186 | 180 | 177770 | 20.647 | ng | 99 |
| 31) Naphthalene | 8.269 | 128 | 601316 | 20.881 | ng | 100 |
| 32) Benzoic acid | 7.963 | 122 | 64517m | 18.587 | ng | |
| 33) 4-Chloroaniline | 8.310 | 127 | 239869 | 20.714 | ng | 98 |
| 34) Hexachlorobutadiene | 8.392 | 225 | 106359 | 20.655 | ng | 100 |
| 35) Caprolactam | 8.657 | 113 | 45418 | 19.171 | ng | 98 |
| 36) 4-Chloro-3-methylphenol | 8.781 | 107 | 171615 | 20.307 | ng | 99 |
| 37) 2-Methylnaphthalene | 8.963 | 142 | 356378 | 20.744 | ng | 100 |
| 38) 1-Methylnaphthalene | 9.063 | 142 | 365700 | 20.524 | ng | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 9.128 | 216 | 170678 | 20.673 | ng | 98 |
| 41) Hexachlorocyclopentadiene | 9.122 | 237 | 86596 | 18.918 | ng | 98 |
| 43) 2,4,6-Trichlorophenol | 9.233 | 196 | 106780 | 20.815 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143096.D
Acq On : 15 Jul 2025 14:36
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

Quant Time: Jul 15 17:39:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 9.269 | 196 | 110466 | 20.537 | ng | 99 |
| 46) 1,1'-Biphenyl | 9.422 | 154 | 473419 | 20.955 | ng | 100 |
| 47) 2-Chloronaphthalene | 9.451 | 162 | 341636 | 20.584 | ng | 100 |
| 48) 2-Nitroaniline | 9.533 | 65 | 83805 | 18.969 | ng | 99 |
| 49) Acenaphthylene | 9.863 | 152 | 576265 | 20.837 | ng | 100 |
| 50) Dimethylphthalate | 9.716 | 163 | 366019 | 20.536 | ng | 100 |
| 51) 2,6-Dinitrotoluene | 9.775 | 165 | 62776 | 19.583 | ng | 96 |
| 52) Acenaphthene | 10.039 | 154 | 332427 | 20.369 | ng | 100 |
| 53) 3-Nitroaniline | 9.939 | 138 | 80094 | 20.797 | ng | 100 |
| 54) 2,4-Dinitrophenol | 10.045 | 184 | 15826 | 19.749 | ng | 71 |
| 55) Dibenzofuran | 10.210 | 168 | 505784 | 20.792 | ng | 100 |
| 56) 4-Nitrophenol | 10.086 | 139 | 57072 | 18.647 | ng | 98 |
| 57) 2,4-Dinitrotoluene | 10.175 | 165 | 71762 | 18.561 | ng | 95 |
| 58) Fluorene | 10.551 | 166 | 376169 | 20.626 | ng | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 10.322 | 232 | 85620 | 20.030 | ng | 98 |
| 60) Diethylphthalate | 10.422 | 149 | 356111 | 20.710 | ng | 100 |
| 61) 4-Chlorophenyl-phenyle... | 10.545 | 204 | 178641 | 20.436 | ng | 98 |
| 62) 4-Nitroaniline | 10.551 | 138 | 67264 | 20.084 | ng | 98 |
| 63) Azobenzene | 10.704 | 77 | 389419 | 20.478 | ng | 100 |
| 65) 4,6-Dinitro-2-methylph... | 10.580 | 198 | 23973 | 19.643 | ng | 94 |
| 66) n-Nitrosodiphenylamine | 10.657 | 169 | 332248 | 20.948 | ng | 98 |
| 67) 4-Bromophenyl-phenylether | 11.033 | 248 | 105446 | 20.535 | ng | 98 |
| 68) Hexachlorobenzene | 11.104 | 284 | 111727 | 20.507 | ng | 98 |
| 69) Atrazine | 11.180 | 200 | 83181 | 20.499 | ng | 99 |
| 70) Pentachlorophenol | 11.292 | 266 | 56380 | 18.862 | ng | 98 |
| 71) Phenanthrene | 11.516 | 178 | 501521 | 20.581 | ng | 99 |
| 72) Anthracene | 11.569 | 178 | 508525 | 20.752 | ng | 100 |
| 73) Carbazole | 11.716 | 167 | 455047 | 20.728 | ng | 99 |
| 74) Di-n-butylphthalate | 12.051 | 149 | 412206 | 20.456 | ng | 99 |
| 75) Fluoranthene | 12.704 | 202 | 455455 | 20.556 | ng | 100 |
| 77) Benzidine | 12.816 | 184 | 186115 | 23.083 | ng | 99 |
| 78) Pyrene | 12.933 | 202 | 462715 | 21.072 | ng | 99 |
| 80) Butylbenzylphthalate | 13.545 | 149 | 73439 | 18.875 | ng | 96 |
| 81) Benzo(a)anthracene | 14.115 | 228 | 332795 | 21.102 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 14.074 | 252 | 96174 | 21.000 | ng | 99 |
| 83) Chrysene | 14.157 | 228 | 286610 | 19.732 | ng | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 14.110 | 149 | 113349 | 19.236 | ng | 98 |
| 85) Di-n-octyl phthalate | 14.727 | 149 | 192002 | 19.817 | ng | 99 |
| 87) Indeno(1,2,3-cd)pyrene | 17.174 | 276 | 400422 | 20.516 | ng | 99 |
| 88) Benzo(b)fluoranthene | 15.186 | 252 | 307312 | 20.302 | ng | 99 |
| 89) Benzo(k)fluoranthene | 15.215 | 252 | 291618 | 20.194 | ng | 99 |
| 90) Benzo(a)pyrene | 15.568 | 252 | 290809 | 20.313 | ng | 99 |
| 91) Dibenzo(a,h)anthracene | 17.192 | 278 | 332152 | 20.818 | ng | 99 |
| 92) Benzo(g,h,i)perylene | 17.639 | 276 | 331447 | 20.376 | ng | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143096.D
Acq On : 15 Jul 2025 14:36
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jul 15 17:39:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.969 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:152 Resp: 15323

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 158.3 | 126.5 | 189.7 |
| 115 | 60.6 | 47.0 | 70.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#2
1,4-Dioxane
Concen: 20.301 ng
RT: 2.810 min Scan# 105
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 88 Resp: 98010

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 87.1 | 70.5 | 105.7 |
| 43 | 34.3 | 27.3 | 40.9 |






#3
Pyridine
Concen: 20.259 ng
RT: 3.581 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 79 Resp: 24767

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 87.3 | 69.2 | 103.8 |
| 51 | 42.8 | 33.3 | 49.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#4
n-Nitrosodimethylamine
Concen: 19.948 ng
RT: 3.516 min Scan# 225
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 42 Resp: 126355

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 116.0 | 92.8 | 139.2 |
| 44 | 7.5 | 5.8 | 8.6 |






#5
2-Fluorophenol
Concen: 41.549 ng
RT: 5.587 min Scan# 5
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:112 Resp: 40297

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 67.1 | 54.6 | 82.0 |
| 63 | 33.1 | 27.3 | 40.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#6
Aniline
Concen: 20.349 ng
RT: 6.628 min Scan# 754
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 93 Resp: 349337

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 41.3 | 33.1 | 49.7 |
| 65 | 20.9 | 16.9 | 25.3 |






#7
Phenol-d6
Concen: 41.231 ng
RT: 6.581 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 99 Resp: 50265

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 21.8 | 17.2 | 25.8 |
| 71 | 33.8 | 27.4 | 41.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#8
2-Chlorophenol
Concen: 20.775 ng
RT: 6.751 min Scan# 775
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:128 Resp: 202437

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.6 | 12.8 | 52.8 |
| 64 | 54.3 | 36.1 | 76.1 |






#9
Benzaldehyde
Concen: 18.952 ng
RT: 6.516 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 77 Resp: 17605

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 95.6 | 76.3 | 116.3 |
| 106 | 93.1 | 72.8 | 112.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#10
Phenol
Concen: 20.693 ng m
RT: 6.592 min Scan# 748
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 94 Resp: 273102

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 29.5 | 9.8 | 49.8 |
| 66 | 40.0 | 20.5 | 60.5 |






#11
bis(2-Chloroethyl)ether
Concen: 20.491 ng
RT: 6.692 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 93 Resp: 20936

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 84.1 | 61.9 | 101.9 |
| 95 | 32.2 | 12.2 | 52.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#12
1,3-Dichlorobenzene
Concen: 20.458 ng
RT: 6.910 min Scan# 802
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:146 Resp: 225218

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.7 | 50.1 | 75.1 |
| 75 | 34.5 | 27.4 | 41.0 |






#13
1,4-Dichlorobenzene
Concen: 20.804 ng
RT: 6.987 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:146 Resp: 23040

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.7 | 50.6 | 76.0 |
| 111 | 42.7 | 35.0 | 52.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#14
1,2-Dichlorobenzene
Concen: 20.745 ng
RT: 7.139 min Scan# 841
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:146 Resp: 219102

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.6 | 49.8 | 74.6 |
| 111 | 44.5 | 35.4 | 53.0 |






#15
Benzyl Alcohol
Concen: 20.428 ng
RT: 7.098 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 79 Resp: 18453

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 75.2 | 61.1 | 91.7 |
| 77 | 60.8 | 49.9 | 74.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 20.571 ng
RT: 7.239 min Scan# 858
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 45 Resp: 387425

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 11.1 | 0.0 | 31.1 |
| 79 | 8.4 | 0.0 | 28.4 |






#17
2-Methylphenol
Concen: 20.480 ng
RT: 7.204 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:107 Resp: 17221

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 114.5 | 90.9 | 136.3 |
| 77 | 44.9 | 37.5 | 56.3 |
| 79 | 45.1 | 36.9 | 55.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#18
Hexachloroethane
Concen: 20.654 ng
RT: 7.486 min Scan# 900
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:117 Resp: 74466

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 96.2 | 76.7 | 115.1 |
| 201 | 93.7 | 75.0 | 112.4 |






#19
n-Nitroso-di-n-propylamine
Concen: 20.559 ng
RT: 7.369 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 70 Resp: 15355

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 68.3 | 53.2 | 79.8 |
| 101 | 9.7 | 7.5 | 11.3 |
| 130 | 19.4 | 15.7 | 23.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#20
3+4-Methylphenols
Concen: 21.080 ng
RT: 7.357 min Scan# 878
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:107 Resp: 217393

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 91.8 | 71.3 | 111.3 |
| 77 | 78.6 | 77.2 | 117.2 |
| 79 | 29.8 | 10.7 | 50.7 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.251 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:136 Resp: 57990

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.7 | 8.9 | 13.3 |
| 54 | 10.2 | 8.3 | 12.5 |
| 68 | 6.8 | 5.3 | 7.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#22
Acetophenone
Concen: 20.725 ng
RT: 7.369 min Scan# 880
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:105 Resp: 289716

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 8.5 | 7.2 | 10.8 |
| 51 | 36.1 | 27.0 | 40.4 |
| 120 | 22.7 | 18.2 | 27.2 |






#23
Nitrobenzene-d5
Concen: 40.693 ng m
RT: 7.522 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 82 Resp: 42067

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 40.4 | 33.3 | 49.9 |
| 54 | 60.2 | 47.4 | 71.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#24
Nitrobenzene
Concen: 20.721 ng
RT: 7.539 min Scan# 909
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 77 Resp: 206781

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 42.0 | 34.5 | 51.7 |
| 65 | 13.5 | 11.3 | 16.9 |






#25
Isophorone
Concen: 20.239 ng
RT: 7.781 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 82 Resp: 40438

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.1 | 5.6 | 8.4 |
| 138 | 17.7 | 14.3 | 21.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#26
2-Nitrophenol
Concen: 18.099 ng
RT: 7.857 min Scan# 963
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:139 Resp: 61544

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 29.7 | 23.4 | 35.0 |
| 65 | 56.3 | 43.4 | 65.2 |






#27
2,4-Dimethylphenol
Concen: 20.363 ng
RT: 7.886 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:122 Resp: 18884

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 115.1 | 91.3 | 136.9 |
| 121 | 58.6 | 45.5 | 68.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#28
bis(2-Chloroethoxy)methane
Concen: 20.814 ng
RT: 7.986 min Scan# 985
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 93 Resp: 252183

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.4 | 26.2 | 39.4 |
| 123 | 11.1 | 9.6 | 14.4 |






#29
2,4-Dichlorophenol
Concen: 20.427 ng
RT: 8.098 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:162 Resp: 15497

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 65.2 | 43.1 | 83.1 |
| 98 | 40.0 | 19.2 | 59.2 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#30
1,2,4-Trichlorobenzene
Concen: 20.647 ng
RT: 8.186 min Scan# 1019
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:180 Resp: 177770

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 94.2 | 74.7 | 112.1 |
| 145 | 31.9 | 24.5 | 36.7 |






#31
Naphthalene
Concen: 20.881 ng
RT: 8.269 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:128 Resp: 60131

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.2 | 9.0 | 13.4 |
| 127 | 13.0 | 10.5 | 15.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#32
Benzoic acid
Concen: 18.587 ng m
RT: 7.963 min Scan# 981
Delta R.T. -0.024 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:122 Resp: 64517

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 130.0 | 105.4 | 145.4 |
| 77 | 100.9 | 78.6 | 118.6 |






#33
4-Chloroaniline
Concen: 20.714 ng
RT: 8.310 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:127 Resp: 23986

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 31.7 | 26.2 | 39.4 |
| 65 | 34.8 | 28.6 | 42.8 |
| 92 | 20.9 | 17.1 | 25.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#34
Hexachlorobutadiene
Concen: 20.655 ng
RT: 8.392 min Scan# 1054
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:225 Resp: 106359

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 63.3 | 50.3 | 75.5 |
| 227 | 62.9 | 50.1 | 75.1 |






#35
Caprolactam
Concen: 19.171 ng
RT: 8.657 min Scan# 1
Delta R.T. -0.012 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:113 Resp: 4541

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 203.9 | 186.2 | 226.2 |
| 56 | 150.9 | 134.1 | 174.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#36
4-Chloro-3-methylphenol
Concen: 20.307 ng
RT: 8.781 min Scan# 1120
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:107 Resp: 171615

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 24.6 | 20.2 | 30.2 |
| 142 | 77.0 | 62.4 | 93.6 |






#37
2-Methylnaphthalene
Concen: 20.744 ng
RT: 8.963 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:142 Resp: 35637

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 87.4 | 70.2 | 105.2 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#38
1-Methylnaphthalene
Concen: 20.524 ng
RT: 9.063 min Scan# 1168
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:142 Resp: 365700

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 91.4 | 72.2 | 108.4 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:164 Resp: 28330

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 98.6 | 79.0 | 118.6 |
| 160 | 44.2 | 35.8 | 53.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.673 ng
RT: 9.128 min Scan# 1179
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:216 Resp: 170678

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 79.7 | 62.6 | 94.0 |
| 179 | 21.1 | 16.2 | 24.2 |
| 108 | 20.3 | 15.8 | 23.6 |






#41
Hexachlorocyclopentadiene
Concen: 18.918 ng
RT: 9.122 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:237 Resp: 8659

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 63.3 | 41.6 | 81.6 |
| 272 | 13.4 | 0.0 | 32.9 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#42
2,4,6-Tribromophenol
Concen: 40.489 ng
RT: 10.792 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:330 Resp: 102277

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 96.8 | 77.9 | 116.9 |
| 141 | 32.4 | 25.9 | 38.9 |






#43
2,4,6-Trichlorophenol
Concen: 20.815 ng
RT: 9.233 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:196 Resp: 10678

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.5 | 76.9 | 115.3 |
| 200 | 30.6 | 23.5 | 35.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#44
2,4,5-Trichlorophenol
Concen: 20.537 ng
RT: 9.269 min Scan# 1203
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:196 Resp: 110466

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.0 | 76.4 | 114.6 |
| 97 | 52.0 | 42.4 | 63.6 |
| 132 | 30.1 | 23.9 | 35.9 |






#45
2-Fluorobiphenyl
Concen: 42.581 ng
RT: 9.322 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:172 Resp: 90749

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 34.6 | 28.2 | 42.4 |
| 170 | 22.9 | 18.6 | 28.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#46
1,1'-Biphenyl
Concen: 20.955 ng
RT: 9.422 min Scan# 1229
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:154 Resp: 473419

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.1 | 20.1 | 60.1 |
| 76 | 14.2 | 0.0 | 33.7 |






#47
2-Chloronaphthalene
Concen: 20.584 ng
RT: 9.451 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:162 Resp: 34163

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 41.2 | 33.0 | 49.6 |
| 164 | 32.8 | 26.1 | 39.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#48
2-Nitroaniline
Concen: 18.969 ng
RT: 9.533 min Scan# 1248
Delta R.T. 0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 65 Resp: 83805

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 63.1 | 51.3 | 76.9 |
| 138 | 95.8 | 77.8 | 116.6 |






#49
Acenaphthylene
Concen: 20.837 ng
RT: 9.863 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:152 Resp: 57626

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 19.7 | 15.8 | 23.8 |
| 153 | 13.1 | 10.4 | 15.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#50
Dimethylphthalate
Concen: 20.536 ng
RT: 9.716 min Scan# 1279
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:163 Resp: 366019

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.2 | 3.4 | 5.2 |
| 164 | 10.2 | 8.2 | 12.2 |






#51
2,6-Dinitrotoluene
Concen: 19.583 ng
RT: 9.775 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:165 Resp: 6277

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 68.1 | 57.4 | 86.0 |
| 89 | 61.7 | 51.8 | 77.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#52
Acenaphthene
Concen: 20.369 ng
RT: 10.039 min Scan# 1334
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:154 Resp: 332427

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 108.9 | 86.7 | 130.1 |
| 152 | 51.9 | 41.4 | 62.2 |






#53
3-Nitroaniline
Concen: 20.797 ng
RT: 9.939 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:138 Resp: 8009

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 13.1 | 10.1 | 15.1 |
| 92 | 120.9 | 96.4 | 144.6 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#54
2,4-Dinitrophenol
Concen: 19.749 ng
RT: 10.045 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:184 Resp: 15826

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 193.5 | 149.3 | 223.9 |
| 154 | 1488.1 | 1048.0 | 1572.0 |






#55
Dibenzofuran
Concen: 20.792 ng
RT: 10.210 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:168 Resp: 50578

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 37.3 | 29.8 | 44.6 |
| 169 | 13.3 | 10.8 | 16.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#56
4-Nitrophenol
Concen: 18.647 ng
RT: 10.086 min Scan# 1342
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:139 Resp: 57072

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 69.7 | 49.4 | 89.4 |
| 65 | 105.6 | 88.2 | 128.2 |






#57
2,4-Dinitrotoluene
Concen: 18.561 ng
RT: 10.175 min Scan# 1
Delta R.T. -0.005 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:165 Resp: 7176

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 53.8 | 40.6 | 61.0 |
| 89 | 82.6 | 62.3 | 93.5 |
| 182 | 3.2 | 2.6 | 3.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#58
Fluorene
Concen: 20.626 ng
RT: 10.551 min Scan# 1421
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:166 Resp: 376169

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 91.8 | 73.7 | 110.5 |
| 167 | 13.6 | 10.9 | 16.3 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 20.030 ng
RT: 10.322 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:232 Resp: 8562

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 46.3 | 36.9 | 55.3 |
| 130 | 2.1 | 1.8 | 2.8 |
| 166 | 31.3 | 23.0 | 34.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#60
Diethylphthalate
Concen: 20.710 ng
RT: 10.422 min Scan# 1399
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:149 Resp: 356111

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 21.7 | 17.3 | 25.9 |
| 150 | 12.1 | 9.9 | 14.9 |






#61
4-Chlorophenyl-phenylether
Concen: 20.436 ng
RT: 10.545 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:204 Resp: 17864

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 33.3 | 25.8 | 38.6 |
| 141 | 59.8 | 46.8 | 70.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#62
4-Nitroaniline
Concen: 20.084 ng
RT: 10.551 min Scan# 1421
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:138 Resp: 67264

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 53.2 | 30.4 | 70.4 |
| 108 | 70.4 | 49.5 | 89.5 |






#63
Azobenzene
Concen: 20.478 ng
RT: 10.704 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion: 77 Resp: 38941

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 24.3 | 4.0 | 44.0 |
| 105 | 21.8 | 1.4 | 41.4 |
| 51 | 34.3 | 14.3 | 54.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.492 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:188 Resp: 451524

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 10.5 | 8.6 | 13.0 |
| 80 | 11.6 | 9.3 | 13.9 |






#65
4,6-Dinitro-2-methylphenol
Concen: 19.643 ng
RT: 10.580 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:198 Resp: 2397

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 69.3 | 44.6 | 84.6 |
| 105 | 52.7 | 29.3 | 69.3 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#66
n-Nitrosodiphenylamine
Concen: 20.948 ng
RT: 10.657 min Scan# 1439
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:169 Resp: 332248

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 64.9 | 53.0 | 79.6 |
| 167 | 34.8 | 28.4 | 42.6 |






#67
4-Bromophenyl-phenylether
Concen: 20.535 ng
RT: 11.033 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:248 Resp: 10544

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 99.2 | 76.7 | 115.1 |
| 141 | 76.9 | 61.7 | 92.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#68
Hexachlorobenzene
Concen: 20.507 ng
RT: 11.104 min Scan# 1515
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:284 Resp: 111727

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 34.2 | 28.9 | 43.3 |
| 249 | 30.4 | 24.6 | 37.0 |






#69
Atrazine
Concen: 20.499 ng
RT: 11.180 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:200 Resp: 8318

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 27.8 | 9.1 | 49.1 |
| 215 | 50.3 | 31.0 | 71.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#70
Pentachlorophenol
Concen: 18.862 ng
RT: 11.292 min Scan# 1547
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:266 Resp: 56380

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 66.2 | 51.8 | 77.6 |
| 264 | 61.0 | 49.7 | 74.5 |






#71
Phenanthrene
Concen: 20.581 ng
RT: 11.516 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:178 Resp: 50152

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.4 | 15.1 | 22.7 |
| 179 | 15.1 | 12.2 | 18.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#72
Anthracene
Concen: 20.752 ng
RT: 11.569 min Scan# 1594
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:178 Resp: 508525

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.5 | 14.7 | 22.1 |
| 179 | 15.3 | 12.4 | 18.6 |






#73
Carbazole
Concen: 20.728 ng
RT: 11.716 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:167 Resp: 45504

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.0 | 17.0 | 25.4 |
| 139 | 12.3 | 10.2 | 15.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#74
Di-n-butylphthalate
Concen: 20.456 ng
RT: 12.051 min Scan# 1676
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:149 Resp: 412206

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 8.8 | 7.3 | 10.9 |
| 104 | 4.9 | 3.8 | 5.8 |






#75
Fluoranthene
Concen: 20.556 ng
RT: 12.704 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:202 Resp: 45545

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 12.4 | 0.0 | 32.4 |
| 203 | 17.2 | 0.0 | 37.4 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.127 min Scan# 2029
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:240 Resp: 236667

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 12.4 | 10.0 | 15.0 |
| 236 | 24.9 | 20.0 | 30.0 |






#77
Benzidine
Concen: 23.083 ng
RT: 12.816 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:184 Resp: 18611

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 16.1 | 12.4 | 18.6 |
| 183 | 10.9 | 8.6 | 12.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#78
Pyrene
Concen: 21.072 ng
RT: 12.933 min Scan# 1826
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:202 Resp: 462715

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 19.3 | 15.5 | 23.3 |
| 203 | 17.1 | 14.0 | 21.0 |






#79
Terphenyl-d14
Concen: 42.012 ng
RT: 13.074 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:244 Resp: 68017

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.3 | 5.2 | 7.8 |
| 122 | 11.0 | 8.8 | 13.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#80
Butylbenzylphthalate
Concen: 18.875 ng
RT: 13.545 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:149 Resp: 73439

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 74.6 | 57.0 | 85.4 |
| 206 | 19.6 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 21.102 ng
RT: 14.115 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:228 Resp: 33279

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 26.4 | 21.3 | 31.9 |
| 229 | 19.6 | 15.5 | 23.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#82
3,3'-Dichlorobenzidine
Concen: 21.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:252 Resp: 96174

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 64.3 | 51.5 | 77.3 |
| 126 | 12.7 | 9.4 | 14.0 |






#83
Chrysene
Concen: 19.732 ng
RT: 14.157 min Scan# 2(
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:228 Resp: 28661(

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 29.3 | 23.4 | 35.2 |
| 229 | 19.2 | 15.3 | 22.9 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#84
Bis(2-ethylhexyl)phthalate
Concen: 19.236 ng
RT: 14.110 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:149 Resp: 113349

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 28.8 | 21.9 | 32.9 |
| 279 | 4.0 | 3.4 | 5.0 |






#85
Di-n-octyl phthalate
Concen: 19.817 ng
RT: 14.727 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:149 Resp: 19200

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.5 | 1.2 | 1.8 |
| 43 | 13.7 | 10.6 | 16.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.633 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:264 Resp: 262419

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 22.9 | 18.5 | 27.7 |
| 265 | 21.0 | 17.5 | 26.3 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 20.516 ng
RT: 17.174 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:276 Resp: 40042

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 29.1 | 22.6 | 34.0 |
| 277 | 24.8 | 20.2 | 30.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#88
Benzo(b)fluoranthene
Concen: 20.302 ng
RT: 15.186 min Scan# 2209
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:252 Resp: 307312

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.7 | 17.5 | 26.3 |
| 125 | 11.3 | 8.6 | 13.0 |






#89
Benzo(k)fluoranthene
Concen: 20.194 ng
RT: 15.215 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:252 Resp: 29161

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.7 | 17.4 | 26.2 |
| 125 | 11.1 | 8.5 | 12.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#90
Benzo(a)pyrene
Concen: 20.313 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:252 Resp: 290809

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.4 | 17.5 | 26.3 |
| 125 | 11.7 | 9.4 | 14.0 |






#91
Dibenzo(a,h)anthracene
Concen: 20.818 ng
RT: 17.192 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:278 Resp: 33215

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 17.7 | 14.6 | 22.0 |
| 279 | 23.1 | 18.8 | 28.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#92
Benzo(g,h,i)perylene
Concen: 20.376 ng
RT: 17.639 min Scan# 2626
Delta R.T. -0.006 min
Lab File: BF143096.D
Acq: 15 Jul 2025 14:36

Tgt Ion:276 Resp: 331447

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.0 | 19.2 | 28.8 |
| 138 | 24.3 | 19.8 | 29.6 |



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143097.D
Acq On : 15 Jul 2025 15:05
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jul 15 17:40:13 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| **Internal Standards** | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 6.969 | 152 | 123440 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 8.251 | 136 | 466428 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 10.004 | 164 | 227730 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 11.492 | 188 | 369241 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 14.127 | 240 | 192349 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 15.639 | 264 | 226517 | 20.000 | ng | 0.00 |
| **System Monitoring Compounds** | | | | | | |
| 5) 2-Fluorophenol | 5.587 | 112 | 619840 | 79.335 | ng | 0.00 |
| 7) Phenol-d6 | 6.587 | 99 | 772641 | 78.673 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 7.528 | 82 | 694498m | 83.526 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 10.792 | 330 | 170969 | 84.198 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 9.328 | 172 | 1326482 | 77.428 | ng | 0.00 |
| 79) Terphenyl-d14 | 13.074 | 244 | 1033540 | 78.547 | ng | 0.00 |
| **Target Compounds** | | | | | | Qvalue |
| 2) 1,4-Dioxane | 2.810 | 88 | 154739 | 39.787 | ng | 100 |
| 3) Pyridine | 3.581 | 79 | 393514 | 39.957 | ng | 100 |
| 4) n-Nitrosodimethylamine | 3.522 | 42 | 201552 | 39.500 | ng | 100 |
| 6) Aniline | 6.628 | 93 | 544405 | 39.365 | ng | 100 |
| 8) 2-Chlorophenol | 6.751 | 128 | 314192 | 40.025 | ng | 100 |
| 9) Benzaldehyde | 6.516 | 77 | 358216 | 47.869 | ng | 100 |
| 10) Phenol | 6.598 | 94 | 421819m | 39.676 | ng | |
| 11) bis(2-Chloroethyl)ether | 6.698 | 93 | 323129 | 39.260 | ng | 100 |
| 12) 1,3-Dichlorobenzene | 6.910 | 146 | 349154 | 39.371 | ng | 100 |
| 13) 1,4-Dichlorobenzene | 6.987 | 146 | 348706 | 39.084 | ng | 100 |
| 14) 1,2-Dichlorobenzene | 7.140 | 146 | 336818 | 39.587 | ng | 100 |
| 15) Benzyl Alcohol | 7.098 | 79 | 288056 | 39.585 | ng | 100 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.239 | 45 | 593735 | 39.134 | ng | 100 |
| 17) 2-Methylphenol | 7.204 | 107 | 267472 | 39.484 | ng | 100 |
| 18) Hexachloroethane | 7.487 | 117 | 116640 | 40.159 | ng | 100 |
| 19) n-Nitroso-di-n-propyla... | 7.375 | 70 | 237958 | 39.549 | ng | 100 |
| 20) 3+4-Methylphenols | 7.363 | 107 | 334719 | 40.291 | ng | 100 |
| 22) Acetophenone | 7.375 | 105 | 437447 | 38.906 | ng | 100 |
| 24) Nitrobenzene | 7.545 | 77 | 332013 | 41.364 | ng | 100 |
| 25) Isophorone | 7.787 | 82 | 632327 | 39.346 | ng | 100 |
| 26) 2-Nitrophenol | 7.863 | 139 | 112314 | 36.079 | ng | 100 |
| 27) 2,4-Dimethylphenol | 7.892 | 122 | 298241 | 39.984 | ng | 100 |
| 28) bis(2-Chloroethoxy)met... | 7.992 | 93 | 381576 | 39.155 | ng | 100 |
| 29) 2,4-Dichlorophenol | 8.098 | 162 | 248676 | 40.752 | ng | 100 |
| 30) 1,2,4-Trichlorobenzene | 8.192 | 180 | 274944 | 39.701 | ng | 100 |
| 31) Naphthalene | 8.275 | 128 | 900967 | 38.898 | ng | 100 |
| 32) Benzoic acid | 7.987 | 122 | 133968m | 37.684 | ng | |
| 33) 4-Chloroaniline | 8.310 | 127 | 371249 | 39.858 | ng | 100 |
| 34) Hexachlorobutadiene | 8.392 | 225 | 165576 | 39.978 | ng | 100 |
| 35) Caprolactam | 8.669 | 113 | 77472 | 40.657 | ng | 100 |
| 36) 4-Chloro-3-methylphenol | 8.786 | 107 | 272885 | 40.145 | ng | 100 |
| 37) 2-Methylnaphthalene | 8.963 | 142 | 540844 | 39.141 | ng | 100 |
| 38) 1-Methylnaphthalene | 9.063 | 142 | 562696 | 39.263 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 9.128 | 216 | 263160 | 39.653 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 9.122 | 237 | 147221 | 40.011 | ng | 100 |
| 43) 2,4,6-Trichlorophenol | 9.233 | 196 | 171419 | 41.569 | ng | 100 |

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143097.D
Acq On : 15 Jul 2025 15:05
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jul 15 17:40:13 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 9.269 | 196 | 179101 | 41.422 | ng | 100 |
| 46) 1,1'-Biphenyl | 9.428 | 154 | 713513 | 39.288 | ng | 100 |
| 47) 2-Chloronaphthalene | 9.451 | 162 | 527800 | 39.561 | ng | 100 |
| 48) 2-Nitroaniline | 9.533 | 65 | 149322 | 38.182 | ng | 100 |
| 49) Acenaphthylene | 9.869 | 152 | 892054 | 40.126 | ng | 100 |
| 50) Dimethylphthalate | 9.722 | 163 | 572658 | 39.971 | ng | 100 |
| 51) 2,6-Dinitrotoluene | 9.775 | 165 | 106468 | 38.213 | ng | 100 |
| 52) Acenaphthene | 10.039 | 154 | 513370 | 39.131 | ng | 100 |
| 53) 3-Nitroaniline | 9.945 | 138 | 130953 | 42.301 | ng | 100 |
| 54) 2,4-Dinitrophenol | 10.045 | 184 | 31236 | 37.582 | ng | 100 |
| 55) Dibenzofuran | 10.210 | 168 | 772862 | 39.524 | ng | 100 |
| 56) 4-Nitrophenol | 10.086 | 139 | 101948 | 41.436 | ng | 100 |
| 57) 2,4-Dinitrotoluene | 10.180 | 165 | 134243 | 38.682 | ng | 100 |
| 58) Fluorene | 10.551 | 166 | 567511 | 38.710 | ng | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 10.322 | 232 | 142027 | 41.334 | ng | 100 |
| 60) Diethylphthalate | 10.422 | 149 | 557712 | 40.349 | ng | 100 |
| 61) 4-Chlorophenyl-phenyle... | 10.545 | 204 | 279817 | 39.822 | ng | 100 |
| 62) 4-Nitroaniline | 10.557 | 138 | 113355 | 42.106 | ng | 100 |
| 63) Azobenzene | 10.704 | 77 | 604710 | 39.558 | ng | 100 |
| 65) 4,6-Dinitro-2-methylph... | 10.586 | 198 | 47349 | 37.223 | ng | 100 |
| 66) n-Nitrosodiphenylamine | 10.657 | 169 | 503590 | 38.827 | ng | 100 |
| 67) 4-Bromophenyl-phenylether | 11.033 | 248 | 164133 | 39.088 | ng | 100 |
| 68) Hexachlorobenzene | 11.104 | 284 | 173911 | 39.034 | ng | 100 |
| 69) Atrazine | 11.180 | 200 | 133416 | 40.206 | ng | 100 |
| 70) Pentachlorophenol | 11.292 | 266 | 97625 | 39.938 | ng | 100 |
| 71) Phenanthrene | 11.516 | 178 | 769372 | 38.609 | ng | 100 |
| 72) Anthracene | 11.569 | 178 | 780754 | 38.961 | ng | 100 |
| 73) Carbazole | 11.716 | 167 | 692468 | 38.572 | ng | 100 |
| 74) Di-n-butylphthalate | 12.051 | 149 | 680394 | 41.288 | ng | 100 |
| 75) Fluoranthene | 12.704 | 202 | 708736 | 39.115 | ng | 100 |
| 77) Benzidine | 12.816 | 184 | 321029 | 48.989 | ng | 100 |
| 78) Pyrene | 12.933 | 202 | 708631 | 39.706 | ng | 100 |
| 80) Butylbenzylphthalate | 13.551 | 149 | 132061 | 37.625 | ng | 100 |
| 81) Benzo(a)anthracene | 14.121 | 228 | 518137 | 40.424 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 14.080 | 252 | 148294 | 39.841 | ng | 100 |
| 83) Chrysene | 14.157 | 228 | 461979 | 39.134 | ng | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 14.110 | 149 | 198176 | 37.454 | ng | 100 |
| 85) Di-n-octyl phthalate | 14.727 | 149 | 354747 | 36.531 | ng | 100 |
| 87) Indeno(1,2,3-cd)pyrene | 17.180 | 276 | 689277 | 40.912 | ng | 100 |
| 88) Benzo(b)fluoranthene | 15.192 | 252 | 490240 | 37.520 | ng | 100 |
| 89) Benzo(k)fluoranthene | 15.221 | 252 | 520058 | 41.721 | ng | 100 |
| 90) Benzo(a)pyrene | 15.574 | 252 | 509307 | 41.215 | ng | 100 |
| 91) Dibenzo(a,h)anthracene | 17.198 | 278 | 561661 | 40.782 | ng | 100 |
| 92) Benzo(g,h,i)perylene | 17.645 | 276 | 569850 | 40.584 | ng | 100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143097.D
Acq On : 15 Jul 2025 15:05
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jul 15 17:40:13 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.969 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:152 Resp: 12344

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 158.1 | 126.5 | 189.7 |
| 115 | 58.8 | 47.0 | 70.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#2
1,4-Dioxane
Concen: 39.787 ng
RT: 2.810 min Scan# 105
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 88 Resp: 154739

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 88.1 | 70.5 | 105.7 |
| 43 | 34.1 | 27.3 | 40.9 |






#3
Pyridine
Concen: 39.957 ng
RT: 3.581 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 79 Resp: 39351

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 86.5 | 69.2 | 103.8 |
| 51 | 41.6 | 33.3 | 49.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#4
n-Nitrosodimethylamine
Concen: 39.500 ng
RT: 3.522 min Scan# 226
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 42 Resp: 201552

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 116.0 | 92.8 | 139.2 |
| 44 | 7.2 | 5.8 | 8.6 |






#5
2-Fluorophenol
Concen: 79.335 ng
RT: 5.587 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:112 Resp: 619840

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 68.3 | 54.6 | 82.0 |
| 63 | 34.1 | 27.3 | 40.9 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#6
Aniline
Concen: 39.365 ng
RT: 6.628 min Scan# 754
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 93 Resp: 544405

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 41.4 | 33.1 | 49.7 |
| 65 | 21.1 | 16.9 | 25.3 |






#7
Phenol-d6
Concen: 78.673 ng
RT: 6.587 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 99 Resp: 77264

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 21.5 | 17.2 | 25.8 |
| 71 | 34.2 | 27.4 | 41.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#8
2-Chlorophenol
Concen: 40.025 ng
RT: 6.751 min Scan# 775
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:128 Resp: 314192

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.8 | 12.8 | 52.8 |
| 64 | 56.1 | 36.1 | 76.1 |






#9
Benzaldehyde
Concen: 47.869 ng
RT: 6.516 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 77 Resp: 35821

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 96.3 | 76.3 | 116.3 |
| 106 | 92.8 | 72.8 | 112.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#10
Phenol
Concen: 39.676 ng m
RT: 6.598 min Scan# 749
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 94 Resp: 421819

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 29.8 | 9.8 | 49.8 |
| 66 | 40.5 | 20.5 | 60.5 |






#11
bis(2-Chloroethyl)ether
Concen: 39.260 ng
RT: 6.698 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 93 Resp: 32312

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 81.9 | 61.9 | 101.9 |
| 95 | 32.2 | 12.2 | 52.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#12
1,3-Dichlorobenzene
Concen: 39.371 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:146 Resp: 349154

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.6 | 50.1 | 75.1 |
| 75 | 34.2 | 27.4 | 41.0 |






#13
1,4-Dichlorobenzene
Concen: 39.084 ng
RT: 6.987 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:146 Resp: 34870

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.3 | 50.6 | 76.0 |
| 111 | 43.8 | 35.0 | 52.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#14
1,2-Dichlorobenzene
Concen: 39.587 ng
RT: 7.140 min Scan# 841
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:146 Resp: 336818

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.2 | 49.8 | 74.6 |
| 111 | 44.2 | 35.4 | 53.0 |






#15
Benzyl Alcohol
Concen: 39.585 ng
RT: 7.098 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 79 Resp: 28805

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 76.4 | 61.1 | 91.7 |
| 77 | 62.4 | 49.9 | 74.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 39.134 ng
RT: 7.239 min Scan# 858
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 45 Resp: 593735

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 11.1 | 0.0 | 31.1 |
| 79 | 8.4 | 0.0 | 28.4 |






#17
2-Methylphenol
Concen: 39.484 ng
RT: 7.204 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:107 Resp: 26747

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 113.6 | 90.9 | 136.3 |
| 77 | 46.9 | 37.5 | 56.3 |
| 79 | 46.1 | 36.9 | 55.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#18
Hexachloroethane
Concen: 40.159 ng
RT: 7.487 min Scan# 900
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:117 Resp: 116640

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 95.9 | 76.7 | 115.1 |
| 201 | 93.7 | 75.0 | 112.4 |






#19
n-Nitroso-di-n-propylamine
Concen: 39.549 ng
RT: 7.375 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 70 Resp: 23795

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 66.5 | 53.2 | 79.8 |
| 101 | 9.4 | 7.5 | 11.3 |
| 130 | 19.6 | 15.7 | 23.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#20
3+4-Methylphenols
Concen: 40.291 ng
RT: 7.363 min Scan# 879
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:107 Resp: 334719

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 91.3 | 71.3 | 111.3 |
| 77 | 97.2 | 77.2 | 117.2 |
| 79 | 30.7 | 10.7 | 50.7 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.251 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:136 Resp: 46642

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 11.1 | 8.9 | 13.3 |
| 54 | 10.4 | 8.3 | 12.5 |
| 68 | 6.6 | 5.3 | 7.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#22
Acetophenone
Concen: 38.906 ng
RT: 7.375 min Scan# 881
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:105 Resp: 437447

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 9.0 | 7.2 | 10.8 |
| 51 | 33.7 | 27.0 | 40.4 |
| 120 | 22.7 | 18.2 | 27.2 |






#23
Nitrobenzene-d5
Concen: 83.526 ng m
RT: 7.528 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 82 Resp: 69449

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 41.6 | 33.3 | 49.9 |
| 54 | 59.3 | 47.4 | 71.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#24
Nitrobenzene
Concen: 41.364 ng
RT: 7.545 min Scan# 910
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 77 Resp: 332013

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 43.1 | 34.5 | 51.7 |
| 65 | 14.1 | 11.3 | 16.9 |






#25
Isophorone
Concen: 39.346 ng
RT: 7.787 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 82 Resp: 63232

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.0 | 5.6 | 8.4 |
| 138 | 17.9 | 14.3 | 21.5 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#26
2-Nitrophenol
Concen: 36.079 ng
RT: 7.863 min Scan# 964
Delta R.T. -0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:139 Resp: 112314

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 29.2 | 23.4 | 35.0 |
| 65 | 54.3 | 43.4 | 65.2 |






#27
2,4-Dimethylphenol
Concen: 39.984 ng
RT: 7.892 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:122 Resp: 29824

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 114.1 | 91.3 | 136.9 |
| 121 | 56.9 | 45.5 | 68.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#28
bis(2-Chloroethoxy)methane
Concen: 39.155 ng
RT: 7.992 min Scan# 986
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 93 Resp: 381576

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.8 | 26.2 | 39.4 |
| 123 | 12.0 | 9.6 | 14.4 |






#29
2,4-Dichlorophenol
Concen: 40.752 ng
RT: 8.098 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:162 Resp: 24867

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 63.1 | 43.1 | 83.1 |
| 98 | 39.2 | 19.2 | 59.2 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#30
1,2,4-Trichlorobenzene
Concen: 39.701 ng
RT: 8.192 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:180 Resp: 274944

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 93.4 | 74.7 | 112.1 |
| 145 | 30.6 | 24.5 | 36.7 |






#31
Naphthalene
Concen: 38.898 ng
RT: 8.275 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:128 Resp: 90096

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.2 | 9.0 | 13.4 |
| 127 | 13.1 | 10.5 | 15.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#32
Benzoic acid
Concen: 37.684 ng m
RT: 7.987 min Scan# 985
Delta R.T. -0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:122 Resp: 133968

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 125.4 | 105.4 | 145.4 |
| 77 | 98.6 | 78.6 | 118.6 |






#33
4-Chloroaniline
Concen: 39.858 ng
RT: 8.310 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:127 Resp: 37124

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.8 | 26.2 | 39.4 |
| 65 | 35.7 | 28.6 | 42.8 |
| 92 | 21.4 | 17.1 | 25.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#34
Hexachlorobutadiene
Concen: 39.978 ng
RT: 8.392 min Scan# 1054
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:225 Resp: 165576

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 62.9 | 50.3 | 75.5 |
| 227 | 62.6 | 50.1 | 75.1 |






#35
Caprolactam
Concen: 40.657 ng
RT: 8.669 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:113 Resp: 7747

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 206.2 | 186.2 | 226.2 |
| 56 | 154.1 | 134.1 | 174.1 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#36
4-Chloro-3-methylphenol
Concen: 40.145 ng
RT: 8.786 min Scan# 1121
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:107 Resp: 272885

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 25.2 | 20.2 | 30.2 |
| 142 | 78.0 | 62.4 | 93.6 |






#37
2-Methylnaphthalene
Concen: 39.141 ng
RT: 8.963 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:142 Resp: 54084

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 87.7 | 70.2 | 105.2 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#38
1-Methylnaphthalene
Concen: 39.263 ng
RT: 9.063 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:142 Resp: 562696

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 90.3 | 72.2 | 108.4 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:164 Resp: 22773

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 98.8 | 79.0 | 118.6 |
| 160 | 44.7 | 35.8 | 53.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 39.653 ng
RT: 9.128 min Scan# 1179
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:216 Resp: 263160

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.3 | 62.6 | 94.0 |
| 179 | 20.2 | 16.2 | 24.2 |
| 108 | 19.7 | 15.8 | 23.6 |






#41
Hexachlorocyclopentadiene
Concen: 40.011 ng
RT: 9.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:237 Resp: 14722

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 61.6 | 41.6 | 81.6 |
| 272 | 12.9 | 0.0 | 32.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#42
2,4,6-Tribromophenol
Concen: 84.198 ng
RT: 10.792 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:330 Resp: 170969

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 97.4 | 77.9 | 116.9 |
| 141 | 32.4 | 25.9 | 38.9 |






#43
2,4,6-Trichlorophenol
Concen: 41.569 ng
RT: 9.233 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:196 Resp: 17141

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.1 | 76.9 | 115.3 |
| 200 | 29.4 | 23.5 | 35.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#44
2,4,5-Trichlorophenol
Concen: 41.422 ng
RT: 9.269 min Scan# 1203
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:196 Resp: 179101

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.5 | 76.4 | 114.6 |
| 97 | 53.0 | 42.4 | 63.6 |
| 132 | 29.9 | 23.9 | 35.9 |






#45
2-Fluorobiphenyl
Concen: 77.428 ng
RT: 9.328 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:172 Resp: 132648

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.3 | 28.2 | 42.4 |
| 170 | 23.3 | 18.6 | 28.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#46
1,1'-Biphenyl
Concen: 39.288 ng
RT: 9.428 min Scan# 1230
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:154 Resp: 713513

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.1 | 20.1 | 60.1 |
| 76 | 13.7 | 0.0 | 33.7 |






#47
2-Chloronaphthalene
Concen: 39.561 ng
RT: 9.451 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:162 Resp: 52780

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 41.3 | 33.0 | 49.6 |
| 164 | 32.6 | 26.1 | 39.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#48
2-Nitroaniline
Concen: 38.182 ng
RT: 9.533 min Scan# 1248
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 65 Resp: 149322

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 64.1 | 51.3 | 76.9 |
| 138 | 97.2 | 77.8 | 116.6 |






#49
Acenaphthylene
Concen: 40.126 ng
RT: 9.869 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:152 Resp: 89205

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 19.8 | 15.8 | 23.8 |
| 153 | 13.0 | 10.4 | 15.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#50
Dimethylphthalate
Concen: 39.971 ng
RT: 9.722 min Scan# 1280
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:163 Resp: 572658

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.3 | 3.4 | 5.2 |
| 164 | 10.2 | 8.2 | 12.2 |






#51
2,6-Dinitrotoluene
Concen: 38.213 ng
RT: 9.775 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:165 Resp: 10646

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 71.7 | 57.4 | 86.0 |
| 89 | 64.8 | 51.8 | 77.8 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#52
Acenaphthene
Concen: 39.131 ng
RT: 10.039 min Scan# 1334
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:154 Resp: 513370

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 108.4 | 86.7 | 130.1 |
| 152 | 51.8 | 41.4 | 62.2 |






#53
3-Nitroaniline
Concen: 42.301 ng
RT: 9.945 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:138 Resp: 13095

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 12.6 | 10.1 | 15.1 |
| 92 | 120.5 | 96.4 | 144.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#54
2,4-Dinitrophenol
Concen: 37.582 ng
RT: 10.045 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:184 Resp: 31236

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 186.6 | 149.3 | 223.9 |
| 154 | 1310.0 | 1048.0 | 1572.0 |






#55
Dibenzofuran
Concen: 39.524 ng
RT: 10.210 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:168 Resp: 77286

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 37.2 | 29.8 | 44.6 |
| 169 | 13.5 | 10.8 | 16.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#56
4-Nitrophenol
Concen: 41.436 ng
RT: 10.086 min Scan# 1342
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:139 Resp: 101948

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 69.4 | 49.4 | 89.4 |
| 65 | 108.2 | 88.2 | 128.2 |






#57
2,4-Dinitrotoluene
Concen: 38.682 ng
RT: 10.180 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:165 Resp: 13424

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 50.8 | 40.6 | 61.0 |
| 89 | 77.9 | 62.3 | 93.5 |
| 182 | 3.2 | 2.6 | 3.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#58
Fluorene
Concen: 38.710 ng
RT: 10.551 min Scan# 1421
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:166 Resp: 567511

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 92.1 | 73.7 | 110.5 |
| 167 | 13.6 | 10.9 | 16.3 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 41.334 ng
RT: 10.322 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:232 Resp: 14202

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 46.1 | 36.9 | 55.3 |
| 130 | 2.3 | 1.8 | 2.8 |
| 166 | 28.8 | 23.0 | 34.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#60
Diethylphthalate
Concen: 40.349 ng
RT: 10.422 min Scan# 1399
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:149 Resp: 557712

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 21.6 | 17.3 | 25.9 |
| 150 | 12.4 | 9.9 | 14.9 |






#61
4-Chlorophenyl-phenylether
Concen: 39.822 ng
RT: 10.545 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:204 Resp: 27981

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 32.2 | 25.8 | 38.6 |
| 141 | 58.5 | 46.8 | 70.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#62
4-Nitroaniline
Concen: 42.106 ng
RT: 10.557 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:138 Resp: 113355

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 50.4 | 30.4 | 70.4 |
| 108 | 69.5 | 49.5 | 89.5 |






#63
Azobenzene
Concen: 39.558 ng
RT: 10.704 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion: 77 Resp: 60471

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 24.0 | 4.0 | 44.0 |
| 105 | 21.4 | 1.4 | 41.4 |
| 51 | 34.3 | 14.3 | 54.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.492 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:188 Resp: 369241

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 10.8 | 8.6 | 13.0 |
| 80 | 11.6 | 9.3 | 13.9 |






#65
4,6-Dinitro-2-methylphenol
Concen: 37.223 ng
RT: 10.586 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:198 Resp: 4734

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 64.6 | 44.6 | 84.6 |
| 105 | 49.3 | 29.3 | 69.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#66
n-Nitrosodiphenylamine
Concen: 38.827 ng
RT: 10.657 min Scan# 1439
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:169 Resp: 503590

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 66.3 | 53.0 | 79.6 |
| 167 | 35.5 | 28.4 | 42.6 |






#67
4-Bromophenyl-phenylether
Concen: 39.088 ng
RT: 11.033 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:248 Resp: 16413

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 95.9 | 76.7 | 115.1 |
| 141 | 77.1 | 61.7 | 92.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#68
Hexachlorobenzene
Concen: 39.034 ng
RT: 11.104 min Scan# 1515
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:284 Resp: 173911

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 36.1 | 28.9 | 43.3 |
| 249 | 30.8 | 24.6 | 37.0 |






#69
Atrazine
Concen: 40.206 ng
RT: 11.180 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:200 Resp: 13341

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 29.1 | 9.1 | 49.1 |
| 215 | 51.0 | 31.0 | 71.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#70
Pentachlorophenol
Concen: 39.938 ng
RT: 11.292 min Scan# 1547
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:266 Resp: 97625

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 64.7 | 51.8 | 77.6 |
| 264 | 62.1 | 49.7 | 74.5 |






#71
Phenanthrene
Concen: 38.609 ng
RT: 11.516 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:178 Resp: 76937

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.9 | 15.1 | 22.7 |
| 179 | 15.3 | 12.2 | 18.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#72
Anthracene
Concen: 38.961 ng
RT: 11.569 min Scan# 1594
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:178 Resp: 780754

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.4 | 14.7 | 22.1 |
| 179 | 15.5 | 12.4 | 18.6 |






#73
Carbazole
Concen: 38.572 ng
RT: 11.716 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:167 Resp: 69246

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.2 | 17.0 | 25.4 |
| 139 | 12.8 | 10.2 | 15.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#74
Di-n-butylphthalate
Concen: 41.288 ng
RT: 12.051 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:149 Resp: 680394

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.1 | 7.3 | 10.9 |
| 104 | 4.8 | 3.8 | 5.8 |






#75
Fluoranthene
Concen: 39.115 ng
RT: 12.704 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

Tgt Ion:202 Resp: 70873

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 12.4 | 0.0 | 32.4 |
| 203 | 17.4 | 0.0 | 37.4 |

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.127 min Scan# 2029
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:240 Resp: 192349

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 12.5 | 10.0 | 15.0 |
| 236 | 25.0 | 20.0 | 30.0 |






#77
Benzidine
Concen: 48.989 ng
RT: 12.816 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:184 Resp: 32102

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 15.5 | 12.4 | 18.6 |
| 183 | 10.7 | 8.6 | 12.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#78
Pyrene
Concen: 39.706 ng
RT: 12.933 min Scan# 1826
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:202 Resp: 708631

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 19.4 | 15.5 | 23.3 |
| 203 | 17.5 | 14.0 | 21.0 |






#79
Terphenyl-d14
Concen: 78.547 ng
RT: 13.074 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:244 Resp: 103354

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.5 | 5.2 | 7.8 |
| 122 | 11.0 | 8.8 | 13.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#80
Butylbenzylphthalate
Concen: 37.625 ng
RT: 13.551 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:149 Resp: 132061

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 71.2 | 57.0 | 85.4 |
| 206 | 21.5 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 40.424 ng
RT: 14.121 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:228 Resp: 51813

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 26.6 | 21.3 | 31.9 |
| 229 | 19.4 | 15.5 | 23.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#82
3,3'-Dichlorobenzidine
Concen: 39.841 ng
RT: 14.080 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:252 Resp: 148294

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 64.4 | 51.5 | 77.3 |
| 126 | 11.7 | 9.4 | 14.0 |






#83
Chrysene
Concen: 39.134 ng
RT: 14.157 min Scan# 2(
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:228 Resp: 46197(

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 29.3 | 23.4 | 35.2 |
| 229 | 19.1 | 15.3 | 22.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 37.454 ng
RT: 14.110 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:149 Resp: 198176

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 27.4 | 21.9 | 32.9 |
| 279 | 4.2 | 3.4 | 5.0 |






#85
Di-n-octyl phthalate
Concen: 36.531 ng
RT: 14.727 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:149 Resp: 35474

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.5 | 1.2 | 1.8 |
| 43 | 13.3 | 10.6 | 16.0 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.639 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:264 Resp: 226517

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.1 | 18.5 | 27.7 |
| 265 | 21.9 | 17.5 | 26.3 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 40.912 ng
RT: 17.180 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:276 Resp: 68927

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 28.3 | 22.6 | 34.0 |
| 277 | 25.2 | 20.2 | 30.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#88
Benzo(b)fluoranthene
Concen: 37.520 ng
RT: 15.192 min Scan# 2210
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:252 Resp: 490240

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.9 | 17.5 | 26.3 |
| 125 | 10.8 | 8.6 | 13.0 |






#89
Benzo(k)fluoranthene
Concen: 41.721 ng
RT: 15.221 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:252 Resp: 52005

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.8 | 17.4 | 26.2 |
| 125 | 10.6 | 8.5 | 12.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#90
Benzo(a)pyrene
Concen: 41.215 ng
RT: 15.574 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:252 Resp: 509307

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.9 | 17.5 | 26.3 |
| 125 | 11.7 | 9.4 | 14.0 |






#91
Dibenzo(a,h)anthracene
Concen: 40.782 ng
RT: 17.198 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:278 Resp: 56166

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 18.3 | 14.6 | 22.0 |
| 279 | 23.5 | 18.8 | 28.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#92
Benzo(g,h,i)perylene
Concen: 40.584 ng
RT: 17.645 min Scan# 2627
Delta R.T. 0.000 min
Lab File: BF143097.D
Acq: 15 Jul 2025 15:05

Tgt Ion:276 Resp: 569850

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 24.0 | 19.2 | 28.8 |
| 138 | 24.7 | 19.8 | 29.6 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143098.D
Acq On : 15 Jul 2025 15:35
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jul 15 17:41:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| **Internal Standards** | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 6.969 | 152 | 119035 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 8.251 | 136 | 456501 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 10.004 | 164 | 223126 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 11.492 | 188 | 354075 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 14.127 | 240 | 183692 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 15.633 | 264 | 221502 | 20.000 | ng | 0.00 |
| **System Monitoring Compounds** | | | | | | |
| 5) 2-Fluorophenol | 5.593 | 112 | 776953 | 103.124 | ng | 0.00 |
| 7) Phenol-d6 | 6.587 | 99 | 978105 | 103.280 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 7.528 | 82 | 911721 | 112.035 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 10.792 | 330 | 227678 | 114.439 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 9.328 | 172 | 1628659 | 97.028 | ng | 0.00 |
| 79) Terphenyl-d14 | 13.075 | 244 | 1271626 | 101.195 | ng | 0.00 |
| **Target Compounds** | | | | | | Qvalue |
| 2) 1,4-Dioxane | 2.811 | 88 | 196652 | 52.435 | ng | 99 |
| 3) Pyridine | 3.581 | 79 | 500208 | 52.670 | ng | 99 |
| 4) n-Nitrosodimethylamine | 3.528 | 42 | 261539 | 53.153 | ng | 100 |
| 6) Aniline | 6.628 | 93 | 699680 | 52.465 | ng | 100 |
| 8) 2-Chlorophenol | 6.751 | 128 | 405741 | 53.601 | ng | 100 |
| 9) Benzaldehyde | 6.516 | 77 | 431994 | 59.864 | ng | 98 |
| 10) Phenol | 6.598 | 94 | 529842m | 51.680 | ng | |
| 11) bis(2-Chloroethyl)ether | 6.698 | 93 | 408091 | 51.417 | ng | 98 |
| 12) 1,3-Dichlorobenzene | 6.910 | 146 | 437876 | 51.203 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 6.987 | 146 | 441012 | 51.260 | ng | 100 |
| 14) 1,2-Dichlorobenzene | 7.140 | 146 | 421724 | 51.401 | ng | 98 |
| 15) Benzyl Alcohol | 7.104 | 79 | 373560 | 53.234 | ng | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.240 | 45 | 748892 | 51.187 | ng | 100 |
| 17) 2-Methylphenol | 7.210 | 107 | 342357 | 52.409 | ng | 99 |
| 18) Hexachloroethane | 7.487 | 117 | 148761 | 53.114 | ng | 98 |
| 19) n-Nitroso-di-n-propyla... | 7.381 | 70 | 302685 | 52.169 | ng | 99 |
| 20) 3+4-Methylphenols | 7.363 | 107 | 422794 | 52.776 | ng | 93 |
| 22) Acetophenone | 7.375 | 105 | 549763 | 49.959 | ng | 97 |
| 24) Nitrobenzene | 7.545 | 77 | 439273 | 55.918 | ng | 100 |
| 25) Isophorone | 7.787 | 82 | 819274 | 52.088 | ng | 100 |
| 26) 2-Nitrophenol | 7.863 | 139 | 160386 | 50.837 | ng | 99 |
| 27) 2,4-Dimethylphenol | 7.893 | 122 | 379432 | 51.976 | ng | 100 |
| 28) bis(2-Chloroethoxy)met... | 7.993 | 93 | 483608 | 50.704 | ng | 99 |
| 29) 2,4-Dichlorophenol | 8.098 | 162 | 319968 | 53.575 | ng | 99 |
| 30) 1,2,4-Trichlorobenzene | 8.193 | 180 | 345389 | 50.958 | ng | 99 |
| 31) Naphthalene | 8.275 | 128 | 1126659 | 49.700 | ng | 100 |
| 32) Benzoic acid | 7.998 | 122 | 180966 | 49.535 | ng | 98 |
| 33) 4-Chloroaniline | 8.310 | 127 | 467074 | 51.237 | ng | 99 |
| 34) Hexachlorobutadiene | 8.392 | 225 | 207797 | 51.263 | ng | 98 |
| 35) Caprolactam | 8.681 | 113 | 99144 | 53.162 | ng | 96 |
| 36) 4-Chloro-3-methylphenol | 8.787 | 107 | 350906 | 52.746 | ng | 98 |
| 37) 2-Methylnaphthalene | 8.963 | 142 | 681726 | 50.410 | ng | 99 |
| 38) 1-Methylnaphthalene | 9.063 | 142 | 701548 | 50.016 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 9.128 | 216 | 335206 | 51.551 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 9.122 | 237 | 198811 | 55.147 | ng | 98 |
| 43) 2,4,6-Trichlorophenol | 9.234 | 196 | 223131 | 55.226 | ng | 99 |

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143098.D
Acq On : 15 Jul 2025 15:35
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jul 15 17:41:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 9.269 | 196 | 232905 | 54.977 | ng | | 99 |
| 46) 1,1'-Biphenyl | 9.428 | 154 | 897834 | 50.458 | ng | | 100 |
| 47) 2-Chloronaphthalene | 9.451 | 162 | 666760 | 51.007 | ng | | 100 |
| 48) 2-Nitroaniline | 9.539 | 65 | 207291 | 52.775 | ng | | 98 |
| 49) Acenaphthylene | 9.869 | 152 | 1109586 | 50.940 | ng | | 99 |
| 50) Dimethylphthalate | 9.722 | 163 | 726121 | 51.728 | ng | | 100 |
| 51) 2,6-Dinitrotoluene | 9.781 | 165 | 146385 | 52.497 | ng | | 92 |
| 52) Acenaphthene | 10.039 | 154 | 651291 | 50.669 | ng | | 100 |
| 53) 3-Nitroaniline | 9.945 | 138 | 177602 | 58.553 | ng | | 99 |
| 54) 2,4-Dinitrophenol | 10.051 | 184 | 47800 | 51.931 | ng | # | 1 |
| 55) Dibenzofuran | 10.210 | 168 | 964545 | 50.344 | ng | | 100 |
| 56) 4-Nitrophenol | 10.092 | 139 | 133073 | 55.203 | ng | | 98 |
| 57) 2,4-Dinitrotoluene | 10.181 | 165 | 181610 | 52.116 | ng | | 99 |
| 58) Fluorene | 10.557 | 166 | 716173 | 49.859 | ng | | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 10.322 | 232 | 189211 | 56.202 | ng | | 98 |
| 60) Diethylphthalate | 10.422 | 149 | 702995 | 51.909 | ng | | 99 |
| 61) 4-Chlorophenyl-phenyle... | 10.545 | 204 | 348876 | 50.675 | ng | | 99 |
| 62) 4-Nitroaniline | 10.557 | 138 | 150470 | 57.045 | ng | | 99 |
| 63) Azobenzene | 10.704 | 77 | 770603 | 51.451 | ng | | 99 |
| 65) 4,6-Dinitro-2-methylph... | 10.592 | 198 | 70324 | 51.847 | ng | | 92 |
| 66) n-Nitrosodiphenylamine | 10.663 | 169 | 634539 | 51.019 | ng | | 100 |
| 67) 4-Bromophenyl-phenylether | 11.034 | 248 | 211371 | 52.494 | ng | | 99 |
| 68) Hexachlorobenzene | 11.104 | 284 | 221245 | 51.785 | ng | | 98 |
| 69) Atrazine | 11.181 | 200 | 172972 | 54.359 | ng | | 98 |
| 70) Pentachlorophenol | 11.292 | 266 | 131545 | 56.120 | ng | | 99 |
| 71) Phenanthrene | 11.516 | 178 | 963568 | 50.425 | ng | | 100 |
| 72) Anthracene | 11.569 | 178 | 979118 | 50.953 | ng | | 99 |
| 73) Carbazole | 11.716 | 167 | 883827 | 51.340 | ng | | 100 |
| 74) Di-n-butylphthalate | 12.051 | 149 | 870362 | 55.079 | ng | | 100 |
| 75) Fluoranthene | 12.704 | 202 | 883558 | 50.852 | ng | | 100 |
| 77) Benzidine | 12.816 | 184 | 309313 | 49.425 | ng | | 99 |
| 78) Pyrene | 12.933 | 202 | 880826 | 51.681 | ng | | 99 |
| 80) Butylbenzylphthalate | 13.545 | 149 | 176615 | 51.324 | ng | | 96 |
| 81) Benzo(a)anthracene | 14.122 | 228 | 644353 | 52.641 | ng | | 99 |
| 82) 3,3'-Dichlorobenzidine | 14.080 | 252 | 196472 | 55.273 | ng | | 100 |
| 83) Chrysene | 14.157 | 228 | 583302 | 51.740 | ng | | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 14.110 | 149 | 265225 | 51.119 | ng | | 99 |
| 85) Di-n-octyl phthalate | 14.727 | 149 | 492840 | 50.100 | ng | | 99 |
| 87) Indeno(1,2,3-cd)pyrene | 17.180 | 276 | 896872 | 54.440 | ng | | 100 |
| 88) Benzo(b)fluoranthene | 15.192 | 252 | 718106 | 56.204 | ng | | 100 |
| 89) Benzo(k)fluoranthene | 15.221 | 252 | 602261 | 49.410 | ng | | 100 |
| 90) Benzo(a)pyrene | 15.574 | 252 | 662193 | 54.800 | ng | | 99 |
| 91) Dibenzo(a,h)anthracene | 17.204 | 278 | 736672 | 54.700 | ng | | 99 |
| 92) Benzo(g,h,i)perylene | 17.651 | 276 | 746789 | 54.390 | ng | | 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143098.D
Acq On : 15 Jul 2025 15:35
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jul 15 17:41:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.969 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:152 Resp: 11903

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 160.4 | 126.5 | 189.7 |
| 115 | 61.5 | 47.0 | 70.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#2
1,4-Dioxane
Concen: 52.435 ng
RT: 2.811 min Scan# 105
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 88 Resp: 196652

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 87.3 | 70.5 | 105.7 |
| 43 | 34.6 | 27.3 | 40.9 |






#3
Pyridine
Concen: 52.670 ng
RT: 3.581 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 79 Resp: 50020

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 85.7 | 69.2 | 103.8 |
| 51 | 40.9 | 33.3 | 49.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#4
n-Nitrosodimethylamine
Concen: 53.153 ng
RT: 3.528 min Scan# 227
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 42 Resp: 261539

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 115.8 | 92.8 | 139.2 |
| 44 | 7.3 | 5.8 | 8.6 |






#5
2-Fluorophenol
Concen: 103.124 ng
RT: 5.593 min Scan# 5
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:112 Resp: 77695

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 65.4 | 54.6 | 82.0 |
| 63 | 32.8 | 27.3 | 40.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#6
Aniline
Concen: 52.465 ng
RT: 6.628 min Scan# 754
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 93 Resp: 699680

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 41.5 | 33.1 | 49.7 |
| 65 | 21.0 | 16.9 | 25.3 |






#7
Phenol-d6
Concen: 103.280 ng
RT: 6.587 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 99 Resp: 97810

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 21.7 | 17.2 | 25.8 |
| 71 | 34.4 | 27.4 | 41.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#8
2-Chlorophenol
Concen: 53.601 ng
RT: 6.751 min Scan# 775
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:128 Resp: 405741

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.5 | 12.8 | 52.8 |
| 64 | 56.3 | 36.1 | 76.1 |






#9
Benzaldehyde
Concen: 59.864 ng
RT: 6.516 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 77 Resp: 43199

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 92.9 | 76.3 | 116.3 |
| 106 | 91.5 | 72.8 | 112.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#10
Phenol
Concen: 51.680 ng m
RT: 6.598 min Scan# 749
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 94 Resp: 529842

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 30.7 | 9.8 | 49.8 |
| 66 | 42.9 | 20.5 | 60.5 |






#11
bis(2-Chloroethyl)ether
Concen: 51.417 ng
RT: 6.698 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 93 Resp: 40809

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 84.0 | 61.9 | 101.9 |
| 95 | 32.4 | 12.2 | 52.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#12
1,3-Dichlorobenzene
Concen: 51.203 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:146 Resp: 437876

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.5 | 50.1 | 75.1 |
| 75 | 34.8 | 27.4 | 41.0 |






#13
1,4-Dichlorobenzene
Concen: 51.260 ng
RT: 6.987 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:146 Resp: 44101

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.3 | 50.6 | 76.0 |
| 111 | 43.3 | 35.0 | 52.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#14
1,2-Dichlorobenzene
Concen: 51.401 ng
RT: 7.140 min Scan# 841
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:146 Resp: 421724

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.1 | 49.8 | 74.6 |
| 111 | 45.8 | 35.4 | 53.0 |






#15
Benzyl Alcohol
Concen: 53.234 ng
RT: 7.104 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 79 Resp: 37356

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 76.1 | 61.1 | 91.7 |
| 77 | 60.9 | 49.9 | 74.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 51.187 ng
RT: 7.240 min Scan# 858
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 45 Resp: 748892

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 11.2 | 0.0 | 31.1 |
| 79 | 8.4 | 0.0 | 28.4 |






#17
2-Methylphenol
Concen: 52.409 ng
RT: 7.210 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:107 Resp: 34235

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 114.3 | 90.9 | 136.3 |
| 77 | 45.5 | 37.5 | 56.3 |
| 79 | 45.1 | 36.9 | 55.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#18
Hexachloroethane
Concen: 53.114 ng
RT: 7.487 min Scan# 900
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:117 Resp: 148761

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 97.2 | 76.7 | 115.1 |
| 201 | 95.5 | 75.0 | 112.4 |






#19
n-Nitroso-di-n-propylamine
Concen: 52.169 ng
RT: 7.381 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 70 Resp: 30268

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 65.5 | 53.2 | 79.8 |
| 101 | 9.6 | 7.5 | 11.3 |
| 130 | 20.5 | 15.7 | 23.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#20
3+4-Methylphenols
Concen: 52.776 ng
RT: 7.363 min Scan# 879
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:107 Resp: 422794

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 90.7 | 71.3 | 111.3 |
| 77 | 81.5 | 77.2 | 117.2 |
| 79 | 30.2 | 10.7 | 50.7 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.251 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:136 Resp: 45650

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.9 | 8.9 | 13.3 |
| 54 | 9.9 | 8.3 | 12.5 |
| 68 | 6.7 | 5.3 | 7.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#22
Acetophenone
Concen: 49.959 ng
RT: 7.375 min Scan# 881
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:105 Resp: 549763

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 7.8 | 7.2 | 10.8 |
| 51 | 36.0 | 27.0 | 40.4 |
| 120 | 23.4 | 18.2 | 27.2 |






#23
Nitrobenzene-d5
Concen: 112.035 ng
RT: 7.528 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 82 Resp: 91172

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 41.4 | 33.3 | 49.9 |
| 54 | 59.4 | 47.4 | 71.2 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#24
Nitrobenzene
Concen: 55.918 ng
RT: 7.545 min Scan# 910
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 77 Resp: 439273

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 43.2 | 34.5 | 51.7 |
| 65 | 14.3 | 11.3 | 16.9 |






#25
Isophorone
Concen: 52.088 ng
RT: 7.787 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 82 Resp: 81927

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.0 | 5.6 | 8.4 |
| 138 | 17.8 | 14.3 | 21.5 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#26
2-Nitrophenol
Concen: 50.837 ng
RT: 7.863 min Scan# 964
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:139 Resp: 160386

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 29.7 | 23.4 | 35.0 |
| 65 | 53.7 | 43.4 | 65.2 |






#27
2,4-Dimethylphenol
Concen: 51.976 ng
RT: 7.893 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:122 Resp: 37943

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 114.5 | 91.3 | 136.9 |
| 121 | 57.1 | 45.5 | 68.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#28
bis(2-Chloroethoxy)methane
Concen: 50.704 ng
RT: 7.993 min Scan# 986
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 93 Resp: 483608

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.4 | 26.2 | 39.4 |
| 123 | 12.2 | 9.6 | 14.4 |






#29
2,4-Dichlorophenol
Concen: 53.575 ng
RT: 8.098 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:162 Resp: 31996

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 63.7 | 43.1 | 83.1 |
| 98 | 40.4 | 19.2 | 59.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#30
1,2,4-Trichlorobenzene
Concen: 50.958 ng
RT: 8.193 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:180 Resp: 345389

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 94.3 | 74.7 | 112.1 |
| 145 | 31.8 | 24.5 | 36.7 |






#31
Naphthalene
Concen: 49.700 ng
RT: 8.275 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:128 Resp: 112665

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.1 | 9.0 | 13.4 |
| 127 | 13.3 | 10.5 | 15.7 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#32
Benzoic acid
Concen: 49.535 ng
RT: 7.998 min Scan# 987
Delta R.T. 0.011 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:122 Resp: 180966

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 122.8 | 105.4 | 145.4 |
| 77 | 99.4 | 78.6 | 118.6 |






#33
4-Chloroaniline
Concen: 51.237 ng
RT: 8.310 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:127 Resp: 46707

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 31.3 | 26.2 | 39.4 |
| 65 | 35.6 | 28.6 | 42.8 |
| 92 | 21.1 | 17.1 | 25.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#34
Hexachlorobutadiene
Concen: 51.263 ng
RT: 8.392 min Scan# 1054
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:225 Resp: 207797

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 63.6 | 50.3 | 75.5 |
| 227 | 64.7 | 50.1 | 75.1 |






#35
Caprolactam
Concen: 53.162 ng
RT: 8.681 min Scan# 11
Delta R.T. 0.012 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:113 Resp: 9914

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 199.4 | 186.2 | 226.2 |
| 56 | 151.0 | 134.1 | 174.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#36
4-Chloro-3-methylphenol
Concen: 52.746 ng
RT: 8.787 min Scan# 1121
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:107 Resp: 350906

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 24.2 | 20.2 | 30.2 |
| 142 | 76.8 | 62.4 | 93.6 |






#37
2-Methylnaphthalene
Concen: 50.410 ng
RT: 8.963 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:142 Resp: 68172

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 88.5 | 70.2 | 105.2 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#38
1-Methylnaphthalene
Concen: 50.016 ng
RT: 9.063 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:142 Resp: 701548

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 90.1 | 72.2 | 108.4 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:164 Resp: 22312

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 99.3 | 79.0 | 118.6 |
| 160 | 44.1 | 35.8 | 53.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 51.551 ng
RT: 9.128 min Scan# 1179
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:216 Resp: 335206

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.5 | 62.6 | 94.0 |
| 179 | 20.5 | 16.2 | 24.2 |
| 108 | 19.4 | 15.8 | 23.6 |






#41
Hexachlorocyclopentadiene
Concen: 55.147 ng
RT: 9.122 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:237 Resp: 19881

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 63.4 | 41.6 | 81.6 |
| 272 | 13.1 | 0.0 | 32.9 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#42
2,4,6-Tribromophenol
Concen: 114.439 ng
RT: 10.792 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:330 Resp: 227678

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 96.3 | 77.9 | 116.9 |
| 141 | 30.5 | 25.9 | 38.9 |






#43
2,4,6-Trichlorophenol
Concen: 55.226 ng
RT: 9.234 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:196 Resp: 22313

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 94.9 | 76.9 | 115.3 |
| 200 | 30.2 | 23.5 | 35.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#44
2,4,5-Trichlorophenol
Concen: 54.977 ng
RT: 9.269 min Scan# 1203
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:196 Resp: 232905

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.5 | 76.4 | 114.6 |
| 97 | 54.3 | 42.4 | 63.6 |
| 132 | 30.6 | 23.9 | 35.9 |






#45
2-Fluorobiphenyl
Concen: 97.028 ng
RT: 9.328 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:172 Resp: 162865

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.6 | 28.2 | 42.4 |
| 170 | 23.6 | 18.6 | 28.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#46
1,1'-Biphenyl
Concen: 50.458 ng
RT: 9.428 min Scan# 1230
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:154 Resp: 897834

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.4 | 20.1 | 60.1 |
| 76 | 13.7 | 0.0 | 33.7 |






#47
2-Chloronaphthalene
Concen: 51.007 ng
RT: 9.451 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:162 Resp: 66676

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 41.2 | 33.0 | 49.6 |
| 164 | 33.0 | 26.1 | 39.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#48
2-Nitroaniline
Concen: 52.775 ng
RT: 9.539 min Scan# 1249
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 65 Resp: 207291

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 65.1 | 51.3 | 76.9 |
| 138 | 100.2 | 77.8 | 116.6 |






#49
Acenaphthylene
Concen: 50.940 ng
RT: 9.869 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:152 Resp: 110958

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.2 | 15.8 | 23.8 |
| 153 | 13.4 | 10.4 | 15.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#50
Dimethylphthalate
Concen: 51.728 ng
RT: 9.722 min Scan# 1280
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:163 Resp: 726121

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.3 | 3.4 | 5.2 |
| 164 | 10.1 | 8.2 | 12.2 |






#51
2,6-Dinitrotoluene
Concen: 52.497 ng
RT: 9.781 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:165 Resp: 14638

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 63.7 | 57.4 | 86.0 |
| 89 | 60.4 | 51.8 | 77.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#52
Acenaphthene
Concen: 50.669 ng
RT: 10.039 min Scan# 1334
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:154 Resp: 651291

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 108.6 | 86.7 | 130.1 |
| 152 | 51.1 | 41.4 | 62.2 |






#53
3-Nitroaniline
Concen: 58.553 ng
RT: 9.945 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:138 Resp: 17760

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 12.1 | 10.1 | 15.1 |
| 92 | 119.6 | 96.4 | 144.6 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#54
2,4-Dinitrophenol
Concen: 51.931 ng
RT: 10.051 min Scan# 1336
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:184 Resp: 47800

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 113.4 | 149.3 | 223.9# |
| 154 | 651.1 | 1048.0 | 1572.0# |






#55
Dibenzofuran
Concen: 50.344 ng
RT: 10.210 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:168 Resp: 96454

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 37.0 | 29.8 | 44.6 |
| 169 | 13.5 | 10.8 | 16.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#56
4-Nitrophenol
Concen: 55.203 ng
RT: 10.092 min Scan# 1343
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:139 Resp: 133073

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 70.9 | 49.4 | 89.4 |
| 65 | 105.2 | 88.2 | 128.2 |






#57
2,4-Dinitrotoluene
Concen: 52.116 ng
RT: 10.181 min Scan# 1
Delta R.T. 0.001 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:165 Resp: 18161

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 52.2 | 40.6 | 61.0 |
| 89 | 78.6 | 62.3 | 93.5 |
| 182 | 2.9 | 2.6 | 3.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#58
Fluorene
Concen: 49.859 ng
RT: 10.557 min Scan# 1422
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:166 Resp: 716173

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 91.7 | 73.7 | 110.5 |
| 167 | 13.3 | 10.9 | 16.3 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 56.202 ng
RT: 10.322 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:232 Resp: 18921

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 43.8 | 36.9 | 55.3 |
| 130 | 2.3 | 1.8 | 2.8 |
| 166 | 28.8 | 23.0 | 34.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#60
Diethylphthalate
Concen: 51.909 ng
RT: 10.422 min Scan# 1399
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:149 Resp: 702995

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 21.7 | 17.3 | 25.9 |
| 150 | 12.8 | 9.9 | 14.9 |






#61
4-Chlorophenyl-phenylether
Concen: 50.675 ng
RT: 10.545 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:204 Resp: 34887

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 33.0 | 25.8 | 38.6 |
| 141 | 59.6 | 46.8 | 70.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#62
4-Nitroaniline
Concen: 57.045 ng
RT: 10.557 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:138 Resp: 150470

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 50.9 | 30.4 | 70.4 |
| 108 | 70.3 | 49.5 | 89.5 |






#63
Azobenzene
Concen: 51.451 ng
RT: 10.704 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion: 77 Resp: 77060

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 23.5 | 4.0 | 44.0 |
| 105 | 21.1 | 1.4 | 41.4 |
| 51 | 33.7 | 14.3 | 54.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.492 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:188 Resp: 354075

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 10.3 | 8.6 | 13.0 |
| 80 | 11.3 | 9.3 | 13.9 |






#65
4,6-Dinitro-2-methylphenol
Concen: 51.847 ng
RT: 10.592 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:198 Resp: 7032

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 56.6 | 44.6 | 84.6 |
| 105 | 46.4 | 29.3 | 69.3 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#66
n-Nitrosodiphenylamine
Concen: 51.019 ng
RT: 10.663 min Scan# 1440
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:169 Resp: 634539

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 66.1 | 53.0 | 79.6 |
| 167 | 35.3 | 28.4 | 42.6 |






#67
4-Bromophenyl-phenylether
Concen: 52.494 ng
RT: 11.034 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:248 Resp: 21137

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 96.9 | 76.7 | 115.1 |
| 141 | 77.2 | 61.7 | 92.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#68
Hexachlorobenzene
Concen: 51.785 ng
RT: 11.104 min Scan# 1515
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:284 Resp: 221245

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 34.5 | 28.9 | 43.3 |
| 249 | 31.1 | 24.6 | 37.0 |






#69
Atrazine
Concen: 54.359 ng
RT: 11.181 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:200 Resp: 17297

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 29.0 | 9.1 | 49.1 |
| 215 | 49.4 | 31.0 | 71.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#70
Pentachlorophenol
Concen: 56.120 ng
RT: 11.292 min Scan# 1547
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:266 Resp: 131545

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 63.3 | 51.8 | 77.6 |
| 264 | 61.7 | 49.7 | 74.5 |






#71
Phenanthrene
Concen: 50.425 ng
RT: 11.516 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:178 Resp: 96356

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.1 | 15.1 | 22.7 |
| 179 | 15.5 | 12.2 | 18.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#72
Anthracene
Concen: 50.953 ng
RT: 11.569 min Scan# 1594
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:178 Resp: 979118

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.9 | 14.7 | 22.1 |
| 179 | 15.5 | 12.4 | 18.6 |






#73
Carbazole
Concen: 51.340 ng
RT: 11.716 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:167 Resp: 88382

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.0 | 17.0 | 25.4 |
| 139 | 12.7 | 10.2 | 15.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#74
Di-n-butylphthalate
Concen: 55.079 ng
RT: 12.051 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:149 Resp: 870362

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.2 | 7.3 | 10.9 |
| 104 | 4.8 | 3.8 | 5.8 |






#75
Fluoranthene
Concen: 50.852 ng
RT: 12.704 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:202 Resp: 88355

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 12.2 | 0.0 | 32.4 |
| 203 | 17.5 | 0.0 | 37.4 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.127 min Scan# 2029
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:240 Resp: 183692

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 12.4 | 10.0 | 15.0 |
| 236 | 24.5 | 20.0 | 30.0 |






#77
Benzidine
Concen: 49.425 ng
RT: 12.816 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:184 Resp: 30931

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 15.6 | 12.4 | 18.6 |
| 183 | 11.1 | 8.6 | 12.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#78
Pyrene
Concen: 51.681 ng
RT: 12.933 min Scan# 1826
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:202 Resp: 880826

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 19.8 | 15.5 | 23.3 |
| 203 | 17.6 | 14.0 | 21.0 |






#79
Terphenyl-d14
Concen: 101.195 ng
RT: 13.075 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:244 Resp: 127162

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.6 | 5.2 | 7.8 |
| 122 | 10.7 | 8.8 | 13.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#80
Butylbenzylphthalate
Concen: 51.324 ng
RT: 13.545 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:149 Resp: 176615

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 74.8 | 57.0 | 85.4 |
| 206 | 20.7 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 52.641 ng
RT: 14.122 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:228 Resp: 64435

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 26.9 | 21.3 | 31.9 |
| 229 | 20.0 | 15.5 | 23.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#82
3,3'-Dichlorobenzidine
Concen: 55.273 ng
RT: 14.080 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:252 Resp: 196472

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 64.6 | 51.5 | 77.3 |
| 126 | 11.7 | 9.4 | 14.0 |






#83
Chrysene
Concen: 51.740 ng
RT: 14.157 min Scan# 2(
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:228 Resp: 58330

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 29.5 | 23.4 | 35.2 |
| 229 | 19.9 | 15.3 | 22.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 51.119 ng
RT: 14.110 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:149 Resp: 265225

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 28.2 | 21.9 | 32.9 |
| 279 | 4.1 | 3.4 | 5.0 |






#85
Di-n-octyl phthalate
Concen: 50.100 ng
RT: 14.727 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:149 Resp: 49284

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.4 | 1.2 | 1.8 |
| 43 | 13.0 | 10.6 | 16.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.633 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:264 Resp: 221502

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 22.6 | 18.5 | 27.7 |
| 265 | 21.7 | 17.5 | 26.3 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 54.440 ng
RT: 17.180 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:276 Resp: 89687

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 28.1 | 22.6 | 34.0 |
| 277 | 25.4 | 20.2 | 30.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#88
Benzo(b)fluoranthene
Concen: 56.204 ng
RT: 15.192 min Scan# 2210
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:252 Resp: 718106

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.1 | 17.5 | 26.3 |
| 125 | 10.7 | 8.6 | 13.0 |






#89
Benzo(k)fluoranthene
Concen: 49.410 ng
RT: 15.221 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:252 Resp: 60226

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.8 | 17.4 | 26.2 |
| 125 | 10.5 | 8.5 | 12.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#90
Benzo(a)pyrene
Concen: 54.800 ng
RT: 15.574 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:252 Resp: 662193

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.3 | 17.5 | 26.3 |
| 125 | 11.6 | 9.4 | 14.0 |






#91
Dibenzo(a,h)anthracene
Concen: 54.700 ng
RT: 17.204 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:278 Resp: 73667

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 17.4 | 14.6 | 22.0 |
| 279 | 23.3 | 18.8 | 28.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#92
Benzo(g,h,i)perylene
Concen: 54.390 ng
RT: 17.651 min Scan# 2628
Delta R.T. 0.006 min
Lab File: BF143098.D
Acq: 15 Jul 2025 15:35

Tgt Ion:276 Resp: 746789

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.2 | 19.2 | 28.8 |
| 138 | 23.8 | 19.8 | 29.6 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143099.D
Acq On : 15 Jul 2025 16:05
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jul 15 17:41:54 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| **Internal Standards** | | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 6.969 | 152 | 133949 | 20.000 | ng | | 0.00 |
| 21) Naphthalene-d8 | 8.251 | 136 | 501735 | 20.000 | ng | | 0.00 |
| 39) Acenaphthene-d10 | 10.010 | 164 | 246002 | 20.000 | ng | | 0.00 |
| 64) Phenanthrene-d10 | 11.492 | 188 | 380313 | 20.000 | ng | | 0.00 |
| 76) Chrysene-d12 | 14.133 | 240 | 205942 | 20.000 | ng | | 0.00 |
| 86) Perylene-d12 | 15.639 | 264 | 260121 | 20.000 | ng | | 0.00 |
| **System Monitoring Compounds** | | | | | | | |
| 5) 2-Fluorophenol | 5.593 | 112 | 990040 | 116.776 | ng | | 0.00 |
| 7) Phenol-d6 | 6.593 | 99 | 1243617 | 116.695 | ng | | 0.00 |
| 23) Nitrobenzene-d5 | 7.534 | 82 | 1172838 | 131.129 | ng | | 0.00 |
| 42) 2,4,6-Tribromophenol | 10.798 | 330 | 286830 | 130.765 | ng | | 0.00 |
| 45) 2-Fluorobiphenyl | 9.328 | 172 | 1999612 | 108.050 | ng | | 0.00 |
| 79) Terphenyl-d14 | 13.075 | 244 | 1539271 | 109.260 | ng | | 0.00 |
| **Target Compounds** | | | | | | | Qvalue |
| 2) 1,4-Dioxane | 2.810 | 88 | 251636 | 59.625 | ng | | 100 |
| 3) Pyridine | 3.587 | 79 | 647877 | 60.624 | ng | | 99 |
| 4) n-Nitrosodimethylamine | 3.534 | 42 | 337438 | 60.943 | ng | | 100 |
| 6) Aniline | 6.634 | 93 | 882633 | 58.815 | ng | | 100 |
| 8) 2-Chlorophenol | 6.757 | 128 | 516965 | 60.690 | ng | | 98 |
| 9) Benzaldehyde | 6.522 | 77 | 467499 | 57.571 | ng | | 99 |
| 10) Phenol | 6.604 | 94 | 680847m | 59.015 | ng | | |
| 11) bis(2-Chloroethyl)ether | 6.704 | 93 | 518695 | 58.076 | ng | | 99 |
| 12) 1,3-Dichlorobenzene | 6.916 | 146 | 557813 | 57.965 | ng | | 99 |
| 13) 1,4-Dichlorobenzene | 6.987 | 146 | 556805 | 57.513 | ng | | 99 |
| 14) 1,2-Dichlorobenzene | 7.140 | 146 | 530965 | 57.510 | ng | | 98 |
| 15) Benzyl Alcohol | 7.104 | 79 | 477773 | 60.505 | ng | | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.245 | 45 | 941009 | 57.157 | ng | | 99 |
| 17) 2-Methylphenol | 7.210 | 107 | 432787 | 58.876 | ng | | 99 |
| 18) Hexachloroethane | 7.487 | 117 | 191457 | 60.747 | ng | | 99 |
| 19) n-Nitroso-di-n-propyla... | 7.387 | 70 | 379882 | 58.184 | ng | | 98 |
| 20) 3+4-Methylphenols | 7.369 | 107 | 519658 | 57.645 | ng | | 98 |
| 22) Acetophenone | 7.375 | 105 | 679007 | 56.141 | ng | # | 92 |
| 24) Nitrobenzene | 7.551 | 77 | 550668 | 63.778 | ng | | 99 |
| 25) Isophorone | 7.793 | 82 | 1033650 | 59.793 | ng | | 100 |
| 26) 2-Nitrophenol | 7.863 | 139 | 219665 | 62.382 | ng | | 98 |
| 27) 2,4-Dimethylphenol | 7.898 | 122 | 474411 | 59.127 | ng | | 98 |
| 28) bis(2-Chloroethoxy)met... | 7.992 | 93 | 608314 | 58.029 | ng | | 100 |
| 29) 2,4-Dichlorophenol | 8.104 | 162 | 408655 | 62.256 | ng | | 99 |
| 30) 1,2,4-Trichlorobenzene | 8.192 | 180 | 431427 | 57.913 | ng | | 98 |
| 31) Naphthalene | 8.275 | 128 | 1410509 | 56.612 | ng | | 100 |
| 32) Benzoic acid | 8.010 | 122 | 256755m | 62.050 | ng | | |
| 33) 4-Chloroaniline | 8.316 | 127 | 574196 | 57.309 | ng | | 99 |
| 34) Hexachlorobutadiene | 8.398 | 225 | 265081 | 59.499 | ng | | 99 |
| 35) Caprolactam | 8.687 | 113 | 129583 | 63.219 | ng | | 97 |
| 36) 4-Chloro-3-methylphenol | 8.792 | 107 | 440907 | 60.299 | ng | | 98 |
| 37) 2-Methylnaphthalene | 8.963 | 142 | 846190 | 56.930 | ng | | 100 |
| 38) 1-Methylnaphthalene | 9.063 | 142 | 869460 | 56.399 | ng | | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 9.134 | 216 | 414099 | 57.762 | ng | | 99 |
| 41) Hexachlorocyclopentadiene | 9.122 | 237 | 262731 | 66.101 | ng | | 98 |
| 43) 2,4,6-Trichlorophenol | 9.239 | 196 | 290245 | 65.157 | ng | | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143099.D
Acq On : 15 Jul 2025 16:05
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jul 15 17:41:54 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 9.275 | 196 | 277922 | 59.502 | ng | | 99 |
| 46) 1,1'-Biphenyl | 9.428 | 154 | 1103990 | 56.274 | ng | | 100 |
| 47) 2-Chloronaphthalene | 9.451 | 162 | 826528 | 57.350 | ng | | 98 |
| 48) 2-Nitroaniline | 9.539 | 65 | 267636 | 61.259 | ng | | 98 |
| 49) Acenaphthylene | 9.869 | 152 | 1378196 | 57.388 | ng | | 100 |
| 50) Dimethylphthalate | 9.728 | 163 | 899456 | 58.118 | ng | | 100 |
| 51) 2,6-Dinitrotoluene | 9.781 | 165 | 193185 | 62.287 | ng | | 93 |
| 52) Acenaphthene | 10.045 | 154 | 802583 | 56.633 | ng | | 99 |
| 53) 3-Nitroaniline | 9.951 | 138 | 230773 | 69.008 | ng | | 96 |
| 54) 2,4-Dinitrophenol | 10.051 | 184 | 66479 | 61.314 | ng | # | 4 |
| 55) Dibenzofuran | 10.216 | 168 | 1184002 | 56.052 | ng | | 100 |
| 56) 4-Nitrophenol | 10.098 | 139 | 174657 | 65.716 | ng | | 96 |
| 57) 2,4-Dinitrotoluene | 10.186 | 165 | 235488 | 60.695 | ng | | 99 |
| 58) Fluorene | 10.557 | 166 | 875218 | 55.265 | ng | | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 10.328 | 232 | 238428 | 64.235 | ng | | 98 |
| 60) Diethylphthalate | 10.428 | 149 | 876047 | 58.672 | ng | | 99 |
| 61) 4-Chlorophenyl-phenyle... | 10.545 | 204 | 432659 | 57.000 | ng | | 99 |
| 62) 4-Nitroaniline | 10.563 | 138 | 194271 | 66.802 | ng | | 97 |
| 63) Azobenzene | 10.704 | 77 | 946204 | 57.300 | ng | | 99 |
| 65) 4,6-Dinitro-2-methylph... | 10.592 | 198 | 96305 | 62.285 | ng | | 98 |
| 66) n-Nitrosodiphenylamine | 10.663 | 169 | 781087 | 58.469 | ng | | 100 |
| 67) 4-Bromophenyl-phenylether | 11.039 | 248 | 260491 | 60.229 | ng | | 96 |
| 68) Hexachlorobenzene | 11.110 | 284 | 273112 | 59.515 | ng | | 95 |
| 69) Atrazine | 11.186 | 200 | 211351 | 61.838 | ng | | 99 |
| 70) Pentachlorophenol | 11.292 | 266 | 165676 | 65.805 | ng | | 99 |
| 71) Phenanthrene | 11.516 | 178 | 1165047 | 56.762 | ng | | 99 |
| 72) Anthracene | 11.569 | 178 | 1188635 | 57.589 | ng | | 100 |
| 73) Carbazole | 11.716 | 167 | 1057291 | 57.179 | ng | | 100 |
| 74) Di-n-butylphthalate | 12.051 | 149 | 1074246 | 63.291 | ng | | 100 |
| 75) Fluoranthene | 12.704 | 202 | 1069898 | 57.328 | ng | | 100 |
| 77) Benzidine | 12.816 | 184 | 380587 | 54.244 | ng | | 100 |
| 78) Pyrene | 12.933 | 202 | 1060216 | 55.485 | ng | | 99 |
| 80) Butylbenzylphthalate | 13.551 | 149 | 243805 | 62.406 | ng | | 98 |
| 81) Benzo(a)anthracene | 14.122 | 228 | 819246 | 59.697 | ng | | 99 |
| 82) 3,3'-Dichlorobenzidine | 14.080 | 252 | 254086 | 63.758 | ng | | 99 |
| 83) Chrysene | 14.157 | 228 | 758916 | 60.044 | ng | | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 14.110 | 149 | 364902 | 61.957 | ng | | 100 |
| 85) Di-n-octyl phthalate | 14.727 | 149 | 688940 | 60.817 | ng | | 98 |
| 87) Indeno(1,2,3-cd)pyrene | 17.192 | 276 | 1185702 | 61.286 | ng | | 99 |
| 88) Benzo(b)fluoranthene | 15.192 | 252 | 924866 | 61.640 | ng | | 100 |
| 89) Benzo(k)fluoranthene | 15.227 | 252 | 827274 | 57.794 | ng | | 99 |
| 90) Benzo(a)pyrene | 15.574 | 252 | 876856 | 61.791 | ng | | 100 |
| 91) Dibenzo(a,h)anthracene | 17.210 | 278 | 953666 | 60.300 | ng | | 99 |
| 92) Benzo(g,h,i)perylene | 17.657 | 276 | 982960 | 60.962 | ng | | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143099.D
Acq On : 15 Jul 2025 16:05
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jul 15 17:41:54 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.969 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:152 Resp: 13394

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 160.4 | 126.5 | 189.7 |
| 115 | 60.4 | 47.0 | 70.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#2
1,4-Dioxane
Concen: 59.625 ng
RT: 2.810 min Scan# 105
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 88 Resp: 251636

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 87.7 | 70.5 | 105.7 |
| 43 | 34.0 | 27.3 | 40.9 |






#3
Pyridine
Concen: 60.624 ng
RT: 3.587 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 79 Resp: 64787

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 85.3 | 69.2 | 103.8 |
| 51 | 41.3 | 33.3 | 49.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#4
n-Nitrosodimethylamine
Concen: 60.943 ng
RT: 3.534 min Scan# 228
Delta R.T. 0.012 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 42 Resp: 337438

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 115.8 | 92.8 | 139.2 |
| 44 | 7.4 | 5.8 | 8.6 |






#5
2-Fluorophenol
Concen: 116.776 ng
RT: 5.593 min Scan# 5
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:112 Resp: 99004

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 66.2 | 54.6 | 82.0 |
| 63 | 33.2 | 27.3 | 40.9 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#6
Aniline
Concen: 58.815 ng
RT: 6.634 min Scan# 755
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 93 Resp: 882633

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 41.3 | 33.1 | 49.7 |
| 65 | 20.8 | 16.9 | 25.3 |






#7
Phenol-d6
Concen: 116.695 ng
RT: 6.593 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 99 Resp: 124361

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 21.2 | 17.2 | 25.8 |
| 71 | 34.7 | 27.4 | 41.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#8
2-Chlorophenol
Concen: 60.690 ng
RT: 6.757 min Scan# 776
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:128 Resp: 516965

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.5 | 12.8 | 52.8 |
| 64 | 53.9 | 36.1 | 76.1 |






#9
Benzaldehyde
Concen: 57.571 ng
RT: 6.522 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 77 Resp: 46749

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 95.1 | 76.3 | 116.3 |
| 106 | 92.7 | 72.8 | 112.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#10
Phenol
Concen: 59.015 ng m
RT: 6.604 min Scan# 750
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 94 Resp: 680847

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 30.9 | 9.8 | 49.8 |
| 66 | 42.4 | 20.5 | 60.5 |






#11
bis(2-Chloroethyl)ether
Concen: 58.076 ng
RT: 6.704 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 93 Resp: 51869

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 82.4 | 61.9 | 101.9 |
| 95 | 32.6 | 12.2 | 52.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#12
1,3-Dichlorobenzene
Concen: 57.965 ng
RT: 6.916 min Scan# 803
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:146 Resp: 557813

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.8 | 50.1 | 75.1 |
| 75 | 33.1 | 27.4 | 41.0 |






#13
1,4-Dichlorobenzene
Concen: 57.513 ng
RT: 6.987 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:146 Resp: 55680

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.8 | 50.6 | 76.0 |
| 111 | 44.6 | 35.0 | 52.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#14
1,2-Dichlorobenzene
Concen: 57.510 ng
RT: 7.140 min Scan# 841
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:146 Resp: 530965

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.1 | 49.8 | 74.6 |
| 111 | 46.1 | 35.4 | 53.0 |






#15
Benzyl Alcohol
Concen: 60.505 ng
RT: 7.104 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 79 Resp: 47777

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 75.3 | 61.1 | 91.7 |
| 77 | 61.2 | 49.9 | 74.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 57.157 ng
RT: 7.245 min Scan# 859
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 45 Resp: 941009

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 11.5 | 0.0 | 31.1 |
| 79 | 8.9 | 0.0 | 28.4 |






#17
2-Methylphenol
Concen: 58.876 ng
RT: 7.210 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:107 Resp: 43278

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 114.9 | 90.9 | 136.3 |
| 77 | 47.2 | 37.5 | 56.3 |
| 79 | 46.9 | 36.9 | 55.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#18
Hexachloroethane
Concen: 60.747 ng
RT: 7.487 min Scan# 900
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:117 Resp: 191457

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 96.2 | 76.7 | 115.1 |
| 201 | 92.9 | 75.0 | 112.4 |






#19
n-Nitroso-di-n-propylamine
Concen: 58.184 ng
RT: 7.387 min Scan# 8
Delta R.T. 0.012 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 70 Resp: 37988

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 64.6 | 53.2 | 79.8 |
| 101 | 9.9 | 7.5 | 11.3 |
| 130 | 20.5 | 15.7 | 23.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#20
3+4-Methylphenols
Concen: 57.645 ng
RT: 7.369 min Scan# 880
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:107 Resp: 519658

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 90.6 | 71.3 | 111.3 |
| 77 | 101.8 | 77.2 | 117.2 |
| 79 | 30.8 | 10.7 | 50.7 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.251 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:136 Resp: 50173

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 11.1 | 8.9 | 13.3 |
| 54 | 10.4 | 8.3 | 12.5 |
| 68 | 6.8 | 5.3 | 7.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#22
Acetophenone
Concen: 56.141 ng
RT: 7.375 min Scan# 881
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:105 Resp: 679007

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 5.6 | 7.2 | 10.8# |
| 51 | 40.7 | 27.0 | 40.4# |
| 120 | 23.0 | 18.2 | 27.2 |






#23
Nitrobenzene-d5
Concen: 131.129 ng
RT: 7.534 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 82 Resp: 117283

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 42.2 | 33.3 | 49.9 |
| 54 | 58.7 | 47.4 | 71.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#24
Nitrobenzene
Concen: 63.778 ng
RT: 7.551 min Scan# 911
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 77 Resp: 550668

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 43.6 | 34.5 | 51.7 |
| 65 | 14.3 | 11.3 | 16.9 |






#25
Isophorone
Concen: 59.793 ng
RT: 7.793 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 82 Resp: 103365

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.0 | 5.6 | 8.4 |
| 138 | 17.6 | 14.3 | 21.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#26
2-Nitrophenol
Concen: 62.382 ng
RT: 7.863 min Scan# 964
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:139 Resp: 219665

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 30.4 | 23.4 | 35.0 |
| 65 | 56.1 | 43.4 | 65.2 |






#27
2,4-Dimethylphenol
Concen: 59.127 ng
RT: 7.898 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:122 Resp: 47441

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 116.4 | 91.3 | 136.9 |
| 121 | 58.4 | 45.5 | 68.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#28
bis(2-Chloroethoxy)methane
Concen: 58.029 ng
RT: 7.992 min Scan# 986
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 93 Resp: 608314

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.6 | 26.2 | 39.4 |
| 123 | 11.8 | 9.6 | 14.4 |






#29
2,4-Dichlorophenol
Concen: 62.256 ng
RT: 8.104 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:162 Resp: 40865

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 63.4 | 43.1 | 83.1 |
| 98 | 38.5 | 19.2 | 59.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#30
1,2,4-Trichlorobenzene
Concen: 57.913 ng
RT: 8.192 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:180 Resp: 431427

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 94.3 | 74.7 | 112.1 |
| 145 | 32.6 | 24.5 | 36.7 |






#31
Naphthalene
Concen: 56.612 ng
RT: 8.275 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:128 Resp: 141050

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.3 | 9.0 | 13.4 |
| 127 | 13.3 | 10.5 | 15.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#32
Benzoic acid
Concen: 62.050 ng m
RT: 8.010 min Scan# 989
Delta R.T. 0.023 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:122 Resp: 256755

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 128.8 | 105.4 | 145.4 |
| 77 | 101.6 | 78.6 | 118.6 |






#33
4-Chloroaniline
Concen: 57.309 ng
RT: 8.316 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:127 Resp: 57419

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.4 | 26.2 | 39.4 |
| 65 | 35.1 | 28.6 | 42.8 |
| 92 | 21.3 | 17.1 | 25.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#34
Hexachlorobutadiene
Concen: 59.499 ng
RT: 8.398 min Scan# 1055
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:225 Resp: 265081

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 62.1 | 50.3 | 75.5 |
| 227 | 63.7 | 50.1 | 75.1 |






#35
Caprolactam
Concen: 63.219 ng
RT: 8.687 min Scan# 1
Delta R.T. 0.018 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:113 Resp: 12958

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 199.1 | 186.2 | 226.2 |
| 56 | 151.7 | 134.1 | 174.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#36
4-Chloro-3-methylphenol
Concen: 60.299 ng
RT: 8.792 min Scan# 1122
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:107 Resp: 440907

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 24.9 | 20.2 | 30.2 |
| 142 | 76.3 | 62.4 | 93.6 |






#37
2-Methylnaphthalene
Concen: 56.930 ng
RT: 8.963 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:142 Resp: 84619

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 87.5 | 70.2 | 105.2 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#38
1-Methylnaphthalene
Concen: 56.399 ng
RT: 9.063 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:142 Resp: 869460

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 90.7 | 72.2 | 108.4 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.010 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:164 Resp: 24600

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 98.6 | 79.0 | 118.6 |
| 160 | 44.5 | 35.8 | 53.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 57.762 ng
RT: 9.134 min Scan# 1180
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:216 Resp: 414099

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.8 | 62.6 | 94.0 |
| 179 | 20.6 | 16.2 | 24.2 |
| 108 | 19.9 | 15.8 | 23.6 |






#41
Hexachlorocyclopentadiene
Concen: 66.101 ng
RT: 9.122 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:237 Resp: 26273

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 63.5 | 41.6 | 81.6 |
| 272 | 12.8 | 0.0 | 32.9 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#42
2,4,6-Tribromophenol
Concen: 130.765 ng
RT: 10.798 min Scan# 1463
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:330 Resp: 286830

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 97.0 | 77.9 | 116.9 |
| 141 | 30.8 | 25.9 | 38.9 |






#43
2,4,6-Trichlorophenol
Concen: 65.157 ng
RT: 9.239 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:196 Resp: 29024

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.5 | 76.9 | 115.3 |
| 200 | 30.8 | 23.5 | 35.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#44
2,4,5-Trichlorophenol
Concen: 59.502 ng
RT: 9.275 min Scan# 1204
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:196 Resp: 277922

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 97.1 | 76.4 | 114.6 |
| 97 | 53.2 | 42.4 | 63.6 |
| 132 | 29.8 | 23.9 | 35.9 |






#45
2-Fluorobiphenyl
Concen: 108.050 ng
RT: 9.328 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:172 Resp: 199961

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.6 | 28.2 | 42.4 |
| 170 | 23.4 | 18.6 | 28.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#46
1,1'-Biphenyl
Concen: 56.274 ng
RT: 9.428 min Scan# 1230
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:154 Resp: 1103990

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.1 | 20.1 | 60.1 |
| 76 | 14.0 | 0.0 | 33.7 |






#47
2-Chloronaphthalene
Concen: 57.350 ng
RT: 9.451 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:162 Resp: 82652

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 43.7 | 33.0 | 49.6 |
| 164 | 33.0 | 26.1 | 39.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#48
2-Nitroaniline
Concen: 61.259 ng
RT: 9.539 min Scan# 1249
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 65 Resp: 267636

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 65.9 | 51.3 | 76.9 |
| 138 | 99.5 | 77.8 | 116.6 |






#49
Acenaphthylene
Concen: 57.388 ng
RT: 9.869 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:152 Resp: 137819

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 19.9 | 15.8 | 23.8 |
| 153 | 13.2 | 10.4 | 15.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#50
Dimethylphthalate
Concen: 58.118 ng
RT: 9.728 min Scan# 1281
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:163 Resp: 899456

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.4 | 3.4 | 5.2 |
| 164 | 10.3 | 8.2 | 12.2 |






#51
2,6-Dinitrotoluene
Concen: 62.287 ng
RT: 9.781 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:165 Resp: 19318

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 65.2 | 57.4 | 86.0 |
| 89 | 59.9 | 51.8 | 77.8 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#52
Acenaphthene
Concen: 56.633 ng
RT: 10.045 min Scan# 1335
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:154 Resp: 802583

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 107.1 | 86.7 | 130.1 |
| 152 | 50.6 | 41.4 | 62.2 |






#53
3-Nitroaniline
Concen: 69.008 ng
RT: 9.951 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:138 Resp: 23077

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 11.6 | 10.1 | 15.1 |
| 92 | 115.8 | 96.4 | 144.6 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#54
2,4-Dinitrophenol
Concen: 61.314 ng
RT: 10.051 min Scan# 1336
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:184 Resp: 66479

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 133.6 | 149.3 | 223.9# |
| 154 | 743.3 | 1048.0 | 1572.0# |






#55
Dibenzofuran
Concen: 56.052 ng
RT: 10.216 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:168 Resp: 118400

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 37.1 | 29.8 | 44.6 |
| 169 | 13.3 | 10.8 | 16.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#56
4-Nitrophenol
Concen: 65.716 ng
RT: 10.098 min Scan# 1344
Delta R.T. 0.012 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:139 Resp: 174657

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 67.9 | 49.4 | 89.4 |
| 65 | 102.8 | 88.2 | 128.2 |






#57
2,4-Dinitrotoluene
Concen: 60.695 ng
RT: 10.186 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:165 Resp: 23548

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 49.9 | 40.6 | 61.0 |
| 89 | 76.9 | 62.3 | 93.5 |
| 182 | 3.0 | 2.6 | 3.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#58
Fluorene
Concen: 55.265 ng
RT: 10.557 min Scan# 1422
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:166 Resp: 875218

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 91.9 | 73.7 | 110.5 |
| 167 | 13.6 | 10.9 | 16.3 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 64.235 ng
RT: 10.328 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:232 Resp: 23842

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 43.9 | 36.9 | 55.3 |
| 130 | 2.1 | 1.8 | 2.8 |
| 166 | 28.8 | 23.0 | 34.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#60
Diethylphthalate
Concen: 58.672 ng
RT: 10.428 min Scan# 1400
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:149 Resp: 876047

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 22.1 | 17.3 | 25.9 |
| 150 | 12.6 | 9.9 | 14.9 |






#61
4-Chlorophenyl-phenylether
Concen: 57.000 ng
RT: 10.545 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:204 Resp: 43265

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 32.5 | 25.8 | 38.6 |
| 141 | 59.9 | 46.8 | 70.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#62
4-Nitroaniline
Concen: 66.802 ng
RT: 10.563 min Scan# 1423
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:138 Resp: 194271

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 51.7 | 30.4 | 70.4 |
| 108 | 72.5 | 49.5 | 89.5 |






#63
Azobenzene
Concen: 57.300 ng
RT: 10.704 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion: 77 Resp: 94620

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 23.2 | 4.0 | 44.0 |
| 105 | 21.1 | 1.4 | 41.4 |
| 51 | 33.8 | 14.3 | 54.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.492 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:188 Resp: 380313

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 10.9 | 8.6 | 13.0 |
| 80 | 11.4 | 9.3 | 13.9 |






#65
4,6-Dinitro-2-methylphenol
Concen: 62.285 ng
RT: 10.592 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:198 Resp: 9630

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 62.6 | 44.6 | 84.6 |
| 105 | 48.6 | 29.3 | 69.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#66
n-Nitrosodiphenylamine
Concen: 58.469 ng
RT: 10.663 min Scan# 1440
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:169 Resp: 781087

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 66.5 | 53.0 | 79.6 |
| 167 | 35.8 | 28.4 | 42.6 |






#67
4-Bromophenyl-phenylether
Concen: 60.229 ng
RT: 11.039 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:248 Resp: 26049

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 98.7 | 76.7 | 115.1 |
| 141 | 72.6 | 61.7 | 92.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#68
Hexachlorobenzene
Concen: 59.515 ng
RT: 11.110 min Scan# 1516
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:284 Resp: 273112

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 31.3 | 28.9 | 43.3 |
| 249 | 30.5 | 24.6 | 37.0 |






#69
Atrazine
Concen: 61.838 ng
RT: 11.186 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:200 Resp: 21135

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 27.9 | 9.1 | 49.1 |
| 215 | 51.1 | 31.0 | 71.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#70
Pentachlorophenol
Concen: 65.805 ng
RT: 11.292 min Scan# 1547
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:266 Resp: 165676

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 65.2 | 51.8 | 77.6 |
| 264 | 63.7 | 49.7 | 74.5 |






#71
Phenanthrene
Concen: 56.762 ng
RT: 11.516 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:178 Resp: 116504

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.3 | 15.1 | 22.7 |
| 179 | 15.7 | 12.2 | 18.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#72
Anthracene
Concen: 57.589 ng
RT: 11.569 min Scan# 1594
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:178 Resp: 1188635

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.5 | 14.7 | 22.1 |
| 179 | 15.8 | 12.4 | 18.6 |






#73
Carbazole
Concen: 57.179 ng
RT: 11.716 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:167 Resp: 105729

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.4 | 17.0 | 25.4 |
| 139 | 13.1 | 10.2 | 15.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#74
Di-n-butylphthalate
Concen: 63.291 ng
RT: 12.051 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:149 Resp: 1074246

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.2 | 7.3 | 10.9 |
| 104 | 4.7 | 3.8 | 5.8 |






#75
Fluoranthene
Concen: 57.328 ng
RT: 12.704 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:202 Resp: 106989

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 12.2 | 0.0 | 32.4 |
| 203 | 17.4 | 0.0 | 37.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.133 min Scan# 2030
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:240 Resp: 205942

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.3 | 10.0 | 15.0 |
| 236 | 25.2 | 20.0 | 30.0 |






#77
Benzidine
Concen: 54.244 ng
RT: 12.816 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:184 Resp: 38058

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 15.6 | 12.4 | 18.6 |
| 183 | 10.5 | 8.6 | 12.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#78
Pyrene
Concen: 55.485 ng
RT: 12.933 min Scan# 1826
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:202 Resp: 1060216

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 19.9 | 15.5 | 23.3 |
| 203 | 17.8 | 14.0 | 21.0 |






#79
Terphenyl-d14
Concen: 109.260 ng
RT: 13.075 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:244 Resp: 153927

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.5 | 5.2 | 7.8 |
| 122 | 10.7 | 8.8 | 13.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#80
Butylbenzylphthalate
Concen: 62.406 ng
RT: 13.551 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:149 Resp: 243805

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 69.6 | 57.0 | 85.4 |
| 206 | 21.3 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 59.697 ng
RT: 14.122 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

Tgt Ion:228 Resp: 81924

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 27.0 | 21.3 | 31.9 |
| 229 | 19.5 | 15.5 | 23.3 |

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#82
3,3'-Dichlorobenzidine
Concen: 63.758 ng
RT: 14.080 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:252 Resp: 254086

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 65.3 | 51.5 | 77.3 |
| 126 | 10.9 | 9.4 | 14.0 |






#83
Chrysene
Concen: 60.044 ng
RT: 14.157 min Scan# 2(
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:228 Resp: 75891(

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 29.6 | 23.4 | 35.2 |
| 229 | 19.7 | 15.3 | 22.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 61.957 ng
RT: 14.110 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:149 Resp: 364902

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 27.4 | 21.9 | 32.9 |
| 279 | 4.2 | 3.4 | 5.0 |






#85
Di-n-octyl phthalate
Concen: 60.817 ng
RT: 14.727 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:149 Resp: 68894

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.5 | 1.2 | 1.8 |
| 43 | 12.4 | 10.6 | 16.0 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.639 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:264 Resp: 260121

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.4 | 18.5 | 27.7 |
| 265 | 21.4 | 17.5 | 26.3 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 61.286 ng
RT: 17.192 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:276 Resp: 118570

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 27.6 | 22.6 | 34.0 |
| 277 | 25.4 | 20.2 | 30.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#88
Benzo(b)fluoranthene
Concen: 61.640 ng
RT: 15.192 min Scan# 2210
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:252 Resp: 924866

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.0 | 17.5 | 26.3 |
| 125 | 10.5 | 8.6 | 13.0 |






#89
Benzo(k)fluoranthene
Concen: 57.794 ng
RT: 15.227 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:252 Resp: 82727

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.7 | 17.4 | 26.2 |
| 125 | 10.2 | 8.5 | 12.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#90
Benzo(a)pyrene
Concen: 61.791 ng
RT: 15.574 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:252 Resp: 876856

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.1 | 17.5 | 26.3 |
| 125 | 11.6 | 9.4 | 14.0 |






#91
Dibenzo(a,h)anthracene
Concen: 60.300 ng
RT: 17.210 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:278 Resp: 95366

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 17.4 | 14.6 | 22.0 |
| 279 | 23.9 | 18.8 | 28.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#92
Benzo(g,h,i)perylene
Concen: 60.962 ng
RT: 17.657 min Scan# 2629
Delta R.T. 0.012 min
Lab File: BF143099.D
Acq: 15 Jul 2025 16:05

Tgt Ion:276 Resp: 982960

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.5 | 19.2 | 28.8 |
| 138 | 24.8 | 19.8 | 29.6 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143100.D
Acq On : 15 Jul 2025 16:35
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Jul 15 17:42:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| **Internal Standards** | | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 6.969 | 152 | 129576 | 20.000 | ng | | 0.00 |
| 21) Naphthalene-d8 | 8.251 | 136 | 487361 | 20.000 | ng | | 0.00 |
| 39) Acenaphthene-d10 | 10.010 | 164 | 236260 | 20.000 | ng | | 0.00 |
| 64) Phenanthrene-d10 | 11.492 | 188 | 362117 | 20.000 | ng | | 0.00 |
| 76) Chrysene-d12 | 14.133 | 240 | 196411 | 20.000 | ng | | 0.00 |
| 86) Perylene-d12 | 15.639 | 264 | 256691 | 20.000 | ng | | 0.00 |
| **System Monitoring Compounds** | | | | | | | |
| 5) 2-Fluorophenol | 5.593 | 112 | 1189824 | 145.077 | ng | | 0.00 |
| 7) Phenol-d6 | 6.593 | 99 | 1493965 | 144.917 | ng | | 0.00 |
| 23) Nitrobenzene-d5 | 7.534 | 82 | 1441614 | 165.933 | ng | | 0.00 |
| 42) 2,4,6-Tribromophenol | 10.798 | 330 | 355830 | 168.911 | ng | | 0.00 |
| 45) 2-Fluorobiphenyl | 9.328 | 172 | 2339151 | 131.609 | ng | | 0.00 |
| 79) Terphenyl-d14 | 13.075 | 244 | 1779829 | 132.466 | ng | | 0.00 |
| **Target Compounds** | | | | | | | Qvalue |
| 2) 1,4-Dioxane | 2.805 | 88 | 303852 | 74.427 | ng | | 100 |
| 3) Pyridine | 3.581 | 79 | 787323 | 76.158 | ng | | 99 |
| 4) n-Nitrosodimethylamine | 3.534 | 42 | 421072 | 78.614 | ng | | 100 |
| 6) Aniline | 6.634 | 93 | 1072162 | 73.855 | ng | | 100 |
| 8) 2-Chlorophenol | 6.757 | 128 | 628294 | 76.249 | ng | | 99 |
| 9) Benzaldehyde | 6.522 | 77 | 447861 | 57.014 | ng | | 99 |
| 10) Phenol | 6.610 | 94 | 820918m | 73.558 | ng | | |
| 11) bis(2-Chloroethyl)ether | 6.704 | 93 | 626821 | 72.551 | ng | | 98 |
| 12) 1,3-Dichlorobenzene | 6.916 | 146 | 666489 | 71.595 | ng | | 100 |
| 13) 1,4-Dichlorobenzene | 6.987 | 146 | 669586 | 71.496 | ng | | 99 |
| 14) 1,2-Dichlorobenzene | 7.145 | 146 | 643168 | 72.014 | ng | | 99 |
| 15) Benzyl Alcohol | 7.104 | 79 | 586953 | 76.840 | ng | | 98 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.245 | 45 | 1125357 | 70.661 | ng | | 98 |
| 17) 2-Methylphenol | 7.216 | 107 | 536428 | 75.438 | ng | | 99 |
| 18) Hexachloroethane | 7.487 | 117 | 235713 | 77.313 | ng | | 99 |
| 19) n-Nitroso-di-n-propyla... | 7.393 | 70 | 464705 | 73.578 | ng | | 98 |
| 20) 3+4-Methylphenols | 7.369 | 107 | 616275 | 70.670 | ng | | 95 |
| 22) Acetophenone | 7.381 | 105 | 817243 | 69.563 | ng | # | 95 |
| 24) Nitrobenzene | 7.551 | 77 | 676569 | 80.671 | ng | | 99 |
| 25) Isophorone | 7.798 | 82 | 1269015 | 75.573 | ng | | 100 |
| 26) 2-Nitrophenol | 7.863 | 139 | 287169 | 82.598 | ng | | 98 |
| 27) 2,4-Dimethylphenol | 7.898 | 122 | 582059 | 74.683 | ng | | 98 |
| 28) bis(2-Chloroethoxy)met... | 7.998 | 93 | 731977 | 71.885 | ng | | 100 |
| 29) 2,4-Dichlorophenol | 8.104 | 162 | 496027 | 77.795 | ng | | 99 |
| 30) 1,2,4-Trichlorobenzene | 8.192 | 180 | 520808 | 71.974 | ng | | 98 |
| 31) Naphthalene | 8.275 | 128 | 1711163 | 70.704 | ng | | 100 |
| 32) Benzoic acid | 8.022 | 122 | 336178 | 81.374 | ng | | 99 |
| 33) 4-Chloroaniline | 8.316 | 127 | 688999 | 70.795 | ng | | 99 |
| 34) Hexachlorobutadiene | 8.398 | 225 | 318274 | 73.546 | ng | | 98 |
| 35) Caprolactam | 8.698 | 113 | 161466 | 81.097 | ng | | 94 |
| 36) 4-Chloro-3-methylphenol | 8.792 | 107 | 539340 | 75.937 | ng | | 97 |
| 37) 2-Methylnaphthalene | 8.963 | 142 | 1021312 | 70.738 | ng | | 99 |
| 38) 1-Methylnaphthalene | 9.063 | 142 | 1047560 | 69.955 | ng | | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 9.134 | 216 | 493246 | 71.639 | ng | | 99 |
| 41) Hexachlorocyclopentadiene | 9.128 | 237 | 324865 | 85.103 | ng | | 98 |
| 43) 2,4,6-Trichlorophenol | 9.239 | 196 | 355839 | 83.176 | ng | | 98 |

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143100.D
Acq On : 15 Jul 2025 16:35
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Jul 15 17:42:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 9.275 | 196 | 357298 | 79.651 | ng | | 99 |
| 46) 1,1'-Biphenyl | 9.428 | 154 | 1307947 | 69.419 | ng | | 99 |
| 47) 2-Chloronaphthalene | 9.457 | 162 | 990436 | 71.557 | ng | | 99 |
| 48) 2-Nitroaniline | 9.539 | 65 | 335875 | 79.074 | ng | | 99 |
| 49) Acenaphthylene | 9.875 | 152 | 1650097 | 71.544 | ng | | 99 |
| 50) Dimethylphthalate | 9.728 | 163 | 1104756 | 74.327 | ng | | 100 |
| 51) 2,6-Dinitrotoluene | 9.787 | 165 | 233478 | 77.660 | ng | | 91 |
| 52) Acenaphthene | 10.045 | 154 | 975285 | 71.657 | ng | | 99 |
| 53) 3-Nitroaniline | 9.951 | 138 | 279161 | 86.920 | ng | | 99 |
| 54) 2,4-Dinitrophenol | 10.057 | 184 | 91751 | 79.064 | ng | # | 1 |
| 55) Dibenzofuran | 10.216 | 168 | 1432458 | 70.610 | ng | | 100 |
| 56) 4-Nitrophenol | 10.098 | 139 | 219224 | 85.886 | ng | | 97 |
| 57) 2,4-Dinitrotoluene | 10.186 | 165 | 302558 | 80.048 | ng | | 99 |
| 58) Fluorene | 10.557 | 166 | 1043990 | 68.640 | ng | | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 10.328 | 232 | 291933 | 81.893 | ng | | 99 |
| 60) Diethylphthalate | 10.428 | 149 | 1067815 | 74.464 | ng | | 99 |
| 61) 4-Chlorophenyl-phenyle... | 10.545 | 204 | 524910 | 72.005 | ng | | 98 |
| 62) 4-Nitroaniline | 10.569 | 138 | 244461 | 87.527 | ng | | 97 |
| 63) Azobenzene | 10.710 | 77 | 1141613 | 71.985 | ng | | 98 |
| 65) 4,6-Dinitro-2-methylph... | 10.598 | 198 | 125251 | 78.600 | ng | | 96 |
| 66) n-Nitrosodiphenylamine | 10.663 | 169 | 947890 | 74.520 | ng | | 99 |
| 67) 4-Bromophenyl-phenylether | 11.039 | 248 | 316619 | 76.885 | ng | | 97 |
| 68) Hexachlorobenzene | 11.110 | 284 | 333899 | 76.417 | ng | | 96 |
| 69) Atrazine | 11.186 | 200 | 266668 | 81.943 | ng | | 97 |
| 70) Pentachlorophenol | 11.292 | 266 | 210729 | 87.905 | ng | | 99 |
| 71) Phenanthrene | 11.522 | 178 | 1401500 | 71.714 | ng | | 99 |
| 72) Anthracene | 11.569 | 178 | 1423259 | 72.421 | ng | | 99 |
| 73) Carbazole | 11.716 | 167 | 1264053 | 71.796 | ng | | 100 |
| 74) Di-n-butylphthalate | 12.051 | 149 | 1299711 | 80.422 | ng | | 100 |
| 75) Fluoranthene | 12.704 | 202 | 1258479 | 70.821 | ng | | 99 |
| 77) Benzidine | 12.816 | 184 | 403599 | 60.315 | ng | | 99 |
| 78) Pyrene | 12.933 | 202 | 1244953 | 68.315 | ng | | 99 |
| 80) Butylbenzylphthalate | 13.551 | 149 | 301525 | 79.914 | ng | | 99 |
| 81) Benzo(a)anthracene | 14.122 | 228 | 979909 | 74.870 | ng | | 99 |
| 82) 3,3'-Dichlorobenzidine | 14.080 | 252 | 296061 | 77.896 | ng | | 99 |
| 83) Chrysene | 14.157 | 228 | 913035 | 75.743 | ng | | 98 |
| 84) Bis(2-ethylhexyl)phtha... | 14.110 | 149 | 458210 | 80.495 | ng | | 99 |
| 85) Di-n-octyl phthalate | 14.727 | 149 | 916748 | 82.209 | ng | | 98 |
| 87) Indeno(1,2,3-cd)pyrene | 17.192 | 276 | 1509813 | 79.082 | ng | | 98 |
| 88) Benzo(b)fluoranthene | 15.198 | 252 | 1115815 | 75.360 | ng | | 99 |
| 89) Benzo(k)fluoranthene | 15.227 | 252 | 1083939 | 76.736 | ng | | 99 |
| 90) Benzo(a)pyrene | 15.580 | 252 | 1107280 | 79.071 | ng | | 99 |
| 91) Dibenzo(a,h)anthracene | 17.215 | 278 | 1209971 | 77.528 | ng | | 98 |
| 92) Benzo(g,h,i)perylene | 17.668 | 276 | 1246206 | 78.321 | ng | | 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143100.D
Acq On : 15 Jul 2025 16:35
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Jul 15 17:42:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:08:07 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.969 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:152 Resp: 12957

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 160.9 | 126.5 | 189.7 |
| 115 | 60.0 | 47.0 | 70.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#2
1,4-Dioxane
Concen: 74.427 ng
RT: 2.805 min Scan# 104
Delta R.T. -0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 88 Resp: 303852

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 87.8 | 70.5 | 105.7 |
| 43 | 33.9 | 27.3 | 40.9 |






#3
Pyridine
Concen: 76.158 ng
RT: 3.581 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 79 Resp: 78732

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 84.9 | 69.2 | 103.8 |
| 51 | 42.0 | 33.3 | 49.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#4
n-Nitrosodimethylamine
Concen: 78.614 ng
RT: 3.534 min Scan# 228
Delta R.T. 0.012 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 42 Resp: 421072

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 115.9 | 92.8 | 139.2 |
| 44 | 7.0 | 5.8 | 8.6 |






#5
2-Fluorophenol
Concen: 145.077 ng
RT: 5.593 min Scan# 5
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:112 Resp: 118982

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 66.4 | 54.6 | 82.0 |
| 63 | 33.7 | 27.3 | 40.9 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#6
Aniline
Concen: 73.855 ng
RT: 6.634 min Scan# 755
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 93 Resp: 1072162

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 41.6 | 33.1 | 49.7 |
| 65 | 21.1 | 16.9 | 25.3 |






#7
Phenol-d6
Concen: 144.917 ng
RT: 6.593 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 99 Resp: 149396

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 21.3 | 17.2 | 25.8 |
| 71 | 35.4 | 27.4 | 41.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#8
2-Chlorophenol
Concen: 76.249 ng
RT: 6.757 min Scan# 776
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:128 Resp: 628294

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.7 | 12.8 | 52.8 |
| 64 | 54.7 | 36.1 | 76.1 |






#9
Benzaldehyde
Concen: 57.014 ng
RT: 6.522 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 77 Resp: 44786

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 95.6 | 76.3 | 116.3 |
| 106 | 93.6 | 72.8 | 112.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#10
Phenol
Concen: 73.558 ng m
RT: 6.610 min Scan# 751
Delta R.T. 0.012 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 94 Resp: 820918

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 32.1 | 9.8 | 49.8 |
| 66 | 44.8 | 20.5 | 60.5 |






#11
bis(2-Chloroethyl)ether
Concen: 72.551 ng
RT: 6.704 min Scan# 7
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 93 Resp: 62682

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 84.0 | 61.9 | 101.9 |
| 95 | 32.3 | 12.2 | 52.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#12
1,3-Dichlorobenzene
Concen: 71.595 ng
RT: 6.916 min Scan# 803
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:146 Resp: 666489

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.7 | 50.1 | 75.1 |
| 75 | 33.5 | 27.4 | 41.0 |






#13
1,4-Dichlorobenzene
Concen: 71.496 ng
RT: 6.987 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:146 Resp: 66958

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.0 | 50.6 | 76.0 |
| 111 | 44.3 | 35.0 | 52.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#14
1,2-Dichlorobenzene
Concen: 72.014 ng
RT: 7.145 min Scan# 842
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:146 Resp: 643168

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.6 | 49.8 | 74.6 |
| 111 | 44.8 | 35.4 | 53.0 |






#15
Benzyl Alcohol
Concen: 76.840 ng
RT: 7.104 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 79 Resp: 58695

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 74.6 | 61.1 | 91.7 |
| 77 | 60.9 | 49.9 | 74.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 70.661 ng
RT: 7.245 min Scan# 859
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 45 Resp: 1125357

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 11.8 | 0.0 | 31.1 |
| 79 | 8.9 | 0.0 | 28.4 |






#17
2-Methylphenol
Concen: 75.438 ng
RT: 7.216 min Scan# 8
Delta R.T. 0.012 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:107 Resp: 53642

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 113.8 | 90.9 | 136.3 |
| 77 | 45.6 | 37.5 | 56.3 |
| 79 | 46.4 | 36.9 | 55.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#18
Hexachloroethane
Concen: 77.313 ng
RT: 7.487 min Scan# 900
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:117 Resp: 235713

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 95.2 | 76.7 | 115.1 |
| 201 | 92.5 | 75.0 | 112.4 |






#19
n-Nitroso-di-n-propylamine
Concen: 73.578 ng
RT: 7.393 min Scan# 8
Delta R.T. 0.018 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 70 Resp: 46470

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 64.5 | 53.2 | 79.8 |
| 101 | 9.6 | 7.5 | 11.3 |
| 130 | 21.1 | 15.7 | 23.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#20
3+4-Methylphenols
Concen: 70.670 ng
RT: 7.369 min Scan# 880
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:107 Resp: 616275

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 89.3 | 71.3 | 111.3 |
| 77 | 88.3 | 77.2 | 117.2 |
| 79 | 30.3 | 10.7 | 50.7 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.251 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:136 Resp: 48736

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.9 | 8.9 | 13.3 |
| 54 | 10.3 | 8.3 | 12.5 |
| 68 | 6.8 | 5.3 | 7.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#22
Acetophenone
Concen: 69.563 ng
RT: 7.381 min Scan# 882
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:105 Resp: 817243

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 6.4 | 7.2 | 10.8# |
| 51 | 37.7 | 27.0 | 40.4 |
| 120 | 23.3 | 18.2 | 27.2 |






#23
Nitrobenzene-d5
Concen: 165.933 ng
RT: 7.534 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 82 Resp: 144161

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 41.5 | 33.3 | 49.9 |
| 54 | 58.4 | 47.4 | 71.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#24
Nitrobenzene
Concen: 80.671 ng
RT: 7.551 min Scan# 911
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 77 Resp: 676569

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 42.8 | 34.5 | 51.7 |
| 65 | 14.3 | 11.3 | 16.9 |






#25
Isophorone
Concen: 75.573 ng
RT: 7.798 min Scan# 9
Delta R.T. 0.012 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 82 Resp: 126901

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.1 | 5.6 | 8.4 |
| 138 | 18.2 | 14.3 | 21.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#26
2-Nitrophenol
Concen: 82.598 ng
RT: 7.863 min Scan# 964
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:139 Resp: 287169

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 29.8 | 23.4 | 35.0 |
| 65 | 55.8 | 43.4 | 65.2 |






#27
2,4-Dimethylphenol
Concen: 74.683 ng
RT: 7.898 min Scan# 9
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:122 Resp: 58205

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 116.8 | 91.3 | 136.9 |
| 121 | 58.4 | 45.5 | 68.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#28
bis(2-Chloroethoxy)methane
Concen: 71.885 ng
RT: 7.998 min Scan# 987
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 93 Resp: 731977

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 33.0 | 26.2 | 39.4 |
| 123 | 12.1 | 9.6 | 14.4 |






#29
2,4-Dichlorophenol
Concen: 77.795 ng
RT: 8.104 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:162 Resp: 49602

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 63.9 | 43.1 | 83.1 |
| 98 | 39.4 | 19.2 | 59.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#30
1,2,4-Trichlorobenzene
Concen: 71.974 ng
RT: 8.192 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:180 Resp: 520808

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 95.0 | 74.7 | 112.1 |
| 145 | 32.0 | 24.5 | 36.7 |






#31
Naphthalene
Concen: 70.704 ng
RT: 8.275 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:128 Resp: 171116

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.1 | 9.0 | 13.4 |
| 127 | 13.2 | 10.5 | 15.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#32
Benzoic acid
Concen: 81.374 ng
RT: 8.022 min Scan# 991
Delta R.T. 0.035 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:122 Resp: 336178

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 127.0 | 105.4 | 145.4 |
| 77 | 98.6 | 78.6 | 118.6 |






#33
4-Chloroaniline
Concen: 70.795 ng
RT: 8.316 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:127 Resp: 68899

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.5 | 26.2 | 39.4 |
| 65 | 35.0 | 28.6 | 42.8 |
| 92 | 21.3 | 17.1 | 25.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#34
Hexachlorobutadiene
Concen: 73.546 ng
RT: 8.398 min Scan# 1055
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:225 Resp: 318274

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 63.9 | 50.3 | 75.5 |
| 227 | 65.0 | 50.1 | 75.1 |






#35
Caprolactam
Concen: 81.097 ng
RT: 8.698 min Scan# 1
Delta R.T. 0.030 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:113 Resp: 16146

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 195.6 | 186.2 | 226.2 |
| 56 | 148.8 | 134.1 | 174.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#36
4-Chloro-3-methylphenol
Concen: 75.937 ng
RT: 8.792 min Scan# 1122
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:107 Resp: 539340

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 23.8 | 20.2 | 30.2 |
| 142 | 75.5 | 62.4 | 93.6 |






#37
2-Methylnaphthalene
Concen: 70.738 ng
RT: 8.963 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:142 Resp: 102131

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 86.4 | 70.2 | 105.2 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#38
1-Methylnaphthalene
Concen: 69.955 ng
RT: 9.063 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:142 Resp: 1047560

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 90.7 | 72.2 | 108.4 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.010 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:164 Resp: 23626

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 98.7 | 79.0 | 118.6 |
| 160 | 44.7 | 35.8 | 53.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 71.639 ng
RT: 9.134 min Scan# 1180
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:216 Resp: 493246

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 79.8 | 62.6 | 94.0 |
| 179 | 20.5 | 16.2 | 24.2 |
| 108 | 20.0 | 15.8 | 23.6 |






#41
Hexachlorocyclopentadiene
Concen: 85.103 ng
RT: 9.128 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:237 Resp: 32486

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 62.7 | 41.6 | 81.6 |
| 272 | 13.8 | 0.0 | 32.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#42
2,4,6-Tribromophenol
Concen: 168.911 ng
RT: 10.798 min Scan# 1463
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:330 Resp: 355830

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 96.3 | 77.9 | 116.9 |
| 141 | 29.2 | 25.9 | 38.9 |






#43
2,4,6-Trichlorophenol
Concen: 83.176 ng
RT: 9.239 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:196 Resp: 35583

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 94.1 | 76.9 | 115.3 |
| 200 | 30.9 | 23.5 | 35.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#44
2,4,5-Trichlorophenol
Concen: 79.651 ng
RT: 9.275 min Scan# 1204
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:196 Resp: 357298

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 94.5 | 76.4 | 114.6 |
| 97 | 52.7 | 42.4 | 63.6 |
| 132 | 29.7 | 23.9 | 35.9 |






#45
2-Fluorobiphenyl
Concen: 131.609 ng
RT: 9.328 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:172 Resp: 233915

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.9 | 28.2 | 42.4 |
| 170 | 23.7 | 18.6 | 28.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#46
1,1'-Biphenyl
Concen: 69.419 ng
RT: 9.428 min Scan# 1230
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:154 Resp: 1307947

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.6 | 20.1 | 60.1 |
| 76 | 13.9 | 0.0 | 33.7 |






#47
2-Chloronaphthalene
Concen: 71.557 ng
RT: 9.457 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:162 Resp: 99043

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 41.8 | 33.0 | 49.6 |
| 164 | 33.2 | 26.1 | 39.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#48
2-Nitroaniline
Concen: 79.074 ng
RT: 9.539 min Scan# 1249
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 65 Resp: 335875

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 65.2 | 51.3 | 76.9 |
| 138 | 98.6 | 77.8 | 116.6 |






#49
Acenaphthylene
Concen: 71.544 ng
RT: 9.875 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:152 Resp: 165097

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.2 | 15.8 | 23.8 |
| 153 | 13.2 | 10.4 | 15.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#50
Dimethylphthalate
Concen: 74.327 ng
RT: 9.728 min Scan# 1281
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:163 Resp: 1104756

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.4 | 3.4 | 5.2 |
| 164 | 10.3 | 8.2 | 12.2 |






#51
2,6-Dinitrotoluene
Concen: 77.660 ng
RT: 9.787 min Scan# 1
Delta R.T. 0.011 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:165 Resp: 23347

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 63.3 | 57.4 | 86.0 |
| 89 | 59.4 | 51.8 | 77.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#52
Acenaphthene
Concen: 71.657 ng
RT: 10.045 min Scan# 1335
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:154 Resp: 975285

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 107.3 | 86.7 | 130.1 |
| 152 | 50.5 | 41.4 | 62.2 |






#53
3-Nitroaniline
Concen: 86.920 ng
RT: 9.951 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:138 Resp: 27916

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 11.6 | 10.1 | 15.1 |
| 92 | 121.4 | 96.4 | 144.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#54
2,4-Dinitrophenol
Concen: 79.064 ng
RT: 10.057 min Scan# 1337
Delta R.T. 0.012 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:184 Resp: 91751

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 89.3 | 149.3 | 223.9# |
| 154 | 352.7 | 1048.0 | 1572.0# |






#55
Dibenzofuran
Concen: 70.610 ng
RT: 10.216 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:168 Resp: 143245

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 37.4 | 29.8 | 44.6 |
| 169 | 13.5 | 10.8 | 16.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#56
4-Nitrophenol
Concen: 85.886 ng
RT: 10.098 min Scan# 1344
Delta R.T. 0.012 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:139 Resp: 219224

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 70.2 | 49.4 | 89.4 |
| 65 | 103.6 | 88.2 | 128.2 |






#57
2,4-Dinitrotoluene
Concen: 80.048 ng
RT: 10.186 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:165 Resp: 30255

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 50.4 | 40.6 | 61.0 |
| 89 | 76.3 | 62.3 | 93.5 |
| 182 | 2.9 | 2.6 | 3.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#58
Fluorene
Concen: 68.640 ng
RT: 10.557 min Scan# 1422
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:166 Resp: 1043990

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 92.9 | 73.7 | 110.5 |
| 167 | 13.6 | 10.9 | 16.3 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 81.893 ng
RT: 10.328 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:232 Resp: 29193

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 44.7 | 36.9 | 55.3 |
| 130 | 2.2 | 1.8 | 2.8 |
| 166 | 29.0 | 23.0 | 34.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#60
Diethylphthalate
Concen: 74.464 ng
RT: 10.428 min Scan# 1400
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:149 Resp: 1067815

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 22.2 | 17.3 | 25.9 |
| 150 | 12.8 | 9.9 | 14.9 |






#61
4-Chlorophenyl-phenylether
Concen: 72.005 ng
RT: 10.545 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:204 Resp: 52491

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 32.7 | 25.8 | 38.6 |
| 141 | 60.2 | 46.8 | 70.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#62
4-Nitroaniline
Concen: 87.527 ng
RT: 10.569 min Scan# 1424
Delta R.T. 0.012 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:138 Resp: 244461

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 51.4 | 30.4 | 70.4 |
| 108 | 72.9 | 49.5 | 89.5 |






#63
Azobenzene
Concen: 71.985 ng
RT: 10.710 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion: 77 Resp: 114161

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 25.4 | 4.0 | 44.0 |
| 105 | 21.7 | 1.4 | 41.4 |
| 51 | 33.3 | 14.3 | 54.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.492 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:188 Resp: 362117

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 11.0 | 8.6 | 13.0 |
| 80 | 12.0 | 9.3 | 13.9 |






#65
4,6-Dinitro-2-methylphenol
Concen: 78.600 ng
RT: 10.598 min Scan# 14
Delta R.T. 0.012 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:198 Resp: 12525

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 59.8 | 44.6 | 84.6 |
| 105 | 47.9 | 29.3 | 69.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#66
n-Nitrosodiphenylamine
Concen: 74.520 ng
RT: 10.663 min Scan# 1440
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:169 Resp: 947890

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 65.5 | 53.0 | 79.6 |
| 167 | 35.8 | 28.4 | 42.6 |






#67
4-Bromophenyl-phenylether
Concen: 76.885 ng
RT: 11.039 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:248 Resp: 31661

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 98.3 | 76.7 | 115.1 |
| 141 | 73.1 | 61.7 | 92.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#68
Hexachlorobenzene
Concen: 76.417 ng
RT: 11.110 min Scan# 1516
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:284 Resp: 333899

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 31.8 | 28.9 | 43.3 |
| 249 | 30.5 | 24.6 | 37.0 |






#69
Atrazine
Concen: 81.943 ng
RT: 11.186 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:200 Resp: 26666

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 27.6 | 9.1 | 49.1 |
| 215 | 49.2 | 31.0 | 71.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#70
Pentachlorophenol
Concen: 87.905 ng
RT: 11.292 min Scan# 1547
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:266 Resp: 210729

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 63.0 | 51.8 | 77.6 |
| 264 | 61.8 | 49.7 | 74.5 |






#71
Phenanthrene
Concen: 71.714 ng
RT: 11.522 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:178 Resp: 140150

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.3 | 15.1 | 22.7 |
| 179 | 15.6 | 12.2 | 18.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#72
Anthracene
Concen: 72.421 ng
RT: 11.569 min Scan# 1594
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:178 Resp: 1423259

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.7 | 14.7 | 22.1 |
| 179 | 15.6 | 12.4 | 18.6 |






#73
Carbazole
Concen: 71.796 ng
RT: 11.716 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:167 Resp: 126405

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.1 | 17.0 | 25.4 |
| 139 | 13.0 | 10.2 | 15.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#74
Di-n-butylphthalate
Concen: 80.422 ng
RT: 12.051 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:149 Resp: 1299711

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.3 | 7.3 | 10.9 |
| 104 | 4.8 | 3.8 | 5.8 |






#75
Fluoranthene
Concen: 70.821 ng
RT: 12.704 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:202 Resp: 125847

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 12.1 | 0.0 | 32.4 |
| 203 | 17.9 | 0.0 | 37.4 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.133 min Scan# 2030
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:240 Resp: 196411

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.3 | 10.0 | 15.0 |
| 236 | 25.0 | 20.0 | 30.0 |






#77
Benzidine
Concen: 60.315 ng
RT: 12.816 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:184 Resp: 40359

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 15.9 | 12.4 | 18.6 |
| 183 | 11.1 | 8.6 | 12.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#78
Pyrene
Concen: 68.315 ng
RT: 12.933 min Scan# 1826
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:202 Resp: 1244953

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 19.9 | 15.5 | 23.3 |
| 203 | 17.7 | 14.0 | 21.0 |






#79
Terphenyl-d14
Concen: 132.466 ng
RT: 13.075 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:244 Resp: 177982

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.7 | 5.2 | 7.8 |
| 122 | 11.0 | 8.8 | 13.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#80
Butylbenzylphthalate
Concen: 79.914 ng
RT: 13.551 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:149 Resp: 301525

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 70.6 | 57.0 | 85.4 |
| 206 | 21.9 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 74.870 ng
RT: 14.122 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:228 Resp: 97990

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 27.0 | 21.3 | 31.9 |
| 229 | 19.7 | 15.5 | 23.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#82
3,3'-Dichlorobenzidine
Concen: 77.896 ng
RT: 14.080 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:252 Resp: 296061

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 65.0 | 51.5 | 77.3 |
| 126 | 11.4 | 9.4 | 14.0 |






#83
Chrysene
Concen: 75.743 ng
RT: 14.157 min Scan# 2(
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:228 Resp: 91303

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 30.3 | 23.4 | 35.2 |
| 229 | 19.8 | 15.3 | 22.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 80.495 ng
RT: 14.110 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:149 Resp: 458210

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 27.9 | 21.9 | 32.9 |
| 279 | 4.3 | 3.4 | 5.0 |






#85
Di-n-octyl phthalate
Concen: 82.209 ng
RT: 14.727 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:149 Resp: 91674

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.5 | 1.2 | 1.8 |
| 43 | 12.3 | 10.6 | 16.0 |

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.639 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:264 Resp: 256691

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 22.6 | 18.5 | 27.7 |
| 265 | 21.4 | 17.5 | 26.3 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 79.082 ng
RT: 17.192 min Scan# 2
Delta R.T. 0.012 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:276 Resp: 150981

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 26.8 | 22.6 | 34.0 |
| 277 | 25.3 | 20.2 | 30.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#88
Benzo(b)fluoranthene
Concen: 75.360 ng
RT: 15.198 min Scan# 2211
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:252 Resp: 1115815

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.1 | 17.5 | 26.3 |
| 125 | 9.8 | 8.6 | 13.0 |






#89
Benzo(k)fluoranthene
Concen: 76.736 ng
RT: 15.227 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:252 Resp: 108393

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.8 | 17.4 | 26.2 |
| 125 | 10.0 | 8.5 | 12.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#90
Benzo(a)pyrene
Concen: 79.071 ng
RT: 15.580 min Scan# 2276
Delta R.T. 0.006 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:252 Resp: 1107280

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.0 | 17.5 | 26.3 |
| 125 | 11.3 | 9.4 | 14.0 |






#91
Dibenzo(a,h)anthracene
Concen: 77.528 ng
RT: 17.215 min Scan# 2
Delta R.T. 0.018 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:278 Resp: 120997

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 16.7 | 14.6 | 22.0 |
| 279 | 23.5 | 18.8 | 28.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#92
Benzo(g,h,i)perylene
Concen: 78.321 ng
RT: 17.668 min Scan# 2631
Delta R.T. 0.024 min
Lab File: BF143100.D
Acq: 15 Jul 2025 16:35

Tgt Ion:276 Resp: 1246206

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.7 | 19.2 | 28.8 |
| 138 | 22.8 | 19.8 | 29.6 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143101.D
Acq On : 15 Jul 2025 17:27
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

Quant Time: Jul 15 17:54:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 6.969 | 152 | 130235 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 8.251 | 136 | 496378 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 10.004 | 164 | 248221 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 11.492 | 188 | 398969 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 14.133 | 240 | 215152 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 15.639 | 264 | 235493 | 20.000 | ng | 0.00 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.587 | 112 | 649544 | 78.799 | ng | 0.00 |
| 7) Phenol-d6 | 6.587 | 99 | 814393 | 78.598 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 7.528 | 82 | 692735m | 78.289 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 10.792 | 330 | 190277 | 85.971 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 9.328 | 172 | 1290148 | 69.090 | ng | 0.00 |
| 79) Terphenyl-d14 | 13.074 | 244 | 1064372 | 72.317 | ng | 0.00 |
| Target Compounds | | | | | | Qvalue |
| 2) 1,4-Dioxane | 2.810 | 88 | 156917 | 38.242 | ng | 99 |
| 3) Pyridine | 3.581 | 79 | 426106 | 41.009 | ng | 99 |
| 4) n-Nitrosodimethylamine | 3.522 | 42 | 223240 | 41.468 | ng | 98 |
| 6) Aniline | 6.628 | 93 | 598787 | 41.038 | ng | 100 |
| 8) 2-Chlorophenol | 6.751 | 128 | 351129 | 42.397 | ng | 99 |
| 9) Benzaldehyde | 6.516 | 77 | 293033 | 37.115 | ng | 99 |
| 10) Phenol | 6.598 | 94 | 467536m | 41.681 | ng | |
| 11) bis(2-Chloroethyl)ether | 6.698 | 93 | 354713 | 40.848 | ng | 99 |
| 12) 1,3-Dichlorobenzene | 6.910 | 146 | 380756 | 40.694 | ng | 100 |
| 13) 1,4-Dichlorobenzene | 6.987 | 146 | 381590 | 40.539 | ng | 100 |
| 14) 1,2-Dichlorobenzene | 7.139 | 146 | 366082 | 40.782 | ng | 99 |
| 15) Benzyl Alcohol | 7.098 | 79 | 325035 | 42.336 | ng | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.239 | 45 | 644651 | 40.273 | ng | 100 |
| 17) 2-Methylphenol | 7.204 | 107 | 296835 | 41.532 | ng | 99 |
| 18) Hexachloroethane | 7.486 | 117 | 132499 | 43.239 | ng | 100 |
| 19) n-Nitroso-di-n-propyla... | 7.375 | 70 | 266499 | 41.982 | ng | 100 |
| 20) 3+4-Methylphenols | 7.363 | 107 | 368618 | 42.057 | ng | 99 |
| 22) Acetophenone | 7.375 | 105 | 482823 | 40.351 | ng | 99 |
| 24) Nitrobenzene | 7.545 | 77 | 381374 | 44.647 | ng | 99 |
| 25) Isophorone | 7.786 | 82 | 704424 | 41.188 | ng | 100 |
| 26) 2-Nitrophenol | 7.863 | 139 | 150156 | 44.317 | ng | 100 |
| 27) 2,4-Dimethylphenol | 7.892 | 122 | 333899 | 42.064 | ng | 100 |
| 28) bis(2-Chloroethoxy)met... | 7.992 | 93 | 423022 | 40.789 | ng | 100 |
| 29) 2,4-Dichlorophenol | 8.098 | 162 | 283011 | 43.580 | ng | 99 |
| 30) 1,2,4-Trichlorobenzene | 8.192 | 180 | 303602 | 41.194 | ng | 99 |
| 31) Naphthalene | 8.275 | 128 | 999363 | 40.543 | ng | 99 |
| 32) Benzoic acid | 7.986 | 122 | 169020m | 43.449 | ng | |
| 33) 4-Chloroaniline | 8.310 | 127 | 408741 | 41.236 | ng | 99 |
| 34) Hexachlorobutadiene | 8.392 | 225 | 182140 | 41.324 | ng | 99 |
| 35) Caprolactam | 8.675 | 113 | 88286 | 43.536 | ng | 95 |
| 36) 4-Chloro-3-methylphenol | 8.786 | 107 | 306624 | 42.387 | ng | 99 |
| 37) 2-Methylnaphthalene | 8.963 | 142 | 597819 | 40.654 | ng | 100 |
| 38) 1-Methylnaphthalene | 9.063 | 142 | 624130 | 40.922 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 9.128 | 216 | 293187 | 40.531 | ng | 99 |
| 41) Hexachlorocyclopentadiene | 9.122 | 237 | 180819 | 45.086 | ng | 98 |
| 43) 2,4,6-Trichlorophenol | 9.233 | 196 | 197565 | 43.955 | ng | 98 |

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143101.D
Acq On : 15 Jul 2025 17:27
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Jul 15 17:54:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 9.269 | 196 | 202658 | 43.001 | ng | | 99 |
| 46) 1,1'-Biphenyl | 9.428 | 154 | 799467 | 40.387 | ng | | 100 |
| 47) 2-Chloronaphthalene | 9.451 | 162 | 584687 | 40.207 | ng | | 100 |
| 48) 2-Nitroaniline | 9.539 | 65 | 185847 | 43.148 | ng | | 96 |
| 49) Acenaphthylene | 9.869 | 152 | 987106 | 40.736 | ng | | 99 |
| 50) Dimethylphthalate | 9.722 | 163 | 644618 | 41.279 | ng | | 100 |
| 51) 2,6-Dinitrotoluene | 9.780 | 165 | 132241 | 43.156 | ng | | 92 |
| 52) Acenaphthene | 10.039 | 154 | 571724 | 39.982 | ng | | 99 |
| 53) 3-Nitroaniline | 9.945 | 138 | 160978 | 47.707 | ng | | 99 |
| 54) 2,4-Dinitrophenol | 10.051 | 184 | 46465 | 47.113 | ng | # | 1 |
| 55) Dibenzofuran | 10.210 | 168 | 858944 | 40.300 | ng | | 100 |
| 56) 4-Nitrophenol | 10.092 | 139 | 122610 | 45.720 | ng | | 96 |
| 57) 2,4-Dinitrotoluene | 10.180 | 165 | 168227 | 43.964 | ng | | 99 |
| 58) Fluorene | 10.557 | 166 | 634897 | 39.732 | ng | | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 10.327 | 232 | 168172 | 44.902 | ng | | 98 |
| 60) Diethylphthalate | 10.422 | 149 | 637247 | 42.297 | ng | | 99 |
| 61) 4-Chlorophenyl-phenyle... | 10.545 | 204 | 313111 | 40.882 | ng | | 98 |
| 62) 4-Nitroaniline | 10.557 | 138 | 139342 | 47.486 | ng | | 98 |
| 63) Azobenzene | 10.704 | 77 | 680354 | 40.833 | ng | | 99 |
| 65) 4,6-Dinitro-2-methylph... | 10.586 | 198 | 65266 | 44.754 | ng | | 98 |
| 66) n-Nitrosodiphenylamine | 10.663 | 169 | 563086 | 40.179 | ng | | 100 |
| 67) 4-Bromophenyl-phenylether | 11.033 | 248 | 189573 | 41.782 | ng | | 100 |
| 68) Hexachlorobenzene | 11.104 | 284 | 196969 | 40.915 | ng | | 99 |
| 69) Atrazine | 11.180 | 200 | 158741 | 44.273 | ng | | 98 |
| 70) Pentachlorophenol | 11.292 | 266 | 118797 | 44.978 | ng | | 99 |
| 71) Phenanthrene | 11.516 | 178 | 871369 | 40.469 | ng | | 100 |
| 72) Anthracene | 11.569 | 178 | 887738 | 40.999 | ng | | 100 |
| 73) Carbazole | 11.716 | 167 | 791886 | 40.823 | ng | | 99 |
| 74) Di-n-butylphthalate | 12.051 | 149 | 820833 | 46.099 | ng | | 100 |
| 75) Fluoranthene | 12.704 | 202 | 824497 | 42.113 | ng | | 100 |
| 77) Benzidine | 12.816 | 184 | 350573 | 47.827 | ng | | 98 |
| 78) Pyrene | 12.933 | 202 | 820201 | 41.087 | ng | | 100 |
| 80) Butylbenzylphthalate | 13.551 | 149 | 172390 | 43.340 | ng | | 100 |
| 81) Benzo(a)anthracene | 14.121 | 228 | 583794 | 40.719 | ng | | 100 |
| 82) 3,3'-Dichlorobenzidine | 14.080 | 252 | 183371 | 44.044 | ng | | 99 |
| 83) Chrysene | 14.157 | 228 | 530018 | 40.139 | ng | | 98 |
| 84) Bis(2-ethylhexyl)phtha... | 14.110 | 149 | 239329 | 40.166 | ng | | 99 |
| 85) Di-n-octyl phthalate | 14.727 | 149 | 424959 | 38.648 | ng | | 100 |
| 87) Indeno(1,2,3-cd)pyrene | 17.186 | 276 | 744447 | 42.503 | ng | | 100 |
| 88) Benzo(b)fluoranthene | 15.192 | 252 | 594727 | 43.782 | ng | | 100 |
| 89) Benzo(k)fluoranthene | 15.221 | 252 | 513977 | 39.662 | ng | | 99 |
| 90) Benzo(a)pyrene | 15.574 | 252 | 548875 | 42.724 | ng | | 99 |
| 91) Dibenzo(a,h)anthracene | 17.203 | 278 | 611881 | 42.735 | ng | | 99 |
| 92) Benzo(g,h,i)perylene | 17.651 | 276 | 632939 | 43.359 | ng | | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143101.D
Acq On : 15 Jul 2025 17:27
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Jul 15 17:54:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.969 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:152 Resp: 13023

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 159.9 | 126.5 | 189.7 |
| 115 | 58.9 | 47.0 | 70.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#2
1,4-Dioxane
Concen: 38.242 ng
RT: 2.810 min Scan# 105
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 88 Resp: 156917

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 87.3 | 70.5 | 105.7 |
| 43 | 33.7 | 27.3 | 40.9 |






#3
Pyridine
Concen: 41.009 ng
RT: 3.581 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 79 Resp: 42610

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 86.0 | 69.2 | 103.8 |
| 51 | 41.1 | 33.3 | 49.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#4
n-Nitrosodimethylamine
Concen: 41.468 ng
RT: 3.522 min Scan# 226
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 42 Resp: 223240

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 114.0 | 92.8 | 139.2 |
| 44 | 7.3 | 5.8 | 8.6 |






#5
2-Fluorophenol
Concen: 78.799 ng
RT: 5.587 min Scan# 5
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:112 Resp: 64954

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 67.0 | 54.6 | 82.0 |
| 63 | 33.4 | 27.3 | 40.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#6
Aniline
Concen: 41.038 ng
RT: 6.628 min Scan# 754
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 93 Resp: 598787

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 41.5 | 33.1 | 49.7 |
| 65 | 20.7 | 16.9 | 25.3 |






#7
Phenol-d6
Concen: 78.598 ng
RT: 6.587 min Scan# 74
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 99 Resp: 81439

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 21.1 | 17.2 | 25.8 |
| 71 | 34.2 | 27.4 | 41.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#8
2-Chlorophenol
Concen: 42.397 ng
RT: 6.751 min Scan# 775
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:128 Resp: 351129

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.8 | 12.8 | 52.8 |
| 64 | 55.5 | 36.1 | 76.1 |






#9
Benzaldehyde
Concen: 37.115 ng
RT: 6.516 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 77 Resp: 29303

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 94.0 | 76.3 | 116.3 |
| 106 | 92.5 | 72.8 | 112.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#10
Phenol
Concen: 41.681 ng m
RT: 6.598 min Scan# 749
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 94 Resp: 467536

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 29.6 | 9.8 | 49.8 |
| 66 | 39.9 | 20.5 | 60.5 |






#11
bis(2-Chloroethyl)ether
Concen: 40.848 ng
RT: 6.698 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 93 Resp: 35471

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 82.7 | 61.9 | 101.9 |
| 95 | 32.3 | 12.2 | 52.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#12
1,3-Dichlorobenzene
Concen: 40.694 ng
RT: 6.910 min Scan# 802
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:146 Resp: 380756

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.7 | 50.1 | 75.1 |
| 75 | 34.2 | 27.4 | 41.0 |






#13
1,4-Dichlorobenzene
Concen: 40.539 ng
RT: 6.987 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:146 Resp: 38159

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.2 | 50.6 | 76.0 |
| 111 | 43.4 | 35.0 | 52.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#14
1,2-Dichlorobenzene
Concen: 40.782 ng
RT: 7.139 min Scan# 841
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:146 Resp: 366082

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.9 | 49.8 | 74.6 |
| 111 | 45.7 | 35.4 | 53.0 |






#15
Benzyl Alcohol
Concen: 42.336 ng
RT: 7.098 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 79 Resp: 32503

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 75.2 | 61.1 | 91.7 |
| 77 | 61.8 | 49.9 | 74.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.273 ng
RT: 7.239 min Scan# 858
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 45 Resp: 644651

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 11.3 | 0.0 | 31.1 |
| 79 | 8.4 | 0.0 | 28.4 |






#17
2-Methylphenol
Concen: 41.532 ng
RT: 7.204 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:107 Resp: 29683

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 115.5 | 90.9 | 136.3 |
| 77 | 46.3 | 37.5 | 56.3 |
| 79 | 46.8 | 36.9 | 55.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#18
Hexachloroethane
Concen: 43.239 ng
RT: 7.486 min Scan# 900
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:117 Resp: 132499

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 96.3 | 76.7 | 115.1 |
| 201 | 93.9 | 75.0 | 112.4 |






#19
n-Nitroso-di-n-propylamine
Concen: 41.982 ng
RT: 7.375 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 70 Resp: 26649

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 66.7 | 53.2 | 79.8 |
| 101 | 9.5 | 7.5 | 11.3 |
| 130 | 20.0 | 15.7 | 23.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#20
3+4-Methylphenols
Concen: 42.057 ng
RT: 7.363 min Scan# 879
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:107 Resp: 368618

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 92.5 | 71.3 | 111.3 |
| 77 | 98.3 | 77.2 | 117.2 |
| 79 | 30.7 | 10.7 | 50.7 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.251 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:136 Resp: 49637

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.9 | 8.9 | 13.3 |
| 54 | 9.9 | 8.3 | 12.5 |
| 68 | 6.7 | 5.3 | 7.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#22
Acetophenone
Concen: 40.351 ng
RT: 7.375 min Scan# 881
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:105 Resp: 482823

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 9.0 | 7.2 | 10.8 |
| 51 | 34.6 | 27.0 | 40.4 |
| 120 | 23.1 | 18.2 | 27.2 |






#23
Nitrobenzene-d5
Concen: 78.289 ng m
RT: 7.528 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 82 Resp: 69273

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 42.8 | 33.3 | 49.9 |
| 54 | 59.4 | 47.4 | 71.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#24
Nitrobenzene
Concen: 44.647 ng
RT: 7.545 min Scan# 910
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 77 Resp: 381374

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 43.9 | 34.5 | 51.7 |
| 65 | 13.9 | 11.3 | 16.9 |






#25
Isophorone
Concen: 41.188 ng
RT: 7.786 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 82 Resp: 70442

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.0 | 5.6 | 8.4 |
| 138 | 17.8 | 14.3 | 21.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#26
2-Nitrophenol
Concen: 44.317 ng
RT: 7.863 min Scan# 964
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:139 Resp: 150156

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 29.0 | 23.4 | 35.0 |
| 65 | 54.1 | 43.4 | 65.2 |






#27
2,4-Dimethylphenol
Concen: 42.064 ng
RT: 7.892 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:122 Resp: 33389

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 114.2 | 91.3 | 136.9 |
| 121 | 57.5 | 45.5 | 68.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#28
bis(2-Chloroethoxy)methane
Concen: 40.789 ng
RT: 7.992 min Scan# 986
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 93 Resp: 423022

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.6 | 26.2 | 39.4 |
| 123 | 11.9 | 9.6 | 14.4 |






#29
2,4-Dichlorophenol
Concen: 43.580 ng
RT: 8.098 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:162 Resp: 28301

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 63.7 | 43.1 | 83.1 |
| 98 | 39.6 | 19.2 | 59.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#30
1,2,4-Trichlorobenzene
Concen: 41.194 ng
RT: 8.192 min Scan# 1020
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:180 Resp: 303602

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 93.2 | 74.7 | 112.1 |
| 145 | 31.4 | 24.5 | 36.7 |






#31
Naphthalene
Concen: 40.543 ng
RT: 8.275 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:128 Resp: 99936

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.0 | 9.0 | 13.4 |
| 127 | 13.3 | 10.5 | 15.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#32
Benzoic acid
Concen: 43.449 ng m
RT: 7.986 min Scan# 985
Delta R.T. -0.001 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:122 Resp: 169020

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 128.4 | 105.4 | 145.4 |
| 77 | 100.3 | 78.6 | 118.6 |






#33
4-Chloroaniline
Concen: 41.236 ng
RT: 8.310 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:127 Resp: 40874

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.5 | 26.2 | 39.4 |
| 65 | 36.2 | 28.6 | 42.8 |
| 92 | 21.8 | 17.1 | 25.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#34
Hexachlorobutadiene
Concen: 41.324 ng
RT: 8.392 min Scan# 1054
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:225 Resp: 182140

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 63.6 | 50.3 | 75.5 |
| 227 | 63.1 | 50.1 | 75.1 |






#35
Caprolactam
Concen: 43.536 ng
RT: 8.675 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:113 Resp: 8828

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 197.8 | 186.2 | 226.2 |
| 56 | 148.9 | 134.1 | 174.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#36
4-Chloro-3-methylphenol
Concen: 42.387 ng
RT: 8.786 min Scan# 1121
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:107 Resp: 306624

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 24.4 | 20.2 | 30.2 |
| 142 | 77.2 | 62.4 | 93.6 |






#37
2-Methylnaphthalene
Concen: 40.654 ng
RT: 8.963 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:142 Resp: 59781

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 88.1 | 70.2 | 105.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#38
1-Methylnaphthalene
Concen: 40.922 ng
RT: 9.063 min Scan# 1168
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:142 Resp: 624130

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 89.9 | 72.2 | 108.4 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:164 Resp: 24822

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 98.1 | 79.0 | 118.6 |
| 160 | 43.6 | 35.8 | 53.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.531 ng
RT: 9.128 min Scan# 1179
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:216 Resp: 293187

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.6 | 62.6 | 94.0 |
| 179 | 20.8 | 16.2 | 24.2 |
| 108 | 19.6 | 15.8 | 23.6 |






#41
Hexachlorocyclopentadiene
Concen: 45.086 ng
RT: 9.122 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:237 Resp: 18081

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 63.5 | 41.6 | 81.6 |
| 272 | 13.0 | 0.0 | 32.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#42
2,4,6-Tribromophenol
Concen: 85.971 ng
RT: 10.792 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:330 Resp: 190277

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 98.2 | 77.9 | 116.9 |
| 141 | 31.8 | 25.9 | 38.9 |






#43
2,4,6-Trichlorophenol
Concen: 43.955 ng
RT: 9.233 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:196 Resp: 19756

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 94.2 | 76.9 | 115.3 |
| 200 | 30.9 | 23.5 | 35.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#44
2,4,5-Trichlorophenol
Concen: 43.001 ng
RT: 9.269 min Scan# 1203
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:196 Resp: 202658

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.1 | 76.4 | 114.6 |
| 97 | 55.6 | 42.4 | 63.6 |
| 132 | 29.9 | 23.9 | 35.9 |






#45
2-Fluorobiphenyl
Concen: 69.090 ng
RT: 9.328 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:172 Resp: 129014

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.3 | 28.2 | 42.4 |
| 170 | 23.4 | 18.6 | 28.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#46
1,1'-Biphenyl
Concen: 40.387 ng
RT: 9.428 min Scan# 1230
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:154 Resp: 799467

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.1 | 20.1 | 60.1 |
| 76 | 13.4 | 0.0 | 33.7 |






#47
2-Chloronaphthalene
Concen: 40.207 ng
RT: 9.451 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:162 Resp: 58468

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 41.3 | 33.0 | 49.6 |
| 164 | 33.0 | 26.1 | 39.1 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#48
2-Nitroaniline
Concen: 43.148 ng
RT: 9.539 min Scan# 1249
Delta R.T. 0.006 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 65 Resp: 185847

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 65.6 | 51.3 | 76.9 |
| 138 | 102.7 | 77.8 | 116.6 |






#49
Acenaphthylene
Concen: 40.736 ng
RT: 9.869 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:152 Resp: 98710

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.2 | 15.8 | 23.8 |
| 153 | 13.1 | 10.4 | 15.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#50
Dimethylphthalate
Concen: 41.279 ng
RT: 9.722 min Scan# 1280
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:163 Resp: 644618

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.3 | 3.4 | 5.2 |
| 164 | 10.3 | 8.2 | 12.2 |






#51
2,6-Dinitrotoluene
Concen: 43.156 ng
RT: 9.780 min Scan# 1
Delta R.T. 0.005 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:165 Resp: 13224

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 63.9 | 57.4 | 86.0 |
| 89 | 59.5 | 51.8 | 77.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#52
Acenaphthene
Concen: 39.982 ng
RT: 10.039 min Scan# 1334
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:154 Resp: 571724

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 109.2 | 86.7 | 130.1 |
| 152 | 51.3 | 41.4 | 62.2 |






#53
3-Nitroaniline
Concen: 47.707 ng
RT: 9.945 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:138 Resp: 16097

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 11.7 | 10.1 | 15.1 |
| 92 | 119.9 | 96.4 | 144.6 |

**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#54
2,4-Dinitrophenol
Concen: 47.113 ng
RT: 10.051 min Scan# 1336
Delta R.T. 0.006 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:184 Resp: 46465

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 106.9 | 149.3 | 223.9# |
| 154 | 573.8 | 1048.0 | 1572.0# |






#55
Dibenzofuran
Concen: 40.300 ng
RT: 10.210 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:168 Resp: 85894

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 37.2 | 29.8 | 44.6 |
| 169 | 13.5 | 10.8 | 16.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#56
4-Nitrophenol
Concen: 45.720 ng
RT: 10.092 min Scan# 1343
Delta R.T. 0.006 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:139 Resp: 122610

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 68.4 | 49.4 | 89.4 |
| 65 | 102.5 | 88.2 | 128.2 |






#57
2,4-Dinitrotoluene
Concen: 43.964 ng
RT: 10.180 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:165 Resp: 16822

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 51.2 | 40.6 | 61.0 |
| 89 | 76.6 | 62.3 | 93.5 |
| 182 | 2.9 | 2.6 | 3.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#58
Fluorene
Concen: 39.732 ng
RT: 10.557 min Scan# 1422
Delta R.T. 0.006 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:166 Resp: 634897

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 92.6 | 73.7 | 110.5 |
| 167 | 13.3 | 10.9 | 16.3 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 44.902 ng
RT: 10.327 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:232 Resp: 16817

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 44.4 | 36.9 | 55.3 |
| 130 | 2.1 | 1.8 | 2.8 |
| 166 | 29.4 | 23.0 | 34.6 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#60
Diethylphthalate
Concen: 42.297 ng
RT: 10.422 min Scan# 1399
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:149 Resp: 637247

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 21.5 | 17.3 | 25.9 |
| 150 | 12.9 | 9.9 | 14.9 |






#61
4-Chlorophenyl-phenylether
Concen: 40.882 ng
RT: 10.545 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:204 Resp: 31311

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 32.5 | 25.8 | 38.6 |
| 141 | 60.4 | 46.8 | 70.2 |

**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025







#62
4-Nitroaniline
Concen: 47.486 ng
RT: 10.557 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:138 Resp: 139342

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 49.0 | 30.4 | 70.4 |
| 108 | 70.8 | 49.5 | 89.5 |






#63
Azobenzene
Concen: 40.833 ng
RT: 10.704 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion: 77 Resp: 68035

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 23.4 | 4.0 | 44.0 |
| 105 | 21.3 | 1.4 | 41.4 |
| 51 | 34.1 | 14.3 | 54.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.492 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:188 Resp: 398969

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 11.0 | 8.6 | 13.0 |
| 80 | 11.9 | 9.3 | 13.9 |






#65
4,6-Dinitro-2-methylphenol
Concen: 44.754 ng
RT: 10.586 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:198 Resp: 6526

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 66.2 | 44.6 | 84.6 |
| 105 | 50.9 | 29.3 | 69.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#66
n-Nitrosodiphenylamine
Concen: 40.179 ng
RT: 10.663 min Scan# 1440
Delta R.T. 0.006 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:169 Resp: 563086

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 66.4 | 53.0 | 79.6 |
| 167 | 35.5 | 28.4 | 42.6 |






#67
4-Bromophenyl-phenylether
Concen: 41.782 ng
RT: 11.033 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:248 Resp: 18957

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 96.0 | 76.7 | 115.1 |
| 141 | 77.4 | 61.7 | 92.5 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#68
Hexachlorobenzene
Concen: 40.915 ng
RT: 11.104 min Scan# 1515
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:284 Resp: 196969

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 35.1 | 28.9 | 43.3 |
| 249 | 30.9 | 24.6 | 37.0 |






#69
Atrazine
Concen: 44.273 ng
RT: 11.180 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:200 Resp: 15874

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 28.1 | 9.1 | 49.1 |
| 215 | 49.8 | 31.0 | 71.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#70
Pentachlorophenol
Concen: 44.978 ng
RT: 11.292 min Scan# 1547
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:266 Resp: 118797

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 64.1 | 51.8 | 77.6 |
| 264 | 63.7 | 49.7 | 74.5 |






#71
Phenanthrene
Concen: 40.469 ng
RT: 11.516 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:178 Resp: 87136

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.1 | 15.1 | 22.7 |
| 179 | 15.3 | 12.2 | 18.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#72
Anthracene
Concen: 40.999 ng
RT: 11.569 min Scan# 1594
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:178 Resp: 887738

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.6 | 14.7 | 22.1 |
| 179 | 15.5 | 12.4 | 18.6 |






#73
Carbazole
Concen: 40.823 ng
RT: 11.716 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:167 Resp: 79188

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.4 | 17.0 | 25.4 |
| 139 | 13.0 | 10.2 | 15.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#74
Di-n-butylphthalate
Concen: 46.099 ng
RT: 12.051 min Scan# 1676
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:149 Resp: 820833

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.0 | 7.3 | 10.9 |
| 104 | 4.8 | 3.8 | 5.8 |






#75
Fluoranthene
Concen: 42.113 ng
RT: 12.704 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:202 Resp: 82449

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 12.7 | 0.0 | 32.4 |
| 203 | 17.5 | 0.0 | 37.4 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.133 min Scan# 2030
Delta R.T. 0.006 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:240 Resp: 215152

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.6 | 10.0 | 15.0 |
| 236 | 24.9 | 20.0 | 30.0 |






#77
Benzidine
Concen: 47.827 ng
RT: 12.816 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:184 Resp: 35057

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 16.3 | 12.4 | 18.6 |
| 183 | 11.1 | 8.6 | 12.8 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#78
Pyrene
Concen: 41.087 ng
RT: 12.933 min Scan# 1826
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:202 Resp: 820201

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 19.7 | 15.5 | 23.3 |
| 203 | 17.6 | 14.0 | 21.0 |






#79
Terphenyl-d14
Concen: 72.317 ng
RT: 13.074 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:244 Resp: 106437

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.5 | 5.2 | 7.8 |
| 122 | 10.9 | 8.8 | 13.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#80
Butylbenzylphthalate
Concen: 43.340 ng
RT: 13.551 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:149 Resp: 172390

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 71.0 | 57.0 | 85.4 |
| 206 | 21.3 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 40.719 ng
RT: 14.121 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:228 Resp: 58379

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 26.6 | 21.3 | 31.9 |
| 229 | 19.6 | 15.5 | 23.3 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#82
3,3'-Dichlorobenzidine
Concen: 44.044 ng
RT: 14.080 min Scan# 2021
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:252 Resp: 183371

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 65.4 | 51.5 | 77.3 |
| 126 | 11.7 | 9.4 | 14.0 |






#83
Chrysene
Concen: 40.139 ng
RT: 14.157 min Scan# 2(
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:228 Resp: 530018

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 30.3 | 23.4 | 35.2 |
| 229 | 19.8 | 15.3 | 22.9 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 40.166 ng
RT: 14.110 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:149 Resp: 239329

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 27.9 | 21.9 | 32.9 |
| 279 | 4.2 | 3.4 | 5.0 |






#85
Di-n-octyl phthalate
Concen: 38.648 ng
RT: 14.727 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:149 Resp: 42495

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.4 | 1.2 | 1.8 |
| 43 | 13.5 | 10.6 | 16.0 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.639 min Scan# 2286
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:264 Resp: 235493

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 22.5 | 18.5 | 27.7 |
| 265 | 21.3 | 17.5 | 26.3 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 42.503 ng
RT: 17.186 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:276 Resp: 74444

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 28.1 | 22.6 | 34.0 |
| 277 | 25.4 | 20.2 | 30.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#88
Benzo(b)fluoranthene
Concen: 43.782 ng
RT: 15.192 min Scan# 2210
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:252 Resp: 594727

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.6 | 17.5 | 26.3 |
| 125 | 10.7 | 8.6 | 13.0 |






#89
Benzo(k)fluoranthene
Concen: 39.662 ng
RT: 15.221 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:252 Resp: 51397

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.5 | 17.4 | 26.2 |
| 125 | 10.7 | 8.5 | 12.7 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#90
Benzo(a)pyrene
Concen: 42.724 ng
RT: 15.574 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:252 Resp: 548875

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.5 | 17.5 | 26.3 |
| 125 | 11.4 | 9.4 | 14.0 |






#91
Dibenzo(a,h)anthracene
Concen: 42.735 ng
RT: 17.203 min Scan# 2
Delta R.T. 0.006 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:278 Resp: 61188

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 18.1 | 14.6 | 22.0 |
| 279 | 24.1 | 18.8 | 28.2 |




**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/16/2025
Supervised By :Jagrut Upadhyay 07/16/2025




#92
Benzo(g,h,i)perylene
Concen: 43.359 ng
RT: 17.651 min Scan# 2628
Delta R.T. 0.006 min
Lab File: BF143101.D
Acq: 15 Jul 2025 17:27

Tgt Ion:276 Resp: 632939

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.7 | 19.2 | 28.8 |
| 138 | 23.9 | 19.8 | 29.6 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143101.D
Acq On : 15 Jul 2025 17:27
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

Quant Time: Jul 15 17:54:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 1 | I | 1,4-Dichlorobenzene-d4 | 1.000 | 1.000 | 0.0 | 106 | 0.00 |
| 2 | | 1,4-Dioxane | 0.630 | 0.602 | 4.4 | 101 | 0.00 |
| 3 | | Pyridine | 1.596 | 1.636 | -2.5 | 108 | 0.00 |
| 4 | | n-Nitrosodimethylamine | 0.827 | 0.857 | -3.6 | 111 | 0.00 |
| 5 | S | 2-Fluorophenol | 1.266 | 1.247 | 1.5 | 105 | 0.00 |
| 6 | | Aniline | 2.241 | 2.299 | -2.6 | 110 | 0.00 |
| 7 | S | Phenol-d6 | 1.591 | 1.563 | 1.8 | 105 | 0.00 |
| 8 | | 2-Chlorophenol | 1.272 | 1.348 | -6.0 | 112 | 0.00 |
| 9 | | Benzaldehyde | 1.212 | 1.125 | 7.2 | 82 | 0.00 |
| 10 | C | Phenol | 1.723 | 1.795 | -4.2 | 111 | 0.00 |
| 11 | | bis(2-Chloroethyl)ether | 1.334 | 1.362 | -2.1 | 110 | 0.00 |
| 12 | | 1,3-Dichlorobenzene | 1.437 | 1.462 | -1.7 | 109 | 0.00 |
| 13 | C | 1,4-Dichlorobenzene | 1.446 | 1.465 | -1.3 | 109 | 0.00 |
| 14 | | 1,2-Dichlorobenzene | 1.379 | 1.405 | -1.9 | 109 | 0.00 |
| 15 | | Benzyl Alcohol | 1.179 | 1.248 | -5.9 | 113 | 0.00 |
| 16 | | 2,2'-oxybis(1-Chloropropane | 2.458 | 2.475 | -0.7 | 109 | 0.00 |
| 17 | | 2-Methylphenol | 1.098 | 1.140 | -3.8 | 111 | 0.00 |
| 18 | | Hexachloroethane | 0.471 | 0.509 | -8.1 | 114 | 0.00 |
| 19 | P | n-Nitroso-di-n-propylamine | 0.975 | 1.023 | -4.9 | 112 | 0.00 |
| 20 | | 3+4-Methylphenols | 1.346 | 1.415 | -5.1 | 110 | 0.00 |
| 21 | I | Naphthalene-d8 | 1.000 | 1.000 | 0.0 | 106 | 0.00 |
| 22 | | Acetophenone | 0.482 | 0.486 | -0.8 | 110 | 0.00 |
| 23 | S | Nitrobenzene-d5 | 0.357 | 0.349 | 2.2 | 100 | 0.00 |
| 24 | | Nitrobenzene | 0.344 | 0.384 | -11.6 | 115 | 0.00 |
| 25 | | Isophorone | 0.689 | 0.710 | -3.0 | 111 | 0.00 |
| 26 | C | 2-Nitrophenol | 0.114 | 0.151 | -32.5# | 134 | 0.00 |
| 27 | | 2,4-Dimethylphenol | 0.320 | 0.336 | -5.0 | 112 | 0.00 |
| 28 | | bis(2-Chloroethoxy)methane | 0.418 | 0.426 | -1.9 | 111 | 0.00 |
| 29 | C | 2,4-Dichlorophenol | 0.262 | 0.285 | -8.8 | 114 | 0.00 |
| 30 | | 1,2,4-Trichlorobenzene | 0.297 | 0.306 | -3.0 | 110 | 0.00 |
| 31 | | Naphthalene | 0.993 | 1.007 | -1.4 | 111 | 0.00 |
| 32 | | Benzoic acid | 0.139 | 0.170 | -22.3 | 126 | 0.00 |
| 33 | | 4-Chloroaniline | 0.399 | 0.412 | -3.3 | 110 | 0.00 |
| 34 | C | Hexachlorobutadiene | 0.178 | 0.183 | -2.8 | 110 | 0.00 |
| 35 | | Caprolactam | 0.082 | 0.089 | -8.5 | 114 | 0.00 |
| 36 | C | 4-Chloro-3-methylphenol | 0.291 | 0.309 | -6.2 | 112 | 0.00 |
| 37 | | 2-Methylnaphthalene | 0.592 | 0.602 | -1.7 | 111 | 0.00 |
| 38 | | 1-Methylnaphthalene | 0.615 | 0.629 | -2.3 | 111 | 0.00 |
| 39 | I | Acenaphthene-d10 | 1.000 | 1.000 | 0.0 | 109 | 0.00 |
| 40 | | 1,2,4,5-Tetrachlorobenzene | 0.583 | 0.591 | -1.4 | 111 | 0.00 |
| 41 | P | Hexachlorocyclopentadiene | 0.323 | 0.364 | -12.7 | 123 | 0.00 |
| 42 | S | 2,4,6-Tribromophenol | 0.178 | 0.192 | -7.9 | 111 | 0.00 |
| 43 | C | 2,4,6-Trichlorophenol | 0.362 | 0.398 | -9.9 | 115 | 0.00 |
| 44 | | 2,4,5-Trichlorophenol | 0.380 | 0.408 | -7.4 | 113 | 0.00 |
| 45 | S | 2-Fluorobiphenyl | 1.505 | 1.299 | 13.7 | 97 | 0.00 |
| 46 | | 1,1'-Biphenyl | 1.595 | 1.610 | -0.9 | 112 | 0.00 |
| 47 | | 2-Chloronaphthalene | 1.172 | 1.178 | -0.5 | 111 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143101.D
Acq On : 15 Jul 2025 17:27
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

Quant Time: Jul 15 17:54:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 48 | | 2-Nitroaniline | 0.303 | 0.374 | -23.4 | 124 | 0.00 |
| 49 | | Acenaphthylene | 1.952 | 1.988 | -1.8 | 111 | 0.00 |
| 50 | | Dimethylphthalate | 1.258 | 1.298 | -3.2 | 113 | 0.00 |
| 51 | | 2,6-Dinitrotoluene | 0.219 | 0.266 | -21.5 | 124 | 0.00 |
| 52 | C | Acenaphthene | 1.152 | 1.152 | 0.0 | 111 | 0.00 |
| 53 | | 3-Nitroaniline | 0.272 | 0.324 | -19.1 | 123 | 0.00 |
| 54 | P | 2,4-Dinitrophenol | 0.073 | 0.094 | -28.8# | 149 | 0.00 |
| 55 | | Dibenzofuran | 1.717 | 1.730 | -0.8 | 111 | 0.00 |
| 56 | P | 4-Nitrophenol | 0.216 | 0.247 | -14.4 | 120 | 0.00 |
| 57 | | 2,4-Dinitrotoluene | 0.265 | 0.339 | -27.9# | 125 | 0.00 |
| 58 | | Fluorene | 1.288 | 1.279 | 0.7 | 112 | 0.00 |
| 59 | | 2,3,4,6-Tetrachlorophenol | 0.302 | 0.339 | -12.3 | 118 | 0.00 |
| 60 | | Diethylphthalate | 1.214 | 1.284 | -5.8 | 114 | 0.00 |
| 61 | | 4-Chlorophenyl-phenylether | 0.617 | 0.631 | -2.3 | 112 | 0.00 |
| 62 | | 4-Nitroaniline | 0.236 | 0.281 | -19.1 | 123 | 0.00 |
| 63 | | Azobenzene | 1.343 | 1.370 | -2.0 | 113 | 0.00 |
| 64 | I | Phenanthrene-d10 | 1.000 | 1.000 | 0.0 | 108 | 0.00 |
| 65 | | 4,6-Dinitro-2-methylphenol | 0.067 | 0.082 | -22.4 | 138 | 0.00 |
| 66 | c | n-Nitrosodiphenylamine | 0.703 | 0.706 | -0.4 | 112 | 0.00 |
| 67 | | 4-Bromophenyl-phenylether | 0.227 | 0.238 | -4.8 | 115 | 0.00 |
| 68 | | Hexachlorobenzene | 0.241 | 0.247 | -2.5 | 113 | 0.00 |
| 69 | | Atrazine | 0.180 | 0.199 | -10.6 | 119 | 0.00 |
| 70 | C | Pentachlorophenol | 0.132 | 0.149 | -12.9 | 122 | 0.00 |
| 71 | | Phenanthrene | 1.079 | 1.092 | -1.2 | 113 | 0.00 |
| 72 | | Anthracene | 1.085 | 1.113 | -2.6 | 114 | 0.00 |
| 73 | | Carbazole | 0.972 | 0.992 | -2.1 | 114 | 0.00 |
| 74 | | Di-n-butylphthalate | 0.893 | 1.029 | -15.2 | 121 | 0.00 |
| 75 | C | Fluoranthene | 0.981 | 1.033 | -5.3 | 116 | 0.00 |
| 76 | I | Chrysene-d12 | 1.000 | 1.000 | 0.0 | 112 | 0.00 |
| 77 | | Benzidine | 0.681 | 0.815 | -19.7 | 109 | 0.00 |
| 78 | | Pyrene | 1.856 | 1.906 | -2.7 | 116 | 0.00 |
| 79 | S | Terphenyl-d14 | 1.368 | 1.237 | 9.6 | 103 | 0.00 |
| 80 | | Butylbenzylphthalate | 0.318 | 0.401 | -26.1# | 131 | 0.00 |
| 81 | | Benzo(a)anthracene | 1.333 | 1.357 | -1.8 | 113 | 0.00 |
| 82 | | 3,3'-Dichlorobenzidine | 0.387 | 0.426 | -10.1 | 124 | 0.00 |
| 83 | | Chrysene | 1.227 | 1.232 | -0.4 | 115 | 0.00 |
| 84 | | Bis(2-ethylhexyl)phthalate | 0.480 | 0.556 | -15.8 | 121 | 0.00 |
| 85 | c | Di-n-octyl phthalate | 0.927 | 0.988 | -6.6 | 120 | 0.00 |
| 86 | I | Perylene-d12 | 1.000 | 1.000 | 0.0 | 104 | 0.00 |
| 87 | | Indeno(1,2,3-cd)pyrene | 1.488 | 1.581 | -6.2 | 108 | 0.00 |
| 88 | | Benzo(b)fluoranthene | 1.154 | 1.263 | -9.4 | 121 | 0.00 |
| 89 | | Benzo(k)fluoranthene | 1.101 | 1.091 | 0.9 | 99 | 0.00 |
| 90 | C | Benzo(a)pyrene | 1.091 | 1.165 | -6.8 | 108 | 0.00 |
| 91 | | Dibenzo(a,h)anthracene | 1.216 | 1.299 | -6.8 | 109 | 0.00 |
| 92 | | Benzo(g,h,i)perylene | 1.240 | 1.344 | -8.4 | 111 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143101.D
Acq On : 15 Jul 2025 17:27
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

Quant Time: Jul 15 17:54:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | AvgRF | CCRF | %Dev Area% | Dev(min) |
|---|---|---|---|---|

(#) = Out of Range SPCC's out = 0 CCC's out = 1

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143101.D
Acq On : 15 Jul 2025 17:27
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

Quant Time: Jul 15 17:54:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 1 | I | 1,4-Dichlorobenzene-d4 | 20.000 | 20.000 | 0.0 | 106 | 0.00 |
| 2 | | 1,4-Dioxane | 40.000 | 38.242 | 4.4 | 101 | 0.00 |
| 3 | | Pyridine | 40.000 | 41.009 | -2.5 | 108 | 0.00 |
| 4 | | n-Nitrosodimethylamine | 40.000 | 41.468 | -3.7 | 111 | 0.00 |
| 5 | S | 2-Fluorophenol | 80.000 | 78.799 | 1.5 | 105 | 0.00 |
| 6 | | Aniline | 40.000 | 41.038 | -2.6 | 110 | 0.00 |
| 7 | S | Phenol-d6 | 80.000 | 78.598 | 1.8 | 105 | 0.00 |
| 8 | | 2-Chlorophenol | 40.000 | 42.397 | -6.0 | 112 | 0.00 |
| 9 | | Benzaldehyde | 40.000 | 37.115 | 7.2 | 82 | 0.00 |
| 10 | C | Phenol | 40.000 | 41.681 | -4.2 | 111 | 0.00 |
| 11 | | bis(2-Chloroethyl)ether | 40.000 | 40.848 | -2.1 | 110 | 0.00 |
| 12 | | 1,3-Dichlorobenzene | 40.000 | 40.694 | -1.7 | 109 | 0.00 |
| 13 | C | 1,4-Dichlorobenzene | 40.000 | 40.539 | -1.3 | 109 | 0.00 |
| 14 | | 1,2-Dichlorobenzene | 40.000 | 40.782 | -2.0 | 109 | 0.00 |
| 15 | | Benzyl Alcohol | 40.000 | 42.336 | -5.8 | 113 | 0.00 |
| 16 | | 2,2'-oxybis(1-Chloropropane | 40.000 | 40.273 | -0.7 | 109 | 0.00 |
| 17 | | 2-Methylphenol | 40.000 | 41.532 | -3.8 | 111 | 0.00 |
| 18 | | Hexachloroethane | 40.000 | 43.239 | -8.1 | 114 | 0.00 |
| 19 | P | n-Nitroso-di-n-propylamine | 40.000 | 41.982 | -5.0 | 112 | 0.00 |
| 20 | | 3+4-Methylphenols | 40.000 | 42.057 | -5.1 | 110 | 0.00 |
| 21 | I | Naphthalene-d8 | 20.000 | 20.000 | 0.0 | 106 | 0.00 |
| 22 | | Acetophenone | 40.000 | 40.351 | -0.9 | 110 | 0.00 |
| 23 | S | Nitrobenzene-d5 | 80.000 | 78.289 | 2.1 | 100 | 0.00 |
| 24 | | Nitrobenzene | 40.000 | 44.647 | -11.6 | 115 | 0.00 |
| 25 | | Isophorone | 40.000 | 41.188 | -3.0 | 111 | 0.00 |
| 26 | C | 2-Nitrophenol | 40.000 | 44.317 | -10.8 | 134 | 0.00 |
| 27 | | 2,4-Dimethylphenol | 40.000 | 42.064 | -5.2 | 112 | 0.00 |
| 28 | | bis(2-Chloroethoxy)methane | 40.000 | 40.789 | -2.0 | 111 | 0.00 |
| 29 | C | 2,4-Dichlorophenol | 40.000 | 43.580 | -8.9 | 114 | 0.00 |
| 30 | | 1,2,4-Trichlorobenzene | 40.000 | 41.194 | -3.0 | 110 | 0.00 |
| 31 | | Naphthalene | 40.000 | 40.543 | -1.4 | 111 | 0.00 |
| 32 | | Benzoic acid | 40.000 | 43.449 | -8.6 | 126 | 0.00 |
| 33 | | 4-Chloroaniline | 40.000 | 41.236 | -3.1 | 110 | 0.00 |
| 34 | C | Hexachlorobutadiene | 40.000 | 41.324 | -3.3 | 110 | 0.00 |
| 35 | | Caprolactam | 40.000 | 43.536 | -8.8 | 114 | 0.00 |
| 36 | C | 4-Chloro-3-methylphenol | 40.000 | 42.387 | -6.0 | 112 | 0.00 |
| 37 | | 2-Methylnaphthalene | 40.000 | 40.654 | -1.6 | 111 | 0.00 |
| 38 | | 1-Methylnaphthalene | 40.000 | 40.922 | -2.3 | 111 | 0.00 |
| 39 | I | Acenaphthene-d10 | 20.000 | 20.000 | 0.0 | 109 | 0.00 |
| 40 | | 1,2,4,5-Tetrachlorobenzene | 40.000 | 40.531 | -1.3 | 111 | 0.00 |
| 41 | P | Hexachlorocyclopentadiene | 40.000 | 45.086 | -12.7 | 123 | 0.00 |
| 42 | S | 2,4,6-Tribromophenol | 80.000 | 85.971 | -7.5 | 111 | 0.00 |
| 43 | C | 2,4,6-Trichlorophenol | 40.000 | 43.955 | -9.9 | 115 | 0.00 |
| 44 | | 2,4,5-Trichlorophenol | 40.000 | 43.001 | -7.5 | 113 | 0.00 |
| 45 | S | 2-Fluorobiphenyl | 80.000 | 69.090 | 13.6 | 97 | 0.00 |
| 46 | | 1,1'-Biphenyl | 40.000 | 40.387 | -1.0 | 112 | 0.00 |
| 47 | | 2-Chloronaphthalene | 40.000 | 40.207 | -0.5 | 111 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143101.D
Acq On : 15 Jul 2025 17:27
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

Quant Time: Jul 15 17:54:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 48 | | 2-Nitroaniline | 40.000 | 43.148 | -7.9 | 124 | 0.00 |
| 49 | | Acenaphthylene | 40.000 | 40.736 | -1.8 | 111 | 0.00 |
| 50 | | Dimethylphthalate | 40.000 | 41.279 | -3.2 | 113 | 0.00 |
| 51 | | 2,6-Dinitrotoluene | 40.000 | 43.156 | -7.9 | 124 | 0.00 |
| 52 | C | Acenaphthene | 40.000 | 39.982 | 0.0 | 111 | 0.00 |
| 53 | | 3-Nitroaniline | 40.000 | 47.707 | -19.3 | 123 | 0.00 |
| 54 | P | 2,4-Dinitrophenol | 40.000 | 47.113 | -17.8 | 149 | 0.00 |
| 55 | | Dibenzofuran | 40.000 | 40.300 | -0.7 | 111 | 0.00 |
| 56 | P | 4-Nitrophenol | 40.000 | 45.720 | -14.3 | 120 | 0.00 |
| 57 | | 2,4-Dinitrotoluene | 40.000 | 43.964 | -9.9 | 125 | 0.00 |
| 58 | | Fluorene | 40.000 | 39.732 | 0.7 | 112 | 0.00 |
| 59 | | 2,3,4,6-Tetrachlorophenol | 40.000 | 44.902 | -12.3 | 118 | 0.00 |
| 60 | | Diethylphthalate | 40.000 | 42.297 | -5.7 | 114 | 0.00 |
| 61 | | 4-Chlorophenyl-phenylether | 40.000 | 40.882 | -2.2 | 112 | 0.00 |
| 62 | | 4-Nitroaniline | 40.000 | 47.486 | -18.7 | 123 | 0.00 |
| 63 | | Azobenzene | 40.000 | 40.833 | -2.1 | 113 | 0.00 |
| 64 | I | Phenanthrene-d10 | 20.000 | 20.000 | 0.0 | 108 | 0.00 |
| 65 | | 4,6-Dinitro-2-methylphenol | 40.000 | 44.754 | -11.9 | 138 | 0.00 |
| 66 | c | n-Nitrosodiphenylamine | 40.000 | 40.179 | -0.4 | 112 | 0.00 |
| 67 | | 4-Bromophenyl-phenylether | 40.000 | 41.782 | -4.5 | 115 | 0.00 |
| 68 | | Hexachlorobenzene | 40.000 | 40.915 | -2.3 | 113 | 0.00 |
| 69 | | Atrazine | 40.000 | 44.273 | -10.7 | 119 | 0.00 |
| 70 | C | Pentachlorophenol | 40.000 | 44.978 | -12.4 | 122 | 0.00 |
| 71 | | Phenanthrene | 40.000 | 40.469 | -1.2 | 113 | 0.00 |
| 72 | | Anthracene | 40.000 | 40.999 | -2.5 | 114 | 0.00 |
| 73 | | Carbazole | 40.000 | 40.823 | -2.1 | 114 | 0.00 |
| 74 | | Di-n-butylphthalate | 40.000 | 46.099 | -15.2 | 121 | 0.00 |
| 75 | C | Fluoranthene | 40.000 | 42.113 | -5.3 | 116 | 0.00 |
| 76 | I | Chrysene-d12 | 20.000 | 20.000 | 0.0 | 112 | 0.00 |
| 77 | | Benzidine | 40.000 | 47.827 | -19.6 | 109 | 0.00 |
| 78 | | Pyrene | 40.000 | 41.087 | -2.7 | 116 | 0.00 |
| 79 | S | Terphenyl-d14 | 80.000 | 72.317 | 9.6 | 103 | 0.00 |
| 80 | | Butylbenzylphthalate | 40.000 | 43.340 | -8.4 | 131 | 0.00 |
| 81 | | Benzo(a)anthracene | 40.000 | 40.719 | -1.8 | 113 | 0.00 |
| 82 | | 3,3'-Dichlorobenzidine | 40.000 | 44.044 | -10.1 | 124 | 0.00 |
| 83 | | Chrysene | 40.000 | 40.139 | -0.3 | 115 | 0.00 |
| 84 | | Bis(2-ethylhexyl)phthalate | 40.000 | 40.166 | -0.4 | 121 | 0.00 |
| 85 | c | Di-n-octyl phthalate | 40.000 | 38.648 | 3.4 | 120 | 0.00 |
| 86 | I | Perylene-d12 | 20.000 | 20.000 | 0.0 | 104 | 0.00 |
| 87 | | Indeno(1,2,3-cd)pyrene | 40.000 | 42.503 | -6.3 | 108 | 0.00 |
| 88 | | Benzo(b)fluoranthene | 40.000 | 43.782 | -9.5 | 121 | 0.00 |
| 89 | | Benzo(k)fluoranthene | 40.000 | 39.662 | 0.8 | 99 | 0.00 |
| 90 | C | Benzo(a)pyrene | 40.000 | 42.724 | -6.8 | 108 | 0.00 |
| 91 | | Dibenzo(a,h)anthracene | 40.000 | 42.735 | -6.8 | 109 | 0.00 |
| 92 | | Benzo(g,h,i)perylene | 40.000 | 43.359 | -8.4 | 111 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143101.D
Acq On : 15 Jul 2025 17:27
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
ICVBF071525

Quant Time: Jul 15 17:54:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Response Factor Report MSBNA_P

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP070325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jul 03 16:05:44 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BP025071.D 5 =BP025072.D 10 =BP025073.D 20 =BP025074.D 40 =BP025075.D 50 =BP025076.D 60 =BP025077.D 80 =BP0250
78.D

| | | Compound | 2.5 | 5 | 10 | 20 | 40 | 50 | 60 | 80 | Avg | %RSD |
|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1) | I | 1,4-Dichlorobenzen... | ---------- | ISTD | ---------- | | | | | | | |
| 2) | | 1,4-Dioxane | | 0.470 | 0.452 | 0.457 | 0.435 | 0.420 | 0.428 | 0.409 | 0.439 | 4.99 |
| 3) | | Pyridine | | 1.176 | 1.153 | 1.243 | 1.228 | 1.171 | 1.239 | 1.203 | 1.202 | 3.00 |
| 4) | | n-Nitrosodimet... | | | 0.500 | 0.531 | 0.506 | 0.493 | 0.522 | 0.500 | 0.509 | 2.85 |
| 5) | S | 2-Fluorophenol | | 1.070 | 1.098 | 1.179 | 1.134 | 1.127 | 1.168 | 1.127 | 1.129 | 3.33 |
| 6) | | Aniline | | 1.334 | 1.324 | 1.430 | 1.388 | 1.371 | 1.428 | 1.359 | 1.376 | 3.05 |
| 7) | S | Phenol-d6 | | 1.428 | 1.407 | 1.520 | 1.458 | 1.472 | 1.543 | 1.482 | 1.473 | 3.26 |
| 8) | | 2-Chlorophenol | | 1.254 | 1.207 | 1.320 | 1.256 | 1.250 | 1.294 | 1.259 | 1.263 | 2.83 |
| 9) | | Benzaldehyde | | | 0.903 | 0.894 | 0.724 | 0.684 | 0.641 | | 0.769 | 15.84 |
| 10) | C | Phenol | | 1.473 | 1.461 | 1.551 | 1.472 | 1.485 | 1.540 | 1.493 | 1.497 | 2.36 |
| 11) | | bis(2-Chloroet... | | 1.187 | 1.139 | 1.200 | 1.122 | 1.129 | 1.172 | 1.115 | 1.152 | 2.95 |
| 12) | | 1,3-Dichlorobe... | | 1.461 | 1.434 | 1.485 | 1.379 | 1.366 | 1.413 | 1.356 | 1.413 | 3.48 |
| 13) | C | 1,4-Dichlorobe... | | 1.500 | 1.432 | 1.505 | 1.399 | 1.384 | 1.424 | 1.374 | 1.431 | 3.70 |
| 14) | | 1,2-Dichlorobe... | | 1.454 | 1.388 | 1.445 | 1.349 | 1.329 | 1.368 | 1.303 | 1.377 | 4.12 |
| 15) | | Benzyl Alcohol | | | 0.887 | 0.991 | 0.957 | 0.990 | 1.058 | 1.015 | 0.983 | 5.86 |
| 16) | | 2,2'-oxybis(1-... | | 1.587 | 1.520 | 1.570 | 1.433 | 1.462 | 1.497 | 1.415 | 1.498 | 4.40 |
| 17) | | 2-Methylphenol | | 0.887 | 0.908 | 1.007 | 0.969 | 0.974 | 1.021 | 0.983 | 0.964 | 5.11 |
| 18) | | Hexachloroethane | | 0.476 | 0.441 | 0.480 | 0.460 | 0.463 | 0.481 | 0.462 | 0.466 | 3.04 |
| 19) | P | n-Nitroso-di-n... | 0.815 | 0.862 | 0.854 | 0.938 | 0.865 | 0.871 | 0.912 | 0.872 | 0.874 | 4.26 |
| 20) | | 3+4-Methylphenols | | | 1.246 | 1.383 | 1.334 | 1.341 | 1.428 | 1.370 | 1.350 | 4.55 |
| 21) | I | Naphthalene-d8 | ---------- | ISTD | ---------- | | | | | | | |
| 22) | | Acetophenone | | 0.480 | 0.481 | 0.502 | 0.465 | 0.456 | 0.480 | 0.447 | 0.473 | 3.88 |
| 23) | S | Nitrobenzene-d5 | | 0.267 | 0.277 | 0.314 | 0.307 | 0.308 | 0.323 | 0.308 | 0.301 | 6.84 |
| 24) | | Nitrobenzene | | 0.280 | 0.293 | 0.319 | 0.310 | 0.307 | 0.327 | 0.308 | 0.306 | 5.13 |
| 25) | | Isophorone | | 0.525 | 0.534 | 0.587 | 0.558 | 0.571 | 0.597 | 0.561 | 0.562 | 4.67 |
| 26) | C | 2-Nitrophenol | | 0.086 | 0.093 | 0.117 | 0.133 | 0.146 | 0.160 | 0.159 | 0.128 | 23.51 |
| 27) | | 2,4-Dimethylph... | | 0.199 | 0.195 | 0.218 | 0.208 | 0.213 | 0.224 | 0.210 | 0.210 | 4.79 |
| 28) | | bis(2-Chloroet... | | 0.378 | 0.371 | 0.401 | 0.374 | 0.379 | 0.397 | 0.371 | 0.382 | 3.25 |
| 29) | C | 2,4-Dichloroph... | | 0.242 | 0.263 | 0.298 | 0.299 | 0.304 | 0.321 | 0.306 | 0.291 | 9.48 |
| 30) | | 1,2,4-Trichlor... | | 0.319 | 0.319 | 0.331 | 0.319 | 0.316 | 0.335 | 0.314 | 0.322 | 2.53 |
| 31) | | Naphthalene | | 1.057 | 1.030 | 1.081 | 1.011 | 1.005 | 1.049 | 0.975 | 1.030 | 3.47 |
| 32) | | Benzoic acid | | | 0.115 | 0.152 | 0.186 | 0.199 | 0.225 | 0.231 | 0.185 | 24.07 |
| 33) | | 4-Chloroaniline | | 0.358 | 0.370 | 0.399 | 0.390 | 0.381 | 0.411 | 0.381 | 0.384 | 4.58 |
| 34) | C | Hexachlorobuta... | | 0.187 | 0.180 | 0.188 | 0.182 | 0.184 | 0.191 | 0.180 | 0.185 | 2.31 |
| 35) | | Caprolactam | | | 0.087 | 0.101 | 0.101 | 0.098 | 0.108 | 0.102 | 0.099 | 7.23 |
| 36) | C | 4-Chloro-3-met... | | 0.259 | 0.279 | 0.321 | 0.309 | 0.309 | 0.332 | 0.316 | 0.304 | 8.35 |
| 37) | | 2-Methylnaphth... | | 0.724 | 0.711 | 0.766 | 0.718 | 0.715 | 0.747 | 0.706 | 0.727 | 2.99 |
| 38) | | 1-Methylnaphth... | | 0.707 | 0.704 | 0.753 | 0.692 | 0.691 | 0.736 | 0.678 | 0.709 | 3.74 |

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP070325.M

| | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|
| 39) | I | Acenaphthene-d10 | ----------ISTD---------- | | | | | | | | |
| 40) | | 1,2,4,5-Tetrac... | 0.554 | 0.549 | 0.584 | 0.580 | 0.578 | 0.610 | 0.568 | 0.575 | 3.56 |
| 41) | P | Hexachlorocycl... | | 0.120 | 0.135 | 0.149 | 0.162 | 0.171 | 0.164 | 0.150 | 13.13 |
| 42) | S | 2,4,6-Tribromo... | 0.194 | 0.214 | 0.240 | 0.246 | 0.244 | 0.266 | 0.249 | 0.236 | 10.22 |
| 43) | C | 2,4,6-Trichlor... | 0.271 | 0.295 | 0.355 | 0.370 | 0.368 | 0.399 | 0.379 | 0.348 | 13.51 |
| 44) | | 2,4,5-Trichlor... | 0.326 | 0.349 | 0.406 | 0.411 | 0.412 | 0.441 | 0.416 | 0.394 | 10.41 |
| 45) | S | 2-Fluorobiphenyl | 1.338 | 1.310 | 1.361 | 1.279 | 1.260 | 1.311 | 1.177 | 1.291 | 4.70 |
| 46) | | 1,1'-Biphenyl | 1.482 | 1.436 | 1.510 | 1.454 | 1.435 | 1.490 | 1.380 | 1.455 | 2.99 |
| 47) | | 2-Chloronaphth... | 1.097 | 1.089 | 1.141 | 1.091 | 1.069 | 1.121 | 1.048 | 1.094 | 2.83 |
| 48) | | 2-Nitroaniline | 0.157 | 0.185 | 0.235 | 0.259 | 0.259 | 0.289 | 0.273 | 0.237 | 20.49 |
| 49) | | Acenaphthylene | 1.582 | 1.622 | 1.771 | 1.701 | 1.660 | 1.761 | 1.620 | 1.674 | 4.36 |
| 50) | | Dimethylphthalate | 1.364 | 1.376 | 1.469 | 1.408 | 1.365 | 1.482 | 1.337 | 1.400 | 3.99 |
| 51) | | 2,6-Dinitrotol... | 0.199 | 0.238 | 0.281 | 0.295 | 0.291 | 0.313 | 0.294 | 0.273 | 14.58 |
| 52) | C | Acenaphthene | 1.085 | 1.074 | 1.142 | 1.066 | 1.047 | 1.111 | 1.027 | 1.079 | 3.59 |
| 53) | | 3-Nitroaniline | 0.204 | 0.248 | 0.297 | 0.321 | 0.310 | 0.345 | 0.323 | 0.293 | 16.81 |
| 54) | P | 2,4-Dinitrophenol | | 0.077 | 0.098 | 0.117 | 0.120 | 0.137 | 0.137 | 0.114 | 20.49 |
| 55) | | Dibenzofuran | 1.829 | 1.814 | 1.866 | 1.768 | 1.720 | 1.813 | 1.655 | 1.781 | 4.07 |
| 56) | P | 4-Nitrophenol | | 0.212 | 0.260 | 0.285 | 0.274 | 0.301 | 0.289 | 0.270 | 11.73 |
| 57) | | 2,4-Dinitrotol... | 0.231 | 0.285 | 0.364 | 0.392 | 0.387 | 0.430 | 0.406 | 0.356 | 20.13 |
| 58) | | Fluorene | 1.382 | 1.384 | 1.467 | 1.393 | 1.335 | 1.403 | 1.278 | 1.377 | 4.27 |
| 59) | | 2,3,4,6-Tetrac... | 0.305 | 0.318 | 0.370 | 0.370 | 0.363 | 0.397 | 0.377 | 0.357 | 9.32 |
| 60) | | Diethylphthalate | 1.368 | 1.376 | 1.468 | 1.429 | 1.366 | 1.477 | 1.358 | 1.406 | 3.64 |
| 61) | | 4-Chlorophenyl... | 0.711 | 0.687 | 0.710 | 0.679 | 0.659 | 0.700 | 0.626 | 0.682 | 4.47 |
| 62) | | 4-Nitroaniline | 0.221 | 0.267 | 0.321 | 0.342 | 0.319 | 0.354 | 0.324 | 0.307 | 15.26 |
| 63) | | Azobenzene | 1.251 | 1.275 | 1.328 | 1.265 | 1.211 | 1.265 | 1.171 | 1.252 | 3.98 |
| 64) | I | Phenanthrene-d10 | ----------ISTD---------- | | | | | | | | |
| 65) | | 4,6-Dinitro-2-... | | 0.054 | 0.069 | 0.081 | 0.089 | 0.097 | 0.098 | 0.082 | 21.25 |
| 66) | c | n-Nitrosodiphe... | 0.557 | 0.548 | 0.597 | 0.573 | 0.574 | 0.592 | 0.565 | 0.572 | 3.11 |
| 67) | | 4-Bromophenyl-... | 0.191 | 0.190 | 0.213 | 0.205 | 0.208 | 0.214 | 0.213 | 0.205 | 5.14 |
| 68) | | Hexachlorobenzene | 0.255 | 0.240 | 0.255 | 0.247 | 0.248 | 0.257 | 0.249 | 0.250 | 2.28 |
| 69) | | Atrazine | 0.155 | 0.167 | 0.183 | 0.172 | 0.171 | 0.172 | 0.154 | 0.168 | 6.10 |
| 70) | C | Pentachlorophenol | | 0.131 | 0.163 | 0.172 | 0.178 | 0.186 | 0.184 | 0.169 | 12.03 |
| 71) | | Phenanthrene | 1.102 | 1.084 | 1.104 | 1.068 | 1.030 | 1.082 | 1.011 | 1.069 | 3.34 |
| 72) | | Anthracene | 0.998 | 0.996 | 1.092 | 1.036 | 1.021 | 1.062 | 0.971 | 1.025 | 4.10 |
| 73) | | Carbazole | 0.954 | 0.994 | 1.057 | 1.029 | 0.990 | 1.017 | 0.976 | 1.002 | 3.45 |
| 74) | | Di-n-butylphth... | 0.929 | 0.994 | 1.147 | 1.187 | 1.177 | 1.228 | 1.152 | 1.116 | 9.89 |
| 75) | C | Fluoranthene | 1.170 | 1.222 | 1.330 | 1.290 | 1.249 | 1.268 | 1.215 | 1.249 | 4.22 |
| 76) | I | Chrysene-d12 | ----------ISTD---------- | | | | | | | | |
| 77) | | Benzidine | | 0.177 | 0.354 | 0.571 | 0.420 | 0.288 | 0.310 | 0.353 | 37.82 |
| 78) | | Pyrene | 1.177 | 1.184 | 1.288 | 1.231 | 1.251 | 1.302 | 1.238 | 1.239 | 3.81 |
| 79) | S | Terphenyl-d14 | 0.943 | 0.919 | 0.978 | 0.914 | 0.915 | 0.936 | 0.764 | 0.910 | 7.50 |
| 80) | | Butylbenzylpht... | 0.228 | 0.276 | 0.377 | 0.440 | 0.471 | 0.519 | 0.499 | 0.402 | 28.01 |
| 81) | | Benzo(a)anthra... | 1.146 | 1.175 | 1.293 | 1.217 | 1.207 | 1.258 | 1.189 | 1.212 | 4.11 |
| 82) | | 3,3'-Dichlorob... | | 0.249 | 0.340 | 0.390 | 0.399 | 0.414 | 0.397 | 0.365 | 16.97 |
| 83) | | Chrysene | 1.169 | 1.147 | 1.223 | 1.159 | 1.157 | 1.191 | 1.118 | 1.166 | 2.88 |
| 84) | | Bis(2-ethylhex... | 0.460 | 0.520 | 0.646 | 0.703 | 0.737 | 0.764 | 0.720 | 0.650 | 17.89 |
| 85) | c | Di-n-octyl pht... | | 0.609 | 0.876 | 1.094 | 1.206 | 1.292 | 1.248 | 1.054 | 25.04 |

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP070325.M

| | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|
| 86) | I | Perylene-d12 | ----------ISTD---------- | | | | | | | |
| 87) | | Indeno(1,2,3-c... | 1.003 | 1.101 | 1.318 | 1.331 | 1.366 | 1.433 | 1.366 | 1.274 | 12.44 |
| 88) | | Benzo(b)fluora... | 1.029 | 1.063 | 1.215 | 1.201 | 1.188 | 1.266 | 1.187 | 1.164 | 7.36 |
| 89) | | Benzo(k)fluora... | 1.070 | 1.129 | 1.249 | 1.153 | 1.157 | 1.218 | 1.133 | 1.158 | 5.11 |
| 90) | C | Benzo(a)pyrene | 0.817 | 0.868 | 1.035 | 1.033 | 1.035 | 1.108 | 1.040 | 0.991 | 10.70 |
| 91) | | Dibenzo(a,h)an... | 0.840 | 0.921 | 1.098 | 1.097 | 1.126 | 1.172 | 1.118 | 1.053 | 11.66 |
| 92) | | Benzo(g,h,i)pe... | 0.915 | 0.967 | 1.126 | 1.127 | 1.142 | 1.197 | 1.142 | 1.088 | 9.58 |

(#) = Out of Range




Response = 1.596e-001 * Amt - 2.635e-002
Coef of Det (r^2) = 0.992912  Curve Fit: wlr(1/a)
Method: \\chfs1srv\HPCHEM1\BNA P\Methods\8270E-BP070325.M
Calibration Table Last Updated: Thu Jul 03 16:05:44 2025




Response = 2.843e-001 * Amt - 3.930e-002
Coef of Det (r^2) = 0.996989  Curve Fit: wlr(1/a)
Method: Q:\250NSVOC_Group1\srv\HPCHEM1\BNA P\Methods\8270E-BP070325.M
Calibration Table Last Updated: Thu Jul 03 16:05:44 2025




Response = 4.240e-001 * Amt - 5.591e-002
Coef of Det (r^2) = 0.997730  Curve Fit: wlr(1/a)
Method: \\Chromsrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M
Calibration Table Last Updated: Thu Jul 03 16:05:44 2025




Response = 1.431e-001 * Amt - 3.852e-002
Coef of Det (r^2) = 0.995026  Curve Fit: wlr(1/a)
Method: \\nsvcms\srv\HPCHEM1\BNA P\Methods\8270E-BP070325.M
Calibration Table Last Updated: Thu Jul 03 16:05:44 2025




Response = 1.030e-001 * Amt - 2.873e-002
Coef of Det (r^2) = 0.995315  Curve Fit: wlr(1/a)
Method: \\nsvocms\chemsrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M
Calibration Table Last Updated: Thu Jul 03 16:05:44 2025




Response = 5.169e-001 * Amt - 9.516e-002
Coef of Det (r^2) = 0.994647  Curve Fit: wlr(1/a)
Method: \\nt-svecms\chemsrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M
Calibration Table Last Updated: Thu Jul 03 16:05:44 2025




Response = 1.366e+000 * Amt - 4.164e-001
Coef of Det (r^2) = 0.996792  Curve Fit: wlr(1/a)
Method: \\nsvcsrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M
Calibration Table Last Updated: Thu Jul 03 16:05:44 2025

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025071.D
Acq On : 03 Jul 2025 09:39
Operator : CG/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC2.5

Quant Time: Jul 03 15:48:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.437 | 152 | 360375 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.184 | 136 | 1444138 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.090 | 164 | 947160 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 16.913 | 188 | 1996879 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.348 | 240 | 2099311 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 24.465 | 264 | 2125718 | 20.000 | ng | -0.03 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 0.000 | 112 | 0d | 0.000 | ng | |
| 7) Phenol-d6 | 0.000 | 99 | 0d | 0.000 | ng | |
| 23) Nitrobenzene-d5 | 0.000 | 82 | 0d | 0.000 | ng | |
| 42) 2,4,6-Tribromophenol | 0.000 | 330 | 0d | 0.000 | ng | |
| 45) 2-Fluorobiphenyl | 0.000 | 172 | 0d | 0.000 | ng | |
| 79) Terphenyl-d14 | 0.000 | 244 | 0d | 0.000 | ng | |
| Target Compounds | | | | | | Qvalue |
| 19) n-Nitroso-di-n-propyla... | 8.231 | 70 | 36693 | 2.331 | ng | 91 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025071.D
Acq On : 03 Jul 2025 09:39
Operator : CG/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC2.5

Quant Time: Jul 03 15:48:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.437 min Scan# 7
Delta R.T. -0.006 min
Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC2.5

Tgt Ion:152 Resp: 360375

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 155.8 | 126.1 | 189.1 |
| 115 | 56.3 | 44.3 | 66.5 |







#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.10 min

Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

Tgt Ion: 112

| Sig | Exp Ratio |
|---|---|
| 112 | 100 |
| 64 | 52.5 |
| 63 | 27.7 |






#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.63 min

Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC2.5

Tgt Ion: 99

| Sig | Exp Ratio |
|---|---|
| 99 | 100 |
| 42 | 17.0 |
| 71 | 30.5 |







#19
n-Nitroso-di-n-propylamine
Concen: 2.331 ng
RT: 8.231 min Scan# 900
Delta R.T. -0.006 min
Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

Tgt Ion: 70 Resp: 36693

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 67.0 | 46.5 | 69.7 |
| 101 | 10.3 | 7.9 | 11.9 |
| 130 | 22.5 | 19.3 | 28.9 |












#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.184 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC2.5

Tgt Ion:136 Resp: 1444138

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.7 | 8.7 | 13.1 |
| 54 | 7.0 | 5.4 | 8.2 |
| 68 | 5.3 | 4.5 | 6.7 |







#23
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 8.57 min

Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

Tgt Ion: 82

| Sig | Exp Ratio |
|---|---|
| 82 | 100 |
| 128 | 48.0 |
| 54 | 47.2 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.090 min Scan# 1
Delta R.T. 0.001 min
Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC2.5

Tgt Ion:164 Resp: 947160

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 101.4 | 80.3 | 120.5 |
| 160 | 46.7 | 35.9 | 53.9 |







#42
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 15.62 min

Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

Tgt Ion: 330

| Sig | Exp Ratio |
|---|---|
| 330 | 100 |
| 332 | 97.1 |
| 141 | 32.5 |






#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 12.70 min

Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC2.5

Tgt Ion: 172

| Sig | Exp Ratio |
|---|---|
| 172 | 100 |
| 171 | 35.2 |
| 170 | 23.4 |







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.913 min Scan# 2376
Delta R.T. 0.006 min
Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

Tgt Ion:188 Resp: 1996879

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 8.6 | 7.3 | 10.9 |
| 80 | 9.4 | 7.8 | 11.6 |






#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.348 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC2.5

Tgt Ion:240 Resp: 2099311

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.8 | 9.0 | 13.4 |
| 236 | 25.9 | 20.3 | 30.5 |







#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 19.67 min

Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

Tgt Ion: 244

| Sig | Exp Ratio |
|---|---|
| 244 | 100 |
| 212 | 7.4 |
| 122 | 11.8 |






#86
Perylene-d12
Concen: 20.000 ng
RT: 24.465 min Scan# 3
Delta R.T. -0.029 min
Lab File: BP025071.D
Acq: 03 Jul 2025 09:39

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC2.5

Tgt Ion:264 Resp: 2125718

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.8 | 18.6 | 28.0 |
| 265 | 21.7 | 17.7 | 26.5 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025072.D
Acq On : 03 Jul 2025 10:21
Operator : CG/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:49:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| Internal Standards | | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.443 | 152 | 354349 | 20.000 | ng | | 0.00 |
| 21) Naphthalene-d8 | 10.190 | 136 | 1455964 | 20.000 | ng | | 0.00 |
| 39) Acenaphthene-d10 | 14.095 | 164 | 982759 | 20.000 | ng | | 0.00 |
| 64) Phenanthrene-d10 | 16.907 | 188 | 1998531 | 20.000 | ng | | 0.00 |
| 76) Chrysene-d12 | 21.348 | 240 | 2143770 | 20.000 | ng | | 0.00 |
| 86) Perylene-d12 | 24.495 | 264 | 2178150 | 20.000 | ng | | 0.00 |
| System Monitoring Compounds | | | | | | | |
| 5) 2-Fluorophenol | 5.102 | 112 | 189521 | 9.477 | ng | | 0.00 |
| 7) Phenol-d6 | 6.631 | 99 | 252937 | 9.692 | ng | | 0.00 |
| 23) Nitrobenzene-d5 | 8.566 | 82 | 194218 | 8.871 | ng | | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.625 | 330 | 95353 | 8.213 | ng | | 0.00 |
| 45) 2-Fluorobiphenyl | 12.695 | 172 | 657518 | 10.365 | ng | | 0.00 |
| 79) Terphenyl-d14 | 19.683 | 244 | 1010825 | 10.363 | ng | | 0.01 |
| Target Compounds | | | | | | | Qvalue |
| 2) 1,4-Dioxane | 3.125 | 88 | 41667 | 5.358 | ng | | 98 |
| 3) Pyridine | 3.496 | 79 | 104170 | 4.896 | ng | | 97 |
| 6) Aniline | 6.784 | 93 | 118174 | 4.846 | ng | | 99 |
| 8) 2-Chlorophenol | 7.019 | 128 | 111121 | 4.966 | ng | | 99 |
| 10) Phenol | 6.655 | 94 | 130527 | 4.922 | ng | | 99 |
| 11) bis(2-Chloroethyl)ether | 6.884 | 93 | 105128 | 5.151 | ng | | 98 |
| 12) 1,3-Dichlorobenzene | 7.337 | 146 | 129394 | 5.167 | ng | | 98 |
| 13) 1,4-Dichlorobenzene | 7.478 | 146 | 132919 | 5.242 | ng | | 99 |
| 14) 1,2-Dichlorobenzene | 7.784 | 146 | 128817 | 5.282 | ng | | 100 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.978 | 45 | 140624 | 5.299 | ng | | 97 |
| 17) 2-Methylphenol | 7.878 | 107 | 78577 | 4.600 | ng | | 96 |
| 18) Hexachloroethane | 8.501 | 117 | 42157 | 5.104 | ng | | 96 |
| 19) n-Nitroso-di-n-propyla... | 8.237 | 70 | 76382 | 4.935 | ng | | 96 |
| 22) Acetophenone | 8.243 | 105 | 174562 | 5.070 | ng | # | 98 |
| 24) Nitrobenzene | 8.607 | 77 | 102014 | 4.575 | ng | | 97 |
| 25) Isophorone | 9.137 | 82 | 190993 | 4.668 | ng | | 98 |
| 26) 2-Nitrophenol | 9.319 | 139 | 31408 | 6.004 | ng | | 98 |
| 27) 2,4-Dimethylphenol | 9.384 | 122 | 72542 | 4.753 | ng | | 99 |
| 28) bis(2-Chloroethoxy)met... | 9.619 | 93 | 137622 | 4.954 | ng | | 97 |
| 29) 2,4-Dichlorophenol | 9.837 | 162 | 88262 | 4.171 | ng | | 98 |
| 30) 1,2,4-Trichlorobenzene | 10.060 | 180 | 116080 | 4.953 | ng | | 94 |
| 31) Naphthalene | 10.237 | 128 | 384889 | 5.134 | ng | | 100 |
| 33) 4-Chloroaniline | 10.348 | 127 | 130394 | 4.661 | ng | | 99 |
| 34) Hexachlorobutadiene | 10.543 | 225 | 67982 | 5.055 | ng | | 97 |
| 36) 4-Chloro-3-methylphenol | 11.484 | 107 | 94437 | 4.270 | ng | | 98 |
| 37) 2-Methylnaphthalene | 11.866 | 142 | 263643 | 4.984 | ng | | 100 |
| 38) 1-Methylnaphthalene | 12.090 | 142 | 257270 | 4.987 | ng | | 98 |
| 40) 1,2,4,5-Tetrachloroben... | 12.242 | 216 | 136193 | 4.822 | ng | | 99 |
| 43) 2,4,6-Trichlorophenol | 12.495 | 196 | 66524 | 3.890 | ng | | 96 |
| 44) 2,4,5-Trichlorophenol | 12.560 | 196 | 80137 | 4.135 | ng | | 96 |
| 46) 1,1'-Biphenyl | 12.907 | 154 | 364077 | 5.092 | ng | | 99 |
| 47) 2-Chloronaphthalene | 12.942 | 162 | 269579 | 5.016 | ng | | 97 |
| 48) 2-Nitroaniline | 13.142 | 65 | 38515 | 5.521 | ng | | 94 |
| 49) Acenaphthylene | 13.807 | 152 | 388781 | 4.726 | ng | | 100 |
| 50) Dimethylphthalate | 13.554 | 163 | 335065 | 4.871 | ng | | 99 |
| 51) 2,6-Dinitrotoluene | 13.654 | 165 | 48991 | 3.651 | ng | | 88 |

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025072.D
Acq On : 03 Jul 2025 10:21
Operator : CG/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:49:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 52) Acenaphthene | 14.166 | 154 | 266690m | 5.030 | ng | | |
| 53) 3-Nitroaniline | 13.984 | 138 | 50202 | 3.491 | ng | # | 87 |
| 55) Dibenzofuran | 14.507 | 168 | 449302 | 5.135 | ng | | 100 |
| 57) 2,4-Dinitrotoluene | 14.466 | 165 | 56730 | 5.360 | ng | # | 97 |
| 58) Fluorene | 15.166 | 166 | 339444 | 5.016 | ng | | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 14.731 | 232 | 74947 | 4.270 | ng | | 97 |
| 60) Diethylphthalate | 14.960 | 149 | 336119 | 4.865 | ng | | 98 |
| 61) 4-Chlorophenyl-phenyle... | 15.178 | 204 | 174567 | 5.212 | ng | | 97 |
| 62) 4-Nitroaniline | 15.172 | 138 | 54190 | 3.594 | ng | | 90 |
| 63) Azobenzene | 15.472 | 77 | 307304 | 4.995 | ng | | 96 |
| 66) n-Nitrosodiphenylamine | 15.389 | 169 | 278113 | 4.863 | ng | | 98 |
| 67) 4-Bromophenyl-phenylether | 16.095 | 248 | 95310 | 4.658 | ng | | 93 |
| 68) Hexachlorobenzene | 16.195 | 284 | 127389 | 5.093 | ng | | 94 |
| 69) Atrazine | 16.389 | 200 | 77424 | 4.620 | ng | | 99 |
| 71) Phenanthrene | 16.948 | 178 | 550598 | 5.156 | ng | | 99 |
| 72) Anthracene | 17.054 | 178 | 498692 | 4.868 | ng | | 99 |
| 73) Carbazole | 17.330 | 167 | 476619 | 4.758 | ng | | 99 |
| 74) Di-n-butylphthalate | 17.966 | 149 | 464272 | 4.162 | ng | | 99 |
| 75) Fluoranthene | 19.054 | 202 | 584660 | 4.684 | ng | | 99 |
| 78) Pyrene | 19.430 | 202 | 630754 | 4.750 | ng | | 98 |
| 80) Butylbenzylphthalate | 20.442 | 149 | 122350 | 5.890 | ng | | 93 |
| 81) Benzo(a)anthracene | 21.330 | 228 | 614350 | 4.729 | ng | | 99 |
| 83) Chrysene | 21.395 | 228 | 626717 | 5.013 | ng | | 98 |
| 84) Bis(2-ethylhexyl)phtha... | 21.342 | 149 | 246749 | 3.541 | ng | | 99 |
| 87) Indeno(1,2,3-cd)pyrene | 28.024 | 276 | 546143m | 3.935 | ng | | |
| 88) Benzo(b)fluoranthene | 23.507 | 252 | 560057 | 4.418 | ng | | 98 |
| 89) Benzo(k)fluoranthene | 23.571 | 252 | 582662 | 4.619 | ng | | 99 |
| 90) Benzo(a)pyrene | 24.354 | 252 | 444644 | 4.121 | ng | | 99 |
| 91) Dibenzo(a,h)anthracene | 28.112 | 278 | 457436 | 3.988 | ng | | 99 |
| 92) Benzo(g,h,i)perylene | 29.112 | 276 | 498456 | 4.207 | ng | | 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025072.D
Acq On : 03 Jul 2025 10:21
Operator : CG/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jul 03 15:49:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.443 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:152 Resp: 35434

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 155.3 | 126.1 | 189.1 |
| 115 | 58.2 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#2
1,4-Dioxane
Concen: 5.358 ng
RT: 3.125 min Scan# 32
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 88 Resp: 41667

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 69.7 | 55.0 | 82.4 |
| 43 | 28.9 | 21.9 | 32.9 |






#3
Pyridine
Concen: 4.896 ng
RT: 3.496 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 79 Resp: 10417

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 64.7 | 52.2 | 78.4 |
| 51 | 36.8 | 26.8 | 40.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#5
2-Fluorophenol
Concen: 9.477 ng
RT: 5.102 min Scan# 368
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:112 Resp: 189521

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 54.2 | 42.0 | 63.0 |
| 63 | 29.9 | 22.2 | 33.2 |






#6
Aniline
Concen: 4.846 ng
RT: 6.784 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 93 Resp: 11817

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 37.0 | 28.9 | 43.3 |
| 65 | 18.4 | 15.2 | 22.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#7
Phenol-d6
Concen: 9.692 ng
RT: 6.631 min Scan# 628
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 99 Resp: 252937

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 18.3 | 13.6 | 20.4 |
| 71 | 30.1 | 24.4 | 36.6 |






#8
2-Chlorophenol
Concen: 4.966 ng
RT: 7.019 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:128 Resp: 11112

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.0 | 12.9 | 52.9 |
| 64 | 39.5 | 20.3 | 60.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#10
Phenol
Concen: 4.922 ng
RT: 6.655 min Scan# 632
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 94 Resp: 130527

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 27.2 | 6.9 | 46.9 |
| 66 | 34.2 | 14.6 | 54.6 |






#11
bis(2-Chloroethyl)ether
Concen: 5.151 ng
RT: 6.884 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 93 Resp: 10512

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 74.3 | 51.9 | 91.9 |
| 95 | 32.3 | 12.4 | 52.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#12
1,3-Dichlorobenzene
Concen: 5.167 ng
RT: 7.337 min Scan# 748
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:146 Resp: 129394

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 66.0 | 51.9 | 77.9 |
| 75 | 28.2 | 21.8 | 32.6 |






#13
1,4-Dichlorobenzene
Concen: 5.242 ng
RT: 7.478 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:146 Resp: 13291

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 65.7 | 51.1 | 76.7 |
| 111 | 38.9 | 31.2 | 46.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#14
1,2-Dichlorobenzene
Concen: 5.282 ng
RT: 7.784 min Scan# 824
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:146 Resp: 128817

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.4 | 51.1 | 76.7 |
| 111 | 39.0 | 31.2 | 46.8 |






#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.299 ng
RT: 7.978 min Scan# 8
Delta R.T. 0.012 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 45 Resp: 14062

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 15.4 | 0.0 | 34.8 |
| 79 | 12.2 | 0.0 | 30.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#17
2-Methylphenol
Concen: 4.600 ng
RT: 7.878 min Scan# 840
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:107 Resp: 78577

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 117.3 | 89.7 | 134.5 |
| 77 | 46.0 | 34.2 | 51.4 |
| 79 | 43.1 | 33.4 | 50.2 |






#18
Hexachloroethane
Concen: 5.104 ng
RT: 8.501 min Scan# 94
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:117 Resp: 4215

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 96.9 | 76.6 | 115.0 |
| 201 | 100.4 | 85.4 | 128.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#19
n-Nitroso-di-n-propylamine
Concen: 4.935 ng
RT: 8.237 min Scan# 901
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 70 Resp: 76382

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 62.2 | 46.5 | 69.7 |
| 101 | 10.2 | 7.9 | 11.9 |
| 130 | 22.9 | 19.3 | 28.9 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.190 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:136 Resp: 145596

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 11.0 | 8.7 | 13.1 |
| 54 | 7.2 | 5.4 | 8.2 |
| 68 | 5.7 | 4.5 | 6.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#22
Acetophenone
Concen: 5.070 ng
RT: 8.243 min Scan# 902
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:105 Resp: 174562

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 5.3 | 5.4 | 8.2# |
| 51 | 26.3 | 19.6 | 29.4 |
| 120 | 23.6 | 18.8 | 28.2 |






#23
Nitrobenzene-d5
Concen: 8.871 ng
RT: 8.566 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 82 Resp: 19421

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 43.9 | 38.4 | 57.6 |
| 54 | 46.9 | 37.8 | 56.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#24
Nitrobenzene
Concen: 4.575 ng
RT: 8.607 min Scan# 964
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 77 Resp: 102014

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 46.2 | 39.0 | 58.6 |
| 65 | 13.1 | 10.7 | 16.1 |






#25
Isophorone
Concen: 4.668 ng
RT: 9.137 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 82 Resp: 19099

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.0 | 5.8 | 8.6 |
| 138 | 18.6 | 15.7 | 23.5 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#26
2-Nitrophenol
Concen: 6.004 ng
RT: 9.319 min Scan# 1085
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:139 Resp: 31408

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 28.5 | 22.3 | 33.5 |
| 65 | 44.4 | 33.8 | 50.8 |






#27
2,4-Dimethylphenol
Concen: 4.753 ng
RT: 9.384 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:122 Resp: 7254

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 109.2 | 88.2 | 132.2 |
| 121 | 56.8 | 46.1 | 69.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#28
bis(2-Chloroethoxy)methane
Concen: 4.954 ng
RT: 9.619 min Scan# 1136
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 93 Resp: 137622

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 31.0 | 26.4 | 39.6 |
| 123 | 11.7 | 10.1 | 15.1 |






#29
2,4-Dichlorophenol
Concen: 4.171 ng
RT: 9.837 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:162 Resp: 8826

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 62.8 | 44.1 | 84.1 |
| 98 | 32.1 | 13.1 | 53.1 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#30
1,2,4-Trichlorobenzene
Concen: 4.953 ng
RT: 10.060 min Scan# 1211
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:180 Resp: 116080

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 98.8 | 75.1 | 112.7 |
| 145 | 33.3 | 23.3 | 34.9 |






#31
Naphthalene
Concen: 5.134 ng
RT: 10.237 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:128 Resp: 38488

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.0 | 8.9 | 13.3 |
| 127 | 12.9 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#33
4-Chloroaniline
Concen: 4.661 ng
RT: 10.348 min Scan# 1260
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:127 Resp: 130394

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 31.3 | 25.4 | 38.2 |
| 65 | 27.0 | 21.0 | 31.6 |
| 92 | 17.3 | 13.4 | 20.0 |






#34
Hexachlorobutadiene
Concen: 5.055 ng
RT: 10.543 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:225 Resp: 6798

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 59.6 | 50.3 | 75.5 |
| 227 | 62.5 | 51.3 | 76.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#36
4-Chloro-3-methylphenol
Concen: 4.270 ng
RT: 11.484 min Scan# 1453
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:107 Resp: 94437

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 29.0 | 22.4 | 33.6 |
| 142 | 89.0 | 69.6 | 104.4 |






#37
2-Methylnaphthalene
Concen: 4.984 ng
RT: 11.866 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:142 Resp: 26364

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 89.4 | 71.2 | 106.8 |
| 115 | 32.3 | 25.9 | 38.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#38
1-Methylnaphthalene
Concen: 4.987 ng
RT: 12.090 min Scan# 1556
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:142 Resp: 257270

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 91.1 | 74.1 | 111.1 |
| 116 | 3.7 | 2.7 | 4.1 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.095 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:164 Resp: 98275

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 99.9 | 80.3 | 120.5 |
| 160 | 46.1 | 35.9 | 53.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 4.822 ng
RT: 12.242 min Scan# 1582
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:216 Resp: 136193

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 77.4 | 62.4 | 93.6 |
| 179 | 21.4 | 16.5 | 24.7 |
| 108 | 18.8 | 13.7 | 20.5 |






#42
2,4,6-Tribromophenol
Concen: 8.213 ng
RT: 15.625 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:330 Resp: 9535

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 97.4 | 77.7 | 116.5 |
| 141 | 34.7 | 26.0 | 39.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#43
2,4,6-Trichlorophenol
Concen: 3.890 ng
RT: 12.495 min Scan# 1625
Delta R.T. 0.012 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:196 Resp: 66524

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 98.8 | 76.2 | 114.2 |
| 200 | 31.9 | 24.3 | 36.5 |






#44
2,4,5-Trichlorophenol
Concen: 4.135 ng
RT: 12.560 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:196 Resp: 8013

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 98.8 | 76.6 | 115.0 |
| 97 | 47.2 | 36.6 | 55.0 |
| 132 | 30.7 | 21.4 | 32.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#45
2-Fluorobiphenyl
Concen: 10.365 ng
RT: 12.695 min Scan# 1659
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:172 Resp: 657518

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.3 | 28.2 | 42.2 |
| 170 | 23.0 | 18.7 | 28.1 |






#46
1,1'-Biphenyl
Concen: 5.092 ng
RT: 12.907 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:154 Resp: 36407

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 39.6 | 20.2 | 60.2 |
| 76 | 11.5 | 0.0 | 31.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#47
2-Chloronaphthalene
Concen: 5.016 ng
RT: 12.942 min Scan# 1701
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:162 Resp: 269579

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 39.5 | 29.4 | 44.2 |
| 164 | 32.4 | 26.4 | 39.6 |






#48
2-Nitroaniline
Concen: 5.521 ng
RT: 13.142 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 65 Resp: 3851

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 71.1 | 58.7 | 88.1 |
| 138 | 115.3 | 99.0 | 148.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#49
Acenaphthylene
Concen: 4.726 ng
RT: 13.807 min Scan# 1848
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:152 Resp: 388781

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.2 | 16.1 | 24.1 |
| 153 | 13.7 | 10.7 | 16.1 |






#50
Dimethylphthalate
Concen: 4.871 ng
RT: 13.554 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:163 Resp: 33506

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 3.8 | 3.4 | 5.0 |
| 164 | 9.7 | 8.0 | 12.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#51
2,6-Dinitrotoluene
Concen: 3.651 ng
RT: 13.654 min Scan# 1822
Delta R.T. -0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:165 Resp: 48991

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 54.7 | 38.5 | 57.7 |
| 89 | 60.1 | 40.6 | 61.0 |






#52
Acenaphthene
Concen: 5.030 ng m
RT: 14.166 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:154 Resp: 26669

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 115.8 | 90.6 | 136.0 |
| 152 | 57.5 | 43.8 | 65.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#53
3-Nitroaniline
Concen: 3.491 ng
RT: 13.984 min Scan# 1878
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:138 Resp: 50202

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 15.4 | 7.8 | 11.8# |
| 92 | 114.2 | 81.0 | 121.4 |






#55
Dibenzofuran
Concen: 5.135 ng
RT: 14.507 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:168 Resp: 44930

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 35.6 | 28.6 | 43.0 |
| 169 | 13.4 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#57
2,4-Dinitrotoluene
Concen: 5.360 ng
RT: 14.466 min Scan# 1960
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:165 Resp: 56730

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 40.8 | 31.0 | 46.4 |
| 89 | 68.6 | 53.4 | 80.0 |
| 182 | 3.3 | 2.0 | 3.0# |






#58
Fluorene
Concen: 5.016 ng
RT: 15.166 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:166 Resp: 33944

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 97.3 | 78.6 | 117.8 |
| 167 | 13.7 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#59
2,3,4,6-Tetrachlorophenol
Concen: 4.270 ng
RT: 14.731 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:232 Resp: 74947

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 47.0 | 35.7 | 53.5 |
| 130 | 2.5 | 1.8 | 2.6 |
| 166 | 29.9 | 24.4 | 36.6 |






#60
Diethylphthalate
Concen: 4.865 ng
RT: 14.960 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:149 Resp: 33611

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 21.0 | 17.8 | 26.8 |
| 150 | 12.4 | 10.2 | 15.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#61
4-Chlorophenyl-phenylether
Concen: 5.212 ng
RT: 15.178 min Scan# 2081
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:204 Resp: 174567

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 33.4 | 25.8 | 38.6 |
| 141 | 59.2 | 45.1 | 67.7 |






#62
4-Nitroaniline
Concen: 3.594 ng
RT: 15.172 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:138 Resp: 5419

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 50.4 | 26.0 | 66.0 |
| 108 | 80.2 | 50.0 | 90.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#63
Azobenzene
Concen: 4.995 ng
RT: 15.472 min Scan# 2131
Delta R.T. -0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion: 77 Resp: 307304

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 24.3 | 7.7 | 47.7 |
| 105 | 20.6 | 1.9 | 41.9 |
| 51 | 28.9 | 8.2 | 48.2 |






#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.907 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:188 Resp: 199853

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 9.0 | 7.3 | 10.9 |
| 80 | 9.5 | 7.8 | 11.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#66
n-Nitrosodiphenylamine
Concen: 4.863 ng
RT: 15.389 min Scan# 2117
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:169 Resp: 278113

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 66.1 | 54.1 | 81.1 |
| 167 | 35.7 | 29.0 | 43.6 |






#67
4-Bromophenyl-phenylether
Concen: 4.658 ng
RT: 16.095 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:248 Resp: 9531

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 103.6 | 79.4 | 119.2 |
| 141 | 72.3 | 51.4 | 77.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#68
Hexachlorobenzene
Concen: 5.093 ng
RT: 16.195 min Scan# 2254
Delta R.T. -0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:284 Resp: 127389

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 36.8 | 25.9 | 38.9 |
| 249 | 31.9 | 23.8 | 35.8 |






#69
Atrazine
Concen: 4.620 ng
RT: 16.389 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:200 Resp: 7742

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 25.3 | 6.6 | 46.6 |
| 215 | 48.2 | 28.0 | 68.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#71
Phenanthrene
Concen: 5.156 ng
RT: 16.948 min Scan# 2382
Delta R.T. -0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:178 Resp: 550598

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.5 | 15.4 | 23.2 |
| 179 | 15.1 | 12.6 | 19.0 |






#72
Anthracene
Concen: 4.868 ng
RT: 17.054 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:178 Resp: 49869

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.3 | 15.0 | 22.6 |
| 179 | 14.9 | 12.5 | 18.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#73
Carbazole
Concen: 4.758 ng
RT: 17.330 min Scan# 2447
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:167 Resp: 476619

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.5 | 17.4 | 26.0 |
| 139 | 12.2 | 10.4 | 15.6 |






#74
Di-n-butylphthalate
Concen: 4.162 ng
RT: 17.966 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:149 Resp: 46427

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 8.6 | 7.3 | 10.9 |
| 104 | 5.2 | 3.8 | 5.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#75
Fluoranthene
Concen: 4.684 ng
RT: 19.054 min Scan# 2740
Delta R.T. -0.012 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:202 Resp: 584660

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 11.3 | 0.0 | 30.4 |
| 203 | 17.6 | 0.0 | 37.6 |






#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.348 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:240 Resp: 214377

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.2 | 9.0 | 13.4 |
| 236 | 25.6 | 20.3 | 30.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#78
Pyrene
Concen: 4.750 ng
RT: 19.430 min Scan# 2804
Delta R.T. -0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:202 Resp: 630754

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 19.5 | 17.1 | 25.7 |
| 203 | 17.5 | 14.1 | 21.1 |






#79
Terphenyl-d14
Concen: 10.363 ng
RT: 19.683 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:244 Resp: 101082

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.4 | 5.9 | 8.9 |
| 122 | 10.6 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#80
Butylbenzylphthalate
Concen: 5.890 ng
RT: 20.442 min Scan# 2976
Delta R.T. 0.012 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:149 Resp: 122350

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 69.9 | 50.6 | 76.0 |
| 206 | 19.4 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 4.729 ng
RT: 21.330 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:228 Resp: 61435

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 26.8 | 22.1 | 33.1 |
| 229 | 19.7 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#83
Chrysene
Concen: 5.013 ng
RT: 21.395 min Scan# 3138
Delta R.T. 0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:228 Resp: 626717

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 29.2 | 24.4 | 36.6 |
| 229 | 19.5 | 16.0 | 24.0 |






#84
Bis(2-ethylhexyl)phthalate
Concen: 3.541 ng
RT: 21.342 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:149 Resp: 24674

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 28.0 | 21.9 | 32.9 |
| 279 | 4.3 | 3.5 | 5.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 24.495 min Scan# 3665
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:264 Resp: 2178150

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.6 | 18.6 | 28.0 |
| 265 | 21.5 | 17.7 | 26.5 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 3.935 ng m
RT: 28.024 min Scan# 4
Delta R.T. -0.012 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:276 Resp: 54614

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 14.0 | 13.0 | 19.6 |
| 277 | 17.8 | 15.9 | 23.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#88
Benzo(b)fluoranthene
Concen: 4.418 ng
RT: 23.507 min Scan# 3497
Delta R.T. -0.006 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:252 Resp: 560057

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.1 | 17.7 | 26.5 |
| 125 | 10.5 | 8.1 | 12.1 |






#89
Benzo(k)fluoranthene
Concen: 4.619 ng
RT: 23.571 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:252 Resp: 58266

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.8 | 17.7 | 26.5 |
| 125 | 10.2 | 8.2 | 12.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#90
Benzo(a)pyrene
Concen: 4.121 ng
RT: 24.354 min Scan# 3641
Delta R.T. -0.000 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:252 Resp: 444644

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.8 | 17.6 | 26.4 |
| 125 | 11.8 | 9.0 | 13.4 |






#91
Dibenzo(a,h)anthracene
Concen: 3.988 ng
RT: 28.112 min Scan# 4
Delta R.T. -0.024 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:278 Resp: 45743

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 15.5 | 13.1 | 19.7 |
| 279 | 23.4 | 18.9 | 28.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC005

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#92
Benzo(g,h,i)perylene
Concen: 4.207 ng
RT: 29.112 min Scan# 4450
Delta R.T. -0.012 min
Lab File: BP025072.D
Acq: 03 Jul 2025 10:21

Tgt Ion:276 Resp: 498456

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.4 | 19.3 | 28.9 |
| 138 | 21.9 | 16.9 | 25.3 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025073.D
Acq On : 03 Jul 2025 11:02
Operator : CG/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:51:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| **Internal Standards** | | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.449 | 152 | 342992 | 20.000 | ng | | 0.00 |
| 21) Naphthalene-d8 | 10.196 | 136 | 1376560 | 20.000 | ng | | 0.01 |
| 39) Acenaphthene-d10 | 14.096 | 164 | 924609 | 20.000 | ng | | 0.00 |
| 64) Phenanthrene-d10 | 16.913 | 188 | 1970410 | 20.000 | ng | | 0.00 |
| 76) Chrysene-d12 | 21.342 | 240 | 2172955 | 20.000 | ng | | 0.00 |
| 86) Perylene-d12 | 24.489 | 264 | 2220570 | 20.000 | ng | | 0.00 |
| **System Monitoring Compounds** | | | | | | | |
| 5) 2-Fluorophenol | 5.102 | 112 | 376746 | 19.462 | ng | | 0.00 |
| 7) Phenol-d6 | 6.631 | 99 | 482699 | 19.108 | ng | | 0.00 |
| 23) Nitrobenzene-d5 | 8.566 | 82 | 381978 | 18.454 | ng | | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.619 | 330 | 198235 | 18.149 | ng | | 0.00 |
| 45) 2-Fluorobiphenyl | 12.701 | 172 | 1211432 | 20.297 | ng | | 0.00 |
| 79) Terphenyl-d14 | 19.672 | 244 | 1997838 | 20.207 | ng | | 0.00 |
| **Target Compounds** | | | | | | | Qvalue |
| 2) 1,4-Dioxane | 3.126 | 88 | 77575 | 10.306 | ng | | 98 |
| 3) Pyridine | 3.490 | 79 | 197662 | 9.598 | ng | | 97 |
| 4) n-Nitrosodimethylamine | 3.408 | 42 | 85827 | 9.836 | ng | # | 92 |
| 6) Aniline | 6.790 | 93 | 227023 | 9.619 | ng | | 98 |
| 8) 2-Chlorophenol | 7.014 | 128 | 206980 | 9.556 | ng | | 96 |
| 9) Benzaldehyde | 6.602 | 77 | 154921 | 11.743 | ng | | 98 |
| 10) Phenol | 6.655 | 94 | 250593 | 9.763 | ng | | 99 |
| 11) bis(2-Chloroethyl)ether | 6.890 | 93 | 195263 | 9.885 | ng | | 98 |
| 12) 1,3-Dichlorobenzene | 7.343 | 146 | 245857 | 10.143 | ng | | 98 |
| 13) 1,4-Dichlorobenzene | 7.478 | 146 | 245523 | 10.004 | ng | | 98 |
| 14) 1,2-Dichlorobenzene | 7.790 | 146 | 237961 | 10.079 | ng | | 98 |
| 15) Benzyl Alcohol | 7.678 | 79 | 152168 | 9.025 | ng | | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.978 | 45 | 260708 | 10.150 | ng | | 97 |
| 17) 2-Methylphenol | 7.878 | 107 | 155778 | 9.422 | ng | | 98 |
| 18) Hexachloroethane | 8.502 | 117 | 75576 | 9.453 | ng | | 96 |
| 19) n-Nitroso-di-n-propyla... | 8.237 | 70 | 146424 | 9.774 | ng | | 97 |
| 20) 3+4-Methylphenols | 8.196 | 107 | 213641 | 9.225 | ng | | 97 |
| 22) Acetophenone | 8.243 | 105 | 330725 | 10.159 | ng | | 98 |
| 24) Nitrobenzene | 8.608 | 77 | 201551 | 9.560 | ng | | 99 |
| 25) Isophorone | 9.137 | 82 | 367740 | 9.507 | ng | | 100 |
| 26) 2-Nitrophenol | 9.313 | 139 | 63769 | 9.106 | ng | | 92 |
| 27) 2,4-Dimethylphenol | 9.390 | 122 | 134296 | 9.307 | ng | | 97 |
| 28) bis(2-Chloroethoxy)met... | 9.625 | 93 | 255336 | 9.722 | ng | | 99 |
| 29) 2,4-Dichlorophenol | 9.843 | 162 | 181263 | 9.059 | ng | | 99 |
| 30) 1,2,4-Trichlorobenzene | 10.066 | 180 | 219650 | 9.913 | ng | | 99 |
| 31) Naphthalene | 10.243 | 128 | 709141 | 10.005 | ng | | 99 |
| 32) Benzoic acid | 9.455 | 122 | 79225 | 10.455 | ng | | 95 |
| 33) 4-Chloroaniline | 10.343 | 127 | 254489 | 9.621 | ng | | 98 |
| 34) Hexachlorobutadiene | 10.549 | 225 | 124190 | 9.767 | ng | | 99 |
| 35) Caprolactam | 11.084 | 113 | 59547 | 8.700 | ng | | 94 |
| 36) 4-Chloro-3-methylphenol | 11.478 | 107 | 192224 | 9.193 | ng | | 99 |
| 37) 2-Methylnaphthalene | 11.872 | 142 | 489608 | 9.790 | ng | | 99 |
| 38) 1-Methylnaphthalene | 12.084 | 142 | 484403 | 9.931 | ng | | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 12.249 | 216 | 253871 | 9.554 | ng | | 99 |
| 41) Hexachlorocyclopentadiene | 12.243 | 237 | 55283 | 7.965 | ng | | 96 |
| 43) 2,4,6-Trichlorophenol | 12.490 | 196 | 136269 | 8.469 | ng | | 99 |

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025073.D
Acq On : 03 Jul 2025 11:02
Operator : CG/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:51:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 12.554 | 196 | 161200 | 8.840 | ng | | 98 |
| 46) 1,1'-Biphenyl | 12.913 | 154 | 663919 | 9.869 | ng | | 99 |
| 47) 2-Chloronaphthalene | 12.949 | 162 | 503580 | 9.959 | ng | | 98 |
| 48) 2-Nitroaniline | 13.143 | 65 | 85669 | 9.283 | ng | | 97 |
| 49) Acenaphthylene | 13.807 | 152 | 750027 | 9.692 | ng | | 99 |
| 50) Dimethylphthalate | 13.554 | 163 | 636102 | 9.828 | ng | | 99 |
| 51) 2,6-Dinitrotoluene | 13.660 | 165 | 110177 | 8.728 | ng | | 95 |
| 52) Acenaphthene | 14.160 | 154 | 496424m | 9.953 | ng | | |
| 53) 3-Nitroaniline | 13.990 | 138 | 114723 | 8.481 | ng | # | 95 |
| 54) 2,4-Dinitrophenol | 14.201 | 184 | 35399 | 10.737 | ng | | 96 |
| 55) Dibenzofuran | 14.507 | 168 | 838464 | 10.186 | ng | | 100 |
| 56) 4-Nitrophenol | 14.301 | 139 | 98038 | 7.853 | ng | | 92 |
| 57) 2,4-Dinitrotoluene | 14.466 | 165 | 131885 | 9.365 | ng | | 98 |
| 58) Fluorene | 15.172 | 166 | 639683 | 10.047 | ng | | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 14.731 | 232 | 146800 | 8.890 | ng | | 97 |
| 60) Diethylphthalate | 14.960 | 149 | 636288 | 9.788 | ng | | 98 |
| 61) 4-Chlorophenyl-phenyle... | 15.184 | 204 | 317737 | 10.084 | ng | | 98 |
| 62) 4-Nitroaniline | 15.178 | 138 | 123493 | 8.706 | ng | | 93 |
| 63) Azobenzene | 15.478 | 77 | 589360 | 10.181 | ng | | 97 |
| 65) 4,6-Dinitro-2-methylph... | 15.243 | 198 | 52978 | 10.800 | ng | | 87 |
| 66) n-Nitrosodiphenylamine | 15.390 | 169 | 539820 | 9.573 | ng | | 99 |
| 67) 4-Bromophenyl-phenylether | 16.090 | 248 | 186736 | 9.256 | ng | | 93 |
| 68) Hexachlorobenzene | 16.195 | 284 | 236838 | 9.604 | ng | | 98 |
| 69) Atrazine | 16.384 | 200 | 164479 | 9.954 | ng | | 100 |
| 70) Pentachlorophenol | 16.554 | 266 | 129459 | 7.774 | ng | | 98 |
| 71) Phenanthrene | 16.960 | 178 | 1068228 | 10.146 | ng | | 99 |
| 72) Anthracene | 17.042 | 178 | 980773 | 9.710 | ng | | 99 |
| 73) Carbazole | 17.331 | 167 | 979377 | 9.917 | ng | | 99 |
| 74) Di-n-butylphthalate | 17.966 | 149 | 979540 | 8.907 | ng | | 100 |
| 75) Fluoranthene | 19.060 | 202 | 1203752 | 9.781 | ng | | 99 |
| 77) Benzidine | 19.266 | 184 | 192008 | 5.003 | ng | | 99 |
| 78) Pyrene | 19.442 | 202 | 1286495 | 9.559 | ng | | 99 |
| 80) Butylbenzylphthalate | 20.430 | 149 | 300028 | 9.024 | ng | | 95 |
| 81) Benzo(a)anthracene | 21.325 | 228 | 1276106 | 9.691 | ng | | 99 |
| 82) 3,3'-Dichlorobenzidine | 21.272 | 252 | 270931 | 6.834 | ng | | 99 |
| 83) Chrysene | 21.389 | 228 | 1246003 | 9.832 | ng | | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 21.325 | 149 | 564998 | 7.999 | ng | | 97 |
| 85) Di-n-octyl phthalate | 22.554 | 149 | 662011 | 10.561 | ng | | 97 |
| 87) Indeno(1,2,3-cd)pyrene | 28.042 | 276 | 1222863 | 8.642 | ng | | 98 |
| 88) Benzo(b)fluoranthene | 23.495 | 252 | 1180548 | 9.134 | ng | | 100 |
| 89) Benzo(k)fluoranthene | 23.560 | 252 | 1252989 | 9.743 | ng | | 99 |
| 90) Benzo(a)pyrene | 24.342 | 252 | 963455 | 8.758 | ng | | 99 |
| 91) Dibenzo(a,h)anthracene | 28.118 | 278 | 1022797 | 8.747 | ng | | 100 |
| 92) Benzo(g,h,i)perylene | 29.095 | 276 | 1073371 | 8.886 | ng | | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025073.D
Acq On : 03 Jul 2025 11:02
Operator : CG/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jul 03 15:51:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.449 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:152 Resp: 34299

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 154.9 | 126.1 | 189.1 |
| 115 | 55.2 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#2
1,4-Dioxane
Concen: 10.306 ng
RT: 3.126 min Scan# 32
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 88 Resp: 77575

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 70.0 | 55.0 | 82.4 |
| 43 | 28.7 | 21.9 | 32.9 |






#3
Pyridine
Concen: 9.598 ng
RT: 3.490 min Scan# 94
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 79 Resp: 19766

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 67.3 | 52.2 | 78.4 |
| 51 | 36.3 | 26.8 | 40.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#4
n-Nitrosodimethylamine
Concen: 9.836 ng
RT: 3.408 min Scan# 80
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 42 Resp: 85827

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 123.2 | 105.2 | 157.8 |
| 44 | 13.8 | 6.6 | 9.8# |






#5
2-Fluorophenol
Concen: 19.462 ng
RT: 5.102 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:112 Resp: 37674

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 51.6 | 42.0 | 63.0 |
| 63 | 27.7 | 22.2 | 33.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#6
Aniline
Concen: 9.619 ng
RT: 6.790 min Scan# 655
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 93 Resp: 227023

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 35.1 | 28.9 | 43.3 |
| 65 | 18.2 | 15.2 | 22.8 |






#7
Phenol-d6
Concen: 19.108 ng
RT: 6.631 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 99 Resp: 48269

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 17.4 | 13.6 | 20.4 |
| 71 | 29.7 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#8
2-Chlorophenol
Concen: 9.556 ng
RT: 7.014 min Scan# 693
Delta R.T. -0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:128 Resp: 206980

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 31.6 | 12.9 | 52.9 |
| 64 | 43.7 | 20.3 | 60.3 |






#9
Benzaldehyde
Concen: 11.743 ng
RT: 6.602 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 77 Resp: 15492

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 100.1 | 82.4 | 122.4 |
| 106 | 97.4 | 78.3 | 118.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#10
Phenol
Concen: 9.763 ng
RT: 6.655 min Scan# 632
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 94 Resp: 250593

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 26.3 | 6.9 | 46.9 |
| 66 | 34.9 | 14.6 | 54.6 |






#11
bis(2-Chloroethyl)ether
Concen: 9.885 ng
RT: 6.890 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 93 Resp: 19526

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 70.2 | 51.9 | 91.9 |
| 95 | 31.8 | 12.4 | 52.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#12
1,3-Dichlorobenzene
Concen: 10.143 ng
RT: 7.343 min Scan# 749
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:146 Resp: 245857

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.7 | 51.9 | 77.9 |
| 75 | 26.5 | 21.8 | 32.6 |






#13
1,4-Dichlorobenzene
Concen: 10.004 ng
RT: 7.478 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:146 Resp: 24552

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 65.4 | 51.1 | 76.7 |
| 111 | 37.7 | 31.2 | 46.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#14
1,2-Dichlorobenzene
Concen: 10.079 ng
RT: 7.790 min Scan# 825
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:146 Resp: 237961

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.0 | 51.1 | 76.7 |
| 111 | 39.5 | 31.2 | 46.8 |






#15
Benzyl Alcohol
Concen: 9.025 ng
RT: 7.678 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 79 Resp: 15216

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 80.7 | 66.0 | 99.0 |
| 77 | 67.4 | 53.4 | 80.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.150 ng
RT: 7.978 min Scan# 857
Delta R.T. 0.012 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 45 Resp: 260708

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 13.6 | 0.0 | 34.8 |
| 79 | 11.6 | 0.0 | 30.7 |






#17
2-Methylphenol
Concen: 9.422 ng
RT: 7.878 min Scan# 84
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:107 Resp: 15577

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 111.3 | 89.7 | 134.5 |
| 77 | 44.8 | 34.2 | 51.4 |
| 79 | 44.6 | 33.4 | 50.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#18
Hexachloroethane
Concen: 9.453 ng
RT: 8.502 min Scan# 946
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:117 Resp: 75576

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 100.9 | 76.6 | 115.0 |
| 201 | 104.6 | 85.4 | 128.2 |






#19
n-Nitroso-di-n-propylamine
Concen: 9.774 ng
RT: 8.237 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 70 Resp: 14642

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 59.7 | 46.5 | 69.7 |
| 101 | 9.3 | 7.9 | 11.9 |
| 130 | 21.9 | 19.3 | 28.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#20
3+4-Methylphenols
Concen: 9.225 ng
RT: 8.196 min Scan# 894
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:107 Resp: 213641

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 86.1 | 69.5 | 109.5 |
| 77 | 30.6 | 11.4 | 51.4 |
| 79 | 26.1 | 6.9 | 46.9 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.196 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:136 Resp: 137656

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.8 | 8.7 | 13.1 |
| 54 | 6.9 | 5.4 | 8.2 |
| 68 | 5.6 | 4.5 | 6.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#22
Acetophenone
Concen: 10.159 ng
RT: 8.243 min Scan# 902
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:105 Resp: 330725

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 6.5 | 5.4 | 8.2 |
| 51 | 25.4 | 19.6 | 29.4 |
| 120 | 22.7 | 18.8 | 28.2 |






#23
Nitrobenzene-d5
Concen: 18.454 ng
RT: 8.566 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 82 Resp: 38197

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 45.3 | 38.4 | 57.6 |
| 54 | 49.3 | 37.8 | 56.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#24
Nitrobenzene
Concen: 9.560 ng
RT: 8.608 min Scan# 964
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 77 Resp: 201551

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 49.2 | 39.0 | 58.6 |
| 65 | 14.0 | 10.7 | 16.1 |






#25
Isophorone
Concen: 9.507 ng
RT: 9.137 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 82 Resp: 3677

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.3 | 5.8 | 8.6 |
| 138 | 19.6 | 15.7 | 23.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#26
2-Nitrophenol
Concen: 9.106 ng
RT: 9.313 min Scan# 1084
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:139 Resp: 63769

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 29.8 | 22.3 | 33.5 |
| 65 | 48.7 | 33.8 | 50.8 |






#27
2,4-Dimethylphenol
Concen: 9.307 ng
RT: 9.390 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:122 Resp: 13429

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 113.9 | 88.2 | 132.2 |
| 121 | 58.7 | 46.1 | 69.1 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#28
bis(2-Chloroethoxy)methane
Concen: 9.722 ng
RT: 9.625 min Scan# 1137
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 93 Resp: 255336

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.4 | 26.4 | 39.6 |
| 123 | 12.4 | 10.1 | 15.1 |






#29
2,4-Dichlorophenol
Concen: 9.059 ng
RT: 9.843 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:162 Resp: 18126

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 63.9 | 44.1 | 84.1 |
| 98 | 31.9 | 13.1 | 53.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#30
1,2,4-Trichlorobenzene
Concen: 9.913 ng
RT: 10.066 min Scan# 1212
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:180 Resp: 219650

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 94.5 | 75.1 | 112.7 |
| 145 | 28.6 | 23.3 | 34.9 |






#31
Naphthalene
Concen: 10.005 ng
RT: 10.243 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:128 Resp: 70914

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.0 | 8.9 | 13.3 |
| 127 | 12.6 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#32
Benzoic acid
Concen: 10.455 ng
RT: 9.455 min Scan# 1108
Delta R.T. -0.046 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:122 Resp: 79225

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 114.4 | 101.0 | 141.0 |
| 77 | 84.3 | 67.6 | 107.6 |






#33
4-Chloroaniline
Concen: 9.621 ng
RT: 10.343 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:127 Resp: 25448

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.5 | 25.4 | 38.2 |
| 65 | 27.3 | 21.0 | 31.6 |
| 92 | 17.6 | 13.4 | 20.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#34
Hexachlorobutadiene
Concen: 9.767 ng
RT: 10.549 min Scan# 1294
Delta R.T. 0.012 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:225 Resp: 124190

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 60.9 | 50.3 | 75.5 |
| 227 | 63.8 | 51.3 | 76.9 |






#35
Caprolactam
Concen: 8.700 ng
RT: 11.084 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:113 Resp: 5954

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 157.7 | 128.1 | 168.1 |
| 56 | 114.6 | 90.7 | 130.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#36
4-Chloro-3-methylphenol
Concen: 9.193 ng
RT: 11.478 min Scan# 1452
Delta R.T. -0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:107 Resp: 192224

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 27.6 | 22.4 | 33.6 |
| 142 | 87.9 | 69.6 | 104.4 |






#37
2-Methylnaphthalene
Concen: 9.790 ng
RT: 11.872 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:142 Resp: 48960

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 88.4 | 71.2 | 106.8 |
| 115 | 32.6 | 25.9 | 38.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#38
1-Methylnaphthalene
Concen: 9.931 ng
RT: 12.084 min Scan# 1555
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:142 Resp: 484403

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 92.7 | 74.1 | 111.1 |
| 116 | 3.5 | 2.7 | 4.1 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.096 min Scan# 1
Delta R.T. 0.007 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:164 Resp: 92460

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 102.0 | 80.3 | 120.5 |
| 160 | 46.5 | 35.9 | 53.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 9.554 ng
RT: 12.249 min Scan# 1583
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:216 Resp: 253871

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.5 | 62.4 | 93.6 |
| 179 | 21.3 | 16.5 | 24.7 |
| 108 | 18.1 | 13.7 | 20.5 |






#41
Hexachlorocyclopentadiene
Concen: 7.965 ng
RT: 12.243 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:237 Resp: 5528

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 59.0 | 41.6 | 81.6 |
| 272 | 11.8 | 0.0 | 33.6 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#42
2,4,6-Tribromophenol
Concen: 18.149 ng
RT: 15.619 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:330 Resp: 198235

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 97.6 | 77.7 | 116.5 |
| 141 | 33.5 | 26.0 | 39.0 |






#43
2,4,6-Trichlorophenol
Concen: 8.469 ng
RT: 12.490 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:196 Resp: 13626

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 94.8 | 76.2 | 114.2 |
| 200 | 30.9 | 24.3 | 36.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#44
2,4,5-Trichlorophenol
Concen: 8.840 ng
RT: 12.554 min Scan# 1635
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:196 Resp: 161200

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 97.5 | 76.6 | 115.0 |
| 97 | 47.6 | 36.6 | 55.0 |
| 132 | 27.0 | 21.4 | 32.0 |






#45
2-Fluorobiphenyl
Concen: 20.297 ng
RT: 12.701 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:172 Resp: 121143

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 34.6 | 28.2 | 42.2 |
| 170 | 23.2 | 18.7 | 28.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#46
1,1'-Biphenyl
Concen: 9.869 ng
RT: 12.913 min Scan# 1696
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:154 Resp: 663919

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.5 | 20.2 | 60.2 |
| 76 | 12.1 | 0.0 | 31.5 |






#47
2-Chloronaphthalene
Concen: 9.959 ng
RT: 12.949 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:162 Resp: 50358

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 36.1 | 29.4 | 44.2 |
| 164 | 31.6 | 26.4 | 39.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#48
2-Nitroaniline
Concen: 9.283 ng
RT: 13.143 min Scan# 1735
Delta R.T. 0.001 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 65 Resp: 85669

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 75.7 | 58.7 | 88.1 |
| 138 | 119.3 | 99.0 | 148.4 |






#49
Acenaphthylene
Concen: 9.692 ng
RT: 13.807 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:152 Resp: 75002

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 19.6 | 16.1 | 24.1 |
| 153 | 13.2 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#50
Dimethylphthalate
Concen: 9.828 ng
RT: 13.554 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:163 Resp: 636102

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.1 | 3.4 | 5.0 |
| 164 | 10.4 | 8.0 | 12.0 |






#51
2,6-Dinitrotoluene
Concen: 8.728 ng
RT: 13.660 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:165 Resp: 11017

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 51.6 | 38.5 | 57.7 |
| 89 | 53.7 | 40.6 | 61.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#52
Acenaphthene
Concen: 9.953 ng m
RT: 14.160 min Scan# 1908
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:154 Resp: 496424

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 114.3 | 90.6 | 136.0 |
| 152 | 57.0 | 43.8 | 65.8 |






#53
3-Nitroaniline
Concen: 8.481 ng
RT: 13.990 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:138 Resp: 11472

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 13.6 | 7.8 | 11.8 |
| 92 | 105.5 | 81.0 | 121.4 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#54
2,4-Dinitrophenol
Concen: 10.737 ng
RT: 14.201 min Scan# 1915
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:184 Resp: 35399

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 54.7 | 47.4 | 71.0 |
| 154 | 61.2 | 47.8 | 71.8 |






#55
Dibenzofuran
Concen: 10.186 ng
RT: 14.507 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:168 Resp: 83846

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 36.0 | 28.6 | 43.0 |
| 169 | 13.8 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#56
4-Nitrophenol
Concen: 7.853 ng
RT: 14.301 min Scan# 1932
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:139 Resp: 98038

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 72.6 | 45.9 | 85.9 |
| 65 | 88.2 | 61.4 | 101.4 |






#57
2,4-Dinitrotoluene
Concen: 9.365 ng
RT: 14.466 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:165 Resp: 13188

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 37.5 | 31.0 | 46.4 |
| 89 | 67.9 | 53.4 | 80.0 |
| 182 | 2.5 | 2.0 | 3.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#58
Fluorene
Concen: 10.047 ng
RT: 15.172 min Scan# 2080
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:166 Resp: 639683

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 98.9 | 78.6 | 117.8 |
| 167 | 13.5 | 10.7 | 16.1 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 8.890 ng
RT: 14.731 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:232 Resp: 14680

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 46.4 | 35.7 | 53.5 |
| 130 | 2.5 | 1.8 | 2.6 |
| 166 | 32.5 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#60
Diethylphthalate
Concen: 9.788 ng
RT: 14.960 min Scan# 2044
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:149 Resp: 636288

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 21.3 | 17.8 | 26.8 |
| 150 | 12.1 | 10.2 | 15.2 |




#61
4-Chlorophenyl-phenylether
Concen: 10.084 ng
RT: 15.184 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:204 Resp: 31773

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 33.6 | 25.8 | 38.6 |
| 141 | 57.4 | 45.1 | 67.7 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#62
4-Nitroaniline
Concen: 8.706 ng
RT: 15.178 min Scan# 2081
Delta R.T. -0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:138 Resp: 123493

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 50.2 | 26.0 | 66.0 |
| 108 | 75.9 | 50.0 | 90.0 |






#63
Azobenzene
Concen: 10.181 ng
RT: 15.478 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion: 77 Resp: 58936

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 25.5 | 7.7 | 47.7 |
| 105 | 20.6 | 1.9 | 41.9 |
| 51 | 29.6 | 8.2 | 48.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.913 min Scan# 2376
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:188 Resp: 1970410

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 9.5 | 7.3 | 10.9 |
| 80 | 10.0 | 7.8 | 11.6 |






#65
4,6-Dinitro-2-methylphenol
Concen: 10.800 ng
RT: 15.243 min Scan# 2
Delta R.T. -0.005 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:198 Resp: 5297

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 46.1 | 16.8 | 56.8 |
| 105 | 46.9 | 20.8 | 60.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#66
n-Nitrosodiphenylamine
Concen: 9.573 ng
RT: 15.390 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:169 Resp: 539820

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 67.9 | 54.1 | 81.1 |
| 167 | 37.0 | 29.0 | 43.6 |






#67
4-Bromophenyl-phenylether
Concen: 9.256 ng
RT: 16.090 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:248 Resp: 18673

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 97.2 | 79.4 | 119.2 |
| 141 | 75.5 | 51.4 | 77.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#68
Hexachlorobenzene
Concen: 9.604 ng
RT: 16.195 min Scan# 2254
Delta R.T. -0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:284 Resp: 236838

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 33.3 | 25.9 | 38.9 |
| 249 | 28.9 | 23.8 | 35.8 |






#69
Atrazine
Concen: 9.954 ng
RT: 16.384 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:200 Resp: 16447

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 26.6 | 6.6 | 46.6 |
| 215 | 47.9 | 28.0 | 68.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#70
Pentachlorophenol
Concen: 7.774 ng
RT: 16.554 min Scan# 2315
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:266 Resp: 129459

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 65.5 | 50.4 | 75.6 |
| 264 | 63.0 | 50.2 | 75.2 |






#71
Phenanthrene
Concen: 10.146 ng
RT: 16.960 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:178 Resp: 106822

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.0 | 15.4 | 23.2 |
| 179 | 14.8 | 12.6 | 19.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#72
Anthracene
Concen: 9.710 ng
RT: 17.042 min Scan# 2398
Delta R.T. -0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:178 Resp: 980773

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.3 | 15.0 | 22.6 |
| 179 | 15.3 | 12.5 | 18.7 |






#73
Carbazole
Concen: 9.917 ng
RT: 17.331 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:167 Resp: 97937

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 20.9 | 17.4 | 26.0 |
| 139 | 12.5 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#74
Di-n-butylphthalate
Concen: 8.907 ng
RT: 17.966 min Scan# 2555
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:149 Resp: 979540

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.0 | 7.3 | 10.9 |
| 104 | 4.9 | 3.8 | 5.6 |






#75
Fluoranthene
Concen: 9.781 ng
RT: 19.060 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:202 Resp: 120375

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 11.5 | 0.0 | 30.4 |
| 203 | 17.4 | 0.0 | 37.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.342 min Scan# 3129
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:240 Resp: 2172955

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 10.7 | 9.0 | 13.4 |
| 236 | 25.7 | 20.3 | 30.5 |






#77
Benzidine
Concen: 5.003 ng
RT: 19.266 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:184 Resp: 19200

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 15.1 | 12.6 | 18.8 |
| 183 | 11.4 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#78
Pyrene
Concen: 9.559 ng
RT: 19.442 min Scan# 2806
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:202 Resp: 1286495

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 21.0 | 17.1 | 25.7 |
| 203 | 17.2 | 14.1 | 21.1 |






#79
Terphenyl-d14
Concen: 20.207 ng
RT: 19.672 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:244 Resp: 199783

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.5 | 5.9 | 8.9 |
| 122 | 11.0 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#80
Butylbenzylphthalate
Concen: 9.024 ng
RT: 20.430 min Scan# 2974
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:149 Resp: 300028

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 67.2 | 50.6 | 76.0 |
| 206 | 18.7 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 9.691 ng
RT: 21.325 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:228 Resp: 127610

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 27.0 | 22.1 | 33.1 |
| 229 | 19.6 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#82
3,3'-Dichlorobenzidine
Concen: 6.834 ng
RT: 21.272 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:252 Resp: 270931

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 64.1 | 51.9 | 77.9 |
| 126 | 11.0 | 8.1 | 12.1 |






#83
Chrysene
Concen: 9.832 ng
RT: 21.389 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:228 Resp: 124600

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 29.5 | 24.4 | 36.6 |
| 229 | 19.9 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 7.999 ng
RT: 21.325 min Scan# 3126
Delta R.T. -0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:149 Resp: 564998

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 25.7 | 21.9 | 32.9 |
| 279 | 3.9 | 3.5 | 5.3 |






#85
Di-n-octyl phthalate
Concen: 10.561 ng
RT: 22.554 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:149 Resp: 66201

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.5 | 1.3 | 1.9 |
| 43 | 9.9 | 6.9 | 10.3 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 24.489 min Scan# 3664
Delta R.T. -0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:264 Resp: 2220570

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 24.0 | 18.6 | 28.0 |
| 265 | 22.0 | 17.7 | 26.5 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 8.642 ng
RT: 28.042 min Scan# 4
Delta R.T. 0.006 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:276 Resp: 122286

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 15.3 | 13.0 | 19.6 |
| 277 | 18.8 | 15.9 | 23.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#88
Benzo(b)fluoranthene
Concen: 9.134 ng
RT: 23.495 min Scan# 3495
Delta R.T. -0.017 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:252 Resp: 1180548

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.9 | 17.7 | 26.5 |
| 125 | 10.3 | 8.1 | 12.1 |






#89
Benzo(k)fluoranthene
Concen: 9.743 ng
RT: 23.560 min Scan# 3
Delta R.T. -0.023 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:252 Resp: 125298

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.0 | 17.7 | 26.5 |
| 125 | 11.0 | 8.2 | 12.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#90
Benzo(a)pyrene
Concen: 8.758 ng
RT: 24.342 min Scan# 3639
Delta R.T. -0.012 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:252 Resp: 963455

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.8 | 17.6 | 26.4 |
| 125 | 11.5 | 9.0 | 13.4 |






#91
Dibenzo(a,h)anthracene
Concen: 8.747 ng
RT: 28.118 min Scan# 4
Delta R.T. -0.017 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:278 Resp: 102279

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 16.0 | 13.1 | 19.7 |
| 279 | 23.5 | 18.9 | 28.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC010

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#92
Benzo(g,h,i)perylene
Concen: 8.886 ng
RT: 29.095 min Scan# 4447
Delta R.T. -0.029 min
Lab File: BP025073.D
Acq: 03 Jul 2025 11:02

Tgt Ion:276 Resp: 1073371

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 24.5 | 19.3 | 28.9 |
| 138 | 21.8 | 16.9 | 25.3 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025074.D
Acq On : 03 Jul 2025 11:44
Operator : CG/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jul 03 15:52:25 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| **Internal Standards** | | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.443 | 152 | 357536 | 20.000 | ng | | 0.00 |
| 21) Naphthalene-d8 | 10.184 | 136 | 1460146 | 20.000 | ng | | 0.00 |
| 39) Acenaphthene-d10 | 14.089 | 164 | 975973 | 20.000 | ng | | 0.00 |
| 64) Phenanthrene-d10 | 16.913 | 188 | 1992018 | 20.000 | ng | | 0.00 |
| 76) Chrysene-d12 | 21.348 | 240 | 2159758 | 20.000 | ng | | 0.00 |
| 86) Perylene-d12 | 24.483 | 264 | 2243019 | 20.000 | ng | | -0.01 |
| **System Monitoring Compounds** | | | | | | | |
| 5) 2-Fluorophenol | 5.102 | 112 | 842975 | 41.776 | ng | | 0.00 |
| 7) Phenol-d6 | 6.625 | 99 | 1087166 | 41.286 | ng | | 0.00 |
| 23) Nitrobenzene-d5 | 8.566 | 82 | 918035 | 41.814 | ng | | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.607 | 330 | 468488 | 40.634 | ng | | -0.01 |
| 45) 2-Fluorobiphenyl | 12.690 | 172 | 2657285 | 42.179 | ng | | 0.00 |
| 79) Terphenyl-d14 | 19.672 | 244 | 4224886 | 42.993 | ng | | 0.00 |
| **Target Compounds** | | | | | | | Qvalue |
| 2) 1,4-Dioxane | 3.125 | 88 | 163489 | 20.836 | ng | | 98 |
| 3) Pyridine | 3.490 | 79 | 444402 | 20.702 | ng | | 98 |
| 4) n-Nitrosodimethylamine | 3.408 | 42 | 189864 | 20.874 | ng | # | 98 |
| 6) Aniline | 6.784 | 93 | 511196 | 20.777 | ng | | 99 |
| 8) 2-Chlorophenol | 7.013 | 128 | 472043 | 20.908 | ng | | 98 |
| 9) Benzaldehyde | 6.596 | 77 | 319740 | 23.250 | ng | | 99 |
| 10) Phenol | 6.655 | 94 | 554695 | 20.731 | ng | | 99 |
| 11) bis(2-Chloroethyl)ether | 6.884 | 93 | 429103 | 20.838 | ng | | 99 |
| 12) 1,3-Dichlorobenzene | 7.331 | 146 | 531035 | 21.016 | ng | | 99 |
| 13) 1,4-Dichlorobenzene | 7.484 | 146 | 538001 | 21.030 | ng | | 98 |
| 14) 1,2-Dichlorobenzene | 7.784 | 146 | 516666 | 20.995 | ng | | 100 |
| 15) Benzyl Alcohol | 7.666 | 79 | 354316 | 20.160 | ng | | 98 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.966 | 45 | 561269 | 20.963 | ng | | 99 |
| 17) 2-Methylphenol | 7.878 | 107 | 359925 | 20.883 | ng | | 99 |
| 18) Hexachloroethane | 8.502 | 117 | 171635 | 20.594 | ng | | 96 |
| 19) n-Nitroso-di-n-propyla... | 8.237 | 70 | 335331 | 21.473 | ng | | 97 |
| 20) 3+4-Methylphenols | 8.202 | 107 | 494507 | 20.484 | ng | | 99 |
| 22) Acetophenone | 8.243 | 105 | 733244 | 21.234 | ng | | 99 |
| 24) Nitrobenzene | 8.613 | 77 | 465912 | 20.834 | ng | | 98 |
| 25) Isophorone | 9.131 | 82 | 857174 | 20.891 | ng | | 100 |
| 26) 2-Nitrophenol | 9.313 | 139 | 171063 | 17.980 | ng | | 97 |
| 27) 2,4-Dimethylphenol | 9.384 | 122 | 318872 | 20.834 | ng | | 99 |
| 28) bis(2-Chloroethoxy)met... | 9.619 | 93 | 585299 | 21.009 | ng | | 99 |
| 29) 2,4-Dichlorophenol | 9.837 | 162 | 435473 | 20.519 | ng | | 99 |
| 30) 1,2,4-Trichlorobenzene | 10.060 | 180 | 483959 | 20.592 | ng | | 99 |
| 31) Naphthalene | 10.237 | 128 | 1578374 | 20.993 | ng | | 99 |
| 32) Benzoic acid | 9.472 | 122 | 222091 | 18.881 | ng | | 99 |
| 33) 4-Chloroaniline | 10.348 | 127 | 583020 | 20.779 | ng | | 98 |
| 34) Hexachlorobutadiene | 10.537 | 225 | 275058 | 20.394 | ng | | 99 |
| 35) Caprolactam | 11.096 | 113 | 148067 | 20.394 | ng | | 99 |
| 36) 4-Chloro-3-methylphenol | 11.478 | 107 | 468989 | 21.146 | ng | | 100 |
| 37) 2-Methylnaphthalene | 11.866 | 142 | 1118020 | 21.075 | ng | | 99 |
| 38) 1-Methylnaphthalene | 12.090 | 142 | 1099060 | 21.243 | ng | | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 12.243 | 216 | 570143 | 20.326 | ng | | 100 |
| 41) Hexachlorocyclopentadiene | 12.237 | 237 | 131777 | 17.986 | ng | | 97 |
| 43) 2,4,6-Trichlorophenol | 12.490 | 196 | 346793 | 20.418 | ng | | 98 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025074.D
Acq On : 03 Jul 2025 11:44
Operator : CG/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:52:25 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 12.560 | 196 | 395864 | 20.566 | ng | 99 |
| 46) 1,1'-Biphenyl | 12.907 | 154 | 1473578 | 20.751 | ng | 99 |
| 47) 2-Chloronaphthalene | 12.942 | 162 | 1113594 | 20.865 | ng | 99 |
| 48) 2-Nitroaniline | 13.137 | 65 | 228962 | 19.268 | ng | 97 |
| 49) Acenaphthylene | 13.795 | 152 | 1728416 | 21.159 | ng | 99 |
| 50) Dimethylphthalate | 13.554 | 163 | 1434045 | 20.991 | ng | 100 |
| 51) 2,6-Dinitrotoluene | 13.660 | 165 | 274371 | 20.592 | ng | 97 |
| 52) Acenaphthene | 14.154 | 154 | 1114579m | 21.170 | ng | |
| 53) 3-Nitroaniline | 13.989 | 138 | 290162 | 20.320 | ng | 95 |
| 54) 2,4-Dinitrophenol | 14.189 | 184 | 95858 | 19.115 | ng | 97 |
| 55) Dibenzofuran | 14.501 | 168 | 1821290 | 20.961 | ng | 99 |
| 56) 4-Nitrophenol | 14.307 | 139 | 253402 | 19.230 | ng | 97 |
| 57) 2,4-Dinitrotoluene | 14.460 | 165 | 354786 | 19.783 | ng | 95 |
| 58) Fluorene | 15.166 | 166 | 1431605 | 21.301 | ng | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 14.742 | 232 | 361259 | 20.726 | ng | 100 |
| 60) Diethylphthalate | 14.960 | 149 | 1433207 | 20.887 | ng | 99 |
| 61) 4-Chlorophenyl-phenyle... | 15.172 | 204 | 692727 | 20.828 | ng | 99 |
| 62) 4-Nitroaniline | 15.172 | 138 | 313433 | 20.933 | ng | 99 |
| 63) Azobenzene | 15.472 | 77 | 1295934 | 21.209 | ng | 100 |
| 65) 4,6-Dinitro-2-methylph... | 15.242 | 198 | 138316 | 19.064 | ng | 99 |
| 66) n-Nitrosodiphenylamine | 15.395 | 169 | 1189868 | 20.872 | ng | 99 |
| 67) 4-Bromophenyl-phenylether | 16.095 | 248 | 425191 | 20.847 | ng | 95 |
| 68) Hexachlorobenzene | 16.201 | 284 | 507935 | 20.375 | ng | 97 |
| 69) Atrazine | 16.378 | 200 | 363960 | 21.787 | ng | 99 |
| 70) Pentachlorophenol | 16.548 | 266 | 324553 | 19.277 | ng | 99 |
| 71) Phenanthrene | 16.948 | 178 | 2198840 | 20.658 | ng | 100 |
| 72) Anthracene | 17.042 | 178 | 2176257 | 21.311 | ng | 100 |
| 73) Carbazole | 17.319 | 167 | 2105955 | 21.092 | ng | 99 |
| 74) Di-n-butylphthalate | 17.960 | 149 | 2284685 | 20.548 | ng | 100 |
| 75) Fluoranthene | 19.060 | 202 | 2649288 | 21.293 | ng | 100 |
| 77) Benzidine | 19.260 | 184 | 764455 | 20.041 | ng | 100 |
| 78) Pyrene | 19.436 | 202 | 2780845 | 20.788 | ng | 100 |
| 80) Butylbenzylphthalate | 20.430 | 149 | 815246 | 18.287 | ng | 99 |
| 81) Benzo(a)anthracene | 21.330 | 228 | 2791625 | 21.329 | ng | 99 |
| 82) 3,3'-Dichlorobenzidine | 21.260 | 252 | 734695 | 18.647 | ng | 100 |
| 83) Chrysene | 21.395 | 228 | 2642166 | 20.977 | ng | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 21.330 | 149 | 1396109 | 19.886 | ng | 99 |
| 85) Di-n-octyl phthalate | 22.548 | 149 | 1892919 | 18.935 | ng | 98 |
| 87) Indeno(1,2,3-cd)pyrene | 28.018 | 276 | 2955341 | 20.677 | ng | 97 |
| 88) Benzo(b)fluoranthene | 23.495 | 252 | 2724905 | 20.872 | ng | 100 |
| 89) Benzo(k)fluoranthene | 23.560 | 252 | 2800667 | 21.560 | ng | 99 |
| 90) Benzo(a)pyrene | 24.330 | 252 | 2321462 | 20.892 | ng | 99 |
| 91) Dibenzo(a,h)anthracene | 28.101 | 278 | 2461745 | 20.842 | ng | 99 |
| 92) Benzo(g,h,i)perylene | 29.095 | 276 | 2525145 | 20.695 | ng | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025074.D
Acq On : 03 Jul 2025 11:44
Operator : CG/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Jul 03 15:52:25 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.443 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:152 Resp: 35753

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 158.3 | 126.1 | 189.1 |
| 115 | 56.4 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#2
1,4-Dioxane
Concen: 20.836 ng
RT: 3.125 min Scan# 32
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 88 Resp: 163489

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 70.1 | 55.0 | 82.4 |
| 43 | 28.4 | 21.9 | 32.9 |






#3
Pyridine
Concen: 20.702 ng
RT: 3.490 min Scan# 94
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 79 Resp: 44440

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 67.2 | 52.2 | 78.4 |
| 51 | 33.9 | 26.8 | 40.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#4
n-Nitrosodimethylamine
Concen: 20.874 ng
RT: 3.408 min Scan# 80
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 42 Resp: 189864

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 130.0 | 105.2 | 157.8 |
| 44 | 10.1 | 6.6 | 9.8# |






#5
2-Fluorophenol
Concen: 41.776 ng
RT: 5.102 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:112 Resp: 84297

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 52.2 | 42.0 | 63.0 |
| 63 | 27.3 | 22.2 | 33.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#6
Aniline
Concen: 20.777 ng
RT: 6.784 min Scan# 654
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 93 Resp: 511196

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 37.0 | 28.9 | 43.3 |
| 65 | 18.9 | 15.2 | 22.8 |






#7
Phenol-d6
Concen: 41.286 ng
RT: 6.625 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 99 Resp: 108716

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 17.7 | 13.6 | 20.4 |
| 71 | 31.3 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#8
2-Chlorophenol
Concen: 20.908 ng
RT: 7.013 min Scan# 693
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:128 Resp: 472043

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 31.9 | 12.9 | 52.9 |
| 64 | 41.7 | 20.3 | 60.3 |






#9
Benzaldehyde
Concen: 23.250 ng
RT: 6.596 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 77 Resp: 31974

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 102.6 | 82.4 | 122.4 |
| 106 | 99.5 | 78.3 | 118.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#10
Phenol
Concen: 20.731 ng
RT: 6.655 min Scan# 632
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 94 Resp: 554695

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 26.2 | 6.9 | 46.9 |
| 66 | 34.3 | 14.6 | 54.6 |






#11
bis(2-Chloroethyl)ether
Concen: 20.838 ng
RT: 6.884 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 93 Resp: 42910

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 72.6 | 51.9 | 91.9 |
| 95 | 32.4 | 12.4 | 52.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#12
1,3-Dichlorobenzene
Concen: 21.016 ng
RT: 7.331 min Scan# 747
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:146 Resp: 531035

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.4 | 51.9 | 77.9 |
| 75 | 27.9 | 21.8 | 32.6 |






#13
1,4-Dichlorobenzene
Concen: 21.030 ng
RT: 7.484 min Scan# 7
Delta R.T. 0.012 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:146 Resp: 53800

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.9 | 51.1 | 76.7 |
| 111 | 36.2 | 31.2 | 46.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#14
1,2-Dichlorobenzene
Concen: 20.995 ng
RT: 7.784 min Scan# 824
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:146 Resp: 516666

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.0 | 51.1 | 76.7 |
| 111 | 39.5 | 31.2 | 46.8 |






#15
Benzyl Alcohol
Concen: 20.160 ng
RT: 7.666 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 79 Resp: 35431

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 84.7 | 66.0 | 99.0 |
| 77 | 67.4 | 53.4 | 80.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 20.963 ng
RT: 7.966 min Scan# 855
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 45 Resp: 561269

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 14.0 | 0.0 | 34.8 |
| 79 | 10.5 | 0.0 | 30.7 |






#17
2-Methylphenol
Concen: 20.883 ng
RT: 7.878 min Scan# 84
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:107 Resp: 35992

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 113.8 | 89.7 | 134.5 |
| 77 | 42.8 | 34.2 | 51.4 |
| 79 | 42.1 | 33.4 | 50.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#18
Hexachloroethane
Concen: 20.594 ng
RT: 8.502 min Scan# 946
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:117 Resp: 171635

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 94.5 | 76.6 | 115.0 |
| 201 | 100.0 | 85.4 | 128.2 |






#19
n-Nitroso-di-n-propylamine
Concen: 21.473 ng
RT: 8.237 min Scan# 9(
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

Tgt Ion: 70 Resp: 33533

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 60.5 | 46.5 | 69.7 |
| 101 | 10.1 | 7.9 | 11.9 |
| 130 | 22.8 | 19.3 | 28.9 |

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#20
3+4-Methylphenols
Concen: 20.484 ng
RT: 8.202 min Scan# 895
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:107 Resp: 494507

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 89.5 | 69.5 | 109.5 |
| 77 | 30.3 | 11.4 | 51.4 |
| 79 | 27.2 | 6.9 | 46.9 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.184 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:136 Resp: 146014

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 11.3 | 8.7 | 13.1 |
| 54 | 6.8 | 5.4 | 8.2 |
| 68 | 5.7 | 4.5 | 6.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#22
Acetophenone
Concen: 21.234 ng
RT: 8.243 min Scan# 902
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:105 Resp: 733244

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 7.5 | 5.4 | 8.2 |
| 51 | 24.6 | 19.6 | 29.4 |
| 120 | 23.0 | 18.8 | 28.2 |






#23
Nitrobenzene-d5
Concen: 41.814 ng
RT: 8.566 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 82 Resp: 91803

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 44.8 | 38.4 | 57.6 |
| 54 | 47.7 | 37.8 | 56.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#24
Nitrobenzene
Concen: 20.834 ng
RT: 8.613 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 77 Resp: 465912

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 50.0 | 39.0 | 58.6 |
| 65 | 12.8 | 10.7 | 16.1 |






#25
Isophorone
Concen: 20.891 ng
RT: 9.131 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 82 Resp: 85717

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.2 | 5.8 | 8.6 |
| 138 | 19.5 | 15.7 | 23.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#26
2-Nitrophenol
Concen: 17.980 ng
RT: 9.313 min Scan# 1084
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:139 Resp: 171063

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 29.1 | 22.3 | 33.5 |
| 65 | 44.4 | 33.8 | 50.8 |






#27
2,4-Dimethylphenol
Concen: 20.834 ng
RT: 9.384 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:122 Resp: 31887

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 109.0 | 88.2 | 132.2 |
| 121 | 58.2 | 46.1 | 69.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#28
bis(2-Chloroethoxy)methane
Concen: 21.009 ng
RT: 9.619 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 93 Resp: 585299

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.3 | 26.4 | 39.6 |
| 123 | 11.9 | 10.1 | 15.1 |






#29
2,4-Dichlorophenol
Concen: 20.519 ng
RT: 9.837 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:162 Resp: 43547

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 63.5 | 44.1 | 84.1 |
| 98 | 33.7 | 13.1 | 53.1 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#30
1,2,4-Trichlorobenzene
Concen: 20.592 ng
RT: 10.060 min Scan# 1211
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:180 Resp: 483959

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 94.9 | 75.1 | 112.7 |
| 145 | 28.5 | 23.3 | 34.9 |






#31
Naphthalene
Concen: 20.993 ng
RT: 10.237 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:128 Resp: 157837

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 10.7 | 8.9 | 13.3 |
| 127 | 12.9 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#32
Benzoic acid
Concen: 18.881 ng
RT: 9.472 min Scan# 1111
Delta R.T. -0.029 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:122 Resp: 222091

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 122.3 | 101.0 | 141.0 |
| 77 | 87.8 | 67.6 | 107.6 |






#33
4-Chloroaniline
Concen: 20.779 ng
RT: 10.348 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:127 Resp: 58302

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.9 | 25.4 | 38.2 |
| 65 | 27.6 | 21.0 | 31.6 |
| 92 | 17.3 | 13.4 | 20.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#34
Hexachlorobutadiene
Concen: 20.394 ng
RT: 10.537 min Scan# 1292
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:225 Resp: 275058

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 62.2 | 50.3 | 75.5 |
| 227 | 63.5 | 51.3 | 76.9 |






#35
Caprolactam
Concen: 20.394 ng
RT: 11.096 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:113 Resp: 14806

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 149.7 | 128.1 | 168.1 |
| 56 | 109.2 | 90.7 | 130.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#36
4-Chloro-3-methylphenol
Concen: 21.146 ng
RT: 11.478 min Scan# 1452
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:107 Resp: 468989

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 28.8 | 22.4 | 33.6 |
| 142 | 87.0 | 69.6 | 104.4 |






#37
2-Methylnaphthalene
Concen: 21.075 ng
RT: 11.866 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:142 Resp: 111802

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 89.7 | 71.2 | 106.8 |
| 115 | 31.5 | 25.9 | 38.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#38
1-Methylnaphthalene
Concen: 21.243 ng
RT: 12.090 min Scan# 1556
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:142 Resp: 1099060

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 91.6 | 74.1 | 111.1 |
| 116 | 3.6 | 2.7 | 4.1 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.089 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:164 Resp: 97597

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 101.9 | 80.3 | 120.5 |
| 160 | 45.4 | 35.9 | 53.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.326 ng
RT: 12.243 min Scan# 1582
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:216 Resp: 570143

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.4 | 62.4 | 93.6 |
| 179 | 20.9 | 16.5 | 24.7 |
| 108 | 17.0 | 13.7 | 20.5 |






#41
Hexachlorocyclopentadiene
Concen: 17.986 ng
RT: 12.237 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:237 Resp: 13177

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 64.3 | 41.6 | 81.6 |
| 272 | 14.0 | 0.0 | 33.6 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#42
2,4,6-Tribromophenol
Concen: 40.634 ng
RT: 15.607 min Scan# 2154
Delta R.T. -0.012 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:330 Resp: 468488

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 97.7 | 77.7 | 116.5 |
| 141 | 33.2 | 26.0 | 39.0 |






#43
2,4,6-Trichlorophenol
Concen: 20.418 ng
RT: 12.490 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:196 Resp: 34679

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 97.0 | 76.2 | 114.2 |
| 200 | 31.7 | 24.3 | 36.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#44
2,4,5-Trichlorophenol
Concen: 20.566 ng
RT: 12.560 min Scan# 1636
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:196 Resp: 395864

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.1 | 76.6 | 115.0 |
| 97 | 44.7 | 36.6 | 55.0 |
| 132 | 27.5 | 21.4 | 32.0 |






#45
2-Fluorobiphenyl
Concen: 42.179 ng
RT: 12.690 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:172 Resp: 265728

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.2 | 28.2 | 42.2 |
| 170 | 23.6 | 18.7 | 28.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#46
1,1'-Biphenyl
Concen: 20.751 ng
RT: 12.907 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:154 Resp: 1473578

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 39.8 | 20.2 | 60.2 |
| 76 | 11.7 | 0.0 | 31.5 |






#47
2-Chloronaphthalene
Concen: 20.865 ng
RT: 12.942 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:162 Resp: 111359

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 36.2 | 29.4 | 44.2 |
| 164 | 32.4 | 26.4 | 39.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#48
2-Nitroaniline
Concen: 19.268 ng
RT: 13.137 min Scan# 1734
Delta R.T. -0.005 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 65 Resp: 228962

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 70.6 | 58.7 | 88.1 |
| 138 | 119.6 | 99.0 | 148.4 |






#49
Acenaphthylene
Concen: 21.159 ng
RT: 13.795 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:152 Resp: 172841

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 19.8 | 16.1 | 24.1 |
| 153 | 13.2 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#50
Dimethylphthalate
Concen: 20.991 ng
RT: 13.554 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:163 Resp: 1434045

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 3.9 | 3.4 | 5.0 |
| 164 | 10.0 | 8.0 | 12.0 |






#51
2,6-Dinitrotoluene
Concen: 20.592 ng
RT: 13.660 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:165 Resp: 27437

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 49.3 | 38.5 | 57.7 |
| 89 | 53.1 | 40.6 | 61.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#52
Acenaphthene
Concen: 21.170 ng m
RT: 14.154 min Scan# 1907
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:154 Resp: 1114579

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 112.7 | 90.6 | 136.0 |
| 152 | 55.2 | 43.8 | 65.8 |






#53
3-Nitroaniline
Concen: 20.320 ng
RT: 13.989 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:138 Resp: 29016

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 11.3 | 7.8 | 11.8 |
| 92 | 106.9 | 81.0 | 121.4 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#54
2,4-Dinitrophenol
Concen: 19.115 ng
RT: 14.189 min Scan# 1913
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:184 Resp: 95858

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 57.3 | 47.4 | 71.0 |
| 154 | 61.9 | 47.8 | 71.8 |






#55
Dibenzofuran
Concen: 20.961 ng
RT: 14.501 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:168 Resp: 182129

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 36.2 | 28.6 | 43.0 |
| 169 | 13.2 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#56
4-Nitrophenol
Concen: 19.230 ng
RT: 14.307 min Scan# 1933
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:139 Resp: 253402

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 68.2 | 45.9 | 85.9 |
| 65 | 84.0 | 61.4 | 101.4 |






#57
2,4-Dinitrotoluene
Concen: 19.783 ng
RT: 14.460 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:165 Resp: 35478

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 35.3 | 31.0 | 46.4 |
| 89 | 63.5 | 53.4 | 80.0 |
| 182 | 2.7 | 2.0 | 3.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#58
Fluorene
Concen: 21.301 ng
RT: 15.166 min Scan# 2079
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:166 Resp: 1431605

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 97.4 | 78.6 | 117.8 |
| 167 | 13.6 | 10.7 | 16.1 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 20.726 ng
RT: 14.742 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:232 Resp: 36125

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 44.2 | 35.7 | 53.5 |
| 130 | 2.3 | 1.8 | 2.6 |
| 166 | 30.3 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#60
Diethylphthalate
Concen: 20.887 ng
RT: 14.960 min Scan# 2044
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:149 Resp: 1433207

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 21.9 | 17.8 | 26.8 |
| 150 | 12.4 | 10.2 | 15.2 |






#61
4-Chlorophenyl-phenylether
Concen: 20.828 ng
RT: 15.172 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:204 Resp: 69272

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 33.3 | 25.8 | 38.6 |
| 141 | 56.6 | 45.1 | 67.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#62
4-Nitroaniline
Concen: 20.933 ng
RT: 15.172 min Scan# 2080
Delta R.T. -0.012 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:138 Resp: 313433

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 46.8 | 26.0 | 66.0 |
| 108 | 70.9 | 50.0 | 90.0 |






#63
Azobenzene
Concen: 21.209 ng
RT: 15.472 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion: 77 Resp: 129593

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 27.4 | 7.7 | 47.7 |
| 105 | 21.9 | 1.9 | 41.9 |
| 51 | 28.1 | 8.2 | 48.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.913 min Scan# 2376
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:188 Resp: 1992018

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 8.5 | 7.3 | 10.9 |
| 80 | 9.0 | 7.8 | 11.6 |






#65
4,6-Dinitro-2-methylphenol
Concen: 19.064 ng
RT: 15.242 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:198 Resp: 13831

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 36.5 | 16.8 | 56.8 |
| 105 | 39.9 | 20.8 | 60.8 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#66
n-Nitrosodiphenylamine
Concen: 20.872 ng
RT: 15.395 min Scan# 2118
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:169 Resp: 1189868

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 67.6 | 54.1 | 81.1 |
| 167 | 37.2 | 29.0 | 43.6 |






#67
4-Bromophenyl-phenylether
Concen: 20.847 ng
RT: 16.095 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:248 Resp: 42519

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 99.7 | 79.4 | 119.2 |
| 141 | 73.5 | 51.4 | 77.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#68
Hexachlorobenzene
Concen: 20.375 ng
RT: 16.201 min Scan# 2255
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:284 Resp: 507935

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 34.4 | 25.9 | 38.9 |
| 249 | 31.4 | 23.8 | 35.8 |






#69
Atrazine
Concen: 21.787 ng
RT: 16.378 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:200 Resp: 36396

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 26.8 | 6.6 | 46.6 |
| 215 | 48.5 | 28.0 | 68.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#70
Pentachlorophenol
Concen: 19.277 ng
RT: 16.548 min Scan# 2314
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:266 Resp: 324553

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 62.4 | 50.4 | 75.6 |
| 264 | 61.9 | 50.2 | 75.2 |






#71
Phenanthrene
Concen: 20.658 ng
RT: 16.948 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:178 Resp: 219884

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.3 | 15.4 | 23.2 |
| 179 | 15.6 | 12.6 | 19.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#72
Anthracene
Concen: 21.311 ng
RT: 17.042 min Scan# 2398
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:178 Resp: 2176257

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.6 | 15.0 | 22.6 |
| 179 | 15.7 | 12.5 | 18.7 |






#73
Carbazole
Concen: 21.092 ng
RT: 17.319 min Scan# 24
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:167 Resp: 210595

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.3 | 17.4 | 26.0 |
| 139 | 13.0 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#74
Di-n-butylphthalate
Concen: 20.548 ng
RT: 17.960 min Scan# 2554
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:149 Resp: 2284685

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.3 | 7.3 | 10.9 |
| 104 | 4.8 | 3.8 | 5.6 |






#75
Fluoranthene
Concen: 21.293 ng
RT: 19.060 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:202 Resp: 264928

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 10.5 | 0.0 | 30.4 |
| 203 | 17.5 | 0.0 | 37.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.348 min Scan# 3130
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:240 Resp: 2159758

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.2 | 9.0 | 13.4 |
| 236 | 25.9 | 20.3 | 30.5 |






#77
Benzidine
Concen: 20.041 ng
RT: 19.260 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:184 Resp: 76445

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 15.4 | 12.6 | 18.8 |
| 183 | 11.9 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#78
Pyrene
Concen: 20.788 ng
RT: 19.436 min Scan# 2805
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:202 Resp: 2780845

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 21.1 | 17.1 | 25.7 |
| 203 | 17.6 | 14.1 | 21.1 |






#79
Terphenyl-d14
Concen: 42.993 ng
RT: 19.672 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:244 Resp: 422488

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.7 | 5.9 | 8.9 |
| 122 | 10.3 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#80
Butylbenzylphthalate
Concen: 18.287 ng
RT: 20.430 min Scan# 2974
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:149 Resp: 815246

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 62.3 | 50.6 | 76.0 |
| 206 | 21.1 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 21.329 ng
RT: 21.330 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:228 Resp: 279162

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 27.1 | 22.1 | 33.1 |
| 229 | 19.6 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#82
3,3'-Dichlorobenzidine
Concen: 18.647 ng
RT: 21.260 min Scan# 3115
Delta R.T. -0.012 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:252 Resp: 734695

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 65.1 | 51.9 | 77.9 |
| 126 | 10.7 | 8.1 | 12.1 |






#83
Chrysene
Concen: 20.977 ng
RT: 21.395 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:228 Resp: 264216

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 30.2 | 24.4 | 36.6 |
| 229 | 19.8 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 19.886 ng
RT: 21.330 min Scan# 3127
Delta R.T. 0.000 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:149 Resp: 1396109

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 27.1 | 21.9 | 32.9 |
| 279 | 4.4 | 3.5 | 5.3 |






#85
Di-n-octyl phthalate
Concen: 18.935 ng
RT: 22.548 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:149 Resp: 189291

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.6 | 1.3 | 1.9 |
| 43 | 9.2 | 6.9 | 10.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 24.483 min Scan# 3663
Delta R.T. -0.012 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:264 Resp: 2243019

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.6 | 18.6 | 28.0 |
| 265 | 21.1 | 17.7 | 26.5 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 20.677 ng
RT: 28.018 min Scan# 4
Delta R.T. -0.018 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:276 Resp: 295534

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 14.7 | 13.0 | 19.6 |
| 277 | 18.4 | 15.9 | 23.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#88
Benzo(b)fluoranthene
Concen: 20.872 ng
RT: 23.495 min Scan# 3495
Delta R.T. -0.018 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:252 Resp: 2724905

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.0 | 17.7 | 26.5 |
| 125 | 10.5 | 8.1 | 12.1 |






#89
Benzo(k)fluoranthene
Concen: 21.560 ng
RT: 23.560 min Scan# 3
Delta R.T. -0.024 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:252 Resp: 280066

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.6 | 17.7 | 26.5 |
| 125 | 9.9 | 8.2 | 12.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#90
Benzo(a)pyrene
Concen: 20.892 ng
RT: 24.330 min Scan# 3637
Delta R.T. -0.024 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:252 Resp: 2321462

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.6 | 17.6 | 26.4 |
| 125 | 11.2 | 9.0 | 13.4 |






#91
Dibenzo(a,h)anthracene
Concen: 20.842 ng
RT: 28.101 min Scan# 4
Delta R.T. -0.035 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:278 Resp: 246174

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 16.0 | 13.1 | 19.7 |
| 279 | 24.1 | 18.9 | 28.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC020

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#92
Benzo(g,h,i)perylene
Concen: 20.695 ng
RT: 29.095 min Scan# 4447
Delta R.T. -0.029 min
Lab File: BP025074.D
Acq: 03 Jul 2025 11:44

Tgt Ion:276 Resp: 2525145

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.9 | 19.3 | 28.9 |
| 138 | 22.0 | 16.9 | 25.3 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025075.D
Acq On : 03 Jul 2025 12:25
Operator : CG/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:53:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.443 | 152 | 390810 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.184 | 136 | 1560152 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.089 | 164 | 1003103 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 16.907 | 188 | 2062708 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.342 | 240 | 2262594 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 24.495 | 264 | 2431657 | 20.000 | ng | 0.00 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.096 | 112 | 1772070 | 80.342 | ng | 0.00 |
| 7) Phenol-d6 | 6.631 | 99 | 2279753 | 79.204 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.566 | 82 | 1914639 | 81.616 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.619 | 330 | 987112 | 83.301 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 12.695 | 172 | 5132522 | 79.265 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.672 | 244 | 8276493 | 80.394 | ng | 0.00 |
| Target Compounds | | | | | | Qvalue |
| 2) 1,4-Dioxane | 3.125 | 88 | 339998 | 39.642 | ng | 100 |
| 3) Pyridine | 3.490 | 79 | 959760 | 40.903 | ng | 100 |
| 4) n-Nitrosodimethylamine | 3.408 | 42 | 395498 | 39.780 | ng | 100 |
| 6) Aniline | 6.784 | 93 | 1084746 | 40.335 | ng | 100 |
| 8) 2-Chlorophenol | 7.019 | 128 | 981576 | 39.775 | ng | 100 |
| 9) Benzaldehyde | 6.602 | 77 | 565720 | 37.634 | ng | 100 |
| 10) Phenol | 6.655 | 94 | 1150792 | 39.348 | ng | 100 |
| 11) bis(2-Chloroethyl)ether | 6.890 | 93 | 876880 | 38.958 | ng | 100 |
| 12) 1,3-Dichlorobenzene | 7.337 | 146 | 1077716 | 39.020 | ng | 100 |
| 13) 1,4-Dichlorobenzene | 7.472 | 146 | 1093628 | 39.109 | ng | 100 |
| 14) 1,2-Dichlorobenzene | 7.784 | 146 | 1054645 | 39.206 | ng | 100 |
| 15) Benzyl Alcohol | 7.672 | 79 | 748226 | 38.947 | ng | 100 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.966 | 45 | 1120385 | 38.282 | ng | 100 |
| 17) 2-Methylphenol | 7.872 | 107 | 757253 | 40.195 | ng | 100 |
| 18) Hexachloroethane | 8.496 | 117 | 359924 | 39.510 | ng | 100 |
| 19) n-Nitroso-di-n-propyla... | 8.237 | 70 | 676070 | 39.605 | ng | 100 |
| 20) 3+4-Methylphenols | 8.196 | 107 | 1042619 | 39.511 | ng | 100 |
| 22) Acetophenone | 8.243 | 105 | 1451215 | 39.332 | ng | 100 |
| 24) Nitrobenzene | 8.607 | 77 | 966724 | 40.457 | ng | 100 |
| 25) Isophorone | 9.131 | 82 | 1742244 | 39.740 | ng | 100 |
| 26) 2-Nitrophenol | 9.313 | 139 | 416304 | 36.734 | ng | 100 |
| 27) 2,4-Dimethylphenol | 9.384 | 122 | 649221 | 39.698 | ng | 100 |
| 28) bis(2-Chloroethoxy)met... | 9.619 | 93 | 1166063 | 39.172 | ng | 100 |
| 29) 2,4-Dichlorophenol | 9.837 | 162 | 932551 | 41.124 | ng | 100 |
| 30) 1,2,4-Trichlorobenzene | 10.060 | 180 | 995067 | 39.625 | ng | 100 |
| 31) Naphthalene | 10.231 | 128 | 3155257 | 39.276 | ng | 100 |
| 32) Benzoic acid | 9.501 | 122 | 581334 | 38.533 | ng | 100 |
| 33) 4-Chloroaniline | 10.343 | 127 | 1217057 | 40.595 | ng | 100 |
| 34) Hexachlorobutadiene | 10.537 | 225 | 567921 | 39.409 | ng | 100 |
| 35) Caprolactam | 11.107 | 113 | 313947 | 40.469 | ng | 100 |
| 36) 4-Chloro-3-methylphenol | 11.484 | 107 | 965263 | 40.733 | ng | 100 |
| 37) 2-Methylnaphthalene | 11.860 | 142 | 2240648 | 39.530 | ng | 100 |
| 38) 1-Methylnaphthalene | 12.084 | 142 | 2158945 | 39.054 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 12.243 | 216 | 1164588 | 40.396 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 12.237 | 237 | 298913 | 39.695 | ng | 100 |
| 43) 2,4,6-Trichlorophenol | 12.484 | 196 | 741839 | 42.497 | ng | 100 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025075.D
Acq On : 03 Jul 2025 12:25
Operator : CG/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:53:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 12.554 | 196 | 824324 | 41.668 | ng | 100 |
| 46) 1,1'-Biphenyl | 12.907 | 154 | 2916222 | 39.957 | ng | 100 |
| 47) 2-Chloronaphthalene | 12.942 | 162 | 2188234 | 39.891 | ng | 100 |
| 48) 2-Nitroaniline | 13.142 | 65 | 519267 | 39.182 | ng | 100 |
| 49) Acenaphthylene | 13.801 | 152 | 3413485 | 40.657 | ng | 100 |
| 50) Dimethylphthalate | 13.554 | 163 | 2825141 | 40.234 | ng | 100 |
| 51) 2,6-Dinitrotoluene | 13.660 | 165 | 592325 | 43.253 | ng | 100 |
| 52) Acenaphthene | 14.154 | 154 | 2139419 | 39.537 | ng | 100 |
| 53) 3-Nitroaniline | 13.984 | 138 | 644835 | 43.937 | ng | 100 |
| 54) 2,4-Dinitrophenol | 14.195 | 184 | 234827 | 38.110 | ng | 100 |
| 55) Dibenzofuran | 14.507 | 168 | 3547100 | 39.719 | ng | 100 |
| 56) 4-Nitrophenol | 14.301 | 139 | 572405 | 42.263 | ng | 100 |
| 57) 2,4-Dinitrotoluene | 14.466 | 165 | 785486 | 39.572 | ng | 100 |
| 58) Fluorene | 15.166 | 166 | 2794372 | 40.454 | ng | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 14.731 | 232 | 742726 | 41.458 | ng | 100 |
| 60) Diethylphthalate | 14.960 | 149 | 2866041 | 40.639 | ng | 100 |
| 61) 4-Chlorophenyl-phenyle... | 15.178 | 204 | 1361531 | 39.829 | ng | 100 |
| 62) 4-Nitroaniline | 15.184 | 138 | 686633 | 44.618 | ng | 100 |
| 63) Azobenzene | 15.478 | 77 | 2537142 | 40.400 | ng | 100 |
| 65) 4,6-Dinitro-2-methylph... | 15.248 | 198 | 335740 | 37.190 | ng | 100 |
| 66) n-Nitrosodiphenylamine | 15.389 | 169 | 2365317 | 40.069 | ng | 100 |
| 67) 4-Bromophenyl-phenylether | 16.095 | 248 | 843967 | 39.961 | ng | 100 |
| 68) Hexachlorobenzene | 16.201 | 284 | 1020715 | 39.541 | ng | 100 |
| 69) Atrazine | 16.383 | 200 | 711320 | 41.121 | ng | 100 |
| 70) Pentachlorophenol | 16.548 | 266 | 707874 | 40.604 | ng | 100 |
| 71) Phenanthrene | 16.954 | 178 | 4407982 | 39.993 | ng | 100 |
| 72) Anthracene | 17.048 | 178 | 4275492 | 40.434 | ng | 100 |
| 73) Carbazole | 17.325 | 167 | 4245502 | 41.064 | ng | 100 |
| 74) Di-n-butylphthalate | 17.960 | 149 | 4897263 | 42.536 | ng | 100 |
| 75) Fluoranthene | 19.066 | 202 | 5320900 | 41.300 | ng | 100 |
| 77) Benzidine | 19.266 | 184 | 2584906 | 64.686 | ng | 100 |
| 78) Pyrene | 19.436 | 202 | 5572652 | 39.765 | ng | 100 |
| 80) Butylbenzylphthalate | 20.430 | 149 | 1993151 | 37.765 | ng | 100 |
| 81) Benzo(a)anthracene | 21.324 | 228 | 5507840 | 40.169 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 21.271 | 252 | 1763991 | 42.735 | ng | 100 |
| 83) Chrysene | 21.389 | 228 | 5244045 | 39.742 | ng | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 21.330 | 149 | 3179611 | 43.232 | ng | 100 |
| 85) Di-n-octyl phthalate | 22.560 | 149 | 4951836 | 38.152 | ng | 100 |
| 87) Indeno(1,2,3-cd)pyrene | 28.036 | 276 | 6474467m | 41.785 | ng | |
| 88) Benzo(b)fluoranthene | 23.512 | 252 | 5839482 | 41.258 | ng | 100 |
| 89) Benzo(k)fluoranthene | 23.583 | 252 | 5609534 | 39.833 | ng | 100 |
| 90) Benzo(a)pyrene | 24.354 | 252 | 5022829 | 41.697 | ng | 100 |
| 91) Dibenzo(a,h)anthracene | 28.136 | 278 | 5336147 | 41.673 | ng | 100 |
| 92) Benzo(g,h,i)perylene | 29.124 | 276 | 5482072 | 41.443 | ng | 100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025075.D
Acq On : 03 Jul 2025 12:25
Operator : CG/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jul 03 15:53:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.443 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:152 Resp: 39081

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 157.6 | 126.1 | 189.1 |
| 115 | 55.4 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#2
1,4-Dioxane
Concen: 39.642 ng
RT: 3.125 min Scan# 32
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 88 Resp: 339998

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 68.7 | 55.0 | 82.4 |
| 43 | 27.4 | 21.9 | 32.9 |






#3
Pyridine
Concen: 40.903 ng
RT: 3.490 min Scan# 94
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 79 Resp: 95976

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 65.3 | 52.2 | 78.4 |
| 51 | 33.5 | 26.8 | 40.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#4
n-Nitrosodimethylamine
Concen: 39.780 ng
RT: 3.408 min Scan# 80
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 42 Resp: 395498

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 131.5 | 105.2 | 157.8 |
| 44 | 8.2 | 6.6 | 9.8 |






#5
2-Fluorophenol
Concen: 80.342 ng
RT: 5.096 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:112 Resp: 177207

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 52.5 | 42.0 | 63.0 |
| 63 | 27.7 | 22.2 | 33.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#6
Aniline
Concen: 40.335 ng
RT: 6.784 min Scan# 654
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 93 Resp: 1084746

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 36.1 | 28.9 | 43.3 |
| 65 | 19.0 | 15.2 | 22.8 |






#7
Phenol-d6
Concen: 79.204 ng
RT: 6.631 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 99 Resp: 227975

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 17.0 | 13.6 | 20.4 |
| 71 | 30.5 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#8
2-Chlorophenol
Concen: 39.775 ng
RT: 7.019 min Scan# 694
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:128 Resp: 981576

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.9 | 12.9 | 52.9 |
| 64 | 40.3 | 20.3 | 60.3 |






#9
Benzaldehyde
Concen: 37.634 ng
RT: 6.602 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 77 Resp: 56572

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 102.4 | 82.4 | 122.4 |
| 106 | 98.3 | 78.3 | 118.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#10
Phenol
Concen: 39.348 ng
RT: 6.655 min Scan# 632
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 94 Resp: 1150792

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 26.9 | 6.9 | 46.9 |
| 66 | 34.6 | 14.6 | 54.6 |






#11
bis(2-Chloroethyl)ether
Concen: 38.958 ng
RT: 6.890 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 93 Resp: 87688

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 71.9 | 51.9 | 91.9 |
| 95 | 32.4 | 12.4 | 52.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#12
1,3-Dichlorobenzene
Concen: 39.020 ng
RT: 7.337 min Scan# 748
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:146 Resp: 1077716

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.9 | 51.9 | 77.9 |
| 75 | 27.2 | 21.8 | 32.6 |






#13
1,4-Dichlorobenzene
Concen: 39.109 ng
RT: 7.472 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:146 Resp: 109362

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.9 | 51.1 | 76.7 |
| 111 | 39.0 | 31.2 | 46.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#14
1,2-Dichlorobenzene
Concen: 39.206 ng
RT: 7.784 min Scan# 824
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:146 Resp: 1054645

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.9 | 51.1 | 76.7 |
| 111 | 39.0 | 31.2 | 46.8 |






#15
Benzyl Alcohol
Concen: 38.947 ng
RT: 7.672 min Scan# 8(
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 79 Resp: 74822(

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 82.5 | 66.0 | 99.0 |
| 77 | 66.8 | 53.4 | 80.2 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.282 ng
RT: 7.966 min Scan# 855
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 45 Resp: 1120385

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 14.8 | 0.0 | 34.8 |
| 79 | 10.7 | 0.0 | 30.7 |






#17
2-Methylphenol
Concen: 40.195 ng
RT: 7.872 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:107 Resp: 75725

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 112.1 | 89.7 | 134.5 |
| 77 | 42.8 | 34.2 | 51.4 |
| 79 | 41.8 | 33.4 | 50.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#18
Hexachloroethane
Concen: 39.510 ng
RT: 8.496 min Scan# 945
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:117 Resp: 359924

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 95.8 | 76.6 | 115.0 |
| 201 | 106.8 | 85.4 | 128.2 |






#19
n-Nitroso-di-n-propylamine
Concen: 39.605 ng
RT: 8.237 min Scan# 9(
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 70 Resp: 67607(

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 58.1 | 46.5 | 69.7 |
| 101 | 9.9 | 7.9 | 11.9 |
| 130 | 24.1 | 19.3 | 28.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#20
3+4-Methylphenols
Concen: 39.511 ng
RT: 8.196 min Scan# 894
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:107 Resp: 1042619

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 89.5 | 69.5 | 109.5 |
| 77 | 31.4 | 11.4 | 51.4 |
| 79 | 26.9 | 6.9 | 46.9 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.184 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:136 Resp: 156015

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.9 | 8.7 | 13.1 |
| 54 | 6.8 | 5.4 | 8.2 |
| 68 | 5.6 | 4.5 | 6.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#22
Acetophenone
Concen: 39.332 ng
RT: 8.243 min Scan# 902
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:105 Resp: 1451215

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 6.8 | 5.4 | 8.2 |
| 51 | 24.5 | 19.6 | 29.4 |
| 120 | 23.5 | 18.8 | 28.2 |






#23
Nitrobenzene-d5
Concen: 81.616 ng
RT: 8.566 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 82 Resp: 191463

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 48.0 | 38.4 | 57.6 |
| 54 | 47.2 | 37.8 | 56.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#24
Nitrobenzene
Concen: 40.457 ng
RT: 8.607 min Scan# 964
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 77 Resp: 966724

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 48.8 | 39.0 | 58.6 |
| 65 | 13.4 | 10.7 | 16.1 |






#25
Isophorone
Concen: 39.740 ng
RT: 9.131 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 82 Resp: 174224

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.2 | 5.8 | 8.6 |
| 138 | 19.6 | 15.7 | 23.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#26
2-Nitrophenol
Concen: 36.734 ng
RT: 9.313 min Scan# 1084
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:139 Resp: 416304

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 27.9 | 22.3 | 33.5 |
| 65 | 42.3 | 33.8 | 50.8 |






#27
2,4-Dimethylphenol
Concen: 39.698 ng
RT: 9.384 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:122 Resp: 64922

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 110.2 | 88.2 | 132.2 |
| 121 | 57.6 | 46.1 | 69.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#28
bis(2-Chloroethoxy)methane
Concen: 39.172 ng
RT: 9.619 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 93 Resp: 1166063

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 33.0 | 26.4 | 39.6 |
| 123 | 12.6 | 10.1 | 15.1 |






#29
2,4-Dichlorophenol
Concen: 41.124 ng
RT: 9.837 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:162 Resp: 93255

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 64.1 | 44.1 | 84.1 |
| 98 | 33.1 | 13.1 | 53.1 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#30
1,2,4-Trichlorobenzene
Concen: 39.625 ng
RT: 10.060 min Scan# 1211
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:180 Resp: 995067

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 93.9 | 75.1 | 112.7 |
| 145 | 29.1 | 23.3 | 34.9 |






#31
Naphthalene
Concen: 39.276 ng
RT: 10.231 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:128 Resp: 315525

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.1 | 8.9 | 13.3 |
| 127 | 13.0 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#32
Benzoic acid
Concen: 38.533 ng
RT: 9.501 min Scan# 1116
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:122 Resp: 581334

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 121.0 | 101.0 | 141.0 |
| 77 | 87.6 | 67.6 | 107.6 |






#33
4-Chloroaniline
Concen: 40.595 ng
RT: 10.343 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:127 Resp: 121705

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 31.8 | 25.4 | 38.2 |
| 65 | 26.3 | 21.0 | 31.6 |
| 92 | 16.7 | 13.4 | 20.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#34
Hexachlorobutadiene
Concen: 39.409 ng
RT: 10.537 min Scan# 1292
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:225 Resp: 567921

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 62.9 | 50.3 | 75.5 |
| 227 | 64.1 | 51.3 | 76.9 |






#35
Caprolactam
Concen: 40.469 ng
RT: 11.107 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:113 Resp: 31394

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 148.1 | 128.1 | 168.1 |
| 56 | 110.7 | 90.7 | 130.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#36
4-Chloro-3-methylphenol
Concen: 40.733 ng
RT: 11.484 min Scan# 1453
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:107 Resp: 965263

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 28.0 | 22.4 | 33.6 |
| 142 | 87.0 | 69.6 | 104.4 |






#37
2-Methylnaphthalene
Concen: 39.530 ng
RT: 11.860 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:142 Resp: 224064

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 89.0 | 71.2 | 106.8 |
| 115 | 32.4 | 25.9 | 38.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#38
1-Methylnaphthalene
Concen: 39.054 ng
RT: 12.084 min Scan# 1555
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:142 Resp: 2158945

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 92.6 | 74.1 | 111.1 |
| 116 | 3.4 | 2.7 | 4.1 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.089 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:164 Resp: 100310

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 100.4 | 80.3 | 120.5 |
| 160 | 44.9 | 35.9 | 53.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.396 ng
RT: 12.243 min Scan# 1582
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:216 Resp: 1164588

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.0 | 62.4 | 93.6 |
| 179 | 20.6 | 16.5 | 24.7 |
| 108 | 17.1 | 13.7 | 20.5 |






#41
Hexachlorocyclopentadiene
Concen: 39.695 ng
RT: 12.237 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

Tgt Ion:237 Resp: 29891

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 61.6 | 41.6 | 81.6 |
| 272 | 13.6 | 0.0 | 33.6 |

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#42
2,4,6-Tribromophenol
Concen: 83.301 ng
RT: 15.619 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:330 Resp: 987112

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 97.1 | 77.7 | 116.5 |
| 141 | 32.5 | 26.0 | 39.0 |






#43
2,4,6-Trichlorophenol
Concen: 42.497 ng
RT: 12.484 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:196 Resp: 74183

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.2 | 76.2 | 114.2 |
| 200 | 30.4 | 24.3 | 36.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#44
2,4,5-Trichlorophenol
Concen: 41.668 ng
RT: 12.554 min Scan# 1635
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:196 Resp: 824324

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.8 | 76.6 | 115.0 |
| 97 | 45.8 | 36.6 | 55.0 |
| 132 | 26.7 | 21.4 | 32.0 |






#45
2-Fluorobiphenyl
Concen: 79.265 ng
RT: 12.695 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:172 Resp: 513252

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.2 | 28.2 | 42.2 |
| 170 | 23.4 | 18.7 | 28.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#46
1,1'-Biphenyl
Concen: 39.957 ng
RT: 12.907 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:154 Resp: 2916222

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.2 | 20.2 | 60.2 |
| 76 | 11.5 | 0.0 | 31.5 |






#47
2-Chloronaphthalene
Concen: 39.891 ng
RT: 12.942 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:162 Resp: 218823

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 36.8 | 29.4 | 44.2 |
| 164 | 33.0 | 26.4 | 39.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#48
2-Nitroaniline
Concen: 39.182 ng
RT: 13.142 min Scan# 1735
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 65 Resp: 519267

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 73.4 | 58.7 | 88.1 |
| 138 | 123.7 | 99.0 | 148.4 |






#49
Acenaphthylene
Concen: 40.657 ng
RT: 13.801 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:152 Resp: 341348

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.1 | 16.1 | 24.1 |
| 153 | 13.4 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#50
Dimethylphthalate
Concen: 40.234 ng
RT: 13.554 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:163 Resp: 2825141

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.2 | 3.4 | 5.0 |
| 164 | 10.0 | 8.0 | 12.0 |






#51
2,6-Dinitrotoluene
Concen: 43.253 ng
RT: 13.660 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:165 Resp: 59232

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 48.1 | 38.5 | 57.7 |
| 89 | 50.8 | 40.6 | 61.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#52
Acenaphthene
Concen: 39.537 ng
RT: 14.154 min Scan# 1907
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:154 Resp: 2139419

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 113.3 | 90.6 | 136.0 |
| 152 | 54.8 | 43.8 | 65.8 |






#53
3-Nitroaniline
Concen: 43.937 ng
RT: 13.984 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:138 Resp: 64483

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 9.8 | 7.8 | 11.8 |
| 92 | 101.2 | 81.0 | 121.4 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#54
2,4-Dinitrophenol
Concen: 38.110 ng
RT: 14.195 min Scan# 1914
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:184 Resp: 234827

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 59.2 | 47.4 | 71.0 |
| 154 | 59.8 | 47.8 | 71.8 |






#55
Dibenzofuran
Concen: 39.719 ng
RT: 14.507 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:168 Resp: 354710

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 35.8 | 28.6 | 43.0 |
| 169 | 13.4 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#56
4-Nitrophenol
Concen: 42.263 ng
RT: 14.301 min Scan# 1932
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:139 Resp: 572405

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 65.9 | 45.9 | 85.9 |
| 65 | 81.4 | 61.4 | 101.4 |






#57
2,4-Dinitrotoluene
Concen: 39.572 ng
RT: 14.466 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:165 Resp: 78548

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 38.7 | 31.0 | 46.4 |
| 89 | 66.7 | 53.4 | 80.0 |
| 182 | 2.5 | 2.0 | 3.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#58
Fluorene
Concen: 40.454 ng
RT: 15.166 min Scan# 2079
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:166 Resp: 2794372

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 98.2 | 78.6 | 117.8 |
| 167 | 13.4 | 10.7 | 16.1 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 41.458 ng
RT: 14.731 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:232 Resp: 74272

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 44.6 | 35.7 | 53.5 |
| 130 | 2.2 | 1.8 | 2.6 |
| 166 | 30.5 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#60
Diethylphthalate
Concen: 40.639 ng
RT: 14.960 min Scan# 2044
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:149 Resp: 2866041

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 22.3 | 17.8 | 26.8 |
| 150 | 12.7 | 10.2 | 15.2 |






#61
4-Chlorophenyl-phenylether
Concen: 39.829 ng
RT: 15.178 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:204 Resp: 136153

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 32.2 | 25.8 | 38.6 |
| 141 | 56.4 | 45.1 | 67.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#62
4-Nitroaniline
Concen: 44.618 ng
RT: 15.184 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:138 Resp: 686633

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 46.0 | 26.0 | 66.0 |
| 108 | 70.0 | 50.0 | 90.0 |






#63
Azobenzene
Concen: 40.400 ng
RT: 15.478 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion: 77 Resp: 253714

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 27.7 | 7.7 | 47.7 |
| 105 | 21.9 | 1.9 | 41.9 |
| 51 | 28.2 | 8.2 | 48.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.907 min Scan# 2375
Delta R.T. -0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:188 Resp: 2062708

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 9.1 | 7.3 | 10.9 |
| 80 | 9.7 | 7.8 | 11.6 |






#65
4,6-Dinitro-2-methylphenol
Concen: 37.190 ng
RT: 15.248 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:198 Resp: 33574

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 36.8 | 16.8 | 56.8 |
| 105 | 40.8 | 20.8 | 60.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#66
n-Nitrosodiphenylamine
Concen: 40.069 ng
RT: 15.389 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:169 Resp: 2365317

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 67.6 | 54.1 | 81.1 |
| 167 | 36.3 | 29.0 | 43.6 |






#67
4-Bromophenyl-phenylether
Concen: 39.961 ng
RT: 16.095 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:248 Resp: 84396

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 99.3 | 79.4 | 119.2 |
| 141 | 64.2 | 51.4 | 77.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#68
Hexachlorobenzene
Concen: 39.541 ng
RT: 16.201 min Scan# 2255
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:284 Resp: 1020715

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 32.4 | 25.9 | 38.9 |
| 249 | 29.8 | 23.8 | 35.8 |






#69
Atrazine
Concen: 41.121 ng
RT: 16.383 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:200 Resp: 71132

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 26.6 | 6.6 | 46.6 |
| 215 | 48.0 | 28.0 | 68.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#70
Pentachlorophenol
Concen: 40.604 ng
RT: 16.548 min Scan# 2314
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:266 Resp: 707874

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 63.0 | 50.4 | 75.6 |
| 264 | 62.7 | 50.2 | 75.2 |






#71
Phenanthrene
Concen: 39.993 ng
RT: 16.954 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:178 Resp: 440798

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.3 | 15.4 | 23.2 |
| 179 | 15.8 | 12.6 | 19.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#72
Anthracene
Concen: 40.434 ng
RT: 17.048 min Scan# 2399
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:178 Resp: 4275492

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.8 | 15.0 | 22.6 |
| 179 | 15.6 | 12.5 | 18.7 |






#73
Carbazole
Concen: 41.064 ng
RT: 17.325 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:167 Resp: 424550

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.7 | 17.4 | 26.0 |
| 139 | 13.0 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#74
Di-n-butylphthalate
Concen: 42.536 ng
RT: 17.960 min Scan# 2554
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:149 Resp: 4897263

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.1 | 7.3 | 10.9 |
| 104 | 4.7 | 3.8 | 5.6 |






#75
Fluoranthene
Concen: 41.300 ng
RT: 19.066 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:202 Resp: 532090

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 10.4 | 0.0 | 30.4 |
| 203 | 17.6 | 0.0 | 37.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.342 min Scan# 3129
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:240 Resp: 2262594

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.2 | 9.0 | 13.4 |
| 236 | 25.4 | 20.3 | 30.5 |






#77
Benzidine
Concen: 64.686 ng
RT: 19.266 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:184 Resp: 258490

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 15.7 | 12.6 | 18.8 |
| 183 | 11.8 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#78
Pyrene
Concen: 39.765 ng
RT: 19.436 min Scan# 2805
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:202 Resp: 5572652

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 21.4 | 17.1 | 25.7 |
| 203 | 17.6 | 14.1 | 21.1 |






#79
Terphenyl-d14
Concen: 80.394 ng
RT: 19.672 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:244 Resp: 827649

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 7.4 | 5.9 | 8.9 |
| 122 | 11.8 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#80
Butylbenzylphthalate
Concen: 37.765 ng
RT: 20.430 min Scan# 2974
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:149 Resp: 1993151

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 63.3 | 50.6 | 76.0 |
| 206 | 21.5 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 40.169 ng
RT: 21.324 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:228 Resp: 550784

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 27.6 | 22.1 | 33.1 |
| 229 | 20.0 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#82
3,3'-Dichlorobenzidine
Concen: 42.735 ng
RT: 21.271 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:252 Resp: 1763991

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 64.9 | 51.9 | 77.9 |
| 126 | 10.1 | 8.1 | 12.1 |






#83
Chrysene
Concen: 39.742 ng
RT: 21.389 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:228 Resp: 524404

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 30.5 | 24.4 | 36.6 |
| 229 | 20.0 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 43.232 ng
RT: 21.330 min Scan# 3127
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:149 Resp: 3179611

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 27.4 | 21.9 | 32.9 |
| 279 | 4.4 | 3.5 | 5.3 |






#85
Di-n-octyl phthalate
Concen: 38.152 ng
RT: 22.560 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:149 Resp: 495183

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.6 | 1.3 | 1.9 |
| 43 | 8.6 | 6.9 | 10.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 24.495 min Scan# 3665
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:264 Resp: 2431657

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.3 | 18.6 | 28.0 |
| 265 | 22.1 | 17.7 | 26.5 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 41.785 ng m
RT: 28.036 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:276 Resp: 647446

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 16.3 | 13.0 | 19.6 |
| 277 | 19.9 | 15.9 | 23.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#88
Benzo(b)fluoranthene
Concen: 41.258 ng
RT: 23.512 min Scan# 3498
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:252 Resp: 5839482

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.1 | 17.7 | 26.5 |
| 125 | 10.1 | 8.1 | 12.1 |






#89
Benzo(k)fluoranthene
Concen: 39.833 ng
RT: 23.583 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:252 Resp: 560953

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.1 | 17.7 | 26.5 |
| 125 | 10.2 | 8.2 | 12.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#90
Benzo(a)pyrene
Concen: 41.697 ng
RT: 24.354 min Scan# 3641
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:252 Resp: 5022829

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.0 | 17.6 | 26.4 |
| 125 | 11.2 | 9.0 | 13.4 |






#91
Dibenzo(a,h)anthracene
Concen: 41.673 ng
RT: 28.136 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:278 Resp: 533614

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 16.4 | 13.1 | 19.7 |
| 279 | 23.6 | 18.9 | 28.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#92
Benzo(g,h,i)perylene
Concen: 41.443 ng
RT: 29.124 min Scan# 4452
Delta R.T. 0.000 min
Lab File: BP025075.D
Acq: 03 Jul 2025 12:25

Tgt Ion:276 Resp: 5482072

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 24.1 | 19.3 | 28.9 |
| 138 | 21.1 | 16.9 | 25.3 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025076.D
Acq On : 03 Jul 2025 13:07
Operator : CG/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:55:08 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.443 | 152 | 426915 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.195 | 136 | 1726448 | 20.000 | ng | 0.01 |
| 39) Acenaphthene-d10 | 14.095 | 164 | 1117104 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 16.907 | 188 | 2198784 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.354 | 240 | 2246809 | 20.000 | ng | 0.01 |
| 86) Perylene-d12 | 24.518 | 264 | 2460402 | 20.000 | ng | 0.02 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.102 | 112 | 2404869 | 99.811 | ng | 0.00 |
| 7) Phenol-d6 | 6.631 | 99 | 3142540 | 99.945 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.572 | 82 | 2657612 | 102.375 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.625 | 330 | 1362624 | 103.255 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 12.701 | 172 | 7037498 | 97.594 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.683 | 244 | 10280025 | 100.557 | ng | 0.01 |
| Target Compounds | | | | | | Qvalue |
| 2) 1,4-Dioxane | 3.125 | 88 | 448307 | 47.849 | ng | 100 |
| 3) Pyridine | 3.490 | 79 | 1249873m | 48.762 | ng | |
| 4) n-Nitrosodimethylamine | 3.408 | 42 | 526435 | 48.471 | ng | 98 |
| 6) Aniline | 6.790 | 93 | 1463015 | 49.800 | ng | 99 |
| 8) 2-Chlorophenol | 7.019 | 128 | 1334192 | 49.491 | ng | 99 |
| 9) Benzaldehyde | 6.602 | 77 | 730343 | 44.476 | ng | 98 |
| 10) Phenol | 6.660 | 94 | 1584515 | 49.596 | ng | 98 |
| 11) bis(2-Chloroethyl)ether | 6.884 | 93 | 1204826 | 49.001 | ng | 100 |
| 12) 1,3-Dichlorobenzene | 7.331 | 146 | 1457893 | 48.321 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 7.478 | 146 | 1476720 | 48.343 | ng | 98 |
| 14) 1,2-Dichlorobenzene | 7.790 | 146 | 1418407 | 48.270 | ng | 100 |
| 15) Benzyl Alcohol | 7.672 | 79 | 1056771 | 50.356 | ng | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.966 | 45 | 1560041 | 48.796 | ng | 100 |
| 17) 2-Methylphenol | 7.878 | 107 | 1039628 | 50.517 | ng | 99 |
| 18) Hexachloroethane | 8.501 | 117 | 494292 | 49.671 | ng | 98 |
| 19) n-Nitroso-di-n-propyla... | 8.243 | 70 | 930000 | 49.874 | ng | 99 |
| 20) 3+4-Methylphenols | 8.201 | 107 | 1431288 | 49.653 | ng | 99 |
| 22) Acetophenone | 8.243 | 105 | 1967270 | 48.182 | ng | 99 |
| 24) Nitrobenzene | 8.607 | 77 | 1323155 | 50.040 | ng | 96 |
| 25) Isophorone | 9.137 | 82 | 2464795 | 50.806 | ng | 100 |
| 26) 2-Nitrophenol | 9.313 | 139 | 629107 | 48.958 | ng | 99 |
| 27) 2,4-Dimethylphenol | 9.384 | 122 | 918599 | 50.760 | ng | 98 |
| 28) bis(2-Chloroethoxy)met... | 9.625 | 93 | 1636120 | 49.669 | ng | 99 |
| 29) 2,4-Dichlorophenol | 9.843 | 162 | 1314014 | 52.364 | ng | 99 |
| 30) 1,2,4-Trichlorobenzene | 10.054 | 180 | 1365048 | 49.122 | ng | 98 |
| 31) Naphthalene | 10.243 | 128 | 4337583 | 48.793 | ng | 100 |
| 32) Benzoic acid | 9.531 | 122 | 859495 | 49.693 | ng | 97 |
| 33) 4-Chloroaniline | 10.343 | 127 | 1645871 | 49.611 | ng | 99 |
| 34) Hexachlorobutadiene | 10.543 | 225 | 793648 | 49.768 | ng | 99 |
| 35) Caprolactam | 11.107 | 113 | 422268 | 49.189 | ng | 99 |
| 36) 4-Chloro-3-methylphenol | 11.484 | 107 | 1335155 | 50.915 | ng | 99 |
| 37) 2-Methylnaphthalene | 11.860 | 142 | 3084171 | 49.170 | ng | 100 |
| 38) 1-Methylnaphthalene | 12.089 | 142 | 2983697 | 48.774 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 12.242 | 216 | 1612893 | 50.237 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 12.231 | 237 | 452638 | 53.975 | ng | 100 |
| 43) 2,4,6-Trichlorophenol | 12.489 | 196 | 1026903 | 52.823 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025076.D
Acq On : 03 Jul 2025 13:07
Operator : CG/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:55:08 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 12.554 | 196 | 1151711 | 52.275 | ng | 99 |
| 46) 1,1'-Biphenyl | 12.907 | 154 | 4006544 | 49.294 | ng | 100 |
| 47) 2-Chloronaphthalene | 12.942 | 162 | 2986246 | 48.883 | ng | 100 |
| 48) 2-Nitroaniline | 13.148 | 65 | 722237 | 48.247 | ng | 97 |
| 49) Acenaphthylene | 13.813 | 152 | 4634923 | 49.571 | ng | 100 |
| 50) Dimethylphthalate | 13.560 | 163 | 3810769 | 48.733 | ng | 99 |
| 51) 2,6-Dinitrotoluene | 13.660 | 165 | 811379 | 53.202 | ng | 98 |
| 52) Acenaphthene | 14.160 | 154 | 2923605 | 48.515 | ng | 99 |
| 53) 3-Nitroaniline | 13.989 | 138 | 865670 | 52.965 | ng | 96 |
| 54) 2,4-Dinitrophenol | 14.207 | 184 | 335631 | 47.384 | ng | 96 |
| 55) Dibenzofuran | 14.507 | 168 | 4802343 | 48.287 | ng | 99 |
| 56) 4-Nitrophenol | 14.313 | 139 | 764396 | 50.679 | ng | 98 |
| 57) 2,4-Dinitrotoluene | 14.472 | 165 | 1081273 | 48.291 | ng | 97 |
| 58) Fluorene | 15.172 | 166 | 3727642 | 48.458 | ng | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 14.731 | 232 | 1014960 | 50.872 | ng | 99 |
| 60) Diethylphthalate | 14.972 | 149 | 3816042 | 48.587 | ng | 100 |
| 61) 4-Chlorophenyl-phenyle... | 15.178 | 204 | 1839971 | 48.332 | ng | 99 |
| 62) 4-Nitroaniline | 15.189 | 138 | 890284 | 51.948 | ng | 100 |
| 63) Azobenzene | 15.472 | 77 | 3381849 | 48.355 | ng | 98 |
| 65) 4,6-Dinitro-2-methylph... | 15.254 | 198 | 488912 | 48.762 | ng | 98 |
| 66) n-Nitrosodiphenylamine | 15.389 | 169 | 3156476 | 50.162 | ng | 99 |
| 67) 4-Bromophenyl-phenylether | 16.089 | 248 | 1145496 | 50.881 | ng | 98 |
| 68) Hexachlorobenzene | 16.195 | 284 | 1365447 | 49.621 | ng | 100 |
| 69) Atrazine | 16.395 | 200 | 937888 | 50.863 | ng | 99 |
| 70) Pentachlorophenol | 16.554 | 266 | 975866 | 52.511 | ng | 99 |
| 71) Phenanthrene | 16.954 | 178 | 5660551 | 48.179 | ng | 100 |
| 72) Anthracene | 17.054 | 178 | 5614698 | 49.812 | ng | 100 |
| 73) Carbazole | 17.325 | 167 | 5444196 | 49.399 | ng | 100 |
| 74) Di-n-butylphthalate | 17.966 | 149 | 6469226 | 52.713 | ng | 99 |
| 75) Fluoranthene | 19.054 | 202 | 6865989 | 49.995 | ng | 99 |
| 77) Benzidine | 19.271 | 184 | 2357157 | 59.401 | ng | 99 |
| 78) Pyrene | 19.436 | 202 | 7028090 | 50.503 | ng | 99 |
| 80) Butylbenzylphthalate | 20.442 | 149 | 2645413 | 49.237 | ng | 99 |
| 81) Benzo(a)anthracene | 21.336 | 228 | 6780561 | 49.798 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 21.271 | 252 | 2241436 | 54.684 | ng | 98 |
| 83) Chrysene | 21.395 | 228 | 6500368 | 49.609 | ng | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 21.336 | 149 | 4140331 | 56.690 | ng | 99 |
| 85) Di-n-octyl phthalate | 22.565 | 149 | 6776549 | 50.271 | ng | 100 |
| 87) Indeno(1,2,3-cd)pyrene | 28.059 | 276 | 8400646m | 53.583 | ng | |
| 88) Benzo(b)fluoranthene | 23.518 | 252 | 7308592 | 51.035 | ng | 100 |
| 89) Benzo(k)fluoranthene | 23.595 | 252 | 7114695 | 49.931 | ng | 100 |
| 90) Benzo(a)pyrene | 24.359 | 252 | 6366556 | 52.234 | ng | 100 |
| 91) Dibenzo(a,h)anthracene | 28.136 | 278 | 6926985 | 53.464 | ng | 99 |
| 92) Benzo(g,h,i)perylene | 29.159 | 276 | 7023994 | 52.480 | ng | 100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025076.D
Acq On : 03 Jul 2025 13:07
Operator : CG/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jul 03 15:55:08 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.443 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:152 Resp: 42691

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 155.7 | 126.1 | 189.1 |
| 115 | 57.8 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#2
1,4-Dioxane
Concen: 47.849 ng
RT: 3.125 min Scan# 32
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 88 Resp: 448307

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 68.8 | 55.0 | 82.4 |
| 43 | 27.3 | 21.9 | 32.9 |






#3
Pyridine
Concen: 48.762 ng m
RT: 3.490 min Scan# 94
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 79 Resp: 124987

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 66.4 | 52.2 | 78.4 |
| 51 | 34.0 | 26.8 | 40.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#4
n-Nitrosodimethylamine
Concen: 48.471 ng
RT: 3.408 min Scan# 80
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 42 Resp: 526435

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 133.4 | 105.2 | 157.8 |
| 44 | 7.2 | 6.6 | 9.8 |






#5
2-Fluorophenol
Concen: 99.811 ng
RT: 5.102 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:112 Resp: 240486

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 51.9 | 42.0 | 63.0 |
| 63 | 26.6 | 22.2 | 33.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#6
Aniline
Concen: 49.800 ng
RT: 6.790 min Scan# 655
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 93 Resp: 1463015

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 36.0 | 28.9 | 43.3 |
| 65 | 18.2 | 15.2 | 22.8 |






#7
Phenol-d6
Concen: 99.945 ng
RT: 6.631 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 99 Resp: 314254

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 17.4 | 13.6 | 20.4 |
| 71 | 31.2 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#8
2-Chlorophenol
Concen: 49.491 ng
RT: 7.019 min Scan# 694
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:128 Resp: 1334192

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 33.4 | 12.9 | 52.9 |
| 64 | 40.6 | 20.3 | 60.3 |






#9
Benzaldehyde
Concen: 44.476 ng
RT: 6.602 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 77 Resp: 73034

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 103.5 | 82.4 | 122.4 |
| 106 | 100.9 | 78.3 | 118.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#10
Phenol
Concen: 49.596 ng
RT: 6.660 min Scan# 633
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 94 Resp: 1584515

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 25.6 | 6.9 | 46.9 |
| 66 | 33.2 | 14.6 | 54.6 |






#11
bis(2-Chloroethyl)ether
Concen: 49.001 ng
RT: 6.884 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 93 Resp: 120482

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 71.9 | 51.9 | 91.9 |
| 95 | 32.6 | 12.4 | 52.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#12
1,3-Dichlorobenzene
Concen: 48.321 ng
RT: 7.331 min Scan# 747
Delta R.T. -0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:146 Resp: 1457893

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.8 | 51.9 | 77.9 |
| 75 | 27.5 | 21.8 | 32.6 |






#13
1,4-Dichlorobenzene
Concen: 48.343 ng
RT: 7.478 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:146 Resp: 147672

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.4 | 51.1 | 76.7 |
| 111 | 36.9 | 31.2 | 46.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#14
1,2-Dichlorobenzene
Concen: 48.270 ng
RT: 7.790 min Scan# 825
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:146 Resp: 1418407

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.3 | 51.1 | 76.7 |
| 111 | 39.0 | 31.2 | 46.8 |






#15
Benzyl Alcohol
Concen: 50.356 ng
RT: 7.672 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 79 Resp: 105677

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 82.7 | 66.0 | 99.0 |
| 77 | 65.3 | 53.4 | 80.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 48.796 ng
RT: 7.966 min Scan# 855
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 45 Resp: 1560041

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 14.9 | 0.0 | 34.8 |
| 79 | 10.9 | 0.0 | 30.7 |






#17
2-Methylphenol
Concen: 50.517 ng
RT: 7.878 min Scan# 84
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:107 Resp: 103962

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 111.4 | 89.7 | 134.5 |
| 77 | 43.2 | 34.2 | 51.4 |
| 79 | 41.5 | 33.4 | 50.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#18
Hexachloroethane
Concen: 49.671 ng
RT: 8.501 min Scan# 946
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:117 Resp: 494292

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 96.2 | 76.6 | 115.0 |
| 201 | 102.8 | 85.4 | 128.2 |






#19
n-Nitroso-di-n-propylamine
Concen: 49.874 ng
RT: 8.243 min Scan# 9(
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 70 Resp: 93000(

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 58.5 | 46.5 | 69.7 |
| 101 | 10.2 | 7.9 | 11.9 |
| 130 | 24.4 | 19.3 | 28.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#20
3+4-Methylphenols
Concen: 49.653 ng
RT: 8.201 min Scan# 895
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:107 Resp: 1431288

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 90.3 | 69.5 | 109.5 |
| 77 | 31.3 | 11.4 | 51.4 |
| 79 | 26.7 | 6.9 | 46.9 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.195 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:136 Resp: 172644

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.9 | 8.7 | 13.1 |
| 54 | 6.9 | 5.4 | 8.2 |
| 68 | 5.5 | 4.5 | 6.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#22
Acetophenone
Concen: 48.182 ng
RT: 8.243 min Scan# 902
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:105 Resp: 1967270

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 7.9 | 5.4 | 8.2 |
| 51 | 24.1 | 19.6 | 29.4 |
| 120 | 23.4 | 18.8 | 28.2 |






#23
Nitrobenzene-d5
Concen: 102.375 ng
RT: 8.572 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 82 Resp: 265761

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 49.6 | 38.4 | 57.6 |
| 54 | 46.9 | 37.8 | 56.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#24
Nitrobenzene
Concen: 50.040 ng
RT: 8.607 min Scan# 964
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 77 Resp: 1323155

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 52.0 | 39.0 | 58.6 |
| 65 | 13.1 | 10.7 | 16.1 |






#25
Isophorone
Concen: 50.806 ng
RT: 9.137 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 82 Resp: 246479

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.0 | 5.8 | 8.6 |
| 138 | 19.8 | 15.7 | 23.5 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#26
2-Nitrophenol
Concen: 48.958 ng
RT: 9.313 min Scan# 1084
Delta R.T. 0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:139 Resp: 629107

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 26.6 | 22.3 | 33.5 |
| 65 | 41.8 | 33.8 | 50.8 |






#27
2,4-Dimethylphenol
Concen: 50.760 ng
RT: 9.384 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:122 Resp: 91859

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 107.6 | 88.2 | 132.2 |
| 121 | 57.9 | 46.1 | 69.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#28
bis(2-Chloroethoxy)methane
Concen: 49.669 ng
RT: 9.625 min Scan# 1137
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 93 Resp: 1636120

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.7 | 26.4 | 39.6 |
| 123 | 12.2 | 10.1 | 15.1 |






#29
2,4-Dichlorophenol
Concen: 52.364 ng
RT: 9.843 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:162 Resp: 131401

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 64.8 | 44.1 | 84.1 |
| 98 | 32.0 | 13.1 | 53.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#30
1,2,4-Trichlorobenzene
Concen: 49.122 ng
RT: 10.054 min Scan# 1210
Delta R.T. -0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:180 Resp: 1365048

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 96.1 | 75.1 | 112.7 |
| 145 | 30.2 | 23.3 | 34.9 |






#31
Naphthalene
Concen: 48.793 ng
RT: 10.243 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:128 Resp: 433758

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 10.9 | 8.9 | 13.3 |
| 127 | 12.9 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#32
Benzoic acid
Concen: 49.693 ng
RT: 9.531 min Scan# 1121
Delta R.T. 0.030 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:122 Resp: 859495

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 118.2 | 101.0 | 141.0 |
| 77 | 84.7 | 67.6 | 107.6 |






#33
4-Chloroaniline
Concen: 49.611 ng
RT: 10.343 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:127 Resp: 164587

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.5 | 25.4 | 38.2 |
| 65 | 26.5 | 21.0 | 31.6 |
| 92 | 17.5 | 13.4 | 20.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#34
Hexachlorobutadiene
Concen: 49.768 ng
RT: 10.543 min Scan# 1293
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:225 Resp: 793648

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 62.0 | 50.3 | 75.5 |
| 227 | 65.0 | 51.3 | 76.9 |






#35
Caprolactam
Concen: 49.189 ng
RT: 11.107 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:113 Resp: 42226

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 149.2 | 128.1 | 168.1 |
| 56 | 111.6 | 90.7 | 130.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#36
4-Chloro-3-methylphenol
Concen: 50.915 ng
RT: 11.484 min Scan# 1453
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:107 Resp: 1335155

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 28.4 | 22.4 | 33.6 |
| 142 | 87.5 | 69.6 | 104.4 |






#37
2-Methylnaphthalene
Concen: 49.170 ng
RT: 11.860 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:142 Resp: 308417

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 88.6 | 71.2 | 106.8 |
| 115 | 32.3 | 25.9 | 38.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#38
1-Methylnaphthalene
Concen: 48.774 ng
RT: 12.089 min Scan# 1556
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:142 Resp: 2983697

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 92.4 | 74.1 | 111.1 |
| 116 | 3.5 | 2.7 | 4.1 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.095 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:164 Resp: 111710

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 101.0 | 80.3 | 120.5 |
| 160 | 44.9 | 35.9 | 53.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 50.237 ng
RT: 12.242 min Scan# 1582
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:216 Resp: 1612893

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.4 | 62.4 | 93.6 |
| 179 | 20.8 | 16.5 | 24.7 |
| 108 | 16.9 | 13.7 | 20.5 |






#41
Hexachlorocyclopentadiene
Concen: 53.975 ng
RT: 12.231 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:237 Resp: 45263

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 61.9 | 41.6 | 81.6 |
| 272 | 13.4 | 0.0 | 33.6 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#42
2,4,6-Tribromophenol
Concen: 103.255 ng
RT: 15.625 min Scan# 2157
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:330 Resp: 1362624

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 96.6 | 77.7 | 116.5 |
| 141 | 31.6 | 26.0 | 39.0 |






#43
2,4,6-Trichlorophenol
Concen: 52.823 ng
RT: 12.489 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:196 Resp: 102690

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.9 | 76.2 | 114.2 |
| 200 | 30.6 | 24.3 | 36.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#44
2,4,5-Trichlorophenol
Concen: 52.275 ng
RT: 12.554 min Scan# 1635
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:196 Resp: 1151711

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 94.8 | 76.6 | 115.0 |
| 97 | 45.0 | 36.6 | 55.0 |
| 132 | 27.5 | 21.4 | 32.0 |






#45
2-Fluorobiphenyl
Concen: 97.594 ng
RT: 12.701 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:172 Resp: 703749

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.7 | 28.2 | 42.2 |
| 170 | 23.7 | 18.7 | 28.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#46
1,1'-Biphenyl
Concen: 49.294 ng
RT: 12.907 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:154 Resp: 4006544

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.4 | 20.2 | 60.2 |
| 76 | 11.6 | 0.0 | 31.5 |






#47
2-Chloronaphthalene
Concen: 48.883 ng
RT: 12.942 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:162 Resp: 298624

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 36.9 | 29.4 | 44.2 |
| 164 | 32.7 | 26.4 | 39.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#48
2-Nitroaniline
Concen: 48.247 ng
RT: 13.148 min Scan# 1736
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 65 Resp: 722237

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 72.6 | 58.7 | 88.1 |
| 138 | 128.1 | 99.0 | 148.4 |






#49
Acenaphthylene
Concen: 49.571 ng
RT: 13.813 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:152 Resp: 463492

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.2 | 16.1 | 24.1 |
| 153 | 13.3 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#50
Dimethylphthalate
Concen: 48.733 ng
RT: 13.560 min Scan# 1806
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:163 Resp: 3810769

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.0 | 3.4 | 5.0 |
| 164 | 10.2 | 8.0 | 12.0 |






#51
2,6-Dinitrotoluene
Concen: 53.202 ng
RT: 13.660 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:165 Resp: 81137

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 46.7 | 38.5 | 57.7 |
| 89 | 50.1 | 40.6 | 61.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#52
Acenaphthene
Concen: 48.515 ng
RT: 14.160 min Scan# 1908
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:154 Resp: 2923605

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 112.3 | 90.6 | 136.0 |
| 152 | 53.4 | 43.8 | 65.8 |






#53
3-Nitroaniline
Concen: 52.965 ng
RT: 13.989 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:138 Resp: 8656

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 10.1 | 7.8 | 11.8 |
| 92 | 105.7 | 81.0 | 121.4 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#54
2,4-Dinitrophenol
Concen: 47.384 ng
RT: 14.207 min Scan# 1916
Delta R.T. 0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:184 Resp: 335631

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 54.5 | 47.4 | 71.0 |
| 154 | 59.1 | 47.8 | 71.8 |






#55
Dibenzofuran
Concen: 48.287 ng
RT: 14.507 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:168 Resp: 480234

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 36.4 | 28.6 | 43.0 |
| 169 | 13.5 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#56
4-Nitrophenol
Concen: 50.679 ng
RT: 14.313 min Scan# 1934
Delta R.T. 0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:139 Resp: 764396

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 67.8 | 45.9 | 85.9 |
| 65 | 83.0 | 61.4 | 101.4 |






#57
2,4-Dinitrotoluene
Concen: 48.291 ng
RT: 14.472 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:165 Resp: 108127

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 37.1 | 31.0 | 46.4 |
| 89 | 64.5 | 53.4 | 80.0 |
| 182 | 2.7 | 2.0 | 3.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#58
Fluorene
Concen: 48.458 ng
RT: 15.172 min Scan# 2080
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:166 Resp: 3727642

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 97.5 | 78.6 | 117.8 |
| 167 | 13.4 | 10.7 | 16.1 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 50.872 ng
RT: 14.731 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:232 Resp: 101496

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 44.2 | 35.7 | 53.5 |
| 130 | 2.2 | 1.8 | 2.6 |
| 166 | 31.1 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#60
Diethylphthalate
Concen: 48.587 ng
RT: 14.972 min Scan# 2046
Delta R.T. 0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:149 Resp: 3816042

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 22.2 | 17.8 | 26.8 |
| 150 | 12.4 | 10.2 | 15.2 |






#61
4-Chlorophenyl-phenylether
Concen: 48.332 ng
RT: 15.178 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:204 Resp: 183997

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 32.7 | 25.8 | 38.6 |
| 141 | 56.1 | 45.1 | 67.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#62
4-Nitroaniline
Concen: 51.948 ng
RT: 15.189 min Scan# 2083
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:138 Resp: 890284

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 46.1 | 26.0 | 66.0 |
| 108 | 70.5 | 50.0 | 90.0 |






#63
Azobenzene
Concen: 48.355 ng
RT: 15.472 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion: 77 Resp: 338184

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 30.2 | 7.7 | 47.7 |
| 105 | 22.3 | 1.9 | 41.9 |
| 51 | 27.7 | 8.2 | 48.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.907 min Scan# 2375
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:188 Resp: 2198784

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 9.0 | 7.3 | 10.9 |
| 80 | 9.4 | 7.8 | 11.6 |






#65
4,6-Dinitro-2-methylphenol
Concen: 48.762 ng
RT: 15.254 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:198 Resp: 48891

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 39.3 | 16.8 | 56.8 |
| 105 | 41.3 | 20.8 | 60.8 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#66
n-Nitrosodiphenylamine
Concen: 50.162 ng
RT: 15.389 min Scan# 2117
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:169 Resp: 3156476

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 66.8 | 54.1 | 81.1 |
| 167 | 36.0 | 29.0 | 43.6 |






#67
4-Bromophenyl-phenylether
Concen: 50.881 ng
RT: 16.089 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:248 Resp: 114549

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 98.7 | 79.4 | 119.2 |
| 141 | 66.9 | 51.4 | 77.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#68
Hexachlorobenzene
Concen: 49.621 ng
RT: 16.195 min Scan# 2254
Delta R.T. -0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:284 Resp: 1365447

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 32.0 | 25.9 | 38.9 |
| 249 | 29.9 | 23.8 | 35.8 |






#69
Atrazine
Concen: 50.863 ng
RT: 16.395 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:200 Resp: 93788

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 27.0 | 6.6 | 46.6 |
| 215 | 48.6 | 28.0 | 68.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#70
Pentachlorophenol
Concen: 52.511 ng
RT: 16.554 min Scan# 2315
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:266 Resp: 975866

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 64.4 | 50.4 | 75.6 |
| 264 | 62.2 | 50.2 | 75.2 |






#71
Phenanthrene
Concen: 48.179 ng
RT: 16.954 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:178 Resp: 566055

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.2 | 15.4 | 23.2 |
| 179 | 15.6 | 12.6 | 19.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#72
Anthracene
Concen: 49.812 ng
RT: 17.054 min Scan# 2400
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:178 Resp: 5614698

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.9 | 15.0 | 22.6 |
| 179 | 15.9 | 12.5 | 18.7 |






#73
Carbazole
Concen: 49.399 ng
RT: 17.325 min Scan# 24
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:167 Resp: 544419

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.8 | 17.4 | 26.0 |
| 139 | 13.1 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#74
Di-n-butylphthalate
Concen: 52.713 ng
RT: 17.966 min Scan# 2555
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:149 Resp: 6469226

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.6 | 7.3 | 10.9 |
| 104 | 4.7 | 3.8 | 5.6 |






#75
Fluoranthene
Concen: 49.995 ng
RT: 19.054 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:202 Resp: 686598

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 11.4 | 0.0 | 30.4 |
| 203 | 18.0 | 0.0 | 37.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.354 min Scan# 3131
Delta R.T. 0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:240 Resp: 2246809

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 10.2 | 9.0 | 13.4 |
| 236 | 26.0 | 20.3 | 30.5 |






#77
Benzidine
Concen: 59.401 ng
RT: 19.271 min Scan# 2
Delta R.T. 0.005 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:184 Resp: 235715

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 15.4 | 12.6 | 18.8 |
| 183 | 11.6 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#78
Pyrene
Concen: 50.503 ng
RT: 19.436 min Scan# 2805
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:202 Resp: 7028090

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 21.5 | 17.1 | 25.7 |
| 203 | 17.9 | 14.1 | 21.1 |






#79
Terphenyl-d14
Concen: 100.557 ng
RT: 19.683 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:244 Resp:1028002

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 7.5 | 5.9 | 8.9 |
| 122 | 11.2 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#80
Butylbenzylphthalate
Concen: 49.237 ng
RT: 20.442 min Scan# 2976
Delta R.T. 0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:149 Resp: 2645413

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 64.1 | 50.6 | 76.0 |
| 206 | 21.5 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 49.798 ng
RT: 21.336 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:228 Resp: 678056

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 27.7 | 22.1 | 33.1 |
| 229 | 20.2 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#82
3,3'-Dichlorobenzidine
Concen: 54.684 ng
RT: 21.271 min Scan# 3117
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:252 Resp: 2241436

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 66.1 | 51.9 | 77.9 |
| 126 | 10.8 | 8.1 | 12.1 |






#83
Chrysene
Concen: 49.609 ng
RT: 21.395 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:228 Resp: 650036

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 30.4 | 24.4 | 36.6 |
| 229 | 20.3 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 56.690 ng
RT: 21.336 min Scan# 3128
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:149 Resp: 4140331

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 28.1 | 21.9 | 32.9 |
| 279 | 4.5 | 3.5 | 5.3 |






#85
Di-n-octyl phthalate
Concen: 50.271 ng
RT: 22.565 min Scan# 3
Delta R.T. 0.005 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:149 Resp: 677654

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.6 | 1.3 | 1.9 |
| 43 | 8.4 | 6.9 | 10.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 24.518 min Scan# 3669
Delta R.T. 0.023 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:264 Resp: 2460402

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 24.3 | 18.6 | 28.0 |
| 265 | 21.9 | 17.7 | 26.5 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 53.583 ng m
RT: 28.059 min Scan# 4
Delta R.T. 0.023 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:276 Resp: 840064

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 16.1 | 13.0 | 19.6 |
| 277 | 20.3 | 15.9 | 23.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#88
Benzo(b)fluoranthene
Concen: 51.035 ng
RT: 23.518 min Scan# 3499
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:252 Resp: 7308592

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.0 | 17.7 | 26.5 |
| 125 | 10.4 | 8.1 | 12.1 |






#89
Benzo(k)fluoranthene
Concen: 49.931 ng
RT: 23.595 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:252 Resp: 711469

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.9 | 17.7 | 26.5 |
| 125 | 10.1 | 8.2 | 12.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#90
Benzo(a)pyrene
Concen: 52.234 ng
RT: 24.359 min Scan# 3642
Delta R.T. 0.006 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:252 Resp: 6366556

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.0 | 17.6 | 26.4 |
| 125 | 10.8 | 9.0 | 13.4 |






#91
Dibenzo(a,h)anthracene
Concen: 53.464 ng
RT: 28.136 min Scan# 4
Delta R.T. -0.000 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:278 Resp: 692698

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 16.1 | 13.1 | 19.7 |
| 279 | 24.4 | 18.9 | 28.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC050

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#92
Benzo(g,h,i)perylene
Concen: 52.480 ng
RT: 29.159 min Scan# 4458
Delta R.T. 0.035 min
Lab File: BP025076.D
Acq: 03 Jul 2025 13:07

Tgt Ion:276 Resp: 7023994

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 24.0 | 19.3 | 28.9 |
| 138 | 21.1 | 16.9 | 25.3 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025077.D
Acq On : 03 Jul 2025 13:49
Operator : CG/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:56:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.443 | 152 | 400654 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.184 | 136 | 1612661 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.095 | 164 | 1045830 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 16.907 | 188 | 2142509 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.360 | 240 | 2162992 | 20.000 | ng | 0.02 |
| 86) Perylene-d12 | 24.512 | 264 | 2316428 | 20.000 | ng | 0.02 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.096 | 112 | 2806806 | 124.128 | ng | 0.00 |
| 7) Phenol-d6 | 6.631 | 99 | 3710318 | 125.737 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.572 | 82 | 3129724 | 129.067 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.613 | 330 | 1671151 | 135.264 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 12.695 | 172 | 8229134 | 121.896 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.683 | 244 | 12149101 | 123.446 | ng | 0.01 |
| Target Compounds | | | | | | Qvalue |
| 2) 1,4-Dioxane | 3.125 | 88 | 514997 | 58.570 | ng | 100 |
| 3) Pyridine | 3.490 | 79 | 1488833m | 61.892 | ng | |
| 4) n-Nitrosodimethylamine | 3.408 | 42 | 627147 | 61.529 | ng | 96 |
| 6) Aniline | 6.790 | 93 | 1716823 | 62.270 | ng | 100 |
| 8) 2-Chlorophenol | 7.019 | 128 | 1555133 | 61.468 | ng | 99 |
| 9) Benzaldehyde | 6.596 | 77 | 770087 | 49.971 | ng | 100 |
| 10) Phenol | 6.660 | 94 | 1851536 | 61.753 | ng | 100 |
| 11) bis(2-Chloroethyl)ether | 6.884 | 93 | 1408699 | 61.048 | ng | 100 |
| 12) 1,3-Dichlorobenzene | 7.331 | 146 | 1698963 | 60.002 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 7.478 | 146 | 1711649 | 59.706 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 7.784 | 146 | 1644605 | 59.636 | ng | 99 |
| 15) Benzyl Alcohol | 7.678 | 79 | 1271626 | 64.566 | ng | 98 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.978 | 45 | 1799327 | 59.970 | ng | 99 |
| 17) 2-Methylphenol | 7.878 | 107 | 1227158 | 63.538 | ng | 99 |
| 18) Hexachloroethane | 8.501 | 117 | 577842 | 61.873 | ng | 99 |
| 19) n-Nitroso-di-n-propyla... | 8.243 | 70 | 1096243 | 62.642 | ng | 99 |
| 20) 3+4-Methylphenols | 8.201 | 107 | 1716929 | 63.466 | ng | 99 |
| 22) Acetophenone | 8.249 | 105 | 2324403 | 60.946 | ng | 98 |
| 24) Nitrobenzene | 8.613 | 77 | 1583678 | 64.118 | ng | 98 |
| 25) Isophorone | 9.137 | 82 | 2889727 | 63.768 | ng | 98 |
| 26) 2-Nitrophenol | 9.313 | 139 | 773866 | 63.426 | ng | 99 |
| 27) 2,4-Dimethylphenol | 9.384 | 122 | 1082275 | 64.024 | ng | 98 |
| 28) bis(2-Chloroethoxy)met... | 9.625 | 93 | 1922438 | 62.478 | ng | 100 |
| 29) 2,4-Dichlorophenol | 9.837 | 162 | 1554368 | 66.313 | ng | 99 |
| 30) 1,2,4-Trichlorobenzene | 10.060 | 180 | 1621972 | 62.486 | ng | 98 |
| 31) Naphthalene | 10.231 | 128 | 5073090 | 61.093 | ng | 100 |
| 32) Benzoic acid | 9.537 | 122 | 1088519 | 65.480 | ng | 95 |
| 33) 4-Chloroaniline | 10.348 | 127 | 1986181 | 64.093 | ng | 99 |
| 34) Hexachlorobutadiene | 10.537 | 225 | 926172 | 62.176 | ng | 99 |
| 35) Caprolactam | 11.125 | 113 | 523393 | 65.271 | ng | 97 |
| 36) 4-Chloro-3-methylphenol | 11.478 | 107 | 1604130 | 65.488 | ng | 99 |
| 37) 2-Methylnaphthalene | 11.860 | 142 | 3613084 | 61.667 | ng | 99 |
| 38) 1-Methylnaphthalene | 12.090 | 142 | 3560508 | 62.310 | ng | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 12.237 | 216 | 1914890 | 63.709 | ng | 99 |
| 41) Hexachlorocyclopentadiene | 12.231 | 237 | 537315 | 68.439 | ng | 99 |
| 43) 2,4,6-Trichlorophenol | 12.495 | 196 | 1250359 | 68.701 | ng | 98 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025077.D
Acq On : 03 Jul 2025 13:49
Operator : CG/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:56:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 12.560 | 196 | 1383790 | 67.090 | ng | 99 |
| 46) 1,1'-Biphenyl | 12.907 | 154 | 4675752 | 61.447 | ng | 100 |
| 47) 2-Chloronaphthalene | 12.937 | 162 | 3517185 | 61.497 | ng | 98 |
| 48) 2-Nitroaniline | 13.148 | 65 | 907242 | 63.792 | ng | 99 |
| 49) Acenaphthylene | 13.807 | 152 | 5526428 | 63.134 | ng | 100 |
| 50) Dimethylphthalate | 13.566 | 163 | 4648592 | 63.498 | ng | 100 |
| 51) 2,6-Dinitrotoluene | 13.660 | 165 | 982034 | 68.780 | ng | 99 |
| 52) Acenaphthene | 14.160 | 154 | 3486609 | 61.801 | ng | 99 |
| 53) 3-Nitroaniline | 13.995 | 138 | 1081057 | 70.651 | ng | 97 |
| 54) 2,4-Dinitrophenol | 14.207 | 184 | 430100 | 62.873 | ng | 94 |
| 55) Dibenzofuran | 14.501 | 168 | 5689473 | 61.105 | ng | 99 |
| 56) 4-Nitrophenol | 14.319 | 139 | 944200 | 66.867 | ng | 99 |
| 57) 2,4-Dinitrotoluene | 14.460 | 165 | 1348673 | 63.463 | ng | 98 |
| 58) Fluorene | 15.160 | 166 | 4402726 | 61.134 | ng | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 14.736 | 232 | 1246432 | 66.732 | ng | 99 |
| 60) Diethylphthalate | 14.972 | 149 | 4634770 | 63.033 | ng | 99 |
| 61) 4-Chlorophenyl-phenyle... | 15.172 | 204 | 2194839 | 61.583 | ng | 99 |
| 62) 4-Nitroaniline | 15.183 | 138 | 1111684 | 69.287 | ng | 99 |
| 63) Azobenzene | 15.472 | 77 | 3969523 | 60.626 | ng | 98 |
| 65) 4,6-Dinitro-2-methylph... | 15.254 | 198 | 625497 | 62.277 | ng | 99 |
| 66) n-Nitrosodiphenylamine | 15.395 | 169 | 3803576 | 62.033 | ng | 99 |
| 67) 4-Bromophenyl-phenylether | 16.095 | 248 | 1375012 | 62.680 | ng | 96 |
| 68) Hexachlorobenzene | 16.201 | 284 | 1649772 | 61.529 | ng | 97 |
| 69) Atrazine | 16.389 | 200 | 1107895 | 61.661 | ng | 99 |
| 70) Pentachlorophenol | 16.548 | 266 | 1197597 | 66.135 | ng | 99 |
| 71) Phenanthrene | 16.948 | 178 | 6952893 | 60.733 | ng | 100 |
| 72) Anthracene | 17.048 | 178 | 6826544 | 62.154 | ng | 100 |
| 73) Carbazole | 17.325 | 167 | 6534571 | 60.850 | ng | 100 |
| 74) Di-n-butylphthalate | 17.966 | 149 | 7890575 | 65.983 | ng | 99 |
| 75) Fluoranthene | 19.054 | 202 | 8151085 | 60.911 | ng | 98 |
| 77) Benzidine | 19.260 | 184 | 1867570 | 48.887 | ng | 100 |
| 78) Pyrene | 19.436 | 202 | 8447118 | 63.052 | ng | 99 |
| 80) Butylbenzylphthalate | 20.436 | 149 | 3368485 | 63.936 | ng | 99 |
| 81) Benzo(a)anthracene | 21.342 | 228 | 8160252 | 62.254 | ng | 99 |
| 82) 3,3'-Dichlorobenzidine | 21.271 | 252 | 2685016 | 68.044 | ng | 100 |
| 83) Chrysene | 21.401 | 228 | 7729618 | 61.276 | ng | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 21.342 | 149 | 4960480 | 70.551 | ng | 98 |
| 85) Di-n-octyl phthalate | 22.565 | 149 | 8384283 | 62.869 | ng | 99 |
| 87) Indeno(1,2,3-cd)pyrene | 28.071 | 276 | 9961748m | 67.489 | ng | |
| 88) Benzo(b)fluoranthene | 23.518 | 252 | 8800482 | 65.272 | ng | 100 |
| 89) Benzo(k)fluoranthene | 23.583 | 252 | 8461854 | 63.076 | ng | 99 |
| 90) Benzo(a)pyrene | 24.359 | 252 | 7703067 | 67.128 | ng | 100 |
| 91) Dibenzo(a,h)anthracene | 28.153 | 278 | 8144621 | 66.770 | ng | 99 |
| 92) Benzo(g,h,i)perylene | 29.159 | 276 | 8319243 | 66.020 | ng | 100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025077.D
Acq On : 03 Jul 2025 13:49
Operator : CG/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Jul 03 15:56:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.443 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:152 Resp: 40065

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 156.3 | 126.1 | 189.1 |
| 115 | 56.1 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#2
1,4-Dioxane
Concen: 58.570 ng
RT: 3.125 min Scan# 32
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 88 Resp: 514997

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 68.7 | 55.0 | 82.4 |
| 43 | 27.4 | 21.9 | 32.9 |






#3
Pyridine
Concen: 61.892 ng m
RT: 3.490 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 79 Resp: 148883

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 64.6 | 52.2 | 78.4 |
| 51 | 33.6 | 26.8 | 40.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#4
n-Nitrosodimethylamine
Concen: 61.529 ng
RT: 3.408 min Scan# 80
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 42 Resp: 627147

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 136.8 | 105.2 | 157.8 |
| 44 | 7.2 | 6.6 | 9.8 |






#5
2-Fluorophenol
Concen: 124.128 ng
RT: 5.096 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:112 Resp: 280680

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 53.0 | 42.0 | 63.0 |
| 63 | 27.4 | 22.2 | 33.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#6
Aniline
Concen: 62.270 ng
RT: 6.790 min Scan# 655
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 93 Resp: 1716823

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 36.1 | 28.9 | 43.3 |
| 65 | 18.5 | 15.2 | 22.8 |






#7
Phenol-d6
Concen: 125.737 ng
RT: 6.631 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 99 Resp: 371031

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 17.1 | 13.6 | 20.4 |
| 71 | 31.5 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#8
2-Chlorophenol
Concen: 61.468 ng
RT: 7.019 min Scan# 694
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:128 Resp: 1555133

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.8 | 12.9 | 52.9 |
| 64 | 41.0 | 20.3 | 60.3 |






#9
Benzaldehyde
Concen: 49.971 ng
RT: 6.596 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 77 Resp: 77008

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 102.7 | 82.4 | 122.4 |
| 106 | 97.8 | 78.3 | 118.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#10
Phenol
Concen: 61.753 ng
RT: 6.660 min Scan# 633
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 94 Resp: 1851536

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 26.8 | 6.9 | 46.9 |
| 66 | 34.6 | 14.6 | 54.6 |






#11
bis(2-Chloroethyl)ether
Concen: 61.048 ng
RT: 6.884 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 93 Resp: 140869

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 71.9 | 51.9 | 91.9 |
| 95 | 32.4 | 12.4 | 52.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#12
1,3-Dichlorobenzene
Concen: 60.002 ng
RT: 7.331 min Scan# 747
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:146 Resp: 1698963

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.8 | 51.9 | 77.9 |
| 75 | 27.7 | 21.8 | 32.6 |






#13
1,4-Dichlorobenzene
Concen: 59.706 ng
RT: 7.478 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:146 Resp: 171164

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.7 | 51.1 | 76.7 |
| 111 | 37.8 | 31.2 | 46.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#14
1,2-Dichlorobenzene
Concen: 59.636 ng
RT: 7.784 min Scan# 824
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:146 Resp: 1644605

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.7 | 51.1 | 76.7 |
| 111 | 39.4 | 31.2 | 46.8 |






#15
Benzyl Alcohol
Concen: 64.566 ng
RT: 7.678 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 79 Resp: 127162

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 84.2 | 66.0 | 99.0 |
| 77 | 65.6 | 53.4 | 80.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 59.970 ng
RT: 7.978 min Scan# 857
Delta R.T. 0.012 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 45 Resp: 1799327

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 14.5 | 0.0 | 34.8 |
| 79 | 11.6 | 0.0 | 30.7 |






#17
2-Methylphenol
Concen: 63.538 ng
RT: 7.878 min Scan# 84
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:107 Resp: 122715

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 112.7 | 89.7 | 134.5 |
| 77 | 42.9 | 34.2 | 51.4 |
| 79 | 41.1 | 33.4 | 50.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#18
Hexachloroethane
Concen: 61.873 ng
RT: 8.501 min Scan# 946
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:117 Resp: 577842

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 96.5 | 76.6 | 115.0 |
| 201 | 109.0 | 85.4 | 128.2 |






#19
n-Nitroso-di-n-propylamine
Concen: 62.642 ng
RT: 8.243 min Scan# 9(
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 70 Resp: 109624

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 59.6 | 46.5 | 69.7 |
| 101 | 10.6 | 7.9 | 11.9 |
| 130 | 24.2 | 19.3 | 28.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#20
3+4-Methylphenols
Concen: 63.466 ng
RT: 8.201 min Scan# 895
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:107 Resp: 1716929

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 88.7 | 69.5 | 109.5 |
| 77 | 30.9 | 11.4 | 51.4 |
| 79 | 26.5 | 6.9 | 46.9 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.184 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:136 Resp: 161266

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.9 | 8.7 | 13.1 |
| 54 | 6.9 | 5.4 | 8.2 |
| 68 | 5.6 | 4.5 | 6.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#22
Acetophenone
Concen: 60.946 ng
RT: 8.249 min Scan# 903
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:105 Resp: 2324403

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 7.4 | 5.4 | 8.2 |
| 51 | 23.4 | 19.6 | 29.4 |
| 120 | 23.0 | 18.8 | 28.2 |






#23
Nitrobenzene-d5
Concen: 129.067 ng
RT: 8.572 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 82 Resp: 312972

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 49.3 | 38.4 | 57.6 |
| 54 | 47.5 | 37.8 | 56.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#24
Nitrobenzene
Concen: 64.118 ng
RT: 8.613 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 77 Resp: 1583678

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 50.6 | 39.0 | 58.6 |
| 65 | 13.2 | 10.7 | 16.1 |






#25
Isophorone
Concen: 63.768 ng
RT: 9.137 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 82 Resp: 288972

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 6.9 | 5.8 | 8.6 |
| 138 | 20.4 | 15.7 | 23.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#26
2-Nitrophenol
Concen: 63.426 ng
RT: 9.313 min Scan# 1084
Delta R.T. 0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:139 Resp: 773866

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 27.8 | 22.3 | 33.5 |
| 65 | 42.8 | 33.8 | 50.8 |






#27
2,4-Dimethylphenol
Concen: 64.024 ng
RT: 9.384 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:122 Resp: 108227

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 112.2 | 88.2 | 132.2 |
| 121 | 59.0 | 46.1 | 69.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#28
bis(2-Chloroethoxy)methane
Concen: 62.478 ng
RT: 9.625 min Scan# 1137
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 93 Resp: 1922438

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 33.0 | 26.4 | 39.6 |
| 123 | 12.1 | 10.1 | 15.1 |






#29
2,4-Dichlorophenol
Concen: 66.313 ng
RT: 9.837 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:162 Resp: 155436

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 65.1 | 44.1 | 84.1 |
| 98 | 32.7 | 13.1 | 53.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#30
1,2,4-Trichlorobenzene
Concen: 62.486 ng
RT: 10.060 min Scan# 1211
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:180 Resp: 1621972

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 95.7 | 75.1 | 112.7 |
| 145 | 28.9 | 23.3 | 34.9 |






#31
Naphthalene
Concen: 61.093 ng
RT: 10.231 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:128 Resp: 507309

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.1 | 8.9 | 13.3 |
| 127 | 13.2 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#32
Benzoic acid
Concen: 65.480 ng
RT: 9.537 min Scan# 1122
Delta R.T. 0.036 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:122 Resp: 1088519

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 115.5 | 101.0 | 141.0 |
| 77 | 82.7 | 67.6 | 107.6 |






#33
4-Chloroaniline
Concen: 64.093 ng
RT: 10.348 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:127 Resp: 198618

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.5 | 25.4 | 38.2 |
| 65 | 26.6 | 21.0 | 31.6 |
| 92 | 17.4 | 13.4 | 20.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#34
Hexachlorobutadiene
Concen: 62.176 ng
RT: 10.537 min Scan# 1292
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:225 Resp: 926172

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 62.7 | 50.3 | 75.5 |
| 227 | 65.4 | 51.3 | 76.9 |






#35
Caprolactam
Concen: 65.271 ng
RT: 11.125 min Scan# 1
Delta R.T. 0.018 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:113 Resp: 52339

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 144.5 | 128.1 | 168.1 |
| 56 | 106.1 | 90.7 | 130.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#36
4-Chloro-3-methylphenol
Concen: 65.488 ng
RT: 11.478 min Scan# 1452
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:107 Resp: 1604130

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 28.0 | 22.4 | 33.6 |
| 142 | 86.0 | 69.6 | 104.4 |






#37
2-Methylnaphthalene
Concen: 61.667 ng
RT: 11.860 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:142 Resp: 361308

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 88.7 | 71.2 | 106.8 |
| 115 | 33.0 | 25.9 | 38.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#38
1-Methylnaphthalene
Concen: 62.310 ng
RT: 12.090 min Scan# 1556
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:142 Resp: 3560508

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 91.7 | 74.1 | 111.1 |
| 116 | 3.6 | 2.7 | 4.1 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.095 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:164 Resp: 104583

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 103.0 | 80.3 | 120.5 |
| 160 | 45.8 | 35.9 | 53.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 63.709 ng
RT: 12.237 min Scan# 1581
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:216 Resp: 1914890

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.5 | 62.4 | 93.6 |
| 179 | 20.7 | 16.5 | 24.7 |
| 108 | 16.6 | 13.7 | 20.5 |






#41
Hexachlorocyclopentadiene
Concen: 68.439 ng
RT: 12.231 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:237 Resp: 53731

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 63.0 | 41.6 | 81.6 |
| 272 | 13.6 | 0.0 | 33.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#42
2,4,6-Tribromophenol
Concen: 135.264 ng
RT: 15.613 min Scan# 2155
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:330 Resp: 1671151

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 97.1 | 77.7 | 116.5 |
| 141 | 31.6 | 26.0 | 39.0 |






#43
2,4,6-Trichlorophenol
Concen: 68.701 ng
RT: 12.495 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:196 Resp: 125035

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.8 | 76.2 | 114.2 |
| 200 | 31.2 | 24.3 | 36.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#44
2,4,5-Trichlorophenol
Concen: 67.090 ng
RT: 12.560 min Scan# 1636
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:196 Resp: 1383790

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 97.3 | 76.6 | 115.0 |
| 97 | 46.1 | 36.6 | 55.0 |
| 132 | 27.5 | 21.4 | 32.0 |






#45
2-Fluorobiphenyl
Concen: 121.896 ng
RT: 12.695 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:172 Resp: 822913

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.5 | 28.2 | 42.2 |
| 170 | 23.9 | 18.7 | 28.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#46
1,1'-Biphenyl
Concen: 61.447 ng
RT: 12.907 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:154 Resp: 4675752

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.4 | 20.2 | 60.2 |
| 76 | 11.4 | 0.0 | 31.5 |






#47
2-Chloronaphthalene
Concen: 61.497 ng
RT: 12.937 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:162 Resp: 351718

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 38.7 | 29.4 | 44.2 |
| 164 | 32.9 | 26.4 | 39.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#48
2-Nitroaniline
Concen: 63.792 ng
RT: 13.148 min Scan# 1736
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 65 Resp: 907242

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 73.3 | 58.7 | 88.1 |
| 138 | 125.1 | 99.0 | 148.4 |






#49
Acenaphthylene
Concen: 63.134 ng
RT: 13.807 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:152 Resp: 552642

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.1 | 16.1 | 24.1 |
| 153 | 13.5 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#50
Dimethylphthalate
Concen: 63.498 ng
RT: 13.566 min Scan# 1807
Delta R.T. 0.012 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:163 Resp: 4648592

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.0 | 3.4 | 5.0 |
| 164 | 10.2 | 8.0 | 12.0 |






#51
2,6-Dinitrotoluene
Concen: 68.780 ng
RT: 13.660 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:165 Resp: 98203

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 49.0 | 38.5 | 57.7 |
| 89 | 51.9 | 40.6 | 61.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#52
Acenaphthene
Concen: 61.801 ng
RT: 14.160 min Scan# 1908
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:154 Resp: 3486609

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 111.7 | 90.6 | 136.0 |
| 152 | 53.7 | 43.8 | 65.8 |






#53
3-Nitroaniline
Concen: 70.651 ng
RT: 13.995 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:138 Resp: 108105

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 9.9 | 7.8 | 11.8 |
| 92 | 103.9 | 81.0 | 121.4 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#54
2,4-Dinitrophenol
Concen: 62.873 ng
RT: 14.207 min Scan# 1916
Delta R.T. 0.012 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:184 Resp: 430100

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 51.5 | 47.4 | 71.0 |
| 154 | 58.2 | 47.8 | 71.8 |






#55
Dibenzofuran
Concen: 61.105 ng
RT: 14.501 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:168 Resp: 568947

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 36.4 | 28.6 | 43.0 |
| 169 | 13.3 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#56
4-Nitrophenol
Concen: 66.867 ng
RT: 14.319 min Scan# 1935
Delta R.T. 0.018 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:139 Resp: 944200

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 66.7 | 45.9 | 85.9 |
| 65 | 80.4 | 61.4 | 101.4 |






#57
2,4-Dinitrotoluene
Concen: 63.463 ng
RT: 14.460 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:165 Resp: 134867

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 37.9 | 31.0 | 46.4 |
| 89 | 65.1 | 53.4 | 80.0 |
| 182 | 2.7 | 2.0 | 3.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#58
Fluorene
Concen: 61.134 ng
RT: 15.160 min Scan# 2078
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:166 Resp: 4402726

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 97.0 | 78.6 | 117.8 |
| 167 | 13.6 | 10.7 | 16.1 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 66.732 ng
RT: 14.736 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:232 Resp: 124643

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 43.7 | 35.7 | 53.5 |
| 130 | 2.3 | 1.8 | 2.6 |
| 166 | 30.5 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#60
Diethylphthalate
Concen: 63.033 ng
RT: 14.972 min Scan# 2046
Delta R.T. 0.012 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:149 Resp: 4634770

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 22.9 | 17.8 | 26.8 |
| 150 | 12.4 | 10.2 | 15.2 |






#61
4-Chlorophenyl-phenylether
Concen: 61.583 ng
RT: 15.172 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:204 Resp: 219483

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 33.0 | 25.8 | 38.6 |
| 141 | 56.4 | 45.1 | 67.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#62
4-Nitroaniline
Concen: 69.287 ng
RT: 15.183 min Scan# 2082
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:138 Resp: 1111684

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 45.0 | 26.0 | 66.0 |
| 108 | 69.6 | 50.0 | 90.0 |






#63
Azobenzene
Concen: 60.626 ng
RT: 15.472 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion: 77 Resp: 396952

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 30.5 | 7.7 | 47.7 |
| 105 | 22.3 | 1.9 | 41.9 |
| 51 | 28.1 | 8.2 | 48.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.907 min Scan# 2375
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:188 Resp: 2142509

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 8.7 | 7.3 | 10.9 |
| 80 | 9.3 | 7.8 | 11.6 |






#65
4,6-Dinitro-2-methylphenol
Concen: 62.277 ng
RT: 15.254 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:198 Resp: 62549

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 36.5 | 16.8 | 56.8 |
| 105 | 39.6 | 20.8 | 60.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#66
n-Nitrosodiphenylamine
Concen: 62.033 ng
RT: 15.395 min Scan# 2118
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:169 Resp: 3803576

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 66.4 | 54.1 | 81.1 |
| 167 | 36.3 | 29.0 | 43.6 |






#67
4-Bromophenyl-phenylether
Concen: 62.680 ng
RT: 16.095 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:248 Resp: 137501

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 98.3 | 79.4 | 119.2 |
| 141 | 70.2 | 51.4 | 77.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#68
Hexachlorobenzene
Concen: 61.529 ng
RT: 16.201 min Scan# 2255
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:284 Resp: 1649772

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 29.7 | 25.9 | 38.9 |
| 249 | 30.2 | 23.8 | 35.8 |






#69
Atrazine
Concen: 61.661 ng
RT: 16.389 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:200 Resp: 110789

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 25.9 | 6.6 | 46.6 |
| 215 | 47.6 | 28.0 | 68.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#70
Pentachlorophenol
Concen: 66.135 ng
RT: 16.548 min Scan# 2314
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:266 Resp: 1197597

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 64.7 | 50.4 | 75.6 |
| 264 | 63.2 | 50.2 | 75.2 |






#71
Phenanthrene
Concen: 60.733 ng
RT: 16.948 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:178 Resp: 695289

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.4 | 15.4 | 23.2 |
| 179 | 15.8 | 12.6 | 19.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#72
Anthracene
Concen: 62.154 ng
RT: 17.048 min Scan# 2399
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:178 Resp: 6826544

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.6 | 15.0 | 22.6 |
| 179 | 15.6 | 12.5 | 18.7 |






#73
Carbazole
Concen: 60.850 ng
RT: 17.325 min Scan# 24
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:167 Resp: 653457

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.9 | 17.4 | 26.0 |
| 139 | 12.8 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#74
Di-n-butylphthalate
Concen: 65.983 ng
RT: 17.966 min Scan# 2555
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:149 Resp: 7890575

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.5 | 7.3 | 10.9 |
| 104 | 4.7 | 3.8 | 5.6 |






#75
Fluoranthene
Concen: 60.911 ng
RT: 19.054 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:202 Resp: 815108

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 11.6 | 0.0 | 30.4 |
| 203 | 18.4 | 0.0 | 37.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.360 min Scan# 3132
Delta R.T. 0.018 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:240 Resp: 2162992

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.2 | 9.0 | 13.4 |
| 236 | 25.5 | 20.3 | 30.5 |






#77
Benzidine
Concen: 48.887 ng
RT: 19.260 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:184 Resp: 186757

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 15.7 | 12.6 | 18.8 |
| 183 | 11.6 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#78
Pyrene
Concen: 63.052 ng
RT: 19.436 min Scan# 2805
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:202 Resp: 8447118

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 21.8 | 17.1 | 25.7 |
| 203 | 17.9 | 14.1 | 21.1 |






#79
Terphenyl-d14
Concen: 123.446 ng
RT: 19.683 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:244 Resp:1214910

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 8.0 | 5.9 | 8.9 |
| 122 | 11.8 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#80
Butylbenzylphthalate
Concen: 63.936 ng
RT: 20.436 min Scan# 2975
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:149 Resp: 3368485

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 64.5 | 50.6 | 76.0 |
| 206 | 21.0 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 62.254 ng
RT: 21.342 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:228 Resp: 816025

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 28.0 | 22.1 | 33.1 |
| 229 | 20.5 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#82
3,3'-Dichlorobenzidine
Concen: 68.044 ng
RT: 21.271 min Scan# 3117
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:252 Resp: 2685016

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 64.9 | 51.9 | 77.9 |
| 126 | 9.9 | 8.1 | 12.1 |






#83
Chrysene
Concen: 61.276 ng
RT: 21.401 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:228 Resp: 772961

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 31.1 | 24.4 | 36.6 |
| 229 | 20.7 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 70.551 ng
RT: 21.342 min Scan# 3129
Delta R.T. 0.012 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:149 Resp: 4960480

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 28.5 | 21.9 | 32.9 |
| 279 | 4.8 | 3.5 | 5.3 |






#85
Di-n-octyl phthalate
Concen: 62.869 ng
RT: 22.565 min Scan# 3
Delta R.T. 0.005 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:149 Resp: 838428

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.6 | 1.3 | 1.9 |
| 43 | 8.3 | 6.9 | 10.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 24.512 min Scan# 3668
Delta R.T. 0.018 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:264 Resp: 2316428

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.9 | 18.6 | 28.0 |
| 265 | 21.8 | 17.7 | 26.5 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 67.489 ng m
RT: 28.071 min Scan# 4
Delta R.T. 0.035 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:276 Resp: 996174

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 16.5 | 13.0 | 19.6 |
| 277 | 19.8 | 15.9 | 23.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#88
Benzo(b)fluoranthene
Concen: 65.272 ng
RT: 23.518 min Scan# 3499
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:252 Resp: 8800482

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.3 | 17.7 | 26.5 |
| 125 | 10.2 | 8.1 | 12.1 |






#89
Benzo(k)fluoranthene
Concen: 63.076 ng
RT: 23.583 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:252 Resp: 846185

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.5 | 17.7 | 26.5 |
| 125 | 9.9 | 8.2 | 12.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#90
Benzo(a)pyrene
Concen: 67.128 ng
RT: 24.359 min Scan# 3642
Delta R.T. 0.006 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:252 Resp: 7703067

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.1 | 17.6 | 26.4 |
| 125 | 10.9 | 9.0 | 13.4 |






#91
Dibenzo(a,h)anthracene
Concen: 66.770 ng
RT: 28.153 min Scan# 4
Delta R.T. 0.018 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:278 Resp: 814462

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 16.1 | 13.1 | 19.7 |
| 279 | 24.3 | 18.9 | 28.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC060

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#92
Benzo(g,h,i)perylene
Concen: 66.020 ng
RT: 29.159 min Scan# 4458
Delta R.T. 0.035 min
Lab File: BP025077.D
Acq: 03 Jul 2025 13:49

Tgt Ion:276 Resp: 8319243

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 24.0 | 19.3 | 28.9 |
| 138 | 21.2 | 16.9 | 25.3 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025078.D
Acq On : 03 Jul 2025 14:31
Operator : CG/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:57:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.443 | 152 | 405609 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.195 | 136 | 1662149 | 20.000 | ng | 0.01 |
| 39) Acenaphthene-d10 | 14.095 | 164 | 1084696 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 16.919 | 188 | 2129464 | 20.000 | ng | 0.01 |
| 76) Chrysene-d12 | 21.354 | 240 | 2139587 | 20.000 | ng | 0.01 |
| 86) Perylene-d12 | 24.483 | 264 | 2356318 | 20.000 | ng | -0.01 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.102 | 112 | 3655581 | 159.689 | ng | 0.00 |
| 7) Phenol-d6 | 6.637 | 99 | 4808315 | 160.956 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.572 | 82 | 4100042 | 164.048 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.625 | 330 | 2161722 | 168.702 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 12.701 | 172 | 10212772 | 145.859 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.677 | 244 | 13073627 | 134.293 | ng | 0.00 |
| Target Compounds | | | | | | Qvalue |
| 2) 1,4-Dioxane | 3.125 | 88 | 663607 | 74.549 | ng | 100 |
| 3) Pyridine | 3.490 | 79 | 1951300m | 80.126 | ng | |
| 4) n-Nitrosodimethylamine | 3.408 | 42 | 811711 | 78.664 | ng | 98 |
| 6) Aniline | 6.796 | 93 | 2205608 | 79.022 | ng | 99 |
| 8) 2-Chlorophenol | 7.019 | 128 | 2043001 | 79.765 | ng | 99 |
| 10) Phenol | 6.666 | 94 | 2423035 | 79.827 | ng | 99 |
| 11) bis(2-Chloroethyl)ether | 6.890 | 93 | 1808927 | 77.435 | ng | 100 |
| 12) 1,3-Dichlorobenzene | 7.337 | 146 | 2200490 | 76.765 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 7.478 | 146 | 2228669 | 76.792 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 7.790 | 146 | 2114072 | 75.723 | ng | 99 |
| 15) Benzyl Alcohol | 7.684 | 79 | 1647105 | 82.609 | ng | 97 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.972 | 45 | 2295253 | 75.564 | ng | 99 |
| 17) 2-Methylphenol | 7.878 | 107 | 1594735 | 81.561 | ng | 100 |
| 18) Hexachloroethane | 8.507 | 117 | 750162 | 79.344 | ng | 96 |
| 19) n-Nitroso-di-n-propyla... | 8.243 | 70 | 1414389 | 79.834 | ng | 99 |
| 20) 3+4-Methylphenols | 8.207 | 107 | 2223238 | 81.178 | ng | 99 |
| 22) Acetophenone | 8.248 | 105 | 2974255 | 75.663 | ng | 99 |
| 24) Nitrobenzene | 8.613 | 77 | 2049856 | 80.522 | ng | 96 |
| 25) Isophorone | 9.143 | 82 | 3731607 | 79.894 | ng | 99 |
| 26) 2-Nitrophenol | 9.313 | 139 | 1054507 | 82.791 | ng | 99 |
| 27) 2,4-Dimethylphenol | 9.390 | 122 | 1396982 | 80.180 | ng | 99 |
| 28) bis(2-Chloroethoxy)met... | 9.625 | 93 | 2467631 | 77.809 | ng | 99 |
| 29) 2,4-Dichlorophenol | 9.848 | 162 | 2035842 | 84.268 | ng | 99 |
| 30) 1,2,4-Trichlorobenzene | 10.054 | 180 | 2084432 | 77.912 | ng | 98 |
| 31) Naphthalene | 10.242 | 128 | 6485251 | 75.773 | ng | 100 |
| 32) Benzoic acid | 9.560 | 122 | 1535999 | 87.682 | ng | 99 |
| 33) 4-Chloroaniline | 10.348 | 127 | 2533698 | 79.326 | ng | 99 |
| 34) Hexachlorobutadiene | 10.542 | 225 | 1198440 | 78.059 | ng | 99 |
| 35) Caprolactam | 11.137 | 113 | 679020 | 82.158 | ng | 99 |
| 36) 4-Chloro-3-methylphenol | 11.484 | 107 | 2102953 | 83.296 | ng | 99 |
| 37) 2-Methylnaphthalene | 11.866 | 142 | 4691171 | 77.683 | ng | 99 |
| 38) 1-Methylnaphthalene | 12.089 | 142 | 4509252 | 76.564 | ng | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 12.248 | 216 | 2462639 | 78.996 | ng | 99 |
| 41) Hexachlorocyclopentadiene | 12.236 | 237 | 711159 | 87.336 | ng | 99 |
| 43) 2,4,6-Trichlorophenol | 12.489 | 196 | 1646539 | 87.227 | ng | 99 |
| 44) 2,4,5-Trichlorophenol | 12.566 | 196 | 1806219 | 84.432 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025078.D
Acq On : 03 Jul 2025 14:31
Operator : CG/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 15:57:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 46) 1,1'-Biphenyl | 12.913 | 154 | 5987203 | 75.863 | ng | | 100 |
| 47) 2-Chloronaphthalene | 12.948 | 162 | 4544953 | 76.620 | ng | | 99 |
| 48) 2-Nitroaniline | 13.142 | 65 | 1186338 | 79.706 | ng | | 98 |
| 49) Acenaphthylene | 13.807 | 152 | 7027119 | 77.402 | ng | | 99 |
| 50) Dimethylphthalate | 13.566 | 163 | 5799322 | 76.378 | ng | | 99 |
| 51) 2,6-Dinitrotoluene | 13.672 | 165 | 1274236 | 86.048 | ng | | 99 |
| 52) Acenaphthene | 14.166 | 154 | 4454631 | 76.130 | ng | | 98 |
| 53) 3-Nitroaniline | 14.001 | 138 | 1399587 | 88.191 | ng | | 99 |
| 54) 2,4-Dinitrophenol | 14.207 | 184 | 592801 | 81.781 | ng | | 98 |
| 55) Dibenzofuran | 14.513 | 168 | 7178793 | 74.338 | ng | | 99 |
| 56) 4-Nitrophenol | 14.325 | 139 | 1251984 | 85.486 | ng | | 98 |
| 57) 2,4-Dinitrotoluene | 14.478 | 165 | 1759928 | 79.166 | ng | # | 96 |
| 58) Fluorene | 15.178 | 166 | 5543506 | 74.216 | ng | | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 14.742 | 232 | 1634442 | 84.370 | ng | | 98 |
| 60) Diethylphthalate | 14.978 | 149 | 5891551 | 77.254 | ng | | 99 |
| 61) 4-Chlorophenyl-phenyle... | 15.183 | 204 | 2717790 | 73.523 | ng | | 99 |
| 62) 4-Nitroaniline | 15.195 | 138 | 1404085 | 84.376 | ng | | 99 |
| 63) Azobenzene | 15.489 | 77 | 5079474 | 74.798 | ng | | 97 |
| 65) 4,6-Dinitro-2-methylph... | 15.254 | 198 | 836918 | 81.907 | ng | | 97 |
| 66) n-Nitrosodiphenylamine | 15.401 | 169 | 4815876 | 79.024 | ng | | 99 |
| 67) 4-Bromophenyl-phenylether | 16.095 | 248 | 1812737 | 83.140 | ng | | 96 |
| 68) Hexachlorobenzene | 16.207 | 284 | 2122928 | 79.660 | ng | | 97 |
| 69) Atrazine | 16.389 | 200 | 1312003 | 73.469 | ng | | 100 |
| 70) Pentachlorophenol | 16.566 | 266 | 1571184 | 87.298 | ng | | 98 |
| 71) Phenanthrene | 16.966 | 178 | 8609082 | 75.660 | ng | | 99 |
| 72) Anthracene | 17.048 | 178 | 8269798 | 75.756 | ng | | 99 |
| 73) Carbazole | 17.336 | 167 | 8311609 | 77.872 | ng | | 99 |
| 74) Di-n-butylphthalate | 17.966 | 149 | 9814307 | 82.572 | ng | | 99 |
| 75) Fluoranthene | 19.066 | 202 | 10352176 | 77.833 | ng | | 98 |
| 77) Benzidine | 19.266 | 184 | 2653175 | 70.212 | ng | | 100 |
| 78) Pyrene | 19.448 | 202 | 10597728 | 79.970 | ng | | 99 |
| 80) Butylbenzylphthalate | 20.436 | 149 | 4268020 | 80.862 | ng | | 99 |
| 81) Benzo(a)anthracene | 21.336 | 228 | 10175638 | 78.478 | ng | | 99 |
| 82) 3,3'-Dichlorobenzidine | 21.271 | 252 | 3397776 | 87.049 | ng | | 99 |
| 83) Chrysene | 21.407 | 228 | 9566614 | 76.668 | ng | | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 21.330 | 149 | 6160265 | 88.574 | ng | | 98 |
| 85) Di-n-octyl phthalate | 22.559 | 149 | 10681173 | 79.212 | ng | | 99 |
| 87) Indeno(1,2,3-cd)pyrene | 28.047 | 276 | 12878282m | 85.771 | ng | | |
| 88) Benzo(b)fluoranthene | 23.518 | 252 | 11185819 | 81.559 | ng | | 99 |
| 89) Benzo(k)fluoranthene | 23.583 | 252 | 10678946 | 78.255 | ng | | 99 |
| 90) Benzo(a)pyrene | 24.354 | 252 | 9800149 | 83.957 | ng | | 99 |
| 91) Dibenzo(a,h)anthracene | 28.147 | 278 | 10538537 | 84.932 | ng | | 99 |
| 92) Benzo(g,h,i)perylene | 29.147 | 276 | 10759787 | 83.943 | ng | | 100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025078.D
Acq On : 03 Jul 2025 14:31
Operator : CG/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Jul 03 15:57:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 15:38:19 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.443 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:152 Resp: 40560

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 155.0 | 126.1 | 189.1 |
| 115 | 56.6 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#2
1,4-Dioxane
Concen: 74.549 ng
RT: 3.125 min Scan# 32
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 88 Resp: 663607

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 68.5 | 55.0 | 82.4 |
| 43 | 27.4 | 21.9 | 32.9 |






#3
Pyridine
Concen: 80.126 ng m
RT: 3.490 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 79 Resp: 195130

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 65.4 | 52.2 | 78.4 |
| 51 | 33.1 | 26.8 | 40.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#4
n-Nitrosodimethylamine
Concen: 78.664 ng
RT: 3.408 min Scan# 80
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 42 Resp: 811711

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 134.1 | 105.2 | 157.8 |
| 44 | 6.7 | 6.6 | 9.8 |






#5
2-Fluorophenol
Concen: 159.689 ng
RT: 5.102 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:112 Resp: 365558

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 52.0 | 42.0 | 63.0 |
| 63 | 27.1 | 22.2 | 33.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#6
Aniline
Concen: 79.022 ng
RT: 6.796 min Scan# 656
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 93 Resp: 2205608

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 36.4 | 28.9 | 43.3 |
| 65 | 18.7 | 15.2 | 22.8 |






#7
Phenol-d6
Concen: 160.956 ng
RT: 6.637 min Scan# 6
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 99 Resp: 480831

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 17.0 | 13.6 | 20.4 |
| 71 | 31.3 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#8
2-Chlorophenol
Concen: 79.765 ng
RT: 7.019 min Scan# 694
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:128 Resp: 2043001

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.5 | 12.9 | 52.9 |
| 64 | 40.7 | 20.3 | 60.3 |






#10
Phenol
Concen: 79.827 ng
RT: 6.666 min Scan# 6
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 94 Resp: 242303

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 26.1 | 6.9 | 46.9 |
| 66 | 34.1 | 14.6 | 54.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#11
bis(2-Chloroethyl)ether
Concen: 77.435 ng
RT: 6.890 min Scan# 672
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 93 Resp: 1808927

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 71.5 | 51.9 | 91.9 |
| 95 | 32.4 | 12.4 | 52.4 |






#12
1,3-Dichlorobenzene
Concen: 76.765 ng
RT: 7.337 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:146 Resp: 220049

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.1 | 51.9 | 77.9 |
| 75 | 26.9 | 21.8 | 32.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#13
1,4-Dichlorobenzene
Concen: 76.792 ng
RT: 7.478 min Scan# 772
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:146 Resp: 2228669

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.8 | 51.1 | 76.7 |
| 111 | 38.4 | 31.2 | 46.8 |






#14
1,2-Dichlorobenzene
Concen: 75.723 ng
RT: 7.790 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:146 Resp: 211407

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.6 | 51.1 | 76.7 |
| 111 | 40.3 | 31.2 | 46.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#15
Benzyl Alcohol
Concen: 82.609 ng
RT: 7.684 min Scan# 807
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 79 Resp: 1647105

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 85.6 | 66.0 | 99.0 |
| 77 | 65.5 | 53.4 | 80.2 |






#16
2,2'-oxybis(1-Chloropropane)
Concen: 75.564 ng
RT: 7.972 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 45 Resp: 229525

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 14.7 | 0.0 | 34.8 |
| 79 | 11.3 | 0.0 | 30.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#17
2-Methylphenol
Concen: 81.561 ng
RT: 7.878 min Scan# 840
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:107 Resp: 1594735

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 112.0 | 89.7 | 134.5 |
| 77 | 43.2 | 34.2 | 51.4 |
| 79 | 41.2 | 33.4 | 50.2 |






#18
Hexachloroethane
Concen: 79.344 ng
RT: 8.507 min Scan# 94
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:117 Resp: 75016

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 97.7 | 76.6 | 115.0 |
| 201 | 112.0 | 85.4 | 128.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#19
n-Nitroso-di-n-propylamine
Concen: 79.834 ng
RT: 8.243 min Scan# 902
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 70 Resp: 1414389

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 58.6 | 46.5 | 69.7 |
| 101 | 10.6 | 7.9 | 11.9 |
| 130 | 24.6 | 19.3 | 28.9 |






#20
3+4-Methylphenols
Concen: 81.178 ng
RT: 8.207 min Scan# 8
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:107 Resp: 222323

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 90.8 | 69.5 | 109.5 |
| 77 | 30.9 | 11.4 | 51.4 |
| 79 | 27.7 | 6.9 | 46.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.195 min Scan# 1234
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:136 Resp: 1662149

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 11.0 | 8.7 | 13.1 |
| 54 | 6.7 | 5.4 | 8.2 |
| 68 | 5.4 | 4.5 | 6.7 |






#22
Acetophenone
Concen: 75.663 ng
RT: 8.248 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:105 Resp: 297425

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 7.4 | 5.4 | 8.2 |
| 51 | 23.6 | 19.6 | 29.4 |
| 120 | 23.5 | 18.8 | 28.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#23
Nitrobenzene-d5
Concen: 164.048 ng
RT: 8.572 min Scan# 958
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 82 Resp: 4100042

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 49.0 | 38.4 | 57.6 |
| 54 | 46.7 | 37.8 | 56.6 |






#24
Nitrobenzene
Concen: 80.522 ng
RT: 8.613 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 77 Resp: 204985

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 51.7 | 39.0 | 58.6 |
| 65 | 13.0 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#25
Isophorone
Concen: 79.894 ng
RT: 9.143 min Scan# 1055
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 82 Resp: 3731607

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.1 | 5.8 | 8.6 |
| 138 | 20.0 | 15.7 | 23.5 |






#26
2-Nitrophenol
Concen: 82.791 ng
RT: 9.313 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:139 Resp: 105450

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 27.3 | 22.3 | 33.5 |
| 65 | 42.7 | 33.8 | 50.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#27
2,4-Dimethylphenol
Concen: 80.180 ng
RT: 9.390 min Scan# 1097
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:122 Resp: 1396982

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 110.8 | 88.2 | 132.2 |
| 121 | 59.1 | 46.1 | 69.1 |






#28
bis(2-Chloroethoxy)methane
Concen: 77.809 ng
RT: 9.625 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 93 Resp: 246763

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.5 | 26.4 | 39.6 |
| 123 | 12.1 | 10.1 | 15.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#29
2,4-Dichlorophenol
Concen: 84.268 ng
RT: 9.848 min Scan# 1175
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:162 Resp: 2035842

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 65.1 | 44.1 | 84.1 |
| 98 | 32.4 | 13.1 | 53.1 |






#30
1,2,4-Trichlorobenzene
Concen: 77.912 ng
RT: 10.054 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:180 Resp: 208443

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 95.5 | 75.1 | 112.7 |
| 145 | 29.7 | 23.3 | 34.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#31
Naphthalene
Concen: 75.773 ng
RT: 10.242 min Scan# 1242
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:128 Resp: 6485251

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.0 | 8.9 | 13.3 |
| 127 | 13.2 | 10.4 | 15.6 |






#32
Benzoic acid
Concen: 87.682 ng
RT: 9.560 min Scan# 1
Delta R.T. 0.059 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:122 Resp: 153599

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 120.2 | 101.0 | 141.0 |
| 77 | 86.8 | 67.6 | 107.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#33
4-Chloroaniline
Concen: 79.326 ng
RT: 10.348 min Scan# 1260
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:127 Resp: 2533698

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.3 | 25.4 | 38.2 |
| 65 | 26.1 | 21.0 | 31.6 |
| 92 | 17.5 | 13.4 | 20.0 |






#34
Hexachlorobutadiene
Concen: 78.059 ng
RT: 10.542 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:225 Resp: 119844

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 62.0 | 50.3 | 75.5 |
| 227 | 64.8 | 51.3 | 76.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#35
Caprolactam
Concen: 82.158 ng
RT: 11.137 min Scan# 1394
Delta R.T. 0.029 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:113 Resp: 679020

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 148.6 | 128.1 | 168.1 |
| 56 | 109.9 | 90.7 | 130.7 |






#36
4-Chloro-3-methylphenol
Concen: 83.296 ng
RT: 11.484 min Scan# 14
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:107 Resp: 210295

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 28.2 | 22.4 | 33.6 |
| 142 | 86.3 | 69.6 | 104.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#37
2-Methylnaphthalene
Concen: 77.683 ng
RT: 11.866 min Scan# 1518
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:142 Resp: 4691171

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 88.0 | 71.2 | 106.8 |
| 115 | 33.1 | 25.9 | 38.9 |






#38
1-Methylnaphthalene
Concen: 76.564 ng
RT: 12.089 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:142 Resp: 450925

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 92.1 | 74.1 | 111.1 |
| 116 | 3.6 | 2.7 | 4.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.095 min Scan# 1897
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:164 Resp: 1084696

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 100.6 | 80.3 | 120.5 |
| 160 | 44.6 | 35.9 | 53.9 |






#40
1,2,4,5-Tetrachlorobenzene
Concen: 78.996 ng
RT: 12.248 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:216 Resp: 246263

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.6 | 62.4 | 93.6 |
| 179 | 20.9 | 16.5 | 24.7 |
| 108 | 16.8 | 13.7 | 20.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#41
Hexachlorocyclopentadiene
Concen: 87.336 ng
RT: 12.236 min Scan# 1581
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:237 Resp: 711159

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 62.5 | 41.6 | 81.6 |
| 272 | 13.4 | 0.0 | 33.6 |






#42
2,4,6-Tribromophenol
Concen: 168.702 ng
RT: 15.625 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:330 Resp: 216172

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 96.3 | 77.7 | 116.5 |
| 141 | 30.7 | 26.0 | 39.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#43
2,4,6-Trichlorophenol
Concen: 87.227 ng
RT: 12.489 min Scan# 1624
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:196 Resp: 1646539

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.8 | 76.2 | 114.2 |
| 200 | 31.0 | 24.3 | 36.5 |






#44
2,4,5-Trichlorophenol
Concen: 84.432 ng
RT: 12.566 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:196 Resp: 180621

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.8 | 76.6 | 115.0 |
| 97 | 44.1 | 36.6 | 55.0 |
| 132 | 26.9 | 21.4 | 32.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#45
2-Fluorobiphenyl
Concen: 145.859 ng
RT: 12.701 min Scan# 1660
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:172 Resp:10212772

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 36.3 | 28.2 | 42.2 |
| 170 | 24.2 | 18.7 | 28.1 |






#46
1,1'-Biphenyl
Concen: 75.863 ng
RT: 12.913 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:154 Resp: 598720

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.0 | 20.2 | 60.2 |
| 76 | 11.5 | 0.0 | 31.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#47
2-Chloronaphthalene
Concen: 76.620 ng
RT: 12.948 min Scan# 1702
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:162 Resp: 4544953

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 37.7 | 29.4 | 44.2 |
| 164 | 33.5 | 26.4 | 39.6 |






#48
2-Nitroaniline
Concen: 79.706 ng
RT: 13.142 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 65 Resp: 118633

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 71.8 | 58.7 | 88.1 |
| 138 | 120.9 | 99.0 | 148.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#49
Acenaphthylene
Concen: 77.402 ng
RT: 13.807 min Scan# 1848
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:152 Resp: 7027119

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.5 | 16.1 | 24.1 |
| 153 | 13.7 | 10.7 | 16.1 |






#50
Dimethylphthalate
Concen: 76.378 ng
RT: 13.566 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:163 Resp: 579932

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.1 | 3.4 | 5.0 |
| 164 | 10.3 | 8.0 | 12.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#51
2,6-Dinitrotoluene
Concen: 86.048 ng
RT: 13.672 min Scan# 1825
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:165 Resp: 1274236

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 47.8 | 38.5 | 57.7 |
| 89 | 50.2 | 40.6 | 61.0 |






#52
Acenaphthene
Concen: 76.130 ng
RT: 14.166 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:154 Resp: 445463

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 111.7 | 90.6 | 136.0 |
| 152 | 53.6 | 43.8 | 65.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#53
3-Nitroaniline
Concen: 88.191 ng
RT: 14.001 min Scan# 1881
Delta R.T. 0.018 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:138 Resp: 1399587

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 9.8 | 7.8 | 11.8 |
| 92 | 102.6 | 81.0 | 121.4 |






#54
2,4-Dinitrophenol
Concen: 81.781 ng
RT: 14.207 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:184 Resp: 59280

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 56.4 | 47.4 | 71.0 |
| 154 | 60.2 | 47.8 | 71.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#55
Dibenzofuran
Concen: 74.338 ng
RT: 14.513 min Scan# 1968
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:168 Resp: 7178793

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 36.4 | 28.6 | 43.0 |
| 169 | 13.7 | 10.7 | 16.1 |






#56
4-Nitrophenol
Concen: 85.486 ng
RT: 14.325 min Scan# 1
Delta R.T. 0.023 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:139 Resp: 125198

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 65.3 | 45.9 | 85.9 |
| 65 | 78.5 | 61.4 | 101.4 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#57
2,4-Dinitrotoluene
Concen: 79.166 ng
RT: 14.478 min Scan# 1962
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:165 Resp: 1759928

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 35.9 | 31.0 | 46.4 |
| 89 | 63.4 | 53.4 | 80.0 |
| 182 | 3.0 | 2.0 | 3.0# |






#58
Fluorene
Concen: 74.216 ng
RT: 15.178 min Scan# 2(
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:166 Resp: 554350(

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 97.9 | 78.6 | 117.8 |
| 167 | 13.8 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#59
2,3,4,6-Tetrachlorophenol
Concen: 84.370 ng
RT: 14.742 min Scan# 2007
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:232 Resp: 1634442

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 42.5 | 35.7 | 53.5 |
| 130 | 2.1 | 1.8 | 2.6 |
| 166 | 30.3 | 24.4 | 36.6 |






#60
Diethylphthalate
Concen: 77.254 ng
RT: 14.978 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:149 Resp: 589155

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 22.8 | 17.8 | 26.8 |
| 150 | 12.6 | 10.2 | 15.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#61
4-Chlorophenyl-phenylether
Concen: 73.523 ng
RT: 15.183 min Scan# 2082
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:204 Resp: 2717790

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 33.3 | 25.8 | 38.6 |
| 141 | 56.9 | 45.1 | 67.7 |






#62
4-Nitroaniline
Concen: 84.376 ng
RT: 15.195 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:138 Resp: 140408

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 45.7 | 26.0 | 66.0 |
| 108 | 71.1 | 50.0 | 90.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#63
Azobenzene
Concen: 74.798 ng
RT: 15.489 min Scan# 2134
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion: 77 Resp: 5079474

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 31.2 | 7.7 | 47.7 |
| 105 | 23.1 | 1.9 | 41.9 |
| 51 | 27.9 | 8.2 | 48.2 |






#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.919 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:188 Resp: 212946

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 9.1 | 7.3 | 10.9 |
| 80 | 9.8 | 7.8 | 11.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#65
4,6-Dinitro-2-methylphenol
Concen: 81.907 ng
RT: 15.254 min Scan# 2094
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:198 Resp: 836918

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 35.3 | 16.8 | 56.8 |
| 105 | 39.0 | 20.8 | 60.8 |






#66
n-Nitrosodiphenylamine
Concen: 79.024 ng
RT: 15.401 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:169 Resp: 481587

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 66.8 | 54.1 | 81.1 |
| 167 | 36.7 | 29.0 | 43.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#67
4-Bromophenyl-phenylether
Concen: 83.140 ng
RT: 16.095 min Scan# 2237
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:248 Resp: 1812737

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 97.6 | 79.4 | 119.2 |
| 141 | 71.0 | 51.4 | 77.0 |






#68
Hexachlorobenzene
Concen: 79.660 ng
RT: 16.207 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:284 Resp: 212292

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 30.0 | 25.9 | 38.9 |
| 249 | 30.5 | 23.8 | 35.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#69
Atrazine
Concen: 73.469 ng
RT: 16.389 min Scan# 2287
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:200 Resp: 1312003

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 26.2 | 6.6 | 46.6 |
| 215 | 48.0 | 28.0 | 68.0 |






#70
Pentachlorophenol
Concen: 87.298 ng
RT: 16.566 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:266 Resp: 157118

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 64.5 | 50.4 | 75.6 |
| 264 | 63.7 | 50.2 | 75.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#71
Phenanthrene
Concen: 75.660 ng
RT: 16.966 min Scan# 2385
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:178 Resp: 8609082

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 20.0 | 15.4 | 23.2 |
| 179 | 16.2 | 12.6 | 19.0 |






#72
Anthracene
Concen: 75.756 ng
RT: 17.048 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:178 Resp: 826979

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.3 | 15.0 | 22.6 |
| 179 | 16.2 | 12.5 | 18.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#73
Carbazole
Concen: 77.872 ng
RT: 17.336 min Scan# 2448
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:167 Resp: 8311609

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 22.1 | 17.4 | 26.0 |
| 139 | 12.8 | 10.4 | 15.6 |






#74
Di-n-butylphthalate
Concen: 82.572 ng
RT: 17.966 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:149 Resp: 981430

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.7 | 7.3 | 10.9 |
| 104 | 4.8 | 3.8 | 5.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#75
Fluoranthene
Concen: 77.833 ng
RT: 19.066 min Scan# 2742
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:202 Resp:10352176

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 11.2 | 0.0 | 30.4 |
| 203 | 18.5 | 0.0 | 37.6 |






#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.354 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:240 Resp: 213958

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.2 | 9.0 | 13.4 |
| 236 | 25.8 | 20.3 | 30.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#77
Benzidine
Concen: 70.212 ng
RT: 19.266 min Scan# 2776
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:184 Resp: 2653175

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 15.5 | 12.6 | 18.8 |
| 183 | 12.0 | 9.4 | 14.2 |






#78
Pyrene
Concen: 79.970 ng
RT: 19.448 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:202 Resp:1059772

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 21.9 | 17.1 | 25.7 |
| 203 | 18.1 | 14.1 | 21.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#79
Terphenyl-d14
Concen: 134.293 ng
RT: 19.677 min Scan# 2846
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:244 Resp:13073627

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 11.0 | 5.9 | 8.9# |
| 122 | 15.6 | 9.4 | 14.2# |






#80
Butylbenzylphthalate
Concen: 80.862 ng
RT: 20.436 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:149 Resp: 426802

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 64.1 | 50.6 | 76.0 |
| 206 | 22.1 | 17.2 | 25.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#81
Benzo(a)anthracene
Concen: 78.478 ng
RT: 21.336 min Scan# 3128
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:228 Resp:10175638

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 28.4 | 22.1 | 33.1 |
| 229 | 20.6 | 16.0 | 24.0 |






#82
3,3'-Dichlorobenzidine
Concen: 87.049 ng
RT: 21.271 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:252 Resp: 339777

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 65.3 | 51.9 | 77.9 |
| 126 | 11.3 | 8.1 | 12.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#83
Chrysene
Concen: 76.668 ng
RT: 21.407 min Scan# 3140
Delta R.T. 0.018 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:228 Resp: 9566614

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 31.3 | 24.4 | 36.6 |
| 229 | 20.6 | 16.0 | 24.0 |






#84
Bis(2-ethylhexyl)phthalate
Concen: 88.574 ng
RT: 21.330 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:149 Resp: 616026

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 28.4 | 21.9 | 32.9 |
| 279 | 4.8 | 3.5 | 5.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#85
Di-n-octyl phthalate
Concen: 79.212 ng
RT: 22.559 min Scan# 3336
Delta R.T. -0.001 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:149 Resp:10681173

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.6 | 1.3 | 1.9 |
| 43 | 8.3 | 6.9 | 10.3 |






#86
Perylene-d12
Concen: 20.000 ng
RT: 24.483 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:264 Resp: 235631

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.0 | 18.6 | 28.0 |
| 265 | 22.1 | 17.7 | 26.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#87
Indeno(1,2,3-cd)pyrene
Concen: 85.771 ng m
RT: 28.047 min Scan# 4269
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:276 Resp:12878282

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 16.4 | 13.0 | 19.6 |
| 277 | 20.5 | 15.9 | 23.9 |






#88
Benzo(b)fluoranthene
Concen: 81.559 ng
RT: 23.518 min Scan# 34
Delta R.T. 0.006 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:252 Resp:1118581

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.6 | 17.7 | 26.5 |
| 125 | 10.1 | 8.1 | 12.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#89
Benzo(k)fluoranthene
Concen: 78.255 ng
RT: 23.583 min Scan# 3510
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:252 Resp:10678946

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.6 | 17.7 | 26.5 |
| 125 | 9.6 | 8.2 | 12.2 |






#90
Benzo(a)pyrene
Concen: 83.957 ng
RT: 24.354 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:252 Resp: 980014

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.3 | 17.6 | 26.4 |
| 125 | 10.5 | 9.0 | 13.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#91
Dibenzo(a,h)anthracene
Concen: 84.932 ng
RT: 28.147 min Scan# 4286
Delta R.T. 0.012 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:278 Resp:10538537

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 15.8 | 13.1 | 19.7 |
| 279 | 24.0 | 18.9 | 28.3 |






#92
Benzo(g,h,i)perylene
Concen: 83.943 ng
RT: 29.147 min Scan# 4
Delta R.T. 0.023 min
Lab File: BP025078.D
Acq: 03 Jul 2025 14:31

Tgt Ion:276 Resp:1075978

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 24.2 | 19.3 | 28.9 |
| 138 | 20.8 | 16.9 | 25.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDICC080

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025079.D
Acq On : 03 Jul 2025 16:18
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 16:37:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.437 | 152 | 400883 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.184 | 136 | 1637601 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.090 | 164 | 1071158 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 16.907 | 188 | 2157756 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 21.354 | 240 | 2221808 | 20.000 | ng | 0.01 |
| 86) Perylene-d12 | 24.501 | 264 | 2415004 | 20.000 | ng | 0.00 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.096 | 112 | 1806055 | 79.825 | ng | 0.00 |
| 7) Phenol-d6 | 6.631 | 99 | 2363715 | 80.057 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.572 | 82 | 2040461 | 82.866 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.619 | 330 | 1045523 | 82.624 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 12.690 | 172 | 5442492 | 78.712 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.689 | 244 | 8493260 | 84.014 | ng | 0.02 |
| Target Compounds | | | | | | Qvalue |
| 2) 1,4-Dioxane | 3.120 | 88 | 332686 | 37.814 | ng | 99 |
| 3) Pyridine | 3.484 | 79 | 940029 | 39.027 | ng | 98 |
| 4) n-Nitrosodimethylamine | 3.402 | 42 | 397057 | 38.933 | ng | 97 |
| 6) Aniline | 6.778 | 93 | 1098010 | 39.803 | ng | 100 |
| 8) 2-Chlorophenol | 7.014 | 128 | 1023493 | 40.431 | ng | 99 |
| 9) Benzaldehyde | 6.596 | 77 | 629753 | 40.841 | ng | 98 |
| 10) Phenol | 6.655 | 94 | 1189351 | 39.645 | ng | 99 |
| 11) bis(2-Chloroethyl)ether | 6.884 | 93 | 905215 | 39.206 | ng | 99 |
| 12) 1,3-Dichlorobenzene | 7.331 | 146 | 1115941 | 39.389 | ng | 98 |
| 13) 1,4-Dichlorobenzene | 7.472 | 146 | 1129771 | 39.387 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 7.778 | 146 | 1075162 | 38.965 | ng | 99 |
| 15) Benzyl Alcohol | 7.672 | 79 | 805339 | 40.867 | ng | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.966 | 45 | 1158172 | 38.579 | ng | 98 |
| 17) 2-Methylphenol | 7.872 | 107 | 786921 | 40.721 | ng | 100 |
| 18) Hexachloroethane | 8.496 | 117 | 374849 | 40.115 | ng | 99 |
| 19) n-Nitroso-di-n-propyla... | 8.237 | 70 | 709521 | 40.521 | ng | 99 |
| 20) 3+4-Methylphenols | 8.196 | 107 | 1093973 | 40.415 | ng | 99 |
| 22) Acetophenone | 8.243 | 105 | 1510122 | 38.992 | ng | 99 |
| 24) Nitrobenzene | 8.613 | 77 | 1012505 | 40.369 | ng | 97 |
| 25) Isophorone | 9.125 | 82 | 1857268 | 40.360 | ng | 100 |
| 26) 2-Nitrophenol | 9.308 | 139 | 488721 | 40.694 | ng | 99 |
| 27) 2,4-Dimethylphenol | 9.378 | 122 | 692324 | 40.332 | ng | 99 |
| 28) bis(2-Chloroethoxy)met... | 9.625 | 93 | 1236294 | 39.567 | ng | 99 |
| 29) 2,4-Dichlorophenol | 9.837 | 162 | 985707 | 41.412 | ng | 99 |
| 30) 1,2,4-Trichlorobenzene | 10.060 | 180 | 1043043 | 39.571 | ng | 97 |
| 31) Naphthalene | 10.231 | 128 | 3302943 | 39.170 | ng | 100 |
| 32) Benzoic acid | 9.502 | 122 | 684343 | 40.952 | ng | 97 |
| 33) 4-Chloroaniline | 10.343 | 127 | 1265959 | 40.229 | ng | 99 |
| 34) Hexachlorobutadiene | 10.537 | 225 | 607380 | 40.154 | ng | 99 |
| 35) Caprolactam | 11.107 | 113 | 332744 | 40.864 | ng | 100 |
| 36) 4-Chloro-3-methylphenol | 11.484 | 107 | 1018998 | 40.967 | ng | 98 |
| 37) 2-Methylnaphthalene | 11.866 | 142 | 2340864 | 39.345 | ng | 100 |
| 38) 1-Methylnaphthalene | 12.084 | 142 | 2280444 | 39.301 | ng | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 12.243 | 216 | 1233612 | 40.072 | ng | 100 |
| 41) Hexachlorocyclopentadiene | 12.237 | 237 | 345683 | 42.989 | ng | 99 |
| 43) 2,4,6-Trichlorophenol | 12.490 | 196 | 794957 | 42.646 | ng | 99 |

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025079.D
Acq On : 03 Jul 2025 16:18
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025

Quant Time: Jul 03 16:37:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 12.560 | 196 | 894337 | 42.335 | ng | 98 |
| 46) 1,1'-Biphenyl | 12.907 | 154 | 3042013 | 39.032 | ng | 100 |
| 47) 2-Chloronaphthalene | 12.943 | 162 | 2296829 | 39.210 | ng | 99 |
| 48) 2-Nitroaniline | 13.149 | 65 | 565340 | 39.894 | ng | 97 |
| 49) Acenaphthylene | 13.807 | 152 | 3624825 | 40.431 | ng | 100 |
| 50) Dimethylphthalate | 13.560 | 163 | 3009077 | 40.131 | ng | 99 |
| 51) 2,6-Dinitrotoluene | 13.660 | 165 | 627637 | 42.919 | ng | 99 |
| 52) Acenaphthene | 14.160 | 154 | 2271995m | 39.318 | ng | |
| 53) 3-Nitroaniline | 13.990 | 138 | 680440 | 43.418 | ng | 98 |
| 54) 2,4-Dinitrophenol | 14.201 | 184 | 265751 | 40.066 | ng | 97 |
| 55) Dibenzofuran | 14.507 | 168 | 3740234 | 39.220 | ng | 100 |
| 56) 4-Nitrophenol | 14.313 | 139 | 612774 | 42.369 | ng | 97 |
| 57) 2,4-Dinitrotoluene | 14.466 | 165 | 846954 | 39.932 | ng | 95 |
| 58) Fluorene | 15.166 | 166 | 2925541 | 39.662 | ng | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 14.737 | 232 | 811401 | 42.414 | ng | 99 |
| 60) Diethylphthalate | 14.972 | 149 | 3056611 | 40.587 | ng | 99 |
| 61) 4-Chlorophenyl-phenyle... | 15.178 | 204 | 1430883 | 39.198 | ng | 98 |
| 62) 4-Nitroaniline | 15.184 | 138 | 715671 | 43.550 | ng | 97 |
| 63) Azobenzene | 15.472 | 77 | 2587410 | 38.583 | ng | 99 |
| 65) 4,6-Dinitro-2-methylph... | 15.248 | 198 | 379707 | 39.754 | ng | 97 |
| 66) n-Nitrosodiphenylamine | 15.401 | 169 | 2472464 | 40.039 | ng | 99 |
| 67) 4-Bromophenyl-phenylether | 16.095 | 248 | 894455 | 40.486 | ng | 96 |
| 68) Hexachlorobenzene | 16.201 | 284 | 1066631 | 39.499 | ng | 99 |
| 69) Atrazine | 16.390 | 200 | 739881 | 40.888 | ng | 98 |
| 70) Pentachlorophenol | 16.548 | 266 | 774549 | 42.471 | ng | 98 |
| 71) Phenanthrene | 16.948 | 178 | 4552722 | 39.487 | ng | 100 |
| 72) Anthracene | 17.060 | 178 | 4467088 | 40.385 | ng | 99 |
| 73) Carbazole | 17.331 | 167 | 4335399 | 40.086 | ng | 100 |
| 74) Di-n-butylphthalate | 17.966 | 149 | 5149955 | 42.761 | ng | 100 |
| 75) Fluoranthene | 19.060 | 202 | 5415104 | 40.180 | ng | 99 |
| 77) Benzidine | 19.272 | 184 | 2058569 | 52.461 | ng | 99 |
| 78) Pyrene | 19.436 | 202 | 5698991 | 41.413 | ng | 100 |
| 80) Butylbenzylphthalate | 20.442 | 149 | 2138437 | 40.921 | ng | 97 |
| 81) Benzo(a)anthracene | 21.336 | 228 | 5501070 | 40.856 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 21.266 | 252 | 1808957 | 44.629 | ng | 100 |
| 83) Chrysene | 21.401 | 228 | 5211272 | 40.218 | ng | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 21.342 | 149 | 3299494 | 45.685 | ng | 99 |
| 85) Di-n-octyl phthalate | 22.566 | 149 | 5256676 | 40.750 | ng | 100 |
| 87) Indeno(1,2,3-cd)pyrene | 28.048 | 276 | 6494157m | 42.211 | ng | |
| 88) Benzo(b)fluoranthene | 23.519 | 252 | 5782492 | 41.137 | ng | 100 |
| 89) Benzo(k)fluoranthene | 23.589 | 252 | 5590808 | 39.974 | ng | 100 |
| 90) Benzo(a)pyrene | 24.354 | 252 | 4952570 | 41.397 | ng | 100 |
| 91) Dibenzo(a,h)anthracene | 28.130 | 278 | 5394843 | 42.422 | ng | 99 |
| 92) Benzo(g,h,i)perylene | 29.148 | 276 | 5498609 | 41.855 | ng | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025079.D
Acq On : 03 Jul 2025 16:18
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Jul 03 16:37:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.437 min Scan# 7
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:152 Resp: 40088

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 158.1 | 126.1 | 189.1 |
| 115 | 57.1 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#2
1,4-Dioxane
Concen: 37.814 ng
RT: 3.120 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 88 Resp: 332686

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 69.6 | 55.0 | 82.4 |
| 43 | 27.7 | 21.9 | 32.9 |






#3
Pyridine
Concen: 39.027 ng
RT: 3.484 min Scan# 9
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 79 Resp: 94002

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 67.3 | 52.2 | 78.4 |
| 51 | 34.2 | 26.8 | 40.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#4
n-Nitrosodimethylamine
Concen: 38.933 ng
RT: 3.402 min Scan# 79
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 42 Resp: 397057

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 127.7 | 105.2 | 157.8 |
| 44 | 7.7 | 6.6 | 9.8 |






#5
2-Fluorophenol
Concen: 79.825 ng
RT: 5.096 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:112 Resp: 180605

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 51.7 | 42.0 | 63.0 |
| 63 | 26.6 | 22.2 | 33.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#6
Aniline
Concen: 39.803 ng
RT: 6.778 min Scan# 653
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 93 Resp: 1098010

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 36.2 | 28.9 | 43.3 |
| 65 | 18.5 | 15.2 | 22.8 |






#7
Phenol-d6
Concen: 80.057 ng
RT: 6.631 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 99 Resp: 236371

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 17.0 | 13.6 | 20.4 |
| 71 | 30.9 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#8
2-Chlorophenol
Concen: 40.431 ng
RT: 7.014 min Scan# 693
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:128 Resp: 1023493

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.2 | 12.9 | 52.9 |
| 64 | 40.2 | 20.3 | 60.3 |






#9
Benzaldehyde
Concen: 40.841 ng
RT: 6.596 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 77 Resp: 62975

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 103.8 | 82.4 | 122.4 |
| 106 | 100.8 | 78.3 | 118.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#10
Phenol
Concen: 39.645 ng
RT: 6.655 min Scan# 632
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 94 Resp: 1189351

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 26.3 | 6.9 | 46.9 |
| 66 | 34.2 | 14.6 | 54.6 |






#11
bis(2-Chloroethyl)ether
Concen: 39.206 ng
RT: 6.884 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 93 Resp: 90521

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 72.5 | 51.9 | 91.9 |
| 95 | 32.8 | 12.4 | 52.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#12
1,3-Dichlorobenzene
Concen: 39.389 ng
RT: 7.331 min Scan# 747
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:146 Resp: 1115941

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.5 | 51.9 | 77.9 |
| 75 | 27.8 | 21.8 | 32.6 |






#13
1,4-Dichlorobenzene
Concen: 39.387 ng
RT: 7.472 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:146 Resp: 112977

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.7 | 51.1 | 76.7 |
| 111 | 37.2 | 31.2 | 46.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#14
1,2-Dichlorobenzene
Concen: 38.965 ng
RT: 7.778 min Scan# 823
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:146 Resp: 1075162

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.8 | 51.1 | 76.7 |
| 111 | 39.9 | 31.2 | 46.8 |






#15
Benzyl Alcohol
Concen: 40.867 ng
RT: 7.672 min Scan# 8(
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 79 Resp: 80533(

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 83.2 | 66.0 | 99.0 |
| 77 | 65.7 | 53.4 | 80.2 |

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.579 ng
RT: 7.966 min Scan# 855
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 45 Resp: 1158172

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 15.5 | 0.0 | 34.8 |
| 79 | 11.4 | 0.0 | 30.7 |






#17
2-Methylphenol
Concen: 40.721 ng
RT: 7.872 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:107 Resp: 78692

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 112.2 | 89.7 | 134.5 |
| 77 | 44.0 | 34.2 | 51.4 |
| 79 | 41.8 | 33.4 | 50.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#18
Hexachloroethane
Concen: 40.115 ng
RT: 8.496 min Scan# 945
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:117 Resp: 374849

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 94.5 | 76.6 | 115.0 |
| 201 | 105.8 | 85.4 | 128.2 |






#19
n-Nitroso-di-n-propylamine
Concen: 40.521 ng
RT: 8.237 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 70 Resp: 70952

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 57.6 | 46.5 | 69.7 |
| 101 | 10.2 | 7.9 | 11.9 |
| 130 | 24.5 | 19.3 | 28.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#20
3+4-Methylphenols
Concen: 40.415 ng
RT: 8.196 min Scan# 894
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:107 Resp: 1093973

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 89.8 | 69.5 | 109.5 |
| 77 | 30.5 | 11.4 | 51.4 |
| 79 | 27.0 | 6.9 | 46.9 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.184 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:136 Resp: 163760

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.9 | 8.7 | 13.1 |
| 54 | 6.8 | 5.4 | 8.2 |
| 68 | 5.5 | 4.5 | 6.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#22
Acetophenone
Concen: 38.992 ng
RT: 8.243 min Scan# 902
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:105 Resp: 1510122

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 6.1 | 5.4 | 8.2 |
| 51 | 23.8 | 19.6 | 29.4 |
| 120 | 23.3 | 18.8 | 28.2 |






#23
Nitrobenzene-d5
Concen: 82.866 ng
RT: 8.572 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 82 Resp: 204046

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 48.7 | 38.4 | 57.6 |
| 54 | 46.6 | 37.8 | 56.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#24
Nitrobenzene
Concen: 40.369 ng
RT: 8.613 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 77 Resp: 1012505

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 51.4 | 39.0 | 58.6 |
| 65 | 12.9 | 10.7 | 16.1 |






#25
Isophorone
Concen: 40.360 ng
RT: 9.125 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 82 Resp: 185726

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 6.9 | 5.8 | 8.6 |
| 138 | 19.6 | 15.7 | 23.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#26
2-Nitrophenol
Concen: 40.694 ng
RT: 9.308 min Scan# 1083
Delta R.T. -0.005 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:139 Resp: 488721

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 28.2 | 22.3 | 33.5 |
| 65 | 43.1 | 33.8 | 50.8 |






#27
2,4-Dimethylphenol
Concen: 40.332 ng
RT: 9.378 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:122 Resp: 69232

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 111.6 | 88.2 | 132.2 |
| 121 | 57.5 | 46.1 | 69.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#28
bis(2-Chloroethoxy)methane
Concen: 39.567 ng
RT: 9.625 min Scan# 1137
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 93 Resp: 1236294

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.3 | 26.4 | 39.6 |
| 123 | 11.9 | 10.1 | 15.1 |






#29
2,4-Dichlorophenol
Concen: 41.412 ng
RT: 9.837 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:162 Resp: 98570

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 65.4 | 44.1 | 84.1 |
| 98 | 33.4 | 13.1 | 53.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#30
1,2,4-Trichlorobenzene
Concen: 39.571 ng
RT: 10.060 min Scan# 1211
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:180 Resp: 1043043

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 97.1 | 75.1 | 112.7 |
| 145 | 29.4 | 23.3 | 34.9 |






#31
Naphthalene
Concen: 39.170 ng
RT: 10.231 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:128 Resp: 330294

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.0 | 8.9 | 13.3 |
| 127 | 13.0 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#32
Benzoic acid
Concen: 40.952 ng
RT: 9.502 min Scan# 1116
Delta R.T. 0.001 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:122 Resp: 684343

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 116.7 | 101.0 | 141.0 |
| 77 | 85.5 | 67.6 | 107.6 |






#33
4-Chloroaniline
Concen: 40.229 ng
RT: 10.343 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:127 Resp: 126595

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.2 | 25.4 | 38.2 |
| 65 | 26.4 | 21.0 | 31.6 |
| 92 | 17.4 | 13.4 | 20.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#34
Hexachlorobutadiene
Concen: 40.154 ng
RT: 10.537 min Scan# 1292
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:225 Resp: 607380

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 63.8 | 50.3 | 75.5 |
| 227 | 65.5 | 51.3 | 76.9 |






#35
Caprolactam
Concen: 40.864 ng
RT: 11.107 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:113 Resp: 33274

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 148.1 | 128.1 | 168.1 |
| 56 | 111.3 | 90.7 | 130.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#36
4-Chloro-3-methylphenol
Concen: 40.967 ng
RT: 11.484 min Scan# 1453
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:107 Resp: 1018998

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 29.0 | 22.4 | 33.6 |
| 142 | 89.3 | 69.6 | 104.4 |






#37
2-Methylnaphthalene
Concen: 39.345 ng
RT: 11.866 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:142 Resp: 234086

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 89.4 | 71.2 | 106.8 |
| 115 | 32.3 | 25.9 | 38.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#38
1-Methylnaphthalene
Concen: 39.301 ng
RT: 12.084 min Scan# 1555
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:142 Resp: 2280444

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 91.8 | 74.1 | 111.1 |
| 116 | 3.5 | 2.7 | 4.1 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.090 min Scan# 1
Delta R.T. 0.001 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:164 Resp: 107115

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 100.2 | 80.3 | 120.5 |
| 160 | 44.8 | 35.9 | 53.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.072 ng
RT: 12.243 min Scan# 1582
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:216 Resp: 1233612

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.1 | 62.4 | 93.6 |
| 179 | 20.8 | 16.5 | 24.7 |
| 108 | 16.9 | 13.7 | 20.5 |






#41
Hexachlorocyclopentadiene
Concen: 42.989 ng
RT: 12.237 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:237 Resp: 34568

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 62.3 | 41.6 | 81.6 |
| 272 | 12.9 | 0.0 | 33.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#42
2,4,6-Tribromophenol
Concen: 82.624 ng
RT: 15.619 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:330 Resp: 1045523

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 97.6 | 77.7 | 116.5 |
| 141 | 31.7 | 26.0 | 39.0 |






#43
2,4,6-Trichlorophenol
Concen: 42.646 ng
RT: 12.490 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:196 Resp: 79495

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.4 | 76.2 | 114.2 |
| 200 | 30.7 | 24.3 | 36.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#44
2,4,5-Trichlorophenol
Concen: 42.335 ng
RT: 12.560 min Scan# 1636
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:196 Resp: 894337

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 97.6 | 76.6 | 115.0 |
| 97 | 43.5 | 36.6 | 55.0 |
| 132 | 26.6 | 21.4 | 32.0 |






#45
2-Fluorobiphenyl
Concen: 78.712 ng
RT: 12.690 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:172 Resp: 544249

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.4 | 28.2 | 42.2 |
| 170 | 23.6 | 18.7 | 28.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#46
1,1'-Biphenyl
Concen: 39.032 ng
RT: 12.907 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:154 Resp: 3042013

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.2 | 20.2 | 60.2 |
| 76 | 11.9 | 0.0 | 31.5 |






#47
2-Chloronaphthalene
Concen: 39.210 ng
RT: 12.943 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:162 Resp: 229682

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 37.7 | 29.4 | 44.2 |
| 164 | 32.9 | 26.4 | 39.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#48
2-Nitroaniline
Concen: 39.894 ng
RT: 13.149 min Scan# 1736
Delta R.T. 0.007 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 65 Resp: 565340

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 74.7 | 58.7 | 88.1 |
| 138 | 128.8 | 99.0 | 148.4 |






#49
Acenaphthylene
Concen: 40.431 ng
RT: 13.807 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:152 Resp: 362482

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.0 | 16.1 | 24.1 |
| 153 | 13.2 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#50
Dimethylphthalate
Concen: 40.131 ng
RT: 13.560 min Scan# 1806
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:163 Resp: 3009077

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.2 | 3.4 | 5.0 |
| 164 | 10.3 | 8.0 | 12.0 |






#51
2,6-Dinitrotoluene
Concen: 42.919 ng
RT: 13.660 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:165 Resp: 62763

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 47.6 | 38.5 | 57.7 |
| 89 | 51.7 | 40.6 | 61.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#52
Acenaphthene
Concen: 39.318 ng m
RT: 14.160 min Scan# 1908
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:154 Resp: 2271995

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 113.1 | 90.6 | 136.0 |
| 152 | 54.0 | 43.8 | 65.8 |






#53
3-Nitroaniline
Concen: 43.418 ng
RT: 13.990 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:138 Resp: 68044

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 10.1 | 7.8 | 11.8 |
| 92 | 103.4 | 81.0 | 121.4 |

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#54
2,4-Dinitrophenol
Concen: 40.066 ng
RT: 14.201 min Scan# 1915
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:184 Resp: 265751

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 56.1 | 47.4 | 71.0 |
| 154 | 58.1 | 47.8 | 71.8 |






#55
Dibenzofuran
Concen: 39.220 ng
RT: 14.507 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:168 Resp: 374023

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 35.5 | 28.6 | 43.0 |
| 169 | 13.4 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#56
4-Nitrophenol
Concen: 42.369 ng
RT: 14.313 min Scan# 1934
Delta R.T. 0.012 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:139 Resp: 612774

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 67.3 | 45.9 | 85.9 |
| 65 | 84.2 | 61.4 | 101.4 |






#57
2,4-Dinitrotoluene
Concen: 39.932 ng
RT: 14.466 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:165 Resp: 84695

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 35.8 | 31.0 | 46.4 |
| 89 | 62.7 | 53.4 | 80.0 |
| 182 | 2.8 | 2.0 | 3.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#58
Fluorene
Concen: 39.662 ng
RT: 15.166 min Scan# 2079
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:166 Resp: 2925541

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 97.5 | 78.6 | 117.8 |
| 167 | 13.3 | 10.7 | 16.1 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 42.414 ng
RT: 14.737 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:232 Resp: 81140

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 43.6 | 35.7 | 53.5 |
| 130 | 2.2 | 1.8 | 2.6 |
| 166 | 30.4 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#60
Diethylphthalate
Concen: 40.587 ng
RT: 14.972 min Scan# 2046
Delta R.T. 0.012 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:149 Resp: 3056611

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 22.8 | 17.8 | 26.8 |
| 150 | 12.4 | 10.2 | 15.2 |






#61
4-Chlorophenyl-phenylether
Concen: 39.198 ng
RT: 15.178 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:204 Resp: 143088

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 33.2 | 25.8 | 38.6 |
| 141 | 55.2 | 45.1 | 67.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#62
4-Nitroaniline
Concen: 43.550 ng
RT: 15.184 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:138 Resp: 715671

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 44.5 | 26.0 | 66.0 |
| 108 | 67.8 | 50.0 | 90.0 |






#63
Azobenzene
Concen: 38.583 ng
RT: 15.472 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion: 77 Resp: 258741

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 28.5 | 7.7 | 47.7 |
| 105 | 22.0 | 1.9 | 41.9 |
| 51 | 27.7 | 8.2 | 48.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.907 min Scan# 2375
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:188 Resp: 2157756

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 8.7 | 7.3 | 10.9 |
| 80 | 9.3 | 7.8 | 11.6 |






#65
4,6-Dinitro-2-methylphenol
Concen: 39.754 ng
RT: 15.248 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

Tgt Ion:198 Resp: 37970

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 39.5 | 16.8 | 56.8 |
| 105 | 41.3 | 20.8 | 60.8 |

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025







#66
n-Nitrosodiphenylamine
Concen: 40.039 ng
RT: 15.401 min Scan# 2119
Delta R.T. 0.012 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:169 Resp: 2472464

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 66.7 | 54.1 | 81.1 |
| 167 | 36.2 | 29.0 | 43.6 |






#67
4-Bromophenyl-phenylether
Concen: 40.486 ng
RT: 16.095 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:248 Resp: 89445

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 98.5 | 79.4 | 119.2 |
| 141 | 71.0 | 51.4 | 77.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#68
Hexachlorobenzene
Concen: 39.499 ng
RT: 16.201 min Scan# 2255
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:284 Resp: 1066631

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 32.0 | 25.9 | 38.9 |
| 249 | 30.5 | 23.8 | 35.8 |






#69
Atrazine
Concen: 40.888 ng
RT: 16.390 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:200 Resp: 73988

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 26.2 | 6.6 | 46.6 |
| 215 | 49.4 | 28.0 | 68.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#70
Pentachlorophenol
Concen: 42.471 ng
RT: 16.548 min Scan# 2314
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:266 Resp: 774549

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 65.2 | 50.4 | 75.6 |
| 264 | 63.5 | 50.2 | 75.2 |






#71
Phenanthrene
Concen: 39.487 ng
RT: 16.948 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:178 Resp: 455272

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.5 | 15.4 | 23.2 |
| 179 | 15.8 | 12.6 | 19.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#72
Anthracene
Concen: 40.385 ng
RT: 17.060 min Scan# 2401
Delta R.T. 0.012 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:178 Resp: 4467088

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.5 | 15.0 | 22.6 |
| 179 | 15.3 | 12.5 | 18.7 |






#73
Carbazole
Concen: 40.086 ng
RT: 17.331 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:167 Resp: 433539

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.5 | 17.4 | 26.0 |
| 139 | 12.7 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#74
Di-n-butylphthalate
Concen: 42.761 ng
RT: 17.966 min Scan# 2555
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:149 Resp: 5149955

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.3 | 7.3 | 10.9 |
| 104 | 4.8 | 3.8 | 5.6 |






#75
Fluoranthene
Concen: 40.180 ng
RT: 19.060 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:202 Resp: 541510

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 11.1 | 0.0 | 30.4 |
| 203 | 17.8 | 0.0 | 37.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.354 min Scan# 3131
Delta R.T. 0.012 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:240 Resp: 2221808

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.2 | 9.0 | 13.4 |
| 236 | 25.9 | 20.3 | 30.5 |






#77
Benzidine
Concen: 52.461 ng
RT: 19.272 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:184 Resp: 205856

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 15.4 | 12.6 | 18.8 |
| 183 | 11.6 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#78
Pyrene
Concen: 41.413 ng
RT: 19.436 min Scan# 2805
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:202 Resp: 5698991

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 21.5 | 17.1 | 25.7 |
| 203 | 18.0 | 14.1 | 21.1 |






#79
Terphenyl-d14
Concen: 84.014 ng
RT: 19.689 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:244 Resp: 849326

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 7.1 | 5.9 | 8.9 |
| 122 | 11.1 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#80
Butylbenzylphthalate
Concen: 40.921 ng
RT: 20.442 min Scan# 2976
Delta R.T. 0.012 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:149 Resp: 2138437

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 65.5 | 50.6 | 76.0 |
| 206 | 20.5 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 40.856 ng
RT: 21.336 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:228 Resp: 550107

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 27.8 | 22.1 | 33.1 |
| 229 | 20.2 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#82
3,3'-Dichlorobenzidine
Concen: 44.629 ng
RT: 21.266 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:252 Resp: 1808957

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 65.0 | 51.9 | 77.9 |
| 126 | 10.8 | 8.1 | 12.1 |






#83
Chrysene
Concen: 40.218 ng
RT: 21.401 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:228 Resp: 521127

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 30.0 | 24.4 | 36.6 |
| 229 | 19.7 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 45.685 ng
RT: 21.342 min Scan# 3129
Delta R.T. 0.012 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:149 Resp: 3299494

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 27.8 | 21.9 | 32.9 |
| 279 | 4.4 | 3.5 | 5.3 |






#85
Di-n-octyl phthalate
Concen: 40.750 ng
RT: 22.566 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:149 Resp: 525667

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.5 | 1.3 | 1.9 |
| 43 | 8.5 | 6.9 | 10.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 24.501 min Scan# 3666
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:264 Resp: 2415004

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.4 | 18.6 | 28.0 |
| 265 | 22.1 | 17.7 | 26.5 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 42.211 ng m
RT: 28.048 min Scan# 4
Delta R.T. 0.012 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:276 Resp: 649415

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 16.4 | 13.0 | 19.6 |
| 277 | 19.8 | 15.9 | 23.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#88
Benzo(b)fluoranthene
Concen: 41.137 ng
RT: 23.519 min Scan# 3499
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:252 Resp: 5782492

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.0 | 17.7 | 26.5 |
| 125 | 10.2 | 8.1 | 12.1 |






#89
Benzo(k)fluoranthene
Concen: 39.974 ng
RT: 23.589 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:252 Resp: 559080

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.2 | 17.7 | 26.5 |
| 125 | 10.3 | 8.2 | 12.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#90
Benzo(a)pyrene
Concen: 41.397 ng
RT: 24.354 min Scan# 3641
Delta R.T. 0.000 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:252 Resp: 4952570

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.0 | 17.6 | 26.4 |
| 125 | 11.5 | 9.0 | 13.4 |






#91
Dibenzo(a,h)anthracene
Concen: 42.422 ng
RT: 28.130 min Scan# 4
Delta R.T. -0.006 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:278 Resp: 539484

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 15.9 | 13.1 | 19.7 |
| 279 | 23.8 | 18.9 | 28.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/07/2025
Supervised By :Jagrut Upadhyay 07/07/2025




#92
Benzo(g,h,i)perylene
Concen: 41.855 ng
RT: 29.148 min Scan# 4456
Delta R.T. 0.024 min
Lab File: BP025079.D
Acq: 03 Jul 2025 16:18

Tgt Ion:276 Resp: 5498609

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.9 | 19.3 | 28.9 |
| 138 | 21.9 | 16.9 | 25.3 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025079.D
Acq On : 03 Jul 2025 16:18
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

Quant Time: Jul 03 16:37:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 1 | I | 1,4-Dichlorobenzene-d4 | 1.000 | 1.000 | 0.0 | 103 | 0.00 |
| 2 | | 1,4-Dioxane | 0.439 | 0.415 | 5.5 | 98 | 0.00 |
| 3 | | Pyridine | 1.202 | 1.172 | 2.5 | 98 | 0.00 |
| 4 | | n-Nitrosodimethylamine | 0.509 | 0.495 | 2.8 | 100 | 0.00 |
| 5 | S | 2-Fluorophenol | 1.129 | 1.126 | 0.3 | 102 | 0.00 |
| 6 | | Aniline | 1.376 | 1.369 | 0.5 | 101 | 0.00 |
| 7 | S | Phenol-d6 | 1.473 | 1.474 | -0.1 | 104 | 0.00 |
| 8 | | 2-Chlorophenol | 1.263 | 1.277 | -1.1 | 104 | 0.00 |
| 9 | | Benzaldehyde | 0.769 | 0.785 | -2.1 | 111 | 0.00 |
| 10 | C | Phenol | 1.497 | 1.483 | 0.9 | 103 | 0.00 |
| 11 | | bis(2-Chloroethyl)ether | 1.152 | 1.129 | 2.0 | 103 | 0.00 |
| 12 | | 1,3-Dichlorobenzene | 1.413 | 1.392 | 1.5 | 104 | 0.00 |
| 13 | C | 1,4-Dichlorobenzene | 1.431 | 1.409 | 1.5 | 103 | 0.00 |
| 14 | | 1,2-Dichlorobenzene | 1.377 | 1.341 | 2.6 | 102 | 0.00 |
| 15 | | Benzyl Alcohol | 0.983 | 1.004 | -2.1 | 108 | 0.00 |
| 16 | | 2,2'-oxybis(1-Chloropropane | 1.498 | 1.445 | 3.5 | 103 | 0.00 |
| 17 | | 2-Methylphenol | 0.964 | 0.981 | -1.8 | 104 | 0.00 |
| 18 | | Hexachloroethane | 0.466 | 0.468 | -0.4 | 104 | 0.00 |
| 19 | P | n-Nitroso-di-n-propylamine | 0.874 | 0.885 | -1.3 | 105 | 0.00 |
| 20 | | 3+4-Methylphenols | 1.350 | 1.364 | -1.0 | 105 | 0.00 |
| 21 | I | Naphthalene-d8 | 1.000 | 1.000 | 0.0 | 105 | 0.00 |
| 22 | | Acetophenone | 0.473 | 0.461 | 2.5 | 104 | 0.00 |
| 23 | S | Nitrobenzene-d5 | 0.301 | 0.312 | -3.7 | 107 | 0.00 |
| 24 | | Nitrobenzene | 0.306 | 0.309 | -1.0 | 105 | 0.00 |
| 25 | | Isophorone | 0.562 | 0.567 | -0.9 | 107 | 0.00 |
| 26 | C | 2-Nitrophenol | 0.128 | 0.149 | -16.4 | 117 | 0.00 |
| 27 | | 2,4-Dimethylphenol | 0.210 | 0.211 | -0.5 | 107 | 0.00 |
| 28 | | bis(2-Chloroethoxy)methane | 0.382 | 0.377 | 1.3 | 106 | 0.00 |
| 29 | C | 2,4-Dichlorophenol | 0.291 | 0.301 | -3.4 | 106 | 0.00 |
| 30 | | 1,2,4-Trichlorobenzene | 0.322 | 0.318 | 1.2 | 105 | 0.00 |
| 31 | | Naphthalene | 1.030 | 1.008 | 2.1 | 105 | 0.00 |
| 32 | | Benzoic acid | 0.185 | 0.209 | -13.0 | 118 | 0.00 |
| 33 | | 4-Chloroaniline | 0.384 | 0.387 | -0.8 | 104 | 0.00 |
| 34 | C | Hexachlorobutadiene | 0.185 | 0.185 | 0.0 | 107 | 0.00 |
| 35 | | Caprolactam | 0.099 | 0.102 | -3.0 | 106 | 0.00 |
| 36 | C | 4-Chloro-3-methylphenol | 0.304 | 0.311 | -2.3 | 106 | 0.00 |
| 37 | | 2-Methylnaphthalene | 0.727 | 0.715 | 1.7 | 104 | 0.00 |
| 38 | | 1-Methylnaphthalene | 0.709 | 0.696 | 1.8 | 106 | 0.00 |
| 39 | I | Acenaphthene-d10 | 1.000 | 1.000 | 0.0 | 107 | 0.00 |
| 40 | | 1,2,4,5-Tetrachlorobenzene | 0.575 | 0.576 | -0.2 | 106 | 0.00 |
| 41 | P | Hexachlorocyclopentadiene | 0.150 | 0.161 | -7.3 | 116 | 0.00 |
| 42 | S | 2,4,6-Tribromophenol | 0.236 | 0.244 | -3.4 | 106 | 0.00 |
| 43 | C | 2,4,6-Trichlorophenol | 0.348 | 0.371 | -6.6 | 107 | 0.00 |
| 44 | | 2,4,5-Trichlorophenol | 0.394 | 0.417 | -5.8 | 108 | 0.00 |
| 45 | S | 2-Fluorobiphenyl | 1.291 | 1.270 | 1.6 | 106 | 0.00 |
| 46 | | 1,1'-Biphenyl | 1.455 | 1.420 | 2.4 | 104 | 0.00 |
| 47 | | 2-Chloronaphthalene | 1.094 | 1.072 | 2.0 | 105 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025079.D
Acq On : 03 Jul 2025 16:18
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

Quant Time: Jul 03 16:37:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 48 | | 2-Nitroaniline | 0.237 | 0.264 | -11.4 | 109 | 0.00 |
| 49 | | Acenaphthylene | 1.674 | 1.692 | -1.1 | 106 | 0.00 |
| 50 | | Dimethylphthalate | 1.400 | 1.405 | -0.4 | 107 | 0.00 |
| 51 | | 2,6-Dinitrotoluene | 0.273 | 0.293 | -7.3 | 106 | 0.00 |
| 52 | C | Acenaphthene | 1.079 | 1.061 | 1.7 | 106 | 0.00 |
| 53 | | 3-Nitroaniline | 0.293 | 0.318 | -8.5 | 106 | 0.00 |
| 54 | P | 2,4-Dinitrophenol | 0.114 | 0.124 | -8.8 | 113 | 0.00 |
| 55 | | Dibenzofuran | 1.781 | 1.746 | 2.0 | 105 | 0.00 |
| 56 | P | 4-Nitrophenol | 0.270 | 0.286 | -5.9 | 107 | 0.01 |
| 57 | | 2,4-Dinitrotoluene | 0.356 | 0.395 | -11.0 | 108 | 0.00 |
| 58 | | Fluorene | 1.377 | 1.366 | 0.8 | 105 | 0.00 |
| 59 | | 2,3,4,6-Tetrachlorophenol | 0.357 | 0.379 | -6.2 | 109 | 0.00 |
| 60 | | Diethylphthalate | 1.406 | 1.427 | -1.5 | 107 | 0.01 |
| 61 | | 4-Chlorophenyl-phenylether | 0.682 | 0.668 | 2.1 | 105 | 0.00 |
| 62 | | 4-Nitroaniline | 0.307 | 0.334 | -8.8 | 104 | 0.00 |
| 63 | | Azobenzene | 1.252 | 1.208 | 3.5 | 102 | 0.00 |
| 64 | I | Phenanthrene-d10 | 1.000 | 1.000 | 0.0 | 105 | 0.00 |
| 65 | | 4,6-Dinitro-2-methylphenol | 0.082 | 0.088 | -7.3 | 113 | 0.00 |
| 66 | c | n-Nitrosodiphenylamine | 0.572 | 0.573 | -0.2 | 105 | 0.01 |
| 67 | | 4-Bromophenyl-phenylether | 0.205 | 0.207 | -1.0 | 106 | 0.00 |
| 68 | | Hexachlorobenzene | 0.250 | 0.247 | 1.2 | 104 | 0.00 |
| 69 | | Atrazine | 0.168 | 0.171 | -1.8 | 104 | 0.00 |
| 70 | C | Pentachlorophenol | 0.169 | 0.179 | -5.9 | 109 | 0.00 |
| 71 | | Phenanthrene | 1.069 | 1.055 | 1.3 | 103 | 0.00 |
| 72 | | Anthracene | 1.025 | 1.035 | -1.0 | 104 | 0.01 |
| 73 | | Carbazole | 1.002 | 1.005 | -0.3 | 102 | 0.00 |
| 74 | | Di-n-butylphthalate | 1.116 | 1.193 | -6.9 | 105 | 0.00 |
| 75 | C | Fluoranthene | 1.249 | 1.255 | -0.5 | 102 | 0.00 |
| 76 | I | Chrysene-d12 | 1.000 | 1.000 | 0.0 | 98 | 0.01 |
| 77 | | Benzidine | 0.353 | 0.463 | -31.2# | 80 | 0.00 |
| 78 | | Pyrene | 1.239 | 1.283 | -3.6 | 102 | 0.00 |
| 79 | S | Terphenyl-d14 | 0.910 | 0.956 | -5.1 | 103 | 0.02 |
| 80 | | Butylbenzylphthalate | 0.402 | 0.481 | -19.7 | 107 | 0.01 |
| 81 | | Benzo(a)anthracene | 1.212 | 1.238 | -2.1 | 100 | 0.01 |
| 82 | | 3,3'-Dichlorobenzidine | 0.365 | 0.407 | -11.5 | 103 | 0.00 |
| 83 | | Chrysene | 1.166 | 1.173 | -0.6 | 99 | 0.01 |
| 84 | | Bis(2-ethylhexyl)phthalate | 0.650 | 0.743 | -14.3 | 104 | 0.01 |
| 85 | c | Di-n-octyl phthalate | 1.054 | 1.183 | -12.2 | 106 | 0.00 |
| 86 | I | Perylene-d12 | 1.000 | 1.000 | 0.0 | 99 | 0.00 |
| 87 | | Indeno(1,2,3-cd)pyrene | 1.274 | 1.345 | -5.6 | 100 | 0.01 |
| 88 | | Benzo(b)fluoranthene | 1.164 | 1.197 | -2.8 | 99 | 0.00 |
| 89 | | Benzo(k)fluoranthene | 1.158 | 1.158 | 0.0 | 100 | 0.00 |
| 90 | C | Benzo(a)pyrene | 0.991 | 1.025 | -3.4 | 99 | 0.00 |
| 91 | | Dibenzo(a,h)anthracene | 1.053 | 1.117 | -6.1 | 101 | 0.00 |
| 92 | | Benzo(g,h,i)perylene | 1.088 | 1.138 | -4.6 | 100 | 0.02 |

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025079.D
Acq On : 03 Jul 2025 16:18
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

Quant Time: Jul 03 16:37:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | AvgRF | CCRF | %Dev Area% | Dev(min) |
|---|---|---|---|---|

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025079.D
Acq On : 03 Jul 2025 16:18
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

Quant Time: Jul 03 16:37:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 1 | I | 1,4-Dichlorobenzene-d4 | 20.000 | 20.000 | 0.0 | 103 | 0.00 |
| 2 | | 1,4-Dioxane | 40.000 | 37.814 | 5.5 | 98 | 0.00 |
| 3 | | Pyridine | 40.000 | 39.027 | 2.4 | 98 | 0.00 |
| 4 | | n-Nitrosodimethylamine | 40.000 | 38.933 | 2.7 | 100 | 0.00 |
| 5 | S | 2-Fluorophenol | 80.000 | 79.825 | 0.2 | 102 | 0.00 |
| 6 | | Aniline | 40.000 | 39.803 | 0.5 | 101 | 0.00 |
| 7 | S | Phenol-d6 | 80.000 | 80.057 | -0.1 | 104 | 0.00 |
| 8 | | 2-Chlorophenol | 40.000 | 40.431 | -1.1 | 104 | 0.00 |
| 9 | | Benzaldehyde | 40.000 | 40.841 | -2.1 | 111 | 0.00 |
| 10 | C | Phenol | 40.000 | 39.645 | 0.9 | 103 | 0.00 |
| 11 | | bis(2-Chloroethyl)ether | 40.000 | 39.206 | 2.0 | 103 | 0.00 |
| 12 | | 1,3-Dichlorobenzene | 40.000 | 39.389 | 1.5 | 104 | 0.00 |
| 13 | C | 1,4-Dichlorobenzene | 40.000 | 39.387 | 1.5 | 103 | 0.00 |
| 14 | | 1,2-Dichlorobenzene | 40.000 | 38.965 | 2.6 | 102 | 0.00 |
| 15 | | Benzyl Alcohol | 40.000 | 40.867 | -2.2 | 108 | 0.00 |
| 16 | | 2,2'-oxybis(1-Chloropropane | 40.000 | 38.579 | 3.6 | 103 | 0.00 |
| 17 | | 2-Methylphenol | 40.000 | 40.721 | -1.8 | 104 | 0.00 |
| 18 | | Hexachloroethane | 40.000 | 40.115 | -0.3 | 104 | 0.00 |
| 19 | P | n-Nitroso-di-n-propylamine | 40.000 | 40.521 | -1.3 | 105 | 0.00 |
| 20 | | 3+4-Methylphenols | 40.000 | 40.415 | -1.0 | 105 | 0.00 |
| 21 | I | Naphthalene-d8 | 20.000 | 20.000 | 0.0 | 105 | 0.00 |
| 22 | | Acetophenone | 40.000 | 38.992 | 2.5 | 104 | 0.00 |
| 23 | S | Nitrobenzene-d5 | 80.000 | 82.866 | -3.6 | 107 | 0.00 |
| 24 | | Nitrobenzene | 40.000 | 40.369 | -0.9 | 105 | 0.00 |
| 25 | | Isophorone | 40.000 | 40.360 | -0.9 | 107 | 0.00 |
| 26 | C | 2-Nitrophenol | 40.000 | 40.694 | -1.7 | 117 | 0.00 |
| 27 | | 2,4-Dimethylphenol | 40.000 | 40.332 | -0.8 | 107 | 0.00 |
| 28 | | bis(2-Chloroethoxy)methane | 40.000 | 39.567 | 1.1 | 106 | 0.00 |
| 29 | C | 2,4-Dichlorophenol | 40.000 | 41.412 | -3.5 | 106 | 0.00 |
| 30 | | 1,2,4-Trichlorobenzene | 40.000 | 39.571 | 1.1 | 105 | 0.00 |
| 31 | | Naphthalene | 40.000 | 39.170 | 2.1 | 105 | 0.00 |
| 32 | | Benzoic acid | 40.000 | 40.952 | -2.4 | 118 | 0.00 |
| 33 | | 4-Chloroaniline | 40.000 | 40.229 | -0.6 | 104 | 0.00 |
| 34 | C | Hexachlorobutadiene | 40.000 | 40.154 | -0.4 | 107 | 0.00 |
| 35 | | Caprolactam | 40.000 | 40.864 | -2.2 | 106 | 0.00 |
| 36 | C | 4-Chloro-3-methylphenol | 40.000 | 40.967 | -2.4 | 106 | 0.00 |
| 37 | | 2-Methylnaphthalene | 40.000 | 39.345 | 1.6 | 104 | 0.00 |
| 38 | | 1-Methylnaphthalene | 40.000 | 39.301 | 1.7 | 106 | 0.00 |
| 39 | I | Acenaphthene-d10 | 20.000 | 20.000 | 0.0 | 107 | 0.00 |
| 40 | | 1,2,4,5-Tetrachlorobenzene | 40.000 | 40.072 | -0.2 | 106 | 0.00 |
| 41 | P | Hexachlorocyclopentadiene | 40.000 | 42.989 | -7.5 | 116 | 0.00 |
| 42 | S | 2,4,6-Tribromophenol | 80.000 | 82.624 | -3.3 | 106 | 0.00 |
| 43 | C | 2,4,6-Trichlorophenol | 40.000 | 42.646 | -6.6 | 107 | 0.00 |
| 44 | | 2,4,5-Trichlorophenol | 40.000 | 42.335 | -5.8 | 108 | 0.00 |
| 45 | S | 2-Fluorobiphenyl | 80.000 | 78.712 | 1.6 | 106 | 0.00 |
| 46 | | 1,1'-Biphenyl | 40.000 | 39.032 | 2.4 | 104 | 0.00 |
| 47 | | 2-Chloronaphthalene | 40.000 | 39.210 | 2.0 | 105 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025079.D
Acq On : 03 Jul 2025 16:18
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

Quant Time: Jul 03 16:37:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 48 | | 2-Nitroaniline | 40.000 | 39.894 | 0.3 | 109 | 0.00 |
| 49 | | Acenaphthylene | 40.000 | 40.431 | -1.1 | 106 | 0.00 |
| 50 | | Dimethylphthalate | 40.000 | 40.131 | -0.3 | 107 | 0.00 |
| 51 | | 2,6-Dinitrotoluene | 40.000 | 42.919 | -7.3 | 106 | 0.00 |
| 52 | C | Acenaphthene | 40.000 | 39.318 | 1.7 | 106 | 0.00 |
| 53 | | 3-Nitroaniline | 40.000 | 43.418 | -8.5 | 106 | 0.00 |
| 54 | P | 2,4-Dinitrophenol | 40.000 | 40.066 | -0.2 | 113 | 0.00 |
| 55 | | Dibenzofuran | 40.000 | 39.220 | 2.0 | 105 | 0.00 |
| 56 | P | 4-Nitrophenol | 40.000 | 42.369 | -5.9 | 107 | 0.01 |
| 57 | | 2,4-Dinitrotoluene | 40.000 | 39.932 | 0.2 | 108 | 0.00 |
| 58 | | Fluorene | 40.000 | 39.662 | 0.8 | 105 | 0.00 |
| 59 | | 2,3,4,6-Tetrachlorophenol | 40.000 | 42.414 | -6.0 | 109 | 0.00 |
| 60 | | Diethylphthalate | 40.000 | 40.587 | -1.5 | 107 | 0.01 |
| 61 | | 4-Chlorophenyl-phenylether | 40.000 | 39.198 | 2.0 | 105 | 0.00 |
| 62 | | 4-Nitroaniline | 40.000 | 43.550 | -8.9 | 104 | 0.00 |
| 63 | | Azobenzene | 40.000 | 38.583 | 3.5 | 102 | 0.00 |
| 64 | I | Phenanthrene-d10 | 20.000 | 20.000 | 0.0 | 105 | 0.00 |
| 65 | | 4,6-Dinitro-2-methylphenol | 40.000 | 39.754 | 0.6 | 113 | 0.00 |
| 66 | c | n-Nitrosodiphenylamine | 40.000 | 40.039 | -0.1 | 105 | 0.01 |
| 67 | | 4-Bromophenyl-phenylether | 40.000 | 40.486 | -1.2 | 106 | 0.00 |
| 68 | | Hexachlorobenzene | 40.000 | 39.499 | 1.3 | 104 | 0.00 |
| 69 | | Atrazine | 40.000 | 40.888 | -2.2 | 104 | 0.00 |
| 70 | C | Pentachlorophenol | 40.000 | 42.471 | -6.2 | 109 | 0.00 |
| 71 | | Phenanthrene | 40.000 | 39.487 | 1.3 | 103 | 0.00 |
| 72 | | Anthracene | 40.000 | 40.385 | -1.0 | 104 | 0.01 |
| 73 | | Carbazole | 40.000 | 40.086 | -0.2 | 102 | 0.00 |
| 74 | | Di-n-butylphthalate | 40.000 | 42.761 | -6.9 | 105 | 0.00 |
| 75 | C | Fluoranthene | 40.000 | 40.180 | -0.4 | 102 | 0.00 |
| 76 | I | Chrysene-d12 | 20.000 | 20.000 | 0.0 | 98 | 0.01 |
| 77 | | Benzidine | 40.000 | 52.461 | -31.2# | 80 | 0.00 |
| 78 | | Pyrene | 40.000 | 41.413 | -3.5 | 102 | 0.00 |
| 79 | S | Terphenyl-d14 | 80.000 | 84.014 | -5.0 | 103 | 0.02 |
| 80 | | Butylbenzylphthalate | 40.000 | 40.921 | -2.3 | 107 | 0.01 |
| 81 | | Benzo(a)anthracene | 40.000 | 40.856 | -2.1 | 100 | 0.01 |
| 82 | | 3,3'-Dichlorobenzidine | 40.000 | 44.629 | -11.6 | 103 | 0.00 |
| 83 | | Chrysene | 40.000 | 40.218 | -0.5 | 99 | 0.01 |
| 84 | | Bis(2-ethylhexyl)phthalate | 40.000 | 45.685 | -14.2 | 104 | 0.01 |
| 85 | c | Di-n-octyl phthalate | 40.000 | 40.750 | -1.9 | 106 | 0.00 |
| 86 | I | Perylene-d12 | 20.000 | 20.000 | 0.0 | 99 | 0.00 |
| 87 | | Indeno(1,2,3-cd)pyrene | 40.000 | 42.211 | -5.5 | 100 | 0.01 |
| 88 | | Benzo(b)fluoranthene | 40.000 | 41.137 | -2.8 | 99 | 0.00 |
| 89 | | Benzo(k)fluoranthene | 40.000 | 39.974 | 0.1 | 100 | 0.00 |
| 90 | C | Benzo(a)pyrene | 40.000 | 41.397 | -3.5 | 99 | 0.00 |
| 91 | | Dibenzo(a,h)anthracene | 40.000 | 42.422 | -6.1 | 101 | 0.00 |
| 92 | | Benzo(g,h,i)perylene | 40.000 | 41.855 | -4.6 | 100 | 0.02 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025079.D
Acq On : 03 Jul 2025 16:18
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
ICVBP070325

Quant Time: Jul 03 16:37:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: PARS02

Lab Code: ACE SDG No.: Q2592

Instrument ID: BNA_F Calibration Date/Time: 07/16/2025 19:39

Lab File ID: BF143117.D Init. Calib. Date(s): 07/15/2025 07/15/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 13:04 16:35

GC Column: DB-UI ID: 0.18 (mm)

| COMPOUND | RRF | RRF040 | MIN RRF | %D | MAX%D |
|---|---|---|---|---|---|
| 2-Fluorophenol | 1.266 | 1.249 | | -1.3 | |
| Benzaldehyde | 1.212 | 1.240 | | 2.3 | |
| Phenol-d6 | 1.591 | 1.563 | | -1.8 | |
| Phenol | 1.723 | 1.816 | | 5.4 | 20.0 |
| bis(2-Chloroethyl)ether | 1.334 | 1.298 | | -2.7 | |
| 2-Chlorophenol | 1.272 | 1.316 | | 3.5 | |
| 2-Methylphenol | 1.098 | 1.094 | | -0.4 | |
| 2,2-oxybis(1-Chloropropane) | 2.458 | 2.389 | | -2.8 | |
| Acetophenone | 0.482 | 0.467 | | -3.1 | |
| 3+4-Methylphenols | 1.346 | 1.368 | | 1.6 | |
| n-Nitroso-di-n-propylamine | 0.975 | 0.965 | 0.050 | -1.0 | |
| Nitrobenzene-d5 | 0.357 | 0.402 | | 12.6 | |
| Hexachloroethane | 0.471 | 0.486 | | 3.2 | |
| Nitrobenzene | 0.344 | 0.369 | | 7.3 | |
| Isophorone | 0.689 | 0.675 | | -2.0 | |
| 2-Nitrophenol | 0.114 | 0.145 | | 27.2 | 20.0 |
| 2,4-Dimethylphenol | 0.320 | 0.320 | | 0.0 | |
| bis(2-Chloroethoxy)methane | 0.418 | 0.412 | | -1.4 | |
| 2,4-Dichlorophenol | 0.262 | 0.269 | | 2.7 | 20.0 |
| Naphthalene | 0.993 | 0.972 | | -2.1 | |
| 4-Chloroaniline | 0.399 | 0.393 | | -1.5 | |
| Hexachlorobutadiene | 0.178 | 0.178 | | 0.0 | 20.0 |
| Caprolactam | 0.082 | 0.084 | | 2.4 | |
| 4-Chloro-3-methylphenol | 0.291 | 0.293 | | 0.7 | 20.0 |
| 2-Methylnaphthalene | 0.592 | 0.578 | | -2.4 | |
| Hexachlorocyclopentadiene | 0.323 | 0.341 | 0.050 | 5.6 | |
| 2,4,6-Trichlorophenol | 0.362 | 0.386 | | 6.6 | 20.0 |
| 2-Fluorobiphenyl | 1.505 | 1.421 | | -5.6 | |
| 2,4,5-Trichlorophenol | 0.380 | 0.387 | | 1.8 | |
| 1,1-Biphenyl | 1.595 | 1.534 | | -3.8 | |
| 2-Chloronaphthalene | 1.172 | 1.129 | | -3.7 | |
| 2-Nitroaniline | 0.303 | 0.352 | | 16.2 | |
| Dimethylphthalate | 1.258 | 1.236 | | -1.7 | |
| Acenaphthylene | 1.952 | 1.912 | | -2.0 | |
| 2,6-Dinitrotoluene | 0.219 | 0.258 | | 17.8 | |
| 3-Nitroaniline | 0.272 | 0.299 | | 9.9 | |
| Acenaphthene | 1.152 | 1.115 | | -3.2 | 20.0 |
| 2,4-Dinitrophenol | 0.073 | 0.082 | 0.050 | 12.3 | |
| 4-Nitrophenol | 0.216 | 0.230 | 0.050 | 6.5 | |

Form VII SV-1

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

## SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: PARS02

Lab Code: ACE SDG No.: Q2592

Instrument ID: BNA_F Calibration Date/Time: 07/16/2025 19:39

Lab File ID: BF143117.D Init. Calib. Date(s): 07/15/2025 07/15/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 13:04 16:35

GC Column: DB-UI ID: 0.18 (mm)

| COMPOUND | RRF | RRF040 | MIN RRF | %D | MAX%D |
|---|---|---|---|---|---|
| Dibenzofuran | 1.717 | 1.649 | | -4.0 | |
| 2,4-Dinitrotoluene | 0.265 | 0.307 | | 15.8 | |
| Diethylphthalate | 1.214 | 1.211 | | -0.2 | |
| 4-Chlorophenyl-phenylether | 0.617 | 0.603 | | -2.3 | |
| Fluorene | 1.288 | 1.219 | | -5.4 | |
| 4-Nitroaniline | 0.236 | 0.255 | | 8.1 | |
| 4,6-Dinitro-2-methylphenol | 0.067 | 0.078 | | 16.4 | |
| n-Nitrosodiphenylamine | 0.703 | 0.687 | | -2.3 | 20.0 |
| 2,4,6-Tribromophenol | 0.178 | 0.191 | | 7.3 | |
| 4-Bromophenyl-phenylether | 0.227 | 0.227 | | 0.0 | |
| Hexachlorobenzene | 0.241 | 0.241 | | 0.0 | |
| Atrazine | 0.180 | 0.183 | | 1.7 | |
| Pentachlorophenol | 0.132 | 0.144 | | 9.1 | 20.0 |
| Phenanthrene | 1.079 | 1.046 | | -3.1 | |
| Anthracene | 1.085 | 1.063 | | -2.0 | |
| Carbazole | 0.972 | 0.934 | | -3.9 | |
| Di-n-butylphthalate | 0.893 | 0.937 | | 4.9 | |
| Fluoranthene | 0.981 | 0.947 | | -3.5 | 20.0 |
| Pyrene | 1.856 | 1.773 | | -4.5 | |
| Terphenyl-d14 | 1.368 | 1.302 | | -4.8 | |
| Butylbenzylphthalate | 0.318 | 0.365 | | 14.8 | |
| 3,3-Dichlorobenzidine | 0.387 | 0.391 | | 1.0 | |
| Benzo(a)anthracene | 1.333 | 1.354 | | 1.6 | |
| Chrysene | 1.227 | 1.156 | | -5.8 | |
| Bis(2-ethylhexyl)phthalate | 0.480 | 0.572 | | 19.2 | |
| Di-n-octyl phthalate | 0.927 | 1.043 | | 12.5 | 20.0 |
| Benzo(b)fluoranthene | 1.154 | 1.136 | | -1.6 | |
| Benzo(k)fluoranthene | 1.101 | 1.092 | | -0.8 | |
| Benzo(a)pyrene | 1.091 | 1.116 | | 2.3 | 20.0 |
| Indeno(1,2,3-cd)pyrene | 1.488 | 1.529 | | 2.8 | |
| Dibenzo(a,h)anthracene | 1.216 | 1.243 | | 2.2 | |
| Benzo(g,h,i)perylene | 1.240 | 1.256 | | 1.3 | |
| 1,2,4,5-Tetrachlorobenzene | 0.583 | 0.567 | | -2.7 | |
| 1,4-Dioxane | 0.630 | 0.628 | | -0.3 | 20.0 |
| 2,3,4,6-Tetrachlorophenol | 0.302 | 0.324 | | 7.3 | |

All other compounds must meet a minimum RRF of 0.010.

Form VII SV-1

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143117.D
Acq On : 16 Jul 2025 19:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Quant Time: Jul 17 03:44:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| Internal Standards | | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 6.969 | 152 | 135149 | 20.000 | ng | | 0.00 |
| 21) Naphthalene-d8 | 8.251 | 136 | 510164 | 20.000 | ng | | 0.00 |
| 39) Acenaphthene-d10 | 10.004 | 164 | 254065 | 20.000 | ng | | 0.00 |
| 64) Phenanthrene-d10 | 11.492 | 188 | 397173 | 20.000 | ng | | 0.00 |
| 76) Chrysene-d12 | 14.127 | 240 | 210714 | 20.000 | ng | | 0.00 |
| 86) Perylene-d12 | 15.633 | 264 | 255524 | 20.000 | ng | | 0.00 |
| System Monitoring Compounds | | | | | | | |
| 5) 2-Fluorophenol | 5.587 | 112 | 675027 | 78.913 | ng | | 0.00 |
| 7) Phenol-d6 | 6.587 | 99 | 845178 | 78.603 | ng | | 0.00 |
| 23) Nitrobenzene-d5 | 7.528 | 82 | 820016 | 90.170 | ng | | 0.00 |
| 42) 2,4,6-Tribromophenol | 10.792 | 330 | 193992 | 85.634 | ng | | 0.00 |
| 45) 2-Fluorobiphenyl | 9.322 | 172 | 1444367 | 75.570 | ng | | 0.00 |
| 79) Terphenyl-d14 | 13.075 | 244 | 1097674 | 76.150 | ng | | 0.00 |
| Target Compounds | | | | | | | Qvalue |
| 2) 1,4-Dioxane | 2.811 | 88 | 169746 | 39.864 | ng | | 98 |
| 3) Pyridine | 3.581 | 79 | 425792 | 39.489 | ng | | 98 |
| 4) n-Nitrosodimethylamine | 3.522 | 42 | 217941 | 39.012 | ng | | 99 |
| 6) Aniline | 6.628 | 93 | 595971 | 39.360 | ng | | 99 |
| 8) 2-Chlorophenol | 6.751 | 128 | 355815 | 41.401 | ng | | 98 |
| 9) Benzaldehyde | 6.516 | 77 | 335257 | 40.919 | ng | | 99 |
| 10) Phenol | 6.598 | 94 | 490852 | 42.169 | ng | | 100 |
| 11) bis(2-Chloroethyl)ether | 6.698 | 93 | 350864 | 38.936 | ng | | 99 |
| 12) 1,3-Dichlorobenzene | 6.910 | 146 | 386579 | 39.815 | ng | | 99 |
| 13) 1,4-Dichlorobenzene | 6.987 | 146 | 383281 | 39.238 | ng | | 99 |
| 14) 1,2-Dichlorobenzene | 7.140 | 146 | 363768 | 39.051 | ng | | 98 |
| 15) Benzyl Alcohol | 7.098 | 79 | 319688 | 40.125 | ng | | 99 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.240 | 45 | 645668 | 38.870 | ng | | 100 |
| 17) 2-Methylphenol | 7.204 | 107 | 295753 | 39.876 | ng | | 99 |
| 18) Hexachloroethane | 7.487 | 117 | 131364 | 41.310 | ng | | 98 |
| 19) n-Nitroso-di-n-propyla... | 7.375 | 70 | 260963 | 39.615 | ng | | 99 |
| 20) 3+4-Methylphenols | 7.363 | 107 | 369651 | 40.641 | ng | | 99 |
| 22) Acetophenone | 7.369 | 105 | 476586 | 38.753 | ng | # | 93 |
| 24) Nitrobenzene | 7.545 | 77 | 376400 | 42.874 | ng | | 99 |
| 25) Isophorone | 7.787 | 82 | 688462 | 39.167 | ng | | 100 |
| 26) 2-Nitrophenol | 7.863 | 139 | 148006 | 42.663 | ng | | 100 |
| 27) 2,4-Dimethylphenol | 7.893 | 122 | 326049 | 39.965 | ng | | 99 |
| 28) bis(2-Chloroethoxy)met... | 7.993 | 93 | 420098 | 39.412 | ng | | 100 |
| 29) 2,4-Dichlorophenol | 8.098 | 162 | 274747 | 41.164 | ng | | 99 |
| 30) 1,2,4-Trichlorobenzene | 8.187 | 180 | 297978 | 39.339 | ng | | 98 |
| 31) Naphthalene | 8.269 | 128 | 991571 | 39.140 | ng | | 100 |
| 32) Benzoic acid | 7.987 | 122 | 150277 | 38.466 | ng | | 97 |
| 33) 4-Chloroaniline | 8.310 | 127 | 401159 | 39.377 | ng | | 100 |
| 34) Hexachlorobutadiene | 8.393 | 225 | 181527 | 40.072 | ng | | 99 |
| 35) Caprolactam | 8.675 | 113 | 85847 | 41.190 | ng | | 95 |
| 36) 4-Chloro-3-methylphenol | 8.787 | 107 | 298817 | 40.192 | ng | | 100 |
| 37) 2-Methylnaphthalene | 8.963 | 142 | 589636 | 39.014 | ng | | 100 |
| 38) 1-Methylnaphthalene | 9.063 | 142 | 610968 | 38.976 | ng | | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 9.128 | 216 | 287903 | 38.885 | ng | | 99 |
| 41) Hexachlorocyclopentadiene | 9.122 | 237 | 173333 | 42.225 | ng | | 99 |
| 43) 2,4,6-Trichlorophenol | 9.234 | 196 | 195982 | 42.600 | ng | | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143117.D
Acq On : 16 Jul 2025 19:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Quant Time: Jul 17 03:44:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 9.269 | 196 | 196557 | 40.747 | ng | | 99 |
| 46) 1,1'-Biphenyl | 9.422 | 154 | 779366 | 38.466 | ng | | 100 |
| 47) 2-Chloronaphthalene | 9.451 | 162 | 573731 | 38.546 | ng | | 100 |
| 48) 2-Nitroaniline | 9.534 | 65 | 178618 | 40.710 | ng | | 99 |
| 49) Acenaphthylene | 9.869 | 152 | 971647 | 39.176 | ng | | 99 |
| 50) Dimethylphthalate | 9.722 | 163 | 628010 | 39.291 | ng | | 100 |
| 51) 2,6-Dinitrotoluene | 9.775 | 165 | 130906 | 41.829 | ng | | 95 |
| 52) Acenaphthene | 10.039 | 154 | 566434 | 38.701 | ng | | 99 |
| 53) 3-Nitroaniline | 9.945 | 138 | 151765 | 43.942 | ng | | 97 |
| 54) 2,4-Dinitrophenol | 10.045 | 184 | 41904 | 42.980 | ng | # | 44 |
| 55) Dibenzofuran | 10.210 | 168 | 838011 | 38.413 | ng | | 100 |
| 56) 4-Nitrophenol | 10.086 | 139 | 117068 | 42.650 | ng | | 99 |
| 57) 2,4-Dinitrotoluene | 10.181 | 165 | 155813 | 40.106 | ng | | 96 |
| 58) Fluorene | 10.551 | 166 | 619482 | 37.875 | ng | | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 10.322 | 232 | 164509 | 42.914 | ng | | 98 |
| 60) Diethylphthalate | 10.422 | 149 | 615586 | 39.920 | ng | | 100 |
| 61) 4-Chlorophenyl-phenyle... | 10.545 | 204 | 306293 | 39.072 | ng | | 100 |
| 62) 4-Nitroaniline | 10.557 | 138 | 129761 | 43.204 | ng | | 99 |
| 63) Azobenzene | 10.704 | 77 | 652575 | 38.265 | ng | | 99 |
| 65) 4,6-Dinitro-2-methylph... | 10.586 | 198 | 62192 | 43.295 | ng | | 97 |
| 66) n-Nitrosodiphenylamine | 10.657 | 169 | 545898 | 39.129 | ng | | 99 |
| 67) 4-Bromophenyl-phenylether | 11.034 | 248 | 180440 | 39.949 | ng | | 98 |
| 68) Hexachlorobenzene | 11.104 | 284 | 191457 | 39.950 | ng | | 97 |
| 69) Atrazine | 11.181 | 200 | 145300 | 40.708 | ng | | 99 |
| 70) Pentachlorophenol | 11.292 | 266 | 114049 | 43.376 | ng | | 99 |
| 71) Phenanthrene | 11.516 | 178 | 831246 | 38.780 | ng | | 100 |
| 72) Anthracene | 11.569 | 178 | 844296 | 39.169 | ng | | 99 |
| 73) Carbazole | 11.716 | 167 | 742234 | 38.436 | ng | | 100 |
| 74) Di-n-butylphthalate | 12.051 | 149 | 743944 | 41.970 | ng | | 100 |
| 75) Fluoranthene | 12.704 | 202 | 752396 | 38.604 | ng | | 99 |
| 77) Benzidine | 12.816 | 184 | 274069 | 38.177 | ng | | 99 |
| 78) Pyrene | 12.933 | 202 | 747120 | 38.214 | ng | | 100 |
| 80) Butylbenzylphthalate | 13.545 | 149 | 153672 | 39.754 | ng | | 95 |
| 81) Benzo(a)anthracene | 14.116 | 228 | 570413 | 40.624 | ng | | 100 |
| 82) 3,3'-Dichlorobenzidine | 14.074 | 252 | 164768 | 40.409 | ng | | 99 |
| 83) Chrysene | 14.157 | 228 | 486986 | 37.657 | ng | | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 14.110 | 149 | 241107 | 41.219 | ng | | 100 |
| 85) Di-n-octyl phthalate | 14.727 | 149 | 439529 | 40.440 | ng | | 99 |
| 87) Indeno(1,2,3-cd)pyrene | 17.180 | 276 | 781535 | 41.122 | ng | | 100 |
| 88) Benzo(b)fluoranthene | 15.192 | 252 | 580612 | 39.393 | ng | | 99 |
| 89) Benzo(k)fluoranthene | 15.221 | 252 | 558211 | 39.698 | ng | | 99 |
| 90) Benzo(a)pyrene | 15.574 | 252 | 570084 | 40.896 | ng | | 99 |
| 91) Dibenzo(a,h)anthracene | 17.198 | 278 | 635338 | 40.895 | ng | | 99 |
| 92) Benzo(g,h,i)perylene | 17.645 | 276 | 642024 | 40.534 | ng | | 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143117.D
Acq On : 16 Jul 2025 19:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Quant Time: Jul 17 03:44:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.969 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:152 Resp: 135149

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 160.8 | 126.5 | 189.7 |
| 115 | 59.4 | 47.0 | 70.6 |







#2
1,4-Dioxane
Concen: 39.864 ng
RT: 2.811 min Scan# 105
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion: 88 Resp: 169746

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 86.0 | 70.5 | 105.7 |
| 43 | 32.5 | 27.3 | 40.9 |






#3
Pyridine
Concen: 39.489 ng
RT: 3.581 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion: 79 Resp: 425792

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 84.8 | 69.2 | 103.8 |
| 51 | 40.9 | 33.3 | 49.9 |







#4
n-Nitrosodimethylamine
Concen: 39.012 ng
RT: 3.522 min Scan# 226
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion: 42 Resp: 217941

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 115.4 | 92.8 | 139.2 |
| 44 | 7.2 | 5.8 | 8.6 |






#5
2-Fluorophenol
Concen: 78.913 ng
RT: 5.587 min Scan# 5
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:112 Resp: 675027

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 67.0 | 54.6 | 82.0 |
| 63 | 33.5 | 27.3 | 40.9 |







#6
Aniline
Concen: 39.360 ng
RT: 6.628 min Scan# 754
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion: 93 Resp: 595971

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 41.0 | 33.1 | 49.7 |
| 65 | 20.8 | 16.9 | 25.3 |






#7
Phenol-d6
Concen: 78.603 ng
RT: 6.587 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion: 99 Resp: 845178

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 21.4 | 17.2 | 25.8 |
| 71 | 34.0 | 27.4 | 41.0 |







#8
2-Chlorophenol
Concen: 41.401 ng
RT: 6.751 min Scan# 775
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:128 Resp: 355815

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.1 | 12.8 | 52.8 |
| 64 | 54.7 | 36.1 | 76.1 |






#9
Benzaldehyde
Concen: 40.919 ng
RT: 6.516 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion: 77 Resp: 335257

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 94.2 | 76.3 | 116.3 |
| 106 | 92.7 | 72.8 | 112.8 |







#10
Phenol
Concen: 42.169 ng
RT: 6.598 min Scan# 749
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion: 94 Resp: 490852

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 30.0 | 9.8 | 49.8 |
| 66 | 40.7 | 20.5 | 60.5 |






#11
bis(2-Chloroethyl)ether
Concen: 38.936 ng
RT: 6.698 min Scan# 7
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion: 93 Resp: 350864

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 83.4 | 61.9 | 101.9 |
| 95 | 32.4 | 12.2 | 52.2 |







#12
1,3-Dichlorobenzene
Concen: 39.815 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:146 Resp: 386579

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.3 | 50.1 | 75.1 |
| 75 | 33.5 | 27.4 | 41.0 |






#13
1,4-Dichlorobenzene
Concen: 39.238 ng
RT: 6.987 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:146 Resp: 383281

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.8 | 50.6 | 76.0 |
| 111 | 43.3 | 35.0 | 52.6 |







#14
1,2-Dichlorobenzene
Concen: 39.051 ng
RT: 7.140 min Scan# 841
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:146 Resp: 363768

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.4 | 49.8 | 74.6 |
| 111 | 45.6 | 35.4 | 53.0 |






#15
Benzyl Alcohol
Concen: 40.125 ng
RT: 7.098 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion: 79 Resp: 319688

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 76.1 | 61.1 | 91.7 |
| 77 | 61.7 | 49.9 | 74.9 |







#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.870 ng
RT: 7.240 min Scan# 858
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion: 45 Resp: 645668

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 11.0 | 0.0 | 31.1 |
| 79 | 8.6 | 0.0 | 28.4 |






#17
2-Methylphenol
Concen: 39.876 ng
RT: 7.204 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:107 Resp: 295753

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 114.8 | 90.9 | 136.3 |
| 77 | 45.5 | 37.5 | 56.3 |
| 79 | 47.0 | 36.9 | 55.3 |







#18
Hexachloroethane
Concen: 41.310 ng
RT: 7.487 min Scan# 900
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:117 Resp: 131364

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 97.2 | 76.7 | 115.1 |
| 201 | 95.8 | 75.0 | 112.4 |






#19
n-Nitroso-di-n-propylamine
Concen: 39.615 ng
RT: 7.375 min Scan# 8
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion: 70 Resp: 260963

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 65.6 | 53.2 | 79.8 |
| 101 | 9.2 | 7.5 | 11.3 |
| 130 | 19.8 | 15.7 | 23.5 |







#20
3+4-Methylphenols
Concen: 40.641 ng
RT: 7.363 min Scan# 879
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:107 Resp: 369651

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 91.8 | 71.3 | 111.3 |
| 77 | 98.8 | 77.2 | 117.2 |
| 79 | 30.6 | 10.7 | 50.7 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.251 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:136 Resp: 510164

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 11.0 | 8.9 | 13.3 |
| 54 | 9.9 | 8.3 | 12.5 |
| 68 | 6.4 | 5.3 | 7.9 |







#22
Acetophenone
Concen: 38.753 ng
RT: 7.369 min Scan# 880
Delta R.T. -0.006 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:105 Resp: 476586

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 7.0 | 7.2 | 10.8# |
| 51 | 39.9 | 27.0 | 40.4 |
| 120 | 23.1 | 18.2 | 27.2 |






#23
Nitrobenzene-d5
Concen: 90.170 ng
RT: 7.528 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion: 82 Resp: 820016

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 42.3 | 33.3 | 49.9 |
| 54 | 58.7 | 47.4 | 71.2 |







#24
Nitrobenzene
Concen: 42.874 ng
RT: 7.545 min Scan# 910
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion: 77 Resp: 376400

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 43.6 | 34.5 | 51.7 |
| 65 | 13.8 | 11.3 | 16.9 |






#25
Isophorone
Concen: 39.167 ng
RT: 7.787 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion: 82 Resp: 688462

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 6.9 | 5.6 | 8.4 |
| 138 | 17.7 | 14.3 | 21.5 |







#26
2-Nitrophenol
Concen: 42.663 ng
RT: 7.863 min Scan# 964
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:139 Resp: 148006

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 29.3 | 23.4 | 35.0 |
| 65 | 54.3 | 43.4 | 65.2 |






#27
2,4-Dimethylphenol
Concen: 39.965 ng
RT: 7.893 min Scan# 9
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:122 Resp: 326049

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 114.8 | 91.3 | 136.9 |
| 121 | 57.9 | 45.5 | 68.3 |







#28
bis(2-Chloroethoxy)methane
Concen: 39.412 ng
RT: 7.993 min Scan# 986
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion: 93 Resp: 420098

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.6 | 26.2 | 39.4 |
| 123 | 11.9 | 9.6 | 14.4 |






#29
2,4-Dichlorophenol
Concen: 41.164 ng
RT: 8.098 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:162 Resp: 274747

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 64.6 | 43.1 | 83.1 |
| 98 | 39.5 | 19.2 | 59.2 |







#30
1,2,4-Trichlorobenzene
Concen: 39.339 ng
RT: 8.187 min Scan# 1019
Delta R.T. -0.006 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:180 Resp: 297978

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 94.8 | 74.7 | 112.1 |
| 145 | 32.1 | 24.5 | 36.7 |






#31
Naphthalene
Concen: 39.140 ng
RT: 8.269 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:128 Resp: 991571

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.1 | 9.0 | 13.4 |
| 127 | 13.1 | 10.5 | 15.7 |







#32
Benzoic acid
Concen: 38.466 ng
RT: 7.987 min Scan# 985
Delta R.T. -0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:122 Resp: 150277

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 128.6 | 105.4 | 145.4 |
| 77 | 102.1 | 78.6 | 118.6 |






#33
4-Chloroaniline
Concen: 39.377 ng
RT: 8.310 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:127 Resp: 401159

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.4 | 26.2 | 39.4 |
| 65 | 35.5 | 28.6 | 42.8 |
| 92 | 21.3 | 17.1 | 25.7 |







#34
Hexachlorobutadiene
Concen: 40.072 ng
RT: 8.393 min Scan# 1054
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:225 Resp: 181527

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 62.6 | 50.3 | 75.5 |
| 227 | 63.6 | 50.1 | 75.1 |






#35
Caprolactam
Concen: 41.190 ng
RT: 8.675 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:113 Resp: 85847

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 197.8 | 186.2 | 226.2 |
| 56 | 147.6 | 134.1 | 174.1 |







#36
4-Chloro-3-methylphenol
Concen: 40.192 ng
RT: 8.787 min Scan# 1121
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:107 Resp: 298817

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 24.9 | 20.2 | 30.2 |
| 142 | 77.9 | 62.4 | 93.6 |






#37
2-Methylnaphthalene
Concen: 39.014 ng
RT: 8.963 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:142 Resp: 589636

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 87.6 | 70.2 | 105.2 |







#38
1-Methylnaphthalene
Concen: 38.976 ng
RT: 9.063 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:142 Resp: 610968

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 90.0 | 72.2 | 108.4 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:164 Resp: 254065

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 98.2 | 79.0 | 118.6 |
| 160 | 43.7 | 35.8 | 53.6 |







#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.885 ng
RT: 9.128 min Scan# 1179
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:216 Resp: 287903

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 79.1 | 62.6 | 94.0 |
| 179 | 20.6 | 16.2 | 24.2 |
| 108 | 20.0 | 15.8 | 23.6 |






#41
Hexachlorocyclopentadiene
Concen: 42.225 ng
RT: 9.122 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:237 Resp: 173333

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 62.7 | 41.6 | 81.6 |
| 272 | 13.1 | 0.0 | 32.9 |







#42
2,4,6-Tribromophenol
Concen: 85.634 ng
RT: 10.792 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:330 Resp: 193992

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 97.0 | 77.9 | 116.9 |
| 141 | 30.8 | 25.9 | 38.9 |






#43
2,4,6-Trichlorophenol
Concen: 42.600 ng
RT: 9.234 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:196 Resp: 195982

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.3 | 76.9 | 115.3 |
| 200 | 30.7 | 23.5 | 35.3 |







#44
2,4,5-Trichlorophenol
Concen: 40.747 ng
RT: 9.269 min Scan# 1203
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:196 Resp: 196557

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.9 | 76.4 | 114.6 |
| 97 | 53.9 | 42.4 | 63.6 |
| 132 | 30.6 | 23.9 | 35.9 |






#45
2-Fluorobiphenyl
Concen: 75.570 ng
RT: 9.322 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:172 Resp: 1444367

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.1 | 28.2 | 42.4 |
| 170 | 23.2 | 18.6 | 28.0 |







#46
1,1'-Biphenyl
Concen: 38.466 ng
RT: 9.422 min Scan# 1229
Delta R.T. -0.006 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:154 Resp: 779366

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.0 | 20.1 | 60.1 |
| 76 | 14.3 | 0.0 | 33.7 |






#47
2-Chloronaphthalene
Concen: 38.546 ng
RT: 9.451 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:162 Resp: 573731

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 41.4 | 33.0 | 49.6 |
| 164 | 32.9 | 26.1 | 39.1 |







#48
2-Nitroaniline
Concen: 40.710 ng
RT: 9.534 min Scan# 1248
Delta R.T. 0.001 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion: 65 Resp: 178618

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 64.7 | 51.3 | 76.9 |
| 138 | 95.9 | 77.8 | 116.6 |






#49
Acenaphthylene
Concen: 39.176 ng
RT: 9.869 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:152 Resp: 971647

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.3 | 15.8 | 23.8 |
| 153 | 13.3 | 10.4 | 15.6 |







#50
Dimethylphthalate
Concen: 39.291 ng
RT: 9.722 min Scan# 1280
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:163 Resp: 628010

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.4 | 3.4 | 5.2 |
| 164 | 10.2 | 8.2 | 12.2 |






#51
2,6-Dinitrotoluene
Concen: 41.829 ng
RT: 9.775 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:165 Resp: 130906

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 67.7 | 57.4 | 86.0 |
| 89 | 59.9 | 51.8 | 77.8 |







#52
Acenaphthene
Concen: 38.701 ng
RT: 10.039 min Scan# 1334
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:154 Resp: 566434

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 107.6 | 86.7 | 130.1 |
| 152 | 50.3 | 41.4 | 62.2 |






#53
3-Nitroaniline
Concen: 43.942 ng
RT: 9.945 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:138 Resp: 151765

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 12.2 | 10.1 | 15.1 |
| 92 | 116.8 | 96.4 | 144.6 |







#54
2,4-Dinitrophenol
Concen: 42.980 ng
RT: 10.045 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:184 Resp: 41904

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 149.9 | 149.3 | 223.9 |
| 154 | 980.5 | 1048.0 | 1572.0# |






#55
Dibenzofuran
Concen: 38.413 ng
RT: 10.210 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:168 Resp: 838011

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 37.0 | 29.8 | 44.6 |
| 169 | 13.3 | 10.8 | 16.2 |







#56
4-Nitrophenol
Concen: 42.650 ng
RT: 10.086 min Scan# 1342
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:139 Resp: 117068

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 69.5 | 49.4 | 89.4 |
| 65 | 105.7 | 88.2 | 128.2 |






#57
2,4-Dinitrotoluene
Concen: 40.106 ng
RT: 10.181 min Scan# 1
Delta R.T. 0.001 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:165 Resp: 155813

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 48.6 | 40.6 | 61.0 |
| 89 | 74.4 | 62.3 | 93.5 |
| 182 | 3.1 | 2.6 | 3.8 |







#58
Fluorene
Concen: 37.875 ng
RT: 10.551 min Scan# 1421
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:166 Resp: 619482

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 91.6 | 73.7 | 110.5 |
| 167 | 13.7 | 10.9 | 16.3 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 42.914 ng
RT: 10.322 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:232 Resp: 164509

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 44.1 | 36.9 | 55.3 |
| 130 | 2.3 | 1.8 | 2.8 |
| 166 | 29.4 | 23.0 | 34.6 |







#60
Diethylphthalate
Concen: 39.920 ng
RT: 10.422 min Scan# 1399
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:149 Resp: 615586

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 21.8 | 17.3 | 25.9 |
| 150 | 12.6 | 9.9 | 14.9 |






#61
4-Chlorophenyl-phenylether
Concen: 39.072 ng
RT: 10.545 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:204 Resp: 306293

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 32.7 | 25.8 | 38.6 |
| 141 | 58.6 | 46.8 | 70.2 |







#62
4-Nitroaniline
Concen: 43.204 ng
RT: 10.557 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:138 Resp: 129761

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 49.1 | 30.4 | 70.4 |
| 108 | 68.9 | 49.5 | 89.5 |






#63
Azobenzene
Concen: 38.265 ng
RT: 10.704 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion: 77 Resp: 652575

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 24.9 | 4.0 | 44.0 |
| 105 | 21.3 | 1.4 | 41.4 |
| 51 | 34.1 | 14.3 | 54.3 |







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.492 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:188 Resp: 397173

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 10.4 | 8.6 | 13.0 |
| 80 | 11.1 | 9.3 | 13.9 |






#65
4,6-Dinitro-2-methylphenol
Concen: 43.295 ng
RT: 10.586 min Scan# 14
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:198 Resp: 62192

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 60.8 | 44.6 | 84.6 |
| 105 | 48.2 | 29.3 | 69.3 |







#66
n-Nitrosodiphenylamine
Concen: 39.129 ng
RT: 10.657 min Scan# 1439
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:169 Resp: 545898

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 65.1 | 53.0 | 79.6 |
| 167 | 35.1 | 28.4 | 42.6 |






#67
4-Bromophenyl-phenylether
Concen: 39.949 ng
RT: 11.034 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:248 Resp: 180440

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 97.3 | 76.7 | 115.1 |
| 141 | 75.0 | 61.7 | 92.5 |







#68
Hexachlorobenzene
Concen: 39.950 ng
RT: 11.104 min Scan# 1515
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:284 Resp: 191457

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 32.9 | 28.9 | 43.3 |
| 249 | 30.3 | 24.6 | 37.0 |






#69
Atrazine
Concen: 40.708 ng
RT: 11.181 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:200 Resp: 145300

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 28.4 | 9.1 | 49.1 |
| 215 | 51.4 | 31.0 | 71.0 |







#70
Pentachlorophenol
Concen: 43.376 ng
RT: 11.292 min Scan# 1547
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:266 Resp: 114049

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 63.1 | 51.8 | 77.6 |
| 264 | 62.0 | 49.7 | 74.5 |






#71
Phenanthrene
Concen: 38.780 ng
RT: 11.516 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:178 Resp: 831246

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.2 | 15.1 | 22.7 |
| 179 | 15.4 | 12.2 | 18.4 |







#72
Anthracene
Concen: 39.169 ng
RT: 11.569 min Scan# 1594
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:178 Resp: 844296

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.1 | 14.7 | 22.1 |
| 179 | 15.3 | 12.4 | 18.6 |






#73
Carbazole
Concen: 38.436 ng
RT: 11.716 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:167 Resp: 742234

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.0 | 17.0 | 25.4 |
| 139 | 12.6 | 10.2 | 15.4 |







#74
Di-n-butylphthalate
Concen: 41.970 ng
RT: 12.051 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:149 Resp: 743944

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.0 | 7.3 | 10.9 |
| 104 | 4.6 | 3.8 | 5.8 |






#75
Fluoranthene
Concen: 38.604 ng
RT: 12.704 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:202 Resp: 752396

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 11.7 | 0.0 | 32.4 |
| 203 | 17.4 | 0.0 | 37.4 |







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.127 min Scan# 2029
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:240 Resp: 210714

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 12.3 | 10.0 | 15.0 |
| 236 | 25.1 | 20.0 | 30.0 |






#77
Benzidine
Concen: 38.177 ng
RT: 12.816 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:184 Resp: 274069

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 16.0 | 12.4 | 18.6 |
| 183 | 11.1 | 8.6 | 12.8 |







#78
Pyrene
Concen: 38.214 ng
RT: 12.933 min Scan# 1826
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:202 Resp: 747120

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 19.3 | 15.5 | 23.3 |
| 203 | 17.4 | 14.0 | 21.0 |






#79
Terphenyl-d14
Concen: 76.150 ng
RT: 13.075 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:244 Resp: 1097674

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.5 | 5.2 | 7.8 |
| 122 | 10.8 | 8.8 | 13.2 |







#80
Butylbenzylphthalate
Concen: 39.754 ng
RT: 13.545 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:149 Resp: 153672

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 75.8 | 57.0 | 85.4 |
| 206 | 20.7 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 40.624 ng
RT: 14.116 min Scan# 2
Delta R.T. -0.006 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:228 Resp: 570413

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 26.5 | 21.3 | 31.9 |
| 229 | 19.7 | 15.5 | 23.3 |







#82
3,3'-Dichlorobenzidine
Concen: 40.409 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.006 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:252 Resp: 164768

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 63.9 | 51.5 | 77.3 |
| 126 | 12.6 | 9.4 | 14.0 |






#83
Chrysene
Concen: 37.657 ng
RT: 14.157 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:228 Resp: 486986

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 29.6 | 23.4 | 35.2 |
| 229 | 19.3 | 15.3 | 22.9 |







#84
Bis(2-ethylhexyl)phthalate
Concen: 41.219 ng
RT: 14.110 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:149 Resp: 241107

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 27.2 | 21.9 | 32.9 |
| 279 | 4.4 | 3.4 | 5.0 |






#85
Di-n-octyl phthalate
Concen: 40.440 ng
RT: 14.727 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:149 Resp: 439529

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.6 | 1.2 | 1.8 |
| 43 | 13.1 | 10.6 | 16.0 |







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.633 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:264 Resp: 255524

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.5 | 18.5 | 27.7 |
| 265 | 21.7 | 17.5 | 26.3 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 41.122 ng
RT: 17.180 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:276 Resp: 781535

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 27.9 | 22.6 | 34.0 |
| 277 | 25.2 | 20.2 | 30.2 |







#88
Benzo(b)fluoranthene
Concen: 39.393 ng
RT: 15.192 min Scan# 2210
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:252 Resp: 580612

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.0 | 17.5 | 26.3 |
| 125 | 10.2 | 8.6 | 13.0 |






#89
Benzo(k)fluoranthene
Concen: 39.698 ng
RT: 15.221 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:252 Resp: 558211

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.0 | 17.4 | 26.2 |
| 125 | 10.2 | 8.5 | 12.7 |







#90
Benzo(a)pyrene
Concen: 40.896 ng
RT: 15.574 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:252 Resp: 570084

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.8 | 17.5 | 26.3 |
| 125 | 11.0 | 9.4 | 14.0 |






#91
Dibenzo(a,h)anthracene
Concen: 40.895 ng
RT: 17.198 min Scan# 2
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

**Instrument :**
BNA_F
**ClientSampleId :**
SSTDCCC040

Tgt Ion:278 Resp: 635338

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 17.6 | 14.6 | 22.0 |
| 279 | 23.8 | 18.8 | 28.2 |







#92
Benzo(g,h,i)perylene
Concen: 40.534 ng
RT: 17.645 min Scan# 2627
Delta R.T. 0.000 min
Lab File: BF143117.D
Acq: 16 Jul 2025 19:39

Tgt Ion:276 Resp: 642024

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.1 | 19.2 | 28.8 |
| 138 | 24.0 | 19.8 | 29.6 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143117.D
Acq On : 16 Jul 2025 19:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_F
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 17 03:44:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 1 | I | 1,4-Dichlorobenzene-d4 | 1.000 | 1.000 | 0.0 | 109 | 0.00 |
| 2 | | 1,4-Dioxane | 0.630 | 0.628 | 0.3 | 110 | 0.00 |
| 3 | | Pyridine | 1.596 | 1.575 | 1.3 | 108 | 0.00 |
| 4 | | n-Nitrosodimethylamine | 0.827 | 0.806 | 2.5 | 108 | 0.00 |
| 5 | S | 2-Fluorophenol | 1.266 | 1.249 | 1.3 | 109 | 0.00 |
| 6 | | Aniline | 2.241 | 2.205 | 1.6 | 109 | 0.00 |
| 7 | S | Phenol-d6 | 1.591 | 1.563 | 1.8 | 109 | 0.00 |
| 8 | | 2-Chlorophenol | 1.272 | 1.316 | -3.5 | 113 | 0.00 |
| 9 | | Benzaldehyde | 1.212 | 1.240 | -2.3 | 94 | 0.00 |
| 10 | C | Phenol | 1.723 | 1.816 | -5.4 | 116 | 0.00 |
| 11 | | bis(2-Chloroethyl)ether | 1.334 | 1.298 | 2.7 | 109 | 0.00 |
| 12 | | 1,3-Dichlorobenzene | 1.437 | 1.430 | 0.5 | 111 | 0.00 |
| 13 | C | 1,4-Dichlorobenzene | 1.446 | 1.418 | 1.9 | 110 | 0.00 |
| 14 | | 1,2-Dichlorobenzene | 1.379 | 1.346 | 2.4 | 108 | 0.00 |
| 15 | | Benzyl Alcohol | 1.179 | 1.183 | -0.3 | 111 | 0.00 |
| 16 | | 2,2'-oxybis(1-Chloropropane | 2.458 | 2.389 | 2.8 | 109 | 0.00 |
| 17 | | 2-Methylphenol | 1.098 | 1.094 | 0.4 | 111 | 0.00 |
| 18 | | Hexachloroethane | 0.471 | 0.486 | -3.2 | 113 | 0.00 |
| 19 | P | n-Nitroso-di-n-propylamine | 0.975 | 0.965 | 1.0 | 110 | 0.00 |
| 20 | | 3+4-Methylphenols | 1.346 | 1.368 | -1.6 | 110 | 0.00 |
| 21 | I | Naphthalene-d8 | 1.000 | 1.000 | 0.0 | 109 | 0.00 |
| 22 | | Acetophenone | 0.482 | 0.467 | 3.1 | 109 | 0.00 |
| 23 | S | Nitrobenzene-d5 | 0.357 | 0.402 | -12.6 | 118 | 0.00 |
| 24 | | Nitrobenzene | 0.344 | 0.369 | -7.3 | 113 | 0.00 |
| 25 | | Isophorone | 0.689 | 0.675 | 2.0 | 109 | 0.00 |
| 26 | C | 2-Nitrophenol | 0.114 | 0.145 | -27.2# | 132 | 0.00 |
| 27 | | 2,4-Dimethylphenol | 0.320 | 0.320 | 0.0 | 109 | 0.00 |
| 28 | | bis(2-Chloroethoxy)methane | 0.418 | 0.412 | 1.4 | 110 | 0.00 |
| 29 | C | 2,4-Dichlorophenol | 0.262 | 0.269 | -2.7 | 110 | 0.00 |
| 30 | | 1,2,4-Trichlorobenzene | 0.297 | 0.292 | 1.7 | 108 | 0.00 |
| 31 | | Naphthalene | 0.993 | 0.972 | 2.1 | 110 | 0.00 |
| 32 | | Benzoic acid | 0.139 | 0.147 | -5.8 | 112 | 0.00 |
| 33 | | 4-Chloroaniline | 0.399 | 0.393 | 1.5 | 108 | 0.00 |
| 34 | C | Hexachlorobutadiene | 0.178 | 0.178 | 0.0 | 110 | 0.00 |
| 35 | | Caprolactam | 0.082 | 0.084 | -2.4 | 111 | 0.00 |
| 36 | C | 4-Chloro-3-methylphenol | 0.291 | 0.293 | -0.7 | 110 | 0.00 |
| 37 | | 2-Methylnaphthalene | 0.592 | 0.578 | 2.4 | 109 | 0.00 |
| 38 | | 1-Methylnaphthalene | 0.615 | 0.599 | 2.6 | 109 | 0.00 |
| 39 | I | Acenaphthene-d10 | 1.000 | 1.000 | 0.0 | 112 | 0.00 |
| 40 | | 1,2,4,5-Tetrachlorobenzene | 0.583 | 0.567 | 2.7 | 109 | 0.00 |
| 41 | P | Hexachlorocyclopentadiene | 0.323 | 0.341 | -5.6 | 118 | 0.00 |
| 42 | S | 2,4,6-Tribromophenol | 0.178 | 0.191 | -7.3 | 113 | 0.00 |
| 43 | C | 2,4,6-Trichlorophenol | 0.362 | 0.386 | -6.6 | 114 | 0.00 |
| 44 | | 2,4,5-Trichlorophenol | 0.380 | 0.387 | -1.8 | 110 | 0.00 |
| 45 | S | 2-Fluorobiphenyl | 1.505 | 1.421 | 5.6 | 109 | 0.00 |
| 46 | | 1,1'-Biphenyl | 1.595 | 1.534 | 3.8 | 109 | 0.00 |
| 47 | | 2-Chloronaphthalene | 1.172 | 1.129 | 3.7 | 109 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143117.D
Acq On : 16 Jul 2025 19:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_F
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 17 03:44:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 48 | | 2-Nitroaniline | 0.303 | 0.352 | -16.2 | 120 | 0.00 |
| 49 | | Acenaphthylene | 1.952 | 1.912 | 2.0 | 109 | 0.00 |
| 50 | | Dimethylphthalate | 1.258 | 1.236 | 1.7 | 110 | 0.00 |
| 51 | | 2,6-Dinitrotoluene | 0.219 | 0.258 | -17.8 | 123 | 0.00 |
| 52 | C | Acenaphthene | 1.152 | 1.115 | 3.2 | 110 | 0.00 |
| 53 | | 3-Nitroaniline | 0.272 | 0.299 | -9.9 | 116 | 0.00 |
| 54 | P | 2,4-Dinitrophenol | 0.073 | 0.082 | -12.3 | 134 | 0.00 |
| 55 | | Dibenzofuran | 1.717 | 1.649 | 4.0 | 108 | 0.00 |
| 56 | P | 4-Nitrophenol | 0.216 | 0.230 | -6.5 | 115 | 0.00 |
| 57 | | 2,4-Dinitrotoluene | 0.265 | 0.307 | -15.8 | 116 | 0.00 |
| 58 | | Fluorene | 1.288 | 1.219 | 5.4 | 109 | 0.00 |
| 59 | | 2,3,4,6-Tetrachlorophenol | 0.302 | 0.324 | -7.3 | 116 | 0.00 |
| 60 | | Diethylphthalate | 1.214 | 1.211 | 0.2 | 110 | 0.00 |
| 61 | | 4-Chlorophenyl-phenylether | 0.617 | 0.603 | 2.3 | 109 | 0.00 |
| 62 | | 4-Nitroaniline | 0.236 | 0.255 | -8.1 | 114 | 0.00 |
| 63 | | Azobenzene | 1.343 | 1.284 | 4.4 | 108 | 0.00 |
| 64 | I | Phenanthrene-d10 | 1.000 | 1.000 | 0.0 | 108 | 0.00 |
| 65 | | 4,6-Dinitro-2-methylphenol | 0.067 | 0.078 | -16.4 | 131 | 0.00 |
| 66 | c | n-Nitrosodiphenylamine | 0.703 | 0.687 | 2.3 | 108 | 0.00 |
| 67 | | 4-Bromophenyl-phenylether | 0.227 | 0.227 | 0.0 | 110 | 0.00 |
| 68 | | Hexachlorobenzene | 0.241 | 0.241 | 0.0 | 110 | 0.00 |
| 69 | | Atrazine | 0.180 | 0.183 | -1.7 | 109 | 0.00 |
| 70 | C | Pentachlorophenol | 0.132 | 0.144 | -9.1 | 117 | 0.00 |
| 71 | | Phenanthrene | 1.079 | 1.046 | 3.1 | 108 | 0.00 |
| 72 | | Anthracene | 1.085 | 1.063 | 2.0 | 108 | 0.00 |
| 73 | | Carbazole | 0.972 | 0.934 | 3.9 | 107 | 0.00 |
| 74 | | Di-n-butylphthalate | 0.893 | 0.937 | -4.9 | 109 | 0.00 |
| 75 | C | Fluoranthene | 0.981 | 0.947 | 3.5 | 106 | 0.00 |
| 76 | I | Chrysene-d12 | 1.000 | 1.000 | 0.0 | 110 | 0.00 |
| 77 | | Benzidine | 0.681 | 0.650 | 4.6 | 85 | 0.00 |
| 78 | | Pyrene | 1.856 | 1.773 | 4.5 | 105 | 0.00 |
| 79 | S | Terphenyl-d14 | 1.368 | 1.302 | 4.8 | 106 | 0.00 |
| 80 | | Butylbenzylphthalate | 0.318 | 0.365 | -14.8 | 116 | 0.00 |
| 81 | | Benzo(a)anthracene | 1.333 | 1.354 | -1.6 | 110 | 0.00 |
| 82 | | 3,3'-Dichlorobenzidine | 0.387 | 0.391 | -1.0 | 111 | 0.00 |
| 83 | | Chrysene | 1.227 | 1.156 | 5.8 | 105 | 0.00 |
| 84 | | Bis(2-ethylhexyl)phthalate | 0.480 | 0.572 | -19.2 | 122 | 0.00 |
| 85 | c | Di-n-octyl phthalate | 0.927 | 1.043 | -12.5 | 124 | 0.00 |
| 86 | I | Perylene-d12 | 1.000 | 1.000 | 0.0 | 113 | 0.00 |
| 87 | | Indeno(1,2,3-cd)pyrene | 1.488 | 1.529 | -2.8 | 113 | 0.00 |
| 88 | | Benzo(b)fluoranthene | 1.154 | 1.136 | 1.6 | 118 | 0.00 |
| 89 | | Benzo(k)fluoranthene | 1.101 | 1.092 | 0.8 | 107 | 0.00 |
| 90 | C | Benzo(a)pyrene | 1.091 | 1.116 | -2.3 | 112 | 0.00 |
| 91 | | Dibenzo(a,h)anthracene | 1.216 | 1.243 | -2.2 | 113 | 0.00 |
| 92 | | Benzo(g,h,i)perylene | 1.240 | 1.256 | -1.3 | 113 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143117.D
Acq On : 16 Jul 2025 19:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_F
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 17 03:44:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | AvgRF | CCRF | %Dev Area% | Dev(min) |
|---|---|---|---|---|

(#) = Out of Range SPCC's out = 0 CCC's out = 1

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143117.D
Acq On : 16 Jul 2025 19:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_F
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 17 03:44:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 1 | I | 1,4-Dichlorobenzene-d4 | 20.000 | 20.000 | 0.0 | 109 | 0.00 |
| 2 | | 1,4-Dioxane | 40.000 | 39.864 | 0.3 | 110 | 0.00 |
| 3 | | Pyridine | 40.000 | 39.489 | 1.3 | 108 | 0.00 |
| 4 | | n-Nitrosodimethylamine | 40.000 | 39.012 | 2.5 | 108 | 0.00 |
| 5 | S | 2-Fluorophenol | 80.000 | 78.913 | 1.4 | 109 | 0.00 |
| 6 | | Aniline | 40.000 | 39.360 | 1.6 | 109 | 0.00 |
| 7 | S | Phenol-d6 | 80.000 | 78.603 | 1.7 | 109 | 0.00 |
| 8 | | 2-Chlorophenol | 40.000 | 41.401 | -3.5 | 113 | 0.00 |
| 9 | | Benzaldehyde | 40.000 | 40.919 | -2.3 | 94 | 0.00 |
| 10 | C | Phenol | 40.000 | 42.169 | -5.4 | 116 | 0.00 |
| 11 | | bis(2-Chloroethyl)ether | 40.000 | 38.936 | 2.7 | 109 | 0.00 |
| 12 | | 1,3-Dichlorobenzene | 40.000 | 39.815 | 0.5 | 111 | 0.00 |
| 13 | C | 1,4-Dichlorobenzene | 40.000 | 39.238 | 1.9 | 110 | 0.00 |
| 14 | | 1,2-Dichlorobenzene | 40.000 | 39.051 | 2.4 | 108 | 0.00 |
| 15 | | Benzyl Alcohol | 40.000 | 40.125 | -0.3 | 111 | 0.00 |
| 16 | | 2,2'-oxybis(1-Chloropropane | 40.000 | 38.870 | 2.8 | 109 | 0.00 |
| 17 | | 2-Methylphenol | 40.000 | 39.876 | 0.3 | 111 | 0.00 |
| 18 | | Hexachloroethane | 40.000 | 41.310 | -3.3 | 113 | 0.00 |
| 19 | P | n-Nitroso-di-n-propylamine | 40.000 | 39.615 | 1.0 | 110 | 0.00 |
| 20 | | 3+4-Methylphenols | 40.000 | 40.641 | -1.6 | 110 | 0.00 |
| 21 | I | Naphthalene-d8 | 20.000 | 20.000 | 0.0 | 109 | 0.00 |
| 22 | | Acetophenone | 40.000 | 38.753 | 3.1 | 109 | 0.00 |
| 23 | S | Nitrobenzene-d5 | 80.000 | 90.170 | -12.7 | 118 | 0.00 |
| 24 | | Nitrobenzene | 40.000 | 42.874 | -7.2 | 113 | 0.00 |
| 25 | | Isophorone | 40.000 | 39.167 | 2.1 | 109 | 0.00 |
| 26 | C | 2-Nitrophenol | 40.000 | 42.663 | -6.7 | 132 | 0.00 |
| 27 | | 2,4-Dimethylphenol | 40.000 | 39.965 | 0.1 | 109 | 0.00 |
| 28 | | bis(2-Chloroethoxy)methane | 40.000 | 39.412 | 1.5 | 110 | 0.00 |
| 29 | C | 2,4-Dichlorophenol | 40.000 | 41.164 | -2.9 | 110 | 0.00 |
| 30 | | 1,2,4-Trichlorobenzene | 40.000 | 39.339 | 1.7 | 108 | 0.00 |
| 31 | | Naphthalene | 40.000 | 39.140 | 2.1 | 110 | 0.00 |
| 32 | | Benzoic acid | 40.000 | 38.466 | 3.8 | 112 | 0.00 |
| 33 | | 4-Chloroaniline | 40.000 | 39.377 | 1.6 | 108 | 0.00 |
| 34 | C | Hexachlorobutadiene | 40.000 | 40.072 | -0.2 | 110 | 0.00 |
| 35 | | Caprolactam | 40.000 | 41.190 | -3.0 | 111 | 0.00 |
| 36 | C | 4-Chloro-3-methylphenol | 40.000 | 40.192 | -0.5 | 110 | 0.00 |
| 37 | | 2-Methylnaphthalene | 40.000 | 39.014 | 2.5 | 109 | 0.00 |
| 38 | | 1-Methylnaphthalene | 40.000 | 38.976 | 2.6 | 109 | 0.00 |
| 39 | I | Acenaphthene-d10 | 20.000 | 20.000 | 0.0 | 112 | 0.00 |
| 40 | | 1,2,4,5-Tetrachlorobenzene | 40.000 | 38.885 | 2.8 | 109 | 0.00 |
| 41 | P | Hexachlorocyclopentadiene | 40.000 | 42.225 | -5.6 | 118 | 0.00 |
| 42 | S | 2,4,6-Tribromophenol | 80.000 | 85.634 | -7.0 | 113 | 0.00 |
| 43 | C | 2,4,6-Trichlorophenol | 40.000 | 42.600 | -6.5 | 114 | 0.00 |
| 44 | | 2,4,5-Trichlorophenol | 40.000 | 40.747 | -1.9 | 110 | 0.00 |
| 45 | S | 2-Fluorobiphenyl | 80.000 | 75.570 | 5.5 | 109 | 0.00 |
| 46 | | 1,1'-Biphenyl | 40.000 | 38.466 | 3.8 | 109 | 0.00 |
| 47 | | 2-Chloronaphthalene | 40.000 | 38.546 | 3.6 | 109 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143117.D
Acq On : 16 Jul 2025 19:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_F
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 17 03:44:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 48 | | 2-Nitroaniline | 40.000 | 40.710 | -1.8 | 120 | 0.00 |
| 49 | | Acenaphthylene | 40.000 | 39.176 | 2.1 | 109 | 0.00 |
| 50 | | Dimethylphthalate | 40.000 | 39.291 | 1.8 | 110 | 0.00 |
| 51 | | 2,6-Dinitrotoluene | 40.000 | 41.829 | -4.6 | 123 | 0.00 |
| 52 | C | Acenaphthene | 40.000 | 38.701 | 3.2 | 110 | 0.00 |
| 53 | | 3-Nitroaniline | 40.000 | 43.942 | -9.9 | 116 | 0.00 |
| 54 | P | 2,4-Dinitrophenol | 40.000 | 42.980 | -7.4 | 134 | 0.00 |
| 55 | | Dibenzofuran | 40.000 | 38.413 | 4.0 | 108 | 0.00 |
| 56 | P | 4-Nitrophenol | 40.000 | 42.650 | -6.6 | 115 | 0.00 |
| 57 | | 2,4-Dinitrotoluene | 40.000 | 40.106 | -0.3 | 116 | 0.00 |
| 58 | | Fluorene | 40.000 | 37.875 | 5.3 | 109 | 0.00 |
| 59 | | 2,3,4,6-Tetrachlorophenol | 40.000 | 42.914 | -7.3 | 116 | 0.00 |
| 60 | | Diethylphthalate | 40.000 | 39.920 | 0.2 | 110 | 0.00 |
| 61 | | 4-Chlorophenyl-phenylether | 40.000 | 39.072 | 2.3 | 109 | 0.00 |
| 62 | | 4-Nitroaniline | 40.000 | 43.204 | -8.0 | 114 | 0.00 |
| 63 | | Azobenzene | 40.000 | 38.265 | 4.3 | 108 | 0.00 |
| 64 | I | Phenanthrene-d10 | 20.000 | 20.000 | 0.0 | 108 | 0.00 |
| 65 | | 4,6-Dinitro-2-methylphenol | 40.000 | 43.295 | -8.2 | 131 | 0.00 |
| 66 | c | n-Nitrosodiphenylamine | 40.000 | 39.129 | 2.2 | 108 | 0.00 |
| 67 | | 4-Bromophenyl-phenylether | 40.000 | 39.949 | 0.1 | 110 | 0.00 |
| 68 | | Hexachlorobenzene | 40.000 | 39.950 | 0.1 | 110 | 0.00 |
| 69 | | Atrazine | 40.000 | 40.708 | -1.8 | 109 | 0.00 |
| 70 | C | Pentachlorophenol | 40.000 | 43.376 | -8.4 | 117 | 0.00 |
| 71 | | Phenanthrene | 40.000 | 38.780 | 3.0 | 108 | 0.00 |
| 72 | | Anthracene | 40.000 | 39.169 | 2.1 | 108 | 0.00 |
| 73 | | Carbazole | 40.000 | 38.436 | 3.9 | 107 | 0.00 |
| 74 | | Di-n-butylphthalate | 40.000 | 41.970 | -4.9 | 109 | 0.00 |
| 75 | C | Fluoranthene | 40.000 | 38.604 | 3.5 | 106 | 0.00 |
| 76 | I | Chrysene-d12 | 20.000 | 20.000 | 0.0 | 110 | 0.00 |
| 77 | | Benzidine | 40.000 | 38.177 | 4.6 | 85 | 0.00 |
| 78 | | Pyrene | 40.000 | 38.214 | 4.5 | 105 | 0.00 |
| 79 | S | Terphenyl-d14 | 80.000 | 76.150 | 4.8 | 106 | 0.00 |
| 80 | | Butylbenzylphthalate | 40.000 | 39.754 | 0.6 | 116 | 0.00 |
| 81 | | Benzo(a)anthracene | 40.000 | 40.624 | -1.6 | 110 | 0.00 |
| 82 | | 3,3'-Dichlorobenzidine | 40.000 | 40.409 | -1.0 | 111 | 0.00 |
| 83 | | Chrysene | 40.000 | 37.657 | 5.9 | 105 | 0.00 |
| 84 | | Bis(2-ethylhexyl)phthalate | 40.000 | 41.219 | -3.0 | 122 | 0.00 |
| 85 | c | Di-n-octyl phthalate | 40.000 | 40.440 | -1.1 | 124 | 0.00 |
| 86 | I | Perylene-d12 | 20.000 | 20.000 | 0.0 | 113 | 0.00 |
| 87 | | Indeno(1,2,3-cd)pyrene | 40.000 | 41.122 | -2.8 | 113 | 0.00 |
| 88 | | Benzo(b)fluoranthene | 40.000 | 39.393 | 1.5 | 118 | 0.00 |
| 89 | | Benzo(k)fluoranthene | 40.000 | 39.698 | 0.8 | 107 | 0.00 |
| 90 | C | Benzo(a)pyrene | 40.000 | 40.896 | -2.2 | 112 | 0.00 |
| 91 | | Dibenzo(a,h)anthracene | 40.000 | 40.895 | -2.2 | 113 | 0.00 |
| 92 | | Benzo(g,h,i)perylene | 40.000 | 40.534 | -1.3 | 113 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143117.D
Acq On : 16 Jul 2025 19:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_F
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 17 03:44:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

## SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: PARS02

Lab Code: ACE SDG No.: Q2592

Instrument ID: BNA_P Calibration Date/Time: 07/16/2025 12:42

Lab File ID: BP025155.D Init. Calib. Date(s): 07/03/2025 07/03/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:39 14:31

GC Column: ZB-GR ID: 0.25 (mm)

| COMPOUND | RRF | RRF040 | MIN RRF | %D | MAX%D |
|---|---|---|---|---|---|
| 2-Fluorophenol | 1.129 | 1.127 | | -0.2 | |
| Benzaldehyde | 0.769 | 0.625 | | -18.7 | |
| Phenol-d6 | 1.473 | 1.456 | | -1.2 | |
| Phenol | 1.497 | 1.451 | | -3.1 | 20.0 |
| bis(2-Chloroethyl)ether | 1.152 | 1.106 | | -4.0 | |
| 2-Chlorophenol | 1.263 | 1.281 | | 1.4 | |
| 2-Methylphenol | 0.964 | 0.980 | | 1.7 | |
| 2,2-oxybis(1-Chloropropane) | 1.498 | 1.353 | | -9.7 | |
| Acetophenone | 0.473 | 0.462 | | -2.3 | |
| 3+4-Methylphenols | 1.350 | 1.353 | | 0.2 | |
| n-Nitroso-di-n-propylamine | 0.874 | 0.858 | 0.050 | -1.8 | |
| Nitrobenzene-d5 | 0.301 | 0.315 | | 4.7 | |
| Hexachloroethane | 0.466 | 0.474 | | 1.7 | |
| Nitrobenzene | 0.306 | 0.312 | | 2.0 | |
| Isophorone | 0.562 | 0.566 | | 0.7 | |
| 2-Nitrophenol | 0.128 | 0.181 | | 41.4 | 20.0 |
| 2,4-Dimethylphenol | 0.210 | 0.215 | | 2.4 | |
| bis(2-Chloroethoxy)methane | 0.382 | 0.379 | | -0.8 | |
| 2,4-Dichlorophenol | 0.291 | 0.318 | | 9.3 | 20.0 |
| Naphthalene | 1.030 | 1.012 | | -1.7 | |
| 4-Chloroaniline | 0.384 | 0.391 | | 1.8 | |
| Hexachlorobutadiene | 0.185 | 0.194 | | 4.9 | 20.0 |
| Caprolactam | 0.099 | 0.105 | | 6.1 | |
| 4-Chloro-3-methylphenol | 0.304 | 0.323 | | 6.3 | 20.0 |
| 2-Methylnaphthalene | 0.727 | 0.726 | | -0.1 | |
| Hexachlorocyclopentadiene | 0.150 | 0.183 | 0.050 | 22.0 | |
| 2,4,6-Trichlorophenol | 0.348 | 0.402 | | 15.5 | 20.0 |
| 2-Fluorobiphenyl | 1.291 | 1.243 | | -3.7 | |
| 2,4,5-Trichlorophenol | 0.394 | 0.436 | | 10.7 | |
| 1,1-Biphenyl | 1.455 | 1.427 | | -1.9 | |
| 2-Chloronaphthalene | 1.094 | 1.096 | | 0.2 | |
| 2-Nitroaniline | 0.237 | 0.281 | | 18.6 | |
| Dimethylphthalate | 1.400 | 1.424 | | 1.7 | |
| Acenaphthylene | 1.674 | 1.681 | | 0.4 | |
| 2,6-Dinitrotoluene | 0.273 | 0.307 | | 12.5 | |
| 3-Nitroaniline | 0.293 | 0.331 | | 13.0 | |
| Acenaphthene | 1.079 | 1.056 | | -2.1 | 20.0 |
| 2,4-Dinitrophenol | 0.114 | 0.172 | 0.050 | 50.9 | |
| 4-Nitrophenol | 0.270 | 0.298 | 0.050 | 10.4 | |

Form VII SV-1

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

## SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: PARS02

Lab Code: ACE SDG No.: Q2592

Instrument ID: BNA_P Calibration Date/Time: 07/16/2025 12:42

Lab File ID: BP025155.D Init. Calib. Date(s): 07/03/2025 07/03/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:39 14:31

GC Column: ZB-GR ID: 0.25 (mm)

| COMPOUND | RRF | RRF040 | MIN RRF | %D | MAX%D |
|---|---|---|---|---|---|
| Dibenzofuran | 1.781 | 1.741 | | -2.2 | |
| 2,4-Dinitrotoluene | 0.356 | 0.420 | | 18.0 | |
| Diethylphthalate | 1.406 | 1.422 | | 1.1 | |
| 4-Chlorophenyl-phenylether | 0.682 | 0.683 | | 0.1 | |
| Fluorene | 1.377 | 1.344 | | -2.4 | |
| 4-Nitroaniline | 0.307 | 0.349 | | 13.7 | |
| 4,6-Dinitro-2-methylphenol | 0.082 | 0.120 | | 46.3 | |
| n-Nitrosodiphenylamine | 0.572 | 0.589 | | 3.0 | 20.0 |
| 2,4,6-Tribromophenol | 0.236 | 0.289 | | 22.5 | |
| 4-Bromophenyl-phenylether | 0.205 | 0.226 | | 10.2 | |
| Hexachlorobenzene | 0.250 | 0.266 | | 6.4 | |
| Atrazine | 0.168 | 0.189 | | 12.5 | |
| Pentachlorophenol | 0.169 | 0.198 | | 17.2 | 20.0 |
| Phenanthrene | 1.069 | 1.030 | | -3.6 | |
| Anthracene | 1.025 | 1.026 | | 0.1 | |
| Carbazole | 1.002 | 0.998 | | -0.4 | |
| Di-n-butylphthalate | 1.116 | 1.239 | | 11.0 | |
| Fluoranthene | 1.249 | 1.254 | | 0.4 | 20.0 |
| Pyrene | 1.239 | 1.265 | | 2.1 | |
| Terphenyl-d14 | 0.910 | 0.912 | | 0.2 | |
| Butylbenzylphthalate | 0.402 | 0.541 | | 34.6 | |
| 3,3-Dichlorobenzidine | 0.365 | 0.456 | | 24.9 | |
| Benzo(a)anthracene | 1.212 | 1.209 | | -0.2 | |
| Chrysene | 1.166 | 1.138 | | -2.4 | |
| Bis(2-ethylhexyl)phthalate | 0.650 | 0.792 | | 21.8 | |
| Di-n-octyl phthalate | 1.054 | 1.357 | | 28.7 | 20.0 |
| Benzo(b)fluoranthene | 1.164 | 1.147 | | -1.5 | |
| Benzo(k)fluoranthene | 1.158 | 1.093 | | -5.6 | |
| Benzo(a)pyrene | 0.991 | 1.020 | | 2.9 | 20.0 |
| Indeno(1,2,3-cd)pyrene | 1.274 | 1.358 | | 6.6 | |
| Dibenzo(a,h)anthracene | 1.053 | 1.127 | | 7.0 | |
| Benzo(g,h,i)perylene | 1.088 | 1.140 | | 4.8 | |
| 1,2,4,5-Tetrachlorobenzene | 0.575 | 0.598 | | 4.0 | |
| 1,4-Dioxane | 0.439 | 0.407 | | -7.3 | 20.0 |
| 2,3,4,6-Tetrachlorophenol | 0.357 | 0.403 | | 12.9 | |

All other compounds must meet a minimum RRF of 0.010.

Form VII SV-1

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025155.D
Acq On : 16 Jul 2025 12:42
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jul 16 14:38:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| **Internal Standards** | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.443 | 152 | 564997 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.196 | 136 | 2237579 | 20.000 | ng | 0.01 |
| 39) Acenaphthene-d10 | 14.095 | 164 | 1465074 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 16.930 | 188 | 2922940 | 20.000 | ng | 0.02 |
| 76) Chrysene-d12 | 21.360 | 240 | 2997869 | 20.000 | ng | 0.02 |
| 86) Perylene-d12 | 24.501 | 264 | 3472368 | 20.000 | ng | 0.00 |
| **System Monitoring Compounds** | | | | | | |
| 5) 2-Fluorophenol | 5.102 | 112 | 2547882 | 79.902 | ng | 0.00 |
| 7) Phenol-d6 | 6.637 | 99 | 3291152 | 79.090 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.572 | 82 | 2820767 | 83.838 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.631 | 330 | 1691627 | 97.740 | ng | 0.01 |
| 45) 2-Fluorobiphenyl | 12.695 | 172 | 7286089 | 77.043 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.677 | 244 | 10940897 | 80.210 | ng | 0.00 |
| **Target Compounds** | | | | | | Qvalue |
| 2) 1,4-Dioxane | 3.119 | 88 | 459705 | 37.074 | ng | 98 |
| 3) Pyridine | 3.490 | 79 | 1295219 | 38.154 | ng | 98 |
| 4) n-Nitrosodimethylamine | 3.408 | 42 | 512069 | 35.626 | ng | 90 |
| 6) Aniline | 6.784 | 93 | 1528083 | 39.303 | ng | 99 |
| 8) 2-Chlorophenol | 7.019 | 128 | 1447361 | 40.568 | ng | 98 |
| 9) Benzaldehyde | 6.602 | 77 | 706028 | 32.488 | ng | 96 |
| 10) Phenol | 6.666 | 94 | 1639681 | 38.780 | ng | 96 |
| 11) bis(2-Chloroethyl)ether | 6.884 | 93 | 1250129 | 38.418 | ng | 98 |
| 12) 1,3-Dichlorobenzene | 7.331 | 146 | 1568447 | 39.280 | ng | 98 |
| 13) 1,4-Dichlorobenzene | 7.478 | 146 | 1602132 | 39.630 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 7.784 | 146 | 1522581 | 39.152 | ng | 99 |
| 15) Benzyl Alcohol | 7.678 | 79 | 1127395 | 40.592 | ng | 95 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.972 | 45 | 1528381 | 36.122 | ng | 98 |
| 17) 2-Methylphenol | 7.884 | 107 | 1107049 | 40.646 | ng | 99 |
| 18) Hexachloroethane | 8.502 | 117 | 535091 | 40.630 | ng | 98 |
| 19) n-Nitroso-di-n-propyla... | 8.237 | 70 | 969360 | 39.280 | ng | 97 |
| 20) 3+4-Methylphenols | 8.207 | 107 | 1528669 | 40.071 | ng | 98 |
| 22) Acetophenone | 8.243 | 105 | 2066167 | 39.045 | ng | 98 |
| 24) Nitrobenzene | 8.613 | 77 | 1395349 | 40.716 | ng | 94 |
| 25) Isophorone | 9.143 | 82 | 2532667 | 40.280 | ng | 99 |
| 26) 2-Nitrophenol | 9.313 | 139 | 810527 | 48.687 | ng | 96 |
| 27) 2,4-Dimethylphenol | 9.390 | 122 | 960958 | 40.971 | ng | 99 |
| 28) bis(2-Chloroethoxy)met... | 9.619 | 93 | 1695445 | 39.712 | ng | 98 |
| 29) 2,4-Dichlorophenol | 9.854 | 162 | 1422094 | 43.726 | ng | 98 |
| 30) 1,2,4-Trichlorobenzene | 10.060 | 180 | 1482294 | 41.157 | ng | 99 |
| 31) Naphthalene | 10.249 | 128 | 4529753 | 39.315 | ng | 99 |
| 32) Benzoic acid | 9.543 | 122 | 1191237 | 50.471 | ng | 94 |
| 33) 4-Chloroaniline | 10.348 | 127 | 1749419 | 40.686 | ng | 98 |
| 34) Hexachlorobutadiene | 10.543 | 225 | 868186 | 42.006 | ng | 99 |
| 35) Caprolactam | 11.143 | 113 | 468089 | 42.071 | ng | 94 |
| 36) 4-Chloro-3-methylphenol | 11.501 | 107 | 1445852 | 42.541 | ng | 99 |
| 37) 2-Methylnaphthalene | 11.872 | 142 | 3249124 | 39.967 | ng | 100 |
| 38) 1-Methylnaphthalene | 12.090 | 142 | 3163495 | 39.900 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 12.254 | 216 | 1752724 | 41.626 | ng | 99 |
| 41) Hexachlorocyclopentadiene | 12.243 | 237 | 536542 | 48.784 | ng | 99 |
| 43) 2,4,6-Trichlorophenol | 12.490 | 196 | 1179026 | 46.244 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025155.D
Acq On : 16 Jul 2025 12:42
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jul 16 14:38:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 12.572 | 196 | 1276875 | 44.191 | ng | 98 |
| 46) 1,1'-Biphenyl | 12.913 | 154 | 4181055 | 39.223 | ng | 99 |
| 47) 2-Chloronaphthalene | 12.948 | 162 | 3210239 | 40.068 | ng | 100 |
| 48) 2-Nitroaniline | 13.148 | 65 | 822809 | 42.274 | ng | 94 |
| 49) Acenaphthylene | 13.801 | 152 | 4926382 | 40.174 | ng | 100 |
| 50) Dimethylphthalate | 13.560 | 163 | 4171124 | 40.672 | ng | 99 |
| 51) 2,6-Dinitrotoluene | 13.672 | 165 | 899288 | 44.961 | ng | 94 |
| 52) Acenaphthene | 14.160 | 154 | 3094975 | 39.159 | ng | 99 |
| 53) 3-Nitroaniline | 14.001 | 138 | 969923 | 45.249 | ng | 99 |
| 54) 2,4-Dinitrophenol | 14.213 | 184 | 503837 | 53.458 | ng | 92 |
| 55) Dibenzofuran | 14.513 | 168 | 5100382 | 39.103 | ng | 99 |
| 56) 4-Nitrophenol | 14.337 | 139 | 872559 | 44.110 | ng | 96 |
| 57) 2,4-Dinitrotoluene | 14.478 | 165 | 1229754 | 42.228 | ng | 93 |
| 58) Fluorene | 15.178 | 166 | 3939328 | 39.047 | ng | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 14.754 | 232 | 1182265 | 45.184 | ng | 95 |
| 60) Diethylphthalate | 14.972 | 149 | 4165539 | 40.440 | ng | 99 |
| 61) 4-Chlorophenyl-phenyle... | 15.184 | 204 | 2000871 | 40.075 | ng | 98 |
| 62) 4-Nitroaniline | 15.195 | 138 | 1022502 | 45.492 | ng | 96 |
| 63) Azobenzene | 15.484 | 77 | 3419319 | 37.279 | ng | 97 |
| 65) 4,6-Dinitro-2-methylph... | 15.260 | 198 | 703358 | 52.312 | ng | 91 |
| 66) n-Nitrosodiphenylamine | 15.401 | 169 | 3443735 | 41.168 | ng | 99 |
| 67) 4-Bromophenyl-phenylether | 16.107 | 248 | 1320894 | 44.136 | ng | 98 |
| 68) Hexachlorobenzene | 16.213 | 284 | 1555286 | 42.517 | ng | 93 |
| 69) Atrazine | 16.395 | 200 | 1102501 | 44.978 | ng | 99 |
| 70) Pentachlorophenol | 16.566 | 266 | 1159837 | 46.949 | ng | 99 |
| 71) Phenanthrene | 16.966 | 178 | 6022930 | 38.563 | ng | 99 |
| 72) Anthracene | 17.048 | 178 | 5998866 | 40.035 | ng | 100 |
| 73) Carbazole | 17.336 | 167 | 5833948 | 39.821 | ng | 99 |
| 74) Di-n-butylphthalate | 17.966 | 149 | 7240637 | 44.381 | ng | 99 |
| 75) Fluoranthene | 19.072 | 202 | 7332841 | 40.166 | ng | 99 |
| 77) Benzidine | 19.272 | 184 | 3039209 | 57.401 | ng | 98 |
| 78) Pyrene | 19.448 | 202 | 7585115 | 40.850 | ng | 100 |
| 80) Butylbenzylphthalate | 20.424 | 149 | 3246304 | 45.579 | ng | 99 |
| 81) Benzo(a)anthracene | 21.336 | 228 | 7249296 | 39.902 | ng | 100 |
| 82) 3,3'-Dichlorobenzidine | 21.271 | 252 | 2734732 | 50.003 | ng | 100 |
| 83) Chrysene | 21.401 | 228 | 6824902 | 39.037 | ng | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 21.319 | 149 | 4748475 | 48.728 | ng | 98 |
| 85) Di-n-octyl phthalate | 22.542 | 149 | 8135393 | 45.843 | ng | 97 |
| 87) Indeno(1,2,3-cd)pyrene | 28.030 | 276 | 9428099m | 42.621 | ng | |
| 88) Benzo(b)fluoranthene | 23.501 | 252 | 7966965 | 39.419 | ng | 100 |
| 89) Benzo(k)fluoranthene | 23.571 | 252 | 7588345 | 37.735 | ng | 99 |
| 90) Benzo(a)pyrene | 24.336 | 252 | 7086765 | 41.198 | ng | 99 |
| 91) Dibenzo(a,h)anthracene | 28.101 | 278 | 7827780 | 42.809 | ng | 98 |
| 92) Benzo(g,h,i)perylene | 29.118 | 276 | 7916064 | 41.908 | ng | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025155.D
Acq On : 16 Jul 2025 12:42
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jul 16 14:38:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.443 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:152 Resp: 56499

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 158.0 | 126.1 | 189.1 |
| 115 | 55.9 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#2
1,4-Dioxane
Concen: 37.074 ng
RT: 3.119 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 88 Resp: 459705

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 67.2 | 55.0 | 82.4 |
| 43 | 26.0 | 21.9 | 32.9 |






#3
Pyridine
Concen: 38.154 ng
RT: 3.490 min Scan# 94
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 79 Resp: 129521

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 64.2 | 52.2 | 78.4 |
| 51 | 32.2 | 26.8 | 40.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#4
n-Nitrosodimethylamine
Concen: 35.626 ng
RT: 3.408 min Scan# 80
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 42 Resp: 512069

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 144.0 | 105.2 | 157.8 |
| 44 | 7.4 | 6.6 | 9.8 |






#5
2-Fluorophenol
Concen: 79.902 ng
RT: 5.102 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:112 Resp: 254788

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 50.4 | 42.0 | 63.0 |
| 63 | 26.1 | 22.2 | 33.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#6
Aniline
Concen: 39.303 ng
RT: 6.784 min Scan# 654
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 93 Resp: 1528083

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 35.9 | 28.9 | 43.3 |
| 65 | 18.3 | 15.2 | 22.8 |






#7
Phenol-d6
Concen: 79.090 ng
RT: 6.637 min Scan# 6
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 99 Resp: 329115

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 15.8 | 13.6 | 20.4 |
| 71 | 30.1 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#8
2-Chlorophenol
Concen: 40.568 ng
RT: 7.019 min Scan# 694
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:128 Resp: 1447361

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.7 | 12.9 | 52.9 |
| 64 | 38.0 | 20.3 | 60.3 |






#9
Benzaldehyde
Concen: 32.488 ng
RT: 6.602 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 77 Resp: 70602

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 106.7 | 82.4 | 122.4 |
| 106 | 102.6 | 78.3 | 118.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#10
Phenol
Concen: 38.780 ng
RT: 6.666 min Scan# 634
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 94 Resp: 1639681

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 24.6 | 6.9 | 46.9 |
| 66 | 32.8 | 14.6 | 54.6 |






#11
bis(2-Chloroethyl)ether
Concen: 38.418 ng
RT: 6.884 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 93 Resp: 125012

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 70.1 | 51.9 | 91.9 |
| 95 | 32.7 | 12.4 | 52.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#12
1,3-Dichlorobenzene
Concen: 39.280 ng
RT: 7.331 min Scan# 747
Delta R.T. -0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:146 Resp: 1568447

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.3 | 51.9 | 77.9 |
| 75 | 25.5 | 21.8 | 32.6 |






#13
1,4-Dichlorobenzene
Concen: 39.630 ng
RT: 7.478 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:146 Resp: 160213

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.5 | 51.1 | 76.7 |
| 111 | 37.1 | 31.2 | 46.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#14
1,2-Dichlorobenzene
Concen: 39.152 ng
RT: 7.784 min Scan# 824
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:146 Resp: 1522581

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.7 | 51.1 | 76.7 |
| 111 | 39.0 | 31.2 | 46.8 |






#15
Benzyl Alcohol
Concen: 40.592 ng
RT: 7.678 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 79 Resp: 112739

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 86.3 | 66.0 | 99.0 |
| 77 | 62.7 | 53.4 | 80.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 36.122 ng
RT: 7.972 min Scan# 856
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 45 Resp: 1528381

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 15.3 | 0.0 | 34.8 |
| 79 | 11.8 | 0.0 | 30.7 |






#17
2-Methylphenol
Concen: 40.646 ng
RT: 7.884 min Scan# 84
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:107 Resp: 110704

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 114.0 | 89.7 | 134.5 |
| 77 | 41.4 | 34.2 | 51.4 |
| 79 | 41.9 | 33.4 | 50.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#18
Hexachloroethane
Concen: 40.630 ng
RT: 8.502 min Scan# 946
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:117 Resp: 535091

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 95.3 | 76.6 | 115.0 |
| 201 | 110.0 | 85.4 | 128.2 |






#19
n-Nitroso-di-n-propylamine
Concen: 39.280 ng
RT: 8.237 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 70 Resp: 96936

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 55.0 | 46.5 | 69.7 |
| 101 | 10.5 | 7.9 | 11.9 |
| 130 | 25.2 | 19.3 | 28.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#20
3+4-Methylphenols
Concen: 40.071 ng
RT: 8.207 min Scan# 896
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:107 Resp: 1528669

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 91.1 | 69.5 | 109.5 |
| 77 | 30.1 | 11.4 | 51.4 |
| 79 | 26.5 | 6.9 | 46.9 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.196 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:136 Resp: 223757

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.9 | 8.7 | 13.1 |
| 54 | 6.7 | 5.4 | 8.2 |
| 68 | 5.5 | 4.5 | 6.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#22
Acetophenone
Concen: 39.045 ng
RT: 8.243 min Scan# 902
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:105 Resp: 2066167

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 7.9 | 5.4 | 8.2 |
| 51 | 23.4 | 19.6 | 29.4 |
| 120 | 23.2 | 18.8 | 28.2 |






#23
Nitrobenzene-d5
Concen: 83.838 ng
RT: 8.572 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 82 Resp: 282076

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 51.5 | 38.4 | 57.6 |
| 54 | 44.9 | 37.8 | 56.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#24
Nitrobenzene
Concen: 40.716 ng
RT: 8.613 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 77 Resp: 1395349

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 53.6 | 39.0 | 58.6 |
| 65 | 13.0 | 10.7 | 16.1 |






#25
Isophorone
Concen: 40.280 ng
RT: 9.143 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 82 Resp: 253266

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.1 | 5.8 | 8.6 |
| 138 | 20.4 | 15.7 | 23.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#26
2-Nitrophenol
Concen: 48.687 ng
RT: 9.313 min Scan# 1084
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:139 Resp: 810527

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 26.1 | 22.3 | 33.5 |
| 65 | 39.9 | 33.8 | 50.8 |






#27
2,4-Dimethylphenol
Concen: 40.971 ng
RT: 9.390 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:122 Resp: 96095

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 109.5 | 88.2 | 132.2 |
| 121 | 58.8 | 46.1 | 69.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#28
bis(2-Chloroethoxy)methane
Concen: 39.712 ng
RT: 9.619 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 93 Resp: 1695445

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 31.9 | 26.4 | 39.6 |
| 123 | 12.0 | 10.1 | 15.1 |






#29
2,4-Dichlorophenol
Concen: 43.726 ng
RT: 9.854 min Scan# 1
Delta R.T. 0.018 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:162 Resp: 142209

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 65.0 | 44.1 | 84.1 |
| 98 | 31.2 | 13.1 | 53.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#30
1,2,4-Trichlorobenzene
Concen: 41.157 ng
RT: 10.060 min Scan# 1211
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:180 Resp: 1482294

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 94.6 | 75.1 | 112.7 |
| 145 | 28.2 | 23.3 | 34.9 |






#31
Naphthalene
Concen: 39.315 ng
RT: 10.249 min Scan# 1
Delta R.T. 0.018 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:128 Resp: 452975

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 10.9 | 8.9 | 13.3 |
| 127 | 13.4 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#32
Benzoic acid
Concen: 50.471 ng
RT: 9.543 min Scan# 1123
Delta R.T. 0.042 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:122 Resp: 1191237

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 116.1 | 101.0 | 141.0 |
| 77 | 80.7 | 67.6 | 107.6 |






#33
4-Chloroaniline
Concen: 40.686 ng
RT: 10.348 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:127 Resp: 174941

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 33.1 | 25.4 | 38.2 |
| 65 | 25.1 | 21.0 | 31.6 |
| 92 | 17.1 | 13.4 | 20.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#34
Hexachlorobutadiene
Concen: 42.006 ng
RT: 10.543 min Scan# 1293
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:225 Resp: 868186

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 63.2 | 50.3 | 75.5 |
| 227 | 65.0 | 51.3 | 76.9 |






#35
Caprolactam
Concen: 42.071 ng
RT: 11.143 min Scan# 1
Delta R.T. 0.035 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:113 Resp: 46808

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 140.3 | 128.1 | 168.1 |
| 56 | 104.8 | 90.7 | 130.7 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025







#36
4-Chloro-3-methylphenol
Concen: 42.541 ng
RT: 11.501 min Scan# 1456
Delta R.T. 0.018 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:107 Resp: 1445852

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 28.2 | 22.4 | 33.6 |
| 142 | 87.8 | 69.6 | 104.4 |






#37
2-Methylnaphthalene
Concen: 39.967 ng
RT: 11.872 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:142 Resp: 324912

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 89.1 | 71.2 | 106.8 |
| 115 | 31.7 | 25.9 | 38.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#38
1-Methylnaphthalene
Concen: 39.900 ng
RT: 12.090 min Scan# 1556
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:142 Resp: 3163495

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 92.2 | 74.1 | 111.1 |
| 116 | 3.5 | 2.7 | 4.1 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.095 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:164 Resp: 146507

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 100.2 | 80.3 | 120.5 |
| 160 | 44.1 | 35.9 | 53.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.626 ng
RT: 12.254 min Scan# 1584
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:216 Resp: 1752724

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.2 | 62.4 | 93.6 |
| 179 | 20.5 | 16.5 | 24.7 |
| 108 | 15.7 | 13.7 | 20.5 |






#41
Hexachlorocyclopentadiene
Concen: 48.784 ng
RT: 12.243 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:237 Resp: 53654

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 62.1 | 41.6 | 81.6 |
| 272 | 14.3 | 0.0 | 33.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#42
2,4,6-Tribromophenol
Concen: 97.740 ng
RT: 15.631 min Scan# 2158
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:330 Resp: 1691627

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 96.9 | 77.7 | 116.5 |
| 141 | 27.8 | 26.0 | 39.0 |






#43
2,4,6-Trichlorophenol
Concen: 46.244 ng
RT: 12.490 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:196 Resp: 117902

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.7 | 76.2 | 114.2 |
| 200 | 30.6 | 24.3 | 36.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#44
2,4,5-Trichlorophenol
Concen: 44.191 ng
RT: 12.572 min Scan# 1638
Delta R.T. 0.018 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:196 Resp: 1276875

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.2 | 76.6 | 115.0 |
| 97 | 42.3 | 36.6 | 55.0 |
| 132 | 26.6 | 21.4 | 32.0 |






#45
2-Fluorobiphenyl
Concen: 77.043 ng
RT: 12.695 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:172 Resp: 728608

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.5 | 28.2 | 42.2 |
| 170 | 23.8 | 18.7 | 28.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#46
1,1'-Biphenyl
Concen: 39.223 ng
RT: 12.913 min Scan# 1696
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:154 Resp: 4181055

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.8 | 20.2 | 60.2 |
| 76 | 11.2 | 0.0 | 31.5 |






#47
2-Chloronaphthalene
Concen: 40.068 ng
RT: 12.948 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:162 Resp: 321023

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 36.8 | 29.4 | 44.2 |
| 164 | 32.6 | 26.4 | 39.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#48
2-Nitroaniline
Concen: 42.274 ng
RT: 13.148 min Scan# 1736
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 65 Resp: 822809

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 75.3 | 58.7 | 88.1 |
| 138 | 132.8 | 99.0 | 148.4 |






#49
Acenaphthylene
Concen: 40.174 ng
RT: 13.801 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:152 Resp: 492638

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 19.9 | 16.1 | 24.1 |
| 153 | 13.3 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#50
Dimethylphthalate
Concen: 40.672 ng
RT: 13.560 min Scan# 1806
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:163 Resp: 4171124

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.1 | 3.4 | 5.0 |
| 164 | 10.3 | 8.0 | 12.0 |






#51
2,6-Dinitrotoluene
Concen: 44.961 ng
RT: 13.672 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:165 Resp: 89928

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 43.2 | 38.5 | 57.7 |
| 89 | 47.7 | 40.6 | 61.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025







#52
Acenaphthene
Concen: 39.159 ng
RT: 14.160 min Scan# 1908
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:154 Resp: 3094975

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 112.0 | 90.6 | 136.0 |
| 152 | 53.5 | 43.8 | 65.8 |






#53
3-Nitroaniline
Concen: 45.249 ng
RT: 14.001 min Scan# 1
Delta R.T. 0.018 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:138 Resp: 96992

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 9.6 | 7.8 | 11.8 |
| 92 | 100.6 | 81.0 | 121.4 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025







#54
2,4-Dinitrophenol
Concen: 53.458 ng
RT: 14.213 min Scan# 1917
Delta R.T. 0.018 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:184 Resp: 503837

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 48.8 | 47.4 | 71.0 |
| 154 | 57.9 | 47.8 | 71.8 |






#55
Dibenzofuran
Concen: 39.103 ng
RT: 14.513 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:168 Resp: 510038

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 36.2 | 28.6 | 43.0 |
| 169 | 13.5 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#56
4-Nitrophenol
Concen: 44.110 ng
RT: 14.337 min Scan# 1938
Delta R.T. 0.035 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:139 Resp: 872559

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 64.0 | 45.9 | 85.9 |
| 65 | 77.0 | 61.4 | 101.4 |






#57
2,4-Dinitrotoluene
Concen: 42.228 ng
RT: 14.478 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:165 Resp: 122975

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 34.0 | 31.0 | 46.4 |
| 89 | 60.7 | 53.4 | 80.0 |
| 182 | 2.8 | 2.0 | 3.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#58
Fluorene
Concen: 39.047 ng
RT: 15.178 min Scan# 2081
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:166 Resp: 3939328

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 97.1 | 78.6 | 117.8 |
| 167 | 13.9 | 10.7 | 16.1 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 45.184 ng
RT: 14.754 min Scan# 2
Delta R.T. 0.024 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:232 Resp: 118226

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 40.6 | 35.7 | 53.5 |
| 130 | 2.2 | 1.8 | 2.6 |
| 166 | 28.8 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#60
Diethylphthalate
Concen: 40.440 ng
RT: 14.972 min Scan# 2046
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:149 Resp: 4165539

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 23.1 | 17.8 | 26.8 |
| 150 | 12.4 | 10.2 | 15.2 |






#61
4-Chlorophenyl-phenylether
Concen: 40.075 ng
RT: 15.184 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:204 Resp: 200087

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 33.2 | 25.8 | 38.6 |
| 141 | 55.2 | 45.1 | 67.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#62
4-Nitroaniline
Concen: 45.492 ng
RT: 15.195 min Scan# 2084
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:138 Resp: 1022502

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 45.0 | 26.0 | 66.0 |
| 108 | 65.4 | 50.0 | 90.0 |






#63
Azobenzene
Concen: 37.279 ng
RT: 15.484 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion: 77 Resp: 341931

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 29.8 | 7.7 | 47.7 |
| 105 | 23.7 | 1.9 | 41.9 |
| 51 | 26.9 | 8.2 | 48.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.930 min Scan# 2379
Delta R.T. 0.023 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:188 Resp: 2922940

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 8.7 | 7.3 | 10.9 |
| 80 | 9.0 | 7.8 | 11.6 |






#65
4,6-Dinitro-2-methylphenol
Concen: 52.312 ng
RT: 15.260 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:198 Resp: 70335

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 31.0 | 16.8 | 56.8 |
| 105 | 35.9 | 20.8 | 60.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#66
n-Nitrosodiphenylamine
Concen: 41.168 ng
RT: 15.401 min Scan# 2119
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:169 Resp: 3443735

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 67.1 | 54.1 | 81.1 |
| 167 | 36.7 | 29.0 | 43.6 |






#67
4-Bromophenyl-phenylether
Concen: 44.136 ng
RT: 16.107 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:248 Resp: 132089

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 97.2 | 79.4 | 119.2 |
| 141 | 65.3 | 51.4 | 77.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#68
Hexachlorobenzene
Concen: 42.517 ng
RT: 16.213 min Scan# 2257
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:284 Resp: 1555286

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 26.3 | 25.9 | 38.9 |
| 249 | 28.6 | 23.8 | 35.8 |






#69
Atrazine
Concen: 44.978 ng
RT: 16.395 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:200 Resp: 110250

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 27.1 | 6.6 | 46.6 |
| 215 | 48.7 | 28.0 | 68.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#70
Pentachlorophenol
Concen: 46.949 ng
RT: 16.566 min Scan# 2317
Delta R.T. 0.018 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:266 Resp: 1159837

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 63.5 | 50.4 | 75.6 |
| 264 | 61.6 | 50.2 | 75.2 |






#71
Phenanthrene
Concen: 38.563 ng
RT: 16.966 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:178 Resp: 602293

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.6 | 15.4 | 23.2 |
| 179 | 15.9 | 12.6 | 19.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#72
Anthracene
Concen: 40.035 ng
RT: 17.048 min Scan# 2399
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:178 Resp: 5998866

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.8 | 15.0 | 22.6 |
| 179 | 15.8 | 12.5 | 18.7 |






#73
Carbazole
Concen: 39.821 ng
RT: 17.336 min Scan# 24
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:167 Resp: 583394

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.8 | 17.4 | 26.0 |
| 139 | 12.6 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#74
Di-n-butylphthalate
Concen: 44.381 ng
RT: 17.966 min Scan# 2555
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:149 Resp: 7240637

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.5 | 7.3 | 10.9 |
| 104 | 4.4 | 3.8 | 5.6 |






#75
Fluoranthene
Concen: 40.166 ng
RT: 19.072 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:202 Resp: 733284

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 10.7 | 0.0 | 30.4 |
| 203 | 18.0 | 0.0 | 37.6 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.360 min Scan# 3132
Delta R.T. 0.018 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:240 Resp: 2997869

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 10.1 | 9.0 | 13.4 |
| 236 | 25.8 | 20.3 | 30.5 |






#77
Benzidine
Concen: 57.401 ng
RT: 19.272 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:184 Resp: 303920

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 14.1 | 12.6 | 18.8 |
| 183 | 12.1 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#78
Pyrene
Concen: 40.850 ng
RT: 19.448 min Scan# 2807
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:202 Resp: 7585115

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 21.3 | 17.1 | 25.7 |
| 203 | 17.8 | 14.1 | 21.1 |






#79
Terphenyl-d14
Concen: 80.210 ng
RT: 19.677 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:244 Resp:1094089

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 8.1 | 5.9 | 8.9 |
| 122 | 12.1 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#80
Butylbenzylphthalate
Concen: 45.579 ng
RT: 20.424 min Scan# 2973
Delta R.T. -0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:149 Resp: 3246304

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 64.3 | 50.6 | 76.0 |
| 206 | 22.1 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 39.902 ng
RT: 21.336 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:228 Resp: 724929

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 27.6 | 22.1 | 33.1 |
| 229 | 20.4 | 16.0 | 24.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025







#82
3,3'-Dichlorobenzidine
Concen: 50.003 ng
RT: 21.271 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:252 Resp: 2734732

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 64.6 | 51.9 | 77.9 |
| 126 | 9.6 | 8.1 | 12.1 |






#83
Chrysene
Concen: 39.037 ng
RT: 21.401 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:228 Resp: 682490

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 30.7 | 24.4 | 36.6 |
| 229 | 20.3 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 48.728 ng
RT: 21.319 min Scan# 3125
Delta R.T. -0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:149 Resp: 4748475

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 28.7 | 21.9 | 32.9 |
| 279 | 4.9 | 3.5 | 5.3 |






#85
Di-n-octyl phthalate
Concen: 45.843 ng
RT: 22.542 min Scan# 3
Delta R.T. -0.018 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:149 Resp: 813539

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.6 | 1.3 | 1.9 |
| 43 | 7.5 | 6.9 | 10.3 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 24.501 min Scan# 3666
Delta R.T. 0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:264 Resp: 3472368

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.8 | 18.6 | 28.0 |
| 265 | 21.5 | 17.7 | 26.5 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 42.621 ng m
RT: 28.030 min Scan# 4
Delta R.T. -0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:276 Resp: 942809

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 16.6 | 13.0 | 19.6 |
| 277 | 21.0 | 15.9 | 23.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#88
Benzo(b)fluoranthene
Concen: 39.419 ng
RT: 23.501 min Scan# 3496
Delta R.T. -0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:252 Resp: 7966965

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.4 | 17.7 | 26.5 |
| 125 | 10.0 | 8.1 | 12.1 |






#89
Benzo(k)fluoranthene
Concen: 37.735 ng
RT: 23.571 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:252 Resp: 758834

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.8 | 17.7 | 26.5 |
| 125 | 10.0 | 8.2 | 12.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**
Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#90
Benzo(a)pyrene
Concen: 41.198 ng
RT: 24.336 min Scan# 3638
Delta R.T. -0.018 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:252 Resp: 7086765

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.1 | 17.6 | 26.4 |
| 125 | 10.5 | 9.0 | 13.4 |






#91
Dibenzo(a,h)anthracene
Concen: 42.809 ng
RT: 28.101 min Scan# 4
Delta R.T. -0.035 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:278 Resp: 782778

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 15.6 | 13.1 | 19.7 |
| 279 | 24.3 | 18.9 | 28.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/17/2025
Supervised By :Jagrut Upadhyay 07/17/2025




#92
Benzo(g,h,i)perylene
Concen: 41.908 ng
RT: 29.118 min Scan# 4451
Delta R.T. -0.006 min
Lab File: BP025155.D
Acq: 16 Jul 2025 12:42

Tgt Ion:276 Resp: 7916064

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.3 | 19.3 | 28.9 |
| 138 | 20.7 | 16.9 | 25.3 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025155.D
Acq On : 16 Jul 2025 12:42
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 16 14:38:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 1 | I | 1,4-Dichlorobenzene-d4 | 1.000 | 1.000 | 0.0 | 145 | 0.00 |
| 2 | | 1,4-Dioxane | 0.439 | 0.407 | 7.3 | 135 | 0.00 |
| 3 | | Pyridine | 1.202 | 1.146 | 4.7 | 135 | 0.00 |
| 4 | | n-Nitrosodimethylamine | 0.509 | 0.453 | 11.0 | 129 | 0.00 |
| 5 | S | 2-Fluorophenol | 1.129 | 1.127 | 0.2 | 144 | 0.00 |
| 6 | | Aniline | 1.376 | 1.352 | 1.7 | 141 | 0.00 |
| 7 | S | Phenol-d6 | 1.473 | 1.456 | 1.2 | 144 | 0.00 |
| 8 | | 2-Chlorophenol | 1.263 | 1.281 | -1.4 | 147 | 0.00 |
| 9 | | Benzaldehyde | 0.769 | 0.625 | 18.7 | 125 | 0.00 |
| 10 | C | Phenol | 1.497 | 1.451 | 3.1 | 142 | 0.01 |
| 11 | | bis(2-Chloroethyl)ether | 1.152 | 1.106 | 4.0 | 143 | 0.00 |
| 12 | | 1,3-Dichlorobenzene | 1.413 | 1.388 | 1.8 | 146 | 0.00 |
| 13 | C | 1,4-Dichlorobenzene | 1.431 | 1.418 | 0.9 | 146 | 0.00 |
| 14 | | 1,2-Dichlorobenzene | 1.377 | 1.347 | 2.2 | 144 | 0.00 |
| 15 | | Benzyl Alcohol | 0.983 | 0.998 | -1.5 | 151# | 0.00 |
| 16 | | 2,2'-oxybis(1-Chloropropane | 1.498 | 1.353 | 9.7 | 136 | 0.00 |
| 17 | | 2-Methylphenol | 0.964 | 0.980 | -1.7 | 146 | 0.01 |
| 18 | | Hexachloroethane | 0.466 | 0.474 | -1.7 | 149 | 0.00 |
| 19 | P | n-Nitroso-di-n-propylamine | 0.874 | 0.858 | 1.8 | 143 | 0.00 |
| 20 | | 3+4-Methylphenols | 1.350 | 1.353 | -0.2 | 147 | 0.01 |
| 21 | I | Naphthalene-d8 | 1.000 | 1.000 | 0.0 | 143 | 0.01 |
| 22 | | Acetophenone | 0.473 | 0.462 | 2.3 | 142 | 0.00 |
| 23 | S | Nitrobenzene-d5 | 0.301 | 0.315 | -4.7 | 147 | 0.00 |
| 24 | | Nitrobenzene | 0.306 | 0.312 | -2.0 | 144 | 0.00 |
| 25 | | Isophorone | 0.562 | 0.566 | -0.7 | 145 | 0.01 |
| 26 | C | 2-Nitrophenol | 0.128 | 0.181 | -41.4# | 195# | 0.00 |
| 27 | | 2,4-Dimethylphenol | 0.210 | 0.215 | -2.4 | 148 | 0.00 |
| 28 | | bis(2-Chloroethoxy)methane | 0.382 | 0.379 | 0.8 | 145 | 0.00 |
| 29 | C | 2,4-Dichlorophenol | 0.291 | 0.318 | -9.3 | 152# | 0.02 |
| 30 | | 1,2,4-Trichlorobenzene | 0.322 | 0.331 | -2.8 | 149 | 0.00 |
| 31 | | Naphthalene | 1.030 | 1.012 | 1.7 | 144 | 0.02 |
| 32 | | Benzoic acid | 0.185 | 0.266 | -43.8# | 205# | 0.04 |
| 33 | | 4-Chloroaniline | 0.384 | 0.391 | -1.8 | 144 | 0.00 |
| 34 | C | Hexachlorobutadiene | 0.185 | 0.194 | -4.9 | 153# | 0.00 |
| 35 | | Caprolactam | 0.099 | 0.105 | -6.1 | 149 | 0.04 |
| 36 | C | 4-Chloro-3-methylphenol | 0.304 | 0.323 | -6.3 | 150 | 0.02 |
| 37 | | 2-Methylnaphthalene | 0.727 | 0.726 | 0.1 | 145 | 0.01 |
| 38 | | 1-Methylnaphthalene | 0.709 | 0.707 | 0.3 | 147 | 0.00 |
| 39 | I | Acenaphthene-d10 | 1.000 | 1.000 | 0.0 | 146 | 0.00 |
| 40 | | 1,2,4,5-Tetrachlorobenzene | 0.575 | 0.598 | -4.0 | 151# | 0.01 |
| 41 | P | Hexachlorocyclopentadiene | 0.150 | 0.183 | -22.0 | 179# | 0.00 |
| 42 | S | 2,4,6-Tribromophenol | 0.236 | 0.289 | -22.5 | 171# | 0.01 |
| 43 | C | 2,4,6-Trichlorophenol | 0.348 | 0.402 | -15.5 | 159# | 0.00 |
| 44 | | 2,4,5-Trichlorophenol | 0.394 | 0.436 | -10.7 | 155# | 0.02 |
| 45 | S | 2-Fluorobiphenyl | 1.291 | 1.243 | 3.7 | 142 | 0.00 |
| 46 | | 1,1'-Biphenyl | 1.455 | 1.427 | 1.9 | 143 | 0.00 |
| 47 | | 2-Chloronaphthalene | 1.094 | 1.096 | -0.2 | 147 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025155.D
Acq On : 16 Jul 2025 12:42
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 16 14:38:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 48 | | 2-Nitroaniline | 0.237 | 0.281 | -18.6 | 158# | 0.00 |
| 49 | | Acenaphthylene | 1.674 | 1.681 | -0.4 | 144 | 0.00 |
| 50 | | Dimethylphthalate | 1.400 | 1.424 | -1.7 | 148 | 0.00 |
| 51 | | 2,6-Dinitrotoluene | 0.273 | 0.307 | -12.5 | 152# | 0.01 |
| 52 | C | Acenaphthene | 1.079 | 1.056 | 2.1 | 145 | 0.00 |
| 53 | | 3-Nitroaniline | 0.293 | 0.331 | -13.0 | 150# | 0.02 |
| 54 | P | 2,4-Dinitrophenol | 0.114 | 0.172 | -50.9# | 215# | 0.02 |
| 55 | | Dibenzofuran | 1.781 | 1.741 | 2.2 | 144 | 0.00 |
| 56 | P | 4-Nitrophenol | 0.270 | 0.298 | -10.4 | 152# | 0.04 |
| 57 | | 2,4-Dinitrotoluene | 0.356 | 0.420 | -18.0 | 157# | 0.01 |
| 58 | | Fluorene | 1.377 | 1.344 | 2.4 | 141 | 0.01 |
| 59 | | 2,3,4,6-Tetrachlorophenol | 0.357 | 0.403 | -12.9 | 159# | 0.02 |
| 60 | | Diethylphthalate | 1.406 | 1.422 | -1.1 | 145 | 0.01 |
| 61 | | 4-Chlorophenyl-phenylether | 0.682 | 0.683 | -0.1 | 147 | 0.00 |
| 62 | | 4-Nitroaniline | 0.307 | 0.349 | -13.7 | 149 | 0.01 |
| 63 | | Azobenzene | 1.252 | 1.167 | 6.8 | 135 | 0.00 |
| 64 | I | Phenanthrene-d10 | 1.000 | 1.000 | 0.0 | 142 | 0.02 |
| 65 | | 4,6-Dinitro-2-methylphenol | 0.082 | 0.120 | -46.3# | 209# | 0.01 |
| 66 | c | n-Nitrosodiphenylamine | 0.572 | 0.589 | -3.0 | 146 | 0.01 |
| 67 | | 4-Bromophenyl-phenylether | 0.205 | 0.226 | -10.2 | 157# | 0.01 |
| 68 | | Hexachlorobenzene | 0.250 | 0.266 | -6.4 | 152# | 0.01 |
| 69 | | Atrazine | 0.168 | 0.189 | -12.5 | 155# | 0.01 |
| 70 | C | Pentachlorophenol | 0.169 | 0.198 | -17.2 | 164# | 0.02 |
| 71 | | Phenanthrene | 1.069 | 1.030 | 3.6 | 137 | 0.01 |
| 72 | | Anthracene | 1.025 | 1.026 | -0.1 | 140 | 0.00 |
| 73 | | Carbazole | 1.002 | 0.998 | 0.4 | 137 | 0.01 |
| 74 | | Di-n-butylphthalate | 1.116 | 1.239 | -11.0 | 148 | 0.00 |
| 75 | C | Fluoranthene | 1.249 | 1.254 | -0.4 | 138 | 0.00 |
| 76 | I | Chrysene-d12 | 1.000 | 1.000 | 0.0 | 132 | 0.02 |
| 77 | | Benzidine | 0.353 | 0.507 | -43.6# | 118 | 0.00 |
| 78 | | Pyrene | 1.239 | 1.265 | -2.1 | 136 | 0.01 |
| 79 | S | Terphenyl-d14 | 0.910 | 0.912 | -0.2 | 132 | 0.00 |
| 80 | | Butylbenzylphthalate | 0.402 | 0.541 | -34.6# | 163# | 0.00 |
| 81 | | Benzo(a)anthracene | 1.212 | 1.209 | 0.2 | 132 | 0.01 |
| 82 | | 3,3'-Dichlorobenzidine | 0.365 | 0.456 | -24.9 | 155# | 0.00 |
| 83 | | Chrysene | 1.166 | 1.138 | 2.4 | 130 | 0.01 |
| 84 | | Bis(2-ethylhexyl)phthalate | 0.650 | 0.792 | -21.8 | 149 | -0.01 |
| 85 | c | Di-n-octyl phthalate | 1.054 | 1.357 | -28.7# | 164# | -0.02 |
| 86 | I | Perylene-d12 | 1.000 | 1.000 | 0.0 | 143 | 0.00 |
| 87 | | Indeno(1,2,3-cd)pyrene | 1.274 | 1.358 | -6.6 | 146 | 0.00 |
| 88 | | Benzo(b)fluoranthene | 1.164 | 1.147 | 1.5 | 136 | -0.01 |
| 89 | | Benzo(k)fluoranthene | 1.158 | 1.093 | 5.6 | 135 | -0.01 |
| 90 | C | Benzo(a)pyrene | 0.991 | 1.020 | -2.9 | 141 | -0.02 |
| 91 | | Dibenzo(a,h)anthracene | 1.053 | 1.127 | -7.0 | 147 | -0.04 |
| 92 | | Benzo(g,h,i)perylene | 1.088 | 1.140 | -4.8 | 144 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025155.D
Acq On : 16 Jul 2025 12:42
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 16 14:38:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | AvgRF | CCRF | %Dev Area% | Dev(min) |
|---|---|---|---|---|

(#) = Out of Range SPCC's out = 0 CCC's out = 2

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025155.D
Acq On : 16 Jul 2025 12:42
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 16 14:38:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 1 | I | 1,4-Dichlorobenzene-d4 | 20.000 | 20.000 | 0.0 | 145 | 0.00 |
| 2 | | 1,4-Dioxane | 40.000 | 37.074 | 7.3 | 135 | 0.00 |
| 3 | | Pyridine | 40.000 | 38.154 | 4.6 | 135 | 0.00 |
| 4 | | n-Nitrosodimethylamine | 40.000 | 35.626 | 10.9 | 129 | 0.00 |
| 5 | S | 2-Fluorophenol | 80.000 | 79.902 | 0.1 | 144 | 0.00 |
| 6 | | Aniline | 40.000 | 39.303 | 1.7 | 141 | 0.00 |
| 7 | S | Phenol-d6 | 80.000 | 79.090 | 1.1 | 144 | 0.00 |
| 8 | | 2-Chlorophenol | 40.000 | 40.568 | -1.4 | 147 | 0.00 |
| 9 | | Benzaldehyde | 40.000 | 32.488 | 18.8 | 125 | 0.00 |
| 10 | C | Phenol | 40.000 | 38.780 | 3.0 | 142 | 0.01 |
| 11 | | bis(2-Chloroethyl)ether | 40.000 | 38.418 | 4.0 | 143 | 0.00 |
| 12 | | 1,3-Dichlorobenzene | 40.000 | 39.280 | 1.8 | 146 | 0.00 |
| 13 | C | 1,4-Dichlorobenzene | 40.000 | 39.630 | 0.9 | 146 | 0.00 |
| 14 | | 1,2-Dichlorobenzene | 40.000 | 39.152 | 2.1 | 144 | 0.00 |
| 15 | | Benzyl Alcohol | 40.000 | 40.592 | -1.5 | 151 | 0.00 |
| 16 | | 2,2'-oxybis(1-Chloropropane | 40.000 | 36.122 | 9.7 | 136 | 0.00 |
| 17 | | 2-Methylphenol | 40.000 | 40.646 | -1.6 | 146 | 0.01 |
| 18 | | Hexachloroethane | 40.000 | 40.630 | -1.6 | 149 | 0.00 |
| 19 | P | n-Nitroso-di-n-propylamine | 40.000 | 39.280 | 1.8 | 143 | 0.00 |
| 20 | | 3+4-Methylphenols | 40.000 | 40.071 | -0.2 | 147 | 0.01 |
| 21 | I | Naphthalene-d8 | 20.000 | 20.000 | 0.0 | 143 | 0.01 |
| 22 | | Acetophenone | 40.000 | 39.045 | 2.4 | 142 | 0.00 |
| 23 | S | Nitrobenzene-d5 | 80.000 | 83.838 | -4.8 | 147 | 0.00 |
| 24 | | Nitrobenzene | 40.000 | 40.716 | -1.8 | 144 | 0.00 |
| 25 | | Isophorone | 40.000 | 40.280 | -0.7 | 145 | 0.01 |
| 26 | C | 2-Nitrophenol | 40.000 | 48.687 | -21.7# | 195 | 0.00 |
| 27 | | 2,4-Dimethylphenol | 40.000 | 40.971 | -2.4 | 148 | 0.00 |
| 28 | | bis(2-Chloroethoxy)methane | 40.000 | 39.712 | 0.7 | 145 | 0.00 |
| 29 | C | 2,4-Dichlorophenol | 40.000 | 43.726 | -9.3 | 152 | 0.02 |
| 30 | | 1,2,4-Trichlorobenzene | 40.000 | 41.157 | -2.9 | 149 | 0.00 |
| 31 | | Naphthalene | 40.000 | 39.315 | 1.7 | 144 | 0.02 |
| 32 | | Benzoic acid | 40.000 | 50.471 | -26.2# | 205 | 0.04 |
| 33 | | 4-Chloroaniline | 40.000 | 40.686 | -1.7 | 144 | 0.00 |
| 34 | C | Hexachlorobutadiene | 40.000 | 42.006 | -5.0 | 153 | 0.00 |
| 35 | | Caprolactam | 40.000 | 42.071 | -5.2 | 149 | 0.04 |
| 36 | C | 4-Chloro-3-methylphenol | 40.000 | 42.541 | -6.4 | 150 | 0.02 |
| 37 | | 2-Methylnaphthalene | 40.000 | 39.967 | 0.1 | 145 | 0.01 |
| 38 | | 1-Methylnaphthalene | 40.000 | 39.900 | 0.3 | 147 | 0.00 |
| 39 | I | Acenaphthene-d10 | 20.000 | 20.000 | 0.0 | 146 | 0.00 |
| 40 | | 1,2,4,5-Tetrachlorobenzene | 40.000 | 41.626 | -4.1 | 151 | 0.01 |
| 41 | P | Hexachlorocyclopentadiene | 40.000 | 48.784 | -22.0 | 179 | 0.00 |
| 42 | S | 2,4,6-Tribromophenol | 80.000 | 97.740 | -22.2 | 171 | 0.01 |
| 43 | C | 2,4,6-Trichlorophenol | 40.000 | 46.244 | -15.6 | 159 | 0.00 |
| 44 | | 2,4,5-Trichlorophenol | 40.000 | 44.191 | -10.5 | 155 | 0.02 |
| 45 | S | 2-Fluorobiphenyl | 80.000 | 77.043 | 3.7 | 142 | 0.00 |
| 46 | | 1,1'-Biphenyl | 40.000 | 39.223 | 1.9 | 143 | 0.00 |
| 47 | | 2-Chloronaphthalene | 40.000 | 40.068 | -0.2 | 147 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025155.D
Acq On : 16 Jul 2025 12:42
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 16 14:38:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 48 | | 2-Nitroaniline | 40.000 | 42.274 | -5.7 | 158 | 0.00 |
| 49 | | Acenaphthylene | 40.000 | 40.174 | -0.4 | 144 | 0.00 |
| 50 | | Dimethylphthalate | 40.000 | 40.672 | -1.7 | 148 | 0.00 |
| 51 | | 2,6-Dinitrotoluene | 40.000 | 44.961 | -12.4 | 152 | 0.01 |
| 52 | C | Acenaphthene | 40.000 | 39.159 | 2.1 | 145 | 0.00 |
| 53 | | 3-Nitroaniline | 40.000 | 45.249 | -13.1 | 150 | 0.02 |
| 54 | P | 2,4-Dinitrophenol | 40.000 | 53.458 | -33.6# | 215 | 0.02 |
| 55 | | Dibenzofuran | 40.000 | 39.103 | 2.2 | 144 | 0.00 |
| 56 | P | 4-Nitrophenol | 40.000 | 44.110 | -10.3 | 152 | 0.04 |
| 57 | | 2,4-Dinitrotoluene | 40.000 | 42.228 | -5.6 | 157 | 0.01 |
| 58 | | Fluorene | 40.000 | 39.047 | 2.4 | 141 | 0.01 |
| 59 | | 2,3,4,6-Tetrachlorophenol | 40.000 | 45.184 | -13.0 | 159 | 0.02 |
| 60 | | Diethylphthalate | 40.000 | 40.440 | -1.1 | 145 | 0.01 |
| 61 | | 4-Chlorophenyl-phenylether | 40.000 | 40.075 | -0.2 | 147 | 0.00 |
| 62 | | 4-Nitroaniline | 40.000 | 45.492 | -13.7 | 149 | 0.01 |
| 63 | | Azobenzene | 40.000 | 37.279 | 6.8 | 135 | 0.00 |
| 64 | I | Phenanthrene-d10 | 20.000 | 20.000 | 0.0 | 142 | 0.02 |
| 65 | | 4,6-Dinitro-2-methylphenol | 40.000 | 52.312 | -30.8# | 209 | 0.01 |
| 66 | c | n-Nitrosodiphenylamine | 40.000 | 41.168 | -2.9 | 146 | 0.01 |
| 67 | | 4-Bromophenyl-phenylether | 40.000 | 44.136 | -10.3 | 157 | 0.01 |
| 68 | | Hexachlorobenzene | 40.000 | 42.517 | -6.3 | 152 | 0.01 |
| 69 | | Atrazine | 40.000 | 44.978 | -12.4 | 155 | 0.01 |
| 70 | C | Pentachlorophenol | 40.000 | 46.949 | -17.4 | 164 | 0.02 |
| 71 | | Phenanthrene | 40.000 | 38.563 | 3.6 | 137 | 0.01 |
| 72 | | Anthracene | 40.000 | 40.035 | -0.1 | 140 | 0.00 |
| 73 | | Carbazole | 40.000 | 39.821 | 0.4 | 137 | 0.01 |
| 74 | | Di-n-butylphthalate | 40.000 | 44.381 | -11.0 | 148 | 0.00 |
| 75 | C | Fluoranthene | 40.000 | 40.166 | -0.4 | 138 | 0.00 |
| 76 | I | Chrysene-d12 | 20.000 | 20.000 | 0.0 | 132 | 0.02 |
| 77 | | Benzidine | 40.000 | 57.401 | -43.5# | 118 | 0.00 |
| 78 | | Pyrene | 40.000 | 40.850 | -2.1 | 136 | 0.01 |
| 79 | S | Terphenyl-d14 | 80.000 | 80.210 | -0.3 | 132 | 0.00 |
| 80 | | Butylbenzylphthalate | 40.000 | 45.579 | -13.9 | 163 | 0.00 |
| 81 | | Benzo(a)anthracene | 40.000 | 39.902 | 0.2 | 132 | 0.01 |
| 82 | | 3,3'-Dichlorobenzidine | 40.000 | 50.003 | -25.0# | 155 | 0.00 |
| 83 | | Chrysene | 40.000 | 39.037 | 2.4 | 130 | 0.01 |
| 84 | | Bis(2-ethylhexyl)phthalate | 40.000 | 48.728 | -21.8 | 149 | -0.01 |
| 85 | c | Di-n-octyl phthalate | 40.000 | 45.843 | -14.6 | 164 | -0.02 |
| 86 | I | Perylene-d12 | 20.000 | 20.000 | 0.0 | 143 | 0.00 |
| 87 | | Indeno(1,2,3-cd)pyrene | 40.000 | 42.621 | -6.6 | 146 | 0.00 |
| 88 | | Benzo(b)fluoranthene | 40.000 | 39.419 | 1.5 | 136 | -0.01 |
| 89 | | Benzo(k)fluoranthene | 40.000 | 37.735 | 5.7 | 135 | -0.01 |
| 90 | C | Benzo(a)pyrene | 40.000 | 41.198 | -3.0 | 141 | -0.02 |
| 91 | | Dibenzo(a,h)anthracene | 40.000 | 42.809 | -7.0 | 147 | -0.04 |
| 92 | | Benzo(g,h,i)perylene | 40.000 | 41.908 | -4.8 | 144 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025155.D
Acq On : 16 Jul 2025 12:42
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 16 14:38:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|

(#) = Out of Range SPCC's out = 0 CCC's out = 1

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: PARS02

Lab Code: ACE SDG No.: Q2592

Instrument ID: BNA_P Calibration Date/Time: 07/18/2025 11:04

Lab File ID: BP025187.D Init. Calib. Date(s): 07/03/2025 07/03/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:39 14:31

GC Column: ZB-GR ID: 0.25 (mm)

| COMPOUND | RRF | RRF040 | MIN RRF | %D | MAX%D |
|---|---|---|---|---|---|
| 2-Fluorophenol | 1.129 | 1.060 | | -6.1 | |
| Benzaldehyde | 0.769 | 0.712 | | -7.4 | |
| Phenol-d6 | 1.473 | 1.366 | | -7.3 | |
| Phenol | 1.497 | 1.351 | | -9.8 | 20.0 |
| bis(2-Chloroethyl)ether | 1.152 | 1.027 | | -10.9 | |
| 2-Chlorophenol | 1.263 | 1.211 | | -4.1 | |
| 2-Methylphenol | 0.964 | 0.922 | | -4.4 | |
| 2,2-oxybis(1-Chloropropane) | 1.498 | 1.251 | | -16.5 | |
| Acetophenone | 0.473 | 0.425 | | -10.1 | |
| 3+4-Methylphenols | 1.350 | 1.285 | | -4.8 | |
| n-Nitroso-di-n-propylamine | 0.874 | 0.816 | 0.050 | -6.6 | |
| Nitrobenzene-d5 | 0.301 | 0.288 | | -4.3 | |
| Hexachloroethane | 0.466 | 0.455 | | -2.4 | |
| Nitrobenzene | 0.306 | 0.288 | | -5.9 | |
| Isophorone | 0.562 | 0.529 | | -5.9 | |
| 2-Nitrophenol | 0.128 | 0.172 | | 34.4 | 20.0 |
| 2,4-Dimethylphenol | 0.210 | 0.200 | | -4.8 | |
| bis(2-Chloroethoxy)methane | 0.382 | 0.353 | | -7.6 | |
| 2,4-Dichlorophenol | 0.291 | 0.291 | | 0.0 | 20.0 |
| Naphthalene | 1.030 | 0.936 | | -9.1 | |
| 4-Chloroaniline | 0.384 | 0.359 | | -6.5 | |
| Hexachlorobutadiene | 0.185 | 0.186 | | 0.5 | 20.0 |
| Caprolactam | 0.099 | 0.097 | | -2.0 | |
| 4-Chloro-3-methylphenol | 0.304 | 0.296 | | -2.6 | 20.0 |
| 2-Methylnaphthalene | 0.727 | 0.675 | | -7.2 | |
| Hexachlorocyclopentadiene | 0.150 | 0.168 | 0.050 | 12.0 | |
| 2,4,6-Trichlorophenol | 0.348 | 0.368 | | 5.7 | 20.0 |
| 2-Fluorobiphenyl | 1.291 | 1.166 | | -9.7 | |
| 2,4,5-Trichlorophenol | 0.394 | 0.406 | | 3.0 | |
| 1,1-Biphenyl | 1.455 | 1.318 | | -9.4 | |
| 2-Chloronaphthalene | 1.094 | 0.997 | | -8.9 | |
| 2-Nitroaniline | 0.237 | 0.256 | | 8.0 | |
| Dimethylphthalate | 1.400 | 1.304 | | -6.9 | |
| Acenaphthylene | 1.674 | 1.527 | | -8.8 | |
| 2,6-Dinitrotoluene | 0.273 | 0.286 | | 4.8 | |
| 3-Nitroaniline | 0.293 | 0.306 | | 4.4 | |
| Acenaphthene | 1.079 | 0.968 | | -10.3 | 20.0 |
| 2,4-Dinitrophenol | 0.114 | 0.169 | 0.050 | 48.2 | |
| 4-Nitrophenol | 0.270 | 0.276 | 0.050 | 2.2 | |

Form VII SV-1

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: PARS02

Lab Code: ACE SDG No.: Q2592

Instrument ID: BNA_P Calibration Date/Time: 07/18/2025 11:04

Lab File ID: BP025187.D Init. Calib. Date(s): 07/03/2025 07/03/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:39 14:31

GC Column: ZB-GR ID: 0.25 (mm)

| COMPOUND | RRF | RRF040 | MIN RRF | %D | MAX%D |
|---|---|---|---|---|---|
| Dibenzofuran | 1.781 | 1.614 | | -9.4 | |
| 2,4-Dinitrotoluene | 0.356 | 0.392 | | 10.1 | |
| Diethylphthalate | 1.406 | 1.340 | | -4.7 | |
| 4-Chlorophenyl-phenylether | 0.682 | 0.626 | | -8.2 | |
| Fluorene | 1.377 | 1.237 | | -10.2 | |
| 4-Nitroaniline | 0.307 | 0.328 | | 6.8 | |
| 4,6-Dinitro-2-methylphenol | 0.082 | 0.114 | | 39.0 | |
| n-Nitrosodiphenylamine | 0.572 | 0.539 | | -5.8 | 20.0 |
| 2,4,6-Tribromophenol | 0.236 | 0.273 | | 15.7 | |
| 4-Bromophenyl-phenylether | 0.205 | 0.211 | | 2.9 | |
| Hexachlorobenzene | 0.250 | 0.249 | | -0.4 | |
| Atrazine | 0.168 | 0.146 | | -13.1 | |
| Pentachlorophenol | 0.169 | 0.185 | | 9.5 | 20.0 |
| Phenanthrene | 1.069 | 0.966 | | -9.6 | |
| Anthracene | 1.025 | 0.955 | | -6.8 | |
| Carbazole | 1.002 | 0.943 | | -5.9 | |
| Di-n-butylphthalate | 1.116 | 1.193 | | 6.9 | |
| Fluoranthene | 1.249 | 1.197 | | -4.2 | 20.0 |
| Pyrene | 1.239 | 1.139 | | -8.1 | |
| Terphenyl-d14 | 0.910 | 0.855 | | -6.0 | |
| Butylbenzylphthalate | 0.402 | 0.513 | | 27.6 | |
| 3,3-Dichlorobenzidine | 0.365 | 0.436 | | 19.5 | |
| Benzo(a)anthracene | 1.212 | 1.128 | | -6.9 | |
| Chrysene | 1.166 | 1.080 | | -7.4 | |
| Bis(2-ethylhexyl)phthalate | 0.650 | 0.772 | | 18.8 | |
| Di-n-octyl phthalate | 1.054 | 1.342 | | 27.3 | 20.0 |
| Benzo(b)fluoranthene | 1.164 | 1.062 | | -8.8 | |
| Benzo(k)fluoranthene | 1.158 | 1.036 | | -10.5 | |
| Benzo(a)pyrene | 0.991 | 0.943 | | -4.8 | 20.0 |
| Indeno(1,2,3-cd)pyrene | 1.274 | 1.283 | | 0.7 | |
| Dibenzo(a,h)anthracene | 1.053 | 1.055 | | 0.2 | |
| Benzo(g,h,i)perylene | 1.088 | 1.062 | | -2.4 | |
| 1,2,4,5-Tetrachlorobenzene | 0.575 | 0.552 | | -4.0 | |
| 1,4-Dioxane | 0.439 | 0.381 | | -13.2 | 20.0 |
| 2,3,4,6-Tetrachlorophenol | 0.357 | 0.373 | | 4.5 | |

All other compounds must meet a minimum RRF of 0.010.

Form VII SV-1

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025187.D
Acq On : 18 Jul 2025 11:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jul 18 11:41:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| **Internal Standards** | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.437 | 152 | 531610 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.190 | 136 | 2146082 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.101 | 164 | 1427548 | 20.000 | ng | 0.01 |
| 64) Phenanthrene-d10 | 16.925 | 188 | 2869297 | 20.000 | ng | 0.02 |
| 76) Chrysene-d12 | 21.348 | 240 | 3091728 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 24.495 | 264 | 3620682 | 20.000 | ng | 0.00 |
| **System Monitoring Compounds** | | | | | | |
| 5) 2-Fluorophenol | 5.102 | 112 | 2254632 | 75.147 | ng | 0.00 |
| 7) Phenol-d6 | 6.637 | 99 | 2904926 | 74.193 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 8.572 | 82 | 2469800 | 76.537 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.625 | 330 | 1558969 | 92.443 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 12.701 | 172 | 6660491 | 72.279 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.672 | 244 | 10572248 | 75.154 | ng | 0.00 |
| **Target Compounds** | | | | | | Qvalue |
| 2) 1,4-Dioxane | 3.120 | 88 | 404589 | 34.679 | ng | 97 |
| 3) Pyridine | 3.490 | 79 | 1119432 | 35.046 | ng | 98 |
| 4) n-Nitrosodimethylamine | 3.402 | 42 | 447383 | 33.080 | ng | 90 |
| 6) Aniline | 6.784 | 93 | 1318748 | 36.049 | ng | 99 |
| 8) 2-Chlorophenol | 7.013 | 128 | 1287776 | 38.362 | ng | 98 |
| 9) Benzaldehyde | 6.596 | 77 | 756528 | 36.998 | ng | 97 |
| 10) Phenol | 6.666 | 94 | 1436852 | 36.117 | ng | 97 |
| 11) bis(2-Chloroethyl)ether | 6.878 | 93 | 1091520 | 35.650 | ng | 99 |
| 12) 1,3-Dichlorobenzene | 7.331 | 146 | 1369686 | 36.457 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 7.472 | 146 | 1393382 | 36.631 | ng | 98 |
| 14) 1,2-Dichlorobenzene | 7.784 | 146 | 1352041 | 36.950 | ng | 99 |
| 15) Benzyl Alcohol | 7.678 | 79 | 1004233 | 38.428 | ng | 96 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.960 | 45 | 1330526 | 33.421 | ng | 98 |
| 17) 2-Methylphenol | 7.884 | 107 | 980663 | 38.267 | ng | 99 |
| 18) Hexachloroethane | 8.496 | 117 | 483508 | 39.019 | ng | 98 |
| 19) n-Nitroso-di-n-propyla... | 8.237 | 70 | 867767 | 37.371 | ng | 96 |
| 20) 3+4-Methylphenols | 8.208 | 107 | 1365751 | 38.048 | ng | 99 |
| 22) Acetophenone | 8.243 | 105 | 1825851 | 35.975 | ng | 98 |
| 24) Nitrobenzene | 8.613 | 77 | 1235576 | 37.591 | ng | 95 |
| 25) Isophorone | 9.137 | 82 | 2269615 | 37.635 | ng | 99 |
| 26) 2-Nitrophenol | 9.313 | 139 | 737848 | 46.379 | ng | 96 |
| 27) 2,4-Dimethylphenol | 9.390 | 122 | 858307 | 38.154 | ng | 98 |
| 28) bis(2-Chloroethoxy)met... | 9.619 | 93 | 1516406 | 37.033 | ng | 99 |
| 29) 2,4-Dichlorophenol | 9.849 | 162 | 1250689 | 40.095 | ng | 99 |
| 30) 1,2,4-Trichlorobenzene | 10.054 | 180 | 1315259 | 38.076 | ng | 97 |
| 31) Naphthalene | 10.237 | 128 | 4018592 | 36.365 | ng | 100 |
| 32) Benzoic acid | 9.537 | 122 | 1059596 | 47.258 | ng | 96 |
| 33) 4-Chloroaniline | 10.343 | 127 | 1539023 | 37.319 | ng | 99 |
| 34) Hexachlorobutadiene | 10.537 | 225 | 798092 | 40.261 | ng | 99 |
| 35) Caprolactam | 11.125 | 113 | 416773 | 39.056 | ng | 96 |
| 36) 4-Chloro-3-methylphenol | 11.490 | 107 | 1272371 | 39.033 | ng | 98 |
| 37) 2-Methylnaphthalene | 11.860 | 142 | 2899215 | 37.184 | ng | 99 |
| 38) 1-Methylnaphthalene | 12.090 | 142 | 2804639 | 36.882 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 12.243 | 216 | 1574766 | 38.383 | ng | 99 |
| 41) Hexachlorocyclopentadiene | 12.225 | 237 | 479972 | 44.788 | ng | 98 |
| 43) 2,4,6-Trichlorophenol | 12.496 | 196 | 1051946 | 42.344 | ng | 98 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025187.D
Acq On : 18 Jul 2025 11:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jul 18 11:41:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 12.566 | 196 | 1160384 | 41.215 | ng | | 97 |
| 46) 1,1'-Biphenyl | 12.901 | 154 | 3763681 | 36.236 | ng | | 99 |
| 47) 2-Chloronaphthalene | 12.943 | 162 | 2846065 | 36.457 | ng | | 99 |
| 48) 2-Nitroaniline | 13.148 | 65 | 729633 | 38.721 | ng | | 93 |
| 49) Acenaphthylene | 13.807 | 152 | 4359655 | 36.487 | ng | | 100 |
| 50) Dimethylphthalate | 13.560 | 163 | 3721862 | 37.245 | ng | | 99 |
| 51) 2,6-Dinitrotoluene | 13.672 | 165 | 817897 | 41.967 | ng | | 90 |
| 52) Acenaphthene | 14.160 | 154 | 2763754 | 35.888 | ng | | 99 |
| 53) 3-Nitroaniline | 14.001 | 138 | 872454 | 41.772 | ng | | 98 |
| 54) 2,4-Dinitrophenol | 14.219 | 184 | 483431 | 52.724 | ng | # | 89 |
| 55) Dibenzofuran | 14.507 | 168 | 4608837 | 36.263 | ng | | 99 |
| 56) 4-Nitrophenol | 14.337 | 139 | 787961 | 40.881 | ng | | 96 |
| 57) 2,4-Dinitrotoluene | 14.478 | 165 | 1117932 | 39.574 | ng | | 91 |
| 58) Fluorene | 15.178 | 166 | 3532450 | 35.934 | ng | | 99 |
| 59) 2,3,4,6-Tetrachlorophenol | 14.742 | 232 | 1065574 | 41.795 | ng | | 95 |
| 60) Diethylphthalate | 14.966 | 149 | 3824481 | 38.105 | ng | | 99 |
| 61) 4-Chlorophenyl-phenyle... | 15.184 | 204 | 1786967 | 36.732 | ng | | 96 |
| 62) 4-Nitroaniline | 15.195 | 138 | 937172 | 42.792 | ng | | 94 |
| 63) Azobenzene | 15.484 | 77 | 3086011 | 34.529 | ng | | 96 |
| 65) 4,6-Dinitro-2-methylph... | 15.260 | 198 | 656883 | 50.040 | ng | | 99 |
| 66) n-Nitrosodiphenylamine | 15.395 | 169 | 3093337 | 37.671 | ng | | 99 |
| 67) 4-Bromophenyl-phenylether | 16.095 | 248 | 1210335 | 41.198 | ng | | 99 |
| 68) Hexachlorobenzene | 16.207 | 284 | 1431153 | 39.855 | ng | | 98 |
| 69) Atrazine | 16.395 | 200 | 839410 | 34.885 | ng | | 99 |
| 70) Pentachlorophenol | 16.572 | 266 | 1062970 | 43.832 | ng | | 99 |
| 71) Phenanthrene | 16.966 | 178 | 5542867 | 36.153 | ng | | 100 |
| 72) Anthracene | 17.054 | 178 | 5481414 | 37.266 | ng | | 100 |
| 73) Carbazole | 17.342 | 167 | 5410553 | 37.621 | ng | | 99 |
| 74) Di-n-butylphthalate | 17.966 | 149 | 6847818 | 42.758 | ng | | 99 |
| 75) Fluoranthene | 19.066 | 202 | 6866434 | 38.314 | ng | | 99 |
| 77) Benzidine | 19.272 | 184 | 3389454 | 62.073 | ng | | 97 |
| 78) Pyrene | 19.442 | 202 | 7042387 | 36.776 | ng | | 100 |
| 80) Butylbenzylphthalate | 20.425 | 149 | 3174538 | 43.409 | ng | | 96 |
| 81) Benzo(a)anthracene | 21.330 | 228 | 6977868 | 37.242 | ng | | 99 |
| 82) 3,3'-Dichlorobenzidine | 21.266 | 252 | 2693075 | 47.747 | ng | | 99 |
| 83) Chrysene | 21.395 | 228 | 6675791 | 37.025 | ng | | 100 |
| 84) Bis(2-ethylhexyl)phtha... | 21.313 | 149 | 4774117 | 47.504 | ng | | 98 |
| 85) Di-n-octyl phthalate | 22.530 | 149 | 8298951 | 45.411 | ng | | 97 |
| 87) Indeno(1,2,3-cd)pyrene | 28.042 | 276 | 9288491m | 40.270 | ng | | |
| 88) Benzo(b)fluoranthene | 23.513 | 252 | 7692853 | 36.504 | ng | | 99 |
| 89) Benzo(k)fluoranthene | 23.577 | 252 | 7500297 | 35.769 | ng | | 99 |
| 90) Benzo(a)pyrene | 24.348 | 252 | 6827776 | 38.067 | ng | | 100 |
| 91) Dibenzo(a,h)anthracene | 28.142 | 278 | 7639998 | 40.071 | ng | | 98 |
| 92) Benzo(g,h,i)perylene | 29.136 | 276 | 7688711 | 39.037 | ng | | 100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025187.D
Acq On : 18 Jul 2025 11:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jul 18 11:41:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.437 min Scan# 7
Delta R.T. -0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:152 Resp: 53161

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 155.4 | 126.1 | 189.1 |
| 115 | 53.1 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#2
1,4-Dioxane
Concen: 34.679 ng
RT: 3.120 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 88 Resp: 404589

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 66.7 | 55.0 | 82.4 |
| 43 | 25.6 | 21.9 | 32.9 |






#3
Pyridine
Concen: 35.046 ng
RT: 3.490 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 79 Resp: 111943

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 64.2 | 52.2 | 78.4 |
| 51 | 32.1 | 26.8 | 40.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#4
n-Nitrosodimethylamine
Concen: 33.080 ng
RT: 3.402 min Scan# 79
Delta R.T. -0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 42 Resp: 447383

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 144.0 | 105.2 | 157.8 |
| 44 | 7.4 | 6.6 | 9.8 |






#5
2-Fluorophenol
Concen: 75.147 ng
RT: 5.102 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:112 Resp: 225463

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 49.5 | 42.0 | 63.0 |
| 63 | 25.1 | 22.2 | 33.2 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#6
Aniline
Concen: 36.049 ng
RT: 6.784 min Scan# 654
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 93 Resp: 1318748

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 35.4 | 28.9 | 43.3 |
| 65 | 18.1 | 15.2 | 22.8 |






#7
Phenol-d6
Concen: 74.193 ng
RT: 6.637 min Scan# 6
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 99 Resp: 290492

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 15.7 | 13.6 | 20.4 |
| 71 | 30.5 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#8
2-Chlorophenol
Concen: 38.362 ng
RT: 7.013 min Scan# 693
Delta R.T. -0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:128 Resp: 1287776

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.9 | 12.9 | 52.9 |
| 64 | 38.5 | 20.3 | 60.3 |






#9
Benzaldehyde
Concen: 36.998 ng
RT: 6.596 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 77 Resp: 75652

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 105.8 | 82.4 | 122.4 |
| 106 | 101.9 | 78.3 | 118.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#10
Phenol
Concen: 36.117 ng
RT: 6.666 min Scan# 634
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 94 Resp: 1436852

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 25.0 | 6.9 | 46.9 |
| 66 | 33.3 | 14.6 | 54.6 |






#11
bis(2-Chloroethyl)ether
Concen: 35.650 ng
RT: 6.878 min Scan# 6
Delta R.T. -0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 93 Resp: 109152

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 70.4 | 51.9 | 91.9 |
| 95 | 32.6 | 12.4 | 52.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#12
1,3-Dichlorobenzene
Concen: 36.457 ng
RT: 7.331 min Scan# 747
Delta R.T. -0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:146 Resp: 1369686

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 65.1 | 51.9 | 77.9 |
| 75 | 25.9 | 21.8 | 32.6 |






#13
1,4-Dichlorobenzene
Concen: 36.631 ng
RT: 7.472 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:146 Resp: 139338

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 65.0 | 51.1 | 76.7 |
| 111 | 37.9 | 31.2 | 46.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#14
1,2-Dichlorobenzene
Concen: 36.950 ng
RT: 7.784 min Scan# 824
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:146 Resp: 1352041

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.6 | 51.1 | 76.7 |
| 111 | 39.1 | 31.2 | 46.8 |






#15
Benzyl Alcohol
Concen: 38.428 ng
RT: 7.678 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 79 Resp: 100423

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 85.6 | 66.0 | 99.0 |
| 77 | 63.3 | 53.4 | 80.2 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 33.421 ng
RT: 7.960 min Scan# 854
Delta R.T. -0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 45 Resp: 1330526

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 15.6 | 0.0 | 34.8 |
| 79 | 11.9 | 0.0 | 30.7 |






#17
2-Methylphenol
Concen: 38.267 ng
RT: 7.884 min Scan# 84
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:107 Resp: 98066

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 111.9 | 89.7 | 134.5 |
| 77 | 41.3 | 34.2 | 51.4 |
| 79 | 41.5 | 33.4 | 50.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#18
Hexachloroethane
Concen: 39.019 ng
RT: 8.496 min Scan# 945
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:117 Resp: 483508

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 95.8 | 76.6 | 115.0 |
| 201 | 110.6 | 85.4 | 128.2 |






#19
n-Nitroso-di-n-propylamine
Concen: 37.371 ng
RT: 8.237 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 70 Resp: 86776

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 54.9 | 46.5 | 69.7 |
| 101 | 10.6 | 7.9 | 11.9 |
| 130 | 25.5 | 19.3 | 28.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#20
3+4-Methylphenols
Concen: 38.048 ng
RT: 8.208 min Scan# 896
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:107 Resp: 1365751

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 90.2 | 69.5 | 109.5 |
| 77 | 29.9 | 11.4 | 51.4 |
| 79 | 26.5 | 6.9 | 46.9 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.190 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:136 Resp: 214608

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.9 | 8.7 | 13.1 |
| 54 | 6.5 | 5.4 | 8.2 |
| 68 | 5.3 | 4.5 | 6.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#22
Acetophenone
Concen: 35.975 ng
RT: 8.243 min Scan# 902
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:105 Resp: 1825851

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 7.0 | 5.4 | 8.2 |
| 51 | 23.0 | 19.6 | 29.4 |
| 120 | 23.0 | 18.8 | 28.2 |






#23
Nitrobenzene-d5
Concen: 76.537 ng
RT: 8.572 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 82 Resp: 246980

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 50.7 | 38.4 | 57.6 |
| 54 | 45.3 | 37.8 | 56.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#24
Nitrobenzene
Concen: 37.591 ng
RT: 8.613 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 77 Resp: 1235576

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 53.0 | 39.0 | 58.6 |
| 65 | 12.9 | 10.7 | 16.1 |






#25
Isophorone
Concen: 37.635 ng
RT: 9.137 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 82 Resp: 226961

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.0 | 5.8 | 8.6 |
| 138 | 20.2 | 15.7 | 23.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#26
2-Nitrophenol
Concen: 46.379 ng
RT: 9.313 min Scan# 1084
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:139 Resp: 737848

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 25.6 | 22.3 | 33.5 |
| 65 | 39.5 | 33.8 | 50.8 |






#27
2,4-Dimethylphenol
Concen: 38.154 ng
RT: 9.390 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:122 Resp: 85830

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 108.1 | 88.2 | 132.2 |
| 121 | 58.6 | 46.1 | 69.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#28
bis(2-Chloroethoxy)methane
Concen: 37.033 ng
RT: 9.619 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 93 Resp: 1516406

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.6 | 26.4 | 39.6 |
| 123 | 11.9 | 10.1 | 15.1 |






#29
2,4-Dichlorophenol
Concen: 40.095 ng
RT: 9.849 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:162 Resp: 125068

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 65.2 | 44.1 | 84.1 |
| 98 | 32.3 | 13.1 | 53.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#30
1,2,4-Trichlorobenzene
Concen: 38.076 ng
RT: 10.054 min Scan# 1210
Delta R.T. -0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:180 Resp: 1315259

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 97.4 | 75.1 | 112.7 |
| 145 | 29.4 | 23.3 | 34.9 |






#31
Naphthalene
Concen: 36.365 ng
RT: 10.237 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:128 Resp: 401859

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.1 | 8.9 | 13.3 |
| 127 | 13.0 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#32
Benzoic acid
Concen: 47.258 ng
RT: 9.537 min Scan# 1122
Delta R.T. 0.036 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:122 Resp: 1059596

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 118.4 | 101.0 | 141.0 |
| 77 | 82.3 | 67.6 | 107.6 |






#33
4-Chloroaniline
Concen: 37.319 ng
RT: 10.343 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:127 Resp: 153902

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 31.9 | 25.4 | 38.2 |
| 65 | 25.6 | 21.0 | 31.6 |
| 92 | 17.6 | 13.4 | 20.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#34
Hexachlorobutadiene
Concen: 40.261 ng
RT: 10.537 min Scan# 1292
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:225 Resp: 798092

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 61.7 | 50.3 | 75.5 |
| 227 | 65.0 | 51.3 | 76.9 |






#35
Caprolactam
Concen: 39.056 ng
RT: 11.125 min Scan# 1
Delta R.T. 0.018 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:113 Resp: 41677

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 142.5 | 128.1 | 168.1 |
| 56 | 106.8 | 90.7 | 130.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#36
4-Chloro-3-methylphenol
Concen: 39.033 ng
RT: 11.490 min Scan# 1454
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:107 Resp: 1272371

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 29.0 | 22.4 | 33.6 |
| 142 | 89.1 | 69.6 | 104.4 |






#37
2-Methylnaphthalene
Concen: 37.184 ng
RT: 11.860 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:142 Resp: 289921

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 88.6 | 71.2 | 106.8 |
| 115 | 31.8 | 25.9 | 38.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#38
1-Methylnaphthalene
Concen: 36.882 ng
RT: 12.090 min Scan# 1556
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:142 Resp: 2804639

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 92.7 | 74.1 | 111.1 |
| 116 | 3.5 | 2.7 | 4.1 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.101 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:164 Resp: 142754

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 100.8 | 80.3 | 120.5 |
| 160 | 43.6 | 35.9 | 53.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 38.383 ng
RT: 12.243 min Scan# 1582
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:216 Resp: 1574766

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.0 | 62.4 | 93.6 |
| 179 | 20.2 | 16.5 | 24.7 |
| 108 | 15.8 | 13.7 | 20.5 |






#41
Hexachlorocyclopentadiene
Concen: 44.788 ng
RT: 12.225 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:237 Resp: 47997

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 59.8 | 41.6 | 81.6 |
| 272 | 13.7 | 0.0 | 33.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#42
2,4,6-Tribromophenol
Concen: 92.443 ng
RT: 15.625 min Scan# 2157
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:330 Resp: 1558969

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 96.7 | 77.7 | 116.5 |
| 141 | 27.3 | 26.0 | 39.0 |






#43
2,4,6-Trichlorophenol
Concen: 42.344 ng
RT: 12.496 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:196 Resp: 105194

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.5 | 76.2 | 114.2 |
| 200 | 31.6 | 24.3 | 36.5 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#44
2,4,5-Trichlorophenol
Concen: 41.215 ng
RT: 12.566 min Scan# 1637
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:196 Resp: 1160384

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 94.4 | 76.6 | 115.0 |
| 97 | 41.1 | 36.6 | 55.0 |
| 132 | 25.9 | 21.4 | 32.0 |






#45
2-Fluorobiphenyl
Concen: 72.279 ng
RT: 12.701 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:172 Resp: 666049

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.5 | 28.2 | 42.2 |
| 170 | 23.6 | 18.7 | 28.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#46
1,1'-Biphenyl
Concen: 36.236 ng
RT: 12.901 min Scan# 1694
Delta R.T. -0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:154 Resp: 3763681

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.6 | 20.2 | 60.2 |
| 76 | 11.3 | 0.0 | 31.5 |






#47
2-Chloronaphthalene
Concen: 36.457 ng
RT: 12.943 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:162 Resp: 284606

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 37.3 | 29.4 | 44.2 |
| 164 | 33.2 | 26.4 | 39.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#48
2-Nitroaniline
Concen: 38.721 ng
RT: 13.148 min Scan# 1736
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 65 Resp: 729633

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 76.1 | 58.7 | 88.1 |
| 138 | 133.3 | 99.0 | 148.4 |






#49
Acenaphthylene
Concen: 36.487 ng
RT: 13.807 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:152 Resp: 435965

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.1 | 16.1 | 24.1 |
| 153 | 13.5 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#50
Dimethylphthalate
Concen: 37.245 ng
RT: 13.560 min Scan# 1806
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:163 Resp: 3721862

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.1 | 3.4 | 5.0 |
| 164 | 10.3 | 8.0 | 12.0 |






#51
2,6-Dinitrotoluene
Concen: 41.967 ng
RT: 13.672 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:165 Resp: 81789

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 40.5 | 38.5 | 57.7 |
| 89 | 44.8 | 40.6 | 61.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#52
Acenaphthene
Concen: 35.888 ng
RT: 14.160 min Scan# 1908
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:154 Resp: 2763754

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 113.4 | 90.6 | 136.0 |
| 152 | 53.2 | 43.8 | 65.8 |






#53
3-Nitroaniline
Concen: 41.772 ng
RT: 14.001 min Scan# 1
Delta R.T. 0.018 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:138 Resp: 87245

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 9.7 | 7.8 | 11.8 |
| 92 | 99.3 | 81.0 | 121.4 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#54
2,4-Dinitrophenol
Concen: 52.724 ng
RT: 14.219 min Scan# 1918
Delta R.T. 0.024 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:184 Resp: 483431

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 46.0 | 47.4 | 71.0# |
| 154 | 56.0 | 47.8 | 71.8 |






#55
Dibenzofuran
Concen: 36.263 ng
RT: 14.507 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:168 Resp: 460883

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 36.3 | 28.6 | 43.0 |
| 169 | 13.6 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#56
4-Nitrophenol
Concen: 40.881 ng
RT: 14.337 min Scan# 1938
Delta R.T. 0.035 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:139 Resp: 787961

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 63.8 | 45.9 | 85.9 |
| 65 | 76.5 | 61.4 | 101.4 |






#57
2,4-Dinitrotoluene
Concen: 39.574 ng
RT: 14.478 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:165 Resp: 111793

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 32.8 | 31.0 | 46.4 |
| 89 | 59.6 | 53.4 | 80.0 |
| 182 | 2.9 | 2.0 | 3.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#58
Fluorene
Concen: 35.934 ng
RT: 15.178 min Scan# 2081
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:166 Resp: 3532450

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 97.4 | 78.6 | 117.8 |
| 167 | 13.5 | 10.7 | 16.1 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 41.795 ng
RT: 14.742 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:232 Resp: 106557

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 40.8 | 35.7 | 53.5 |
| 130 | 2.1 | 1.8 | 2.6 |
| 166 | 28.8 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#60
Diethylphthalate
Concen: 38.105 ng
RT: 14.966 min Scan# 2045
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:149 Resp: 3824481

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 23.1 | 17.8 | 26.8 |
| 150 | 12.5 | 10.2 | 15.2 |






#61
4-Chlorophenyl-phenylether
Concen: 36.732 ng
RT: 15.184 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:204 Resp: 178696

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 33.2 | 25.8 | 38.6 |
| 141 | 53.0 | 45.1 | 67.7 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#62
4-Nitroaniline
Concen: 42.792 ng
RT: 15.195 min Scan# 2084
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:138 Resp: 937172

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 43.5 | 26.0 | 66.0 |
| 108 | 64.5 | 50.0 | 90.0 |






#63
Azobenzene
Concen: 34.529 ng
RT: 15.484 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion: 77 Resp: 308601

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 31.0 | 7.7 | 47.7 |
| 105 | 22.8 | 1.9 | 41.9 |
| 51 | 26.7 | 8.2 | 48.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.925 min Scan# 2378
Delta R.T. 0.018 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:188 Resp: 2869297

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 8.5 | 7.3 | 10.9 |
| 80 | 9.2 | 7.8 | 11.6 |






#65
4,6-Dinitro-2-methylphenol
Concen: 50.040 ng
RT: 15.260 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:198 Resp: 65688

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 35.9 | 16.8 | 56.8 |
| 105 | 40.5 | 20.8 | 60.8 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#66
n-Nitrosodiphenylamine
Concen: 37.671 ng
RT: 15.395 min Scan# 2118
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:169 Resp: 3093337

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 66.5 | 54.1 | 81.1 |
| 167 | 36.1 | 29.0 | 43.6 |






#67
4-Bromophenyl-phenylether
Concen: 41.198 ng
RT: 16.095 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:248 Resp: 121033

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 98.8 | 79.4 | 119.2 |
| 141 | 63.0 | 51.4 | 77.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#68
Hexachlorobenzene
Concen: 39.855 ng
RT: 16.207 min Scan# 2256
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:284 Resp: 1431153

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 30.1 | 25.9 | 38.9 |
| 249 | 30.0 | 23.8 | 35.8 |






#69
Atrazine
Concen: 34.885 ng
RT: 16.395 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:200 Resp: 83941

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 27.2 | 6.6 | 46.6 |
| 215 | 48.3 | 28.0 | 68.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#70
Pentachlorophenol
Concen: 43.832 ng
RT: 16.572 min Scan# 2318
Delta R.T. 0.024 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:266 Resp: 1062970

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 64.6 | 50.4 | 75.6 |
| 264 | 62.4 | 50.2 | 75.2 |






#71
Phenanthrene
Concen: 36.153 ng
RT: 16.966 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:178 Resp: 554286

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.3 | 15.4 | 23.2 |
| 179 | 15.6 | 12.6 | 19.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#72
Anthracene
Concen: 37.266 ng
RT: 17.054 min Scan# 2400
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:178 Resp: 5481414

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.7 | 15.0 | 22.6 |
| 179 | 15.5 | 12.5 | 18.7 |






#73
Carbazole
Concen: 37.621 ng
RT: 17.342 min Scan# 24
Delta R.T. 0.018 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:167 Resp: 541055

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.8 | 17.4 | 26.0 |
| 139 | 12.5 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#74
Di-n-butylphthalate
Concen: 42.758 ng
RT: 17.966 min Scan# 2555
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:149 Resp: 6847818

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.3 | 7.3 | 10.9 |
| 104 | 4.4 | 3.8 | 5.6 |






#75
Fluoranthene
Concen: 38.314 ng
RT: 19.066 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:202 Resp: 686643

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 10.4 | 0.0 | 30.4 |
| 203 | 18.1 | 0.0 | 37.6 |

**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.348 min Scan# 3130
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:240 Resp: 3091728

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.4 | 9.0 | 13.4 |
| 236 | 25.9 | 20.3 | 30.5 |






#77
Benzidine
Concen: 62.073 ng
RT: 19.272 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:184 Resp: 338945

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 13.7 | 12.6 | 18.8 |
| 183 | 11.8 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#78
Pyrene
Concen: 36.776 ng
RT: 19.442 min Scan# 2806
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:202 Resp: 7042387

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 21.6 | 17.1 | 25.7 |
| 203 | 17.8 | 14.1 | 21.1 |






#79
Terphenyl-d14
Concen: 75.154 ng
RT: 19.672 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:244 Resp:1057224

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 7.2 | 5.9 | 8.9 |
| 122 | 11.0 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#80
Butylbenzylphthalate
Concen: 43.409 ng
RT: 20.425 min Scan# 2973
Delta R.T. -0.005 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:149 Resp: 3174538

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 60.6 | 50.6 | 76.0 |
| 206 | 23.3 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 37.242 ng
RT: 21.330 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:228 Resp: 697786

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 27.0 | 22.1 | 33.1 |
| 229 | 20.0 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#82
3,3'-Dichlorobenzidine
Concen: 47.747 ng
RT: 21.266 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:252 Resp: 2693075

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 63.6 | 51.9 | 77.9 |
| 126 | 10.1 | 8.1 | 12.1 |






#83
Chrysene
Concen: 37.025 ng
RT: 21.395 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:228 Resp: 667579

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 30.3 | 24.4 | 36.6 |
| 229 | 20.0 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 47.504 ng
RT: 21.313 min Scan# 3124
Delta R.T. -0.018 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:149 Resp: 4774117

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 28.5 | 21.9 | 32.9 |
| 279 | 5.2 | 3.5 | 5.3 |






#85
Di-n-octyl phthalate
Concen: 45.411 ng
RT: 22.530 min Scan# 3
Delta R.T. -0.030 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:149 Resp: 829895

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.6 | 1.3 | 1.9 |
| 43 | 7.3 | 6.9 | 10.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 24.495 min Scan# 3665
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:264 Resp: 3620682

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.7 | 18.6 | 28.0 |
| 265 | 21.8 | 17.7 | 26.5 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 40.270 ng m
RT: 28.042 min Scan# 4
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:276 Resp: 928849

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 15.7 | 13.0 | 19.6 |
| 277 | 20.5 | 15.9 | 23.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#88
Benzo(b)fluoranthene
Concen: 36.504 ng
RT: 23.513 min Scan# 3498
Delta R.T. 0.000 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:252 Resp: 7692853

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.3 | 17.7 | 26.5 |
| 125 | 9.7 | 8.1 | 12.1 |






#89
Benzo(k)fluoranthene
Concen: 35.769 ng
RT: 23.577 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:252 Resp: 750029

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.4 | 17.7 | 26.5 |
| 125 | 9.6 | 8.2 | 12.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#90
Benzo(a)pyrene
Concen: 38.067 ng
RT: 24.348 min Scan# 3640
Delta R.T. -0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:252 Resp: 6827776

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.8 | 17.6 | 26.4 |
| 125 | 11.3 | 9.0 | 13.4 |






#91
Dibenzo(a,h)anthracene
Concen: 40.071 ng
RT: 28.142 min Scan# 4
Delta R.T. 0.006 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:278 Resp: 763999

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 14.9 | 13.1 | 19.7 |
| 279 | 23.6 | 18.9 | 28.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
SSTDCCC040

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#92
Benzo(g,h,i)perylene
Concen: 39.037 ng
RT: 29.136 min Scan# 4454
Delta R.T. 0.012 min
Lab File: BP025187.D
Acq: 18 Jul 2025 11:04

Tgt Ion:276 Resp: 7688711

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.9 | 19.3 | 28.9 |
| 138 | 21.1 | 16.9 | 25.3 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025187.D
Acq On : 18 Jul 2025 11:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 18 11:41:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 1 | I | 1,4-Dichlorobenzene-d4 | 1.000 | 1.000 | 0.0 | 136 | 0.00 |
| 2 | | 1,4-Dioxane | 0.439 | 0.381 | 13.2 | 119 | 0.00 |
| 3 | | Pyridine | 1.202 | 1.053 | 12.4 | 117 | 0.00 |
| 4 | | n-Nitrosodimethylamine | 0.509 | 0.421 | 17.3 | 113 | 0.00 |
| 5 | S | 2-Fluorophenol | 1.129 | 1.060 | 6.1 | 127 | 0.00 |
| 6 | | Aniline | 1.376 | 1.240 | 9.9 | 122 | 0.00 |
| 7 | S | Phenol-d6 | 1.473 | 1.366 | 7.3 | 127 | 0.00 |
| 8 | | 2-Chlorophenol | 1.263 | 1.211 | 4.1 | 131 | 0.00 |
| 9 | | Benzaldehyde | 0.769 | 0.712 | 7.4 | 134 | 0.00 |
| 10 | C | Phenol | 1.497 | 1.351 | 9.8 | 125 | 0.01 |
| 11 | | bis(2-Chloroethyl)ether | 1.152 | 1.027 | 10.9 | 124 | -0.01 |
| 12 | | 1,3-Dichlorobenzene | 1.413 | 1.288 | 8.8 | 127 | 0.00 |
| 13 | C | 1,4-Dichlorobenzene | 1.431 | 1.311 | 8.4 | 127 | 0.00 |
| 14 | | 1,2-Dichlorobenzene | 1.377 | 1.272 | 7.6 | 128 | 0.00 |
| 15 | | Benzyl Alcohol | 0.983 | 0.945 | 3.9 | 134 | 0.00 |
| 16 | | 2,2'-oxybis(1-Chloropropane | 1.498 | 1.251 | 16.5 | 119 | 0.00 |
| 17 | | 2-Methylphenol | 0.964 | 0.922 | 4.4 | 130 | 0.01 |
| 18 | | Hexachloroethane | 0.466 | 0.455 | 2.4 | 134 | 0.00 |
| 19 | P | n-Nitroso-di-n-propylamine | 0.874 | 0.816 | 6.6 | 128 | 0.00 |
| 20 | | 3+4-Methylphenols | 1.350 | 1.285 | 4.8 | 131 | 0.01 |
| 21 | I | Naphthalene-d8 | 1.000 | 1.000 | 0.0 | 138 | 0.00 |
| 22 | | Acetophenone | 0.473 | 0.425 | 10.1 | 126 | 0.00 |
| 23 | S | Nitrobenzene-d5 | 0.301 | 0.288 | 4.3 | 129 | 0.00 |
| 24 | | Nitrobenzene | 0.306 | 0.288 | 5.9 | 128 | 0.00 |
| 25 | | Isophorone | 0.562 | 0.529 | 5.9 | 130 | 0.00 |
| 26 | C | 2-Nitrophenol | 0.128 | 0.172 | -34.4# | 177# | 0.00 |
| 27 | | 2,4-Dimethylphenol | 0.210 | 0.200 | 4.8 | 132 | 0.00 |
| 28 | | bis(2-Chloroethoxy)methane | 0.382 | 0.353 | 7.6 | 130 | 0.00 |
| 29 | C | 2,4-Dichlorophenol | 0.291 | 0.291 | 0.0 | 134 | 0.01 |
| 30 | | 1,2,4-Trichlorobenzene | 0.322 | 0.306 | 5.0 | 132 | 0.00 |
| 31 | | Naphthalene | 1.030 | 0.936 | 9.1 | 127 | 0.00 |
| 32 | | Benzoic acid | 0.185 | 0.247 | -33.5# | 182# | 0.04 |
| 33 | | 4-Chloroaniline | 0.384 | 0.359 | 6.5 | 126 | 0.00 |
| 34 | C | Hexachlorobutadiene | 0.185 | 0.186 | -0.5 | 141 | 0.00 |
| 35 | | Caprolactam | 0.099 | 0.097 | 2.0 | 133 | 0.02 |
| 36 | C | 4-Chloro-3-methylphenol | 0.304 | 0.296 | 2.6 | 132 | 0.00 |
| 37 | | 2-Methylnaphthalene | 0.727 | 0.675 | 7.2 | 129 | 0.00 |
| 38 | | 1-Methylnaphthalene | 0.709 | 0.653 | 7.9 | 130 | 0.00 |
| 39 | I | Acenaphthene-d10 | 1.000 | 1.000 | 0.0 | 142 | 0.01 |
| 40 | | 1,2,4,5-Tetrachlorobenzene | 0.575 | 0.552 | 4.0 | 135 | 0.00 |
| 41 | P | Hexachlorocyclopentadiene | 0.150 | 0.168 | -12.0 | 161# | -0.01 |
| 42 | S | 2,4,6-Tribromophenol | 0.236 | 0.273 | -15.7 | 158# | 0.00 |
| 43 | C | 2,4,6-Trichlorophenol | 0.348 | 0.368 | -5.7 | 142 | 0.01 |
| 44 | | 2,4,5-Trichlorophenol | 0.394 | 0.406 | -3.0 | 141 | 0.01 |
| 45 | S | 2-Fluorobiphenyl | 1.291 | 1.166 | 9.7 | 130 | 0.00 |
| 46 | | 1,1'-Biphenyl | 1.455 | 1.318 | 9.4 | 129 | 0.00 |
| 47 | | 2-Chloronaphthalene | 1.094 | 0.997 | 8.9 | 130 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025187.D
Acq On : 18 Jul 2025 11:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 18 11:41:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 48 | | 2-Nitroaniline | 0.237 | 0.256 | -8.0 | 141 | 0.00 |
| 49 | | Acenaphthylene | 1.674 | 1.527 | 8.8 | 128 | 0.00 |
| 50 | | Dimethylphthalate | 1.400 | 1.304 | 6.9 | 132 | 0.00 |
| 51 | | 2,6-Dinitrotoluene | 0.273 | 0.286 | -4.8 | 138 | 0.01 |
| 52 | C | Acenaphthene | 1.079 | 0.968 | 10.3 | 129 | 0.00 |
| 53 | | 3-Nitroaniline | 0.293 | 0.306 | -4.4 | 135 | 0.02 |
| 54 | P | 2,4-Dinitrophenol | 0.114 | 0.169 | -48.2# | 206# | 0.02 |
| 55 | | Dibenzofuran | 1.781 | 1.614 | 9.4 | 130 | 0.00 |
| 56 | P | 4-Nitrophenol | 0.270 | 0.276 | -2.2 | 138 | 0.04 |
| 57 | | 2,4-Dinitrotoluene | 0.356 | 0.392 | -10.1 | 142 | 0.01 |
| 58 | | Fluorene | 1.377 | 1.237 | 10.2 | 126 | 0.01 |
| 59 | | 2,3,4,6-Tetrachlorophenol | 0.357 | 0.373 | -4.5 | 143 | 0.01 |
| 60 | | Diethylphthalate | 1.406 | 1.340 | 4.7 | 133 | 0.00 |
| 61 | | 4-Chlorophenyl-phenylether | 0.682 | 0.626 | 8.2 | 131 | 0.00 |
| 62 | | 4-Nitroaniline | 0.307 | 0.328 | -6.8 | 136 | 0.01 |
| 63 | | Azobenzene | 1.252 | 1.081 | 13.7 | 122 | 0.00 |
| 64 | I | Phenanthrene-d10 | 1.000 | 1.000 | 0.0 | 139 | 0.02 |
| 65 | | 4,6-Dinitro-2-methylphenol | 0.082 | 0.114 | -39.0# | 196# | 0.01 |
| 66 | c | n-Nitrosodiphenylamine | 0.572 | 0.539 | 5.8 | 131 | 0.00 |
| 67 | | 4-Bromophenyl-phenylether | 0.205 | 0.211 | -2.9 | 143 | 0.00 |
| 68 | | Hexachlorobenzene | 0.250 | 0.249 | 0.4 | 140 | 0.00 |
| 69 | | Atrazine | 0.168 | 0.146 | 13.1 | 118 | 0.01 |
| 70 | C | Pentachlorophenol | 0.169 | 0.185 | -9.5 | 150# | 0.02 |
| 71 | | Phenanthrene | 1.069 | 0.966 | 9.6 | 126 | 0.01 |
| 72 | | Anthracene | 1.025 | 0.955 | 6.8 | 128 | 0.00 |
| 73 | | Carbazole | 1.002 | 0.943 | 5.9 | 127 | 0.02 |
| 74 | | Di-n-butylphthalate | 1.116 | 1.193 | -6.9 | 140 | 0.00 |
| 75 | C | Fluoranthene | 1.249 | 1.197 | 4.2 | 129 | 0.00 |
| 76 | I | Chrysene-d12 | 1.000 | 1.000 | 0.0 | 137 | 0.00 |
| 77 | | Benzidine | 0.353 | 0.548 | -55.2# | 131 | 0.00 |
| 78 | | Pyrene | 1.239 | 1.139 | 8.1 | 126 | 0.00 |
| 79 | S | Terphenyl-d14 | 0.910 | 0.855 | 6.0 | 128 | 0.00 |
| 80 | | Butylbenzylphthalate | 0.402 | 0.513 | -27.6# | 159# | 0.00 |
| 81 | | Benzo(a)anthracene | 1.212 | 1.128 | 6.9 | 127 | 0.00 |
| 82 | | 3,3'-Dichlorobenzidine | 0.365 | 0.436 | -19.5 | 153# | 0.00 |
| 83 | | Chrysene | 1.166 | 1.080 | 7.4 | 127 | 0.00 |
| 84 | | Bis(2-ethylhexyl)phthalate | 0.650 | 0.772 | -18.8 | 150# | -0.02 |
| 85 | c | Di-n-octyl phthalate | 1.054 | 1.342 | -27.3# | 168# | -0.03 |
| 86 | I | Perylene-d12 | 1.000 | 1.000 | 0.0 | 149 | 0.00 |
| 87 | | Indeno(1,2,3-cd)pyrene | 1.274 | 1.283 | -0.7 | 143 | 0.00 |
| 88 | | Benzo(b)fluoranthene | 1.164 | 1.062 | 8.8 | 132 | 0.00 |
| 89 | | Benzo(k)fluoranthene | 1.158 | 1.036 | 10.5 | 134 | 0.00 |
| 90 | C | Benzo(a)pyrene | 0.991 | 0.943 | 4.8 | 136 | 0.00 |
| 91 | | Dibenzo(a,h)anthracene | 1.053 | 1.055 | -0.2 | 143 | 0.00 |
| 92 | | Benzo(g,h,i)perylene | 1.088 | 1.062 | 2.4 | 140 | 0.01 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025187.D
Acq On : 18 Jul 2025 11:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 18 11:41:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|

(#) = Out of Range SPCC's out = 0 CCC's out = 2

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025187.D
Acq On : 18 Jul 2025 11:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 18 11:41:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 1 | I | 1,4-Dichlorobenzene-d4 | 20.000 | 20.000 | 0.0 | 136 | 0.00 |
| 2 | | 1,4-Dioxane | 40.000 | 34.679 | 13.3 | 119 | 0.00 |
| 3 | | Pyridine | 40.000 | 35.046 | 12.4 | 117 | 0.00 |
| 4 | | n-Nitrosodimethylamine | 40.000 | 33.080 | 17.3 | 113 | 0.00 |
| 5 | S | 2-Fluorophenol | 80.000 | 75.147 | 6.1 | 127 | 0.00 |
| 6 | | Aniline | 40.000 | 36.049 | 9.9 | 122 | 0.00 |
| 7 | S | Phenol-d6 | 80.000 | 74.193 | 7.3 | 127 | 0.00 |
| 8 | | 2-Chlorophenol | 40.000 | 38.362 | 4.1 | 131 | 0.00 |
| 9 | | Benzaldehyde | 40.000 | 36.998 | 7.5 | 134 | 0.00 |
| 10 | C | Phenol | 40.000 | 36.117 | 9.7 | 125 | 0.01 |
| 11 | | bis(2-Chloroethyl)ether | 40.000 | 35.650 | 10.9 | 124 | -0.01 |
| 12 | | 1,3-Dichlorobenzene | 40.000 | 36.457 | 8.9 | 127 | 0.00 |
| 13 | C | 1,4-Dichlorobenzene | 40.000 | 36.631 | 8.4 | 127 | 0.00 |
| 14 | | 1,2-Dichlorobenzene | 40.000 | 36.950 | 7.6 | 128 | 0.00 |
| 15 | | Benzyl Alcohol | 40.000 | 38.428 | 3.9 | 134 | 0.00 |
| 16 | | 2,2'-oxybis(1-Chloropropane | 40.000 | 33.421 | 16.4 | 119 | 0.00 |
| 17 | | 2-Methylphenol | 40.000 | 38.267 | 4.3 | 130 | 0.01 |
| 18 | | Hexachloroethane | 40.000 | 39.019 | 2.5 | 134 | 0.00 |
| 19 | P | n-Nitroso-di-n-propylamine | 40.000 | 37.371 | 6.6 | 128 | 0.00 |
| 20 | | 3+4-Methylphenols | 40.000 | 38.048 | 4.9 | 131 | 0.01 |
| 21 | I | Naphthalene-d8 | 20.000 | 20.000 | 0.0 | 138 | 0.00 |
| 22 | | Acetophenone | 40.000 | 35.975 | 10.1 | 126 | 0.00 |
| 23 | S | Nitrobenzene-d5 | 80.000 | 76.537 | 4.3 | 129 | 0.00 |
| 24 | | Nitrobenzene | 40.000 | 37.591 | 6.0 | 128 | 0.00 |
| 25 | | Isophorone | 40.000 | 37.635 | 5.9 | 130 | 0.00 |
| 26 | C | 2-Nitrophenol | 40.000 | 46.379 | -15.9 | 177 | 0.00 |
| 27 | | 2,4-Dimethylphenol | 40.000 | 38.154 | 4.6 | 132 | 0.00 |
| 28 | | bis(2-Chloroethoxy)methane | 40.000 | 37.033 | 7.4 | 130 | 0.00 |
| 29 | C | 2,4-Dichlorophenol | 40.000 | 40.095 | -0.2 | 134 | 0.01 |
| 30 | | 1,2,4-Trichlorobenzene | 40.000 | 38.076 | 4.8 | 132 | 0.00 |
| 31 | | Naphthalene | 40.000 | 36.365 | 9.1 | 127 | 0.00 |
| 32 | | Benzoic acid | 40.000 | 47.258 | -18.1 | 182 | 0.04 |
| 33 | | 4-Chloroaniline | 40.000 | 37.319 | 6.7 | 126 | 0.00 |
| 34 | C | Hexachlorobutadiene | 40.000 | 40.261 | -0.7 | 141 | 0.00 |
| 35 | | Caprolactam | 40.000 | 39.056 | 2.4 | 133 | 0.02 |
| 36 | C | 4-Chloro-3-methylphenol | 40.000 | 39.033 | 2.4 | 132 | 0.00 |
| 37 | | 2-Methylnaphthalene | 40.000 | 37.184 | 7.0 | 129 | 0.00 |
| 38 | | 1-Methylnaphthalene | 40.000 | 36.882 | 7.8 | 130 | 0.00 |
| 39 | I | Acenaphthene-d10 | 20.000 | 20.000 | 0.0 | 142 | 0.01 |
| 40 | | 1,2,4,5-Tetrachlorobenzene | 40.000 | 38.383 | 4.0 | 135 | 0.00 |
| 41 | P | Hexachlorocyclopentadiene | 40.000 | 44.788 | -12.0 | 161 | -0.01 |
| 42 | S | 2,4,6-Tribromophenol | 80.000 | 92.443 | -15.6 | 158 | 0.00 |
| 43 | C | 2,4,6-Trichlorophenol | 40.000 | 42.344 | -5.9 | 142 | 0.01 |
| 44 | | 2,4,5-Trichlorophenol | 40.000 | 41.215 | -3.0 | 141 | 0.01 |
| 45 | S | 2-Fluorobiphenyl | 80.000 | 72.279 | 9.7 | 130 | 0.00 |
| 46 | | 1,1'-Biphenyl | 40.000 | 36.236 | 9.4 | 129 | 0.00 |
| 47 | | 2-Chloronaphthalene | 40.000 | 36.457 | 8.9 | 130 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025187.D
Acq On : 18 Jul 2025 11:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 18 11:41:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| | | Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|---|---|
| 48 | | 2-Nitroaniline | 40.000 | 38.721 | 3.2 | 141 | 0.00 |
| 49 | | Acenaphthylene | 40.000 | 36.487 | 8.8 | 128 | 0.00 |
| 50 | | Dimethylphthalate | 40.000 | 37.245 | 6.9 | 132 | 0.00 |
| 51 | | 2,6-Dinitrotoluene | 40.000 | 41.967 | -4.9 | 138 | 0.01 |
| 52 | C | Acenaphthene | 40.000 | 35.888 | 10.3 | 129 | 0.00 |
| 53 | | 3-Nitroaniline | 40.000 | 41.772 | -4.4 | 135 | 0.02 |
| 54 | P | 2,4-Dinitrophenol | 40.000 | 52.724 | -31.8# | 206 | 0.02 |
| 55 | | Dibenzofuran | 40.000 | 36.263 | 9.3 | 130 | 0.00 |
| 56 | P | 4-Nitrophenol | 40.000 | 40.881 | -2.2 | 138 | 0.04 |
| 57 | | 2,4-Dinitrotoluene | 40.000 | 39.574 | 1.1 | 142 | 0.01 |
| 58 | | Fluorene | 40.000 | 35.934 | 10.2 | 126 | 0.01 |
| 59 | | 2,3,4,6-Tetrachlorophenol | 40.000 | 41.795 | -4.5 | 143 | 0.01 |
| 60 | | Diethylphthalate | 40.000 | 38.105 | 4.7 | 133 | 0.00 |
| 61 | | 4-Chlorophenyl-phenylether | 40.000 | 36.732 | 8.2 | 131 | 0.00 |
| 62 | | 4-Nitroaniline | 40.000 | 42.792 | -7.0 | 136 | 0.01 |
| 63 | | Azobenzene | 40.000 | 34.529 | 13.7 | 122 | 0.00 |
| 64 | I | Phenanthrene-d10 | 20.000 | 20.000 | 0.0 | 139 | 0.02 |
| 65 | | 4,6-Dinitro-2-methylphenol | 40.000 | 50.040 | -25.1# | 196 | 0.01 |
| 66 | c | n-Nitrosodiphenylamine | 40.000 | 37.671 | 5.8 | 131 | 0.00 |
| 67 | | 4-Bromophenyl-phenylether | 40.000 | 41.198 | -3.0 | 143 | 0.00 |
| 68 | | Hexachlorobenzene | 40.000 | 39.855 | 0.4 | 140 | 0.00 |
| 69 | | Atrazine | 40.000 | 34.885 | 12.8 | 118 | 0.01 |
| 70 | C | Pentachlorophenol | 40.000 | 43.832 | -9.6 | 150 | 0.02 |
| 71 | | Phenanthrene | 40.000 | 36.153 | 9.6 | 126 | 0.01 |
| 72 | | Anthracene | 40.000 | 37.266 | 6.8 | 128 | 0.00 |
| 73 | | Carbazole | 40.000 | 37.621 | 5.9 | 127 | 0.02 |
| 74 | | Di-n-butylphthalate | 40.000 | 42.758 | -6.9 | 140 | 0.00 |
| 75 | C | Fluoranthene | 40.000 | 38.314 | 4.2 | 129 | 0.00 |
| 76 | I | Chrysene-d12 | 20.000 | 20.000 | 0.0 | 137 | 0.00 |
| 77 | | Benzidine | 40.000 | 62.073 | -55.2# | 131 | 0.00 |
| 78 | | Pyrene | 40.000 | 36.776 | 8.1 | 126 | 0.00 |
| 79 | S | Terphenyl-d14 | 80.000 | 75.154 | 6.1 | 128 | 0.00 |
| 80 | | Butylbenzylphthalate | 40.000 | 43.409 | -8.5 | 159 | 0.00 |
| 81 | | Benzo(a)anthracene | 40.000 | 37.242 | 6.9 | 127 | 0.00 |
| 82 | | 3,3'-Dichlorobenzidine | 40.000 | 47.747 | -19.4 | 153 | 0.00 |
| 83 | | Chrysene | 40.000 | 37.025 | 7.4 | 127 | 0.00 |
| 84 | | Bis(2-ethylhexyl)phthalate | 40.000 | 47.504 | -18.8 | 150 | -0.02 |
| 85 | c | Di-n-octyl phthalate | 40.000 | 45.411 | -13.5 | 168 | -0.03 |
| 86 | I | Perylene-d12 | 20.000 | 20.000 | 0.0 | 149 | 0.00 |
| 87 | | Indeno(1,2,3-cd)pyrene | 40.000 | 40.270 | -0.7 | 143 | 0.00 |
| 88 | | Benzo(b)fluoranthene | 40.000 | 36.504 | 8.7 | 132 | 0.00 |
| 89 | | Benzo(k)fluoranthene | 40.000 | 35.769 | 10.6 | 134 | 0.00 |
| 90 | C | Benzo(a)pyrene | 40.000 | 38.067 | 4.8 | 136 | 0.00 |
| 91 | | Dibenzo(a,h)anthracene | 40.000 | 40.071 | -0.2 | 143 | 0.00 |
| 92 | | Benzo(g,h,i)perylene | 40.000 | 39.037 | 2.4 | 140 | 0.01 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025187.D
Acq On : 18 Jul 2025 11:04
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

**Instrument :**
BNA_P
**LabSampleId :**
SSTDCCC040

Quant Time: Jul 18 11:41:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|---|---|---|---|---|---|

(#) = Out of Range SPCC's out = 0 CCC's out = 0

# QC SAMPLE DATA

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143092.D
Acq On : 15 Jul 2025 12:35
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Jul 15 17:53:25 2025




AutoFind: Scans 1624, 1625, 1626; Background Corrected with Scan 1617

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|---|---|---|---|---|---|---|
| 68 | 69 | 0.00 | 2 | 2.0 | 1083 | PASS |
| 69 | 69 | 100 | 100 | 100.0 | 55011 | PASS |
| 70 | 69 | 0.00 | 2 | 0.5 | 283 | PASS |
| 197 | 198 | 0.00 | 2 | 0.3 | 479 | PASS |
| 198 | 198 | 100 | 100 | 100.0 | 145821 | PASS |
| 199 | 198 | 5 | 9 | 6.5 | 9539 | PASS |
| 365 | 198 | 1 | 100 | 4.0 | 5805 | PASS |
| 441 | 443 | 0.01 | 150 | 77.7 | 27936 | PASS |
| 442 | 442 | 100 | 100 | 100.0 | 185813 | PASS |
| 443 | 442 | 15 | 24 | 19.4 | 35957 | PASS |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143092.D
Acq On : 15 Jul 2025 12:35
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
DFTPP

Quant Time: Jul 15 18:02:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration










TIC: BF143092.D\data.ms

**(70) Pentachlorophenol (C)**

**11.292min (+ 0.000) 28198.82 ng**

**response 151956**

| Ion | Exp% | Act% |
|---|---|---|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 64.70 | 61.86 |
| 264.00 | 62.10 | 62.56 |
| 0.00 | 0.00 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071525\
Data File : BF143092.D
Acq On : 15 Jul 2025 12:35
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
DFTPP

Quant Time: Jul 15 18:02:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration










TIC: BF143092.D\data.ms

**(77) Benzidine**

**12.822min (+ 0.006) 0.00 ng**

**response 925010**

| Ion | Exp% | Act% |
|---|---|---|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50 | 14.52 |
| 183.00 | 10.70 | 11.25 |
| 0.00 | 0.00 | 0.00 |

**Instrument :**
BNA_F
**ClientSampleId :**
DFTPP

## DDT Breakdown

| Date | Instrument Name | DFTPP Data File |
|---|---|---|
| 7/15/2025 | BNA_F | BF143092.D |
| | | |
| | | |
| **Compound Name** | **Response** | **Retention Time** |
| | | |
| DDT | 237029 | 13.651 |
| DDD | 2265 | 13.386 |
| DDE | 0 | 13.01 |
| | | |
| | | |
| **SUM(DDD+DDE)** | **SUM(DDT+DDD+DDE)** | **% Breakdown Of DDT** |
| | | |
| 2265 | 239294 | 0.95 |
| | | |

Sample Results: **BF143116.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143116.D
Acq On : 16 Jul 2025 19:09
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Jul 15 17:53:25 2025




AutoFind: Scans 1623, 1624, 1625; Background Corrected with Scan 1617

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|---|---|---|---|---|---|---|
| 68 | 69 | 0.00 | 2 | 1.9 | 1058 | PASS |
| 69 | 69 | 100 | 100 | 100.0 | 56128 | PASS |
| 70 | 69 | 0.00 | 2 | 0.6 | 318 | PASS |
| 197 | 198 | 0.00 | 2 | 0.7 | 1000 | PASS |
| 198 | 198 | 100 | 100 | 100.0 | 144192 | PASS |
| 199 | 198 | 5 | 9 | 6.8 | 9864 | PASS |
| 365 | 198 | 1 | 100 | 3.6 | 5196 | PASS |
| 441 | 443 | 0.01 | 150 | 79.1 | 26908 | PASS |
| 442 | 442 | 100 | 100 | 100.0 | 178251 | PASS |
| 443 | 442 | 15 | 24 | 19.1 | 34024 | PASS |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143116.D
Acq On : 16 Jul 2025 19:09
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
DFTPP

Quant Time: Jul 17 03:11:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration










TIC: BF143116.D\data.ms

**(70) Pentachlorophenol (C)**

**11.292min (-0.000) 32920.43 ng**

**response 195924**

| Ion | Exp% | Act% |
|---|---|---|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 64.70 | 63.41 |
| 264.00 | 62.10 | 62.55 |
| 0.00 | 0.00 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143116.D
Acq On : 16 Jul 2025 19:09
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
DFTPP

Quant Time: Jul 17 03:11:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration










TIC: BF143116.D\data.ms

**(77) Benzidine**

**12.821min (+ 0.006) 0.00 ng**

**response 848698**

| Ion | Exp% | Act% |
|---|---|---|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50 | 15.39 |
| 183.00 | 10.70 | 11.10 |
| 0.00 | 0.00 | 0.00 |

**Instrument :**
BNA_F
**ClientSampleId :**
DFTPP

## DDT Breakdown

| Date | Instrument Name | DFTPP Data File |
|---|---|---|
| 7/16/2025 | BNA_F | BF143116.D |
| | | |
| | | |
| **Compound Name** | **Response** | **Retention Time** |
| | | |
| DDT | 269374 | 13.651 |
| DDD | 2925 | 13.38 |
| DDE | 0 | |
| | | |
| | | |
| **SUM(DDD+DDE)** | **SUM(DDT+DDD+DDE)** | **% Breakdown Of DDT** |
| | | |
| 2925 | 272299 | 1.07 |
| | | |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025070.D
Acq On : 03 Jul 2025 08:58
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jul 03 16:05:44 2025




AutoFind: Scans 2463, 2464, 2465; Background Corrected with Scan 2453

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|---|---|---|---|---|---|---|
| 68 | 69 | 0.00 | 2 | 1.7 | 3747 | PASS |
| 69 | 69 | 100 | 100 | 100.0 | 220616 | PASS |
| 70 | 69 | 0.00 | 2 | 0.4 | 974 | PASS |
| 197 | 198 | 0.00 | 2 | 0.3 | 1975 | PASS |
| 198 | 198 | 100 | 100 | 100.0 | 702809 | PASS |
| 199 | 198 | 5 | 9 | 6.8 | 47648 | PASS |
| 365 | 198 | 1 | 100 | 3.7 | 26203 | PASS |
| 441 | 443 | 0.01 | 150 | 79.7 | 100672 | PASS |
| 442 | 442 | 100 | 100 | 100.0 | 651937 | PASS |
| 443 | 442 | 15 | 24 | 19.4 | 126320 | PASS |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025070.D
Acq On : 03 Jul 2025 08:58
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
DFTPP

Quant Time: Jul 03 16:24:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration







TIC: BP025070.D\data.ms

**(77) Benzidine**

**19.272min (+ 0.006) 5891.13 ng**

**response 5273970**

| Ion | Exp% | Act% |
|---|---|---|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.70 | 16.19 |
| 183.00 | 11.80 | 11.79 |
| 0.00 | 0.00 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP070325\
Data File : BP025070.D
Acq On : 03 Jul 2025 08:58
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
DFTPP

Quant Time: Jul 03 16:24:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration







TIC: BP025070.D\data.ms

**(70) Pentachlorophenol (C)**

**16.554min (+ 0.006) 170301.74 ng**

**response 942789**

| Ion | Exp% | Act% |
|---|---|---|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 63.00 | 63.35 |
| 264.00 | 62.70 | 62.60 |
| 0.00 | 0.00 | 0.00 |

**Instrument :**
BNA_P
**ClientSampleId :**
DFTPP

## DDT Breakdown

| Date | Instrument Name | DFTPP Data File |
|---|---|---|
| 7/3/2025 | BNA_P | BP025070.D |
| | | |
| | | |
| **Compound Name** | **Response** | **Retention Time** |
| | | |
| DDT | 2078463 | 20.595 |
| DDD | 21773 | 20.183 |
| DDE | 1831 | 19.595 |
| | | |
| | | |
| **SUM(DDD+DDE)** | **SUM(DDT+DDD+DDE)** | **% Breakdown Of DDT** |
| | | |
| 23604 | 2102067 | 1.12 |
| | | |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025154.D
Acq On : 16 Jul 2025 11:19
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jul 03 16:05:44 2025




AutoFind: Scans 2461, 2462, 2463; Background Corrected with Scan 2451

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|---|---|---|---|---|---|---|
| 68 | 69 | 0.00 | 2 | 1.7 | 9499 | PASS |
| 69 | 69 | 100 | 100 | 100.0 | 547610 | PASS |
| 70 | 69 | 0.00 | 2 | 0.6 | 3175 | PASS |
| 197 | 198 | 0.00 | 2 | 0.2 | 3324 | PASS |
| 198 | 198 | 100 | 100 | 100.0 | 1941472 | PASS |
| 199 | 198 | 5 | 9 | 6.8 | 132421 | PASS |
| 365 | 198 | 1 | 100 | 3.6 | 69917 | PASS |
| 441 | 443 | 0.01 | 150 | 78.5 | 282261 | PASS |
| 442 | 442 | 100 | 100 | 100.0 | 1818101 | PASS |
| 443 | 442 | 15 | 24 | 19.8 | 359579 | PASS |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025154.D
Acq On : 16 Jul 2025 11:19
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
DFTPP

Quant Time: Jul 16 14:37:58 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration










TIC: BP025154.D\data.ms

**(77) Benzidine**

**19.272min (+ 0.006) 56461.29 ng**

**response 14377417**

| Ion | Exp% | Act% |
|---|---|---|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.70 | 16.02 |
| 183.00 | 11.80 | 13.02 |
| 0.00 | 0.00 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071625\
Data File : BP025154.D
Acq On : 16 Jul 2025 11:19
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
DFTPP

Quant Time: Jul 16 14:37:58 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration










TIC: BP025154.D\data.ms

**(77) Benzidine**

**19.272min (+ 0.006) 56461.29 ng**

**response 14377417**

| Ion | Exp% | Act% |
|---|---|---|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.70 | 16.02 |
| 183.00 | 11.80 | 13.02 |
| 0.00 | 0.00 | 0.00 |

**Instrument :**
BNA_P
**ClientSampleId :**
DFTPP

## DDT Breakdown

| Date | Instrument Name | DFTPP Data File |
|---|---|---|
| 7/16/2025 | BNA_P | BP025154.D |
| | | |
| | | |
| **Compound Name** | **Response** | **Retention Time** |
| | | |
| DDT | 7110729 | 20.578 |
| DDD | 166967 | 20.107 |
| DDE | 9571 | 19.583 |
| | | |
| | | |
| **SUM(DDD+DDE)** | **SUM(DDT+DDD+DDE)** | **% Breakdown Of DDT** |
| | | |
| 176538 | 7287267 | 2.42 |
| | | |

Sample Results: **BP025186.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025186.D
Acq On : 18 Jul 2025 09:41
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-Tune.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Mar 29 08:52:24 2022




AutoFind: Scans 2464, 2465, 2466; Background Corrected with Scan 2456

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|---|---|---|---|---|---|---|
| 68 | 69 | 0.00 | 2 | 1.8 | 10952 | PASS |
| 69 | 69 | 100 | 100 | 100.0 | 624035 | PASS |
| 70 | 69 | 0.00 | 2 | 0.5 | 2838 | PASS |
| 197 | 198 | 0.00 | 2 | 0.0 | 0 | PASS |
| 198 | 198 | 100 | 100 | 100.0 | 2422784 | PASS |
| 199 | 198 | 5 | 9 | 6.8 | 164253 | PASS |
| 365 | 198 | 1 | 100 | 4.2 | 101199 | PASS |
| 441 | 443 | 0.01 | 150 | 80.7 | 406912 | PASS |
| 442 | 442 | 100 | 100 | 100.0 | 2631503 | PASS |
| 443 | 442 | 15 | 24 | 19.2 | 504102 | PASS |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025186.D
Acq On : 18 Jul 2025 09:41
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
DFTPP

Quant Time: Jul 18 11:41:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration










TIC: BP025186.D\data.ms

**(70) Pentachlorophenol (C)**

**16.566min (+ 0.017) 169477.45 ng**

**response 4133923**

| Ion | Exp% | Act% |
|---|---|---|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 63.00 | 64.49 |
| 264.00 | 62.70 | 63.53 |
| 0.00 | 0.00 | 0.00 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025186.D
Acq On : 18 Jul 2025 09:41
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

**Instrument :**
BNA_P
**ClientSampleId :**
DFTPP

Quant Time: Jul 18 11:41:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration










TIC: BP025186.D\data.ms

**(77) Benzidine**

**19.277min (+ 0.011) 75417.62 ng**

**response 15500257**

| Ion | Exp% | Act% |
|---|---|---|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.70 | 18.09 |
| 183.00 | 11.80 | 15.05# |
| 0.00 | 0.00 | 0.00 |

**Instrument :**
BNA_P
**ClientSampleId :**
DFTPP

## DDT Breakdown

| Date | Instrument Name | DFTPP Data File |
|---|---|---|
| 7/18/2025 | BNA_P | BP025186.D |
| | | |
| | | |
| **Compound Name** | **Response** | **Retention Time** |
| | | |
| DDT | 9164203 | 20.589 |
| DDD | 214461 | 20.13 |
| DDE | 11445 | 19.595 |
| | | |
| | | |
| **SUM(DDD+DDE)** | **SUM(DDT+DDD+DDE)** | **% Breakdown Of DDT** |
| | | |
| 225906 | 9390109 | 2.41 |
| | | |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | |
| Project: | Con Edison - East River Site 2 | Date Received: | |
| Client Sample ID: | PB168854BL | SDG No.: | Q2592 |
| Lab Sample ID: | PB168854BL | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 100 |
| Sample Wt/Vol: | 30.03 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BF143118.D | 1 | 07/15/25 09:05 | 07/16/25 20:09 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| **TARGETS** | | | | | | |
| 100-52-7 | Benzaldehyde | 330 | U | 160 | 330 | ug/Kg |
| 108-95-2 | Phenol | 170 | U | 22.1 | 170 | ug/Kg |
| 111-44-4 | bis(2-Chloroethyl)ether | 170 | U | 24.3 | 170 | ug/Kg |
| 95-57-8 | 2-Chlorophenol | 170 | U | 24.4 | 170 | ug/Kg |
| 95-48-7 | 2-Methylphenol | 170 | U | 29.9 | 170 | ug/Kg |
| 108-60-1 | 2,2-oxybis(1-Chloropropane) | 170 | U | 37.5 | 170 | ug/Kg |
| 98-86-2 | Acetophenone | 170 | U | 29.5 | 170 | ug/Kg |
| 65794-96-9 | 3+4-Methylphenols | 330 | U | 41.1 | 330 | ug/Kg |
| 621-64-7 | n-Nitroso-di-n-propylamine | 79.9 | U | 47.4 | 79.9 | ug/Kg |
| 67-72-1 | Hexachloroethane | 170 | U | 17.6 | 170 | ug/Kg |
| 98-95-3 | Nitrobenzene | 170 | U | 18.3 | 170 | ug/Kg |
| 78-59-1 | Isophorone | 170 | U | 32.8 | 170 | ug/Kg |
| 88-75-5 | 2-Nitrophenol | 170 | U | 58.1 | 170 | ug/Kg |
| 105-67-9 | 2,4-Dimethylphenol | 170 | U | 64.7 | 170 | ug/Kg |
| 111-91-1 | bis(2-Chloroethoxy)methane | 170 | U | 30.8 | 170 | ug/Kg |
| 120-83-2 | 2,4-Dichlorophenol | 170 | U | 28.3 | 170 | ug/Kg |
| 91-20-3 | Naphthalene | 170 | U | 22.7 | 170 | ug/Kg |
| 106-47-8 | 4-Chloroaniline | 170 | U | 35.4 | 170 | ug/Kg |
| 87-68-3 | Hexachlorobutadiene | 170 | U | 25.3 | 170 | ug/Kg |
| 105-60-2 | Caprolactam | 330 | U | 52.0 | 330 | ug/Kg |
| 59-50-7 | 4-Chloro-3-methylphenol | 170 | U | 28.7 | 170 | ug/Kg |
| 91-57-6 | 2-Methylnaphthalene | 170 | U | 25.6 | 170 | ug/Kg |
| 77-47-4 | Hexachlorocyclopentadiene | 330 | U | 120 | 330 | ug/Kg |
| 88-06-2 | 2,4,6-Trichlorophenol | 170 | U | 19.8 | 170 | ug/Kg |
| 95-95-4 | 2,4,5-Trichlorophenol | 170 | U | 29.1 | 170 | ug/Kg |
| 92-52-4 | 1,1-Biphenyl | 170 | U | 21.8 | 170 | ug/Kg |
| 91-58-7 | 2-Chloronaphthalene | 170 | U | 22.5 | 170 | ug/Kg |
| 88-74-4 | 2-Nitroaniline | 170 | U | 48.1 | 170 | ug/Kg |
| 131-11-3 | Dimethylphthalate | 170 | U | 27.1 | 170 | ug/Kg |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | |
| Project: | Con Edison - East River Site 2 | Date Received: | |
| Client Sample ID: | PB168854BL | SDG No.: | Q2592 |
| Lab Sample ID: | PB168854BL | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 100 |
| Sample Wt/Vol: | 30.03 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BF143118.D | 1 | 07/15/25 09:05 | 07/16/25 20:09 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| 208-96-8 | Acenaphthylene | 170 | U | 28.9 | 170 | ug/Kg |
| 606-20-2 | 2,6-Dinitrotoluene | 170 | U | 33.6 | 170 | ug/Kg |
| 99-09-2 | 3-Nitroaniline | 170 | U | 46.0 | 170 | ug/Kg |
| 83-32-9 | Acenaphthene | 170 | U | 21.3 | 170 | ug/Kg |
| 51-28-5 | 2,4-Dinitrophenol | 330 | U | 230 | 330 | ug/Kg |
| 100-02-7 | 4-Nitrophenol | 330 | U | 110 | 330 | ug/Kg |
| 132-64-9 | Dibenzofuran | 170 | U | 22.7 | 170 | ug/Kg |
| 121-14-2 | 2,4-Dinitrotoluene | 170 | U | 50.0 | 170 | ug/Kg |
| 84-66-2 | Diethylphthalate | 170 | U | 28.3 | 170 | ug/Kg |
| 7005-72-3 | 4-Chlorophenyl-phenylether | 170 | U | 26.7 | 170 | ug/Kg |
| 86-73-7 | Fluorene | 170 | U | 25.3 | 170 | ug/Kg |
| 100-01-6 | 4-Nitroaniline | 170 | U | 64.1 | 170 | ug/Kg |
| 534-52-1 | 4,6-Dinitro-2-methylphenol | 330 | U | 100 | 330 | ug/Kg |
| 86-30-6 | n-Nitrosodiphenylamine | 170 | U | 32.9 | 170 | ug/Kg |
| 101-55-3 | 4-Bromophenyl-phenylether | 170 | U | 27.8 | 170 | ug/Kg |
| 118-74-1 | Hexachlorobenzene | 170 | U | 25.3 | 170 | ug/Kg |
| 1912-24-9 | Atrazine | 170 | U | 34.0 | 170 | ug/Kg |
| 87-86-5 | Pentachlorophenol | 330 | U | 51.2 | 330 | ug/Kg |
| 85-01-8 | Phenanthrene | 170 | U | 20.9 | 170 | ug/Kg |
| 120-12-7 | Anthracene | 170 | U | 33.3 | 170 | ug/Kg |
| 86-74-8 | Carbazole | 170 | U | 31.2 | 170 | ug/Kg |
| 84-74-2 | Di-n-butylphthalate | 170 | U | 47.9 | 170 | ug/Kg |
| 206-44-0 | Fluoranthene | 170 | U | 30.0 | 170 | ug/Kg |
| 129-00-0 | Pyrene | 170 | U | 36.0 | 170 | ug/Kg |
| 85-68-7 | Butylbenzylphthalate | 170 | U | 71.3 | 170 | ug/Kg |
| 91-94-1 | 3,3-Dichlorobenzidine | 330 | U | 36.7 | 330 | ug/Kg |
| 56-55-3 | Benzo(a)anthracene | 170 | U | 23.0 | 170 | ug/Kg |
| 218-01-9 | Chrysene | 170 | U | 19.9 | 170 | ug/Kg |
| 117-81-7 | Bis(2-ethylhexyl)phthalate | 170 | U | 59.1 | 170 | ug/Kg |
| 117-84-0 | Di-n-octyl phthalate | 330 | U | 86.7 | 330 | ug/Kg |
| 205-99-2 | Benzo(b)fluoranthene | 170 | U | 19.0 | 170 | ug/Kg |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | |
| Project: | Con Edison - East River Site 2 | Date Received: | |
| Client Sample ID: | PB168854BL | SDG No.: | Q2592 |
| Lab Sample ID: | PB168854BL | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 100 |
| Sample Wt/Vol: | 30.03 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BF143118.D | 1 | 07/15/25 09:05 | 07/16/25 20:09 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| 207-08-9 | Benzo(k)fluoranthene | 170 | U | 22.4 | 170 | ug/Kg |
| 50-32-8 | Benzo(a)pyrene | 170 | U | 29.5 | 170 | ug/Kg |
| 193-39-5 | Indeno(1,2,3-cd)pyrene | 170 | U | 29.1 | 170 | ug/Kg |
| 53-70-3 | Dibenzo(a,h)anthracene | 170 | U | 27.4 | 170 | ug/Kg |
| 191-24-2 | Benzo(g,h,i)perylene | 170 | U | 25.7 | 170 | ug/Kg |
| 95-94-3 | 1,2,4,5-Tetrachlorobenzene | 170 | U | 25.6 | 170 | ug/Kg |
| 123-91-1 | 1,4-Dioxane | 170 | U | 45.2 | 170 | ug/Kg |
| 58-90-2 | 2,3,4,6-Tetrachlorophenol | 170 | U | 27.4 | 170 | ug/Kg |
| **SURROGATES** | | | | | | |
| 367-12-4 | 2-Fluorophenol | 110 | | 18 - 112 | 73% | SPK: 150 |
| 13127-88-3 | Phenol-d6 | 110 | | 15 - 107 | 73% | SPK: 150 |
| 4165-60-0 | Nitrobenzene-d5 | 82.0 | | 18 - 107 | 82% | SPK: 100 |
| 321-60-8 | 2-Fluorobiphenyl | 70.8 | | 20 - 109 | 71% | SPK: 100 |
| 118-79-6 | 2,4,6-Tribromophenol | 131 | | 10 - 116 | 87% | SPK: 150 |
| 1718-51-0 | Terphenyl-d14 | 77.2 | | 10 - 105 | 77% | SPK: 100 |
| **INTERNAL STANDARDS** | | | | | | |
| 3855-82-1 | 1,4-Dichlorobenzene-d4 | 139000 | 6.963 | | | |
| 1146-65-2 | Naphthalene-d8 | 529000 | 8.245 | | | |
| 15067-26-2 | Acenaphthene-d10 | 273000 | 9.998 | | | |
| 1517-22-2 | Phenanthrene-d10 | 467000 | 11.486 | | | |
| 1719-03-5 | Chrysene-d12 | 263000 | 14.127 | | | |
| 1520-96-3 | Perylene-d12 | 225000 | 15.633 | | | |
| **TENTATIVE IDENTIFIED COMPOUNDS** | | | | | | |
| 000123-42-2 | 2-Pentanone, 4-hydroxy-4-methyl- | 290 | A | | 5.22 | ug/Kg |
| 006311-48-4 | (1,1-Biphenyl)-4,4-diamine, N,N | 180 | J | | 17.5 | ug/Kg |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | | | |
|---|---|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | | | Date Collected: | |
| Project: | Con Edison - East River Site 2 | | | Date Received: | |
| Client Sample ID: | PB168854BL | | | SDG No.: | Q2592 |
| Lab Sample ID: | PB168854BL | | | Matrix: | SOIL |
| Analytical Method: | 8270E | | | % Solid: | 100 |
| Sample Wt/Vol: | 30.03 | Units: | g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | | | uL | Test: | SVOCMS Group1 |
| Extraction Type : | | | Decanted : N | Level : | LOW |
| Injection Volume : | | | GPC Factor : 1.0 | GPC Cleanup : N | PH : |
| Prep Method : | SW3541 | | | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BF143118.D | 1 | 07/15/25 09:05 | 07/16/25 20:09 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143118.D
Acq On : 16 Jul 2025 20:09
Operator : RC/JU
Sample : PB168854BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Quant Time: Jul 17 03:44:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 6.963 | 152 | 138731 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 8.245 | 136 | 528830 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 9.998 | 164 | 273109 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 11.486 | 188 | 466923 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 14.127 | 240 | 262735 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 15.633 | 264 | 225267 | 20.000 | ng | 0.00 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.598 | 112 | 961902 | 109.546 | ng | 0.01 |
| 7) Phenol-d6 | 6.581 | 99 | 1210302 | 109.654 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 7.522 | 82 | 772988 | 81.998 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 10.786 | 330 | 319517 | 131.209 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 9.322 | 172 | 1454255 | 70.782 | ng | 0.00 |
| 79) Terphenyl-d14 | 13.074 | 244 | 1387334 | 77.189 | ng | 0.00 |
| Target Compounds | | | | | | Qvalue |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143118.D
Acq On : 16 Jul 2025 20:09
Operator : RC/JU
Sample : PB168854BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Quant Time: Jul 17 03:44:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.963 min Scan# 8
Delta R.T. -0.006 min
Lab File: BF143118.D
Acq: 16 Jul 2025 20:09

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Tgt Ion:152 Resp: 138731

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 156.1 | 126.5 | 189.7 |
| 115 | 61.2 | 47.0 | 70.6 |







#5
2-Fluorophenol
Concen: 109.546 ng
RT: 5.598 min Scan# 579
Delta R.T. 0.012 min
Lab File: BF143118.D
Acq: 16 Jul 2025 20:09

Tgt Ion:112 Resp: 961902

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 67.2 | 54.6 | 82.0 |
| 63 | 33.7 | 27.3 | 40.9 |






#7
Phenol-d6
Concen: 109.654 ng
RT: 6.581 min Scan# 7
Delta R.T. -0.006 min
Lab File: BF143118.D
Acq: 16 Jul 2025 20:09

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Tgt Ion: 99 Resp: 1210302

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 21.4 | 17.2 | 25.8 |
| 71 | 35.0 | 27.4 | 41.0 |







#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.245 min Scan# 1029
Delta R.T. -0.006 min
Lab File: BF143118.D
Acq: 16 Jul 2025 20:09

Tgt Ion:136 Resp: 528830

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.9 | 8.9 | 13.3 |
| 54 | 10.1 | 8.3 | 12.5 |
| 68 | 6.8 | 5.3 | 7.9 |






#23
Nitrobenzene-d5
Concen: 81.998 ng
RT: 7.522 min Scan# 9
Delta R.T. -0.006 min
Lab File: BF143118.D
Acq: 16 Jul 2025 20:09

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Tgt Ion: 82 Resp: 772988

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 42.7 | 33.3 | 49.9 |
| 54 | 58.3 | 47.4 | 71.2 |







#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.998 min Scan# 1327
Delta R.T. -0.006 min
Lab File: BF143118.D
Acq: 16 Jul 2025 20:09

Tgt Ion:164 Resp: 273109

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 97.9 | 79.0 | 118.6 |
| 160 | 44.2 | 35.8 | 53.6 |






#42
2,4,6-Tribromophenol
Concen: 131.209 ng
RT: 10.786 min Scan# 14
Delta R.T. -0.006 min
Lab File: BF143118.D
Acq: 16 Jul 2025 20:09

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Tgt Ion:330 Resp: 319517

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 96.4 | 77.9 | 116.9 |
| 141 | 28.9 | 25.9 | 38.9 |







#45
2-Fluorobiphenyl
Concen: 70.782 ng
RT: 9.322 min Scan# 1212
Delta R.T. -0.006 min
Lab File: BF143118.D
Acq: 16 Jul 2025 20:09

Tgt Ion:172 Resp: 1454255

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.1 | 28.2 | 42.4 |
| 170 | 23.0 | 18.6 | 28.0 |






#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.486 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143118.D
Acq: 16 Jul 2025 20:09

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Tgt Ion:188 Resp: 466923

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 10.7 | 8.6 | 13.0 |
| 80 | 12.2 | 9.3 | 13.9 |







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.127 min Scan# 2029
Delta R.T. -0.000 min
Lab File: BF143118.D
Acq: 16 Jul 2025 20:09

Tgt Ion:240 Resp: 262735

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.3 | 10.0 | 15.0 |
| 236 | 25.2 | 20.0 | 30.0 |






#79
Terphenyl-d14
Concen: 77.189 ng
RT: 13.074 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143118.D
Acq: 16 Jul 2025 20:09

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Tgt Ion:244 Resp: 1387334

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.4 | 5.2 | 7.8 |
| 122 | 10.7 | 8.8 | 13.2 |







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.633 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BF143118.D
Acq: 16 Jul 2025 20:09

Tgt Ion:264 Resp: 225267

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 22.6 | 18.5 | 27.7 |
| 265 | 20.6 | 17.5 | 26.3 |



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143118.D
Acq On : 16 Jul 2025 20:09
Operator : RC/JU
Sample : PB168854BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Integration Parameters: rteint.p
Integrator: RTE
Smoothing : ON Filtering: 5
Sampling : 1 Min Area: 3 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC: BF143118.D\data.ms

| peak # | R.T. min | first scan | max scan | last scan | PK TY | peak height | corr. area | corr. % max. | % of total |
|---|---|---|---|---|---|---|---|---|---|
| 1 | 2.387 | 23 | 33 | 40 | rVB | 29515 | 54351 | 1.29% | 0.209% |
| 2 | 5.216 | 507 | 514 | 524 | rBV | 236824 | 371332 | 8.81% | 1.429% |
| 3 | 5.598 | 572 | 579 | 585 | rBV | 2565912 | 3568173 | 84.61% | 13.734% |
| 4 | 6.581 | 739 | 746 | 751 | rBV | 2493071 | 3601109 | 85.39% | 13.861% |
| 5 | 6.963 | 806 | 811 | 817 | rVB | 660830 | 839838 | 19.91% | 3.233% |
| 6 | 7.522 | 899 | 906 | 911 | rBV | 1662751 | 2354628 | 55.83% | 9.063% |
| 7 | 8.245 | 1023 | 1029 | 1034 | rBV | 856616 | 1095678 | 25.98% | 4.217% |
| 8 | 9.322 | 1205 | 1212 | 1217 | rBV | 3164376 | 4217254 | 100.00% | 16.233% |
| 9 | 9.998 | 1322 | 1327 | 1333 | rBV | 904180 | 1174324 | 27.85% | 4.520% |
| 10 | 10.786 | 1455 | 1461 | 1480 | rBV | 1835835 | 2385067 | 56.55% | 9.181% |
| 11 | 11.486 | 1575 | 1580 | 1585 | rBV | 960962 | 1171864 | 27.79% | 4.511% |
| 12 | 13.074 | 1844 | 1850 | 1855 | rBV | 2864902 | 3682936 | 87.33% | 14.176% |
| 13 | 14.121 | 2023 | 2028 | 2041 | rBV | 538076 | 708316 | 16.80% | 2.726% |
| 14 | 15.633 | 2279 | 2285 | 2297 | rBV | 381747 | 591901 | 14.04% | 2.278% |
| 15 | 17.539 | 2599 | 2609 | 2636 | rVB3 | 33806 | 162866 | 3.86% | 0.627% |

Sum of corrected areas: 25979637

Sample Results: **BF143118.D**

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143118.D
Acq On : 16 Jul 2025 20:09
Operator : RC/JU
Sample : PB168854BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P









Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143118.D
Acq On : 16 Jul 2025 20:09
Operator : RC/JU
Sample : PB168854BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

**********
Peak Number 1 2-Pentanone, 4-hydroxy-4-me... Concentration Rank 1

| R.T. | EstConc | Area | Relative to ISTD | R.T. |
|---|---|---|---|---|
| 5.216 | 8.84 ng | 371332 | 1,4-Dichlorobenzene-d4 | 6.963 |

| Hit# of 5 | Tentative ID | MW | MolForm | CAS# | Qual |
|---|---|---|---|---|---|
| 1 | 2-Pentanone, 4-hydroxy-4-methyl- | 116 | C6H12O2 | 000123-42-2 | 83 |
| 2 | Acetone | 58 | C3H6O | 000067-64-1 | 16 |
| 3 | 2,3-Butanedione, monooxime | 101 | C4H7NO2 | 000057-71-6 | 12 |
| 4 | 1-Propen-2-ol, acetate | 100 | C5H8O2 | 000108-22-5 | 12 |
| 5 | 4-Penten-2-one, 4-methyl- | 98 | C6H10O | 003744-02-3 | 9 |






Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143118.D
Acq On : 16 Jul 2025 20:09
Operator : RC/JU
Sample : PB168854BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

**********
Peak Number 2 (1,1'-Biphenyl)-4,4'-diamin... Concentration Rank 2

| R.T. | EstConc | Area | Relative to ISTD | R.T. |
|---|---|---|---|---|
| 17.539 | 5.50 ng | 162866 | Perylene-d12 | 15.633 |

| Hit# of 5 | Tentative ID | MW | MolForm | CAS# | Qual |
|---|---|---|---|---|---|
| 1 | (1,1'-Biphenyl)-4,4'-diamine, N,... | 360 | C26H20N2 | 006311-48-4 | 90 |
| 2 | 1,10-Phenanthroline, 2,9-dimethy... | 360 | C26H20N2 | 004733-39-5 | 47 |
| 3 | Benzaldehyde, 4-hydroxy-3,5-dime... | 360 | C18H20N2O6 | 014414-32-5 | 47 |
| 4 | N,N'-di-2-Naphthyl-p-phenylenedi... | 360 | C26H20N2 | 000093-46-9 | 46 |
| 5 | Ethene, 1-(anthracen-9-yl)-2-(4-... | 360 | C27H20O | 1000163-55-1 | 43 |






Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143118.D
Acq On : 16 Jul 2025 20:09
Operator : RC/JU
Sample : PB168854BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BL

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

| | | | | | --Internal Standard--- | | | |
|---|---|---|---|---|---|---|---|---|
| TIC Top Hit name | RT | EstConc | Units | Response | # | RT | Resp | Conc |
| 2-Pentanone, 4-... | 5.216 | 8.8 | ng | 371332 | 1 | 6.963 | 839838 | 20.0 |
| (1,1'-Biphenyl)... | 17.539 | 5.5 | ng | 162866 | 6 | 15.633 | 591901 | 20.0 |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | |
| Project: | Con Edison - East River Site 2 | Date Received: | |
| Client Sample ID: | PB168854BS | SDG No.: | Q2592 |
| Lab Sample ID: | PB168854BS | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 100 |
| Sample Wt/Vol: | 30.01 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BF143119.D | 1 | 07/15/25 09:05 | 07/16/25 20:39 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| **TARGETS** | | | | | | |
| 100-52-7 | Benzaldehyde | 1200 | | 160 | 330 | ug/Kg |
| 108-95-2 | Phenol | 1500 | | 22.1 | 170 | ug/Kg |
| 111-44-4 | bis(2-Chloroethyl)ether | 1400 | | 24.3 | 170 | ug/Kg |
| 95-57-8 | 2-Chlorophenol | 1500 | | 24.4 | 170 | ug/Kg |
| 95-48-7 | 2-Methylphenol | 1500 | | 29.9 | 170 | ug/Kg |
| 108-60-1 | 2,2-oxybis(1-Chloropropane) | 1400 | | 37.5 | 170 | ug/Kg |
| 98-86-2 | Acetophenone | 1400 | | 29.5 | 170 | ug/Kg |
| 65794-96-9 | 3+4-Methylphenols | 1500 | | 41.1 | 330 | ug/Kg |
| 621-64-7 | n-Nitroso-di-n-propylamine | 1500 | | 47.4 | 80.0 | ug/Kg |
| 67-72-1 | Hexachloroethane | 1500 | | 17.6 | 170 | ug/Kg |
| 98-95-3 | Nitrobenzene | 1600 | | 18.3 | 170 | ug/Kg |
| 78-59-1 | Isophorone | 1500 | | 32.8 | 170 | ug/Kg |
| 88-75-5 | 2-Nitrophenol | 1600 | | 58.2 | 170 | ug/Kg |
| 105-67-9 | 2,4-Dimethylphenol | 1500 | | 64.8 | 170 | ug/Kg |
| 111-91-1 | bis(2-Chloroethoxy)methane | 1500 | | 30.8 | 170 | ug/Kg |
| 120-83-2 | 2,4-Dichlorophenol | 1500 | | 28.3 | 170 | ug/Kg |
| 91-20-3 | Naphthalene | 1400 | | 22.7 | 170 | ug/Kg |
| 106-47-8 | 4-Chloroaniline | 620 | | 35.4 | 170 | ug/Kg |
| 87-68-3 | Hexachlorobutadiene | 1500 | | 25.3 | 170 | ug/Kg |
| 105-60-2 | Caprolactam | 1700 | | 52.1 | 330 | ug/Kg |
| 59-50-7 | 4-Chloro-3-methylphenol | 1500 | | 28.7 | 170 | ug/Kg |
| 91-57-6 | 2-Methylnaphthalene | 1500 | | 25.6 | 170 | ug/Kg |
| 77-47-4 | Hexachlorocyclopentadiene | 3100 | E | 120 | 330 | ug/Kg |
| 88-06-2 | 2,4,6-Trichlorophenol | 1600 | | 19.8 | 170 | ug/Kg |
| 95-95-4 | 2,4,5-Trichlorophenol | 1600 | | 29.1 | 170 | ug/Kg |
| 92-52-4 | 1,1-Biphenyl | 1400 | | 21.8 | 170 | ug/Kg |
| 91-58-7 | 2-Chloronaphthalene | 1400 | | 22.5 | 170 | ug/Kg |
| 88-74-4 | 2-Nitroaniline | 1600 | | 48.1 | 170 | ug/Kg |
| 131-11-3 | Dimethylphthalate | 1500 | | 27.1 | 170 | ug/Kg |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | |
| Project: | Con Edison - East River Site 2 | Date Received: | |
| Client Sample ID: | PB168854BS | SDG No.: | Q2592 |
| Lab Sample ID: | PB168854BS | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 100 |
| Sample Wt/Vol: | 30.01 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BF143119.D | 1 | 07/15/25 09:05 | 07/16/25 20:39 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| 208-96-8 | Acenaphthylene | 1500 | | 28.9 | 170 | ug/Kg |
| 606-20-2 | 2,6-Dinitrotoluene | 1600 | | 33.6 | 170 | ug/Kg |
| 99-09-2 | 3-Nitroaniline | 1100 | | 46.0 | 170 | ug/Kg |
| 83-32-9 | Acenaphthene | 1500 | | 21.3 | 170 | ug/Kg |
| 51-28-5 | 2,4-Dinitrophenol | 3200 | E | 230 | 330 | ug/Kg |
| 100-02-7 | 4-Nitrophenol | 3400 | E | 110 | 330 | ug/Kg |
| 132-64-9 | Dibenzofuran | 1500 | | 22.7 | 170 | ug/Kg |
| 121-14-2 | 2,4-Dinitrotoluene | 1700 | | 50.1 | 170 | ug/Kg |
| 84-66-2 | Diethylphthalate | 1600 | | 28.3 | 170 | ug/Kg |
| 7005-72-3 | 4-Chlorophenyl-phenylether | 1500 | | 26.7 | 170 | ug/Kg |
| 86-73-7 | Fluorene | 1400 | | 25.3 | 170 | ug/Kg |
| 100-01-6 | 4-Nitroaniline | 1800 | | 64.2 | 170 | ug/Kg |
| 534-52-1 | 4,6-Dinitro-2-methylphenol | 1700 | | 100 | 330 | ug/Kg |
| 86-30-6 | n-Nitrosodiphenylamine | 1500 | | 32.9 | 170 | ug/Kg |
| 101-55-3 | 4-Bromophenyl-phenylether | 1500 | | 27.8 | 170 | ug/Kg |
| 118-74-1 | Hexachlorobenzene | 1500 | | 25.3 | 170 | ug/Kg |
| 1912-24-9 | Atrazine | 1700 | | 34.0 | 170 | ug/Kg |
| 87-86-5 | Pentachlorophenol | 3300 | E | 51.3 | 330 | ug/Kg |
| 85-01-8 | Phenanthrene | 1400 | | 20.9 | 170 | ug/Kg |
| 120-12-7 | Anthracene | 1500 | | 33.3 | 170 | ug/Kg |
| 86-74-8 | Carbazole | 1500 | | 31.2 | 170 | ug/Kg |
| 84-74-2 | Di-n-butylphthalate | 1700 | | 47.9 | 170 | ug/Kg |
| 206-44-0 | Fluoranthene | 1500 | | 30.0 | 170 | ug/Kg |
| 129-00-0 | Pyrene | 1500 | | 36.0 | 170 | ug/Kg |
| 85-68-7 | Butylbenzylphthalate | 1700 | | 71.4 | 170 | ug/Kg |
| 91-94-1 | 3,3-Dichlorobenzidine | 780 | | 36.7 | 330 | ug/Kg |
| 56-55-3 | Benzo(a)anthracene | 1600 | | 23.0 | 170 | ug/Kg |
| 218-01-9 | Chrysene | 1500 | | 19.9 | 170 | ug/Kg |
| 117-81-7 | Bis(2-ethylhexyl)phthalate | 1600 | | 59.2 | 170 | ug/Kg |
| 117-84-0 | Di-n-octyl phthalate | 1500 | | 86.8 | 330 | ug/Kg |
| 205-99-2 | Benzo(b)fluoranthene | 1500 | | 19.0 | 170 | ug/Kg |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | |
| Project: | Con Edison - East River Site 2 | Date Received: | |
| Client Sample ID: | PB168854BS | SDG No.: | Q2592 |
| Lab Sample ID: | PB168854BS | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 100 |
| Sample Wt/Vol: | 30.01 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BF143119.D | 1 | 07/15/25 09:05 | 07/16/25 20:39 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| 207-08-9 | Benzo(k)fluoranthene | 1700 | | 22.4 | 170 | ug/Kg |
| 50-32-8 | Benzo(a)pyrene | 1600 | | 29.5 | 170 | ug/Kg |
| 193-39-5 | Indeno(1,2,3-cd)pyrene | 1500 | | 29.1 | 170 | ug/Kg |
| 53-70-3 | Dibenzo(a,h)anthracene | 1500 | | 27.4 | 170 | ug/Kg |
| 191-24-2 | Benzo(g,h,i)perylene | 1500 | | 25.7 | 170 | ug/Kg |
| 95-94-3 | 1,2,4,5-Tetrachlorobenzene | 1400 | | 25.6 | 170 | ug/Kg |
| 123-91-1 | 1,4-Dioxane | 1200 | | 45.2 | 170 | ug/Kg |
| 58-90-2 | 2,3,4,6-Tetrachlorophenol | 1700 | | 27.4 | 170 | ug/Kg |
| **SURROGATES** | | | | | | |
| 367-12-4 | 2-Fluorophenol | 112 | | 18 - 112 | 75% | SPK: 150 |
| 13127-88-3 | Phenol-d6 | 114 | | 15 - 107 | 76% | SPK: 150 |
| 4165-60-0 | Nitrobenzene-d5 | 86.0 | | 18 - 107 | 86% | SPK: 100 |
| 321-60-8 | 2-Fluorobiphenyl | 72.4 | | 20 - 109 | 72% | SPK: 100 |
| 118-79-6 | 2,4,6-Tribromophenol | 140 | | 10 - 116 | 94% | SPK: 150 |
| 1718-51-0 | Terphenyl-d14 | 79.7 | | 10 - 105 | 80% | SPK: 100 |
| **INTERNAL STANDARDS** | | | | | | |
| 3855-82-1 | 1,4-Dichlorobenzene-d4 | 140000 | 6.969 | | | |
| 1146-65-2 | Naphthalene-d8 | 539000 | 8.251 | | | |
| 15067-26-2 | Acenaphthene-d10 | 273000 | 10.004 | | | |
| 1517-22-2 | Phenanthrene-d10 | 456000 | 11.492 | | | |
| 1719-03-5 | Chrysene-d12 | 241000 | 14.127 | | | |
| 1520-96-3 | Perylene-d12 | 263000 | 15.633 | | | |

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143119.D
Acq On : 16 Jul 2025 20:39
Operator : RC/JU
Sample : PB168854BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Quant Time: Jul 17 03:44:58 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| Internal Standards | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 6.969 | 152 | 139596 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 8.251 | 136 | 538805 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 10.004 | 164 | 272708 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 11.492 | 188 | 455589 | 20.000 | ng | 0.00 |
| 76) Chrysene-d12 | 14.127 | 240 | 240950 | 20.000 | ng | 0.00 |
| 86) Perylene-d12 | 15.633 | 264 | 262866 | 20.000 | ng | 0.00 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.604 | 112 | 989830 | 112.028 | ng | 0.02 |
| 7) Phenol-d6 | 6.586 | 99 | 1262403 | 113.666 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 7.528 | 82 | 826120 | 86.012 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 10.792 | 330 | 341088 | 140.273 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 9.322 | 172 | 1485452 | 72.406 | ng | 0.00 |
| 79) Terphenyl-d14 | 13.074 | 244 | 1313530 | 79.690 | ng | 0.00 |
| Target Compounds | | | | | | Qvalue |
| 2) 1,4-Dioxane | 2.916 | 88 | 156082 | 35.487 | ng | 99 |
| 3) Pyridine | 3.657 | 79 | 426573 | 38.301 | ng | 100 |
| 4) n-Nitrosodimethylamine | 3.604 | 42 | 242710 | 42.061 | ng | 98 |
| 6) Aniline | 6.628 | 93 | 524410 | 33.531 | ng | 99 |
| 8) 2-Chlorophenol | 6.751 | 128 | 399769 | 45.033 | ng | 99 |
| 9) Benzaldehyde | 6.516 | 77 | 293118 | 34.636 | ng | 98 |
| 10) Phenol | 6.604 | 94 | 540727 | 44.973 | ng | 98 |
| 11) bis(2-Chloroethyl)ether | 6.698 | 93 | 397600 | 42.717 | ng | 99 |
| 12) 1,3-Dichlorobenzene | 6.910 | 146 | 417974 | 41.677 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 6.986 | 146 | 427253 | 42.346 | ng | 100 |
| 14) 1,2-Dichlorobenzene | 7.139 | 146 | 406381 | 42.236 | ng | 99 |
| 15) Benzyl Alcohol | 7.098 | 79 | 372222 | 45.231 | ng | 98 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.239 | 45 | 724325 | 42.216 | ng | 99 |
| 17) 2-Methylphenol | 7.210 | 107 | 337095 | 44.003 | ng | 99 |
| 18) Hexachloroethane | 7.486 | 117 | 145952 | 44.435 | ng | 97 |
| 19) n-Nitroso-di-n-propyla... | 7.375 | 70 | 298994 | 43.942 | ng | 99 |
| 20) 3+4-Methylphenols | 7.363 | 107 | 415512 | 44.228 | ng | 96 |
| 22) Acetophenone | 7.369 | 105 | 550129 | 42.355 | ng | 94 |
| 24) Nitrobenzene | 7.545 | 77 | 432713 | 46.668 | ng | 97 |
| 25) Isophorone | 7.786 | 82 | 828422 | 44.624 | ng | 99 |
| 26) 2-Nitrophenol | 7.863 | 139 | 182190 | 49.075 | ng | 97 |
| 27) 2,4-Dimethylphenol | 7.892 | 122 | 382286 | 44.367 | ng | 100 |
| 28) bis(2-Chloroethoxy)met... | 7.992 | 93 | 493796 | 43.864 | ng | 99 |
| 29) 2,4-Dichlorophenol | 8.098 | 162 | 326258 | 46.284 | ng | 100 |
| 30) 1,2,4-Trichlorobenzene | 8.186 | 180 | 346239 | 43.280 | ng | 99 |
| 31) Naphthalene | 8.269 | 128 | 1150939 | 43.016 | ng | 100 |
| 32) Benzoic acid | 8.004 | 122 | 202294 | 47.238 | ng | 98 |
| 33) 4-Chloroaniline | 8.310 | 127 | 199781 | 18.568 | ng | 98 |
| 34) Hexachlorobutadiene | 8.392 | 225 | 210203 | 43.936 | ng | 99 |
| 35) Caprolactam | 8.675 | 113 | 111579m | 50.690 | ng | |
| 36) 4-Chloro-3-methylphenol | 8.786 | 107 | 361316 | 46.015 | ng | 98 |
| 37) 2-Methylnaphthalene | 8.963 | 142 | 698633 | 43.769 | ng | 100 |
| 38) 1-Methylnaphthalene | 9.063 | 142 | 718711 | 43.413 | ng | 100 |
| 40) 1,2,4,5-Tetrachloroben... | 9.127 | 216 | 333993 | 42.026 | ng | 98 |
| 41) Hexachlorocyclopentadiene | 9.122 | 237 | 411090 | 93.298 | ng | 99 |
| 43) 2,4,6-Trichlorophenol | 9.233 | 196 | 232335 | 47.049 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143119.D
Acq On : 16 Jul 2025 20:39
Operator : RC/JU
Sample : PB168854BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Quant Time: Jul 17 03:44:58 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 9.275 | 196 | 247228 | 47.747 | ng | | 98 |
| 46) 1,1'-Biphenyl | 9.427 | 154 | 936182 | 43.047 | ng | | 100 |
| 47) 2-Chloronaphthalene | 9.451 | 162 | 680991 | 42.624 | ng | | 99 |
| 48) 2-Nitroaniline | 9.539 | 65 | 225544 | 47.330 | ng | | 96 |
| 49) Acenaphthylene | 9.869 | 152 | 1177621 | 44.234 | ng | | 100 |
| 50) Dimethylphthalate | 9.722 | 163 | 795480 | 46.366 | ng | | 100 |
| 51) 2,6-Dinitrotoluene | 9.780 | 165 | 167878 | 49.431 | ng | | 89 |
| 52) Acenaphthene | 10.039 | 154 | 728594 | 46.377 | ng | | 98 |
| 53) 3-Nitroaniline | 9.945 | 138 | 122425 | 33.024 | ng | | 98 |
| 54) 2,4-Dinitrophenol | 10.051 | 184 | 138899 | 95.178 | ng | # | 1 |
| 55) Dibenzofuran | 10.210 | 168 | 1021074 | 43.605 | ng | | 99 |
| 56) 4-Nitrophenol | 10.098 | 139 | 304758 | 103.438 | ng | | 98 |
| 57) 2,4-Dinitrotoluene | 10.186 | 165 | 220603 | 51.817 | ng | | 96 |
| 58) Fluorene | 10.557 | 166 | 761445 | 43.372 | ng | | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 10.327 | 232 | 205668 | 49.983 | ng | | 99 |
| 60) Diethylphthalate | 10.422 | 149 | 783892 | 47.359 | ng | | 99 |
| 61) 4-Chlorophenyl-phenyle... | 10.545 | 204 | 373382 | 44.374 | ng | | 98 |
| 62) 4-Nitroaniline | 10.563 | 138 | 172008 | 53.355 | ng | | 99 |
| 63) Azobenzene | 10.704 | 77 | 828090 | 45.237 | ng | | 99 |
| 65) 4,6-Dinitro-2-methylph... | 10.592 | 198 | 89451 | 51.397 | ng | | 94 |
| 66) n-Nitrosodiphenylamine | 10.663 | 169 | 701985 | 43.865 | ng | | 99 |
| 67) 4-Bromophenyl-phenylether | 11.033 | 248 | 231595 | 44.700 | ng | | 99 |
| 68) Hexachlorobenzene | 11.104 | 284 | 246543 | 44.848 | ng | | 98 |
| 69) Atrazine | 11.186 | 200 | 204382 | 49.918 | ng | | 98 |
| 70) Pentachlorophenol | 11.292 | 266 | 297551 | 98.656 | ng | | 99 |
| 71) Phenanthrene | 11.516 | 178 | 1066550 | 43.378 | ng | | 99 |
| 72) Anthracene | 11.569 | 178 | 1092533 | 44.187 | ng | | 100 |
| 73) Carbazole | 11.716 | 167 | 993920 | 44.870 | ng | | 100 |
| 74) Di-n-butylphthalate | 12.051 | 149 | 1060422 | 52.154 | ng | | 100 |
| 75) Fluoranthene | 12.704 | 202 | 1035606 | 46.322 | ng | | 99 |
| 77) Benzidine | 12.816 | 184 | 325967 | 39.709 | ng | | 99 |
| 78) Pyrene | 12.933 | 202 | 1029906 | 46.068 | ng | | 99 |
| 80) Butylbenzylphthalate | 13.545 | 149 | 231230 | 51.234 | ng | | 96 |
| 81) Benzo(a)anthracene | 14.121 | 228 | 751216 | 46.787 | ng | | 100 |
| 82) 3,3'-Dichlorobenzidine | 14.074 | 252 | 109202 | 23.421 | ng | | 99 |
| 83) Chrysene | 14.157 | 228 | 669052 | 45.243 | ng | | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 14.110 | 149 | 316448 | 46.806 | ng | | 100 |
| 85) Di-n-octyl phthalate | 14.727 | 149 | 586744 | 46.090 | ng | | 98 |
| 87) Indeno(1,2,3-cd)pyrene | 17.180 | 276 | 906388 | 46.360 | ng | | 98 |
| 88) Benzo(b)fluoranthene | 15.192 | 252 | 700298 | 46.186 | ng | | 99 |
| 89) Benzo(k)fluoranthene | 15.221 | 252 | 724358 | 50.075 | ng | | 100 |
| 90) Benzo(a)pyrene | 15.574 | 252 | 706283 | 49.251 | ng | | 99 |
| 91) Dibenzo(a,h)anthracene | 17.198 | 278 | 736917 | 46.108 | ng | | 99 |
| 92) Benzo(g,h,i)perylene | 17.645 | 276 | 748506 | 45.936 | ng | | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071625\
Data File : BF143119.D
Acq On : 16 Jul 2025 20:39
Operator : RC/JU
Sample : PB168854BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Quant Time: Jul 17 03:44:58 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 15 17:53:25 2025
Response via : Initial Calibration






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.969 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:152 Resp: 139596

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 160.7 | 126.5 | 189.7 |
| 115 | 58.0 | 47.0 | 70.6 |







#2
1,4-Dioxane
Concen: 35.487 ng
RT: 2.916 min Scan# 123
Delta R.T. 0.106 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion: 88 Resp: 156082

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 87.7 | 70.5 | 105.7 |
| 43 | 33.2 | 27.3 | 40.9 |






#3
Pyridine
Concen: 38.301 ng
RT: 3.657 min Scan# 24
Delta R.T. 0.076 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion: 79 Resp: 426573

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 86.8 | 69.2 | 103.8 |
| 51 | 41.6 | 33.3 | 49.9 |







#4
n-Nitrosodimethylamine
Concen: 42.061 ng
RT: 3.604 min Scan# 240
Delta R.T. 0.082 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion: 42 Resp: 242710

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 118.2 | 92.8 | 139.2 |
| 44 | 7.5 | 5.8 | 8.6 |






#5
2-Fluorophenol
Concen: 112.028 ng
RT: 5.604 min Scan# 5
Delta R.T. 0.018 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:112 Resp: 989830

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 65.8 | 54.6 | 82.0 |
| 63 | 33.2 | 27.3 | 40.9 |







#6
Aniline
Concen: 33.531 ng
RT: 6.628 min Scan# 754
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion: 93 Resp: 524410

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 40.9 | 33.1 | 49.7 |
| 65 | 20.6 | 16.9 | 25.3 |






#7
Phenol-d6
Concen: 113.666 ng
RT: 6.586 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion: 99 Resp: 1262403

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 21.5 | 17.2 | 25.8 |
| 71 | 35.3 | 27.4 | 41.0 |







#8
2-Chlorophenol
Concen: 45.033 ng
RT: 6.751 min Scan# 775
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:128 Resp: 399769

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.3 | 12.8 | 52.8 |
| 64 | 54.9 | 36.1 | 76.1 |






#9
Benzaldehyde
Concen: 34.636 ng
RT: 6.516 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion: 77 Resp: 293118

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 94.4 | 76.3 | 116.3 |
| 106 | 91.6 | 72.8 | 112.8 |







#10
Phenol
Concen: 44.973 ng
RT: 6.604 min Scan# 750
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion: 94 Resp: 540727

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 30.1 | 9.8 | 49.8 |
| 66 | 41.8 | 20.5 | 60.5 |






#11
bis(2-Chloroethyl)ether
Concen: 42.717 ng
RT: 6.698 min Scan# 7
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion: 93 Resp: 397600

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 82.6 | 61.9 | 101.9 |
| 95 | 32.2 | 12.2 | 52.2 |







#12
1,3-Dichlorobenzene
Concen: 41.677 ng
RT: 6.910 min Scan# 802
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:146 Resp: 417974

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.2 | 50.1 | 75.1 |
| 75 | 33.9 | 27.4 | 41.0 |






#13
1,4-Dichlorobenzene
Concen: 42.346 ng
RT: 6.986 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:146 Resp: 427253

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 63.2 | 50.6 | 76.0 |
| 111 | 43.2 | 35.0 | 52.6 |







#14
1,2-Dichlorobenzene
Concen: 42.236 ng
RT: 7.139 min Scan# 841
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:146 Resp: 406381

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 62.9 | 49.8 | 74.6 |
| 111 | 45.2 | 35.4 | 53.0 |






#15
Benzyl Alcohol
Concen: 45.231 ng
RT: 7.098 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion: 79 Resp: 372222

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 74.9 | 61.1 | 91.7 |
| 77 | 61.3 | 49.9 | 74.9 |







#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.216 ng
RT: 7.239 min Scan# 858
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion: 45 Resp: 724325

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 11.4 | 0.0 | 31.1 |
| 79 | 8.8 | 0.0 | 28.4 |






#17
2-Methylphenol
Concen: 44.003 ng
RT: 7.210 min Scan# 8
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:107 Resp: 337095

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 114.3 | 90.9 | 136.3 |
| 77 | 46.6 | 37.5 | 56.3 |
| 79 | 46.9 | 36.9 | 55.3 |







#18
Hexachloroethane
Concen: 44.435 ng
RT: 7.486 min Scan# 900
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:117 Resp: 145952

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 98.3 | 76.7 | 115.1 |
| 201 | 97.3 | 75.0 | 112.4 |






#19
n-Nitroso-di-n-propylamine
Concen: 43.942 ng
RT: 7.375 min Scan# 8
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion: 70 Resp: 298994

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 66.8 | 53.2 | 79.8 |
| 101 | 9.4 | 7.5 | 11.3 |
| 130 | 20.3 | 15.7 | 23.5 |







#20
3+4-Methylphenols
Concen: 44.228 ng
RT: 7.363 min Scan# 879
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:107 Resp: 415512

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 92.1 | 71.3 | 111.3 |
| 77 | 105.5 | 77.2 | 117.2 |
| 79 | 30.8 | 10.7 | 50.7 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.251 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:136 Resp: 538805

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.8 | 8.9 | 13.3 |
| 54 | 9.7 | 8.3 | 12.5 |
| 68 | 6.7 | 5.3 | 7.9 |







#22
Acetophenone
Concen: 42.355 ng
RT: 7.369 min Scan# 880
Delta R.T. -0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:105 Resp: 550129

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 7.3 | 7.2 | 10.8 |
| 51 | 39.8 | 27.0 | 40.4 |
| 120 | 23.0 | 18.2 | 27.2 |






#23
Nitrobenzene-d5
Concen: 86.012 ng
RT: 7.528 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion: 82 Resp: 826120

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 42.6 | 33.3 | 49.9 |
| 54 | 58.5 | 47.4 | 71.2 |







#24
Nitrobenzene
Concen: 46.668 ng
RT: 7.545 min Scan# 910
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion: 77 Resp: 432713

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 45.2 | 34.5 | 51.7 |
| 65 | 14.4 | 11.3 | 16.9 |






#25
Isophorone
Concen: 44.624 ng
RT: 7.786 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion: 82 Resp: 828422

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 6.9 | 5.6 | 8.4 |
| 138 | 17.6 | 14.3 | 21.5 |







#26
2-Nitrophenol
Concen: 49.075 ng
RT: 7.863 min Scan# 964
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:139 Resp: 182190

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 28.9 | 23.4 | 35.0 |
| 65 | 51.6 | 43.4 | 65.2 |






#27
2,4-Dimethylphenol
Concen: 44.367 ng
RT: 7.892 min Scan# 9
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:122 Resp: 382286

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 113.7 | 91.3 | 136.9 |
| 121 | 57.5 | 45.5 | 68.3 |







#28
bis(2-Chloroethoxy)methane
Concen: 43.864 ng
RT: 7.992 min Scan# 986
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion: 93 Resp: 493796

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.3 | 26.2 | 39.4 |
| 123 | 12.6 | 9.6 | 14.4 |






#29
2,4-Dichlorophenol
Concen: 46.284 ng
RT: 8.098 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:162 Resp: 326258

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 63.5 | 43.1 | 83.1 |
| 98 | 39.4 | 19.2 | 59.2 |







#30
1,2,4-Trichlorobenzene
Concen: 43.280 ng
RT: 8.186 min Scan# 1019
Delta R.T. -0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:180 Resp: 346239

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 92.9 | 74.7 | 112.1 |
| 145 | 31.9 | 24.5 | 36.7 |






#31
Naphthalene
Concen: 43.016 ng
RT: 8.269 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:128 Resp: 1150939

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.3 | 9.0 | 13.4 |
| 127 | 13.0 | 10.5 | 15.7 |







#32
Benzoic acid
Concen: 47.238 ng
RT: 8.004 min Scan# 988
Delta R.T. 0.017 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:122 Resp: 202294

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 124.0 | 105.4 | 145.4 |
| 77 | 96.3 | 78.6 | 118.6 |






#33
4-Chloroaniline
Concen: 18.568 ng
RT: 8.310 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:127 Resp: 199781

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 31.8 | 26.2 | 39.4 |
| 65 | 34.6 | 28.6 | 42.8 |
| 92 | 20.5 | 17.1 | 25.7 |







#34
Hexachlorobutadiene
Concen: 43.936 ng
RT: 8.392 min Scan# 1054
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:225 Resp: 210203

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 62.5 | 50.3 | 75.5 |
| 227 | 63.8 | 50.1 | 75.1 |






#35
Caprolactam
Concen: 50.690 ng m
RT: 8.675 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:113 Resp: 111579

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 193.0 | 186.2 | 226.2 |
| 56 | 141.8 | 134.1 | 174.1 |







#36
4-Chloro-3-methylphenol
Concen: 46.015 ng
RT: 8.786 min Scan# 1121
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:107 Resp: 361316

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 24.5 | 20.2 | 30.2 |
| 142 | 76.5 | 62.4 | 93.6 |






#37
2-Methylnaphthalene
Concen: 43.769 ng
RT: 8.963 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:142 Resp: 698633

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 87.9 | 70.2 | 105.2 |







#38
1-Methylnaphthalene
Concen: 43.413 ng
RT: 9.063 min Scan# 1168
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:142 Resp: 718711

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 90.4 | 72.2 | 108.4 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:164 Resp: 272708

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 99.8 | 79.0 | 118.6 |
| 160 | 43.6 | 35.8 | 53.6 |







#40
1,2,4,5-Tetrachlorobenzene
Concen: 42.026 ng
RT: 9.127 min Scan# 1179
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:216 Resp: 333993

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 79.5 | 62.6 | 94.0 |
| 179 | 20.2 | 16.2 | 24.2 |
| 108 | 21.7 | 15.8 | 23.6 |






#41
Hexachlorocyclopentadiene
Concen: 93.298 ng
RT: 9.122 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:237 Resp: 411090

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 62.9 | 41.6 | 81.6 |
| 272 | 13.0 | 0.0 | 32.9 |







#42
2,4,6-Tribromophenol
Concen: 140.273 ng
RT: 10.792 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:330 Resp: 341088

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 96.6 | 77.9 | 116.9 |
| 141 | 29.6 | 25.9 | 38.9 |






#43
2,4,6-Trichlorophenol
Concen: 47.049 ng
RT: 9.233 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:196 Resp: 232335

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.5 | 76.9 | 115.3 |
| 200 | 30.5 | 23.5 | 35.3 |







#44
2,4,5-Trichlorophenol
Concen: 47.747 ng
RT: 9.275 min Scan# 1204
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:196 Resp: 247228

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.8 | 76.4 | 114.6 |
| 97 | 50.4 | 42.4 | 63.6 |
| 132 | 28.8 | 23.9 | 35.9 |






#45
2-Fluorobiphenyl
Concen: 72.406 ng
RT: 9.322 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:172 Resp: 1485452

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.2 | 28.2 | 42.4 |
| 170 | 23.4 | 18.6 | 28.0 |







#46
1,1'-Biphenyl
Concen: 43.047 ng
RT: 9.427 min Scan# 1230
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:154 Resp: 936182

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.2 | 20.1 | 60.1 |
| 76 | 13.5 | 0.0 | 33.7 |






#47
2-Chloronaphthalene
Concen: 42.624 ng
RT: 9.451 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:162 Resp: 680991

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 41.9 | 33.0 | 49.6 |
| 164 | 32.9 | 26.1 | 39.1 |







#48
2-Nitroaniline
Concen: 47.330 ng
RT: 9.539 min Scan# 1249
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion: 65 Resp: 225544

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 65.4 | 51.3 | 76.9 |
| 138 | 102.3 | 77.8 | 116.6 |






#49
Acenaphthylene
Concen: 44.234 ng
RT: 9.869 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:152 Resp: 1177621

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 19.9 | 15.8 | 23.8 |
| 153 | 13.1 | 10.4 | 15.6 |







#50
Dimethylphthalate
Concen: 46.366 ng
RT: 9.722 min Scan# 1280
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:163 Resp: 795480

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.4 | 3.4 | 5.2 |
| 164 | 10.2 | 8.2 | 12.2 |






#51
2,6-Dinitrotoluene
Concen: 49.431 ng
RT: 9.780 min Scan# 1
Delta R.T. 0.005 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:165 Resp: 167878

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 62.9 | 57.4 | 86.0 |
| 89 | 56.1 | 51.8 | 77.8 |







#52
Acenaphthene
Concen: 46.377 ng
RT: 10.039 min Scan# 1334
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:154 Resp: 728594

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 106.1 | 86.7 | 130.1 |
| 152 | 50.7 | 41.4 | 62.2 |






#53
3-Nitroaniline
Concen: 33.024 ng
RT: 9.945 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:138 Resp: 122425

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 12.2 | 10.1 | 15.1 |
| 92 | 117.7 | 96.4 | 144.6 |







#54
2,4-Dinitrophenol
Concen: 95.178 ng
RT: 10.051 min Scan# 1336
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:184 Resp: 138899

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 85.3 | 149.3 | 223.9# |
| 154 | 247.3 | 1048.0 | 1572.0# |






#55
Dibenzofuran
Concen: 43.605 ng
RT: 10.210 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:168 Resp: 1021074

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 37.6 | 29.8 | 44.6 |
| 169 | 13.6 | 10.8 | 16.2 |







#56
4-Nitrophenol
Concen: 103.438 ng
RT: 10.098 min Scan# 1344
Delta R.T. 0.012 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:139 Resp: 304758

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 69.4 | 49.4 | 89.4 |
| 65 | 105.1 | 88.2 | 128.2 |






#57
2,4-Dinitrotoluene
Concen: 51.817 ng
RT: 10.186 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:165 Resp: 220603

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 48.4 | 40.6 | 61.0 |
| 89 | 73.8 | 62.3 | 93.5 |
| 182 | 3.0 | 2.6 | 3.8 |







#58
Fluorene
Concen: 43.372 ng
RT: 10.557 min Scan# 1422
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:166 Resp: 761445

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 91.9 | 73.7 | 110.5 |
| 167 | 13.5 | 10.9 | 16.3 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 49.983 ng
RT: 10.327 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:232 Resp: 205668

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 45.0 | 36.9 | 55.3 |
| 130 | 2.2 | 1.8 | 2.8 |
| 166 | 29.0 | 23.0 | 34.6 |







#60
Diethylphthalate
Concen: 47.359 ng
RT: 10.422 min Scan# 1399
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:149 Resp: 783892

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 21.3 | 17.3 | 25.9 |
| 150 | 12.6 | 9.9 | 14.9 |






#61
4-Chlorophenyl-phenylether
Concen: 44.374 ng
RT: 10.545 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:204 Resp: 373382

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 33.1 | 25.8 | 38.6 |
| 141 | 59.8 | 46.8 | 70.2 |







#62
4-Nitroaniline
Concen: 53.355 ng
RT: 10.563 min Scan# 1423
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:138 Resp: 172008

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 49.9 | 30.4 | 70.4 |
| 108 | 70.3 | 49.5 | 89.5 |






#63
Azobenzene
Concen: 45.237 ng
RT: 10.704 min Scan# 14
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion: 77 Resp: 828090

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 23.8 | 4.0 | 44.0 |
| 105 | 21.1 | 1.4 | 41.4 |
| 51 | 33.5 | 14.3 | 54.3 |







#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.492 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:188 Resp: 455589

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 10.5 | 8.6 | 13.0 |
| 80 | 11.8 | 9.3 | 13.9 |






#65
4,6-Dinitro-2-methylphenol
Concen: 51.397 ng
RT: 10.592 min Scan# 14
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:198 Resp: 89451

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 58.6 | 44.6 | 84.6 |
| 105 | 46.1 | 29.3 | 69.3 |







#66
n-Nitrosodiphenylamine
Concen: 43.865 ng
RT: 10.663 min Scan# 1440
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:169 Resp: 701985

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 64.7 | 53.0 | 79.6 |
| 167 | 35.3 | 28.4 | 42.6 |






#67
4-Bromophenyl-phenylether
Concen: 44.700 ng
RT: 11.033 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:248 Resp: 231595

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 96.9 | 76.7 | 115.1 |
| 141 | 76.4 | 61.7 | 92.5 |







#68
Hexachlorobenzene
Concen: 44.848 ng
RT: 11.104 min Scan# 1515
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:284 Resp: 246543

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 34.6 | 28.9 | 43.3 |
| 249 | 31.5 | 24.6 | 37.0 |






#69
Atrazine
Concen: 49.918 ng
RT: 11.186 min Scan# 1
Delta R.T. 0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:200 Resp: 204382

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 27.9 | 9.1 | 49.1 |
| 215 | 50.2 | 31.0 | 71.0 |







#70
Pentachlorophenol
Concen: 98.656 ng
RT: 11.292 min Scan# 1547
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:266 Resp: 297551

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 63.5 | 51.8 | 77.6 |
| 264 | 63.2 | 49.7 | 74.5 |






#71
Phenanthrene
Concen: 43.378 ng
RT: 11.516 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:178 Resp: 1066550

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.1 | 15.1 | 22.7 |
| 179 | 15.6 | 12.2 | 18.4 |







#72
Anthracene
Concen: 44.187 ng
RT: 11.569 min Scan# 1594
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:178 Resp: 1092533

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 18.5 | 14.7 | 22.1 |
| 179 | 15.4 | 12.4 | 18.6 |






#73
Carbazole
Concen: 44.870 ng
RT: 11.716 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:167 Resp: 993920

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 21.3 | 17.0 | 25.4 |
| 139 | 12.8 | 10.2 | 15.4 |







#74
Di-n-butylphthalate
Concen: 52.154 ng
RT: 12.051 min Scan# 1676
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:149 Resp: 1060422

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 9.0 | 7.3 | 10.9 |
| 104 | 4.7 | 3.8 | 5.8 |






#75
Fluoranthene
Concen: 46.322 ng
RT: 12.704 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:202 Resp: 1035606

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 11.8 | 0.0 | 32.4 |
| 203 | 17.7 | 0.0 | 37.4 |







#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.127 min Scan# 2029
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:240 Resp: 240950

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 12.3 | 10.0 | 15.0 |
| 236 | 24.8 | 20.0 | 30.0 |






#77
Benzidine
Concen: 39.709 ng
RT: 12.816 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:184 Resp: 325967

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 15.9 | 12.4 | 18.6 |
| 183 | 10.8 | 8.6 | 12.8 |







#78
Pyrene
Concen: 46.068 ng
RT: 12.933 min Scan# 1826
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:202 Resp: 1029906

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 19.7 | 15.5 | 23.3 |
| 203 | 17.7 | 14.0 | 21.0 |






#79
Terphenyl-d14
Concen: 79.690 ng
RT: 13.074 min Scan# 1
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:244 Resp: 1313530

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 6.5 | 5.2 | 7.8 |
| 122 | 10.7 | 8.8 | 13.2 |







#80
Butylbenzylphthalate
Concen: 51.234 ng
RT: 13.545 min Scan# 1930
Delta R.T. -0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:149 Resp: 231230

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 74.6 | 57.0 | 85.4 |
| 206 | 20.2 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 46.787 ng
RT: 14.121 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:228 Resp: 751216

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 26.9 | 21.3 | 31.9 |
| 229 | 19.5 | 15.5 | 23.3 |







#82
3,3'-Dichlorobenzidine
Concen: 23.421 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:252 Resp: 109202

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 64.7 | 51.5 | 77.3 |
| 126 | 12.2 | 9.4 | 14.0 |






#83
Chrysene
Concen: 45.243 ng
RT: 14.157 min Scan# 2(
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:228 Resp: 669052

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 29.8 | 23.4 | 35.2 |
| 229 | 19.9 | 15.3 | 22.9 |







#84
Bis(2-ethylhexyl)phthalate
Concen: 46.806 ng
RT: 14.110 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:149 Resp: 316448

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 27.4 | 21.9 | 32.9 |
| 279 | 4.2 | 3.4 | 5.0 |






#85
Di-n-octyl phthalate
Concen: 46.090 ng
RT: 14.727 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:149 Resp: 586744

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.5 | 1.2 | 1.8 |
| 43 | 12.3 | 10.6 | 16.0 |







#86
Perylene-d12
Concen: 20.000 ng
RT: 15.633 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:264 Resp: 262866

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.4 | 18.5 | 27.7 |
| 265 | 21.6 | 17.5 | 26.3 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 46.360 ng
RT: 17.180 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:276 Resp: 906388

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 27.1 | 22.6 | 34.0 |
| 277 | 24.9 | 20.2 | 30.2 |







#88
Benzo(b)fluoranthene
Concen: 46.186 ng
RT: 15.192 min Scan# 2210
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:252 Resp: 700298

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.7 | 17.5 | 26.3 |
| 125 | 10.2 | 8.6 | 13.0 |






#89
Benzo(k)fluoranthene
Concen: 50.075 ng
RT: 15.221 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:252 Resp: 724358

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.8 | 17.4 | 26.2 |
| 125 | 10.2 | 8.5 | 12.7 |







#90
Benzo(a)pyrene
Concen: 49.251 ng
RT: 15.574 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:252 Resp: 706283

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 21.8 | 17.5 | 26.3 |
| 125 | 11.1 | 9.4 | 14.0 |






#91
Dibenzo(a,h)anthracene
Concen: 46.108 ng
RT: 17.198 min Scan# 2
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

**Instrument :**
BNA_F
**ClientSampleId :**
PB168854BS

Tgt Ion:278 Resp: 736917

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 18.2 | 14.6 | 22.0 |
| 279 | 24.0 | 18.8 | 28.2 |







#92
Benzo(g,h,i)perylene
Concen: 45.936 ng
RT: 17.645 min Scan# 2627
Delta R.T. -0.000 min
Lab File: BF143119.D
Acq: 16 Jul 2025 20:39

Tgt Ion:276 Resp: 748506

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.3 | 19.2 | 28.8 |
| 138 | 24.0 | 19.8 | 29.6 |



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | 07/14/25 |
| Project: | Con Edison - East River Site 2 | Date Received: | 07/14/25 |
| Client Sample ID: | CHRT-26899MS | SDG No.: | Q2592 |
| Lab Sample ID: | Q2597-01MS | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 98.6 |
| Sample Wt/Vol: | 50.06 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BP025191.D | 1 | 07/15/25 09:05 | 07/18/25 13:52 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| **TARGETS** | | | | | | |
| 100-52-7 | Benzaldehyde | 880 | | 94.8 | 200 | ug/Kg |
| 108-95-2 | Phenol | 860 | | 13.4 | 100 | ug/Kg |
| 111-44-4 | bis(2-Chloroethyl)ether | 840 | | 14.8 | 100 | ug/Kg |
| 95-57-8 | 2-Chlorophenol | 910 | | 14.8 | 100 | ug/Kg |
| 95-48-7 | 2-Methylphenol | 930 | | 18.2 | 100 | ug/Kg |
| 108-60-1 | 2,2-oxybis(1-Chloropropane) | 800 | | 22.8 | 100 | ug/Kg |
| 98-86-2 | Acetophenone | 800 | | 17.9 | 100 | ug/Kg |
| 65794-96-9 | 3+4-Methylphenols | 900 | | 25.0 | 200 | ug/Kg |
| 621-64-7 | n-Nitroso-di-n-propylamine | 790 | | 28.8 | 48.6 | ug/Kg |
| 67-72-1 | Hexachloroethane | 920 | | 10.7 | 100 | ug/Kg |
| 98-95-3 | Nitrobenzene | 880 | | 11.1 | 100 | ug/Kg |
| 78-59-1 | Isophorone | 900 | | 19.9 | 100 | ug/Kg |
| 88-75-5 | 2-Nitrophenol | 1000 | | 35.4 | 100 | ug/Kg |
| 105-67-9 | 2,4-Dimethylphenol | 1200 | | 39.4 | 100 | ug/Kg |
| 111-91-1 | bis(2-Chloroethoxy)methane | 850 | | 18.7 | 100 | ug/Kg |
| 120-83-2 | 2,4-Dichlorophenol | 950 | | 17.2 | 100 | ug/Kg |
| 91-20-3 | Naphthalene | 840 | | 13.8 | 100 | ug/Kg |
| 106-47-8 | 4-Chloroaniline | 440 | | 21.5 | 100 | ug/Kg |
| 87-68-3 | Hexachlorobutadiene | 980 | | 15.4 | 100 | ug/Kg |
| 105-60-2 | Caprolactam | 860 | | 31.7 | 200 | ug/Kg |
| 59-50-7 | 4-Chloro-3-methylphenol | 880 | | 17.4 | 100 | ug/Kg |
| 91-57-6 | 2-Methylnaphthalene | 770 | | 15.6 | 100 | ug/Kg |
| 77-47-4 | Hexachlorocyclopentadiene | 2500 | E | 70.5 | 200 | ug/Kg |
| 88-06-2 | 2,4,6-Trichlorophenol | 1100 | | 12.0 | 100 | ug/Kg |
| 95-95-4 | 2,4,5-Trichlorophenol | 1000 | | 17.7 | 100 | ug/Kg |
| 92-52-4 | 1,1-Biphenyl | 880 | | 13.2 | 100 | ug/Kg |
| 91-58-7 | 2-Chloronaphthalene | 910 | | 13.7 | 100 | ug/Kg |
| 88-74-4 | 2-Nitroaniline | 940 | | 29.2 | 100 | ug/Kg |
| 131-11-3 | Dimethylphthalate | 850 | | 16.5 | 100 | ug/Kg |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | 07/14/25 |
| Project: | Con Edison - East River Site 2 | Date Received: | 07/14/25 |
| Client Sample ID: | CHRT-26899MS | SDG No.: | Q2592 |
| Lab Sample ID: | Q2597-01MS | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 98.6 |
| Sample Wt/Vol: | 50.06 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BP025191.D | 1 | 07/15/25 09:05 | 07/18/25 13:52 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| 208-96-8 | Acenaphthylene | 950 | | 17.6 | 100 | ug/Kg |
| 606-20-2 | 2,6-Dinitrotoluene | 1000 | | 20.4 | 100 | ug/Kg |
| 99-09-2 | 3-Nitroaniline | 810 | | 28.0 | 100 | ug/Kg |
| 83-32-9 | Acenaphthene | 850 | | 12.9 | 100 | ug/Kg |
| 51-28-5 | 2,4-Dinitrophenol | 1100 | | 140 | 200 | ug/Kg |
| 100-02-7 | 4-Nitrophenol | 1800 | E | 65.0 | 200 | ug/Kg |
| 132-64-9 | Dibenzofuran | 810 | | 13.8 | 100 | ug/Kg |
| 121-14-2 | 2,4-Dinitrotoluene | 900 | | 30.5 | 100 | ug/Kg |
| 84-66-2 | Diethylphthalate | 810 | | 17.2 | 100 | ug/Kg |
| 7005-72-3 | 4-Chlorophenyl-phenylether | 820 | | 16.2 | 100 | ug/Kg |
| 86-73-7 | Fluorene | 790 | | 15.4 | 100 | ug/Kg |
| 100-01-6 | 4-Nitroaniline | 830 | | 39.0 | 100 | ug/Kg |
| 534-52-1 | 4,6-Dinitro-2-methylphenol | 690 | | 62.6 | 200 | ug/Kg |
| 86-30-6 | n-Nitrosodiphenylamine | 970 | | 20.0 | 100 | ug/Kg |
| 101-55-3 | 4-Bromophenyl-phenylether | 1100 | | 16.9 | 100 | ug/Kg |
| 118-74-1 | Hexachlorobenzene | 1000 | | 15.4 | 100 | ug/Kg |
| 1912-24-9 | Atrazine | 1200 | | 20.7 | 100 | ug/Kg |
| 87-86-5 | Pentachlorophenol | 2000 | E | 31.2 | 200 | ug/Kg |
| 85-01-8 | Phenanthrene | 880 | | 12.7 | 100 | ug/Kg |
| 120-12-7 | Anthracene | 920 | | 20.2 | 100 | ug/Kg |
| 86-74-8 | Carbazole | 870 | | 19.0 | 100 | ug/Kg |
| 84-74-2 | Di-n-butylphthalate | 920 | | 29.1 | 100 | ug/Kg |
| 206-44-0 | Fluoranthene | 910 | | 18.2 | 100 | ug/Kg |
| 129-00-0 | Pyrene | 850 | | 21.9 | 100 | ug/Kg |
| 85-68-7 | Butylbenzylphthalate | 930 | | 43.4 | 100 | ug/Kg |
| 91-94-1 | 3,3-Dichlorobenzidine | 560 | | 22.3 | 200 | ug/Kg |
| 56-55-3 | Benzo(a)anthracene | 930 | | 14.0 | 100 | ug/Kg |
| 218-01-9 | Chrysene | 930 | | 12.1 | 100 | ug/Kg |
| 117-81-7 | Bis(2-ethylhexyl)phthalate | 1100 | | 36.0 | 100 | ug/Kg |
| 117-84-0 | Di-n-octyl phthalate | 1000 | | 52.8 | 200 | ug/Kg |
| 205-99-2 | Benzo(b)fluoranthene | 930 | | 11.5 | 100 | ug/Kg |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | 07/14/25 |
| Project: | Con Edison - East River Site 2 | Date Received: | 07/14/25 |
| Client Sample ID: | CHRT-26899MS | SDG No.: | Q2592 |
| Lab Sample ID: | Q2597-01MS | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 98.6 |
| Sample Wt/Vol: | 50.06 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BP025191.D | 1 | 07/15/25 09:05 | 07/18/25 13:52 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| 207-08-9 | Benzo(k)fluoranthene | 890 | | 13.6 | 100 | ug/Kg |
| 50-32-8 | Benzo(a)pyrene | 1000 | | 17.9 | 100 | ug/Kg |
| 193-39-5 | Indeno(1,2,3-cd)pyrene | 1000 | | 17.7 | 100 | ug/Kg |
| 53-70-3 | Dibenzo(a,h)anthracene | 980 | | 16.7 | 100 | ug/Kg |
| 191-24-2 | Benzo(g,h,i)perylene | 990 | | 15.6 | 100 | ug/Kg |
| 95-94-3 | 1,2,4,5-Tetrachlorobenzene | 950 | | 15.6 | 100 | ug/Kg |
| 123-91-1 | 1,4-Dioxane | 680 | | 27.5 | 100 | ug/Kg |
| 58-90-2 | 2,3,4,6-Tetrachlorophenol | 950 | | 16.7 | 100 | ug/Kg |
| **SURROGATES** | | | | | | |
| 367-12-4 | 2-Fluorophenol | 81.7 | | 18 - 112 | 54% | SPK: 150 |
| 13127-88-3 | Phenol-d6 | 79.9 | | 15 - 107 | 53% | SPK: 150 |
| 4165-60-0 | Nitrobenzene-d5 | 60.7 | | 18 - 107 | 61% | SPK: 100 |
| 321-60-8 | 2-Fluorobiphenyl | 62.3 | | 20 - 109 | 62% | SPK: 100 |
| 118-79-6 | 2,4,6-Tribromophenol | 106 | | 10 - 116 | 71% | SPK: 150 |
| 1718-51-0 | Terphenyl-d14 | 63.9 | | 10 - 105 | 64% | SPK: 100 |
| **INTERNAL STANDARDS** | | | | | | |
| 3855-82-1 | 1,4-Dichlorobenzene-d4 | 725000 | 7.437 | | | |
| 1146-65-2 | Naphthalene-d8 | 2870000 | 10.184 | | | |
| 15067-26-2 | Acenaphthene-d10 | 1760000 | 14.101 | | | |
| 1517-22-2 | Phenanthrene-d10 | 3030000 | 16.919 | | | |
| 1719-03-5 | Chrysene-d12 | 3380000 | 21.36 | | | |
| 1520-96-3 | Perylene-d12 | 4110000 | 24.512 | | | |

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025191.D
Acq On : 18 Jul 2025 13:52
Operator : CG/JU
Sample : Q2597-01MS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jul 18 14:42:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| **Internal Standards** | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.437 | 152 | 725237 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.184 | 136 | 2872256 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.101 | 164 | 1756654 | 20.000 | ng | 0.01 |
| 64) Phenanthrene-d10 | 16.919 | 188 | 3030587 | 20.000 | ng | # 0.01 |
| 76) Chrysene-d12 | 21.360 | 240 | 3384853 | 20.000 | ng | 0.02 |
| 86) Perylene-d12 | 24.512 | 264 | 4114552 | 20.000 | ng | 0.02 |
| **System Monitoring Compounds** | | | | | | |
| 5) 2-Fluorophenol | 5.108 | 112 | 3344729 | 81.716 | ng | 0.01 |
| 7) Phenol-d6 | 6.649 | 99 | 4267319 | 79.891 | ng | 0.02 |
| 23) Nitrobenzene-d5 | 8.578 | 82 | 2621093 | 60.689 | ng | 0.01 |
| 42) 2,4,6-Tribromophenol | 15.636 | 330 | 2206748 | 106.339 | ng | 0.02 |
| 45) 2-Fluorobiphenyl | 12.695 | 172 | 7069182 | 62.342 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.695 | 244 | 9838214 | 63.880 | ng | 0.02 |
| **Target Compounds** | | | | | | Qvalue |
| 2) 1,4-Dioxane | 3.119 | 88 | 530428 | 33.326 | ng | 99 |
| 3) Pyridine | 3.490 | 79 | 1465490 | 33.631 | ng | 97 |
| 4) n-Nitrosodimethylamine | 3.402 | 42 | 635981 | 34.470 | ng | 90 |
| 6) Aniline | 6.784 | 93 | 1455394 | 29.163 | ng | 99 |
| 8) 2-Chlorophenol | 7.019 | 128 | 2066190 | 45.117 | ng | 98 |
| 9) Benzaldehyde | 6.602 | 77 | 1208904 | 43.337 | ng | 96 |
| 10) Phenol | 6.672 | 94 | 2305756 | 42.484 | ng | 98 |
| 11) bis(2-Chloroethyl)ether | 6.884 | 93 | 1730557 | 41.431 | ng | 97 |
| 12) 1,3-Dichlorobenzene | 7.331 | 146 | 2214921 | 43.215 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 7.472 | 146 | 2254625 | 43.448 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 7.784 | 146 | 2161644 | 43.303 | ng | 99 |
| 15) Benzyl Alcohol | 7.678 | 79 | 1547309 | 43.402 | ng | 97 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.966 | 45 | 2155727 | 39.692 | ng | 97 |
| 17) 2-Methylphenol | 7.890 | 107 | 1606309 | 45.946 | ng | 99 |
| 18) Hexachloroethane | 8.496 | 117 | 765719 | 45.295 | ng | 97 |
| 19) n-Nitroso-di-n-propyla... | 8.243 | 70 | 1237318 | 39.060 | ng | 97 |
| 20) 3+4-Methylphenols | 8.213 | 107 | 2165124 | 44.214 | ng | 99 |
| 22) Acetophenone | 8.249 | 105 | 2689995 | 39.601 | ng | 98 |
| 24) Nitrobenzene | 8.619 | 77 | 1904110 | 43.284 | ng | 94 |
| 25) Isophorone | 9.137 | 82 | 3573169 | 44.271 | ng | 99 |
| 26) 2-Nitrophenol | 9.319 | 139 | 1104541 | 51.484 | ng | 97 |
| 27) 2,4-Dimethylphenol | 9.396 | 122 | 1843826 | 61.241 | ng | 99 |
| 28) bis(2-Chloroethoxy)met... | 9.619 | 93 | 2301331 | 41.993 | ng | 99 |
| 29) 2,4-Dichlorophenol | 9.849 | 162 | 1956385 | 46.862 | ng | 99 |
| 30) 1,2,4-Trichlorobenzene | 10.060 | 180 | 2098988 | 45.402 | ng | 99 |
| 31) Naphthalene | 10.237 | 128 | 6137965 | 41.501 | ng | 99 |
| 32) Benzoic acid | 9.554 | 122 | 1560877 | 51.391 | ng | 96 |
| 33) 4-Chloroaniline | 10.354 | 127 | 1207820 | 21.883 | ng | 99 |
| 34) Hexachlorobutadiene | 10.531 | 225 | 1280173 | 48.253 | ng | 100 |
| 35) Caprolactam | 11.148 | 113 | 606849m | 42.491 | ng | |
| 36) 4-Chloro-3-methylphenol | 11.501 | 107 | 1889312 | 43.306 | ng | 97 |
| 37) 2-Methylnaphthalene | 11.872 | 142 | 3941856 | 37.774 | ng | 100 |
| 38) 1-Methylnaphthalene | 12.095 | 142 | 4192599 | 41.195 | ng | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 12.248 | 216 | 2373335 | 47.010 | ng | 99 |
| 41) Hexachlorocyclopentadiene | 12.231 | 237 | 1636353 | 124.087 | ng | 99 |
| 43) 2,4,6-Trichlorophenol | 12.501 | 196 | 1604557 | 52.488 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025191.D
Acq On : 18 Jul 2025 13:52
Operator : CG/JU
Sample : Q2597-01MS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jul 18 14:42:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 12.572 | 196 | 1751985 | 50.570 | ng | | 99 |
| 46) 1,1'-Biphenyl | 12.913 | 154 | 5528311 | 43.253 | ng | | 99 |
| 47) 2-Chloronaphthalene | 12.948 | 162 | 4311680 | 44.883 | ng | | 99 |
| 48) 2-Nitroaniline | 13.160 | 65 | 1095633 | 46.642 | ng | | 91 |
| 49) Acenaphthylene | 13.819 | 152 | 6919582 | 47.062 | ng | | 100 |
| 50) Dimethylphthalate | 13.572 | 163 | 5156677 | 41.936 | ng | | 100 |
| 51) 2,6-Dinitrotoluene | 13.672 | 165 | 1192431 | 49.721 | ng | | 93 |
| 52) Acenaphthene | 14.172 | 154 | 3960058 | 41.788 | ng | | 98 |
| 53) 3-Nitroaniline | 14.007 | 138 | 1024478 | 39.861 | ng | | 99 |
| 54) 2,4-Dinitrophenol | 14.219 | 184 | 621018 | 54.804 | ng | | 94 |
| 55) Dibenzofuran | 14.513 | 168 | 6234201 | 39.862 | ng | | 99 |
| 56) 4-Nitrophenol | 14.354 | 139 | 2135992 | 90.057 | ng | | 95 |
| 57) 2,4-Dinitrotoluene | 14.478 | 165 | 1562642 | 44.595 | ng | | 93 |
| 58) Fluorene | 15.178 | 166 | 4739613 | 39.181 | ng | | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 14.754 | 232 | 1476592 | 47.065 | ng | | 95 |
| 60) Diethylphthalate | 14.984 | 149 | 4928178 | 39.903 | ng | | 98 |
| 61) 4-Chlorophenyl-phenyle... | 15.184 | 204 | 2419173 | 40.411 | ng | | 98 |
| 62) 4-Nitroaniline | 15.201 | 138 | 1104551 | 40.986 | ng | | 99 |
| 63) Azobenzene | 15.484 | 77 | 4154994 | 37.780 | ng | | 94 |
| 65) 4,6-Dinitro-2-methylph... | 15.272 | 198 | 448087 | 34.294 | ng | # | 83 |
| 66) n-Nitrosodiphenylamine | 15.407 | 169 | 4164623 | 48.018 | ng | | 99 |
| 67) 4-Bromophenyl-phenylether | 16.107 | 248 | 1683065 | 54.240 | ng | | 99 |
| 68) Hexachlorobenzene | 16.219 | 284 | 1904013 | 50.202 | ng | | 94 |
| 69) Atrazine | 16.419 | 200 | 1538693 | 60.543 | ng | | 92 |
| 70) Pentachlorophenol | 16.578 | 266 | 2571871 | 100.408 | ng | | 98 |
| 71) Phenanthrene | 16.966 | 178 | 7037848 | 43.460 | ng | | 97 |
| 72) Anthracene | 17.083 | 178 | 7064103 | 45.470 | ng | | 98 |
| 73) Carbazole | 17.348 | 167 | 6507850 | 42.843 | ng | | 98 |
| 74) Di-n-butylphthalate | 17.972 | 149 | 7718219 | 45.628 | ng | | 98 |
| 75) Fluoranthene | 19.077 | 202 | 8459720 | 44.692 | ng | | 97 |
| 77) Benzidine | 19.283 | 184 | 2220748 | 37.148 | ng | # | 91 |
| 78) Pyrene | 19.454 | 202 | 8790194 | 41.928 | ng | | 98 |
| 80) Butylbenzylphthalate | 20.442 | 149 | 3698659 | 45.960 | ng | | 98 |
| 81) Benzo(a)anthracene | 21.348 | 228 | 9398328 | 45.817 | ng | | 98 |
| 82) 3,3'-Dichlorobenzidine | 21.277 | 252 | 1712760 | 27.737 | ng | | 99 |
| 83) Chrysene | 21.407 | 228 | 9091703 | 46.057 | ng | | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 21.330 | 149 | 6024913 | 54.758 | ng | # | 97 |
| 85) Di-n-octyl phthalate | 22.542 | 149 | 10219431 | 50.316 | ng | | 97 |
| 87) Indeno(1,2,3-cd)pyrene | 28.053 | 276 | 13319448m | 50.814 | ng | | |
| 88) Benzo(b)fluoranthene | 23.524 | 252 | 11017325 | 46.004 | ng | | 99 |
| 89) Benzo(k)fluoranthene | 23.595 | 252 | 10410755 | 43.690 | ng | | 99 |
| 90) Benzo(a)pyrene | 24.377 | 252 | 10105828 | 49.580 | ng | | 99 |
| 91) Dibenzo(a,h)anthracene | 28.136 | 278 | 10509553 | 48.505 | ng | | 97 |
| 92) Benzo(g,h,i)perylene | 29.159 | 276 | 10921803 | 48.796 | ng | | 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025191.D
Acq On : 18 Jul 2025 13:52
Operator : CG/JU
Sample : Q2597-01MS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Jul 18 14:42:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.437 min Scan# 7
Delta R.T. -0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:152 Resp: 72523

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 156.2 | 126.1 | 189.1 |
| 115 | 56.0 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#2
1,4-Dioxane
Concen: 33.326 ng
RT: 3.119 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 88 Resp: 530428

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 67.7 | 55.0 | 82.4 |
| 43 | 28.3 | 21.9 | 32.9 |






#3
Pyridine
Concen: 33.631 ng
RT: 3.490 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 79 Resp: 146549

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 63.2 | 52.2 | 78.4 |
| 51 | 32.1 | 26.8 | 40.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#4
n-Nitrosodimethylamine
Concen: 34.470 ng
RT: 3.402 min Scan# 79
Delta R.T. -0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 42 Resp: 635981

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 143.3 | 105.2 | 157.8 |
| 44 | 8.8 | 6.6 | 9.8 |






#5
2-Fluorophenol
Concen: 81.716 ng
RT: 5.108 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:112 Resp: 334472

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 49.5 | 42.0 | 63.0 |
| 63 | 25.6 | 22.2 | 33.2 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#6
Aniline
Concen: 29.163 ng
RT: 6.784 min Scan# 654
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 93 Resp: 1455394

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 35.7 | 28.9 | 43.3 |
| 65 | 18.2 | 15.2 | 22.8 |






#7
Phenol-d6
Concen: 79.891 ng
RT: 6.649 min Scan# 6
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 99 Resp: 426731

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 15.8 | 13.6 | 20.4 |
| 71 | 30.6 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#8
2-Chlorophenol
Concen: 45.117 ng
RT: 7.019 min Scan# 694
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:128 Resp: 2066190

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.8 | 12.9 | 52.9 |
| 64 | 37.8 | 20.3 | 60.3 |






#9
Benzaldehyde
Concen: 43.337 ng
RT: 6.602 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 77 Resp: 120890

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 106.7 | 82.4 | 122.4 |
| 106 | 103.0 | 78.3 | 118.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#10
Phenol
Concen: 42.484 ng
RT: 6.672 min Scan# 635
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 94 Resp: 2305756

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 25.2 | 6.9 | 46.9 |
| 66 | 33.5 | 14.6 | 54.6 |






#11
bis(2-Chloroethyl)ether
Concen: 41.431 ng
RT: 6.884 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 93 Resp: 173055

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 68.8 | 51.9 | 91.9 |
| 95 | 32.3 | 12.4 | 52.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#12
1,3-Dichlorobenzene
Concen: 43.215 ng
RT: 7.331 min Scan# 747
Delta R.T. -0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:146 Resp: 2214921

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.1 | 51.9 | 77.9 |
| 75 | 25.9 | 21.8 | 32.6 |






#13
1,4-Dichlorobenzene
Concen: 43.448 ng
RT: 7.472 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:146 Resp: 225462

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.0 | 51.1 | 76.7 |
| 111 | 37.4 | 31.2 | 46.8 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#14
1,2-Dichlorobenzene
Concen: 43.303 ng
RT: 7.784 min Scan# 824
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:146 Resp: 2161644

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.5 | 51.1 | 76.7 |
| 111 | 39.6 | 31.2 | 46.8 |






#15
Benzyl Alcohol
Concen: 43.402 ng
RT: 7.678 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 79 Resp: 154730

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 83.8 | 66.0 | 99.0 |
| 77 | 62.5 | 53.4 | 80.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#16
2,2'-oxybis(1-Chloropropane)
Concen: 39.692 ng
RT: 7.966 min Scan# 855
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 45 Resp: 2155727

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 15.9 | 0.0 | 34.8 |
| 79 | 12.2 | 0.0 | 30.7 |






#17
2-Methylphenol
Concen: 45.946 ng
RT: 7.890 min Scan# 84
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:107 Resp: 160630

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 112.0 | 89.7 | 134.5 |
| 77 | 41.6 | 34.2 | 51.4 |
| 79 | 40.6 | 33.4 | 50.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#18
Hexachloroethane
Concen: 45.295 ng
RT: 8.496 min Scan# 945
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:117 Resp: 765719

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 98.5 | 76.6 | 115.0 |
| 201 | 110.3 | 85.4 | 128.2 |






#19
n-Nitroso-di-n-propylamine
Concen: 39.060 ng
RT: 8.243 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 70 Resp: 123731

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 55.9 | 46.5 | 69.7 |
| 101 | 11.0 | 7.9 | 11.9 |
| 130 | 25.8 | 19.3 | 28.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#20
3+4-Methylphenols
Concen: 44.214 ng
RT: 8.213 min Scan# 897
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:107 Resp: 2165124

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 89.8 | 69.5 | 109.5 |
| 77 | 29.6 | 11.4 | 51.4 |
| 79 | 26.3 | 6.9 | 46.9 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.184 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:136 Resp: 287225

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 11.1 | 8.7 | 13.1 |
| 54 | 6.4 | 5.4 | 8.2 |
| 68 | 5.4 | 4.5 | 6.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#22
Acetophenone
Concen: 39.601 ng
RT: 8.249 min Scan# 903
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:105 Resp: 2689995

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 7.3 | 5.4 | 8.2 |
| 51 | 23.1 | 19.6 | 29.4 |
| 120 | 24.0 | 18.8 | 28.2 |






#23
Nitrobenzene-d5
Concen: 60.689 ng
RT: 8.578 min Scan# 9
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 82 Resp: 262109

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 51.0 | 38.4 | 57.6 |
| 54 | 45.6 | 37.8 | 56.6 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#24
Nitrobenzene
Concen: 43.284 ng
RT: 8.619 min Scan# 966
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 77 Resp: 1904110

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 53.4 | 39.0 | 58.6 |
| 65 | 12.6 | 10.7 | 16.1 |






#25
Isophorone
Concen: 44.271 ng
RT: 9.137 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 82 Resp: 357316

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.0 | 5.8 | 8.6 |
| 138 | 20.4 | 15.7 | 23.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#26
2-Nitrophenol
Concen: 51.484 ng
RT: 9.319 min Scan# 1085
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:139 Resp: 1104541

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 26.2 | 22.3 | 33.5 |
| 65 | 40.6 | 33.8 | 50.8 |






#27
2,4-Dimethylphenol
Concen: 61.241 ng
RT: 9.396 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:122 Resp: 184382

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 109.6 | 88.2 | 132.2 |
| 121 | 58.6 | 46.1 | 69.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#28
bis(2-Chloroethoxy)methane
Concen: 41.993 ng
RT: 9.619 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 93 Resp: 2301331

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.3 | 26.4 | 39.6 |
| 123 | 12.1 | 10.1 | 15.1 |






#29
2,4-Dichlorophenol
Concen: 46.862 ng
RT: 9.849 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:162 Resp: 195638

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 65.0 | 44.1 | 84.1 |
| 98 | 31.9 | 13.1 | 53.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#30
1,2,4-Trichlorobenzene
Concen: 45.402 ng
RT: 10.060 min Scan# 1211
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:180 Resp: 2098988

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 95.3 | 75.1 | 112.7 |
| 145 | 29.4 | 23.3 | 34.9 |






#31
Naphthalene
Concen: 41.501 ng
RT: 10.237 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:128 Resp: 613796

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.3 | 8.9 | 13.3 |
| 127 | 13.2 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#32
Benzoic acid
Concen: 51.391 ng
RT: 9.554 min Scan# 1125
Delta R.T. 0.053 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:122 Resp: 1560877

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 117.0 | 101.0 | 141.0 |
| 77 | 82.7 | 67.6 | 107.6 |






#33
4-Chloroaniline
Concen: 21.883 ng
RT: 10.354 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:127 Resp: 120782

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.3 | 25.4 | 38.2 |
| 65 | 25.5 | 21.0 | 31.6 |
| 92 | 16.7 | 13.4 | 20.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#34
Hexachlorobutadiene
Concen: 48.253 ng
RT: 10.531 min Scan# 1291
Delta R.T. -0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:225 Resp: 1280173

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 62.8 | 50.3 | 75.5 |
| 227 | 64.0 | 51.3 | 76.9 |






#35
Caprolactam
Concen: 42.491 ng m
RT: 11.148 min Scan# 1
Delta R.T. 0.041 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:113 Resp: 60684

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 136.0 | 128.1 | 168.1 |
| 56 | 102.1 | 90.7 | 130.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#36
4-Chloro-3-methylphenol
Concen: 43.306 ng
RT: 11.501 min Scan# 1456
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:107 Resp: 1889312

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 29.4 | 22.4 | 33.6 |
| 142 | 89.4 | 69.6 | 104.4 |






#37
2-Methylnaphthalene
Concen: 37.774 ng
RT: 11.872 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:142 Resp: 394185

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 89.2 | 71.2 | 106.8 |
| 115 | 32.1 | 25.9 | 38.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#38
1-Methylnaphthalene
Concen: 41.195 ng
RT: 12.095 min Scan# 1557
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:142 Resp: 4192599

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 91.6 | 74.1 | 111.1 |
| 116 | 3.4 | 2.7 | 4.1 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.101 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:164 Resp: 175665

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 100.8 | 80.3 | 120.5 |
| 160 | 44.1 | 35.9 | 53.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 47.010 ng
RT: 12.248 min Scan# 1583
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:216 Resp: 2373335

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.3 | 62.4 | 93.6 |
| 179 | 20.5 | 16.5 | 24.7 |
| 108 | 16.3 | 13.7 | 20.5 |






#41
Hexachlorocyclopentadiene
Concen: 124.087 ng
RT: 12.231 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:237 Resp: 163635

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 62.0 | 41.6 | 81.6 |
| 272 | 13.4 | 0.0 | 33.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#42
2,4,6-Tribromophenol
Concen: 106.339 ng
RT: 15.636 min Scan# 2159
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:330 Resp: 2206748

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 96.4 | 77.7 | 116.5 |
| 141 | 27.4 | 26.0 | 39.0 |






#43
2,4,6-Trichlorophenol
Concen: 52.488 ng
RT: 12.501 min Scan# 1
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:196 Resp: 160455

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 93.5 | 76.2 | 114.2 |
| 200 | 30.2 | 24.3 | 36.5 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#44
2,4,5-Trichlorophenol
Concen: 50.570 ng
RT: 12.572 min Scan# 1638
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:196 Resp: 1751985

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.6 | 76.6 | 115.0 |
| 97 | 43.3 | 36.6 | 55.0 |
| 132 | 27.0 | 21.4 | 32.0 |






#45
2-Fluorobiphenyl
Concen: 62.342 ng
RT: 12.695 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:172 Resp: 706918

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.3 | 28.2 | 42.2 |
| 170 | 23.8 | 18.7 | 28.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#46
1,1'-Biphenyl
Concen: 43.253 ng
RT: 12.913 min Scan# 1696
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:154 Resp: 5528311

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.6 | 20.2 | 60.2 |
| 76 | 11.6 | 0.0 | 31.5 |






#47
2-Chloronaphthalene
Concen: 44.883 ng
RT: 12.948 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:162 Resp: 431168

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 37.6 | 29.4 | 44.2 |
| 164 | 33.0 | 26.4 | 39.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#48
2-Nitroaniline
Concen: 46.642 ng
RT: 13.160 min Scan# 1738
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 65 Resp: 1095633

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 77.0 | 58.7 | 88.1 |
| 138 | 137.0 | 99.0 | 148.4 |






#49
Acenaphthylene
Concen: 47.062 ng
RT: 13.819 min Scan# 1
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:152 Resp: 691958

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.2 | 16.1 | 24.1 |
| 153 | 13.3 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#50
Dimethylphthalate
Concen: 41.936 ng
RT: 13.572 min Scan# 1808
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:163 Resp: 5156677

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.3 | 3.4 | 5.0 |
| 164 | 10.2 | 8.0 | 12.0 |






#51
2,6-Dinitrotoluene
Concen: 49.721 ng
RT: 13.672 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:165 Resp: 119243

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 43.2 | 38.5 | 57.7 |
| 89 | 46.0 | 40.6 | 61.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#52
Acenaphthene
Concen: 41.788 ng
RT: 14.172 min Scan# 1910
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:154 Resp: 3960058

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 112.1 | 90.6 | 136.0 |
| 152 | 53.0 | 43.8 | 65.8 |






#53
3-Nitroaniline
Concen: 39.861 ng
RT: 14.007 min Scan# 1
Delta R.T. 0.024 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:138 Resp: 102447

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 10.6 | 7.8 | 11.8 |
| 92 | 99.9 | 81.0 | 121.4 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#54
2,4-Dinitrophenol
Concen: 54.804 ng
RT: 14.219 min Scan# 1918
Delta R.T. 0.024 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:184 Resp: 621018

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 51.7 | 47.4 | 71.0 |
| 154 | 57.9 | 47.8 | 71.8 |






#55
Dibenzofuran
Concen: 39.862 ng
RT: 14.513 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:168 Resp: 623420

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 36.3 | 28.6 | 43.0 |
| 169 | 13.8 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#56
4-Nitrophenol
Concen: 90.057 ng
RT: 14.354 min Scan# 1941
Delta R.T. 0.053 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:139 Resp: 2135992

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 68.5 | 45.9 | 85.9 |
| 65 | 75.0 | 61.4 | 101.4 |






#57
2,4-Dinitrotoluene
Concen: 44.595 ng
RT: 14.478 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:165 Resp: 156264

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 34.2 | 31.0 | 46.4 |
| 89 | 60.7 | 53.4 | 80.0 |
| 182 | 2.9 | 2.0 | 3.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#58
Fluorene
Concen: 39.181 ng
RT: 15.178 min Scan# 2081
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:166 Resp: 4739613

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 98.3 | 78.6 | 117.8 |
| 167 | 13.8 | 10.7 | 16.1 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 47.065 ng
RT: 14.754 min Scan# 2
Delta R.T. 0.024 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:232 Resp: 147659

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 40.8 | 35.7 | 53.5 |
| 130 | 2.5 | 1.8 | 2.6 |
| 166 | 28.3 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#60
Diethylphthalate
Concen: 39.903 ng
RT: 14.984 min Scan# 2048
Delta R.T. 0.024 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:149 Resp: 4928178

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 23.7 | 17.8 | 26.8 |
| 150 | 13.0 | 10.2 | 15.2 |






#61
4-Chlorophenyl-phenylether
Concen: 40.411 ng
RT: 15.184 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:204 Resp: 241917

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 33.8 | 25.8 | 38.6 |
| 141 | 55.0 | 45.1 | 67.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#62
4-Nitroaniline
Concen: 40.986 ng
RT: 15.201 min Scan# 2085
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:138 Resp: 1104551

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 45.3 | 26.0 | 66.0 |
| 108 | 69.1 | 50.0 | 90.0 |






#63
Azobenzene
Concen: 37.780 ng
RT: 15.484 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion: 77 Resp: 415499

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 31.9 | 7.7 | 47.7 |
| 105 | 25.9 | 1.9 | 41.9 |
| 51 | 26.4 | 8.2 | 48.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.919 min Scan# 2377
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:188 Resp: 3030587

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 11.9 | 7.3 | 10.9# |
| 80 | 11.8 | 7.8 | 11.6# |






#65
4,6-Dinitro-2-methylphenol
Concen: 34.294 ng
RT: 15.272 min Scan# 2
Delta R.T. 0.024 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:198 Resp: 44808

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 37.1 | 16.8 | 56.8 |
| 105 | 61.0 | 20.8 | 60.8 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#66
n-Nitrosodiphenylamine
Concen: 48.018 ng
RT: 15.407 min Scan# 2120
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:169 Resp: 4164623

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 66.5 | 54.1 | 81.1 |
| 167 | 36.5 | 29.0 | 43.6 |






#67
4-Bromophenyl-phenylether
Concen: 54.240 ng
RT: 16.107 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:248 Resp: 168306

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 98.5 | 79.4 | 119.2 |
| 141 | 64.1 | 51.4 | 77.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#68
Hexachlorobenzene
Concen: 50.202 ng
RT: 16.219 min Scan# 2258
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:284 Resp: 1904013

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 27.4 | 25.9 | 38.9 |
| 249 | 28.6 | 23.8 | 35.8 |






#69
Atrazine
Concen: 60.543 ng
RT: 16.419 min Scan# 2
Delta R.T. 0.035 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:200 Resp: 153869

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 31.6 | 6.6 | 46.6 |
| 215 | 52.5 | 28.0 | 68.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#70
Pentachlorophenol
Concen: 100.408 ng
RT: 16.578 min Scan# 2319
Delta R.T. 0.029 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:266 Resp: 2571871

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 64.5 | 50.4 | 75.6 |
| 264 | 64.2 | 50.2 | 75.2 |






#71
Phenanthrene
Concen: 43.460 ng
RT: 16.966 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:178 Resp: 703784

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.8 | 15.4 | 23.2 |
| 179 | 17.8 | 12.6 | 19.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#72
Anthracene
Concen: 45.470 ng
RT: 17.083 min Scan# 2405
Delta R.T. 0.035 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:178 Resp: 7064103

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.1 | 15.0 | 22.6 |
| 179 | 17.3 | 12.5 | 18.7 |






#73
Carbazole
Concen: 42.843 ng
RT: 17.348 min Scan# 24
Delta R.T. 0.024 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:167 Resp: 650785

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 22.2 | 17.4 | 26.0 |
| 139 | 14.4 | 10.4 | 15.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#74
Di-n-butylphthalate
Concen: 45.628 ng
RT: 17.972 min Scan# 2556
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:149 Resp: 7718219

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 10.2 | 7.3 | 10.9 |
| 104 | 4.8 | 3.8 | 5.6 |






#75
Fluoranthene
Concen: 44.692 ng
RT: 19.077 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:202 Resp: 845972

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 9.5 | 0.0 | 30.4 |
| 203 | 18.8 | 0.0 | 37.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.360 min Scan# 3132
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:240 Resp: 3384853

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.5 | 9.0 | 13.4 |
| 236 | 26.1 | 20.3 | 30.5 |






#77
Benzidine
Concen: 37.148 ng
RT: 19.283 min Scan# 2
Delta R.T. 0.017 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:184 Resp: 222074

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 19.2 | 12.6 | 18.8 |
| 183 | 16.0 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#78
Pyrene
Concen: 41.928 ng
RT: 19.454 min Scan# 2808
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:202 Resp: 8790194

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 21.8 | 17.1 | 25.7 |
| 203 | 18.8 | 14.1 | 21.1 |






#79
Terphenyl-d14
Concen: 63.880 ng
RT: 19.695 min Scan# 2
Delta R.T. 0.024 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:244 Resp: 983821

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 7.2 | 5.9 | 8.9 |
| 122 | 10.4 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#80
Butylbenzylphthalate
Concen: 45.960 ng
RT: 20.442 min Scan# 2976
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:149 Resp: 3698659

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 62.2 | 50.6 | 76.0 |
| 206 | 23.1 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 45.817 ng
RT: 21.348 min Scan# 3
Delta R.T. 0.024 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:228 Resp: 939832

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 28.4 | 22.1 | 33.1 |
| 229 | 20.7 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#82
3,3'-Dichlorobenzidine
Concen: 27.737 ng
RT: 21.277 min Scan# 3118
Delta R.T. 0.006 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:252 Resp: 1712760

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 64.5 | 51.9 | 77.9 |
| 126 | 11.6 | 8.1 | 12.1 |






#83
Chrysene
Concen: 46.057 ng
RT: 21.407 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:228 Resp: 909170

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 30.9 | 24.4 | 36.6 |
| 229 | 20.6 | 16.0 | 24.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#84
Bis(2-ethylhexyl)phthalate
Concen: 54.758 ng
RT: 21.330 min Scan# 3127
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:149 Resp: 6024913

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 29.1 | 21.9 | 32.9 |
| 279 | 5.6 | 3.5 | 5.3# |






#85
Di-n-octyl phthalate
Concen: 50.316 ng
RT: 22.542 min Scan# 3
Delta R.T. -0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:149 Resp:1021943

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.6 | 1.3 | 1.9 |
| 43 | 7.3 | 6.9 | 10.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#86
Perylene-d12
Concen: 20.000 ng
RT: 24.512 min Scan# 3668
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:264 Resp: 4114552

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.5 | 18.6 | 28.0 |
| 265 | 21.8 | 17.7 | 26.5 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 50.814 ng m
RT: 28.053 min Scan# 4
Delta R.T. 0.018 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:276 Resp:1331944

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 16.2 | 13.0 | 19.6 |
| 277 | 20.5 | 15.9 | 23.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#88
Benzo(b)fluoranthene
Concen: 46.004 ng
RT: 23.524 min Scan# 3500
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:252 Resp:11017325

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.5 | 17.7 | 26.5 |
| 125 | 10.0 | 8.1 | 12.1 |






#89
Benzo(k)fluoranthene
Concen: 43.690 ng
RT: 23.595 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:252 Resp:1041075

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.5 | 17.7 | 26.5 |
| 125 | 10.3 | 8.2 | 12.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#90
Benzo(a)pyrene
Concen: 49.580 ng
RT: 24.377 min Scan# 3645
Delta R.T. 0.024 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:252 Resp:10105828

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.3 | 17.6 | 26.4 |
| 125 | 10.4 | 9.0 | 13.4 |






#91
Dibenzo(a,h)anthracene
Concen: 48.505 ng
RT: 28.136 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:278 Resp:1050955

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 14.4 | 13.1 | 19.7 |
| 279 | 24.4 | 18.9 | 28.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MS

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#92
Benzo(g,h,i)perylene
Concen: 48.796 ng
RT: 29.159 min Scan# 4458
Delta R.T. 0.035 min
Lab File: BP025191.D
Acq: 18 Jul 2025 13:52

Tgt Ion:276 Resp:10921803

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 24.1 | 19.3 | 28.9 |
| 138 | 20.4 | 16.9 | 25.3 |



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | 07/14/25 |
| Project: | Con Edison - East River Site 2 | Date Received: | 07/14/25 |
| Client Sample ID: | CHRT-26899MSD | SDG No.: | Q2592 |
| Lab Sample ID: | Q2597-01MSD | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 98.6 |
| Sample Wt/Vol: | 50.08 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BP025192.D | 1 | 07/15/25 09:05 | 07/18/25 14:33 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| **TARGETS** | | | | | | |
| 100-52-7 | Benzaldehyde | 870 | | 94.8 | 200 | ug/Kg |
| 108-95-2 | Phenol | 870 | | 13.4 | 100 | ug/Kg |
| 111-44-4 | bis(2-Chloroethyl)ether | 840 | | 14.8 | 100 | ug/Kg |
| 95-57-8 | 2-Chlorophenol | 930 | | 14.8 | 100 | ug/Kg |
| 95-48-7 | 2-Methylphenol | 930 | | 18.2 | 100 | ug/Kg |
| 108-60-1 | 2,2-oxybis(1-Chloropropane) | 800 | | 22.8 | 100 | ug/Kg |
| 98-86-2 | Acetophenone | 800 | | 17.9 | 100 | ug/Kg |
| 65794-96-9 | 3+4-Methylphenols | 900 | | 25.0 | 200 | ug/Kg |
| 621-64-7 | n-Nitroso-di-n-propylamine | 790 | | 28.8 | 48.6 | ug/Kg |
| 67-72-1 | Hexachloroethane | 920 | | 10.7 | 100 | ug/Kg |
| 98-95-3 | Nitrobenzene | 860 | | 11.1 | 100 | ug/Kg |
| 78-59-1 | Isophorone | 890 | | 19.9 | 100 | ug/Kg |
| 88-75-5 | 2-Nitrophenol | 1000 | | 35.4 | 100 | ug/Kg |
| 105-67-9 | 2,4-Dimethylphenol | 1200 | | 39.4 | 100 | ug/Kg |
| 111-91-1 | bis(2-Chloroethoxy)methane | 840 | | 18.7 | 100 | ug/Kg |
| 120-83-2 | 2,4-Dichlorophenol | 960 | | 17.2 | 100 | ug/Kg |
| 91-20-3 | Naphthalene | 840 | | 13.8 | 100 | ug/Kg |
| 106-47-8 | 4-Chloroaniline | 460 | | 21.5 | 100 | ug/Kg |
| 87-68-3 | Hexachlorobutadiene | 980 | | 15.4 | 100 | ug/Kg |
| 105-60-2 | Caprolactam | 850 | | 31.7 | 200 | ug/Kg |
| 59-50-7 | 4-Chloro-3-methylphenol | 880 | | 17.4 | 100 | ug/Kg |
| 91-57-6 | 2-Methylnaphthalene | 760 | | 15.6 | 100 | ug/Kg |
| 77-47-4 | Hexachlorocyclopentadiene | 2300 | E | 70.5 | 200 | ug/Kg |
| 88-06-2 | 2,4,6-Trichlorophenol | 1100 | | 12.0 | 100 | ug/Kg |
| 95-95-4 | 2,4,5-Trichlorophenol | 1000 | | 17.7 | 100 | ug/Kg |
| 92-52-4 | 1,1-Biphenyl | 860 | | 13.2 | 100 | ug/Kg |
| 91-58-7 | 2-Chloronaphthalene | 900 | | 13.7 | 100 | ug/Kg |
| 88-74-4 | 2-Nitroaniline | 940 | | 29.2 | 100 | ug/Kg |
| 131-11-3 | Dimethylphthalate | 830 | | 16.5 | 100 | ug/Kg |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | 07/14/25 |
| Project: | Con Edison - East River Site 2 | Date Received: | 07/14/25 |
| Client Sample ID: | CHRT-26899MSD | SDG No.: | Q2592 |
| Lab Sample ID: | Q2597-01MSD | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 98.6 |
| Sample Wt/Vol: | 50.08 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BP025192.D | 1 | 07/15/25 09:05 | 07/18/25 14:33 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| 208-96-8 | Acenaphthylene | 940 | | 17.6 | 100 | ug/Kg |
| 606-20-2 | 2,6-Dinitrotoluene | 1000 | | 20.4 | 100 | ug/Kg |
| 99-09-2 | 3-Nitroaniline | 800 | | 27.9 | 100 | ug/Kg |
| 83-32-9 | Acenaphthene | 820 | | 12.9 | 100 | ug/Kg |
| 51-28-5 | 2,4-Dinitrophenol | 1100 | | 140 | 200 | ug/Kg |
| 100-02-7 | 4-Nitrophenol | 1800 | E | 65.0 | 200 | ug/Kg |
| 132-64-9 | Dibenzofuran | 790 | | 13.8 | 100 | ug/Kg |
| 121-14-2 | 2,4-Dinitrotoluene | 900 | | 30.4 | 100 | ug/Kg |
| 84-66-2 | Diethylphthalate | 800 | | 17.2 | 100 | ug/Kg |
| 7005-72-3 | 4-Chlorophenyl-phenylether | 810 | | 16.2 | 100 | ug/Kg |
| 86-73-7 | Fluorene | 790 | | 15.4 | 100 | ug/Kg |
| 100-01-6 | 4-Nitroaniline | 820 | | 39.0 | 100 | ug/Kg |
| 534-52-1 | 4,6-Dinitro-2-methylphenol | 670 | | 62.6 | 200 | ug/Kg |
| 86-30-6 | n-Nitrosodiphenylamine | 950 | | 20.0 | 100 | ug/Kg |
| 101-55-3 | 4-Bromophenyl-phenylether | 1100 | | 16.9 | 100 | ug/Kg |
| 118-74-1 | Hexachlorobenzene | 1000 | | 15.4 | 100 | ug/Kg |
| 1912-24-9 | Atrazine | 1200 | | 20.7 | 100 | ug/Kg |
| 87-86-5 | Pentachlorophenol | 2000 | E | 31.2 | 200 | ug/Kg |
| 85-01-8 | Phenanthrene | 850 | | 12.7 | 100 | ug/Kg |
| 120-12-7 | Anthracene | 910 | | 20.2 | 100 | ug/Kg |
| 86-74-8 | Carbazole | 860 | | 19.0 | 100 | ug/Kg |
| 84-74-2 | Di-n-butylphthalate | 890 | | 29.1 | 100 | ug/Kg |
| 206-44-0 | Fluoranthene | 890 | | 18.2 | 100 | ug/Kg |
| 129-00-0 | Pyrene | 860 | | 21.9 | 100 | ug/Kg |
| 85-68-7 | Butylbenzylphthalate | 920 | | 43.4 | 100 | ug/Kg |
| 91-94-1 | 3,3-Dichlorobenzidine | 600 | | 22.3 | 200 | ug/Kg |
| 56-55-3 | Benzo(a)anthracene | 930 | | 14.0 | 100 | ug/Kg |
| 218-01-9 | Chrysene | 930 | | 12.1 | 100 | ug/Kg |
| 117-81-7 | Bis(2-ethylhexyl)phthalate | 1100 | | 36.0 | 100 | ug/Kg |
| 117-84-0 | Di-n-octyl phthalate | 1000 | | 52.7 | 200 | ug/Kg |
| 205-99-2 | Benzo(b)fluoranthene | 960 | | 11.5 | 100 | ug/Kg |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Report of Analysis

| | | | |
|---|---|---|---|
| Client: | PARSONS Engineering of New York, Inc. | Date Collected: | 07/14/25 |
| Project: | Con Edison - East River Site 2 | Date Received: | 07/14/25 |
| Client Sample ID: | CHRT-26899MSD | SDG No.: | Q2592 |
| Lab Sample ID: | Q2597-01MSD | Matrix: | SOIL |
| Analytical Method: | 8270E | % Solid: | 98.6 |
| Sample Wt/Vol: | 50.08 Units: g | Final Vol: | 1000 uL |
| Soil Aliquot Vol: | uL | Test: | SVOCMS Group1 |
| Extraction Type : | Decanted : N | Level : | LOW |
| Injection Volume : | GPC Factor : 1.0 | GPC Cleanup : | N PH : |
| Prep Method : | SW3541 | | |

| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
|---|---|---|---|---|
| BP025192.D | 1 | 07/15/25 09:05 | 07/18/25 14:33 | PB168854 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|---|---|---|---|---|---|---|
| 207-08-9 | Benzo(k)fluoranthene | 850 | | 13.6 | 100 | ug/Kg |
| 50-32-8 | Benzo(a)pyrene | 1000 | | 17.9 | 100 | ug/Kg |
| 193-39-5 | Indeno(1,2,3-cd)pyrene | 1000 | | 17.7 | 100 | ug/Kg |
| 53-70-3 | Dibenzo(a,h)anthracene | 970 | | 16.6 | 100 | ug/Kg |
| 191-24-2 | Benzo(g,h,i)perylene | 980 | | 15.6 | 100 | ug/Kg |
| 95-94-3 | 1,2,4,5-Tetrachlorobenzene | 940 | | 15.6 | 100 | ug/Kg |
| 123-91-1 | 1,4-Dioxane | 670 | | 27.5 | 100 | ug/Kg |
| 58-90-2 | 2,3,4,6-Tetrachlorophenol | 940 | | 16.6 | 100 | ug/Kg |
| **SURROGATES** | | | | | | |
| 367-12-4 | 2-Fluorophenol | 82.8 | | 18 - 112 | 55% | SPK: 150 |
| 13127-88-3 | Phenol-d6 | 80.9 | | 15 - 107 | 54% | SPK: 150 |
| 4165-60-0 | Nitrobenzene-d5 | 59.8 | | 18 - 107 | 60% | SPK: 100 |
| 321-60-8 | 2-Fluorobiphenyl | 60.6 | | 20 - 109 | 61% | SPK: 100 |
| 118-79-6 | 2,4,6-Tribromophenol | 105 | | 10 - 116 | 70% | SPK: 150 |
| 1718-51-0 | Terphenyl-d14 | 62.7 | | 10 - 105 | 63% | SPK: 100 |
| **INTERNAL STANDARDS** | | | | | | |
| 3855-82-1 | 1,4-Dichlorobenzene-d4 | 718000 | 7.443 | | | |
| 1146-65-2 | Naphthalene-d8 | 2880000 | 10.19 | | | |
| 15067-26-2 | Acenaphthene-d10 | 1760000 | 14.095 | | | |
| 1517-22-2 | Phenanthrene-d10 | 3050000 | 16.919 | | | |
| 1719-03-5 | Chrysene-d12 | 3400000 | 21.36 | | | |
| 1520-96-3 | Perylene-d12 | 4140000 | 24.524 | | | |

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025192.D
Acq On : 18 Jul 2025 14:33
Operator : CG/JU
Sample : Q2597-01MSD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jul 18 14:51:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|---|---|---|---|---|---|---|
| **Internal Standards** | | | | | | |
| 1) 1,4-Dichlorobenzene-d4 | 7.443 | 152 | 718069 | 20.000 | ng | 0.00 |
| 21) Naphthalene-d8 | 10.190 | 136 | 2875357 | 20.000 | ng | 0.00 |
| 39) Acenaphthene-d10 | 14.095 | 164 | 1762354 | 20.000 | ng | 0.00 |
| 64) Phenanthrene-d10 | 16.919 | 188 | 3047795 | 20.000 | ng | # 0.01 |
| 76) Chrysene-d12 | 21.360 | 240 | 3398852 | 20.000 | ng | 0.02 |
| 86) Perylene-d12 | 24.524 | 264 | 4139648 | 20.000 | ng | 0.03 |
| **System Monitoring Compounds** | | | | | | |
| 5) 2-Fluorophenol | 5.108 | 112 | 3354831 | 82.781 | ng | 0.01 |
| 7) Phenol-d6 | 6.643 | 99 | 4278992 | 80.909 | ng | 0.01 |
| 23) Nitrobenzene-d5 | 8.572 | 82 | 2584466 | 59.777 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 15.637 | 330 | 2184054 | 104.905 | ng | 0.02 |
| 45) 2-Fluorobiphenyl | 12.701 | 172 | 6896816 | 60.625 | ng | 0.00 |
| 79) Terphenyl-d14 | 19.689 | 244 | 9692693 | 62.676 | ng | 0.02 |
| **Target Compounds** | | | | | | Qvalue |
| 2) 1,4-Dioxane | 3.120 | 88 | 522161 | 33.134 | ng | 99 |
| 3) Pyridine | 3.490 | 79 | 1448416 | 33.571 | ng | 97 |
| 4) n-Nitrosodimethylamine | 3.402 | 42 | 635044 | 34.763 | ng | 91 |
| 6) Aniline | 6.784 | 93 | 1435687 | 29.055 | ng | 99 |
| 8) 2-Chlorophenol | 7.019 | 128 | 2071338 | 45.681 | ng | 98 |
| 9) Benzaldehyde | 6.596 | 77 | 1191263 | 43.131 | ng | 96 |
| 10) Phenol | 6.672 | 94 | 2309080 | 42.970 | ng | 97 |
| 11) bis(2-Chloroethyl)ether | 6.884 | 93 | 1713531 | 41.433 | ng | 97 |
| 12) 1,3-Dichlorobenzene | 7.331 | 146 | 2203625 | 43.423 | ng | 99 |
| 13) 1,4-Dichlorobenzene | 7.472 | 146 | 2220399 | 43.216 | ng | 99 |
| 14) 1,2-Dichlorobenzene | 7.784 | 146 | 2161225 | 43.727 | ng | 99 |
| 15) Benzyl Alcohol | 7.678 | 79 | 1543453 | 43.726 | ng | 97 |
| 16) 2,2'-oxybis(1-Chloropr... | 7.960 | 45 | 2137331 | 39.746 | ng | 97 |
| 17) 2-Methylphenol | 7.890 | 107 | 1587575 | 45.864 | ng | 99 |
| 18) Hexachloroethane | 8.502 | 117 | 759284 | 45.363 | ng | 97 |
| 19) n-Nitroso-di-n-propyla... | 8.237 | 70 | 1228521 | 39.169 | ng | 97 |
| 20) 3+4-Methylphenols | 8.208 | 107 | 2160042 | 44.551 | ng | 97 |
| 22) Acetophenone | 8.243 | 105 | 2675246 | 39.341 | ng | 98 |
| 24) Nitrobenzene | 8.613 | 77 | 1880097 | 42.692 | ng | 95 |
| 25) Isophorone | 9.143 | 82 | 3565539 | 44.129 | ng | 98 |
| 26) 2-Nitrophenol | 9.313 | 139 | 1092514 | 50.908 | ng | 97 |
| 27) 2,4-Dimethylphenol | 9.390 | 122 | 1846201 | 61.254 | ng | 99 |
| 28) bis(2-Chloroethoxy)met... | 9.619 | 93 | 2268251 | 41.345 | ng | 100 |
| 29) 2,4-Dichlorophenol | 9.849 | 162 | 1978308 | 47.336 | ng | 99 |
| 30) 1,2,4-Trichlorobenzene | 10.055 | 180 | 2089765 | 45.153 | ng | 98 |
| 31) Naphthalene | 10.243 | 128 | 6118039 | 41.322 | ng | 100 |
| 32) Benzoic acid | 9.560 | 122 | 1566557 | 51.506 | ng | 96 |
| 33) 4-Chloroaniline | 10.349 | 127 | 1248930 | 22.604 | ng | 99 |
| 34) Hexachlorobutadiene | 10.531 | 225 | 1279328 | 48.169 | ng | 99 |
| 35) Caprolactam | 11.143 | 113 | 601704m | 42.085 | ng | |
| 36) 4-Chloro-3-methylphenol | 11.496 | 107 | 1897199 | 43.440 | ng | 100 |
| 37) 2-Methylnaphthalene | 11.860 | 142 | 3935923 | 37.677 | ng | 100 |
| 38) 1-Methylnaphthalene | 12.096 | 142 | 4079783 | 40.044 | ng | 99 |
| 40) 1,2,4,5-Tetrachloroben... | 12.243 | 216 | 2361542 | 46.625 | ng | 99 |
| 41) Hexachlorocyclopentadiene | 12.225 | 237 | 1532451 | 115.832 | ng | 98 |
| 43) 2,4,6-Trichlorophenol | 12.496 | 196 | 1617632 | 52.744 | ng | 99 |

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025192.D
Acq On : 18 Jul 2025 14:33
Operator : CG/JU
Sample : Q2597-01MSD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jul 18 14:51:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025

| Compound | R.T. | QIon | Response | Conc | Units | | Dev(Min) |
|---|---|---|---|---|---|---|---|
| 44) 2,4,5-Trichlorophenol | 12.572 | 196 | 1741173 | 50.095 | ng | | 98 |
| 46) 1,1'-Biphenyl | 12.907 | 154 | 5453993 | 42.534 | ng | | 100 |
| 47) 2-Chloronaphthalene | 12.943 | 162 | 4280520 | 44.415 | ng | | 99 |
| 48) 2-Nitroaniline | 13.154 | 65 | 1093525 | 46.416 | ng | | 92 |
| 49) Acenaphthylene | 13.813 | 152 | 6827598 | 46.287 | ng | | 100 |
| 50) Dimethylphthalate | 13.566 | 163 | 5063196 | 41.042 | ng | | 99 |
| 51) 2,6-Dinitrotoluene | 13.666 | 165 | 1192272 | 49.554 | ng | | 92 |
| 52) Acenaphthene | 14.166 | 154 | 3864705 | 40.650 | ng | | 99 |
| 53) 3-Nitroaniline | 13.995 | 138 | 1017787 | 39.472 | ng | | 98 |
| 54) 2,4-Dinitrophenol | 14.219 | 184 | 615118 | 54.176 | ng | | 93 |
| 55) Dibenzofuran | 14.507 | 168 | 6117858 | 38.992 | ng | | 99 |
| 56) 4-Nitrophenol | 14.343 | 139 | 2164669 | 90.971 | ng | | 96 |
| 57) 2,4-Dinitrotoluene | 14.484 | 165 | 1558331 | 44.344 | ng | | 88 |
| 58) Fluorene | 15.172 | 166 | 4734515 | 39.012 | ng | | 100 |
| 59) 2,3,4,6-Tetrachlorophenol | 14.737 | 232 | 1465701 | 46.567 | ng | | 96 |
| 60) Diethylphthalate | 14.972 | 149 | 4878323 | 39.371 | ng | | 98 |
| 61) 4-Chlorophenyl-phenyle... | 15.178 | 204 | 2393795 | 39.857 | ng | | 98 |
| 62) 4-Nitroaniline | 15.195 | 138 | 1088146 | 40.246 | ng | | 100 |
| 63) Azobenzene | 15.478 | 77 | 4127404 | 37.408 | ng | | 94 |
| 65) 4,6-Dinitro-2-methylph... | 15.266 | 198 | 429995 | 32.979 | ng | | 85 |
| 66) n-Nitrosodiphenylamine | 15.401 | 169 | 4111370 | 47.136 | ng | | 99 |
| 67) 4-Bromophenyl-phenylether | 16.101 | 248 | 1663783 | 53.316 | ng | | 97 |
| 68) Hexachlorobenzene | 16.207 | 284 | 1881224 | 49.321 | ng | | 96 |
| 69) Atrazine | 16.407 | 200 | 1530515 | 59.881 | ng | | 93 |
| 70) Pentachlorophenol | 16.572 | 266 | 2559186 | 99.349 | ng | | 99 |
| 71) Phenanthrene | 16.966 | 178 | 6836970 | 41.981 | ng | | 96 |
| 72) Anthracene | 17.066 | 178 | 7041601 | 45.069 | ng | | 98 |
| 73) Carbazole | 17.348 | 167 | 6457926 | 42.274 | ng | | 97 |
| 74) Di-n-butylphthalate | 17.966 | 149 | 7440665 | 43.739 | ng | | 98 |
| 75) Fluoranthene | 19.072 | 202 | 8401898 | 44.136 | ng | | 97 |
| 77) Benzidine | 19.283 | 184 | 2380556 | 39.657 | ng | # | 93 |
| 78) Pyrene | 19.454 | 202 | 8893288 | 42.245 | ng | | 98 |
| 80) Butylbenzylphthalate | 20.436 | 149 | 3674497 | 45.510 | ng | | 97 |
| 81) Benzo(a)anthracene | 21.342 | 228 | 9427688 | 45.771 | ng | | 98 |
| 82) 3,3'-Dichlorobenzidine | 21.272 | 252 | 1837810 | 29.639 | ng | # | 99 |
| 83) Chrysene | 21.407 | 228 | 9098431 | 45.901 | ng | | 99 |
| 84) Bis(2-ethylhexyl)phtha... | 21.325 | 149 | 5892349 | 53.333 | ng | | 97 |
| 85) Di-n-octyl phthalate | 22.548 | 149 | 10363365 | 50.755 | ng | | 96 |
| 87) Indeno(1,2,3-cd)pyrene | 28.071 | 276 | 13381638m | 50.742 | ng | | |
| 88) Benzo(b)fluoranthene | 23.536 | 252 | 11384396 | 47.248 | ng | | 99 |
| 89) Benzo(k)fluoranthene | 23.601 | 252 | 10029616 | 41.835 | ng | | 99 |
| 90) Benzo(a)pyrene | 24.377 | 252 | 10226105 | 49.866 | ng | | 99 |
| 91) Dibenzo(a,h)anthracene | 28.159 | 278 | 10419062 | 47.796 | ng | | 99 |
| 92) Benzo(g,h,i)perylene | 29.183 | 276 | 10869264 | 48.267 | ng | | 97 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP071825\
Data File : BP025192.D
Acq On : 18 Jul 2025 14:33
Operator : CG/JU
Sample : Q2597-01MSD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jul 18 14:51:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP070325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 03 16:05:44 2025
Response via : Initial Calibration

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.443 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:152 Resp: 71806

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 150 | 156.2 | 126.1 | 189.1 |
| 115 | 54.3 | 44.3 | 66.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#2
1,4-Dioxane
Concen: 33.134 ng
RT: 3.120 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 88 Resp: 522161

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 88 | 100 | | |
| 58 | 67.1 | 55.0 | 82.4 |
| 43 | 27.2 | 21.9 | 32.9 |






#3
Pyridine
Concen: 33.571 ng
RT: 3.490 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 79 Resp: 144841

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 52 | 63.1 | 52.2 | 78.4 |
| 51 | 31.6 | 26.8 | 40.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#4
n-Nitrosodimethylamine
Concen: 34.763 ng
RT: 3.402 min Scan# 79
Delta R.T. -0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 42 Resp: 635044

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 42 | 100 | | |
| 74 | 142.7 | 105.2 | 157.8 |
| 44 | 8.2 | 6.6 | 9.8 |






#5
2-Fluorophenol
Concen: 82.781 ng
RT: 5.108 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:112 Resp: 335483

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 112 | 100 | | |
| 64 | 49.7 | 42.0 | 63.0 |
| 63 | 25.5 | 22.2 | 33.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#6
Aniline
Concen: 29.055 ng
RT: 6.784 min Scan# 654
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 93 Resp: 1435687

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 66 | 35.9 | 28.9 | 43.3 |
| 65 | 18.0 | 15.2 | 22.8 |






#7
Phenol-d6
Concen: 80.909 ng
RT: 6.643 min Scan# 6
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 99 Resp: 427899

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 99 | 100 | | |
| 42 | 15.9 | 13.6 | 20.4 |
| 71 | 30.6 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#8
2-Chlorophenol
Concen: 45.681 ng
RT: 7.019 min Scan# 694
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:128 Resp: 2071338

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 130 | 32.8 | 12.9 | 52.9 |
| 64 | 38.3 | 20.3 | 60.3 |






#9
Benzaldehyde
Concen: 43.131 ng
RT: 6.596 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 77 Resp: 119126

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 105 | 106.6 | 82.4 | 122.4 |
| 106 | 101.9 | 78.3 | 118.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#10
Phenol
Concen: 42.970 ng
RT: 6.672 min Scan# 635
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 94 Resp: 2309080

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 94 | 100 | | |
| 65 | 25.0 | 6.9 | 46.9 |
| 66 | 33.2 | 14.6 | 54.6 |






#11
bis(2-Chloroethyl)ether
Concen: 41.433 ng
RT: 6.884 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 93 Resp: 171353

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 63 | 68.1 | 51.9 | 91.9 |
| 95 | 32.8 | 12.4 | 52.4 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#12
1,3-Dichlorobenzene
Concen: 43.423 ng
RT: 7.331 min Scan# 747
Delta R.T. -0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:146 Resp: 2203625

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.6 | 51.9 | 77.9 |
| 75 | 26.0 | 21.8 | 32.6 |






#13
1,4-Dichlorobenzene
Concen: 43.216 ng
RT: 7.472 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:146 Resp: 222039

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.2 | 51.1 | 76.7 |
| 111 | 37.1 | 31.2 | 46.8 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#14
1,2-Dichlorobenzene
Concen: 43.727 ng
RT: 7.784 min Scan# 824
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:146 Resp: 2161225

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 146 | 100 | | |
| 148 | 64.3 | 51.1 | 76.7 |
| 111 | 39.7 | 31.2 | 46.8 |






#15
Benzyl Alcohol
Concen: 43.726 ng
RT: 7.678 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 79 Resp: 154345

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 79 | 100 | | |
| 108 | 84.9 | 66.0 | 99.0 |
| 77 | 63.3 | 53.4 | 80.2 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#16
2,2'-oxybis(1-Chloropropane)
Concen: 39.746 ng
RT: 7.960 min Scan# 854
Delta R.T. -0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 45 Resp: 2137331

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 45 | 100 | | |
| 77 | 15.9 | 0.0 | 34.8 |
| 79 | 11.7 | 0.0 | 30.7 |






#17
2-Methylphenol
Concen: 45.864 ng
RT: 7.890 min Scan# 84
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:107 Resp: 158757

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 112.0 | 89.7 | 134.5 |
| 77 | 41.8 | 34.2 | 51.4 |
| 79 | 40.8 | 33.4 | 50.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#18
Hexachloroethane
Concen: 45.363 ng
RT: 8.502 min Scan# 946
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:117 Resp: 759284

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 117 | 100 | | |
| 119 | 96.1 | 76.6 | 115.0 |
| 201 | 113.4 | 85.4 | 128.2 |






#19
n-Nitroso-di-n-propylamine
Concen: 39.169 ng
RT: 8.237 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 70 Resp: 122852

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 70 | 100 | | |
| 42 | 55.4 | 46.5 | 69.7 |
| 101 | 10.6 | 7.9 | 11.9 |
| 130 | 25.4 | 19.3 | 28.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#20
3+4-Methylphenols
Concen: 44.551 ng
RT: 8.208 min Scan# 896
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:107 Resp: 2160042

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 108 | 87.3 | 69.5 | 109.5 |
| 77 | 30.2 | 11.4 | 51.4 |
| 79 | 25.0 | 6.9 | 46.9 |






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.190 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:136 Resp: 287535

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 136 | 100 | | |
| 137 | 10.9 | 8.7 | 13.1 |
| 54 | 6.6 | 5.4 | 8.2 |
| 68 | 5.5 | 4.5 | 6.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#22
Acetophenone
Concen: 39.341 ng
RT: 8.243 min Scan# 902
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:105 Resp: 2675246

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 105 | 100 | | |
| 71 | 7.6 | 5.4 | 8.2 |
| 51 | 22.9 | 19.6 | 29.4 |
| 120 | 23.4 | 18.8 | 28.2 |






#23
Nitrobenzene-d5
Concen: 59.777 ng
RT: 8.572 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 82 Resp: 258446

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 128 | 51.3 | 38.4 | 57.6 |
| 54 | 45.2 | 37.8 | 56.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#24
Nitrobenzene
Concen: 42.692 ng
RT: 8.613 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 77 Resp: 1880097

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 123 | 52.6 | 39.0 | 58.6 |
| 65 | 12.5 | 10.7 | 16.1 |






#25
Isophorone
Concen: 44.129 ng
RT: 9.143 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 82 Resp: 356553

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 82 | 100 | | |
| 95 | 7.2 | 5.8 | 8.6 |
| 138 | 20.6 | 15.7 | 23.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#26
2-Nitrophenol
Concen: 50.908 ng
RT: 9.313 min Scan# 1084
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:139 Resp: 1092514

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 26.8 | 22.3 | 33.5 |
| 65 | 40.2 | 33.8 | 50.8 |






#27
2,4-Dimethylphenol
Concen: 61.254 ng
RT: 9.390 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:122 Resp: 184620

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 107 | 108.9 | 88.2 | 132.2 |
| 121 | 58.3 | 46.1 | 69.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#28
bis(2-Chloroethoxy)methane
Concen: 41.345 ng
RT: 9.619 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 93 Resp: 2268251

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 93 | 100 | | |
| 95 | 32.9 | 26.4 | 39.6 |
| 123 | 12.4 | 10.1 | 15.1 |






#29
2,4-Dichlorophenol
Concen: 47.336 ng
RT: 9.849 min Scan# 11
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:162 Resp: 197830

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 164 | 63.4 | 44.1 | 84.1 |
| 98 | 32.6 | 13.1 | 53.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#30
1,2,4-Trichlorobenzene
Concen: 45.153 ng
RT: 10.055 min Scan# 1210
Delta R.T. -0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:180 Resp: 2089765

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 180 | 100 | | |
| 182 | 96.2 | 75.1 | 112.7 |
| 145 | 29.4 | 23.3 | 34.9 |






#31
Naphthalene
Concen: 41.322 ng
RT: 10.243 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:128 Resp: 611803

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 128 | 100 | | |
| 129 | 11.1 | 8.9 | 13.3 |
| 127 | 13.2 | 10.4 | 15.6 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#32
Benzoic acid
Concen: 51.506 ng
RT: 9.560 min Scan# 1126
Delta R.T. 0.059 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:122 Resp: 1566557

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 122 | 100 | | |
| 105 | 117.0 | 101.0 | 141.0 |
| 77 | 84.4 | 67.6 | 107.6 |






#33
4-Chloroaniline
Concen: 22.604 ng
RT: 10.349 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:127 Resp: 124893

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 127 | 100 | | |
| 129 | 32.6 | 25.4 | 38.2 |
| 65 | 26.1 | 21.0 | 31.6 |
| 92 | 17.3 | 13.4 | 20.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#34
Hexachlorobutadiene
Concen: 48.169 ng
RT: 10.531 min Scan# 1291
Delta R.T. -0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:225 Resp: 1279328

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 225 | 100 | | |
| 223 | 62.0 | 50.3 | 75.5 |
| 227 | 64.4 | 51.3 | 76.9 |






#35
Caprolactam
Concen: 42.085 ng m
RT: 11.143 min Scan# 1
Delta R.T. 0.035 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:113 Resp: 60170

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 113 | 100 | | |
| 55 | 140.4 | 128.1 | 168.1 |
| 56 | 101.7 | 90.7 | 130.7 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#36
4-Chloro-3-methylphenol
Concen: 43.440 ng
RT: 11.496 min Scan# 1455
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:107 Resp: 1897199

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 107 | 100 | | |
| 144 | 28.3 | 22.4 | 33.6 |
| 142 | 87.2 | 69.6 | 104.4 |






#37
2-Methylnaphthalene
Concen: 37.677 ng
RT: 11.860 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:142 Resp: 393592

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 89.1 | 71.2 | 106.8 |
| 115 | 31.9 | 25.9 | 38.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#38
1-Methylnaphthalene
Concen: 40.044 ng
RT: 12.096 min Scan# 1557
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:142 Resp: 4079783

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 142 | 100 | | |
| 141 | 91.8 | 74.1 | 111.1 |
| 116 | 3.4 | 2.7 | 4.1 |






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.095 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:164 Resp: 176235

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 164 | 100 | | |
| 162 | 101.4 | 80.3 | 120.5 |
| 160 | 44.9 | 35.9 | 53.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#40
1,2,4,5-Tetrachlorobenzene
Concen: 46.625 ng
RT: 12.243 min Scan# 1582
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:216 Resp: 2361542

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 216 | 100 | | |
| 214 | 78.5 | 62.4 | 93.6 |
| 179 | 20.2 | 16.5 | 24.7 |
| 108 | 16.4 | 13.7 | 20.5 |






#41
Hexachlorocyclopentadiene
Concen: 115.832 ng
RT: 12.225 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:237 Resp: 153245

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 237 | 100 | | |
| 235 | 62.9 | 41.6 | 81.6 |
| 272 | 13.8 | 0.0 | 33.6 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#42
2,4,6-Tribromophenol
Concen: 104.905 ng
RT: 15.637 min Scan# 2159
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:330 Resp: 2184054

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 330 | 100 | | |
| 332 | 96.7 | 77.7 | 116.5 |
| 141 | 26.9 | 26.0 | 39.0 |






#43
2,4,6-Trichlorophenol
Concen: 52.744 ng
RT: 12.496 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:196 Resp: 161763

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 96.3 | 76.2 | 114.2 |
| 200 | 31.4 | 24.3 | 36.5 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#44
2,4,5-Trichlorophenol
Concen: 50.095 ng
RT: 12.572 min Scan# 1638
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:196 Resp: 1741173

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 196 | 100 | | |
| 198 | 95.4 | 76.6 | 115.0 |
| 97 | 41.4 | 36.6 | 55.0 |
| 132 | 25.6 | 21.4 | 32.0 |






#45
2-Fluorobiphenyl
Concen: 60.625 ng
RT: 12.701 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:172 Resp: 689681

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 172 | 100 | | |
| 171 | 35.4 | 28.2 | 42.2 |
| 170 | 23.5 | 18.7 | 28.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#46
1,1'-Biphenyl
Concen: 42.534 ng
RT: 12.907 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:154 Resp: 5453993

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 40.4 | 20.2 | 60.2 |
| 76 | 11.4 | 0.0 | 31.5 |






#47
2-Chloronaphthalene
Concen: 44.415 ng
RT: 12.943 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:162 Resp: 428052

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 162 | 100 | | |
| 127 | 37.8 | 29.4 | 44.2 |
| 164 | 33.2 | 26.4 | 39.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#48
2-Nitroaniline
Concen: 46.416 ng
RT: 13.154 min Scan# 1737
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 65 Resp: 1093525

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 65 | 100 | | |
| 92 | 75.8 | 58.7 | 88.1 |
| 138 | 136.0 | 99.0 | 148.4 |






#49
Acenaphthylene
Concen: 46.287 ng
RT: 13.813 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:152 Resp: 682759

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 152 | 100 | | |
| 151 | 20.3 | 16.1 | 24.1 |
| 153 | 13.3 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#50
Dimethylphthalate
Concen: 41.042 ng
RT: 13.566 min Scan# 1807
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:163 Resp: 5063196

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 163 | 100 | | |
| 194 | 4.3 | 3.4 | 5.0 |
| 164 | 10.3 | 8.0 | 12.0 |






#51
2,6-Dinitrotoluene
Concen: 49.554 ng
RT: 13.666 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:165 Resp: 119227

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 42.4 | 38.5 | 57.7 |
| 89 | 45.7 | 40.6 | 61.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#52
Acenaphthene
Concen: 40.650 ng
RT: 14.166 min Scan# 1909
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:154 Resp: 3864705

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 154 | 100 | | |
| 153 | 113.4 | 90.6 | 136.0 |
| 152 | 53.7 | 43.8 | 65.8 |






#53
3-Nitroaniline
Concen: 39.472 ng
RT: 13.995 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:138 Resp: 101778

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 108 | 10.6 | 7.8 | 11.8 |
| 92 | 99.7 | 81.0 | 121.4 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#54
2,4-Dinitrophenol
Concen: 54.176 ng
RT: 14.219 min Scan# 1918
Delta R.T. 0.024 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:184 Resp: 615118

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 63 | 50.2 | 47.4 | 71.0 |
| 154 | 58.9 | 47.8 | 71.8 |






#55
Dibenzofuran
Concen: 38.992 ng
RT: 14.507 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:168 Resp: 611785

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 168 | 100 | | |
| 139 | 36.7 | 28.6 | 43.0 |
| 169 | 13.8 | 10.7 | 16.1 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#56
4-Nitrophenol
Concen: 90.971 ng
RT: 14.343 min Scan# 1939
Delta R.T. 0.041 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:139 Resp: 2164669

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 139 | 100 | | |
| 109 | 68.5 | 45.9 | 85.9 |
| 65 | 77.0 | 61.4 | 101.4 |






#57
2,4-Dinitrotoluene
Concen: 44.344 ng
RT: 14.484 min Scan# 1
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:165 Resp: 155833

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 165 | 100 | | |
| 63 | 32.4 | 31.0 | 46.4 |
| 89 | 56.3 | 53.4 | 80.0 |
| 182 | 3.0 | 2.0 | 3.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**
Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#58
Fluorene
Concen: 39.012 ng
RT: 15.172 min Scan# 2080
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:166 Resp: 4734515

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 166 | 100 | | |
| 165 | 98.6 | 78.6 | 117.8 |
| 167 | 13.7 | 10.7 | 16.1 |






#59
2,3,4,6-Tetrachlorophenol
Concen: 46.567 ng
RT: 14.737 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:232 Resp: 146570

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 232 | 100 | | |
| 131 | 41.7 | 35.7 | 53.5 |
| 130 | 2.5 | 1.8 | 2.6 |
| 166 | 28.2 | 24.4 | 36.6 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#60
Diethylphthalate
Concen: 39.371 ng
RT: 14.972 min Scan# 2046
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:149 Resp: 4878323

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 177 | 23.6 | 17.8 | 26.8 |
| 150 | 13.1 | 10.2 | 15.2 |






#61
4-Chlorophenyl-phenylether
Concen: 39.857 ng
RT: 15.178 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:204 Resp: 239379

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 204 | 100 | | |
| 206 | 34.2 | 25.8 | 38.6 |
| 141 | 55.9 | 45.1 | 67.7 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#62
4-Nitroaniline
Concen: 40.246 ng
RT: 15.195 min Scan# 2084
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:138 Resp: 1088146

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 138 | 100 | | |
| 92 | 46.0 | 26.0 | 66.0 |
| 108 | 69.8 | 50.0 | 90.0 |






#63
Azobenzene
Concen: 37.408 ng
RT: 15.478 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion: 77 Resp: 412740

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 77 | 100 | | |
| 182 | 31.6 | 7.7 | 47.7 |
| 105 | 25.8 | 1.9 | 41.9 |
| 51 | 26.6 | 8.2 | 48.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.919 min Scan# 2377
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:188 Resp: 3047795

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 188 | 100 | | |
| 94 | 11.9 | 7.3 | 10.9# |
| 80 | 12.1 | 7.8 | 11.6# |






#65
4,6-Dinitro-2-methylphenol
Concen: 32.979 ng
RT: 15.266 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:198 Resp: 42999

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 198 | 100 | | |
| 51 | 34.3 | 16.8 | 56.8 |
| 105 | 56.7 | 20.8 | 60.8 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#66
n-Nitrosodiphenylamine
Concen: 47.136 ng
RT: 15.401 min Scan# 2119
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:169 Resp: 4111370

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 169 | 100 | | |
| 168 | 67.1 | 54.1 | 81.1 |
| 167 | 36.5 | 29.0 | 43.6 |






#67
4-Bromophenyl-phenylether
Concen: 53.316 ng
RT: 16.101 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:248 Resp: 166378

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 248 | 100 | | |
| 250 | 96.3 | 79.4 | 119.2 |
| 141 | 62.0 | 51.4 | 77.0 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#68
Hexachlorobenzene
Concen: 49.321 ng
RT: 16.207 min Scan# 2256
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:284 Resp: 1881224

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 284 | 100 | | |
| 142 | 28.6 | 25.9 | 38.9 |
| 249 | 29.5 | 23.8 | 35.8 |






#69
Atrazine
Concen: 59.881 ng
RT: 16.407 min Scan# 2
Delta R.T. 0.024 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:200 Resp: 153051

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 200 | 100 | | |
| 173 | 31.3 | 6.6 | 46.6 |
| 215 | 52.4 | 28.0 | 68.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#70
Pentachlorophenol
Concen: 99.349 ng
RT: 16.572 min Scan# 2318
Delta R.T. 0.024 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:266 Resp: 2559186

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 266 | 100 | | |
| 268 | 63.9 | 50.4 | 75.6 |
| 264 | 62.6 | 50.2 | 75.2 |






#71
Phenanthrene
Concen: 41.981 ng
RT: 16.966 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:178 Resp: 683697

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.7 | 15.4 | 23.2 |
| 179 | 18.8 | 12.6 | 19.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#72
Anthracene
Concen: 45.069 ng
RT: 17.066 min Scan# 2402
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:178 Resp: 7041601

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 178 | 100 | | |
| 176 | 19.3 | 15.0 | 22.6 |
| 179 | 17.1 | 12.5 | 18.7 |






#73
Carbazole
Concen: 42.274 ng
RT: 17.348 min Scan# 24
Delta R.T. 0.024 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:167 Resp: 645792(

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 167 | 100 | | |
| 166 | 22.6 | 17.4 | 26.0 |
| 139 | 14.6 | 10.4 | 15.6 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#74
Di-n-butylphthalate
Concen: 43.739 ng
RT: 17.966 min Scan# 2555
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:149 Resp: 7440665

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 150 | 10.4 | 7.3 | 10.9 |
| 104 | 4.8 | 3.8 | 5.6 |






#75
Fluoranthene
Concen: 44.136 ng
RT: 19.072 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:202 Resp: 840189

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 101 | 9.9 | 0.0 | 30.4 |
| 203 | 19.0 | 0.0 | 37.6 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.360 min Scan# 3132
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:240 Resp: 3398852

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 240 | 100 | | |
| 120 | 11.8 | 9.0 | 13.4 |
| 236 | 26.0 | 20.3 | 30.5 |






#77
Benzidine
Concen: 39.657 ng
RT: 19.283 min Scan# 2
Delta R.T. 0.017 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:184 Resp: 238055

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 184 | 100 | | |
| 185 | 17.9 | 12.6 | 18.8 |
| 183 | 15.2 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#78
Pyrene
Concen: 42.245 ng
RT: 19.454 min Scan# 2808
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:202 Resp: 8893288

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 202 | 100 | | |
| 200 | 21.8 | 17.1 | 25.7 |
| 203 | 18.7 | 14.1 | 21.1 |






#79
Terphenyl-d14
Concen: 62.676 ng
RT: 19.689 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:244 Resp: 969269

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 244 | 100 | | |
| 212 | 7.3 | 5.9 | 8.9 |
| 122 | 10.6 | 9.4 | 14.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#80
Butylbenzylphthalate
Concen: 45.510 ng
RT: 20.436 min Scan# 2975
Delta R.T. 0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:149 Resp: 3674497

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 91 | 62.0 | 50.6 | 76.0 |
| 206 | 24.0 | 17.2 | 25.8 |






#81
Benzo(a)anthracene
Concen: 45.771 ng
RT: 21.342 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:228 Resp: 942768

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 28.2 | 22.1 | 33.1 |
| 229 | 20.9 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#82
3,3'-Dichlorobenzidine
Concen: 29.639 ng
RT: 21.272 min Scan# 3117
Delta R.T. 0.000 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:252 Resp: 1837810

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 254 | 64.7 | 51.9 | 77.9 |
| 126 | 12.2 | 8.1 | 12.1# |






#83
Chrysene
Concen: 45.901 ng
RT: 21.407 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:228 Resp: 909843

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 228 | 100 | | |
| 226 | 31.0 | 24.4 | 36.6 |
| 229 | 20.6 | 16.0 | 24.0 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#84
Bis(2-ethylhexyl)phthalate
Concen: 53.333 ng
RT: 21.325 min Scan# 3126
Delta R.T. -0.006 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:149 Resp: 5892349

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 29.0 | 21.9 | 32.9 |
| 279 | 5.2 | 3.5 | 5.3 |






#85
Di-n-octyl phthalate
Concen: 50.755 ng
RT: 22.548 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:149 Resp:1036336

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 149 | 100 | | |
| 167 | 1.6 | 1.3 | 1.9 |
| 43 | 7.0 | 6.9 | 10.3 |

**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025







#86
Perylene-d12
Concen: 20.000 ng
RT: 24.524 min Scan# 3670
Delta R.T. 0.030 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:264 Resp: 4139648

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 264 | 100 | | |
| 260 | 23.6 | 18.6 | 28.0 |
| 265 | 22.1 | 17.7 | 26.5 |






#87
Indeno(1,2,3-cd)pyrene
Concen: 50.742 ng m
RT: 28.071 min Scan# 4
Delta R.T. 0.035 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:276 Resp:1338163

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 138 | 16.6 | 13.0 | 19.6 |
| 277 | 20.9 | 15.9 | 23.9 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#88
Benzo(b)fluoranthene
Concen: 47.248 ng
RT: 23.536 min Scan# 3502
Delta R.T. 0.024 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:252 Resp:11384396

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.4 | 17.7 | 26.5 |
| 125 | 9.8 | 8.1 | 12.1 |






#89
Benzo(k)fluoranthene
Concen: 41.835 ng
RT: 23.601 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:252 Resp:1002961

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.8 | 17.7 | 26.5 |
| 125 | 10.2 | 8.2 | 12.2 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#90
Benzo(a)pyrene
Concen: 49.866 ng
RT: 24.377 min Scan# 3645
Delta R.T. 0.024 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:252 Resp:10226105

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 252 | 100 | | |
| 253 | 22.1 | 17.6 | 26.4 |
| 125 | 10.6 | 9.0 | 13.4 |






#91
Dibenzo(a,h)anthracene
Concen: 47.796 ng
RT: 28.159 min Scan# 4
Delta R.T. 0.024 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:278 Resp:1041906

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 278 | 100 | | |
| 139 | 15.5 | 13.1 | 19.7 |
| 279 | 24.1 | 18.9 | 28.3 |




**Instrument :**
BNA_P
**ClientSampleId :**
CHRT-26899MSD

**Manual Integrations**
**APPROVED**

Reviewed By :Rahul Chavli 07/21/2025
Supervised By :Jagrut Upadhyay 07/21/2025




#92
Benzo(g,h,i)perylene
Concen: 48.267 ng
RT: 29.183 min Scan# 4462
Delta R.T. 0.059 min
Lab File: BP025192.D
Acq: 18 Jul 2025 14:33

Tgt Ion:276 Resp:10869264

| Ion | Ratio | Lower | Upper |
|---|---|---|---|
| 276 | 100 | | |
| 277 | 23.2 | 19.3 | 28.9 |
| 138 | 19.3 | 16.9 | 25.3 |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

# Manual Integration Report

| Sequence: | BF071525 | Instrument | BNA_f |
|---|---|---|---|

| Sample ID | File ID | Parameter | Review By | Review On | Supervised By | Supervised On | Reason |
|---|---|---|---|---|---|---|---|
| SSTDICC005 | BF143094.D | Nitrobenzene-d5 | Rahul | 7/16/2025 11:43:26 AM | Jagrut | 7/16/2025 1:40:56 PM | Peak Integrated by Software |
| SSTDICC010 | BF143095.D | Benzoic acid | Rahul | 7/16/2025 11:43:28 AM | Jagrut | 7/16/2025 1:40:58 PM | Peak Integrated by Software |
| SSTDICC010 | BF143095.D | Nitrobenzene-d5 | Rahul | 7/16/2025 11:43:28 AM | Jagrut | 7/16/2025 1:40:58 PM | Peak Integrated by Software |
| SSTDICC020 | BF143096.D | Benzoic acid | Rahul | 7/16/2025 11:43:30 AM | Jagrut | 7/16/2025 1:41:01 PM | Peak Integrated by Software |
| SSTDICC020 | BF143096.D | Nitrobenzene-d5 | Rahul | 7/16/2025 11:43:30 AM | Jagrut | 7/16/2025 1:41:01 PM | Peak Integrated by Software |
| SSTDICC020 | BF143096.D | Phenol | Rahul | 7/16/2025 11:43:30 AM | Jagrut | 7/16/2025 1:41:01 PM | Peak Integrated by Software |
| SSTDICCC040 | BF143097.D | Benzoic acid | Rahul | 7/16/2025 11:43:33 AM | Jagrut | 7/16/2025 1:41:03 PM | Peak Integrated by Software |
| SSTDICCC040 | BF143097.D | Nitrobenzene-d5 | Rahul | 7/16/2025 11:43:33 AM | Jagrut | 7/16/2025 1:41:03 PM | Peak Integrated by Software |
| SSTDICCC040 | BF143097.D | Phenol | Rahul | 7/16/2025 11:43:33 AM | Jagrut | 7/16/2025 1:41:03 PM | Peak Integrated by Software |
| SSTDICC050 | BF143098.D | Phenol | Rahul | 7/16/2025 11:43:36 AM | Jagrut | 7/16/2025 1:41:06 PM | Peak Integrated by Software |
| SSTDICC060 | BF143099.D | Benzoic acid | Rahul | 7/16/2025 11:43:39 AM | Jagrut | 7/16/2025 1:41:09 PM | Peak Integrated by Software |
| SSTDICC060 | BF143099.D | Phenol | Rahul | 7/16/2025 11:43:39 AM | Jagrut | 7/16/2025 1:41:09 PM | Peak Integrated by Software |
| SSTDICC080 | BF143100.D | Phenol | Rahul | 7/16/2025 11:43:42 AM | Jagrut | 7/16/2025 1:41:12 PM | Peak Integrated by Software |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

# Manual Integration Report

| Sequence: | BF071525 | Instrument | BNA_f |
|---|---|---|---|

| Sample ID | File ID | Parameter | Review By | Review On | Supervised By | Supervised On | Reason |
|---|---|---|---|---|---|---|---|
| SSTDICV040 | BF143101.D | Benzoic acid | Rahul | 7/16/2025 11:43:45 AM | Jagrut | 7/16/2025 1:41:15 PM | Peak Integrated by Software |
| SSTDICV040 | BF143101.D | Nitrobenzene-d5 | Rahul | 7/16/2025 11:43:45 AM | Jagrut | 7/16/2025 1:41:15 PM | Peak Integrated by Software |
| SSTDICV040 | BF143101.D | Phenol | Rahul | 7/16/2025 11:43:45 AM | Jagrut | 7/16/2025 1:41:15 PM | Peak Integrated by Software |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

# Manual Integration Report

| Sequence: | bf071625 | Instrument | BNA_f |
|---|---|---|---|

| Sample ID | File ID | Parameter | Review By | Review On | Supervised By | Supervised On | Reason |
|---|---|---|---|---|---|---|---|
| SSTDCCC040 | BF143115.D | Benzoic acid | Rahul | 7/17/2025 10:01:35 AM | Jagrut | 7/17/2025 11:19:12 AM | Peak Integrated by Software |
| PB168854BS | BF143119.D | Caprolactam | Rahul | 7/17/2025 10:01:37 AM | Jagrut | 7/17/2025 11:19:15 AM | Peak Integrated by Software |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

# Manual Integration Report

| Sequence: | BP070325 | Instrument | BNA_p |
|---|---|---|---|

| Sample ID | File ID | Parameter | Review By | Review On | Supervised By | Supervised On | Reason |
|---|---|---|---|---|---|---|---|
| SSTDICC005 | BP025072.D | Acenaphthene | Rahul | 7/7/2025 11:35:36 AM | Jagrut | 7/7/2025 12:14:15 PM | Peak Integrated by Software |
| SSTDICC005 | BP025072.D | Indeno(1,2,3-cd)pyrene | Rahul | 7/7/2025 11:35:36 AM | Jagrut | 7/7/2025 12:14:15 PM | Peak Integrated by Software |
| SSTDICC010 | BP025073.D | Acenaphthene | Rahul | 7/7/2025 11:35:39 AM | Jagrut | 7/7/2025 12:14:23 PM | Peak Integrated by Software |
| SSTDICC020 | BP025074.D | Acenaphthene | Rahul | 7/7/2025 11:35:42 AM | Jagrut | 7/7/2025 12:14:25 PM | Peak Integrated by Software |
| SSTDICCC040 | BP025075.D | Indeno(1,2,3-cd)pyrene | Rahul | 7/7/2025 11:35:44 AM | Jagrut | 7/7/2025 12:14:28 PM | Peak Integrated by Software |
| SSTDICC050 | BP025076.D | Indeno(1,2,3-cd)pyrene | Rahul | 7/7/2025 11:35:46 AM | Jagrut | 7/7/2025 12:14:34 PM | Peak Integrated by Software |
| SSTDICC050 | BP025076.D | Pyridine | Rahul | 7/7/2025 11:35:46 AM | Jagrut | 7/7/2025 12:14:34 PM | Peak Integrated by Software |
| SSTDICC060 | BP025077.D | Indeno(1,2,3-cd)pyrene | Rahul | 7/7/2025 11:35:49 AM | Jagrut | 7/7/2025 12:14:37 PM | Peak Integrated by Software |
| SSTDICC060 | BP025077.D | Pyridine | Rahul | 7/7/2025 11:35:49 AM | Jagrut | 7/7/2025 12:14:37 PM | Peak Integrated by Software |
| SSTDICC080 | BP025078.D | Indeno(1,2,3-cd)pyrene | Rahul | 7/7/2025 11:35:51 AM | Jagrut | 7/7/2025 12:14:42 PM | Peak Integrated by Software |
| SSTDICC080 | BP025078.D | Pyridine | Rahul | 7/7/2025 11:35:51 AM | Jagrut | 7/7/2025 12:14:42 PM | Peak Integrated by Software |
| SSTDICV040 | BP025079.D | Acenaphthene | Rahul | 7/7/2025 11:35:54 AM | Jagrut | 7/7/2025 12:14:48 PM | Peak Integrated by Software |
| SSTDICV040 | BP025079.D | Indeno(1,2,3-cd)pyrene | Rahul | 7/7/2025 11:35:54 AM | Jagrut | 7/7/2025 12:14:48 PM | Peak Integrated by Software |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

# Manual Integration Report

| Sequence: | BP070325 | Instrument | BNA_p |
|---|---|---|---|

| Sample ID | File ID | Parameter | Review By | Review On | Supervised By | Supervised On | Reason |
|---|---|---|---|---|---|---|---|

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

# Manual Integration Report

| Sequence: | bp071625 | Instrument | BNA_p |
|---|---|---|---|

| Sample ID | File ID | Parameter | Review By | Review On | Supervised By | Supervised On | Reason |
|---|---|---|---|---|---|---|---|
| SSTDCCC040 | BP025155.D | Indeno(1,2,3-cd)pyrene | Rahul | 7/17/2025 11:24:00 AM | Jagrut | 7/17/2025 11:29:25 AM | Peak Integrated by Software |
| Q2592-01 | BP025166.D | Chrysene | Rahul | 7/17/2025 11:24:21 AM | Jagrut | 7/17/2025 11:29:47 AM | Peak Integrated by Software |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

# Manual Integration Report

| Sequence: | BP071825 | Instrument | BNA_p |
|---|---|---|---|

| Sample ID | File ID | Parameter | Review By | Review On | Supervised By | Supervised On | Reason |
|---|---|---|---|---|---|---|---|
| SSTDCCC040 | BP025187.D | Indeno(1,2,3-cd)pyrene | Rahul | 7/21/2025 8:20:24 AM | Jagrut | 7/21/2025 9:29:54 AM | Peak Integrated by Software |
| Q2597-01MS | BP025191.D | Caprolactam | Rahul | 7/21/2025 8:20:30 AM | Jagrut | 7/21/2025 9:29:58 AM | Peak Integrated by Software |
| Q2597-01MS | BP025191.D | Indeno(1,2,3-cd)pyrene | Rahul | 7/21/2025 8:20:30 AM | Jagrut | 7/21/2025 9:29:58 AM | Peak Integrated by Software |
| Q2597-01MSD | BP025192.D | Caprolactam | Rahul | 7/21/2025 8:20:33 AM | Jagrut | 7/21/2025 9:30:00 AM | Peak Integrated by Software |
| Q2597-01MSD | BP025192.D | Indeno(1,2,3-cd)pyrene | Rahul | 7/21/2025 8:20:33 AM | Jagrut | 7/21/2025 9:30:00 AM | Peak Integrated by Software |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Instrument ID: BNA_F**

## Daily Analysis Runlog For Sequence/QCBatch ID # BF071525

| | | | | | |
|---|---|---|---|---|---|
| Review By | Rahul | Review On | 7/16/2025 11:47:09 AM | | |
| Supervise By | Jagrut | Supervise On | 7/16/2025 1:41:28 PM | | |
| SubDirectory | BF071525 | HP Acquire Method | BNA_F | HP Processing Method | Bf071525 |

| STD. NAME | STD REF.# |
|---|---|
| Tune/Reschk | SP6757 |
| Initial Calibration Stds | SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840 |
| CCC | SP6836 |
| Internal Standard/PEM | S12674,10ul/1000ul sample |
| ICV/I.BLK | SP6770 |
| Surrogate Standard | |
| MS/MSD Standard | |
| LCS Standard | |

| Sr# | SampleId | Data File Name | Date-Time | Operator | Status |
|---|---|---|---|---|---|
| 1 | DFTPP | BF143092.D | 15 Jul 2025 12:35 | RC/JU | Ok |
| 2 | SSTDICC2.5 | BF143093.D | 15 Jul 2025 13:04 | RC/JU | Ok |
| 3 | SSTDICC005 | BF143094.D | 15 Jul 2025 13:35 | RC/JU | Ok,M |
| 4 | SSTDICC010 | BF143095.D | 15 Jul 2025 14:05 | RC/JU | Ok,M |
| 5 | SSTDICC020 | BF143096.D | 15 Jul 2025 14:36 | RC/JU | Ok,M |
| 6 | SSTDICCC040 | BF143097.D | 15 Jul 2025 15:05 | RC/JU | Ok,M |
| 7 | SSTDICC050 | BF143098.D | 15 Jul 2025 15:35 | RC/JU | Ok,M |
| 8 | SSTDICC060 | BF143099.D | 15 Jul 2025 16:05 | RC/JU | Ok,M |
| 9 | SSTDICC080 | BF143100.D | 15 Jul 2025 16:35 | RC/JU | Ok,M |
| 10 | SSTDICV040 | BF143101.D | 15 Jul 2025 17:27 | RC/JU | Ok,M |
| 11 | PB168737BL | BF143102.D | 15 Jul 2025 17:57 | RC/JU | Ok |

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Instrument ID: BNA_F**

## Daily Analysis Runlog For Sequence/QCBatch ID # BF071625

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| Review By | Rahul | Review On | 7/17/2025 11:19:37 AM | | | | |
| Supervise By | Jagrut | Supervise On | 7/17/2025 11:19:59 AM | | | | |
| SubDirectory | BF071625 | HP Acquire Method | BNA_F | | HP Processing Method | Bf071525 | |

| STD. NAME | STD REF.# |
|---|---|
| Tune/Reschk | SP6757 |
| Initial Calibration Stds | SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840 |
| CCC | SP6836 |
| Internal Standard/PEM | S12674,10ul/1000ul sample |
| ICV/I.BLK | SP6770 |
| Surrogate Standard | |
| MS/MSD Standard | |
| LCS Standard | |

| Sr# | SampleId | Data File Name | Date-Time | Operator | Status |
|---|---|---|---|---|---|
| 1 | DFTPP | BF143103.D | 16 Jul 2025 09:06 | RC/JU | Ok |
| 2 | SSTDCCC040 | BF143104.D | 16 Jul 2025 10:04 | RC/JU | Ok |
| 3 | PB168787BL | BF143105.D | 16 Jul 2025 10:34 | RC/JU | Ok |
| 4 | Q2126-02 | BF143106.D | 16 Jul 2025 11:03 | RC/JU | Ok |
| 5 | Q2126-02 | BF143107.D | 16 Jul 2025 11:37 | RC/JU | Ok |
| 6 | Q2126-08 | BF143108.D | 16 Jul 2025 12:07 | RC/JU | Ok |
| 7 | Q2126-08 | BF143109.D | 16 Jul 2025 12:36 | RC/JU | Ok |
| 8 | Q2126-01 | BF143110.D | 16 Jul 2025 13:05 | RC/JU | Ok |
| 9 | Q2126-01 | BF143111.D | 16 Jul 2025 13:35 | RC/JU | Ok |
| 10 | Q2126-07 | BF143112.D | 16 Jul 2025 14:05 | RC/JU | Ok |
| 11 | Q2126-07 | BF143113.D | 16 Jul 2025 14:35 | RC/JU | Ok |
| 12 | PB168802BL | BF143114.D | 16 Jul 2025 15:13 | RC/JU | Ok |
| 13 | SSTDCCC040 | BF143115.D | 16 Jul 2025 15:45 | RC/JU | Ok,M |
| 14 | DFTPP | BF143116.D | 16 Jul 2025 19:09 | RC/JU | Ok |
| 15 | SSTDCCC040 | BF143117.D | 16 Jul 2025 19:39 | RC/JU | Ok |
| 16 | PB168854BL | BF143118.D | 16 Jul 2025 20:09 | RC/JU | Ok |
| 17 | PB168854BS | BF143119.D | 16 Jul 2025 20:39 | RC/JU | Ok,M |
| 18 | PB168847TB | BF143120.D | 16 Jul 2025 21:09 | RC/JU | Ok |
| 19 | Q2592-02 | BF143121.D | 16 Jul 2025 21:39 | RC/JU | Ok |
| 20 | Q2592-02MS | BF143122.D | 16 Jul 2025 22:09 | RC/JU | Ok,M |
| 21 | Q2592-02MSD | BF143123.D | 16 Jul 2025 22:38 | RC/JU | Ok,M |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Instrument ID: BNA_F**

## Daily Analysis Runlog For Sequence/QCBatch ID # BF071625

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| Review By | Rahul | Review On | 7/17/2025 11:19:37 AM | | | | |
| Supervise By | Jagrut | Supervise On | 7/17/2025 11:19:59 AM | | | | |
| SubDirectory | BF071625 | HP Acquire Method | BNA_F | | HP Processing Method | Bf071525 | |

| STD. NAME | STD REF.# |
|---|---|
| Tune/Reschk | SP6757 |
| Initial Calibration Stds | SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840 |
| CCC | SP6836 |
| Internal Standard/PEM | S12674,10ul/1000ul sample |
| ICV/I.BLK | SP6770 |
| Surrogate Standard | |
| MS/MSD Standard | |
| LCS Standard | |

| | | | | | |
|---|---|---|---|---|---|
| 22 | Q2600-02 | BF143124.D | 16 Jul 2025 23:08 | RC/JU | Ok |
| 23 | Q2600-06 | BF143125.D | 16 Jul 2025 23:37 | RC/JU | Ok |
| 24 | Q2600-10 | BF143126.D | 17 Jul 2025 00:07 | RC/JU | Ok,M |
| 25 | Q2614-06 | BF143127.D | 17 Jul 2025 00:37 | RC/JU | Ok |
| 26 | Q2605-03 | BF143128.D | 17 Jul 2025 01:05 | RC/JU | Ok |
| 27 | Q2605-04 | BF143129.D | 17 Jul 2025 01:35 | RC/JU | Ok |
| 28 | Q2571-05 | BF143130.D | 17 Jul 2025 02:04 | RC/JU | Ok |
| 29 | Q2571-05MS | BF143131.D | 17 Jul 2025 02:34 | RC/JU | Ok |
| 30 | Q2571-05MSD | BF143132.D | 17 Jul 2025 03:03 | RC/JU | Ok |
| 31 | Q2555-01 | BF143133.D | 17 Jul 2025 03:32 | RC/JU | Ok,M |
| 32 | Q2558-01 | BF143134.D | 17 Jul 2025 04:02 | RC/JU | Ok,M |
| 33 | Q2555-03 | BF143135.D | 17 Jul 2025 04:31 | RC/JU | Ok,M |
| 34 | Q2558-03 | BF143136.D | 17 Jul 2025 05:00 | RC/JU | Ok,M |
| 35 | SSTDCCC040 | BF143137.D | 17 Jul 2025 05:30 | RC/JU | Ok |

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Instrument ID: BNA_P**

## Daily Analysis Runlog For Sequence/QCBatch ID # BP070325

| | | | | | |
|---|---|---|---|---|---|
| Review By | Rahul | Review On | 7/7/2025 11:36:11 AM | | |
| Supervise By | Jagrut | Supervise On | 7/7/2025 12:15:03 PM | | |
| SubDirectory | BP070325 | HP Acquire Method | BNA_P | HP Processing Method | bp070325 |

| STD. NAME | STD REF.# |
|---|---|
| Tune/Reschk | SP6757 |
| Initial Calibration Stds | SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840 |
| CCC | SP6836 |
| Internal Standard/PEM | S12672,10ul/1000ul sample |
| ICV/I.BLK | SP6770 |
| Surrogate Standard | |
| MS/MSD Standard | |
| LCS Standard | |

| Sr# | SampleId | Data File Name | Date-Time | Operator | Status |
|---|---|---|---|---|---|
| 1 | DFTPP | BP025070.D | 03 Jul 2025 08:58 | CG/JU | Ok |
| 2 | SSTDICC2.5 | BP025071.D | 03 Jul 2025 09:39 | CG/JU | Ok |
| 3 | SSTDICC005 | BP025072.D | 03 Jul 2025 10:21 | CG/JU | Ok,M |
| 4 | SSTDICC010 | BP025073.D | 03 Jul 2025 11:02 | CG/JU | Ok,M |
| 5 | SSTDICC020 | BP025074.D | 03 Jul 2025 11:44 | CG/JU | Ok,M |
| 6 | SSTDICCC040 | BP025075.D | 03 Jul 2025 12:25 | CG/JU | Ok,M |
| 7 | SSTDICC050 | BP025076.D | 03 Jul 2025 13:07 | CG/JU | Ok,M |
| 8 | SSTDICC060 | BP025077.D | 03 Jul 2025 13:49 | CG/JU | Ok,M |
| 9 | SSTDICC080 | BP025078.D | 03 Jul 2025 14:31 | CG/JU | Ok,M |
| 10 | SSTDICV040 | BP025079.D | 03 Jul 2025 16:18 | CG/JU | Ok,M |
| 11 | PB168694BL | BP025080.D | 03 Jul 2025 16:59 | CG/JU | Not Ok |

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Instrument ID: BNA_P**

## Daily Analysis Runlog For Sequence/QCBatch ID # BP071625

| | | | | | |
|---|---|---|---|---|---|
| Review By | Rahul | Review On | 7/17/2025 11:25:01 AM | | |
| Supervise By | Jagrut | Supervise On | 7/17/2025 11:34:34 AM | | |
| SubDirectory | BP071625 | HP Acquire Method | BNA_P | HP Processing Method | bp070325 |

| STD. NAME | STD REF.# |
|---|---|
| Tune/Reschk | SP6757 |
| Initial Calibration Stds | SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840 |
| CCC | SP6836 |
| Internal Standard/PEM | S12673,10ul/1000ul sample |
| ICV/I.BLK | SP6770 |
| Surrogate Standard | |
| MS/MSD Standard | |
| LCS Standard | |

| Sr# | SampleId | Data File Name | Date-Time | Operator | Status |
|---|---|---|---|---|---|
| 1 | DFTPP | BP025154.D | 16 Jul 2025 11:19 | CG/JU | Ok |
| 2 | SSTDCCC040 | BP025155.D | 16 Jul 2025 12:42 | CG/JU | Ok,M |
| 3 | PB168832BL | BP025156.D | 16 Jul 2025 13:23 | CG/JU | Ok |
| 4 | PB168832BS | BP025157.D | 16 Jul 2025 14:05 | CG/JU | Ok,M |
| 5 | Q2586-01 | BP025158.D | 16 Jul 2025 14:49 | CG/JU | Ok |
| 6 | Q2586-01MS | BP025159.D | 16 Jul 2025 15:31 | CG/JU | Ok,M |
| 7 | Q2586-01MSD | BP025160.D | 16 Jul 2025 16:12 | CG/JU | Ok,M |
| 8 | Q2600-05 | BP025161.D | 16 Jul 2025 16:54 | CG/JU | Ok |
| 9 | Q2529-02 | BP025162.D | 16 Jul 2025 17:36 | CG/JU | Ok,M |
| 10 | Q2600-01 | BP025163.D | 16 Jul 2025 18:17 | CG/JU | Ok,M |
| 11 | Q2600-09 | BP025164.D | 16 Jul 2025 18:59 | CG/JU | Ok,M |
| 12 | Q2598-01 | BP025165.D | 16 Jul 2025 19:40 | CG/JU | Dilution |
| 13 | Q2592-01 | BP025166.D | 16 Jul 2025 20:22 | CG/JU | Ok,M |
| 14 | Q2589-01 | BP025167.D | 16 Jul 2025 21:04 | CG/JU | Ok |
| 15 | Q2565-04 | BP025168.D | 16 Jul 2025 21:45 | CG/JU | Ok |
| 16 | Q2611-01 | BP025169.D | 16 Jul 2025 22:26 | CG/JU | Ok,M |

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Instrument ID: BNA_P**

## Daily Analysis Runlog For Sequence/QCBatch ID # BP071825

| | | | | | |
|---|---|---|---|---|---|
| Review By | Rahul | Review On | 7/21/2025 8:38:03 AM | | |
| Supervise By | Jagrut | Supervise On | 7/21/2025 9:30:13 AM | | |
| SubDirectory | BP071825 | HP Acquire Method | BNA_P | HP Processing Method | bp070325 |

| STD. NAME | STD REF.# |
|---|---|
| Tune/Reschk | SP6757 |
| Initial Calibration Stds | SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840 |
| CCC | SP6836 |
| Internal Standard/PEM | S12674,10ul/1000ul sample |
| ICV/I.BLK | SP6770 |
| Surrogate Standard | |
| MS/MSD Standard | |
| LCS Standard | |

| Sr# | SampleId | Data File Name | Date-Time | Operator | Status |
|---|---|---|---|---|---|
| 1 | DFTPP | BP025186.D | 18 Jul 2025 09:41 | CG/JU | Ok |
| 2 | SSTDCCC040 | BP025187.D | 18 Jul 2025 11:04 | CG/JU | Ok,M |
| 3 | PB168899BL | BP025188.D | 18 Jul 2025 11:45 | CG/JU | Ok |
| 4 | PB168899BS | BP025189.D | 18 Jul 2025 12:27 | CG/JU | Ok |
| 5 | Q2614-01 | BP025190.D | 18 Jul 2025 13:11 | CG/JU | Ok,M |
| 6 | Q2597-01MS | BP025191.D | 18 Jul 2025 13:52 | CG/JU | Ok,M |
| 7 | Q2597-01MSD | BP025192.D | 18 Jul 2025 14:33 | CG/JU | Ok,M |
| 8 | Q2621-01 | BP025193.D | 18 Jul 2025 15:14 | CG/JU | Ok |

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Instrument ID:** BNA_F

## Daily Analysis Runlog For Sequence/QCBatch ID # BF071525

| | | | | | |
|---|---|---|---|---|---|
| Review By | Rahul | Review On | 7/16/2025 11:47:09 AM | | |
| Supervise By | Jagrut | Supervise On | 7/16/2025 1:41:28 PM | | |
| SubDirectory | BF071525 | HP Acquire Method | BNA_F | HP Processing Method | Bf071525 |

| STD. NAME | STD REF.# |
|---|---|
| Tune/Reschk | SP6757 |
| Initial Calibration Stds | SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840 |
| CCC | SP6836 |
| Internal Standard/PEM | S12674,10ul/1000ul sample |
| ICV/I.BLK | SP6770 |
| Surrogate Standard | |
| MS/MSD Standard | |
| LCS Standard | |

| Sr# | SampleId | ClientID | Data File Name | Date-Time | Comment | Operator | Status |
|---|---|---|---|---|---|---|---|
| 1 | DFTPP | DFTPP | BF143092.D | 15 Jul 2025 12:35 | | RC/JU | Ok |
| 2 | SSTDICC2.5 | SSTDICC2.5 | BF143093.D | 15 Jul 2025 13:04 | | RC/JU | Ok |
| 3 | SSTDICC005 | SSTDICC005 | BF143094.D | 15 Jul 2025 13:35 | | RC/JU | Ok,M |
| 4 | SSTDICC010 | SSTDICC010 | BF143095.D | 15 Jul 2025 14:05 | | RC/JU | Ok,M |
| 5 | SSTDICC020 | SSTDICC020 | BF143096.D | 15 Jul 2025 14:36 | Compound#26,32,48,51,57,80,84,85 Kept on LR | RC/JU | Ok,M |
| 6 | SSTDICCC040 | SSTDICCC040 | BF143097.D | 15 Jul 2025 15:05 | Compound#54,65 Kept on QR | RC/JU | Ok,M |
| 7 | SSTDICC050 | SSTDICC050 | BF143098.D | 15 Jul 2025 15:35 | | RC/JU | Ok,M |
| 8 | SSTDICC060 | SSTDICC060 | BF143099.D | 15 Jul 2025 16:05 | | RC/JU | Ok,M |
| 9 | SSTDICC080 | SSTDICC080 | BF143100.D | 15 Jul 2025 16:35 | | RC/JU | Ok,M |
| 10 | SSTDICV040 | ICVBF071525 | BF143101.D | 15 Jul 2025 17:27 | | RC/JU | Ok,M |
| 11 | PB168737BL | PB168737BL | BF143102.D | 15 Jul 2025 17:57 | | RC/JU | Ok |

M : Manual Integration

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Instrument ID:** BNA_F

## Daily Analysis Runlog For Sequence/QCBatch ID # BF071625

| | | | | | |
|---|---|---|---|---|---|
| Review By | Rahul | Review On | 7/17/2025 11:19:37 AM | | |
| Supervise By | Jagrut | Supervise On | 7/17/2025 11:19:59 AM | | |
| SubDirectory | BF071625 | HP Acquire Method | BNA_F | HP Processing Method | Bf071525 |

| STD. NAME | STD REF.# |
|---|---|
| Tune/Reschk | SP6757 |
| Initial Calibration Stds | SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840 |
| CCC | SP6836 |
| Internal Standard/PEM | S12674,10ul/1000ul sample |
| ICV/I.BLK | SP6770 |
| Surrogate Standard | |
| MS/MSD Standard | |
| LCS Standard | |

| Sr# | SampleId | ClientID | Data File Name | Date-Time | Comment | Operator | Status |
|---|---|---|---|---|---|---|---|
| 1 | DFTPP | DFTPP | BF143103.D | 16 Jul 2025 09:06 | | RC/JU | Ok |
| 2 | SSTDCCC040 | SSTDCCC040 | BF143104.D | 16 Jul 2025 10:04 | | RC/JU | Ok |
| 3 | PB168787BL | PB168787BL | BF143105.D | 16 Jul 2025 10:34 | | RC/JU | Ok |
| 4 | Q2126-02 | LOQ-SOIL-02-QT2-202 | BF143106.D | 16 Jul 2025 11:03 | LOQ-SOIL-5 PPM | RC/JU | Ok |
| 5 | Q2126-02 | LOQ-SOIL-02-QT2-202 | BF143107.D | 16 Jul 2025 11:37 | LOQ-SOIL-10 PPM | RC/JU | Ok |
| 6 | Q2126-08 | LOQ-WATER-02-QT2-2 | BF143108.D | 16 Jul 2025 12:07 | LOQ-WATER-5 PPM | RC/JU | Ok |
| 7 | Q2126-08 | LOQ-WATER-02-QT2-2 | BF143109.D | 16 Jul 2025 12:36 | LOQ-WATER-10 PPM | RC/JU | Ok |
| 8 | Q2126-01 | LOD-MDL-SOIL-03-QT | BF143110.D | 16 Jul 2025 13:05 | LOD-MDL-SOIL-4 PPM | RC/JU | Ok |
| 9 | Q2126-01 | LOD-MDL-SOIL-03-QT | BF143111.D | 16 Jul 2025 13:35 | LOD-MDL-SOIL-8 PPM | RC/JU | Ok |
| 10 | Q2126-07 | LOD-MDL-WATER-01-( | BF143112.D | 16 Jul 2025 14:05 | LOD-MDL-WATER-4 PPM | RC/JU | Ok |
| 11 | Q2126-07 | LOD-MDL-WATER-01-( | BF143113.D | 16 Jul 2025 14:35 | LOD-MDL-WATER-8 PPM | RC/JU | Ok |
| 12 | PB168802BL | PB168802BL | BF143114.D | 16 Jul 2025 15:13 | | RC/JU | Ok |
| 13 | SSTDCCC040 | SSTDCCC040EC | BF143115.D | 16 Jul 2025 15:45 | | RC/JU | Ok,M |
| 14 | DFTPP | DFTPP | BF143116.D | 16 Jul 2025 19:09 | | RC/JU | Ok |
| 15 | SSTDCCC040 | SSTDCCC040 | BF143117.D | 16 Jul 2025 19:39 | | RC/JU | Ok |
| 16 | PB168854BL | PB168854BL | BF143118.D | 16 Jul 2025 20:09 | | RC/JU | Ok |
| 17 | PB168854BS | PB168854BS | BF143119.D | 16 Jul 2025 20:39 | | RC/JU | Ok,M |
| 18 | PB168847TB | PB168847TB | BF143120.D | 16 Jul 2025 21:09 | | RC/JU | Ok |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Instrument ID:** BNA_F

## Daily Analysis Runlog For Sequence/QCBatch ID # BF071625

| | | | | | |
|---|---|---|---|---|---|
| Review By | Rahul | Review On | 7/17/2025 11:19:37 AM | | |
| Supervise By | Jagrut | Supervise On | 7/17/2025 11:19:59 AM | | |
| SubDirectory | BF071625 | HP Acquire Method | BNA_F | HP Processing Method | Bf071525 |

| STD. NAME | STD REF.# |
|---|---|
| Tune/Reschk | SP6757 |
| Initial Calibration Stds | SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840 |
| CCC | SP6836 |
| Internal Standard/PEM | S12674,10ul/1000ul sample |
| ICV/I.BLK | SP6770 |
| Surrogate Standard | |
| MS/MSD Standard | |
| LCS Standard | |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 19 | Q2592-02 | WC-SOIL-20250711 | BF143121.D | 16 Jul 2025 21:39 | | RC/JU | Ok |
| 20 | Q2592-02MS | WC-SOIL-20250711MS | BF143122.D | 16 Jul 2025 22:09 | | RC/JU | Ok,M |
| 21 | Q2592-02MSD | WC-SOIL-20250711MS | BF143123.D | 16 Jul 2025 22:38 | | RC/JU | Ok,M |
| 22 | Q2600-02 | TRENCH | BF143124.D | 16 Jul 2025 23:08 | | RC/JU | Ok |
| 23 | Q2600-06 | STOCK-PILE | BF143125.D | 16 Jul 2025 23:37 | | RC/JU | Ok |
| 24 | Q2600-10 | END-OF-TRENCH | BF143126.D | 17 Jul 2025 00:07 | | RC/JU | Ok,M |
| 25 | Q2614-06 | HR-MCN-COMP-01 | BF143127.D | 17 Jul 2025 00:37 | | RC/JU | Ok |
| 26 | Q2605-03 | VB15061 | BF143128.D | 17 Jul 2025 01:05 | | RC/JU | Ok |
| 27 | Q2605-04 | V897 | BF143129.D | 17 Jul 2025 01:35 | | RC/JU | Ok |
| 28 | Q2571-05 | TP-17 | BF143130.D | 17 Jul 2025 02:04 | | RC/JU | Ok |
| 29 | Q2571-05MS | TP-17MS | BF143131.D | 17 Jul 2025 02:34 | | RC/JU | Ok |
| 30 | Q2571-05MSD | TP-17MSD | BF143132.D | 17 Jul 2025 03:03 | | RC/JU | Ok |
| 31 | Q2555-01 | OU4-TS-29-070925 | BF143133.D | 17 Jul 2025 03:32 | | RC/JU | Ok,M |
| 32 | Q2558-01 | OU4-TS-Denali-070925 | BF143134.D | 17 Jul 2025 04:02 | | RC/JU | Ok,M |
| 33 | Q2555-03 | OU4-TS-30-070925 | BF143135.D | 17 Jul 2025 04:31 | | RC/JU | Ok,M |
| 34 | Q2558-03 | OU4-TS-Grillo-OG-070 | BF143136.D | 17 Jul 2025 05:00 | | RC/JU | Ok,M |
| 35 | SSTDCCC040 | SSTDCCC040EC | BF143137.D | 17 Jul 2025 05:30 | | RC/JU | Ok |

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Instrument ID:** BNA_P

## Daily Analysis Runlog For Sequence/QCBatch ID # BP070325

| | | | | | |
|---|---|---|---|---|---|
| Review By | Rahul | Review On | 7/7/2025 11:36:11 AM | | |
| Supervise By | Jagrut | Supervise On | 7/7/2025 12:15:03 PM | | |
| SubDirectory | BP070325 | HP Acquire Method | BNA_P | HP Processing Method | bp070325 |

| STD. NAME | STD REF.# |
|---|---|
| Tune/Reschk | SP6757 |
| Initial Calibration Stds | SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840 |
| CCC | SP6836 |
| Internal Standard/PEM | S12672,10ul/1000ul sample |
| ICV/I.BLK | SP6770 |
| Surrogate Standard | |
| MS/MSD Standard | |
| LCS Standard | |

| Sr# | SampleId | ClientID | Data File Name | Date-Time | Comment | Operator | Status |
|---|---|---|---|---|---|---|---|
| 1 | DFTPP | DFTPP | BP025070.D | 03 Jul 2025 08:58 | | CG/JU | Ok |
| 2 | SSTDICC2.5 | SSTDICC2.5 | BP025071.D | 03 Jul 2025 09:39 | | CG/JU | Ok |
| 3 | SSTDICC005 | SSTDICC005 | BP025072.D | 03 Jul 2025 10:21 | | CG/JU | Ok,M |
| 4 | SSTDICC010 | SSTDICC010 | BP025073.D | 03 Jul 2025 11:02 | | CG/JU | Ok,M |
| 5 | SSTDICC020 | SSTDICC020 | BP025074.D | 03 Jul 2025 11:44 | | CG/JU | Ok,M |
| 6 | SSTDICCC040 | SSTDICCC040 | BP025075.D | 03 Jul 2025 12:25 | compound #77 failed in the calibration. | CG/JU | Ok,M |
| 7 | SSTDICC050 | SSTDICC050 | BP025076.D | 03 Jul 2025 13:07 | Compound#26,32,48,54,57,65,80,85 Kept on LR | CG/JU | Ok,M |
| 8 | SSTDICC060 | SSTDICC060 | BP025077.D | 03 Jul 2025 13:49 | | CG/JU | Ok,M |
| 9 | SSTDICC080 | SSTDICC080 | BP025078.D | 03 Jul 2025 14:31 | Compound#09 removed from 80 ppm | CG/JU | Ok,M |
| 10 | SSTDICV040 | ICVBP070325 | BP025079.D | 03 Jul 2025 16:18 | | CG/JU | Ok,M |
| 11 | PB168694BL | PB168694BL | BP025080.D | 03 Jul 2025 16:59 | Analyzed for contamination check | CG/JU | Not Ok |

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Instrument ID:** BNA_P

## Daily Analysis Runlog For Sequence/QCBatch ID # BP071625

| | | | | | |
|---|---|---|---|---|---|
| Review By | Rahul | Review On | 7/17/2025 11:25:01 AM | | |
| Supervise By | Jagrut | Supervise On | 7/17/2025 11:34:34 AM | | |
| SubDirectory | BP071625 | HP Acquire Method | BNA_P | HP Processing Method | bp070325 |

| STD. NAME | STD REF.# |
|---|---|
| Tune/Reschk | SP6757 |
| Initial Calibration Stds | SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840 |
| CCC | SP6836 |
| Internal Standard/PEM | S12673,10ul/1000ul sample |
| ICV/I.BLK | SP6770 |
| Surrogate Standard | |
| MS/MSD Standard | |
| LCS Standard | |

| Sr# | SampleId | ClientID | Data File Name | Date-Time | Comment | Operator | Status |
|---|---|---|---|---|---|---|---|
| 1 | DFTPP | DFTPP | BP025154.D | 16 Jul 2025 11:19 | | CG/JU | Ok |
| 2 | SSTDCCC040 | SSTDCCC040 | BP025155.D | 16 Jul 2025 12:42 | | CG/JU | Ok,M |
| 3 | PB168832BL | PB168832BL | BP025156.D | 16 Jul 2025 13:23 | | CG/JU | Ok |
| 4 | PB168832BS | PB168832BS | BP025157.D | 16 Jul 2025 14:05 | | CG/JU | Ok,M |
| 5 | Q2586-01 | TP-16 | BP025158.D | 16 Jul 2025 14:49 | | CG/JU | Ok |
| 6 | Q2586-01MS | TP-16MS | BP025159.D | 16 Jul 2025 15:31 | | CG/JU | Ok,M |
| 7 | Q2586-01MSD | TP-16MSD | BP025160.D | 16 Jul 2025 16:12 | | CG/JU | Ok,M |
| 8 | Q2600-05 | STOCK-PILE | BP025161.D | 16 Jul 2025 16:54 | | CG/JU | Ok |
| 9 | Q2529-02 | TP-80 | BP025162.D | 16 Jul 2025 17:36 | | CG/JU | Ok,M |
| 10 | Q2600-01 | TRENCH | BP025163.D | 16 Jul 2025 18:17 | | CG/JU | Ok,M |
| 11 | Q2600-09 | END-OF-TRENCH | BP025164.D | 16 Jul 2025 18:59 | | CG/JU | Ok,M |
| 12 | Q2598-01 | OK-01-071425 | BP025165.D | 16 Jul 2025 19:40 | Need 5X Dilution | CG/JU | Dilution |
| 13 | Q2592-01 | WC-SOIL-20250711 | BP025166.D | 16 Jul 2025 20:22 | | CG/JU | Ok,M |
| 14 | Q2589-01 | AU-06-071125 | BP025167.D | 16 Jul 2025 21:04 | | CG/JU | Ok |
| 15 | Q2565-04 | MOO-25-0196 | BP025168.D | 16 Jul 2025 21:45 | | CG/JU | Ok |
| 16 | Q2611-01 | EO-02-071525 | BP025169.D | 16 Jul 2025 22:26 | | CG/JU | Ok,M |

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

**Instrument ID:** BNA_P

## Daily Analysis Runlog For Sequence/QCBatch ID # BP071825

| | | | | | |
|---|---|---|---|---|---|
| Review By | Rahul | Review On | 7/21/2025 8:38:03 AM | | |
| Supervise By | Jagrut | Supervise On | 7/21/2025 9:30:13 AM | | |
| SubDirectory | BP071825 | HP Acquire Method | BNA_P | HP Processing Method | bp070325 |

| STD. NAME | STD REF.# |
|---|---|
| Tune/Reschk | SP6757 |
| Initial Calibration Stds | SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840 |
| CCC | SP6836 |
| Internal Standard/PEM | S12674,10ul/1000ul sample |
| ICV/I.BLK | SP6770 |
| Surrogate Standard | |
| MS/MSD Standard | |
| LCS Standard | |

| Sr# | SampleId | ClientID | Data File Name | Date-Time | Comment | Operator | Status |
|---|---|---|---|---|---|---|---|
| 1 | DFTPP | DFTPP | BP025186.D | 18 Jul 2025 09:41 | | CG/JU | Ok |
| 2 | SSTDCCC040 | SSTDCCC040 | BP025187.D | 18 Jul 2025 11:04 | | CG/JU | Ok,M |
| 3 | PB168899BL | PB168899BL | BP025188.D | 18 Jul 2025 11:45 | | CG/JU | Ok |
| 4 | PB168899BS | PB168899BS | BP025189.D | 18 Jul 2025 12:27 | | CG/JU | Ok |
| 5 | Q2614-01 | HR-MCN-COMP-01 | BP025190.D | 18 Jul 2025 13:11 | | CG/JU | Ok,M |
| 6 | Q2597-01MS | CHRT-26899MS | BP025191.D | 18 Jul 2025 13:52 | | CG/JU | Ok,M |
| 7 | Q2597-01MSD | CHRT-26899MSD | BP025192.D | 18 Jul 2025 14:33 | | CG/JU | Ok,M |
| 8 | Q2621-01 | TR-05-07162025 | BP025193.D | 18 Jul 2025 15:14 | | CG/JU | Ok |

M : Manual Integration

# PERCENT SOLID

**Supervisor:** Iwona
**Analyst:** jignesh
**Date:** 7/16/2025

**OVENTEMP IN Celsius(°C):** 107
**Time IN:** 17:15
**In Date:** 07/15/2025
**Weight Check 1.0g:** 1.00
**Weight Check 10g:** 10.00
**OvenID:** M OVEN#1

**OVENTEMP OUT Celsius(°C):** 104
**Time OUT:** 08:25
**Out Date:** 07/16/2025
**Weight Check 1.0g:** 1.00
**Weight Check 10g:** 10.00
**BalanceID:** M SC-4
**Thermometer ID:** % SOLID-OVEN

**QC:** LB136481

| Lab ID | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|---|---|---|---|---|---|---|---|---|
| Q2592-01 | WC-SOIL-20250711 | 1 | 1.15 | 10.84 | 11.99 | 9.46 | 76.7 | |
| Q2605-01 | V908 | 2 | 1.13 | 10.70 | 11.83 | 11.13 | 93.5 | |
| Q2605-02 | VB16135 | 3 | 1.14 | 10.66 | 11.8 | 11.09 | 93.3 | |
| Q2605-03 | VB15061 | 4 | 1.18 | 10.55 | 11.73 | 10.99 | 93.0 | |
| Q2605-04 | V897 | 5 | 1.17 | 10.79 | 11.96 | 11.76 | 98.1 | |
| Q2607-01 | VNJ-257-1 | 6 | 1.00 | 1.00 | 2.00 | 2.00 | 100.0 | WIPE SAMPLE |
| Q2607-02 | VNJ-257-2 | 7 | 1.00 | 1.00 | 2.00 | 2.00 | 100.0 | WIPE SAMPLE |
| Q2607-03 | ETGI-359-1 | 8 | 1.00 | 1.00 | 2.00 | 2.00 | 100.0 | WIPE SAMPLE |
| Q2607-04 | ETGI-359-2 | 9 | 1.00 | 1.00 | 2.00 | 2.00 | 100.0 | WIPE SAMPLE |
| Q2608-04 | 60271 | 10 | 1.14 | 10.65 | 11.79 | 9.61 | 79.5 | |
| Q2609-01 | 710-ABC | 11 | 1.18 | 10.37 | 11.55 | 11.5 | 99.5 | |
| Q2609-03 | 710-C | 12 | 1.13 | 10.29 | 11.42 | 11.37 | 99.5 | |
| Q2609-05 | 709-AB | 13 | 1.16 | 10.26 | 11.42 | 11.39 | 99.7 | |
| Q2609-07 | 709-A | 14 | 1.14 | 10.28 | 11.42 | 11.4 | 99.8 | |
| Q2610-01 | 2010 | 15 | 1.00 | 1.00 | 2.00 | 2.00 | 100.0 | WIPE SAMPLE |
| Q2610-02 | 2011 | 16 | 1.00 | 1.00 | 2.00 | 2.00 | 100.0 | WIPE SAMPLE |
| Q2610-03 | 2012 | 17 | 1.00 | 1.00 | 2.00 | 2.00 | 100.0 | WIPE SAMPLE |
| Q2610-04 | 2013 | 18 | 1.00 | 1.00 | 2.00 | 2.00 | 100.0 | WIPE SAMPLE |
| Q2610-05 | 2014 | 19 | 1.00 | 1.00 | 2.00 | 2.00 | 100.0 | WIPE SAMPLE |
| Q2611-01 | EO-02-071525 | 20 | 1.16 | 10.17 | 11.33 | 10.83 | 95.1 | |
| Q2611-02 | EO-02-071525-E2 | 21 | 1.14 | 10.39 | 11.53 | 10.68 | 91.8 | |
| Q2612-01 | OR-02-071525 | 22 | 1.18 | 10.49 | 11.67 | 10.1 | 85.0 | |
| Q2612-02 | OR-02-071525-E2 | 23 | 1.12 | 10.61 | 11.73 | 9.89 | 82.7 | |
| Q2614-01 | HR-MCN-COMP-01 | 24 | 1.15 | 10.70 | 11.85 | 9.61 | 79.1 | |
| Q2614-02 | HR-MCN-VOC-01 | 25 | 1.15 | 10.29 | 11.44 | 9.14 | 77.6 | |
| Q2614-03 | HR-MCN-01 | 26 | 1.11 | 10.71 | 11.82 | 9.35 | 76.9 | |
| Q2614-04 | HR-MCN-02 | 27 | 1.19 | 10.40 | 11.59 | 8.94 | 74.5 | |
| Q2614-05 | HR-MCN-03 | 28 | 1.14 | 10.58 | 11.72 | 9.89 | 82.7 | |

# PERCENT SOLID

**Supervisor:** Iwona
**Analyst:** jignesh
**Date:** 7/16/2025

**OVENTEMP IN Celsius(°C):** 107
**Time IN:** 17:15
**In Date:** 07/15/2025
**Weight Check 1.0g:** 1.00
**Weight Check 10g:** 10.00
**OvenID:** M OVEN#1

**OVENTEMP OUT Celsius(°C):** 104
**Time OUT:** 08:25
**Out Date:** 07/16/2025
**Weight Check 1.0g:** 1.00
**Weight Check 10g:** 10.00
**BalanceID:** M SC-4
**Thermometer ID:** % SOLID-OVEN

**QC:** LB136481

| Lab ID | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|---|---|---|---|---|---|---|---|---|

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

Jb 136481

**WorkList Name :** %1-071525 **WorkList ID :** 190721 **Department :** Wet-Chemistry **Date :** 07-15-2025 08:40:15

| Sample | Customer Sample | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|---|---|---|---|---|---|---|---|---|
| Q2592-01 | WC-SOIL-20250711 | Solid | Percent Solids | Cool 4 deg C | PARS02 | D51 | 07/11/2025 | Chemtech -SO |
| Q2605-01 | V908 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2605-02 | VB16135 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2605-03 | VB15061 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2605-04 | V897 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2607-01 | VNJ-257-1 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2607-02 | VNJ-257-2 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2607-03 | ETGI-359-1 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2607-04 | ETGI-359-2 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2608-04 | 60271 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2609-01 | 710-ABC | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2609-03 | 710-C | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2609-05 | 709-ABC | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2609-07 | 709-C | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2610-01 | 2010 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2610-02 | 2011 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2610-03 | 2012 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2610-04 | 2013 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2610-05 | 2014 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2611-01 | EO-02-071525 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2611-02 | EO-02-071525-E2 | Solid | Percent Solids | Cool 4 deg C | PSEG05 | D41 | 07/15/2025 | Chemtech -SO |

**Date/Time** 07.15.25 15:15

**Raw Sample Received by:** [signature]

**Raw Sample Relinquished by:** [signature] SM

**Date/Time** 07-15-25 17:30

**Raw Sample Received by:** [signature] SM

**Raw Sample Relinquished by:** [signature]

# WORKLIST(Hardcopy Internal Chain)

VB 136481

**WorkList Name :** %1-071525 **WorkList ID :** 190721 **Department :** Wet-Chemistry **Date :** 07-15-2025 08:40:15

| Sample | Customer Sample | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|---|---|---|---|---|---|---|---|---|
| Q2612-01 | OR-02-071525 | Solid | Percent Solids | Cool 4 deg C | PSEG05 | D41 | 07/15/2025 | Chemtech -SO |
| Q2612-02 | OR-02-071525-E2 | Solid | Percent Solids | Cool 4 deg C | PSEG05 | D41 | 07/15/2025 | Chemtech -SO |
| Q2614-01 | HR-MCN-COMP-01 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2614-02 | HR-MCN-VOC-01 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2614-03 | HR-MCN-01 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2614-04 | HR-MCN-02 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |
| Q2614-05 | HR-MCN-03 | Solid | Percent Solids | Cool 4 deg C | PSEG03 | D41 | 07/15/2025 | Chemtech -SO |

**Date/Time** 07.15.25 15:15

**Raw Sample Received by:** [signature]

**Raw Sample Relinquished by:** [signature] SM

**Date/Time** 07.15.25 17:30

**Raw Sample Received by:** [signature] SM

**Raw Sample Relinquished by:** [signature]

Alliance TECHNICAL GROUP

# EXTRACTION LOGPAGE

PB168854

| | | | | | |
|---|---|---|---|---|---|
| **SOP ID:** | M3541-ASE Extraction-14 | | | | |
| **Clean Up SOP #:** | N/A | | | **Extraction Start Date :** | 07/15/2025 |
| **Matrix :** | Solid | | | **Extraction Start Time :** | 09:05 |
| **Weigh By:** | EH | **Extraction By:** | RJ | **Extraction End Date :** | 07/15/2025 |
| **Balance check:** | RJ | **Filter By:** | RJ | **Extraction End Time :** | 12:25 |
| **Balance ID:** | EX-SC-2 | **pH Meter ID:** | N/A | **Concentration By:** | EH |
| **pH Strip Lot#:** | N/A | **Hood ID:** | 3,7 | **Supervisor By :** | RUPESH |

**Extraction Method:** ☐ Seperatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☑ Soxhlet

| Standared Name | MLS USED | Concentration ug/mL | STD REF. # FROM LOG |
|---|---|---|---|
| Spike Sol 1 | 1.0ML | 50/100 PPM | SP6849 |
| Surrogate | 1.0ML | 100/150 PPM | SP6852 |
| N/A | N/A | N/A | N/A |
| N/A | N/A | N/A | N/A |
| N/A | N/A | N/A | N/A |

| Chemical Used | ML/SAMPLE USED | Lot Number |
|---|---|---|
| MeCl2/Acetone/1:1 | N/A | EP2612 |
| Baked Na2SO4 | N/A | EP2624 |
| Sand | N/A | E3951 |
| Methylene Chloride | N/A | E3954 |
| N/A | N/A | N/A |
| N/A | N/A | N/A |
| N/A | N/A | N/A |
| N/A | N/A | N/A |
| N/A | N/A | N/A |
| N/A | N/A | N/A |
| N/A | N/A | N/A |
| N/A | N/A | N/A |
| N/A | N/A | N/A |
| N/A | N/A | N/A |

**Extraction Conformance/Non-Conformance Comments:**

1.5ML Vail Lot # 2210443.

| | | | |
|---|---|---|---|
| **KD Bath ID:** | N/A | **Envap ID:** | NEVAP-02 |
| **KD Bath Temperature:** | N/A | **Envap Temperature:** | 40 °C |

| Date / Time | Prepped Sample Relinquished By/Location | Received By/Location |
|---|---|---|
| 7/15/25 | RS (Ext lab) | RC/SVOC |
| 12:30 | **Preparation Group** | **Analysis Group** |

Alliance TECHNICAL GROUP

**EXTRACTION LOGPAGE**

**PrepBatch ID : PB168854**

**Analytical Method:** M3541-ASE Extraction-14 **Concentration Date:** 07/15/2025

| Sample ID | Client Sample ID | Test | g/mL | PH | Surr/Spike By: | | Final Vol. (mL) | JarID | Comments | Prep Pos |
|---|---|---|---|---|---|---|---|---|---|---|
| | | | | | AddedBy | VerifiedBy | | | | |
| PB168854BL | SBLK854 | SVOC-TCL BNA -20 | 30.03 | N/A | ritesh | Evelyn | 1 | | | U2-1 |
| PB168854BS | SLCS854 | SVOC-TCL BNA -20 | 30.01 | N/A | ritesh | Evelyn | 1 | | | 2 |
| Q2592-01 | WC-SOIL-20250711 | SVOCMS Group1 | 30.07 | N/A | ritesh | Evelyn | 1 | E | | 3 |
| Q2595-01 | CHRT-24849 | SVOC-TCL BNA -20 | 50.04 | N/A | ritesh | Evelyn | 1 | D | Gel Mat. | 4 |
| Q2597-01 | CHRT-26899 | SVOC-TCL BNA -20 | 50.02 | N/A | ritesh | Evelyn | 1 | D | | 5 |
| Q2597-01MS | CHRT-26899MS | SVOC-TCL BNA -20 | 50.06 | N/A | ritesh | Evelyn | 1 | D | | 6 |
| Q2597-01MSD | CHRT-26899MSD | SVOC-TCL BNA -20 | 50.08 | N/A | ritesh | Evelyn | 1 | D | | U3-1 |
| Q2598-01 | OK-01-071425 | SVOC-TCL BNA -20 | 50.05 | N/A | ritesh | Evelyn | 1 | E | | 2 |
| Q2600-01 | TRENCH | SVOC-TCL BNA -20 | 30.08 | N/A | ritesh | Evelyn | 1 | D | | 3 |
| Q2600-05 | STOCK-PILE | SVOC-TCL BNA -20 | 30.06 | N/A | ritesh | Evelyn | 1 | D | | 4 |
| Q2600-09 | END-OF-TRENCH | SVOC-TCL BNA -20 | 30.03 | N/A | ritesh | Evelyn | 1 | D | | 5 |

RS 7/15

* **Extracts relinquished on the same date as received.**

168654
9:05

# WORKLIST(Hardcopy Internal Chain)

**WorkList Name :** Q2592 **WorkList ID :** 190724 **Department :** Extraction **Date :** 07-15-2025 08:50:21

| Sample | Customer Sample | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|---|---|---|---|---|---|---|---|---|
| Q2592-01 | WC-SOIL-20250711 | Solid | SVOCMS Group1 | Cool 4 deg C | PARS02 | D51 | 07/11/2025 | 8270E |
| Q2595-01 | CHRT-24849 | Solid | SVOC-TCL BNA -20 | Cool 4 deg C | PSEG03 | D41 | 07/14/2025 | 8270E |
| Q2597-01 | CHRT-26899 | Solid | SVOC-TCL BNA -20 | Cool 4 deg C | PSEG03 | D31 | 07/14/2025 | 8270E |
| Q2598-01 | OK-01-071425 | Solid | SVOC-TCL BNA -20 | Cool 4 deg C | PSEG05 | D41 | 07/14/2025 | 8270E |
| Q2600-01 | TRENCH | Solid | SVOC-TCL BNA -20 | Cool 4 deg C | TACO01 | D41 | 07/14/2025 | 8270E |
| Q2600-05 | STOCK-PILE | Solid | SVOC-TCL BNA -20 | Cool 4 deg C | TACO01 | D41 | 07/14/2025 | 8270E |
| Q2600-09 | END-OF-TRENCH | Solid | SVOC-TCL BNA -20 | Cool 4 deg C | TACO01 | D41 | 07/14/2025 | 8270E |

**Date/Time** 07/15/25 9:00
**Raw Sample Received by:** RJ (CET-1446)
**Raw Sample Relinquished by:**

**Date/Time** 07/15/25 9:15
**Raw Sample Received by:**
**Raw Sample Relinquished by:** RJ (CET-1446)

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

# Prep Standard - Chemical Standard Summary

**Order ID :** Q2592

**Test :** SVOCMS Group1

**Prepbatch ID :** PB168854,

**Sequence ID/Qc Batch ID:** BF071625,BP071625,BP071825,

**Standard ID :**
EP2612,EP2624,SP6757,SP6769,SP6770,SP6832,SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840,SP6849,SP6852,

**Chemical ID :**
10ul/1000ul sample,E3551,E3904,E3926,E3930,E3932,E3942,E3946,E3951,E3954,S10105,S11484,S11496,S11651,S11788,S11801,S11802,S11805,S11806,S11807,S12115,S12195,S12196,S12197,S12198,S12216,S12217,S12218,S12219,S12220,S12272,S12277,S12278,S12486,S12499,S12500,S12501,S12502,S12503,S12504,S12505,S12506,S12507,S12534,S12545,S12546,S12547,S12548,S12549,S12550,S12551,S12552,S12577,S12658,S12670,S12673,S12674,S12974,S12987,S12988,S12989,S12990,S12991,S12992,S13057,S13078,S13088,S13089,S13148,S13158,S13159,S13206,S13208,S13209,S13210,S13211,

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# Extractions STANDARD PREPARATION LOG

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 2017 | 1:1 ACETONE/METHYLENE CHLORIDE | EP2612 | 05/09/2025 | 11/05/2025 | RUPESHKUMAR SHAH | None | None | Riteshkumar Patel 05/09/2025 |

**FROM** 8000.00000ml of E3930 + 8000.00000ml of E3932 = Final Quantity: 16000.000 ml

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 3923 | Baked Sodium Sulfate | EP2624 | 06/26/2025 | 12/04/2025 | RUPESHKUMAR SHAH | Extraction_SCALE_2 (EX-SC-2) | None | Riteshkumar Patel 06/26/2025 |

**FROM** 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# SVOC STANDARD PREPARATION LOG

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 3895 | 50 ug/ml DFTPP 8270E | SP6757 | 03/31/2025 | 09/30/2025 | Rahul Chavli | None | None | Jagrut Upadhyay 04/01/2025 |

**FROM** 1.00000ml of S12577 + 19.00000ml of E3904 = Final Quantity: 20.000 ml

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 18 | Second Source Calibration Stock Standard, 100 PPM, (8270/625/CLP) | SP6769 | 04/10/2025 | 09/10/2025 | Jagrut Upadhyay | None | None | Sohil Jodhani 04/16/2025 |

**FROM** 0.04000ml of S12195 + 0.08000ml of S12216 + 0.10000ml of S11788 + 0.20000ml of S12486 + 0.20000ml of S12534 + 0.20000ml of S12974 + 1.18000ml of E3926 = Final Quantity: 2.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# SVOC STANDARD PREPARATION LOG

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 416 | 40 ng BNA ICV, 40 PPM | SP6770 | 04/10/2025 | 09/10/2025 | Jagrut Upadhyay | None | None | Sohil Jodhani 04/16/2025 |

**FROM** 0.01000ml of S12658 + 0.60000ml of E3926 + 0.40000ml of SP6769 = Final Quantity: 1.010 ml

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 4178 | 8270/625.1 Stock Solution-2 100 ng | SP6832 | 06/19/2025 | 10/28/2025 | Jagrut Upadhyay | None | None | Rahul Chavli 07/21/2025 |

**FROM** 0.20000ml of S13211 + 0.26700ml of S10105 + 0.40000ml of S11496 + 0.40000ml of S12278 + 0.50000ml of S12115 + 0.50000ml of S13209 + 0.50000ml of S13210 + 0.60000ml of S12277 + 1.00000ml of S12272 + 1.00000ml of S13057 + 1.00000ml of S13078 + 1.00000ml of S13208 + 2.63300ml of E3942 = Final Quantity: 10.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# SVOC STANDARD PREPARATION LOG

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 4179 | 80 ng BNA ICC | SP6833 | 06/19/2025 | 10/28/2025 | Jagrut Upadhyay | None | None | Rahul Chavli 07/21/2025 |

**FROM** 0.01000ml of S12670 + 0.20000ml of E3942 + 0.80000ml of SP6832 = Final Quantity: 1.010 ml

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 4180 | 60 ng BNA ICC | SP6834 | 06/19/2025 | 10/28/2025 | Jagrut Upadhyay | None | None | Rahul Chavli 07/21/2025 |

**FROM** 0.01000ml of S12670 + 0.40000ml of E3942 + 0.60000ml of SP6832 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# SVOC STANDARD PREPARATION LOG

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 4181 | 50 ng BNA ICC | SP6835 | 06/19/2025 | 10/28/2025 | Jagrut Upadhyay | None | None | Rahul Chavli 07/21/2025 |

**FROM** 0.01000ml of S12670 + 0.50000ml of E3942 + 0.50000ml of SP6832 = Final Quantity: 1.010 ml

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 4182 | 40 ng BNA ICC | SP6836 | 06/19/2025 | 10/28/2025 | Jagrut Upadhyay | None | None | Rahul Chavli 07/21/2025 |

**FROM** 0.01000ml of S12670 + 0.60000ml of E3942 + 0.40000ml of SP6832 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# SVOC STANDARD PREPARATION LOG

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 4183 | 20 ng BNA ICC | SP6837 | 06/19/2025 | 10/28/2025 | Jagrut Upadhyay | None | None | Rahul Chavli 07/21/2025 |

**FROM** 0.01000ml of S12670 + 0.80000ml of E3942 + 0.20000ml of SP6832 = Final Quantity: 1.010 ml

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 4184 | 10 ng BNA ICC | SP6838 | 06/19/2025 | 10/28/2025 | Jagrut Upadhyay | None | None | Rahul Chavli 07/21/2025 |

**FROM** 0.01000ml of S12670 + 0.90000ml of E3942 + 0.10000ml of SP6832 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# SVOC STANDARD PREPARATION LOG

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 4185 | 5 ng BNA ICC | SP6839 | 06/19/2025 | 10/28/2025 | Jagrut Upadhyay | None | None | Rahul Chavli 07/21/2025 |

**FROM** 0.01000ml of S12670 + 0.95000ml of E3942 + 0.05000ml of SP6832 = Final Quantity: 1.010 ml

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 4186 | 2.5 ng BNA ICC | SP6840 | 06/19/2025 | 10/28/2025 | Jagrut Upadhyay | None | None | Rahul Chavli 07/21/2025 |

**FROM** 0.01000ml of S12670 + 0.50000ml of E3942 + 0.50000ml of SP6839 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# SVOC STANDARD PREPARATION LOG

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 171 | 8270/625 Spike Solution, 50/100 PPM | SP6849 | 07/01/2025 | 09/30/2025 | Rahul Chavli | None | None | Jagrut Upadhyay 07/01/2025 |

**FROM** 0.10000ml of S12507 + 0.10000ml of S13159 + 0.30000ml of S13158 + 0.40000ml of S11484 + 0.40000ml of S11651 + 0.40000ml of S11806 + 0.40000ml of S11807 + 0.40000ml of S13148 + 0.40000ml of S13206 + 0.60000ml of S13089 + 0.80000ml of S12499 + 0.90000ml of S12552 + 1.20000ml of S12504 + 1.20000ml of S12506 + 1.20000ml of S12546 + 1.30000ml of S11802 + 1.30000ml of S12501 + 1.30000ml of S12503 + 1.30000ml of S12505 + 1.30000ml of S12545 + 1.30000ml of S12547 + 1.30000ml of S12549 + 1.30000ml of S12550 + 1.30000ml of S12551 + 1.30000ml of S12988 + 1.30000ml of S12989 + 1.30000ml of S12990 + 1.30000ml of S12992 + 1.40000ml of S11805 + 1.40000ml of S12500 + 1.40000ml of S12502 + 1.40000ml of S12548 + 1.40000ml of S12987 + 1.40000ml of S12991 + 1.40000ml of S13088 + 1.50000ml of S11801 + 163.00000ml of E3932 = Final Quantity: 200.000 ml

| Recipe ID | NAME | NO. | Prep Date | Expiration Date | Prepared By | ScaleID | PipetteID | Supervised By |
|---|---|---|---|---|---|---|---|---|
| 19 | 8270/CLP Surrogate Solution, 100 PPM BN/150 PPM ACID | SP6852 | 07/02/2025 | 09/18/2025 | Jagrut Upadhyay | None | None | Rahul Chavli 07/02/2025 |

**FROM** 1.50000ml of S12220 + 2.00000ml of S12195 + 2.00000ml of S12197 + 2.00000ml of S12216 + 5.50000ml of S12196 + 5.50000ml of S12198 + 5.50000ml of S12217 + 5.50000ml of S12218 + 5.50000ml of S12219 + 965.00000ml of E3946 = Final Quantity: 1000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| PCI Scientific Supply, Inc. | PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1 | 313201 | 12/04/2025 | 01/03/2024 / Rajesh | 07/20/2023 / Rajesh | E3551 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Seidler Chemical | BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L) | 24K1762005 | 01/07/2026 | 03/13/2025 / | 12/27/2024 / RUPESH | E3904 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Seidler Chemical | BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L) | 25A0262002 | 10/08/2025 | 04/08/2025 / Rajesh | 02/07/2025 / Rajesh | E3926 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Seidler Chemical | BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L) | 25A0262002 | 02/20/2026 | 05/02/2025 / RUPESH | 03/09/2025 / RUPESH | E3930 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Seidler Chemical | BA-9254-03 / Acetone, Ultra Resi (cs/4x4L) | 24H1462005 | 11/05/2025 | 05/05/2025 / RUPESH | 04/23/2025 / RUPESH | E3932 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Seidler Chemical | BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L) | 25A2862010 | 12/13/2025 | 06/13/2025 / Rajesh | 02/28/2025 / Rajesh | E3942 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Seidler Chemical | BA-9254-03 / Acetone, Ultra Resi (cs/4x4L) | 24H1462005 | 05/24/2027 | 07/01/2025 / RUPESH | 06/19/2025 / RUPESH | E3946 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Seidler Chemical | BA-3382-05 / Sand, Purified (cs/4x2.5kg) | 25A2756718 | 12/31/2028 | 07/09/2025 / RUPESH | 04/28/2020 / RUPESH | E3951 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Seidler Chemical | BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L) | 25B1862001 | 03/19/2026 | 07/14/2025 / RUPESH | 06/11/2025 / RUPESH | E3954 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| CPI International | Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml | 440246 | 12/19/2025 | 06/19/2025 / Jagrut | 12/09/2021 / Christian | S10105 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555870 / Custom Standard, 2,4-dinitrophenol Std [CS 5328-3] | A0200549 | 10/30/2025 | 04/30/2025 / Rahul | 08/10/2023 / yogesh | S11484 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| CPI International | Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml | 506889 | 10/28/2025 | 04/28/2025 / Jagrut | 08/11/2023 / Yogesh | S11496 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555872 / Custom Standard, pentachlorophenol Std [CS 5328-5] | A0201728 | 01/01/2026 | 07/01/2025 / Rahul | 11/09/2023 / Yogesh | S11651 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride | A0196453 | 09/10/2025 | 03/10/2025 / anahy | 11/21/2023 / Rahul | S11788 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride | A0200655 | 01/01/2026 | 07/01/2025 / Rahul | 11/21/2023 / rahul | S11801 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride | A0200655 | 01/01/2026 | 07/01/2025 / Rahul | 11/21/2023 / rahul | S11802 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride | A0200655 | 01/01/2026 | 07/01/2025 / Rahul | 11/21/2023 / rahul | S11805 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride | A0200655 | 12/04/2025 | 06/04/2025 / Jagrut | 11/21/2023 / rahul | S11806 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride | A0200655 | 01/01/2026 | 07/01/2025 / Rahul | 11/21/2023 / rahul | S11807 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| CPI International | z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml | 454157 | 10/28/2025 | 04/28/2025 / Jagrut | 03/08/2024 / Rahul | S12115 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul | A0206206 | 09/18/2025 | 03/18/2025 / anahy | 03/15/2024 / Rahul | S12195 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul | A0206206 | 01/02/2026 | 07/02/2025 / Jagrut | 03/15/2024 / Rahul | S12196 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul | A0206206 | 01/02/2026 | 07/02/2025 / Jagrut | 03/15/2024 / Rahul | S12197 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul | A0206206 | 01/02/2026 | 07/02/2025 / Jagrut | 03/15/2024 / Rahul | S12198 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0206381 | 09/18/2025 | 03/18/2025 / anahy | 03/15/2024 / Rahul | S12216 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0206381 | 01/02/2026 | 07/02/2025 / Jagrut | 03/15/2024 / Rahul | S12217 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0206381 | 01/02/2026 | 07/02/2025 / Jagrut | 03/15/2024 / Rahul | S12218 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0206381 | 01/02/2026 | 07/02/2025 / Jagrut | 03/15/2024 / Rahul | S12219 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml | A0206381 | 01/02/2026 | 07/02/2025 / Jagrut | 03/15/2024 / Rahul | S12220 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| CPI International | z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml | 520963 | 12/03/2025 | 06/03/2025 / Jagrut | 05/24/2024 / Rahul | S12272 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| CPI International | Z-010442-07 / Benzaldehyde Solution, 1000 mg/L, 1.3 ml, (Maximum Expiration: 90 Days) | 495833 | 11/05/2025 | 05/05/2025 / Jagrut | 05/24/2024 / Rahul | S12277 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| CPI International | Z-010442-07 / Benzaldehyde Solution, 1000 mg/L, 1.3 ml, (Maximum Expiration: 90 Days) | 495833 | 12/19/2025 | 06/19/2025 / Jagrut | 05/24/2024 / Rahul | S12278 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1] | A0214021 | 09/10/2025 | 03/10/2025 / anahy | 07/23/2024 / RAHUL | S12486 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1] | A0214021 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12499 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1] | A0214021 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12500 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1] | A0214021 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12501 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1] | A0214021 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12502 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1] | A0214021 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12503 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1] | A0214021 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12504 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1] | A0214021 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12505 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1] | A0214021 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12506 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1] | A0214021 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12507 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2] | A0214017 | 10/10/2025 | 04/10/2025 / Jagrut | 07/23/2024 / RAHUL | S12534 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2] | A0214017 | 12/04/2025 | 06/04/2025 / Jagrut | 07/23/2024 / RAHUL | S12545 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2] | A0214017 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12546 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2] | A0214017 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12547 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2] | A0214017 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12548 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2] | A0214017 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12549 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2] | A0214017 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12550 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2] | A0214017 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12551 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2] | A0214017 | 01/01/2026 | 07/01/2025 / Rahul | 07/23/2024 / RAHUL | S12552 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31615 / SV Mixture, GC/MS Tuning Mixture, CH2Cl2, 1mL, | A0212955 | 06/30/2027 | 03/31/2025 / Rahul | 08/01/2024 / Rahul | S12577 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL | A0212266 | 04/30/2030 | 04/07/2025 / anahy | 09/20/2024 / anahy | S12658 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL | A0212266 | 12/16/2025 | 06/16/2025 / anahy | 09/20/2024 / anahy | S12670 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL | A0212266 | 01/08/2026 | 07/08/2025 / Rahul | 09/20/2024 / anahy | S12673 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL | A0212266 | 01/15/2026 | 07/15/2025 / Rahul | 09/20/2024 / anahy | S12674 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0219438 | 09/10/2025 | 03/10/2025 / anahy | 12/11/2024 / anahy | S12974 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0219438 | 09/30/2025 | 07/01/2025 / Rahul | 12/11/2024 / anahy | S12987 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0219438 | 09/30/2025 | 07/01/2025 / Rahul | 12/11/2024 / anahy | S12988 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0219438 | 09/30/2025 | 07/01/2025 / Rahul | 12/11/2024 / anahy | S12989 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0219438 | 09/30/2025 | 07/01/2025 / Rahul | 12/11/2024 / anahy | S12990 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0219438 | 09/30/2025 | 07/01/2025 / Rahul | 12/11/2024 / anahy | S12991 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0219438 | 09/30/2025 | 07/01/2025 / Rahul | 12/11/2024 / anahy | S12992 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| CPI International | Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days) | 531243 | 12/19/2025 | 06/19/2025 / Jagrut | 01/16/2025 / anahy | S13057 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| CPI International | Z-010074-07 / 3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 ml, (Maximum Expiration: 180 days) | 525551 | 11/05/2025 | 05/05/2025 / Jagrut | 03/10/2025 / anahy | S13078 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0221014 | 11/30/2025 | 07/01/2025 / Rahul | 05/20/2025 / Rahul | S13088 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2] | A0221014 | 11/30/2025 | 07/01/2025 / Rahul | 05/20/2025 / Rahul | S13089 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555869 / Custom Standard, hexachlorocyclopentadiene Std [CS 5328-2] | A0201702 | 12/03/2025 | 06/03/2025 / Jagrut | 11/13/2023 / Rahul | S13148 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555871 / Custom Standard, 4-nitrophenol Std [CS 5238-4] | A0226283 | 12/04/2025 | 06/04/2025 / Jagrut | 06/04/2025 / Rahul | S13158 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555871 / Custom Standard, 4-nitrophenol Std [CS 5238-4] | A0226283 | 01/01/2026 | 07/01/2025 / Rahul | 06/04/2025 / Rahul | S13159 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Restek | 555868 / Custom Standard, Benzidine Std [CS 5328-1] | A0226493 | 01/01/2026 | 07/01/2025 / Rahul | 06/11/2025 / anahy | S13206 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Absolute Standards, Inc. | 70434 / Acetophenone, Single compound solution, 1000 PPM | 121622 | 12/19/2025 | 06/19/2025 / Rahul | 06/19/2025 / Rahul | S13208 |

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Absolute Standards, Inc. | 91126 / Benzoic acid 2000 ug/mL | 060625 | 12/19/2025 | 06/19/2025 / Rahul | 06/19/2025 / Rahul | S13209 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Absolute Standards, Inc. | 90495 / 1,1-Biphenyl 2000 ug/mL | 042325 | 12/19/2025 | 06/19/2025 / Rahul | 06/19/2025 / Rahul | S13210 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|---|---|---|---|---|---|---|
| Absolute Standards, Inc. | 34571 / 1,2,4,5-Tetrachlorobenzene 5000 ug/mL | 092324 | 12/19/2025 | 06/19/2025 / Rahul | 06/19/2025 / Rahul | S13211 |

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

# Certificate of Analysis

Rev 0 

| Catalog No.: | Lot No.: | Storage: | Solvent: | Exp. Date: | Description: |
|---|---|---|---|---|---|
| Z-112090-04 | 440246 | ≤ -10 °C | Methylene Chloride | 2/16/2026 | CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL |

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|---|---|---|---|---|
| 2-chlorophenol-$d_4$ | 93951-73-6 | 99.3 | 248.12.7P | 7487 ± 17.2 |
| 2-fluorophenol | 367-12-4 | 99.8 | 10.7.3.3P | 7513 ± 17.26 |
| phenol-d6 | 13127-88-3 | 99.9 | 949.120.8P | 7481 ± 17.19 |
| 2,4,6-tribromophenol | 118-79-6 | 99.8 | 12.1.6P | 7469 ± 17.17 |

Received on
02/25/21
by
CG

S9236
to
S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: Erica Castiglione

Erica Castiglione
**Chemist**

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

PRODUCTOS QUIMICOS MONTERREY, S.A. DE C.V.

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

# CERTIFICATE OF ANALYSIS

| | | | |
|---|---|---|---|
| PRODUCT : | SODIUM SULFATE CRYSTALS ANHYDROUS | | |
| QUALITY : | ACS (CODE RMB3375) | FORMULA : | $Na_2SO_4$ |
| SPECIFICATION NUMBER : | 6399 | RELEASE DATE: | ABR/21/2023 |
| LOT NUMBER : | 313201 | | |

| TEST | SPECIFICATIONS | LOT VALUES |
|---|---|---|
| Assay ($Na_2SO_4$) | Min. 99.0% | 99.7 % |
| pH of a 5% solution at 25°C | 5.2 - 9.2 | 6.1 |
| Insoluble matter | Max. 0.01% | 0.005 % |
| Loss on ignition | Max. 0.5% | 0.1 % |
| Chloride (Cl) | Max. 0.001% | <0.001 % |
| Nitrogen compounds (as N) | Max. 5 ppm | <5 ppm |
| Phosphate ($PO_4$) | Max. 0.001% | <0.001 % |
| Heavy metals (as Pb) | Max. 5 ppm | <5 ppm |
| Iron (Fe) | Max. 0.001% | <0.001 % |
| Calcium (Ca) | Max. 0.01% | 0.002 % |
| Magnesium (Mg) | Max. 0.005% | 0.001 % |
| Potassium (K) | Max. 0.008% | 0.003 % |
| Extraction-concentration suitability | Passes test | Passes test |
| Appearance | Passes test | Passes test |
| Identification | Passes test | Passes test |
| Solubility and foreing matter | Passes test | Passes test |
| Retained on US Standard No. 10 sieve | Max. 1% | 0.1 % |
| Retained on US Standard No. 60 sieve | Min. 94% | 97.3 % |
| Through US Standard No. 60 sieve | Max. 5% | 2.5 % |
| Through US Standard No. 100 sieve | Max. 10% | 0.1 % |

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RG-02-01, Ed. 3

Recd. by RI on 7/24/23 E3551

PO: PO2-1308 PRODUCT CODE: SHIP DATE: 2/7/25

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor | J.T.Baker

Material No.: 9266-A4
Batch No.: 25A0262002
Manufactured Date: 2024-11-21
Expiration Date:2026-02-20
Revision No.: 0

# Certificate of Analysis

| Test | Specification | Result |
|---|---|---|
| FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL) | <= 5 | 1 |
| ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL) | <= 10 | 4 |
| Assay ($CH_2Cl_2$) (by GC, exclusive of preservative, corrected for water) | >= 99.8 % | 99.9 % |
| Color (APHA) | <= 10 | 10 |
| Residue after Evaporation | <= 1.0 ppm | 0.8 ppm |
| Titrable Acid (µeq/g) | <= 0.3 | <0.1 |
| Chloride (Cl) | <= 10 ppm | <5 ppm |
| Water (by KF, coulometric) | <= 0.02 % | <0.01 % |

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3926

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

PO: PO2-1178.2 PRODUCT CODE: SHIP DATE: 1/20/2025

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor | J.T.Baker

Material No.: 9266-A4
Batch No.: 25A0262002
Manufactured Date: 2024-11-21
Expiration Date:2026-02-20
Revision No.: 0

# Certificate of Analysis

| Test | Specification | Result |
|---|---|---|
| FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL) | <= 5 | 1 |
| ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL) | <= 10 | 4 |
| Assay ($CH_2Cl_2$) (by GC, exclusive of preservative, corrected for water) | >= 99.8 % | 99.9 % |
| Color (APHA) | <= 10 | 10 |
| Residue after Evaporation | <= 1.0 ppm | 0.8 ppm |
| Titrable Acid (µeq/g) | <= 0.3 | <0.1 |
| Chloride (Cl) | <= 10 ppm | <5 ppm |
| Water (by KF, coulometric) | <= 0.02 % | <0.01 % |

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3930

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date:2027-05-24
Revision No.: 0

# Certificate of Analysis

| Test | Specification | Result |
|---|---|---|
| Assay ($(CH_3)_2CO$) (by GC, corrected forwater) | >= 99.4 % | 99.8 % |
| Color (APHA) | <= 10 | 5 |
| Residue after Evaporation | <= 1.0 ppm | 0.2 ppm |
| Substances Reducing Permanganate | Passes Test | Passes Test |
| Titrable Acid (µeq/g) | <= 0.3 | 0.2 |
| Titrable Base (µeq/g) | <= 0.6 | <0.1 |
| Water ($H_2O$) | <= 0.5 % | 0.2 % |
| FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL) | <= 5 | <1 |
| ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL) | <= 10 | 1 |

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD RS

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3932

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor™ | J.T.Baker

Material No.: 9266-A4
Batch No.: 25A2862010
Manufactured Date: 2024-12-18
Expiration Date:2026-03-19
Revision No.: 0

# Certificate of Analysis

| Test | Specification | Result |
|---|---|---|
| FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL) | <= 5 | <1 |
| ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL) | <= 10 | 2 |
| Assay ($CH_2Cl_2$) (by GC, exclusive of preservative, corrected for water) | >= 99.8 % | 99.9 % |
| Color (APHA) | <= 10 | 5 |
| Residue after Evaporation | <= 1.0 ppm | 0.3 ppm |
| Titrable Acid (µeq/g) | <= 0.3 | <0.1 |
| Chloride (Cl) | <= 10 ppm | <5 ppm |
| Water (by KF, coulometric) | <= 0.02 % | <0.01 % |

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3942

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date:2027-05-24
Revision No.: 0

# Certificate of Analysis

| Test | Specification | Result |
|---|---|---|
| Assay ($(CH_3)_2CO$) (by GC, corrected forwater) | >= 99.4 % | 99.8 % |
| Color (APHA) | <= 10 | 5 |
| Residue after Evaporation | <= 1.0 ppm | 0.2 ppm |
| Substances Reducing Permanganate | Passes Test | Passes Test |
| Titrable Acid (µeq/g) | <= 0.3 | 0.2 |
| Titrable Base (µeq/g) | <= 0.6 | <0.1 |
| Water ($H_2O$) | <= 0.5 % | 0.2 % |
| FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL) | <= 5 | <1 |
| ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL) | <= 10 | 1 |

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RS 6/19/25

E3946

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

VWR® BDH® CHEMICALS

# Certificate of Analysis

| | |
|---|---|
| Material | BDH9274-2.5KG |
| Material Description | BDH SAND STDD OTTAWA W+I 2.5KG |
| Grade | NOT APPLICABLE |
| Batch | 25A2756718 |
| Reassay Date | 12/31/2028 |
| CAS Number | 14808-60-7 |
| Molecular Formula | SiO2 |
| Molecular Mass | 60.09 |
| Date of Manufacture | 12/05/2024 |
| Storage | Room Temperature |

| Characteristics | Specifications | Measured Values |
|---|---|---|
| Appearance | Beige granules. | Beige granules. |
| Moisture | <= 0.1 % | 0.1 % |
| Particle Size 30-40 mesh | >= 80 % | 99 % |
| CUSTOMER PART # BDH9274-2.5KG | | |

Received on 7/9/25.

E3951

Internal ID #: 793

| Signature | Additional Information |
|---|---|
| We certify that this batch conforms to the specifications listed above. | Analysis may have been rounded to significant digits in specification limits |
| This document has been electronically produced and is valid without a signature. | Product meets analytical specifications of the grades listed. |
| Leona Edwardson, Quality Control Sr. Manager - Solon<br>VWR Chemicals, LLC.<br>28600 Fountain Parkway, Solon OH 44139 USA | |

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor™ | J.T.Baker

Material No.: 9266-A4
Batch No.: 25B1862001
Manufactured Date: 2024-12-18
Expiration Date:2026-03-19
Revision No.: 0

# Certificate of Analysis

| Test | Specification | Result |
|---|---|---|
| FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL) | <= 5 | <1 |
| ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL) | <= 10 | 2 |
| Assay ($CH_2Cl_2$) (by GC, exclusive of preservative, corrected for water) | >= 99.8 % | 99.9 % |
| Color (APHA) | <= 10 | 5 |
| Residue after Evaporation | <= 1.0 ppm | 0.3 ppm |
| Titrable Acid (µeq/g) | <= 0.3 | <0.1 |
| Chloride (Cl) | <= 10 ppm | <5 ppm |
| Water (by KF, coulometric) | <= 0.02 % | <0.01 % |

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

RS 7/14/25

E3954

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

## CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***gravimetric***

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555870 **Lot No.:** A0200549

**Description :** Custom 2,4-Dinitrophenol Standard

Custom 2,4-Dinitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** August 31, 2026 **Storage:** 10°C or colder

**Ship:** Ambient

S11484 ↓ S11493 Y.P. 08/10/23

**CERTIFIED VALUES**

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 2,4-Dinitrophenol | 51-28-5 | DR230417RSR | 99% | 25,008.0 µg/mL | +/- 777.3323 |

**Solvent:** Methanol
**CAS #** 67-56-1
**Purity** 99%

Tom Suckar - Mix Technician **Date Mixed:** 02-Aug-2023 **Balance:** 1128342314

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

# Certificate of Analysis

Rev 0



| Catalog No.: | Lot No.: | Storage: | Solvent: | Exp. Date: | Description: |
|---|---|---|---|---|---|
| Z-110094-02 | 506889 | ≤ -10 °C | Methylene Chloride | 7/25/2028 | CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml |

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|---|---|---|---|---|
| 1,2-dichlorobenzene-$d_4$ | 2199-69-1 | 99.7 | 247.29.3P | 5035 ± 28.02 |
| 2-fluorobiphenyl | 321-60-8 | 99.69 | 8.286.1.1P | 4999 ± 103.66 |
| nitrobenzene-d5 | 4165-60-0 | 99.67 | 7.9.3P | 4988 ± 27.32 |
| p-terphenyl-d14 | 1718-51-0 | 99.3 | 9.120.8P | 5005 ± 27.85 |

511494 to 511498 } Y.P. 08/11/2023

*Not a certified value

Certified By: _____

Clint Tipton
**Chemist**

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

## CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***gravimetric***

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555872 **Lot No.:** A0201728

**Description :** Custom Pentachlorophenol Standard

Custom Pentachlorophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** September 30, 2026 **Storage:** 10°C or colder

**Ship:** Ambient

511649 ↓ 511658 Y.P. 11/13/23

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Pentachlorophenol | 87-86-5 | RP230530RSR | 99% | 25,000.0 µg/mL | +/- 777.0837 |

**Solvent:** Methanol
**CAS #** 67-56-1
**Purity** 99%

Josh McCloskey - Operations Technician I

**Date Mixed:** 05-Sep-2023 **Balance:** B251644995

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

## CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***chromatographic plus***

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31853 **Lot No.:** A0196453

**Description :** 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** March 31, 2028 **Storage:** 0°C or colder

**Ship:** Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,4-Dioxane | 123-91-1 | SHBN3770 | 99% | 2,013.0 µg/mL | +/- 25.0521 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant flow 1.8 mL/min.

**Temp. Program:**
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
340°C

**Det. Type:**
FID

**Split Vent:**
100 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

**Date Mixed:** 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

**Date Passed:** 31-Mar-2023

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

## CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***chromatographic plus***

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31853 **Lot No.:** A0200655

**Description :** 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** August 31, 2028 **Storage:** 0°C or colder

**Ship:** Ambient

S11795 ↓ S11808 RC 11/30/23

**CERTIFIED VALUES**

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,4-Dioxane | 123-91-1 | SHBQ1693 | 99% | 2,007.0 µg/mL | +/- 24.9775 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant flow 1.8 mL/min.

**Temp. Program:**
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
340°C

**Det. Type:**
FID

**Split Vent:**
100 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

**Date Mixed:** 06-Aug-2023 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

**Date Passed:** 08-Aug-2023

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

## CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***chromatographic plus***

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31853 **Lot No.:** A0200655

**Description :** 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** August 31, 2028 **Storage:** 0°C or colder

**Ship:** Ambient

S11795 ↓ S11808 RC 11/30/23

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,4-Dioxane | 123-91-1 | SHBQ1693 | 99% | 2,007.0 µg/mL | +/- 24.9775 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant flow 1.8 mL/min.

**Temp. Program:**
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
340°C

**Det. Type:**
FID

**Split Vent:**
100 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

**Date Mixed:** 06-Aug-2023 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

**Date Passed:** 08-Aug-2023

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

## CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***chromatographic plus***

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31853 **Lot No.:** A0200655

**Description :** 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** August 31, 2028 **Storage:** 0°C or colder

**Ship:** Ambient

S11795 ↓ S11808 RC 11/30/23

**CERTIFIED VALUES**

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,4-Dioxane | 123-91-1 | SHBQ1693 | 99% | 2,007.0 µg/mL | +/- 24.9775 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant flow 1.8 mL/min.

**Temp. Program:**
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
340°C

**Det. Type:**
FID

**Split Vent:**
100 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

**Date Mixed:** 06-Aug-2023 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

**Date Passed:** 08-Aug-2023

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

## CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***chromatographic plus***

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31853 **Lot No.:** A0200655

**Description :** 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** August 31, 2028 **Storage:** 0°C or colder

**Ship:** Ambient

S11795 ↓ S11808 RC 11/30/23

**CERTIFIED VALUES**

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,4-Dioxane | 123-91-1 | SHBQ1693 | 99% | 2,007.0 µg/mL | +/- 24.9775 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant flow 1.8 mL/min.

**Temp. Program:**
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
340°C

**Det. Type:**
FID

**Split Vent:**
100 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

**Date Mixed:** 06-Aug-2023 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

**Date Passed:** 08-Aug-2023

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

## CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***chromatographic plus***

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31853 **Lot No.:** A0200655

**Description :** 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** August 31, 2028 **Storage:** 0°C or colder

**Ship:** Ambient

S11795 ↓ S11808 RC 11/30/23

**CERTIFIED VALUES**

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,4-Dioxane | 123-91-1 | SHBQ1693 | 99% | 2,007.0 µg/mL | +/- 24.9775 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant flow 1.8 mL/min.

**Temp. Program:**
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
340°C

**Det. Type:**
FID

**Split Vent:**
100 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

**Date Mixed:** 06-Aug-2023 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

**Date Passed:** 08-Aug-2023

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

# Certificate of Analysis

Rev 0



| Catalog No.: | Lot No.: | Storage: | Solvent: | Exp. Date: | Description: |
|---|---|---|---|---|---|
| Z-020223-01 | 454157 | ≤ -10 °C | P/T Methanol | 6/10/2026 | 1,4-Dioxane Solution, 2000 mg/L, 1 mL |

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|---|---|---|---|---|
| 1,4-dioxane | 123-91-1 | 100 | 223.1.3P | 1997 ± 57.08 |

S12112 ↓ S12116 } RC/ 03/08/24

*Not a certified value

Certified By: Melissa Workoff

Melissa Workoff
**Chemist**

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***chromatographic plus***

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31087 **Lot No.:** A0206206

S12187 ↓ S12206 } RC/ 03/18/24

**Description :** Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

**Container Size :** 5 mL **Pkg Amt:** > 5 mL

**Expiration Date :** January 31, 2032 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 2-Fluorophenol | 367-12-4 | STBK1705 | 99% | 10,005.3 µg/mL | +/- 302.5390 |
| 2 | Phenol-d6 | 13127-88-3 | PR-33287A | 99% | 10,005.5 µg/mL | +/- 302.5475 |
| 3 | 2,4,6-Tribromophenol | 118-79-6 | RP230831RSR | 99% | 10,006.6 µg/mL | +/- 302.5783 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methanol
**CAS #** 67-56-1
**Purity** 99%

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant pressure 10 psi.

**Temp. Program:**
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
330°C

**Det. Type:**
FID

**Split Vent:**
2 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

**Date Mixed:** 04-Jan-2024 **Balance Serial #** 1128360905

Christie Mills - Operations Lead Tech - ARM QC

**Date Passed:** 08-Jan-2024

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*chromatographic plus*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31087 **Lot No.:** A0206206

**Description :** Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

**Container Size :** 5 mL **Pkg Amt:** > 5 mL

**Expiration Date :** January 31, 2032 **Storage:** 10°C or colder

**Ship:** Ambient

S12187 ↓ S12206 } RC/ 03/18/24

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 2-Fluorophenol | 367-12-4 | STBK1705 | 99% | 10,005.3 µg/mL | +/- 302.5390 |
| 2 | Phenol-d6 | 13127-88-3 | PR-33287A | 99% | 10,005.5 µg/mL | +/- 302.5475 |
| 3 | 2,4,6-Tribromophenol | 118-79-6 | RP230831RSR | 99% | 10,006.6 µg/mL | +/- 302.5783 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methanol
**CAS #** 67-56-1
**Purity** 99%

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant pressure 10 psi.

**Temp. Program:**
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
330°C

**Det. Type:**
FID

**Split Vent:**
2 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

**Date Mixed:** 04-Jan-2024 **Balance Serial #** 1128360905

Christie Mills - Operations Lead Tech - ARM QC

**Date Passed:** 08-Jan-2024

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***chromatographic plus***

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31087 **Lot No.:** A0206206

S12187 ↓ S12206 } RC/ 03/18/24

**Description :** Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

**Container Size :** 5 mL **Pkg Amt:** > 5 mL

**Expiration Date :** January 31, 2032 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 2-Fluorophenol | 367-12-4 | STBK1705 | 99% | 10,005.3 µg/mL | +/- 302.5390 |
| 2 | Phenol-d6 | 13127-88-3 | PR-33287A | 99% | 10,005.5 µg/mL | +/- 302.5475 |
| 3 | 2,4,6-Tribromophenol | 118-79-6 | RP230831RSR | 99% | 10,006.6 µg/mL | +/- 302.5783 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methanol
**CAS #** 67-56-1
**Purity** 99%

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant pressure 10 psi.

**Temp. Program:**
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
330°C

**Det. Type:**
FID

**Split Vent:**
2 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

**Date Mixed:** 04-Jan-2024 **Balance Serial #** 1128360905

Christie Mills - Operations Lead Tech - ARM QC

**Date Passed:** 08-Jan-2024

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***chromatographic plus***

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31087 **Lot No.:** A0206206

S12187 ↓ S12206 } RC/ 03/18/24

**Description :** Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

**Container Size :** 5 mL **Pkg Amt:** > 5 mL

**Expiration Date :** January 31, 2032 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 2-Fluorophenol | 367-12-4 | STBK1705 | 99% | 10,005.3 µg/mL | +/- 302.5390 |
| 2 | Phenol-d6 | 13127-88-3 | PR-33287A | 99% | 10,005.5 µg/mL | +/- 302.5475 |
| 3 | 2,4,6-Tribromophenol | 118-79-6 | RP230831RSR | 99% | 10,006.6 µg/mL | +/- 302.5783 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methanol
**CAS #** 67-56-1
**Purity** 99%

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant pressure 10 psi.

**Temp. Program:**
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
330°C

**Det. Type:**
FID

**Split Vent:**
2 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

**Date Mixed:** 04-Jan-2024 **Balance Serial #** 1128360905

Christie Mills - Operations Lead Tech - ARM QC

**Date Passed:** 08-Jan-2024

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

# Certificate of Analysis

***chromatographic plus***

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31086 **Lot No.:** A0206381

**Description :** B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

S12207 ↓ S12221 RC/ 03/18/24

**Container Size :** 5 mL **Pkg Amt:** > 5 mL

**Expiration Date :** December 31, 2029 **Storage:** 10°C or colder

**Handling:** Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Nitrobenzene-d5 | 4165-60-0 | I-25158 | 99% | 5,029.3 µg/mL | +/- 226.5204 |
| 2 | 2-Fluorobiphenyl | 321-60-8 | 00021384 | 99% | 5,030.9 µg/mL | +/- 226.5936 |
| 3 | p-Terphenyl-d14 | 1718-51-0 | PR-32599 | 99% | 5,026.4 µg/mL | +/- 226.3909 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant pressure 10 psi.

**Temp. Program:**
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
330°C

**Det. Type:**
FID

**Split Vent:**
2 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

**Date Mixed:** 09-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

**Date Passed:** 11-Jan-2024

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

# Certificate of Analysis

***chromatographic plus***

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31086 **Lot No.:** A0206381

**Description :** B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

S12207 ↓ S12221 RC/ 03/18/24

**Container Size :** 5 mL **Pkg Amt:** > 5 mL

**Expiration Date :** December 31, 2029 **Storage:** 10°C or colder

**Handling:** Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Nitrobenzene-d5 | 4165-60-0 | I-25158 | 99% | 5,029.3 µg/mL | +/- 226.5204 |
| 2 | 2-Fluorobiphenyl | 321-60-8 | 00021384 | 99% | 5,030.9 µg/mL | +/- 226.5936 |
| 3 | p-Terphenyl-d14 | 1718-51-0 | PR-32599 | 99% | 5,026.4 µg/mL | +/- 226.3909 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant pressure 10 psi.

**Temp. Program:**
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
330°C

**Det. Type:**
FID

**Split Vent:**
2 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

**Date Mixed:** 09-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

**Date Passed:** 11-Jan-2024

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*chromatographic plus*

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31086 **Lot No.:** A0206381

**Description :** B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

S12207 ↓ S12221 RC/ 03/18/24

**Container Size :** 5 mL **Pkg Amt:** > 5 mL

**Expiration Date :** December 31, 2029 **Storage:** 10°C or colder

**Handling:** Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Nitrobenzene-d5 | 4165-60-0 | I-25158 | 99% | 5,029.3 µg/mL | +/- 226.5204 |
| 2 | 2-Fluorobiphenyl | 321-60-8 | 00021384 | 99% | 5,030.9 µg/mL | +/- 226.5936 |
| 3 | p-Terphenyl-d14 | 1718-51-0 | PR-32599 | 99% | 5,026.4 µg/mL | +/- 226.3909 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant pressure 10 psi.

**Temp. Program:**
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
330°C

**Det. Type:**
FID

**Split Vent:**
2 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

**Date Mixed:** 09-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

**Date Passed:** 11-Jan-2024

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

# Certificate of Analysis

***chromatographic plus***

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31086 **Lot No.:** A0206381

**Description :** B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

S12207 ↓ S12221 RC/ 03/18/24

**Container Size :** 5 mL **Pkg Amt:** > 5 mL

**Expiration Date :** December 31, 2029 **Storage:** 10°C or colder

**Handling:** Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Nitrobenzene-d5 | 4165-60-0 | I-25158 | 99% | 5,029.3 µg/mL | +/- 226.5204 |
| 2 | 2-Fluorobiphenyl | 321-60-8 | 00021384 | 99% | 5,030.9 µg/mL | +/- 226.5936 |
| 3 | p-Terphenyl-d14 | 1718-51-0 | PR-32599 | 99% | 5,026.4 µg/mL | +/- 226.3909 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant pressure 10 psi.

**Temp. Program:**
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
330°C

**Det. Type:**
FID

**Split Vent:**
2 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

**Date Mixed:** 09-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

**Date Passed:** 11-Jan-2024

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

# Certificate of Analysis

***chromatographic plus***

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31086 **Lot No.:** A0206381

**Description :** B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

S12207 ↓ S12221 RC/ 03/18/24

**Container Size :** 5 mL **Pkg Amt:** > 5 mL

**Expiration Date :** December 31, 2029 **Storage:** 10°C or colder

**Handling:** Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Nitrobenzene-d5 | 4165-60-0 | I-25158 | 99% | 5,029.3 µg/mL | +/- 226.5204 |
| 2 | 2-Fluorobiphenyl | 321-60-8 | 00021384 | 99% | 5,030.9 µg/mL | +/- 226.5936 |
| 3 | p-Terphenyl-d14 | 1718-51-0 | PR-32599 | 99% | 5,026.4 µg/mL | +/- 226.3909 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant pressure 10 psi.

**Temp. Program:**
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
330°C

**Det. Type:**
FID

**Split Vent:**
2 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

**Date Mixed:** 09-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

**Date Passed:** 11-Jan-2024

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

# Certificate of Analysis

Rev 0



| Catalog No.: | Lot No.: | Storage: | Solvent: | Exp. Date: | Description: |
|---|---|---|---|---|---|
| Z-110381-01 | 520963 | ≤ -10 °C | Methylene Chloride | 10/10/2028 | Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL |

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|---|---|---|---|---|
| acenaphthene | 83-32-9 | 99.9 | 13.1.5P | 1010 ± 9.89 |
| acenaphthylene | 208-96-8 | 97.6 | 14.290.1P | 1014 ± 9.93 |
| aniline | 62-53-3 | 99.97 | 64.1.4P | 1001 ± 9.8 |
| anthracene | 120-12-7 | 99.5 | 15.7.1P | 999.6 ± 9.79 |
| azobenzene | 103-33-3 | 98.1 | 252.7.2P | 999.1 ± 9.8 |
| benzo[a]anthracene | 56-55-3 | 100 | 16.7.3P | 1007 ± 9.86 |
| benzo[b]fluoranthene | 205-99-2 | 99.8 | 17.421.3P | 1011 ± 14.11 |
| benzo[k]fluoranthene | 207-08-9 | 98.9 | 18.421.4P | 1001 ± 10.96 |
| benzo[ghi]perylene | 191-24-2 | 93 | 19.286.4P | 999.6 ± 13.95 |
| benzo[a]pyrene | 50-32-8 | 97 | 20.286.2P | 999.9 ± 22.24 |
| benzyl alcohol | 100-51-6 | 99.9 | 65.18.1P | 1001 ± 9.82 |
| bis(2-chloroethoxy)methane | 111-91-1 | 99.1 | 31.3.15P | 1000 ± 14.69 |
| bis(2-chloroethyl)ether | 111-44-4 | 99.8 | 32.7.1P | 1003 ± 13.89 |
| bis(2-chloro-1-methylethyl) ether | 108-60-1 | 99.5 | 34.3.15P | 999.4 ± 14.68 |
| bis(2-ethylhexyl)adipate | 103-23-1 | 99.5 | 874.7.1P | 999.5 ± 9.8 |
| bis(2-ethylhexyl)phthalate | 117-81-7 | 99.4 | 33.29.1P | 998.8 ± 17.03 |
| 4-bromophenyl phenyl ether | 101-55-3 | 99.4 | 35.7.1.1P | 1000 ± 13.85 |
| butyl benzyl phthalate | 85-68-7 | 98.4 | 36.1.6P | 984.7 ± 16.79 |
| carbazole | 86-74-8 | 99.4 | 239.7.2P | 1000 ± 9.8 |

S12270 ↓ S12274 } RC 05/24/24

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
**Chemist**

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

# Certificate of Analysis



**Catalog No.:** Z-110381-01 **Lot No.:** 520963 **Expiration Date:** 10/10/2028

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|---|---|---|---|---|
| 4-chloroaniline | 106-47-8 | 100 | 66.7.1P | 1000 ± 9.79 |
| 4-chlorophenylphenyl ether | 7005-72-3 | 98 | 37.158.2P | 1001 ± 17.07 |
| 4-chloro-3-methylphenol | 59-50-7 | 99 | 102.1.2P | 1006 ± 17.16 |
| 2-chloronaphthalene | 91-58-7 | 99.9 | 42.7.6P | 1000 ± 9.79 |
| 2-chlorophenol | 95-57-8 | 99.8 | 103.7.1P | 1007 ± 13.96 |
| chrysene | 218-01-9 | 96 | 21.286.2P | 998.4 ± 12.85 |
| dibenz[a,h]anthracene | 53-70-3 | 99.44 | 22.286.3P | 1000 ± 9.74 |
| dibenzofuran | 132-64-9 | 100 | 67.7.2.1P | 1002 ± 9.77 |
| di-n-butyl phthalate | 84-74-2 | 99.84 | 40.286.1P | 1007 ± 24.48 |
| 1,2-dichlorobenzene | 95-50-1 | 99.8 | 43.7.1P | 1000 ± 9.79 |
| 1,3-dichlorobenzene | 541-73-1 | 99.5 | 44.1.3P | 999.4 ± 9.79 |
| 1,4-dichlorobenzene | 106-46-7 | 99.9 | 45.29.2P | 1000 ± 9.79 |
| 2,4-dichlorophenol | 120-83-2 | 99.6 | 104.7.1.1P | 1005 ± 13.93 |
| diethyl phthalate | 84-66-2 | 99.8 | 38.7.1P | 1011 ± 14 |
| 2,4-dimethylphenol | 105-67-9 | 99.6 | 105.7.1.1P | 1009 ± 13.98 |
| dimethyl phthalate | 131-11-3 | 99.9 | 39.9.2P | 996.5 ± 13.8 |
| 1,2-dinitrobenzene | 528-29-0 | 99.86 | 86.7.3.1P | 999.5 ± 9.75 |
| 1,3-dinitrobenzene | 99-65-0 | 100 | 313.7.2P | 998 ± 9.79 |
| 1,4-dinitrobenzene | 100-25-4 | 100 | 907.7.1P | 999.5 ± 9.8 |
| 2,4-dinitrophenol | 51-28-5 | 99.9 | 106.1.6DP | 1002 ± 13.89 |
| 2,4-dinitrotoluene | 121-14-2 | 100 | 87.7.3P | 999.8 ± 13.85 |
| 2,6-dinitrotoluene | 606-20-2 | 99.4 | 88.7.2.1P | 999.6 ± 13.85 |
| di-n-octyl phthalate | 117-84-0 | 99.1 | 41.7.5P | 991.6 ± 13.74 |
| diphenylamine | 122-39-4 | 100 | 78.1.6P | 998 ± 13.79 |
| 2,3,5,6-tetrachlorophenol | 935-95-5 | 97 | 1112.286.1P | 1004 ± 14.02 |
| fluoranthene | 206-44-0 | 98.6 | 23.7.4P | 999.6 ± 9.79 |
| fluorene | 86-73-7 | 98.4 | 24.7.1P | 999.7 ± 9.79 |

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
**Chemist**

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

# Certificate of Analysis



**Catalog No.:** Z-110381-01 **Lot No.:** 520963 **Expiration Date:** 10/10/2028

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|---|---|---|---|---|
| hexachlorobenzene | 118-74-1 | 99 | 46.158.4P | 999.9 ± 13.96 |
| hexachlorobutadiene | 87-68-3 | 97.4 | 47.1.4P | 1000 ± 9.79 |
| hexachlorocyclopentadiene | 77-47-4 | 99.2 | 48.2.2P | 1001 ± 9.8 |
| hexachloroethane | 67-72-1 | 99.9 | 49.1.4P | 1003 ± 9.82 |
| indeno[1,2,3-cd]pyrene | 193-39-5 | 98 | 25.286.4P | 999.4 ± 22.23 |
| isophorone | 78-59-1 | 98.9 | 90.1.4P | 999.9 ± 13.85 |
| 2-methyl-4,6-dinitrophenol | 534-52-1 | 99.6 | 107.421.2DP | 991 ± 24.09 |
| 1-methylnaphthalene | 90-12-0 | 97.1 | 249.7.5P | 999.2 ± 13.95 |
| 2-methylnaphthalene | 91-57-6 | 97.4 | 68.7.2P | 1006 ± 22.38 |
| 2-methylphenol | 95-48-7 | 99.6 | 114.7.3P | 1001 ± 13.87 |
| 3-methylphenol | 108-39-4 | 99.1 | 115.7.4P | 499.7 ± 6.92 |
| 4-methylphenol | 106-44-5 | 99.5 | 116.7.1P | 501.2 ± 6.94 |
| naphthalene | 91-20-3 | 99.8 | 26.9.1P | 1018 ± 9.97 |
| 2-nitroaniline | 88-74-4 | 99.7 | 69.29.1P | 999.6 ± 9.79 |
| 3-nitroaniline | 99-09-2 | 100 | 70.7.3P | 1000 ± 9.74 |
| 4-nitroaniline | 100-01-6 | 99.7 | 71.29.1P | 1001 ± 9.8 |
| nitrobenzene | 98-95-3 | 100 | 94.7.1P | 1000 ± 13.85 |
| 2-nitrophenol | 88-75-5 | 99.1 | 108.29.1P | 996.5 ± 13.81 |
| 4-nitrophenol | 100-02-7 | 100 | 109.7.1P | 1000 ± 13.82 |
| N-nitrosodimethylamine | 62-75-9 | 99.5 | 57.3.19P | 998.5 ± 14.67 |
| N-nitrosodi-n-propylamine | 621-64-7 | 99.8 | 59.286.1P | 996.8 ± 17 |
| pentachlorophenol | 87-86-5 | 99 | 110.1.7P | 1004 ± 13.92 |
| phenanthrene | 85-01-8 | 99.7 | 27.1.5P | 999 ± 12.87 |
| phenol | 108-95-2 | 100 | 112.7.1P | 998.5 ± 13.8 |
| pyrene | 129-00-0 | 99.2 | 28.9.2P | 998.9 ± 9.78 |
| pyridine | 110-86-1 | 100 | 101.24.1P | 999 ± 9.73 |
| 2,3,4,6-Tetrachlorophenol | 58-90-2 | 91.8 | 120.421.1P | 996.5 ± 13.92 |

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
**Chemist**

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

# Certificate of Analysis



**Catalog No.:** Z-110381-01 **Lot No.:** 520963 **Expiration Date:** 10/10/2028

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|---|---|---|---|---|
| 1,2,4-trichlorobenzene | 120-82-1 | 99.6 | 54.29.1P | 999.6 ± 9.79 |
| 2,4,5-trichlorophenol | 95-95-4 | 96.5 | 121.7.1.1P | 999.5 ± 13.85 |
| 2,4,6-trichlorophenol | 88-06-2 | 99.6 | 113.7.1P | 996 ± 13.8 |

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
**Chemist**

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

# Certificate of Analysis

Rev 0 

| Catalog No.: | Lot No.: | Storage: | Solvent: | Exp. Date: | Description: |
|---|---|---|---|---|---|
| Z-010442-07 | 495833 | ≤ -10 °C | Methylene Chloride | 1/16/2028 | Benzaldehyde Solution, 1000 mg/L, 1.3 mL |

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|---|---|---|---|---|
| benzaldehyde | 100-52-7 | 98.3 | 442.421.1P | 996.8 ± 11.49 |

S12275 ↓ S12279 } RC/ 05/24/24

*Not a certified value

Certified By: S. Hunter

Scott Hunter
**Chemist**

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricllly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

# Certificate of Analysis

Rev 0 

| Catalog No.: | Lot No.: | Storage: | Solvent: | Exp. Date: | Description: |
|---|---|---|---|---|---|
| Z-010442-07 | 495833 | ≤ -10 °C | Methylene Chloride | 1/16/2028 | Benzaldehyde Solution, 1000 mg/L, 1.3 mL |

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|---|---|---|---|---|
| benzaldehyde | 100-52-7 | 98.3 | 442.421.1P | 996.8 ± 11.49 |

S12275 ↓ S12279 } RC/ 05/24/24

*Not a certified value

Certified By: S. Hunter

Scott Hunter
**Chemist**

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricllly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*gravimetric*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555223 **Lot No.:** A0214021

**Description :** Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Handling:** This product is photosensitive. **Ship:** Ambient

**CERTIFIED VALUES**

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 µg/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 µg/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 µg/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 µg/mL | +/- 22.9569 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12449 ↓ S12508 RC 7/24/24

Rebecca Gingerich - Operations Tech II

**Date Mixed:** 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

## CERTIFIED REFERENCE MATERIAL

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

# Certificate of Analysis

*gravimetric*

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555223 **Lot No.:** A0214021

**Description :** Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Handling:** This product is photosensitive. **Ship:** Ambient

**CERTIFIED VALUES**

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 µg/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 µg/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 µg/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 µg/mL | +/- 22.9569 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12449 ↓ S12508 RC 7/24/24

Rebecca Gingerich - Operations Tech II

**Date Mixed:** 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*gravimetric*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555223 **Lot No.:** A0214021

**Description :** Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Handling:** This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 µg/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 µg/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 µg/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 µg/mL | +/- 22.9569 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12449 ↓ S12508 RC 7/24/24

Rebecca Gingerich - Operations Tech II

**Date Mixed:** 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*gravimetric*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555223 **Lot No.:** A0214021

**Description :** Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Handling:** This product is photosensitive. **Ship:** Ambient

**CERTIFIED VALUES**

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 µg/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 µg/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 µg/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 µg/mL | +/- 22.9569 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12449 ↓ S12508 RC 7/24/24

Rebecca Gingerich - Operations Tech II

**Date Mixed:** 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

# Certificate of Analysis

*gravimetric*

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555223 **Lot No.:** A0214021

**Description :** Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Handling:** This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 µg/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 µg/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 µg/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 µg/mL | +/- 22.9569 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12449 ↓ S12508 RC 7/24/24

Rebecca Gingerich - Operations Tech II

**Date Mixed:** 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*gravimetric*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555223 **Lot No.:** A0214021

**Description :** Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Handling:** This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 µg/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 µg/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 µg/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 µg/mL | +/- 22.9569 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12449 ↓ S12508 RC 7/24/24

Rebecca Gingerich - Operations Tech II

**Date Mixed:** 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

# Certificate of Analysis

*gravimetric*

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555223 **Lot No.:** A0214021

**Description :** Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Handling:** This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 µg/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 µg/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 µg/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 µg/mL | +/- 22.9569 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12449 ↓ S12508 RC 7/24/24

Rebecca Gingerich - Operations Tech II

**Date Mixed:** 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

# Certificate of Analysis

*gravimetric*

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555223 **Lot No.:** A0214021

**Description :** Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Handling:** This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 µg/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 µg/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 µg/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 µg/mL | +/- 22.9569 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12449 ↓ S12508 RC 7/24/24

Rebecca Gingerich - Operations Tech II

**Date Mixed:** 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*gravimetric*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555223 **Lot No.:** A0214021

**Description :** Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Handling:** This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 µg/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 µg/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 µg/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 µg/mL | +/- 22.9569 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12449 ↓ S12508 RC 7/24/24

Rebecca Gingerich - Operations Tech II

**Date Mixed:** 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

## CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***gravimetric***

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555223 **Lot No.:** A0214021

**Description :** Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Handling:** This product is photosensitive. **Ship:** Ambient

**CERTIFIED VALUES**

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 3,3'-Dichlorobenzidine | 91-94-1 | S240326RSR | 99% | 1,004.0 µg/mL | +/- 23.0487 |
| 2 | Atrazine | 1912-24-9 | 5FYWL | 99% | 1,005.0 µg/mL | +/- 23.0717 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 µg/mL | +/- 23.0947 |
| 4 | epsilon-Caprolactam | 105-60-2 | Y16H012 | 99% | 1,000.0 µg/mL | +/- 22.9569 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II **Date Mixed:** 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***gravimetric***

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555224 **Lot No.:** A0214017

**Description :** Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,008.0 µg/mL | +/- 29.630084 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,010.0 µg/mL | +/- 29.688874 |
| 5 | Biphenyl | 92-52-4 | MKCS5928 | 99% | 1,008.0 µg/mL | +/- 29.630084 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

**Date Mixed:** 18-Jul-2024 **Balance:** 1128360905

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***gravimetric***

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555224 **Lot No.:** A0214017

**Description :** Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,008.0 µg/mL | +/- 29.630084 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,010.0 µg/mL | +/- 29.688874 |
| 5 | Biphenyl | 92-52-4 | MKCS5928 | 99% | 1,008.0 µg/mL | +/- 29.630084 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

**Date Mixed:** 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***gravimetric***

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555224 **Lot No.:** A0214017

**Description :** Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,008.0 µg/mL | +/- 29.630084 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,010.0 µg/mL | +/- 29.688874 |
| 5 | Biphenyl | 92-52-4 | MKCS5928 | 99% | 1,008.0 µg/mL | +/- 29.630084 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

**Date Mixed:** 18-Jul-2024 **Balance:** 1128360905

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*gravimetric*

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555224 **Lot No.:** A0214017

**Description :** Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,008.0 µg/mL | +/- 29.630084 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,010.0 µg/mL | +/- 29.688874 |
| 5 | Biphenyl | 92-52-4 | MKCS5928 | 99% | 1,008.0 µg/mL | +/- 29.630084 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

**Date Mixed:** 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*gravimetric*

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555224 **Lot No.:** A0214017

**Description :** Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,008.0 µg/mL | +/- 29.630084 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,010.0 µg/mL | +/- 29.688874 |
| 5 | Biphenyl | 92-52-4 | MKCS5928 | 99% | 1,008.0 µg/mL | +/- 29.630084 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

**Date Mixed:** 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***gravimetric***

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555224 **Lot No.:** A0214017

**Description :** Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,008.0 µg/mL | +/- 29.630084 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,010.0 µg/mL | +/- 29.688874 |
| 5 | Biphenyl | 92-52-4 | MKCS5928 | 99% | 1,008.0 µg/mL | +/- 29.630084 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

**Date Mixed:** 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*gravimetric*

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555224 **Lot No.:** A0214017

**Description :** Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,008.0 µg/mL | +/- 29.630084 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,010.0 µg/mL | +/- 29.688874 |
| 5 | Biphenyl | 92-52-4 | MKCS5928 | 99% | 1,008.0 µg/mL | +/- 29.630084 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

**Date Mixed:** 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*gravimetric*

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555224 **Lot No.:** A0214017

**Description :** Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,008.0 µg/mL | +/- 29.630084 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,010.0 µg/mL | +/- 29.688874 |
| 5 | Biphenyl | 92-52-4 | MKCS5928 | 99% | 1,008.0 µg/mL | +/- 29.630084 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

**Date Mixed:** 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

# Certificate of Analysis

*gravimetric*

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555224 **Lot No.:** A0214017

**Description :** Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2026 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 95-94-3 | MKCT9480 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 2 | Acetophenone | 98-86-2 | STBH8205 | 99% | 1,005.0 µg/mL | +/- 29.541899 |
| 3 | Benzaldehyde | 100-52-7 | RD231129RSRA | 99% | 1,008.0 µg/mL | +/- 29.630084 |
| 4 | Benzoic acid | 65-85-0 | MKCR2694 | 99% | 1,010.0 µg/mL | +/- 29.688874 |
| 5 | Biphenyl | 92-52-4 | MKCS5928 | 99% | 1,008.0 µg/mL | +/- 29.630084 |

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

**Date Mixed:** 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

# General Certified Reference Material Notes

## Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

## Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

## Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  $k$ is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

## Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

## Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

# Certificate of Analysis

*chromatographic plus*

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31615 **Lot No.:** A0212955

**Description :** GC/MS Tuning Mixture

GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** June 30, 2027 **Storage:** 10°C or colder

**Handling:** Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Pentachlorophenol | 87-86-5 | RP240517RSR | 99% | 1,004.5 µg/mL | +/- 44.8902 |
| 2 | DFTPP (Decafluorotriphenylphosphine) | 5074-71-5 | Q117-147 | 99% | 1,004.5 µg/mL | +/- 44.8902 |
| 3 | Benzidine | 92-87-5 | S240430RSR | 99% | 1,006.0 µg/mL | +/- 44.9572 |
| 4 | 4,4'-DDT | 50-29-3 | S240530RSR | 97% | 1,000.1 µg/mL | +/- 44.6922 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12577 ↓ S12579 } RC 8/2/24

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant pressure 10 psi.

**Temp. Program:**
75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
330°C

**Det. Type:**
FID

**Split Vent:**
10 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

**Date Mixed:** 19-Jun-2024 **Balance Serial #** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

**Date Passed:** 26-Jun-2024

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

# Certificate of Analysis

*chromatographic plus*

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31206 **Lot No.:** A0212266

**Description :** SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** April 30, 2030 **Storage:** 10°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,4-Dichlorobenzene-d4 | 3855-82-1 | PR-30447 | 99% | 2,000.6 µg/mL | +/- 90.1075 |
| 2 | Naphthalene-d8 | 1146-65-2 | M-2180 | 99% | 2,000.3 µg/mL | +/- 90.0925 |
| 3 | Acenaphthene-d10 | 15067-26-2 | PR-33507 | 99% | 2,000.4 µg/mL | +/- 90.1000 |
| 4 | Phenanthrene-d10 | 1517-22-2 | PR-34099 | 99% | 2,000.5 µg/mL | +/- 90.1037 |
| 5 | Chrysene-d12 | 1719-03-5 | PR-33506 | 99% | 2,000.7 µg/mL | +/- 90.1112 |
| 6 | Perylene-d12 | 1520-96-3 | PR-33205 | 99% | 2,000.6 µg/mL | +/- 90.1075 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12645 ↓ S12674 AC 10/1/24

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

# Certificate of Analysis

*chromatographic plus*

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31206 **Lot No.:** A0212266

**Description :** SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** April 30, 2030 **Storage:** 10°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,4-Dichlorobenzene-d4 | 3855-82-1 | PR-30447 | 99% | 2,000.6 µg/mL | +/- 90.1075 |
| 2 | Naphthalene-d8 | 1146-65-2 | M-2180 | 99% | 2,000.3 µg/mL | +/- 90.0925 |
| 3 | Acenaphthene-d10 | 15067-26-2 | PR-33507 | 99% | 2,000.4 µg/mL | +/- 90.1000 |
| 4 | Phenanthrene-d10 | 1517-22-2 | PR-34099 | 99% | 2,000.5 µg/mL | +/- 90.1037 |
| 5 | Chrysene-d12 | 1719-03-5 | PR-33506 | 99% | 2,000.7 µg/mL | +/- 90.1112 |
| 6 | Perylene-d12 | 1520-96-3 | PR-33205 | 99% | 2,000.6 µg/mL | +/- 90.1075 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12645 ↓ S12674 AC 10/1/24

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

# Certificate of Analysis

*chromatographic plus*

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31206 **Lot No.:** A0212266

**Description :** SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** April 30, 2030 **Storage:** 10°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,4-Dichlorobenzene-d4 | 3855-82-1 | PR-30447 | 99% | 2,000.6 µg/mL | +/- 90.1075 |
| 2 | Naphthalene-d8 | 1146-65-2 | M-2180 | 99% | 2,000.3 µg/mL | +/- 90.0925 |
| 3 | Acenaphthene-d10 | 15067-26-2 | PR-33507 | 99% | 2,000.4 µg/mL | +/- 90.1000 |
| 4 | Phenanthrene-d10 | 1517-22-2 | PR-34099 | 99% | 2,000.5 µg/mL | +/- 90.1037 |
| 5 | Chrysene-d12 | 1719-03-5 | PR-33506 | 99% | 2,000.7 µg/mL | +/- 90.1112 |
| 6 | Perylene-d12 | 1520-96-3 | PR-33205 | 99% | 2,000.6 µg/mL | +/- 90.1075 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12645 ↓ S12674 AC 10/1/24

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*chromatographic plus*

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31206 **Lot No.:** A0212266

**Description :** SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** April 30, 2030 **Storage:** 10°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 1,4-Dichlorobenzene-d4 | 3855-82-1 | PR-30447 | 99% | 2,000.6 µg/mL | +/- 90.1075 |
| 2 | Naphthalene-d8 | 1146-65-2 | M-2180 | 99% | 2,000.3 µg/mL | +/- 90.0925 |
| 3 | Acenaphthene-d10 | 15067-26-2 | PR-33507 | 99% | 2,000.4 µg/mL | +/- 90.1000 |
| 4 | Phenanthrene-d10 | 1517-22-2 | PR-34099 | 99% | 2,000.5 µg/mL | +/- 90.1037 |
| 5 | Chrysene-d12 | 1719-03-5 | PR-33506 | 99% | 2,000.7 µg/mL | +/- 90.1112 |
| 6 | Perylene-d12 | 1520-96-3 | PR-33205 | 99% | 2,000.6 µg/mL | +/- 90.1075 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

S12645 ↓ S12674 AC 10/1/24

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*chromatographic plus*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31850 **Lot No.:** A0219438

**Description :** 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** September 30, 2025 **Storage:** 0°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

S12963 ↓ S12992 AC 12/17/24

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,008.3 µg/mL | +/- 36.6849 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | S240313RSR | 99% | 1,008.6 µg/mL | +/- 36.6985 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,002.9 µg/mL | +/- 36.4893 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | 002891T24M | 99% | 1,003.0 µg/mL | +/- 36.4938 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,005.6 µg/mL | +/- 36.5894 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,004.1 µg/mL | +/- 36.5348 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,002.1 µg/mL | +/- 36.4620 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBL6287 | 99% | 1,005.3 µg/mL | +/- 36.5757 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,008.4 µg/mL | +/- 36.6894 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,004.6 µg/mL | +/- 36.5530 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 502.1 µg/mL | +/- 18.2697 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.8 µg/mL | +/- 18.3288 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,006.5 µg/mL | +/- 36.6212 |
| 16 | Hexachloroethane | 67-72-1 | DAXRI | 99% | 1,004.5 µg/mL | +/- 36.5484 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,002.5 µg/mL | +/- 36.4757 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 18 | Isophorone | 78-59-1 | MKCR3249 | 99% | 1,003.4 | µg/mL | +/- 36.5075 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,006.5 | µg/mL | +/- 36.6212 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 15705100 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCCK6969 | 99% | 1,001.5 | µg/mL | +/- 36.4393 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,002.1 | µg/mL | +/- 36.4620 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 26 | Hexachlorobutadiene | 87-68-3 | X05J | 98% | 1,002.5 | µg/mL | +/- 36.4771 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,004.5 | µg/mL | +/- 36.5484 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBL3028 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 990.2 | µg/mL | +/- 36.0269 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,001.3 | µg/mL | +/- 36.4325 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBK8870 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | 3YFRE | 97% | 1,004.6 | µg/mL | +/- 36.5505 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,004.3 | µg/mL | +/- 36.5393 |
| 34 | 2-Nitroaniline | 88-74-4 | RP240715RSR | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP240703RSR | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 36 | Acenaphthylene | 208-96-8 | RP241029RSR | 98% | 1,000.0 | µg/mL | +/- 36.3835 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | TRC3-1075941-2-1 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,008.9 | µg/mL | +/- 36.7076 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP240701RSR | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 42 | 3-Nitroaniline | 99-09-2 | RP240708RSR | 99% | 1,004.6 | µg/mL | +/- 36.5530 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | D240927RSR | ----% | 1,005.6 | µg/mL | +/- 36.5894 |
| 44 | Dibenzofuran | 132-64-9 | MKCN1772 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | 102869V26E | 99% | 1,008.3 | µg/mL | +/- 36.6849 |
| 46 | 4-Nitrophenol | 100-02-7 | 20241029-2-AN | 99% | 1,004.8 | µg/mL | +/- 36.5575 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-34476 | 99% | 1,005.8 | µg/mL | +/- 36.5939 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP231219RSR | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 49 | Fluorene | 86-73-7 | 10246250 | 98% | 1,000.7 | µg/mL | +/- 36.4102 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 51 | Diethylphthalate | 84-66-2 | BCCJ6241 | 99% | 1,003.9 | µg/mL | +/- 36.5257 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | S241008RSR | 99% | 1,001.3 | µg/mL | +/- 36.4302 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 54 | Diphenylamine | 122-39-4 | MKCT1512 | 99% | 1,003.0 | µg/mL | +/- 36.4938 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,002.4 | µg/mL | +/- 36.4711 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,008.8 | µg/mL | +/- 36.7031 |
| 57 | Hexachlorobenzene | 118-74-1 | 15458400 | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 58 | Pentachlorophenol | 87-86-5 | RP240517RSR | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 59 | Phenanthrene | 85-01-8 | MKCT3391 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 60 | Anthracene | 120-12-7 | 101492T18R | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 61 | Carbazole | 86-74-8 | 15276700 | 99% | 1,005.4 | µg/mL | +/- 36.5803 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 64 | Pyrene | 129-00-0 | BCCK2592 | 99% | 1,002.0 | µg/mL | +/- 36.4575 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 67 | Benz(a)anthracene | 56-55-3 | I70012022BAA | 99% | 1,005.5 | µg/mL | +/- 36.5848 |
| 68 | Chrysene | 218-01-9 | RP241007RSR | 99% | 1,005.3 | µg/mL | +/- 36.5757 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCS8065 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 15566400 | 99% | 1,002.3 | µg/mL | +/- 36.4666 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 052013B | 99% | 1,004.1 | µg/mL | +/- 36.5348 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 73 | Benzo(a)pyrene | 50-32-8 | NQLXA | 98% | 1,006.2 | µg/mL | +/- 36.6108 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,001.8 | µg/mL | +/- 36.4490 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,003.3 | µg/mL | +/- 36.5029 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP241014RSR | 98% | 1,003.8 | µg/mL | +/- 36.5217 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*chromatographic plus*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31850 **Lot No.:** A0219438

**Description :** 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12963 → S12992 AC 12/17/24

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** September 30, 2025 **Storage:** 0°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,008.3 µg/mL | +/- 36.6849 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | S240313RSR | 99% | 1,008.6 µg/mL | +/- 36.6985 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,002.9 µg/mL | +/- 36.4893 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | 002891T24M | 99% | 1,003.0 µg/mL | +/- 36.4938 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,005.6 µg/mL | +/- 36.5894 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,004.1 µg/mL | +/- 36.5348 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,002.1 µg/mL | +/- 36.4620 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBL6287 | 99% | 1,005.3 µg/mL | +/- 36.5757 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,008.4 µg/mL | +/- 36.6894 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,004.6 µg/mL | +/- 36.5530 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 502.1 µg/mL | +/- 18.2697 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.8 µg/mL | +/- 18.3288 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,006.5 µg/mL | +/- 36.6212 |
| 16 | Hexachloroethane | 67-72-1 | DAXRI | 99% | 1,004.5 µg/mL | +/- 36.5484 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,002.5 µg/mL | +/- 36.4757 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 18 | Isophorone | 78-59-1 | MKCR3249 | 99% | 1,003.4 | µg/mL | +/- 36.5075 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,006.5 | µg/mL | +/- 36.6212 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 15705100 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCCK6969 | 99% | 1,001.5 | µg/mL | +/- 36.4393 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,002.1 | µg/mL | +/- 36.4620 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 26 | Hexachlorobutadiene | 87-68-3 | X05J | 98% | 1,002.5 | µg/mL | +/- 36.4771 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,004.5 | µg/mL | +/- 36.5484 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBL3028 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 990.2 | µg/mL | +/- 36.0269 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,001.3 | µg/mL | +/- 36.4325 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBK8870 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | 3YFRE | 97% | 1,004.6 | µg/mL | +/- 36.5505 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,004.3 | µg/mL | +/- 36.5393 |
| 34 | 2-Nitroaniline | 88-74-4 | RP240715RSR | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP240703RSR | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 36 | Acenaphthylene | 208-96-8 | RP241029RSR | 98% | 1,000.0 | µg/mL | +/- 36.3835 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | TRC3-1075941-2-1 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,008.9 | µg/mL | +/- 36.7076 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP240701RSR | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 42 | 3-Nitroaniline | 99-09-2 | RP240708RSR | 99% | 1,004.6 | µg/mL | +/- 36.5530 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | D240927RSR | ----% | 1,005.6 | µg/mL | +/- 36.5894 |
| 44 | Dibenzofuran | 132-64-9 | MKCN1772 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | 102869V26E | 99% | 1,008.3 | µg/mL | +/- 36.6849 |
| 46 | 4-Nitrophenol | 100-02-7 | 20241029-2-AN | 99% | 1,004.8 | µg/mL | +/- 36.5575 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-34476 | 99% | 1,005.8 | µg/mL | +/- 36.5939 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP231219RSR | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 49 | Fluorene | 86-73-7 | 10246250 | 98% | 1,000.7 | µg/mL | +/- 36.4102 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 51 | Diethylphthalate | 84-66-2 | BCCJ6241 | 99% | 1,003.9 | µg/mL | +/- 36.5257 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | S241008RSR | 99% | 1,001.3 | µg/mL | +/- 36.4302 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 54 | Diphenylamine | 122-39-4 | MKCT1512 | 99% | 1,003.0 | µg/mL | +/- 36.4938 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,002.4 | µg/mL | +/- 36.4711 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,008.8 | µg/mL | +/- 36.7031 |
| 57 | Hexachlorobenzene | 118-74-1 | 15458400 | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 58 | Pentachlorophenol | 87-86-5 | RP240517RSR | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 59 | Phenanthrene | 85-01-8 | MKCT3391 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 60 | Anthracene | 120-12-7 | 101492T18R | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 61 | Carbazole | 86-74-8 | 15276700 | 99% | 1,005.4 | µg/mL | +/- 36.5803 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 64 | Pyrene | 129-00-0 | BCCK2592 | 99% | 1,002.0 | µg/mL | +/- 36.4575 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 67 | Benz(a)anthracene | 56-55-3 | I70012022BAA | 99% | 1,005.5 | µg/mL | +/- 36.5848 |
| 68 | Chrysene | 218-01-9 | RP241007RSR | 99% | 1,005.3 | µg/mL | +/- 36.5757 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCS8065 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 15566400 | 99% | 1,002.3 | µg/mL | +/- 36.4666 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 052013B | 99% | 1,004.1 | µg/mL | +/- 36.5348 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 73 | Benzo(a)pyrene | 50-32-8 | NQLXA | 98% | 1,006.2 | µg/mL | +/- 36.6108 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,001.8 | µg/mL | +/- 36.4490 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,003.3 | µg/mL | +/- 36.5029 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP241014RSR | 98% | 1,003.8 | µg/mL | +/- 36.5217 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*chromatographic plus*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31850 **Lot No.:** A0219438

**Description :** 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** September 30, 2025 **Storage:** 0°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

S12963 ↓ S12992 AC 12/17/24

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,008.3 µg/mL | +/- 36.6849 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | S240313RSR | 99% | 1,008.6 µg/mL | +/- 36.6985 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,002.9 µg/mL | +/- 36.4893 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | 002891T24M | 99% | 1,003.0 µg/mL | +/- 36.4938 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,005.6 µg/mL | +/- 36.5894 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,004.1 µg/mL | +/- 36.5348 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,002.1 µg/mL | +/- 36.4620 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBL6287 | 99% | 1,005.3 µg/mL | +/- 36.5757 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,008.4 µg/mL | +/- 36.6894 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,004.6 µg/mL | +/- 36.5530 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 502.1 µg/mL | +/- 18.2697 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.8 µg/mL | +/- 18.3288 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,006.5 µg/mL | +/- 36.6212 |
| 16 | Hexachloroethane | 67-72-1 | DAXRI | 99% | 1,004.5 µg/mL | +/- 36.5484 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,002.5 µg/mL | +/- 36.4757 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 18 | Isophorone | 78-59-1 | MKCR3249 | 99% | 1,003.4 | µg/mL | +/- 36.5075 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,006.5 | µg/mL | +/- 36.6212 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 15705100 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCCK6969 | 99% | 1,001.5 | µg/mL | +/- 36.4393 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,002.1 | µg/mL | +/- 36.4620 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 26 | Hexachlorobutadiene | 87-68-3 | X05J | 98% | 1,002.5 | µg/mL | +/- 36.4771 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,004.5 | µg/mL | +/- 36.5484 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBL3028 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 990.2 | µg/mL | +/- 36.0269 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,001.3 | µg/mL | +/- 36.4325 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBK8870 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | 3YFRE | 97% | 1,004.6 | µg/mL | +/- 36.5505 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,004.3 | µg/mL | +/- 36.5393 |
| 34 | 2-Nitroaniline | 88-74-4 | RP240715RSR | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP240703RSR | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 36 | Acenaphthylene | 208-96-8 | RP241029RSR | 98% | 1,000.0 | µg/mL | +/- 36.3835 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | TRC3-1075941-2-1 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,008.9 | µg/mL | +/- 36.7076 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP240701RSR | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 42 | 3-Nitroaniline | 99-09-2 | RP240708RSR | 99% | 1,004.6 | µg/mL | +/- 36.5530 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | D240927RSR | ----% | 1,005.6 | µg/mL | +/- 36.5894 |
| 44 | Dibenzofuran | 132-64-9 | MKCN1772 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | 102869V26E | 99% | 1,008.3 | µg/mL | +/- 36.6849 |
| 46 | 4-Nitrophenol | 100-02-7 | 20241029-2-AN | 99% | 1,004.8 | µg/mL | +/- 36.5575 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-34476 | 99% | 1,005.8 | µg/mL | +/- 36.5939 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP231219RSR | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 49 | Fluorene | 86-73-7 | 10246250 | 98% | 1,000.7 | µg/mL | +/- 36.4102 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 51 | Diethylphthalate | 84-66-2 | BCCJ6241 | 99% | 1,003.9 | µg/mL | +/- 36.5257 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | S241008RSR | 99% | 1,001.3 | µg/mL | +/- 36.4302 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 54 | Diphenylamine | 122-39-4 | MKCT1512 | 99% | 1,003.0 | µg/mL | +/- 36.4938 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,002.4 | µg/mL | +/- 36.4711 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,008.8 | µg/mL | +/- 36.7031 |
| 57 | Hexachlorobenzene | 118-74-1 | 15458400 | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 58 | Pentachlorophenol | 87-86-5 | RP240517RSR | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 59 | Phenanthrene | 85-01-8 | MKCT3391 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 60 | Anthracene | 120-12-7 | 101492T18R | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 61 | Carbazole | 86-74-8 | 15276700 | 99% | 1,005.4 | µg/mL | +/- 36.5803 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 64 | Pyrene | 129-00-0 | BCCK2592 | 99% | 1,002.0 | µg/mL | +/- 36.4575 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 67 | Benz(a)anthracene | 56-55-3 | I70012022BAA | 99% | 1,005.5 | µg/mL | +/- 36.5848 |
| 68 | Chrysene | 218-01-9 | RP241007RSR | 99% | 1,005.3 | µg/mL | +/- 36.5757 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCS8065 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 15566400 | 99% | 1,002.3 | µg/mL | +/- 36.4666 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 052013B | 99% | 1,004.1 | µg/mL | +/- 36.5348 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 73 | Benzo(a)pyrene | 50-32-8 | NQLXA | 98% | 1,006.2 | µg/mL | +/- 36.6108 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,001.8 | µg/mL | +/- 36.4490 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,003.3 | µg/mL | +/- 36.5029 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP241014RSR | 98% | 1,003.8 | µg/mL | +/- 36.5217 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*chromatographic plus*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31850 **Lot No.:** A0219438

**Description :** 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** September 30, 2025 **Storage:** 0°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

S12963 ↓ S12992 AC 12/17/24

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,008.3 µg/mL | +/- 36.6849 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | S240313RSR | 99% | 1,008.6 µg/mL | +/- 36.6985 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,002.9 µg/mL | +/- 36.4893 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | 002891T24M | 99% | 1,003.0 µg/mL | +/- 36.4938 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,005.6 µg/mL | +/- 36.5894 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,004.1 µg/mL | +/- 36.5348 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,002.1 µg/mL | +/- 36.4620 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBL6287 | 99% | 1,005.3 µg/mL | +/- 36.5757 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,008.4 µg/mL | +/- 36.6894 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,004.6 µg/mL | +/- 36.5530 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 502.1 µg/mL | +/- 18.2697 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.8 µg/mL | +/- 18.3288 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,006.5 µg/mL | +/- 36.6212 |
| 16 | Hexachloroethane | 67-72-1 | DAXRI | 99% | 1,004.5 µg/mL | +/- 36.5484 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,002.5 µg/mL | +/- 36.4757 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 18 | Isophorone | 78-59-1 | MKCR3249 | 99% | 1,003.4 | µg/mL | +/- 36.5075 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,006.5 | µg/mL | +/- 36.6212 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 15705100 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCCK6969 | 99% | 1,001.5 | µg/mL | +/- 36.4393 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,002.1 | µg/mL | +/- 36.4620 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 26 | Hexachlorobutadiene | 87-68-3 | X05J | 98% | 1,002.5 | µg/mL | +/- 36.4771 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,004.5 | µg/mL | +/- 36.5484 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBL3028 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 990.2 | µg/mL | +/- 36.0269 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,001.3 | µg/mL | +/- 36.4325 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBK8870 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | 3YFRE | 97% | 1,004.6 | µg/mL | +/- 36.5505 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,004.3 | µg/mL | +/- 36.5393 |
| 34 | 2-Nitroaniline | 88-74-4 | RP240715RSR | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP240703RSR | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 36 | Acenaphthylene | 208-96-8 | RP241029RSR | 98% | 1,000.0 | µg/mL | +/- 36.3835 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | TRC3-1075941-2-1 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,008.9 | µg/mL | +/- 36.7076 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP240701RSR | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 42 | 3-Nitroaniline | 99-09-2 | RP240708RSR | 99% | 1,004.6 | µg/mL | +/- 36.5530 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | D240927RSR | ----% | 1,005.6 | µg/mL | +/- 36.5894 |
| 44 | Dibenzofuran | 132-64-9 | MKCN1772 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | 102869V26E | 99% | 1,008.3 | µg/mL | +/- 36.6849 |
| 46 | 4-Nitrophenol | 100-02-7 | 20241029-2-AN | 99% | 1,004.8 | µg/mL | +/- 36.5575 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-34476 | 99% | 1,005.8 | µg/mL | +/- 36.5939 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP231219RSR | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 49 | Fluorene | 86-73-7 | 10246250 | 98% | 1,000.7 | µg/mL | +/- 36.4102 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 51 | Diethylphthalate | 84-66-2 | BCCJ6241 | 99% | 1,003.9 | µg/mL | +/- 36.5257 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | S241008RSR | 99% | 1,001.3 | µg/mL | +/- 36.4302 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 54 | Diphenylamine | 122-39-4 | MKCT1512 | 99% | 1,003.0 | µg/mL | +/- 36.4938 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,002.4 | µg/mL | +/- 36.4711 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,008.8 | µg/mL | +/- 36.7031 |
| 57 | Hexachlorobenzene | 118-74-1 | 15458400 | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 58 | Pentachlorophenol | 87-86-5 | RP240517RSR | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 59 | Phenanthrene | 85-01-8 | MKCT3391 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 60 | Anthracene | 120-12-7 | 101492T18R | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 61 | Carbazole | 86-74-8 | 15276700 | 99% | 1,005.4 | µg/mL | +/- 36.5803 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 64 | Pyrene | 129-00-0 | BCCK2592 | 99% | 1,002.0 | µg/mL | +/- 36.4575 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 67 | Benz(a)anthracene | 56-55-3 | I70012022BAA | 99% | 1,005.5 | µg/mL | +/- 36.5848 |
| 68 | Chrysene | 218-01-9 | RP241007RSR | 99% | 1,005.3 | µg/mL | +/- 36.5757 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCS8065 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 15566400 | 99% | 1,002.3 | µg/mL | +/- 36.4666 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 052013B | 99% | 1,004.1 | µg/mL | +/- 36.5348 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 73 | Benzo(a)pyrene | 50-32-8 | NQLXA | 98% | 1,006.2 | µg/mL | +/- 36.6108 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,001.8 | µg/mL | +/- 36.4490 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,003.3 | µg/mL | +/- 36.5029 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP241014RSR | 98% | 1,003.8 | µg/mL | +/- 36.5217 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*chromatographic plus*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31850 **Lot No.:** A0219438

**Description :** 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12963 → S12992 AC 12/17/24

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** September 30, 2025 **Storage:** 0°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,008.3 µg/mL | +/- 36.6849 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | S240313RSR | 99% | 1,008.6 µg/mL | +/- 36.6985 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,002.9 µg/mL | +/- 36.4893 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | 002891T24M | 99% | 1,003.0 µg/mL | +/- 36.4938 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,005.6 µg/mL | +/- 36.5894 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,004.1 µg/mL | +/- 36.5348 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,002.1 µg/mL | +/- 36.4620 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBL6287 | 99% | 1,005.3 µg/mL | +/- 36.5757 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,008.4 µg/mL | +/- 36.6894 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,004.6 µg/mL | +/- 36.5530 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 502.1 µg/mL | +/- 18.2697 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.8 µg/mL | +/- 18.3288 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,006.5 µg/mL | +/- 36.6212 |
| 16 | Hexachloroethane | 67-72-1 | DAXRI | 99% | 1,004.5 µg/mL | +/- 36.5484 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,002.5 µg/mL | +/- 36.4757 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 18 | Isophorone | 78-59-1 | MKCR3249 | 99% | 1,003.4 | µg/mL | +/- 36.5075 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,006.5 | µg/mL | +/- 36.6212 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 15705100 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCCK6969 | 99% | 1,001.5 | µg/mL | +/- 36.4393 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,002.1 | µg/mL | +/- 36.4620 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 26 | Hexachlorobutadiene | 87-68-3 | X05J | 98% | 1,002.5 | µg/mL | +/- 36.4771 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,004.5 | µg/mL | +/- 36.5484 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBL3028 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 990.2 | µg/mL | +/- 36.0269 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,001.3 | µg/mL | +/- 36.4325 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBK8870 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | 3YFRE | 97% | 1,004.6 | µg/mL | +/- 36.5505 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,004.3 | µg/mL | +/- 36.5393 |
| 34 | 2-Nitroaniline | 88-74-4 | RP240715RSR | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP240703RSR | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 36 | Acenaphthylene | 208-96-8 | RP241029RSR | 98% | 1,000.0 | µg/mL | +/- 36.3835 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | TRC3-1075941-2-1 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,008.9 | µg/mL | +/- 36.7076 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP240701RSR | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 42 | 3-Nitroaniline | 99-09-2 | RP240708RSR | 99% | 1,004.6 | µg/mL | +/- 36.5530 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | D240927RSR | ----% | 1,005.6 | µg/mL | +/- 36.5894 |
| 44 | Dibenzofuran | 132-64-9 | MKCN1772 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | 102869V26E | 99% | 1,008.3 | µg/mL | +/- 36.6849 |
| 46 | 4-Nitrophenol | 100-02-7 | 20241029-2-AN | 99% | 1,004.8 | µg/mL | +/- 36.5575 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-34476 | 99% | 1,005.8 | µg/mL | +/- 36.5939 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP231219RSR | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 49 | Fluorene | 86-73-7 | 10246250 | 98% | 1,000.7 | µg/mL | +/- 36.4102 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 51 | Diethylphthalate | 84-66-2 | BCCJ6241 | 99% | 1,003.9 | µg/mL | +/- 36.5257 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | S241008RSR | 99% | 1,001.3 | µg/mL | +/- 36.4302 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 54 | Diphenylamine | 122-39-4 | MKCT1512 | 99% | 1,003.0 | µg/mL | +/- 36.4938 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,002.4 | µg/mL | +/- 36.4711 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,008.8 | µg/mL | +/- 36.7031 |
| 57 | Hexachlorobenzene | 118-74-1 | 15458400 | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 58 | Pentachlorophenol | 87-86-5 | RP240517RSR | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 59 | Phenanthrene | 85-01-8 | MKCT3391 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 60 | Anthracene | 120-12-7 | 101492T18R | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 61 | Carbazole | 86-74-8 | 15276700 | 99% | 1,005.4 | µg/mL | +/- 36.5803 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 64 | Pyrene | 129-00-0 | BCCK2592 | 99% | 1,002.0 | µg/mL | +/- 36.4575 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 67 | Benz(a)anthracene | 56-55-3 | I70012022BAA | 99% | 1,005.5 | µg/mL | +/- 36.5848 |
| 68 | Chrysene | 218-01-9 | RP241007RSR | 99% | 1,005.3 | µg/mL | +/- 36.5757 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCS8065 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 15566400 | 99% | 1,002.3 | µg/mL | +/- 36.4666 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 052013B | 99% | 1,004.1 | µg/mL | +/- 36.5348 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 73 | Benzo(a)pyrene | 50-32-8 | NQLXA | 98% | 1,006.2 | µg/mL | +/- 36.6108 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,001.8 | µg/mL | +/- 36.4490 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,003.3 | µg/mL | +/- 36.5029 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP241014RSR | 98% | 1,003.8 | µg/mL | +/- 36.5217 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*chromatographic plus*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31850 **Lot No.:** A0219438

**Description :** 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** September 30, 2025 **Storage:** 0°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

S12963 ↓ S12992 AC 12/17/24

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,008.3 µg/mL | +/- 36.6849 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | S240313RSR | 99% | 1,008.6 µg/mL | +/- 36.6985 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,002.9 µg/mL | +/- 36.4893 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | 002891T24M | 99% | 1,003.0 µg/mL | +/- 36.4938 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,005.6 µg/mL | +/- 36.5894 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,004.1 µg/mL | +/- 36.5348 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,002.1 µg/mL | +/- 36.4620 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBL6287 | 99% | 1,005.3 µg/mL | +/- 36.5757 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,008.4 µg/mL | +/- 36.6894 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,004.6 µg/mL | +/- 36.5530 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 502.1 µg/mL | +/- 18.2697 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.8 µg/mL | +/- 18.3288 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,006.5 µg/mL | +/- 36.6212 |
| 16 | Hexachloroethane | 67-72-1 | DAXRI | 99% | 1,004.5 µg/mL | +/- 36.5484 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,002.5 µg/mL | +/- 36.4757 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 18 | Isophorone | 78-59-1 | MKCR3249 | 99% | 1,003.4 | µg/mL | +/- 36.5075 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,006.5 | µg/mL | +/- 36.6212 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 15705100 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCCK6969 | 99% | 1,001.5 | µg/mL | +/- 36.4393 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,002.1 | µg/mL | +/- 36.4620 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 26 | Hexachlorobutadiene | 87-68-3 | X05J | 98% | 1,002.5 | µg/mL | +/- 36.4771 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,004.5 | µg/mL | +/- 36.5484 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBL3028 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 990.2 | µg/mL | +/- 36.0269 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,001.3 | µg/mL | +/- 36.4325 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBK8870 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | 3YFRE | 97% | 1,004.6 | µg/mL | +/- 36.5505 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,004.3 | µg/mL | +/- 36.5393 |
| 34 | 2-Nitroaniline | 88-74-4 | RP240715RSR | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP240703RSR | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 36 | Acenaphthylene | 208-96-8 | RP241029RSR | 98% | 1,000.0 | µg/mL | +/- 36.3835 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | TRC3-1075941-2-1 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,008.9 | µg/mL | +/- 36.7076 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP240701RSR | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 42 | 3-Nitroaniline | 99-09-2 | RP240708RSR | 99% | 1,004.6 | µg/mL | +/- 36.5530 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | D240927RSR | ----% | 1,005.6 | µg/mL | +/- 36.5894 |
| 44 | Dibenzofuran | 132-64-9 | MKCN1772 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | 102869V26E | 99% | 1,008.3 | µg/mL | +/- 36.6849 |
| 46 | 4-Nitrophenol | 100-02-7 | 20241029-2-AN | 99% | 1,004.8 | µg/mL | +/- 36.5575 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-34476 | 99% | 1,005.8 | µg/mL | +/- 36.5939 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP231219RSR | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 49 | Fluorene | 86-73-7 | 10246250 | 98% | 1,000.7 | µg/mL | +/- 36.4102 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 51 | Diethylphthalate | 84-66-2 | BCCJ6241 | 99% | 1,003.9 | µg/mL | +/- 36.5257 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | S241008RSR | 99% | 1,001.3 | µg/mL | +/- 36.4302 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 54 | Diphenylamine | 122-39-4 | MKCT1512 | 99% | 1,003.0 | µg/mL | +/- 36.4938 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,002.4 | µg/mL | +/- 36.4711 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,008.8 | µg/mL | +/- 36.7031 |
| 57 | Hexachlorobenzene | 118-74-1 | 15458400 | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 58 | Pentachlorophenol | 87-86-5 | RP240517RSR | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 59 | Phenanthrene | 85-01-8 | MKCT3391 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 60 | Anthracene | 120-12-7 | 101492T18R | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 61 | Carbazole | 86-74-8 | 15276700 | 99% | 1,005.4 | µg/mL | +/- 36.5803 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 64 | Pyrene | 129-00-0 | BCCK2592 | 99% | 1,002.0 | µg/mL | +/- 36.4575 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 67 | Benz(a)anthracene | 56-55-3 | I70012022BAA | 99% | 1,005.5 | µg/mL | +/- 36.5848 |
| 68 | Chrysene | 218-01-9 | RP241007RSR | 99% | 1,005.3 | µg/mL | +/- 36.5757 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCS8065 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 15566400 | 99% | 1,002.3 | µg/mL | +/- 36.4666 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 052013B | 99% | 1,004.1 | µg/mL | +/- 36.5348 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 73 | Benzo(a)pyrene | 50-32-8 | NQLXA | 98% | 1,006.2 | µg/mL | +/- 36.6108 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,001.8 | µg/mL | +/- 36.4490 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,003.3 | µg/mL | +/- 36.5029 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP241014RSR | 98% | 1,003.8 | µg/mL | +/- 36.5217 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

*chromatographic plus*

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31850 **Lot No.:** A0219438

**Description :** 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

S12963 → S12992 AC 12/17/24

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** September 30, 2025 **Storage:** 0°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Pyridine | 110-86-1 | SHBP6240 | 99% | 1,008.3 µg/mL | +/- 36.6849 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | S240313RSR | 99% | 1,008.6 µg/mL | +/- 36.6985 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,002.9 µg/mL | +/- 36.4893 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | 002891T24M | 99% | 1,003.0 µg/mL | +/- 36.4938 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,005.6 µg/mL | +/- 36.5894 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,004.1 µg/mL | +/- 36.5348 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,002.1 µg/mL | +/- 36.4620 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,003.5 µg/mL | +/- 36.5120 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBL6287 | 99% | 1,005.3 µg/mL | +/- 36.5757 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,008.4 µg/mL | +/- 36.6894 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,004.6 µg/mL | +/- 36.5530 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBJ0710 | 99% | 502.1 µg/mL | +/- 18.2697 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 503.8 µg/mL | +/- 18.3288 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,006.5 µg/mL | +/- 36.6212 |
| 16 | Hexachloroethane | 67-72-1 | DAXRI | 99% | 1,004.5 µg/mL | +/- 36.5484 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,002.5 µg/mL | +/- 36.4757 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 18 | Isophorone | 78-59-1 | MKCR3249 | 99% | 1,003.4 | µg/mL | +/- 36.5075 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | XW5GK | 99% | 1,006.5 | µg/mL | +/- 36.6212 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 15705100 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCCK6969 | 99% | 1,001.5 | µg/mL | +/- 36.4393 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,002.1 | µg/mL | +/- 36.4620 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 26 | Hexachlorobutadiene | 87-68-3 | X05J | 98% | 1,002.5 | µg/mL | +/- 36.4771 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,004.5 | µg/mL | +/- 36.5484 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBL3028 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 990.2 | µg/mL | +/- 36.0269 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063I14L | 98% | 1,001.3 | µg/mL | +/- 36.4325 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBK8870 | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | 3YFRE | 97% | 1,004.6 | µg/mL | +/- 36.5505 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,004.3 | µg/mL | +/- 36.5393 |
| 34 | 2-Nitroaniline | 88-74-4 | RP240715RSR | 99% | 1,004.4 | µg/mL | +/- 36.5439 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP240703RSR | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 36 | Acenaphthylene | 208-96-8 | RP241029RSR | 98% | 1,000.0 | µg/mL | +/- 36.3835 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | TRC3-1075941-2-1 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,008.9 | µg/mL | +/- 36.7076 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP240701RSR | 99% | 1,002.5 | µg/mL | +/- 36.4757 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,000.0 | µg/mL | +/- 36.3847 |
| 42 | 3-Nitroaniline | 99-09-2 | RP240708RSR | 99% | 1,004.6 | µg/mL | +/- 36.5530 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | D240927RSR | ----% | 1,005.6 | µg/mL | +/- 36.5894 |
| 44 | Dibenzofuran | 132-64-9 | MKCN1772 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | 102869V26E | 99% | 1,008.3 | µg/mL | +/- 36.6849 |
| 46 | 4-Nitrophenol | 100-02-7 | 20241029-2-AN | 99% | 1,004.8 | µg/mL | +/- 36.5575 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-34476 | 99% | 1,005.8 | µg/mL | +/- 36.5939 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP231219RSR | 99% | 1,006.4 | µg/mL | +/- 36.6166 |
| 49 | Fluorene | 86-73-7 | 10246250 | 98% | 1,000.7 | µg/mL | +/- 36.4102 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | MKCT7248 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 51 | Diethylphthalate | 84-66-2 | BCCJ6241 | 99% | 1,003.9 | µg/mL | +/- 36.5257 |
| 52 | 4-Nitroaniline | 100-01-6 | RP230111 | 99% | 1,006.6 | µg/mL | +/- 36.6257 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | S241008RSR | 99% | 1,001.3 | µg/mL | +/- 36.4302 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 54 | Diphenylamine | 122-39-4 | MKCT1512 | 99% | 1,003.0 | µg/mL | +/- 36.4938 |
| 55 | Azobenzene | 103-33-3 | BCCK0887 | 99% | 1,002.4 | µg/mL | +/- 36.4711 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,008.8 | µg/mL | +/- 36.7031 |
| 57 | Hexachlorobenzene | 118-74-1 | 15458400 | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 58 | Pentachlorophenol | 87-86-5 | RP240517RSR | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 59 | Phenanthrene | 85-01-8 | MKCT3391 | 99% | 1,004.9 | µg/mL | +/- 36.5621 |
| 60 | Anthracene | 120-12-7 | 101492T18R | 99% | 1,005.1 | µg/mL | +/- 36.5712 |
| 61 | Carbazole | 86-74-8 | 15276700 | 99% | 1,005.4 | µg/mL | +/- 36.5803 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,006.3 | µg/mL | +/- 36.6121 |
| 63 | Fluoranthene | 206-44-0 | MKCQ4728 | 99% | 1,003.5 | µg/mL | +/- 36.5120 |
| 64 | Pyrene | 129-00-0 | BCCK2592 | 99% | 1,002.0 | µg/mL | +/- 36.4575 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCM1988 | 99% | 1,005.9 | µg/mL | +/- 36.5984 |
| 67 | Benz(a)anthracene | 56-55-3 | I70012022BAA | 99% | 1,005.5 | µg/mL | +/- 36.5848 |
| 68 | Chrysene | 218-01-9 | RP241007RSR | 99% | 1,005.3 | µg/mL | +/- 36.5757 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCS8065 | 99% | 1,007.5 | µg/mL | +/- 36.6576 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 15566400 | 99% | 1,002.3 | µg/mL | +/- 36.4666 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 052013B | 99% | 1,004.1 | µg/mL | +/- 36.5348 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 99% | 1,002.8 | µg/mL | +/- 36.4847 |
| 73 | Benzo(a)pyrene | 50-32-8 | NQLXA | 98% | 1,006.2 | µg/mL | +/- 36.6108 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,001.8 | µg/mL | +/- 36.4490 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 2-ASA-59-1 | 99% | 1,003.3 | µg/mL | +/- 36.5029 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP241014RSR | 98% | 1,003.8 | µg/mL | +/- 36.5217 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

# Certificate of Analysis

Rev 0 

| Catalog No.: | Lot No.: | Storage: | Solvent: | Exp. Date: | Description: |
|---|---|---|---|---|---|
| Z-110816-01 | 531243 | ≤ -10 °C | Methylene Chloride | 1/2/2030 | Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL |

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|---|---|---|---|---|
| atrazine | 1912-24-9 | 99.5 | 337.7.4P | 997 ± 5.81 |
| benzidine | 92-87-5 | 99.9 | 124.18.6.2P | 993.8 ± 5.78 |
| caprolactam | 105-60-2 | 99.9 | 271.1.6P | 999 ± 5.82 |

513057 ↓ 513061 AC 1/16/25

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: Melissa Workoff

Melissa Workoff
**Chemist**

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

# Certificate of Analysis

Rev 0 

| Catalog No.: | Lot No.: | Storage: | Solvent: | Exp. Date: | Description: |
|---|---|---|---|---|---|
| Z-010074-07 | 525551 | -18°C +/- 4°C | Methylene Chloride | 9/3/2029 | 3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 mL |

| Compound | CAS No. | Purity (%) | Compound Lot No. | Concentration, mg/L |
|---|---|---|---|---|
| 3,3'-dichlorobenzidine | 91-94-1 | 97.9 | 74.421.2P | 994.7 ± 30.56 |

This RM is intended for use as a calibration standard or a quality control standard for chromatography equipment such as GC, GC/MS, HPLC, and HPLC/MS. It may also be used for various USEPA, NIOSH and ASTM methods.

Recommended storage container for ampuled products after opening is a 12 mm x 32 mm amber vial with screw cap Teflon lined silicon septum. The modeled % change per day can be calculated using the following:

$$\% \text{ Change} = 116192x^{-2.578} + 40.383e^{-0.03y}$$

where x = boiling point of the most volatile analyte in the mix (in degrees K)
y = boiling point of the solvent (in degrees K)

This model assumes the container is stored at -10 °C and is unopened during storage. The user should determine what the acceptable error for their process is and calculate the maximum number of days the opened ampule should be stored. The minimum sample size recommended for use is 1µL.

513078 ↓ 513087 } AC 3/10/25

Manufactured by

Raevyn Steele
**Chemist**

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

*Not a certified valu

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

# Certificate of Analysis

***chromatographic plus***

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31850 **Lot No.:** A0221014

**Description :** 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** November 30, 2025 **Storage:** 0°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

S13088 ↓ S13117 RC 5/20/25.

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Pyridine | 110-86-1 | SHBR4811 | 99% | 1,005.2 µg/mL | +/- 36.5730 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | S241226RSR | 99% | 1,005.5 µg/mL | +/- 36.5848 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,005.3 µg/mL | +/- 36.5780 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,005.4 µg/mL | +/- 36.5816 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | 002891T24M | 99% | 1,004.7 µg/mL | +/- 36.5562 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,005.3 µg/mL | +/- 36.5766 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,004.6 µg/mL | +/- 36.5530 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,006.3 µg/mL | +/- 36.6125 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,005.5 µg/mL | +/- 36.5853 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBL6287 | 99% | 1,004.6 µg/mL | +/- 36.5507 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,004.8 µg/mL | +/- 36.5575 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,005.4 µg/mL | +/- 36.5803 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBL3873 | 99% | 502.7 µg/mL | +/- 18.2888 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 502.6 µg/mL | +/- 18.2869 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,004.6 µg/mL | +/- 36.5502 |
| 16 | Hexachloroethane | 67-72-1 | DAXRI | 99% | 1,004.6 µg/mL | +/- 36.5530 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,005.3 µg/mL | +/- 36.5780 |

| | | | | | | | | |
|---|---|---|---|---|---|---|---|---|
| 18 | Isophorone | 78-59-1 | MKCR3249 | 99% | 1,004.6 | µg/mL | +/- | 36.5511 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,005.7 | µg/mL | +/- | 36.5916 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | DIRAF | 99% | 1,004.9 | µg/mL | +/- | 36.5612 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 15705100 | 99% | 1,004.3 | µg/mL | +/- | 36.5402 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCCK6969 | 99% | 1,004.7 | µg/mL | +/- | 36.5571 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,004.6 | µg/mL | +/- | 36.5516 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,006.8 | µg/mL | +/- | 36.6335 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,005.9 | µg/mL | +/- | 36.5980 |
| 26 | Hexachlorobutadiene | 87-68-3 | X05J | 98% | 1,004.1 | µg/mL | +/- | 36.5328 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,005.4 | µg/mL | +/- | 36.5793 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBL3028 | 99% | 1,006.3 | µg/mL | +/- | 36.6144 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 990.2 | µg/mL | +/- | 36.0269 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063P13G | 99% | 1,005.9 | µg/mL | +/- | 36.5975 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBK8870 | 99% | 1,004.7 | µg/mL | +/- | 36.5566 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | 3YFRE | 97% | 1,004.7 | µg/mL | +/- | 36.5571 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,005.3 | µg/mL | +/- | 36.5757 |
| 34 | 2-Nitroaniline | 88-74-4 | RP240715RSR | 99% | 1,004.8 | µg/mL | +/- | 36.5584 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP240703RSR | 99% | 1,004.5 | µg/mL | +/- | 36.5475 |
| 36 | Acenaphthylene | 208-96-8 | 214935V18H | 95% | 1,000.1 | µg/mL | +/- | 36.3888 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | TRC3-1075941-2-1 | 99% | 1,004.9 | µg/mL | +/- | 36.5616 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,004.5 | µg/mL | +/- | 36.5489 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,005.2 | µg/mL | +/- | 36.5734 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP240701RSR | 99% | 1,005.9 | µg/mL | +/- | 36.6003 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,000.0 | µg/mL | +/- | 36.3847 |
| 42 | 3-Nitroaniline | 99-09-2 | RP240708RSR | 99% | 1,004.6 | µg/mL | +/- | 36.5525 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | D240927RSR | 99% | 1,005.0 | µg/mL | +/- | 36.5657 |
| 44 | Dibenzofuran | 132-64-9 | MKCN1772 | 99% | 1,005.4 | µg/mL | +/- | 36.5803 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | 102869V26E | 99% | 1,005.4 | µg/mL | +/- | 36.5803 |
| 46 | 4-Nitrophenol | 100-02-7 | 20241120-1-AN | 99% | 1,005.7 | µg/mL | +/- | 36.5930 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-34476 | 99% | 1,004.9 | µg/mL | +/- | 36.5616 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP240523RSR | 99% | 1,005.2 | µg/mL | +/- | 36.5739 |
| 49 | Fluorene | 86-73-7 | 10246250 | 98% | 1,005.8 | µg/mL | +/- | 36.5948 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | 002531K02D | 99% | 1,004.8 | µg/mL | +/- | 36.5584 |
| 51 | Diethylphthalate | 84-66-2 | STBL3611 | 99% | 1,005.3 | µg/mL | +/- | 36.5762 |
| 52 | 4-Nitroaniline | 100-01-6 | RP240830RSR | 99% | 1,005.5 | µg/mL | +/- | 36.5844 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | S241008RSR | 99% | 1,004.8 | µg/mL | +/- | 36.5589 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 54 | Diphenylamine | 122-39-4 | MKCT1512 | 99% | 1,005.2 | µg/mL | +/- 36.5725 |
| 55 | Azobenzene | 103-33-3 | BCCL3292 | 99% | 1,004.7 | µg/mL | +/- 36.5566 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,005.6 | µg/mL | +/- 36.5875 |
| 57 | Hexachlorobenzene | 118-74-1 | 15828800 | 99% | 1,005.3 | µg/mL | +/- 36.5789 |
| 58 | Pentachlorophenol | 87-86-5 | RP240517RSR | 99% | 1,005.4 | µg/mL | +/- 36.5825 |
| 59 | Phenanthrene | 85-01-8 | MKCT3391 | 99% | 1,004.8 | µg/mL | +/- 36.5584 |
| 60 | Anthracene | 120-12-7 | MKCW9141 | 99% | 1,005.5 | µg/mL | +/- 36.5834 |
| 61 | Carbazole | 86-74-8 | 15630800 | 99% | 1,005.3 | µg/mL | +/- 36.5789 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,005.5 | µg/mL | +/- 36.5848 |
| 63 | Fluoranthene | 206-44-0 | A0458721 | 99% | 1,005.4 | µg/mL | +/- <0.0001 |
| 64 | Pyrene | 129-00-0 | BCCL8032 | 99% | 1,005.2 | µg/mL | +/- 36.5734 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,004.8 | µg/mL | +/- 36.5584 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCR8567 | 99% | 1,004.5 | µg/mL | +/- 36.5480 |
| 67 | Benz(a)anthracene | 56-55-3 | I60012022BAA | 99% | 1,005.5 | µg/mL | +/- 36.5844 |
| 68 | Chrysene | 218-01-9 | RP241212RSR | 99% | 1,005.5 | µg/mL | +/- 36.5848 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCS8065 | 99% | 1,004.7 | µg/mL | +/- 36.5548 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 15817300 | 99% | 1,004.8 | µg/mL | +/- 36.5607 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 022013B | 99% | 1,005.1 | µg/mL | +/- 36.5716 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 98% | 1,004.6 | µg/mL | +/- 36.5511 |
| 73 | Benzo(a)pyrene | 50-32-8 | NQLXA | 98% | 1,004.2 | µg/mL | +/- 36.5377 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,004.1 | µg/mL | +/- 36.5337 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 712061504-1-1 | 99% | 1,005.7 | µg/mL | +/- 36.5930 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP241014RSR | 98% | 1,005.4 | µg/mL | +/- 36.5814 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant flow 1.8 mL/min.

**Temp. Program:**
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
340°C

**Det. Type:**
FID

**Split Vent:**
100 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

**Date Mixed:** 12-Jan-2025 **Balance Serial #** 1128360905

Dillan Murphy - Operations Technician I

**Date Passed:** 21-Jan-2025

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

# Certificate of Analysis

***chromatographic plus***

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 31850 **Lot No.:** A0221014

**Description :** 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** November 30, 2025 **Storage:** 0°C or colder

**Handling:** Sonication required. Mix is photosensitive. **Ship:** Ambient

S13088 ↓ S13117 RC 5/20/25.

CERTIFIED VALUES

| Elution Order | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Pyridine | 110-86-1 | SHBR4811 | 99% | 1,005.2 µg/mL | +/- 36.5730 |
| 2 | N-Nitrosodimethylamine | 62-75-9 | S241226RSR | 99% | 1,005.5 µg/mL | +/- 36.5848 |
| 3 | Phenol | 108-95-2 | MKCK1120 | 99% | 1,005.3 µg/mL | +/- 36.5780 |
| 4 | Aniline | 62-53-3 | X22F726 | 99% | 1,005.4 µg/mL | +/- 36.5816 |
| 5 | Bis(2-chloroethyl)ether | 111-44-4 | 002891T24M | 99% | 1,004.7 µg/mL | +/- 36.5562 |
| 6 | 2-Chlorophenol | 95-57-8 | STBJ3909 | 99% | 1,005.3 µg/mL | +/- 36.5766 |
| 7 | 1,3-Dichlorobenzene | 541-73-1 | BCCD5315 | 99% | 1,004.6 µg/mL | +/- 36.5530 |
| 8 | 1,4-Dichlorobenzene | 106-46-7 | MKBS7929V | 99% | 1,006.3 µg/mL | +/- 36.6125 |
| 9 | Benzyl alcohol | 100-51-6 | SHBK5469 | 99% | 1,005.5 µg/mL | +/- 36.5853 |
| 10 | 1,2-Dichlorobenzene | 95-50-1 | SHBL6287 | 99% | 1,004.6 µg/mL | +/- 36.5507 |
| 11 | 2-Methylphenol (o-cresol) | 95-48-7 | SHBN7598 | 99% | 1,004.8 µg/mL | +/- 36.5575 |
| 12 | 2,2'-oxybis(1-chloropropane) | 108-60-1 | 29-MAR-45-5 | 99% | 1,005.4 µg/mL | +/- 36.5803 |
| 13 | 3-Methylphenol (m-cresol) | 108-39-4 | STBL3873 | 99% | 502.7 µg/mL | +/- 18.2888 |
| 14 | 4-Methylphenol (p-cresol) | 106-44-5 | SHBN3411 | 99% | 502.6 µg/mL | +/- 18.2869 |
| 15 | N-Nitroso-di-n-propylamine | 621-64-7 | N63MG | 99% | 1,004.6 µg/mL | +/- 36.5502 |
| 16 | Hexachloroethane | 67-72-1 | DAXRI | 99% | 1,004.6 µg/mL | +/- 36.5530 |
| 17 | Nitrobenzene | 98-95-3 | 10224044 | 99% | 1,005.3 µg/mL | +/- 36.5780 |

| | | | | | | | | |
|---|---|---|---|---|---|---|---|---|
| 18 | Isophorone | 78-59-1 | MKCR3249 | 99% | 1,004.6 | µg/mL | +/- | 36.5511 |
| 19 | 2-Nitrophenol | 88-75-5 | RP230710 | 99% | 1,005.7 | µg/mL | +/- | 36.5916 |
| 20 | 2,4-Dimethylphenol | 105-67-9 | DIRAF | 99% | 1,004.9 | µg/mL | +/- | 36.5612 |
| 21 | Bis(2-chloroethoxy)methane | 111-91-1 | 15705100 | 99% | 1,004.3 | µg/mL | +/- | 36.5402 |
| 22 | 2,4-Dichlorophenol | 120-83-2 | BCCK6969 | 99% | 1,004.7 | µg/mL | +/- | 36.5571 |
| 23 | 1,2,4-Trichlorobenzene | 120-82-1 | SHBP5900 | 99% | 1,004.6 | µg/mL | +/- | 36.5516 |
| 24 | Naphthalene | 91-20-3 | STBL1057 | 99% | 1,006.8 | µg/mL | +/- | 36.6335 |
| 25 | 4-Chloroaniline | 106-47-8 | BCCJ3217 | 99% | 1,005.9 | µg/mL | +/- | 36.5980 |
| 26 | Hexachlorobutadiene | 87-68-3 | X05J | 98% | 1,004.1 | µg/mL | +/- | 36.5328 |
| 27 | 4-Chloro-3-methylphenol | 59-50-7 | BCCD4461 | 99% | 1,005.4 | µg/mL | +/- | 36.5793 |
| 28 | 2-Methylnaphthalene | 91-57-6 | STBL3028 | 99% | 1,006.3 | µg/mL | +/- | 36.6144 |
| 29 | 1-Methylnaphthalene | 90-12-0 | 5234.00-8 | 98% | 990.2 | µg/mL | +/- | 36.0269 |
| 30 | Hexachlorocyclopentadiene | 77-47-4 | 099063P13G | 99% | 1,005.9 | µg/mL | +/- | 36.5975 |
| 31 | 2,4,6-Trichlorophenol | 88-06-2 | STBK8870 | 99% | 1,004.7 | µg/mL | +/- | 36.5566 |
| 32 | 2,4,5-Trichlorophenol | 95-95-4 | 3YFRE | 97% | 1,004.7 | µg/mL | +/- | 36.5571 |
| 33 | 2-Chloronaphthalene | 91-58-7 | RPN7O | 99% | 1,005.3 | µg/mL | +/- | 36.5757 |
| 34 | 2-Nitroaniline | 88-74-4 | RP240715RSR | 99% | 1,004.8 | µg/mL | +/- | 36.5584 |
| 35 | 1,4-Dinitrobenzene | 100-25-4 | RP240703RSR | 99% | 1,004.5 | µg/mL | +/- | 36.5475 |
| 36 | Acenaphthylene | 208-96-8 | 214935V18H | 95% | 1,000.1 | µg/mL | +/- | 36.3888 |
| 37 | 1,3-Dinitrobenzene | 99-65-0 | TRC3-1075941-2-1 | 99% | 1,004.9 | µg/mL | +/- | 36.5616 |
| 38 | Dimethylphthalate | 131-11-3 | 358221L17K | 99% | 1,004.5 | µg/mL | +/- | 36.5489 |
| 39 | 2,6-Dinitrotoluene | 606-20-2 | BCCG1833 | 99% | 1,005.2 | µg/mL | +/- | 36.5734 |
| 40 | 1,2-Dinitrobenzene | 528-29-0 | RP240701RSR | 99% | 1,005.9 | µg/mL | +/- | 36.6003 |
| 41 | Acenaphthene | 83-32-9 | MKCR7169 | 99% | 1,000.0 | µg/mL | +/- | 36.3847 |
| 42 | 3-Nitroaniline | 99-09-2 | RP240708RSR | 99% | 1,004.6 | µg/mL | +/- | 36.5525 |
| 43 | 2,4-Dinitrophenol | 51-28-5 | D240927RSR | 99% | 1,005.0 | µg/mL | +/- | 36.5657 |
| 44 | Dibenzofuran | 132-64-9 | MKCN1772 | 99% | 1,005.4 | µg/mL | +/- | 36.5803 |
| 45 | 2,4-Dinitrotoluene | 121-14-2 | 102869V26E | 99% | 1,005.4 | µg/mL | +/- | 36.5803 |
| 46 | 4-Nitrophenol | 100-02-7 | 20241120-1-AN | 99% | 1,005.7 | µg/mL | +/- | 36.5930 |
| 47 | 2,3,4,6-Tetrachlorophenol | 58-90-2 | PR-34476 | 99% | 1,004.9 | µg/mL | +/- | 36.5616 |
| 48 | 2,3,5,6-Tetrachlorophenol | 935-95-5 | RP240523RSR | 99% | 1,005.2 | µg/mL | +/- | 36.5739 |
| 49 | Fluorene | 86-73-7 | 10246250 | 98% | 1,005.8 | µg/mL | +/- | 36.5948 |
| 50 | 4-Chlorophenyl phenyl ether | 7005-72-3 | 002531K02D | 99% | 1,004.8 | µg/mL | +/- | 36.5584 |
| 51 | Diethylphthalate | 84-66-2 | STBL3611 | 99% | 1,005.3 | µg/mL | +/- | 36.5762 |
| 52 | 4-Nitroaniline | 100-01-6 | RP240830RSR | 99% | 1,005.5 | µg/mL | +/- | 36.5844 |
| 53 | 4,6-Dinitro-2-methylphenol (Dinitro-o-cresol) | 534-52-1 | S241008RSR | 99% | 1,004.8 | µg/mL | +/- | 36.5589 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| 54 | Diphenylamine | 122-39-4 | MKCT1512 | 99% | 1,005.2 | µg/mL | +/- 36.5725 |
| 55 | Azobenzene | 103-33-3 | BCCL3292 | 99% | 1,004.7 | µg/mL | +/- 36.5566 |
| 56 | 4-Bromophenyl phenyl ether | 101-55-3 | STBH6361 | 99% | 1,005.6 | µg/mL | +/- 36.5875 |
| 57 | Hexachlorobenzene | 118-74-1 | 15828800 | 99% | 1,005.3 | µg/mL | +/- 36.5789 |
| 58 | Pentachlorophenol | 87-86-5 | RP240517RSR | 99% | 1,005.4 | µg/mL | +/- 36.5825 |
| 59 | Phenanthrene | 85-01-8 | MKCT3391 | 99% | 1,004.8 | µg/mL | +/- 36.5584 |
| 60 | Anthracene | 120-12-7 | MKCW9141 | 99% | 1,005.5 | µg/mL | +/- 36.5834 |
| 61 | Carbazole | 86-74-8 | 15630800 | 99% | 1,005.3 | µg/mL | +/- 36.5789 |
| 62 | Di-n-butylphthalate | 84-74-2 | MKCN4337 | 99% | 1,005.5 | µg/mL | +/- 36.5848 |
| 63 | Fluoranthene | 206-44-0 | A0458721 | 99% | 1,005.4 | µg/mL | +/- <0.0001 |
| 64 | Pyrene | 129-00-0 | BCCL8032 | 99% | 1,005.2 | µg/mL | +/- 36.5734 |
| 65 | Benzyl butyl phthalate | 85-68-7 | X12I018 | 99% | 1,004.8 | µg/mL | +/- 36.5584 |
| 66 | Bis(2-ethylhexyl)adipate | 103-23-1 | MKCR8567 | 99% | 1,004.5 | µg/mL | +/- 36.5480 |
| 67 | Benz(a)anthracene | 56-55-3 | I60012022BAA | 99% | 1,005.5 | µg/mL | +/- 36.5844 |
| 68 | Chrysene | 218-01-9 | RP241212RSR | 99% | 1,005.5 | µg/mL | +/- 36.5848 |
| 69 | Bis(2-ethylhexyl)phthalate | 117-81-7 | MKCS8065 | 99% | 1,004.7 | µg/mL | +/- 36.5548 |
| 70 | Di-n-octyl phthalate | 117-84-0 | 15817300 | 99% | 1,004.8 | µg/mL | +/- 36.5607 |
| 71 | Benzo(b)fluoranthene | 205-99-2 | 022013B | 99% | 1,005.1 | µg/mL | +/- 36.5716 |
| 72 | Benzo(k)fluoranthene | 207-08-9 | 012022K | 98% | 1,004.6 | µg/mL | +/- 36.5511 |
| 73 | Benzo(a)pyrene | 50-32-8 | NQLXA | 98% | 1,004.2 | µg/mL | +/- 36.5377 |
| 74 | Indeno(1,2,3-cd)pyrene | 193-39-5 | 12-JKL-118-9 | 97% | 1,004.1 | µg/mL | +/- 36.5337 |
| 75 | Dibenz(a,h)anthracene | 53-70-3 | 712061504-1-1 | 99% | 1,005.7 | µg/mL | +/- 36.5930 |
| 76 | Benzo(g,h,i)perylene | 191-24-2 | RP241014RSR | 98% | 1,005.4 | µg/mL | +/- 36.5814 |

* Expanded Uncertainty displayed in same units as Grav. Conc.

**Solvent:** Methylene chloride
**CAS #** 75-09-2
**Purity** 99%

**Tech Tips:**
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

# Quality Confirmation Test

**Column:**
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

**Carrier Gas:**
hydrogen-constant flow 1.8 mL/min.

**Temp. Program:**
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

**Inj. Temp:**
250°C

**Det. Temp:**
340°C

**Det. Type:**
FID

**Split Vent:**
100 ml/min.

**Inj. Vol**
1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

**Date Mixed:** 12-Jan-2025 **Balance Serial #** 1128360905

Dillan Murphy - Operations Technician I

**Date Passed:** 21-Jan-2025

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

# Certificate of Analysis

***gravimetric***

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555871 **Lot No.:** A0226283

S13158 ↓ S13167 } RC/ 6/4/25

**Description :** Custom 4-Nitrophenol Standard

Custom 4-Nitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** June 30, 2028 **Storage:** 10°C or colder

**Ship:** Ambient

**CERTIFIED VALUES**

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 4-Nitrophenol | 100-02-7 | 20241120-1-AN | 99% | 25,192.0 µg/mL | +/- 783.0517 |

**Solvent:** Methanol
**CAS #** 67-56-1
**Purity** 99%

Morgan Craighead - Mix Technician

**Date Mixed:** 02-Jun-2025 **Balance:** C322230531

REVIEWED By Dillon Murphy at 11:21 am, Jun 03, 2025

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***gravimetric***

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555871 **Lot No.:** A0226283

**Description :** Custom 4-Nitrophenol Standard

Custom 4-Nitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

S13158 } RC/ 6/4/25
↓
S13167 }

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** June 30, 2028 **Storage:** 10°C or colder

**Ship:** Ambient

**CERTIFIED VALUES**

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | 4-Nitrophenol | 100-02-7 | 20241120-1-AN | 99% | 25,192.0 µg/mL | +/- 783.0517 |

**Solvent:** Methanol
**CAS #** 67-56-1
**Purity** 99%

Morgan Craighead - Mix Technician

**Date Mixed:** 02-Jun-2025 **Balance:** C322230531

REVIEWED

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***gravimetric***

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

S13198 → S13207 AC 6/11/25

**Catalog No. :** 555868 **Lot No.:** A0226493

**Description :** Custom Benzidine Standard

Custom Benzidine Standard 25,000µg/mL, Methanol, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** June 30, 2028 **Storage:** 10°C or colder

**Handling:** Contains carcinogen/reproductive toxin. **Ship:** Ambient

**CERTIFIED VALUES**

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Benzidine | 92-87-5 | S250227ECS | 99% | 25,004.0 µg/mL | +/- 495.8040 |

**Solvent:** Methanol
**CAS #** 67-56-1
**Purity** 99%

Laith Clemente - Operations Technician I

**Date Mixed:** 09-Jun-2025 **Balance:** 1122030677

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

**Absolute Standards, Inc.**
800-368-1131
www.absolutestandards.com

*Certified Reference Material CRM*

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

## CERTIFIED WEIGHT REPORT

**Part Number:** 70434
Lot Number: 121622
Description: Acetophenone

Expiration Date: 121627
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 1000
NIST Test ID#: 6UTB

**Solvent(s):** **Lot#**
Methylene chloride C21F09CAS0000DCM

| | | |
|---|---|---|
| | | 121622 |
| Formulated By: | Benson Chan | DATE |
| | | 121622 |
| Reviewed By: | Pedro L. Rentas | DATE |

5E-05 Balance Uncertainty
Weight(s) shown below were combined and diluted to (mL): 50.0 0.001 Flask Uncertainty

| Compound | RM# | Lot Number | Nominal Conc (µg/mL) | Purity (%) | Uncertainty Purity | Target Weight(g) | Actual Weight(g) | Actual Conc (µg/mL) | Expanded Uncertainty (+/-) (µg/mL) | SDS Information (Solvent Safety Info. On Attached pg.) CAS# | OSHA PEL (TWA) | LD50 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1. Acetophenone | 434 | 04511JX | 1000 | 99 | 0.2 | 0.05055 | 0.05075 | 1004.0 | 4.5 | 98-86-2 | N/A | orl-rat 815mg/kg |

**Method:** GC6MSD-1. **Detector:** MSD. **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (10min.), Temp. 2 = 200°C (8.75 min.), Rate = 4°C/min., Injector Temp.= 200°C, Detector Temp. = 220°C. **Analyst:** Candice Warren. Solvent delay = 25 minutes.









S13208 ✓ RC 6/19/25

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

**Comments**
GC13-M1 Analysis by Melissa Stonier
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3




| Peak No. | Name | FID RT (min.) |
|---|---|---|
| 1 | Acetophenone | 43.05 |

**Absolute Standards, Inc.**
800-368-1131
www.absolutestandards.com

*Certified Reference Material CRM*

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

| | | | |
|---|---|---|---|
| **Part Number:** | **91126** | **Solvent(s):** | **Lot#** |
| Lot Number: | 060625 | Methylene chloride | 23343 |
| Description: | Benzoic acid | | |
| Expiration Date: | 060630 | | |
| Recommended Storage: | Refrigerate (4 °C) | | |
| Nominal Concentration (µg/mL): | 2000 | | |
| NIST Test ID#: | 6UTB | 5E-05 Balance Uncertainty | |
| Weight(s) shown below were combined and diluted to (mL): | 100.0 | 0.031 Flask Uncertainty | |

| | | |
|---|---|---|
| | | 060625 |
| Formulated By: | Anthony Mahoney | DATE |
| | | 060625 |
| Reviewed By: | Pedro L. Rentas | DATE |

| Compound | RM# | Lot Number | Nominal Conc (µg/mL) | Purity (%) | Uncertainty Purity | Target Weight(g) | Actual Weight(g) | Actual Conc (µg/mL) | Expanded Uncertainty (+/-) (µg/mL) | SDS Information (Solvent Safety Info. On Attached pg.) CAS# | OSHA PEL (TWA) | LD50 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1. Benzoic acid | 34 | 11930AW | 2000.0 | 99 | 0.2 | 0.20211 | 0.20227 | 2001.6 | 8.2 | 65-85-0 | N/A | orl-rat 1700mg/kg |

**Method GC8MSD-3.M:** Column:SPB-5 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 50°C (1min.), Temp 2 = 300°C (9min.), Rate = 10°C/min., Injector B= 200°C. Detector B = 275°C, Split Ratio = 100:1, Scan Rate = 2. Analysis performed by: Melissa Stonier.







S132097 RC 6/19/25

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Rev 1.0, 2/25/2025

**Absolute Standards, Inc.**
800-368-1131
www.absolutestandards.com

*Certified Reference Material CRM*

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

## CERTIFIED WEIGHT REPORT

**Part Number:** 34571
Lot Number: 092324
Description: 1,2,4,5-Tetrachlorobenzene

**Solvent(s):** Methanol **Lot#** EI863-US

Expiration Date: 092334
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

5E-05 Balance Uncertainty
0.004 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 25.0

**CAUTION: Sonicate Before Use**

| Formulated By: | Lawrence Barry | 092324 DATE |
|---|---|---|
| **Reviewed By:** | Pedro L. Rentas | 092324 DATE |

| Compound | RM# | Lot Number | Nominal Conc (µg/mL) | Purity (%) | Uncertainty Purity | Target Weight (g) | Actual Weight (g) | Actual Conc(µg/mL) | Expanded Uncertainty (+/-) (µg/mL) | SDS Information (Solvent Safety Info. On Attached pg.) CAS# | OSHA PEL (TWA) | LD50 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1. 1,2,4,5-Tetrachlorobenzene | 274 | 10408AS | 5000 | 98 | 0.20 | 0.12761 | 0.12765 | 5001.4 | 20.8 | 95-94-3 | N/A | orl-rat 1500mg/kg |

**Method GC8MSD-3.M:** Column:SPB-5 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 50°C (1min.), Temp 2 = 300°C (9min.), Rate = 10°C/min., Injector B= 200°C, Detector B = 275°C, Split Ratio = 100:1, Scan Rate = 2. Analysis performed by: Gina McLane.










S132113 RC 6/19/25.

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

# Run 31, "P34571 L092324 [5000µg/mL in MeOH]"

Run Length: 35.00 min, 20999 points at 10 points/second.
Created: Fri, Jan 3, 2025 at 11:24:40 AM.
Sampled: Sequence "010225-GC4M1", Method "GC4-M1".
Analyzed using Method "GC4-M1".

## Comments

GC4-M1 Analysis by Melissa Stonier
Column ID SPB5 L#60062-01A : 30 meter x 0.53mm x 1.5um Film Thickness
Flow rates; Total Flow = 300 ml/min, Helium (carrier) = 6.5 mL, Helium (make-up) = 25 mL, Hydrogen (detector) = 30 mL, Air (detector) =360 mL
Oven Temp 1 = 50°C (1 min), Rate = 10°C/min, Oven Temp 2 = 300°C (9 min), Total Run Time = 35 Minutes.
Injector Temp = 200°C, FID Temp = 300°C, FID Signal = eDaq Channel 1.
Gas Chromatograph = HP 5890, Auto Sampler = HP 7673, Standard Injection = 0.5 uL, Range = 7




| Peak No. | Name | FID RT (min.) |
|---|---|---|
| 1 | 1,2,4,5-Tetrachlorobenzene | 13.24 |

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

# Certificate of Analysis

***gravimetric***

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

**FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.**

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 555869 **Lot No.:** A0201702

**Description :** Custom Hexachlorocyclopentadiene Standard

Custom Hexachlorocyclopentadiene Standard 25,000µg/mL, Methanol, 1mL/ampul

513148 ↓ 513157 RC 11/13/23

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** September 30, 2026 **Storage:** 10°C or colder

**Ship:** Ambient

CERTIFIED VALUES

| Component # | Compound | CAS # | Lot # | Purity | Grav. Conc. (weight/volume) | Expanded Uncertainty * (95% C.L.; K=2) |
|---|---|---|---|---|---|---|
| 1 | Hexachlorocyclopentadiene | 77-47-4 | 099063P13G | 99% | 25,244.0 µg/mL | +/- 450.6896 |

**Solvent:** Methanol
**CAS #** 67-56-1
**Purity** 99%

Brittany Federinko - Operations Tech I

**Date Mixed:** 05-Sep-2023 **Balance:** B707717271

**Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397**

# SHIPPING DOCUMENTS

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

ALLIANCE PROJECT NO. Q 2592
QUOTE NO.
COC Number 2047546

## CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: Parsons
ADDRESS: 301 Plainfield Road
CITY Syracuse STATE: NY ZIP: 13212
ATTENTION: Zohar Lavy
PHONE: (732)796-5536 FAX: —

## CLIENT PROJECT INFORMATION

PROJECT NAME: ConEd East River S1
PROJECT NO.: 454534 LOCATION: Manhattan
PROJECT MANAGER: Zohar Lavy
e-mail: Zohar.lavy@parsons.com
PHONE: (732)-796-5536 FAX: —

## CLIENT BILLING INFORMATION

BILL TO: Parsons PO#: 454534
ADDRESS: 301 Plainfield Road
CITY Syracuse STATE: NY ZIP: 13212
ATTENTION: Zohar Lavy PHONE: (732)-796-5536

## DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): 5-day rush DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

## DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC + Raw Data) ☐ NYS ASP A ☐ NYS ASP B
☐ Other _____
☐ EDD FORMAT _____

## ANALYSIS

1. VOCs, SVOCs
2. TAL metals
3. PCBs, TPH
4. TCLP VOCs, TCLP SVOCs
5. TCLP metals
6. TCLP pest, TCLP herb
7. Ignitability, reactivity
8. corrosivity, pH
9.

| ALLIANCE SAMPLE ID | PROJECT SAMPLE IDENTIFICATION | SAMPLE MATRIX | SAMPLE TYPE COMP | SAMPLE TYPE GRAB | SAMPLE COLLECTION DATE | SAMPLE COLLECTION TIME | # OF BOTTLES | PRESERVATIVES 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | COMMENTS |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| | | | | | | | | E | E | E | E | E | E | E | E | | ← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER |
| 1. | WC-Soil-20250711 | S | X | | 7/11/25 | 1230 | 9 | X | X | X | X | X | X | X | X | | |
| 2. | | | | | | | | | | | | | | | | | |
| 3. | | | | | | | | | | | | | | | | | |
| 4. | | | | | | | | | | | | | | | | | |
| 5. | | | | | | | | | | | | | | | | | |
| 6. | | | | | | | | | | | | | | | | | |
| 7. | | | | | | | | | | | | | | | | | |
| 8. | | | | | | | | | | | | | | | | | |
| 9. | | | | | | | | | | | | | | | | | |
| 10. | | | | | | | | | | | | | | | | | |

## SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

| RELINQUISHED BY SAMPLER: | DATE/TIME: | RECEIVED BY: |
|---|---|---|
| 1. Emma Sauer | 7/11/25 1440 | 1. 7-11-25 1440 |
| 2. | | 2. |
| 3. | 7-11-25 1637 | 3. |

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 3.1 °C
Comments: Include Zohar.lavy@parsons.com and Kirsten.valentini@parsons.com on all data

Page ____ of

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete ☐ YES ☐ NO

Copyright © 2024

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT YELLOW - ALLIANCE COPY PINK - SAMPLER COPY

284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

# Laboratory Certification

| Certified By | License No. |
|---|---|
| | |
| CAS EPA CLP Contract | 68HERH20D0011 |
| | |
| Connecticut | PH-0830 |
| | |
| DOD ELAP (ANAB) | L2219 |
| | |
| Maine | 2024021 |
| | |
| Maryland | 296 |
| | |
| New Hampshire | 255424 Rev 1 |
| | |
| New Jersey | 20012 |
| | |
| New York | 11376 |
| | |
| Pennsylvania | 68-00548 |
| | |
| Soil Permit | 525-24-234-08441 |
| | |
| Texas | T104704488 |

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

# LOGIN REPORT/SAMPLE TRANSFER

**Order ID :** Q2592 PARS02
**Client Name :** PARSONS Engineering of N
**Client Contact :** Zohar Lavy
**Invoice Name :** PARSONS Engineering of N
**Invoice Contact :** Zohar Lavy

**Order Date :** 7/11/2025 3:05:25 PM
**Project Name :** Con Edison - East River Sit
**Receive DateTime :** 7/11/2025 ~~12:00:00 AM~~ 04:39:00 PM
**Purchase Order :**

**Project Mgr :**
**Report Type :** NYS ASP B
**EDD Type :** NYSDEC EDD V-3
**Hard Copy Date :**
**Date Signoff :**

| LAB ID | CLIENT ID | MATRIX | SAMPLE DATE | SAMPLE TIME | TEST | TEST GROUP | METHOD | FAX DATE | DUE DATES |
|---|---|---|---|---|---|---|---|---|---|
| Q2592-01 | WC-SOIL-20250711 | Solid | 07/11/2025 | 12:30 | | | | | |
| | | | | | VOCMS Group1 | | 8260D | | 5 Bus. Days |

DP 07/15/2025

**Relinguished By :**
**Date / Time :** 7/14/25 0725

SAMPLES RECEIVED ON 7/11/25 @ N40
SAMPLES PLACED IN SM-REF-2

**Received By :** Sam
**Date / Time :** 07/14/25 10:00

Ngt 6
FZ2

**Storage Area :** VOA Refridgerator Room