

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2592

Client ID : SSTDCCC040

Date Analyzed: 07/16/2025

Lab File ID: BF143117.D

Time Analyzed: 19:39

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	135149	6.969	510164	8.25	254065	10.00
UPPER LIMIT	270298	7.469	1020330	8.751	508130	10.504
LOWER LIMIT	67574.5	6.469	255082	7.751	127033	9.504
EPA SAMPLE NO.						
01 PB168854BS	139596	6.97	538805	8.25	272708	10.00
02 PB168854BL	138731	6.96	528830	8.25	273109	10.00

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.