

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143092.D
Instrument ID: BNA_F

Contract: PARS02
SDG NO.: Q2592
DFTPP Injection Date: 07/15/2025
DFTPP Injection Time: 12:35

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (2) 1
69	Mass 69 relative abundance	29.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.1
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	15
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.4 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF143093.D	07/15/2025	13:04
SSTDICC005	SSTDICC005	BF143094.D	07/15/2025	13:35
SSTDICC010	SSTDICC010	BF143095.D	07/15/2025	14:05
SSTDICC020	SSTDICC020	BF143096.D	07/15/2025	14:36
SSTDICCC040	SSTDICCC040	BF143097.D	07/15/2025	15:05
SSTDICC050	SSTDICC050	BF143098.D	07/15/2025	15:35
SSTDICC060	SSTDICC060	BF143099.D	07/15/2025	16:05
SSTDICC080	SSTDICC080	BF143100.D	07/15/2025	16:35