

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025186.D
Instrument ID: BNA_P

Contract: PARS02
SDG NO.: Q2592
DFTPP Injection Date: 07/18/2025
DFTPP Injection Time: 09:41

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.4 (1.8) 1
69	Mass 69 relative abundance	23.7
70	Less than 2.0% of mass 69	0.1 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.2
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	15.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.2 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP025187.D	07/18/2025	11:04
CHRT-26899MS	Q2597-01MS	BP025191.D	07/18/2025	13:52
CHRT-26899MSD	Q2597-01MSD	BP025192.D	07/18/2025	14:33