

## Report of Analysis

|                    |                                       |                  |                 |          |      |
|--------------------|---------------------------------------|------------------|-----------------|----------|------|
| Client:            | PARSONS Engineering of New York, Inc. |                  | Date Collected: | 07/11/25 |      |
| Project:           | Con Edison - East River Site 2        |                  | Date Received:  | 07/11/25 |      |
| Client Sample ID:  | WC-SOIL-20250711MSD                   |                  | SDG No.:        | Q2592    |      |
| Lab Sample ID:     | Q2592-02MSD                           |                  | Matrix:         | TCLP     |      |
| Analytical Method: | 8270E                                 |                  | % Solid:        | 0        |      |
| Sample Wt/Vol:     | 100                                   | Units: mL        | Final Vol:      | 1000     | uL   |
| Soil Aliquot Vol:  |                                       | uL               | Test:           | TCLP BNA |      |
| Extraction Type :  |                                       | Decanted : N     | Level :         | LOW      |      |
| Injection Volume : |                                       | GPC Factor : 1.0 | GPC Cleanup :   | N        | PH : |
| Prep Method :      |                                       |                  |                 |          |      |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| BF143123.D        | 1         | 07/16/25 10:14 | 07/16/25 22:38 | PB168885      |

| CAS Number                | Parameter              | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units    |
|---------------------------|------------------------|--------|-----------|----------|------------|----------|
| <b>TARGETS</b>            |                        |        |           |          |            |          |
| 110-86-1                  | Pyridine               | 290    |           | 12.8     | 50.0       | ug/L     |
| 106-46-7                  | 1,4-Dichlorobenzene    | 360    |           | 5.30     | 50.0       | ug/L     |
| 95-48-7                   | 2-Methylphenol         | 400    |           | 11.2     | 50.0       | ug/L     |
| 65794-96-9                | 3+4-Methylphenols      | 400    |           | 11.0     | 100        | ug/L     |
| 67-72-1                   | Hexachloroethane       | 380    |           | 6.50     | 50.0       | ug/L     |
| 98-95-3                   | Nitrobenzene           | 430    |           | 7.60     | 50.0       | ug/L     |
| 87-68-3                   | Hexachlorobutadiene    | 390    |           | 5.40     | 50.0       | ug/L     |
| 88-06-2                   | 2,4,6-Trichlorophenol  | 440    |           | 5.10     | 50.0       | ug/L     |
| 95-95-4                   | 2,4,5-Trichlorophenol  | 430    |           | 6.20     | 50.0       | ug/L     |
| 121-14-2                  | 2,4-Dinitrotoluene     | 470    |           | 12.2     | 50.0       | ug/L     |
| 118-74-1                  | Hexachlorobenzene      | 440    |           | 5.20     | 50.0       | ug/L     |
| 87-86-5                   | Pentachlorophenol      | 710    |           | 15.8     | 100        | ug/L     |
| <b>SURROGATES</b>         |                        |        |           |          |            |          |
| 367-12-4                  | 2-Fluorophenol         | 100    |           | 23 - 138 | 67%        | SPK: 150 |
| 13127-88-3                | Phenol-d6              | 96.3   |           | 10 - 134 | 64%        | SPK: 150 |
| 4165-60-0                 | Nitrobenzene-d5        | 82.6   |           | 67 - 132 | 83%        | SPK: 100 |
| 321-60-8                  | 2-Fluorobiphenyl       | 68.6   |           | 52 - 132 | 69%        | SPK: 100 |
| 118-79-6                  | 2,4,6-Tribromophenol   | 127    |           | 44 - 137 | 85%        | SPK: 150 |
| 1718-51-0                 | Terphenyl-d14          | 68.7   |           | 42 - 152 | 69%        | SPK: 100 |
| <b>INTERNAL STANDARDS</b> |                        |        |           |          |            |          |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4 | 155000 |           | 6.969    |            |          |
| 1146-65-2                 | Naphthalene-d8         | 590000 |           | 8.245    |            |          |
| 15067-26-2                | Acenaphthene-d10       | 297000 |           | 10.004   |            |          |
| 1517-22-2                 | Phenanthrene-d10       | 462000 |           | 11.492   |            |          |
| 1719-03-5                 | Chrysene-d12           | 243000 |           | 14.127   |            |          |
| 1520-96-3                 | Perylene-d12           | 288000 |           | 15.633   |            |          |

## Report of Analysis

|                    |                                       |                  |                 |          |      |
|--------------------|---------------------------------------|------------------|-----------------|----------|------|
| Client:            | PARSONS Engineering of New York, Inc. |                  | Date Collected: | 07/11/25 |      |
| Project:           | Con Edison - East River Site 2        |                  | Date Received:  | 07/11/25 |      |
| Client Sample ID:  | WC-SOIL-20250711MSD                   |                  | SDG No.:        | Q2592    |      |
| Lab Sample ID:     | Q2592-02MSD                           |                  | Matrix:         | TCLP     |      |
| Analytical Method: | 8270E                                 |                  | % Solid:        | 0        |      |
| Sample Wt/Vol:     | 100                                   | Units: mL        | Final Vol:      | 1000     | uL   |
| Soil Aliquot Vol:  |                                       | uL               | Test:           | TCLP BNA |      |
| Extraction Type :  |                                       | Decanted : N     | Level :         | LOW      |      |
| Injection Volume : |                                       | GPC Factor : 1.0 | GPC Cleanup :   | N        | PH : |
| Prep Method :      |                                       |                  |                 |          |      |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| BF143123.D        | 1         | 07/16/25 10:14 | 07/16/25 22:38 | PB168885      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products