

LAB CHRONICLE

OrderID:	Q2592	OrderDate:	7/11/2025 3:05:25 PM
Client:	PARSONS Engineering of New York, Inc.	Project:	Con Edison - East River Site 2
Contact:	Zohar Lavy	Location:	D51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2592-01	WC-SOIL-20250711	SOIL	SVOCMS Group1	8270E	07/11/25	07/15/25	07/16/25	07/11/25
Q2592-02	WC-SOIL-20250711	TCLP	TCLP BNA	8270E	07/11/25	07/16/25	07/16/25	07/11/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2592
Client: PARSONS Engineering of New York, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				0.000				
Total Svoc :					0.00			
Total Concentration:					0.00			



SAMPLE DATA

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	07/16/25	
Project:	Con Edison - East River Site 2		Date Received:	07/16/25	
Client Sample ID:	PB168847TB		SDG No.:	Q2592	
Lab Sample ID:	PB168847TB		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143120.D	1	07/16/25 10:14	07/16/25 21:09	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	50.0	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	50.0	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	50.0	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	100	U	11.0	100	ug/L
67-72-1	Hexachloroethane	50.0	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	50.0	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	50.0	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	50.0	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	50.0	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	50.0	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	50.0	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	100	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	115		23 - 138	77%	SPK: 150
13127-88-3	Phenol-d6	114		10 - 134	76%	SPK: 150
4165-60-0	Nitrobenzene-d5	85.0		67 - 132	85%	SPK: 100
321-60-8	2-Fluorobiphenyl	73.8		52 - 132	74%	SPK: 100
118-79-6	2,4,6-Tribromophenol	134		44 - 137	90%	SPK: 150
1718-51-0	Terphenyl-d14	84.3		42 - 152	84%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	136000	6.963			
1146-65-2	Naphthalene-d8	521000	8.245			
15067-26-2	Acenaphthene-d10	267000	10.004			
1517-22-2	Phenanthrene-d10	451000	11.486			
1719-03-5	Chrysene-d12	236000	14.127			
1520-96-3	Perylene-d12	211000	15.633			

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	07/16/25	
Project:	Con Edison - East River Site 2		Date Received:	07/16/25	
Client Sample ID:	PB168847TB		SDG No.:	Q2592	
Lab Sample ID:	PB168847TB		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143120.D	1	07/16/25 10:14	07/16/25 21:09	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	07/11/25	
Project:	Con Edison - East River Site 2		Date Received:	07/11/25	
Client Sample ID:	WC-SOIL-20250711		SDG No.:	Q2592	
Lab Sample ID:	Q2592-02		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143121.D	1	07/16/25 10:14	07/16/25 21:39	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	50.0	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	50.0	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	50.0	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	100	U	11.0	100	ug/L
67-72-1	Hexachloroethane	50.0	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	50.0	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	50.0	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	50.0	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	50.0	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	50.0	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	50.0	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	100	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	103		23 - 138	68%	SPK: 150
13127-88-3	Phenol-d6	96.1		10 - 134	64%	SPK: 150
4165-60-0	Nitrobenzene-d5	81.1		67 - 132	81%	SPK: 100
321-60-8	2-Fluorobiphenyl	69.8		52 - 132	70%	SPK: 100
118-79-6	2,4,6-Tribromophenol	136		44 - 137	91%	SPK: 150
1718-51-0	Terphenyl-d14	74.4		42 - 152	74%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	137000	6.969			
1146-65-2	Naphthalene-d8	518000	8.245			
15067-26-2	Acenaphthene-d10	266000	9.998			
1517-22-2	Phenanthrene-d10	452000	11.486			
1719-03-5	Chrysene-d12	249000	14.127			
1520-96-3	Perylene-d12	227000	15.633			

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	07/11/25
Project:	Con Edison - East River Site 2		Date Received:	07/11/25
Client Sample ID:	WC-SOIL-20250711		SDG No.:	Q2592
Lab Sample ID:	Q2592-02		Matrix:	TCLP
Analytical Method:	8270E		% Solid:	0
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143121.D	1	07/16/25 10:14	07/16/25 21:39	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q2592

Client: PARSONS Engineering of New York, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168847TB	PB168847TB	2-Fluorophenol	150	115	77		23	138
		Phenol-d6	150	114	76		10	134
		Nitrobenzene-d5	100	85.0	85		67	132
		2-Fluorobiphenyl	100	73.8	74		52	132
		2,4,6-Tribromophenol	150	134	90		44	137
		Terphenyl-d14	100	84.3	84		42	152
PB168885BL	PB168885BL	2-Fluorophenol	150	133	89		23	138
		Phenol-d6	150	129	86		10	134
		Nitrobenzene-d5	100	82.7	83		67	132
		2-Fluorobiphenyl	100	80.1	80		52	132
		2,4,6-Tribromophenol	150	157	105		44	137
		Terphenyl-d14	100	86.6	87		42	152
PB168885BS	PB168885BS	2-Fluorophenol	150	129	86		23	138
		Phenol-d6	150	130	87		10	134
		Nitrobenzene-d5	100	76.8	77		67	132
		2-Fluorobiphenyl	100	73.5	74		52	132
		2,4,6-Tribromophenol	150	150	100		44	137
		Terphenyl-d14	100	76.8	77		42	152
Q2592-02	WC-SOIL-20250711	2-Fluorophenol	150	103	68		23	138
		Phenol-d6	150	96.1	64		10	134
		Nitrobenzene-d5	100	81.1	81		67	132
		2-Fluorobiphenyl	100	69.8	70		52	132
		2,4,6-Tribromophenol	150	136	91		44	137
		Terphenyl-d14	100	74.4	74		42	152
Q2592-02MS	WC-SOIL-20250711MS	2-Fluorophenol	150	99.9	67		23	138
		Phenol-d6	150	95.0	63		10	134
		Nitrobenzene-d5	100	80.6	81		67	132
		2-Fluorobiphenyl	100	68.4	68		52	132
		2,4,6-Tribromophenol	150	126	84		44	137
		Terphenyl-d14	100	69.6	70		42	152
Q2592-02MSD	WC-SOIL-20250711MSD	2-Fluorophenol	150	100	67		23	138
		Phenol-d6	150	96.3	64		10	134
		Nitrobenzene-d5	100	82.6	83		67	132
		2-Fluorobiphenyl	100	68.6	69		52	132
		2,4,6-Tribromophenol	150	127	85		44	137
		Terphenyl-d14	100	68.7	69		42	152

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2592

Analytical Method: SW8270E

Client: PARSONS Engineering of New York, Inc.

DataFile: BF143122.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q2592-02MS		Client Sample ID: WC-SOIL-20250711MS									
Pyridine	500	0	290	ug/L	58				10	109	
1,4-Dichlorobenzene	500	0	360	ug/L	72				55	125	
2-Methylphenol	500	0	390	ug/L	78				60	131	
3+4-Methylphenols	500	0	390	ug/L	78				54	136	
Hexachloroethane	500	0	370	ug/L	74				19	146	
Nitrobenzene	500	0	430	ug/L	86				62	112	
Hexachlorobutadiene	500	0	390	ug/L	78				52	125	
2,4,6-Trichlorophenol	500	0	440	ug/L	88				78	112	
2,4,5-Trichlorophenol	500	0	430	ug/L	86				71	111	
2,4-Dinitrotoluene	500	0	460	ug/L	92				74	137	
Hexachlorobenzene	500	0	430	ug/L	86				72	115	
Pentachlorophenol	1000	0	750	ug/L	75				52	162	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2592

Analytical Method: SW8270E

Client: PARSONS Engineering of New York, Inc.

DataFile: BF143123.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q2592-02MSD		Client Sample ID: WC-SOIL-20250711MSD									
Pyridine	500	0	290	ug/L	58	0			10	109	20
1,4-Dichlorobenzene	500	0	360	ug/L	72	0			55	125	20
2-Methylphenol	500	0	400	ug/L	80	3			60	131	20
3+4-Methylphenols	500	0	400	ug/L	80	3			54	136	20
Hexachloroethane	500	0	380	ug/L	76	3			19	146	20
Nitrobenzene	500	0	430	ug/L	86	0			62	112	20
Hexachlorobutadiene	500	0	390	ug/L	78	0			52	125	20
2,4,6-Trichlorophenol	500	0	440	ug/L	88	0			78	112	20
2,4,5-Trichlorophenol	500	0	430	ug/L	86	0			71	111	20
2,4-Dinitrotoluene	500	0	470	ug/L	94	2			74	137	20
Hexachlorobenzene	500	0	440	ug/L	88	2			72	115	20
Pentachlorophenol	1000	0	710	ug/L	71	5			52	162	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2592</u>	Analytical Method:	<u>8270E</u>
Client:	<u>PARSONS Engineering of New York, Inc.</u>	DataFile:	<u>BM050476.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168885BS	Pyridine	50	39.1	ug/L	78				29	97		
	1,4-Dichlorobenzene	50	43.0	ug/L	86				76	103		
	2-Methylphenol	50	47.4	ug/L	95				69	109		
	3+4-Methylphenols	50	47.8	ug/L	96				67	106		
	Hexachloroethane	50	43.1	ug/L	86				76	118		
	Nitrobenzene	50	44.3	ug/L	89				58	106		
	Hexachlorobutadiene	50	44.8	ug/L	90				69	101		
	2,4,6-Trichlorophenol	50	50.9	ug/L	102				61	110		
	2,4,5-Trichlorophenol	50	51.7	ug/L	103				70	106		
	2,4-Dinitrotoluene	50	49.2	ug/L	98				60	115		
	Hexachlorobenzene	50	47.8	ug/L	96				73	106		
	Pentachlorophenol	100	110	ug/L	110				47	114		

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID

PB168885BL

Lab Name: Alliance

Contract: PARS02

Lab Code: ACE

SDG NO.: Q2592

Lab File ID: BM050475.D

Lab Sample ID: PB168885BL

Instrument ID: BNA_M

Date Extracted: 07/16/2025

Matrix: (soil/water) water

Date Analyzed: 07/17/2025

Level: (low/med) LOW

Time Analyzed: 14:57

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB168885BS	PB168885BS	BM050476.D	07/17/2025
PB168847TB	PB168847TB	BF143120.D	07/16/2025
WC-SOIL-20250711	Q2592-02	BF143121.D	07/16/2025
WC-SOIL-20250711MS	Q2592-02MS	BF143122.D	07/16/2025
WC-SOIL-20250711MSD	Q2592-02MSD	BF143123.D	07/16/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143092.D
Instrument ID: BNA_F

Contract: PARS02
SDG NO.: Q2592
DFTPP Injection Date: 07/15/2025
DFTPP Injection Time: 12:35

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (2) 1
69	Mass 69 relative abundance	29.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.1
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	15
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.4 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF143093.D	07/15/2025	13:04
SSTDICC005	SSTDICC005	BF143094.D	07/15/2025	13:35
SSTDICC010	SSTDICC010	BF143095.D	07/15/2025	14:05
SSTDICC020	SSTDICC020	BF143096.D	07/15/2025	14:36
SSTDICCC040	SSTDICCC040	BF143097.D	07/15/2025	15:05
SSTDICC050	SSTDICC050	BF143098.D	07/15/2025	15:35
SSTDICC060	SSTDICC060	BF143099.D	07/15/2025	16:05
SSTDICC080	SSTDICC080	BF143100.D	07/15/2025	16:35

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143116.D
Instrument ID: BNA_F

Contract: PARS02
SDG NO.: Q2592
DFTPP Injection Date: 07/16/2025
DFTPP Injection Time: 19:09

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.9) 1
69	Mass 69 relative abundance	31.5
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.5
365	Greater than 1% of mass 198	2.9
441	Present, but less than mass 443	15.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF143117.D	07/16/2025	19:39
PB168847TB	PB168847TB	BF143120.D	07/16/2025	21:09
WC-SOIL-20250711	Q2592-02	BF143121.D	07/16/2025	21:39
WC-SOIL-20250711MS	Q2592-02MS	BF143122.D	07/16/2025	22:09
WC-SOIL-20250711MSD	Q2592-02MSD	BF143123.D	07/16/2025	22:38

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BM050376.D
Instrument ID: BNA_M

Contract: PARS02
SDG NO.: Q2592
DFTPP Injection Date: 07/08/2025
DFTPP Injection Time: 11:59

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.5) 1
69	Mass 69 relative abundance	33.9
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	10.5
442	Greater than 50% of mass 198	66.5
443	15.0 - 24.0% of mass 442	12.9 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BM050377.D	07/08/2025	12:39
SSTDICC005	SSTDICC005	BM050378.D	07/08/2025	13:19
SSTDICC010	SSTDICC010	BM050379.D	07/08/2025	14:00
SSTDICC020	SSTDICC020	BM050380.D	07/08/2025	14:40
SSTDICCC040	SSTDICCC040	BM050381.D	07/08/2025	15:20
SSTDICC050	SSTDICC050	BM050382.D	07/08/2025	16:01
SSTDICC060	SSTDICC060	BM050383.D	07/08/2025	16:41
SSTDICC080	SSTDICC080	BM050384.D	07/08/2025	17:22

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BM050473.D
Instrument ID: BNA_M

Contract: PARS02
SDG NO.: Q2592
DFTPP Injection Date: 07/17/2025
DFTPP Injection Time: 13:02

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.4 (1.5) 1
69	Mass 69 relative abundance	25.5
70	Less than 2.0% of mass 69	0.1 (0.5) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	13.3
442	Greater than 50% of mass 198	85.5
443	15.0 - 24.0% of mass 442	16.6 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050474.D	07/17/2025	13:42
PB168885BL	PB168885BL	BM050475.D	07/17/2025	14:57
PB168885BS	PB168885BS	BM050476.D	07/17/2025	15:57

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2592

Client ID : SSTDCCC040

Date Analyzed: 07/16/2025

Lab File ID: BF143117.D

Time Analyzed: 19:39

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	135149	6.969	510164	8.25	254065	10.00
UPPER LIMIT	270298	7.469	1020330	8.751	508130	10.504
LOWER LIMIT	67574.5	6.469	255082	7.751	127033	9.504
EPA SAMPLE NO.						
01 WC-SOIL-20250711	137278	6.97	517786	8.25	265739	10.00
02 WC-SOIL-20250711MS	147018	6.97	556657	8.25	279575	10.00
03 WC-SOIL-20250711MSD	155046	6.97	589623	8.25	296871	10.00
04 PB168847TB	135526	6.96	521315	8.25	267383	10.00

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2592

Client ID: SSTDCCC040

Date Analyzed: 07/16/2025

Lab File ID: BF143117.D

Time Analyzed: 19:39

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	397173	11.492	210714	14.127	255524	15.633
UPPER LIMIT	794346	11.992	421428	14.627	511048	16.133
LOWER LIMIT	198587	10.992	105357	13.627	127762	15.133
EPA SAMPLE NO.						
01 WC-SOIL-20250711	451811	11.49	249471	14.13	227071	15.63
02 WC-SOIL-20250711MS	442443	11.49	237238	14.13	283551	15.63
03 WC-SOIL-20250711MSD	461806	11.49	242870	14.13	288284	15.63
04 PB168847TB	451109	11.49	235930	14.13	210500	15.63

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2592

Client ID : SSTDCCC040

Date Analyzed: 07/17/2025

Lab File ID: BM050474.D

Time Analyzed: 13:42

Instrument ID: BNA_M

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	523700	7.851	1980060	10.64	1283320	14.47
UPPER LIMIT	1047400	8.351	3960120	11.139	2566640	14.974
LOWER LIMIT	261850	7.351	990030	10.139	641660	13.974
EPA SAMPLE NO.						
01 PB168885BL	484190	7.85	1695350	10.64	1085070	14.47
02 PB168885BS	516385	7.85	1948200	10.64	1313340	14.48

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2592

Client ID: SSTDCCC040

Date Analyzed: 07/17/2025

Lab File ID: BM050474.D

Time Analyzed: 13:42

Instrument ID: BNA_M

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	2540660	17.209	2672210	21.433	2895080	24.456
UPPER LIMIT	5081320	17.709	5344420	21.933	5790160	24.956
LOWER LIMIT	1270330	16.709	1336110	20.933	1447540	23.956
EPA SAMPLE NO.						
01 PB168885BL	2246800	17.20	2374390	21.43	2623120	24.45
02 PB168885BS	2583210	17.21	2737580	21.43	2912640	24.46

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Edison - East River Site 2		Date Received:	
Client Sample ID:	PB168885BL		SDG No.:	Q2592
Lab Sample ID:	PB168885BL		Matrix:	TCLP
Analytical Method:	8270E		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050475.D	1	07/16/25 10:14	07/17/25 14:57	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	5.00	U	1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	5.00	U	0.53	5.00	ug/L
95-48-7	2-Methylphenol	5.00	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	10.0	U	1.10	10.0	ug/L
67-72-1	Hexachloroethane	5.00	U	0.65	5.00	ug/L
98-95-3	Nitrobenzene	5.00	U	0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	5.00	U	0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	5.00	U	0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	5.00	U	0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	5.00	U	1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	5.00	U	0.52	5.00	ug/L
87-86-5	Pentachlorophenol	10.0	U	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	133		23 - 138	89%	SPK: 150
13127-88-3	Phenol-d6	129		10 - 134	86%	SPK: 150
4165-60-0	Nitrobenzene-d5	82.7		67 - 132	83%	SPK: 100
321-60-8	2-Fluorobiphenyl	80.1		52 - 132	80%	SPK: 100
118-79-6	2,4,6-Tribromophenol	157		44 - 137	105%	SPK: 150
1718-51-0	Terphenyl-d14	86.6		42 - 152	87%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	484000	7.845			
1146-65-2	Naphthalene-d8	1700000	10.639			
15067-26-2	Acenaphthene-d10	1090000	14.474			
1517-22-2	Phenanthrene-d10	2250000	17.203			
1719-03-5	Chrysene-d12	2370000	21.427			
1520-96-3	Perylene-d12	2620000	24.45			

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Edison - East River Site 2		Date Received:	
Client Sample ID:	PB168885BL		SDG No.:	Q2592
Lab Sample ID:	PB168885BL		Matrix:	TCLP
Analytical Method:	8270E		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050475.D	1	07/16/25 10:14	07/17/25 14:57	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Edison - East River Site 2		Date Received:	
Client Sample ID:	PB168885BS		SDG No.:	Q2592
Lab Sample ID:	PB168885BS		Matrix:	TCLP
Analytical Method:	8270E		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050476.D	1	07/16/25 10:14	07/17/25 15:57	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	39.1		1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	43.0		0.53	5.00	ug/L
95-48-7	2-Methylphenol	47.4		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	47.8		1.10	10.0	ug/L
67-72-1	Hexachloroethane	43.1		0.65	5.00	ug/L
98-95-3	Nitrobenzene	44.3		0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	44.8		0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	50.9		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	51.7		0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	49.2		1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	47.8		0.52	5.00	ug/L
87-86-5	Pentachlorophenol	110	E	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	129		23 - 138	86%	SPK: 150
13127-88-3	Phenol-d6	130		10 - 134	87%	SPK: 150
4165-60-0	Nitrobenzene-d5	76.8		67 - 132	77%	SPK: 100
321-60-8	2-Fluorobiphenyl	73.5		52 - 132	74%	SPK: 100
118-79-6	2,4,6-Tribromophenol	150		44 - 137	100%	SPK: 150
1718-51-0	Terphenyl-d14	76.8		42 - 152	77%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	516000		7.846		
1146-65-2	Naphthalene-d8	1950000		10.64		
15067-26-2	Acenaphthene-d10	1310000		14.475		
1517-22-2	Phenanthrene-d10	2580000		17.21		
1719-03-5	Chrysene-d12	2740000		21.433		
1520-96-3	Perylene-d12	2910000		24.462		

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Edison - East River Site 2		Date Received:	
Client Sample ID:	PB168885BS		SDG No.:	Q2592
Lab Sample ID:	PB168885BS		Matrix:	TCLP
Analytical Method:	8270E		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050476.D	1	07/16/25 10:14	07/17/25 15:57	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	07/11/25	
Project:	Con Edison - East River Site 2		Date Received:	07/11/25	
Client Sample ID:	WC-SOIL-20250711MS		SDG No.:	Q2592	
Lab Sample ID:	Q2592-02MS		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143122.D	1	07/16/25 10:14	07/16/25 22:09	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	290		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	360		5.30	50.0	ug/L
95-48-7	2-Methylphenol	390		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	390		11.0	100	ug/L
67-72-1	Hexachloroethane	370		6.50	50.0	ug/L
98-95-3	Nitrobenzene	430		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	390		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	440		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	430		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	460		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	430		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	750		15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	99.9		23 - 138	67%	SPK: 150
13127-88-3	Phenol-d6	95.0		10 - 134	63%	SPK: 150
4165-60-0	Nitrobenzene-d5	80.6		67 - 132	81%	SPK: 100
321-60-8	2-Fluorobiphenyl	68.4		52 - 132	68%	SPK: 100
118-79-6	2,4,6-Tribromophenol	126		44 - 137	84%	SPK: 150
1718-51-0	Terphenyl-d14	69.6		42 - 152	70%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	147000	6.969			
1146-65-2	Naphthalene-d8	557000	8.245			
15067-26-2	Acenaphthene-d10	280000	10.004			
1517-22-2	Phenanthrene-d10	442000	11.492			
1719-03-5	Chrysene-d12	237000	14.127			
1520-96-3	Perylene-d12	284000	15.633			

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	07/11/25
Project:	Con Edison - East River Site 2		Date Received:	07/11/25
Client Sample ID:	WC-SOIL-20250711MS		SDG No.:	Q2592
Lab Sample ID:	Q2592-02MS		Matrix:	TCLP
Analytical Method:	8270E		% Solid:	0
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143122.D	1	07/16/25 10:14	07/16/25 22:09	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	07/11/25	
Project:	Con Edison - East River Site 2		Date Received:	07/11/25	
Client Sample ID:	WC-SOIL-20250711MSD		SDG No.:	Q2592	
Lab Sample ID:	Q2592-02MSD		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143123.D	1	07/16/25 10:14	07/16/25 22:38	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	290		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	360		5.30	50.0	ug/L
95-48-7	2-Methylphenol	400		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	400		11.0	100	ug/L
67-72-1	Hexachloroethane	380		6.50	50.0	ug/L
98-95-3	Nitrobenzene	430		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	390		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	440		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	430		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	470		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	440		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	710		15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	100		23 - 138	67%	SPK: 150
13127-88-3	Phenol-d6	96.3		10 - 134	64%	SPK: 150
4165-60-0	Nitrobenzene-d5	82.6		67 - 132	83%	SPK: 100
321-60-8	2-Fluorobiphenyl	68.6		52 - 132	69%	SPK: 100
118-79-6	2,4,6-Tribromophenol	127		44 - 137	85%	SPK: 150
1718-51-0	Terphenyl-d14	68.7		42 - 152	69%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	155000		6.969		
1146-65-2	Naphthalene-d8	590000		8.245		
15067-26-2	Acenaphthene-d10	297000		10.004		
1517-22-2	Phenanthrene-d10	462000		11.492		
1719-03-5	Chrysene-d12	243000		14.127		
1520-96-3	Perylene-d12	288000		15.633		

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	07/11/25	
Project:	Con Edison - East River Site 2		Date Received:	07/11/25	
Client Sample ID:	WC-SOIL-20250711MSD		SDG No.:	Q2592	
Lab Sample ID:	Q2592-02MSD		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143123.D	1	07/16/25 10:14	07/16/25 22:38	PB168885

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF071525.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 15 17:53:25 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BF143093.D 5 =BF143094.D 10 =BF143095.D 20 =BF143096.D 40 =BF143097.D 50 =BF143098.D 60 =BF143099.D 80 =BF143100.D

Compound		2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.637	0.635	0.640	0.627	0.661	0.626	0.586	0.630		3.58
3)	Pyridine	1.606	1.541	1.616	1.594	1.681	1.612	1.519	1.596		3.33
4)	n-Nitrosodimet...		0.788	0.825	0.816	0.879	0.840	0.812	0.827		3.70
5) S	2-Fluorophenol	1.334	1.272	1.315	1.255	1.305	1.232	1.148	1.266		4.98
6)	Aniline	2.330	2.254	2.280	2.205	2.351	2.196	2.069	2.241		4.26
7) S	Phenol-d6	1.685	1.616	1.640	1.565	1.643	1.547	1.441	1.591		5.11
8)	2-Chlorophenol	1.239	1.208	1.321	1.273	1.363	1.286	1.212	1.272		4.51
9)	Benzaldehyde		1.196	1.149	1.451	1.452	1.163	0.864	1.212		18.15
10) C	Phenol	1.766	1.742	1.782	1.709	1.780	1.694	1.584	1.723		4.07
11)	bis(2-Chloroet...	1.425	1.363	1.366	1.309	1.371	1.291	1.209	1.334		5.26
12)	1,3-Dichlorobe...	1.547	1.482	1.470	1.414	1.471	1.388	1.286	1.437		5.83
13) C	1,4-Dichlorobe...	1.570	1.473	1.504	1.412	1.482	1.386	1.292	1.446		6.27
14)	1,2-Dichlorobe...	1.468	1.409	1.430	1.364	1.417	1.321	1.241	1.379		5.57
15)	Benzyl Alcohol		1.126	1.204	1.167	1.255	1.189	1.132	1.179		4.09
16)	2,2'-oxybis(1-...	2.713	2.532	2.528	2.405	2.517	2.342	2.171	2.458		6.99
17)	2-Methylphenol	1.132	1.081	1.124	1.083	1.150	1.077	1.035	1.098		3.63
18)	Hexachloroethane	0.465	0.440	0.486	0.472	0.500	0.476	0.455	0.471		4.22
19) P	n-Nitroso-di-n...	0.956	1.029	0.988	1.002	0.964	1.017	0.945	0.897	0.975	4.44
20)	3+4-Methylphenols		1.399	1.419	1.356	1.421	1.293	1.189	1.346		6.74
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.538	0.516	0.500	0.469	0.482	0.451	0.419	0.482		8.33
23) S	Nitrobenzene-d5	0.283	0.319	0.363	0.372	0.399	0.390	0.370	0.357		11.59
24)	Nitrobenzene	0.282	0.317	0.357	0.356	0.385	0.366	0.347	0.344		9.94
25)	Isophorone	0.705	0.688	0.697	0.678	0.718	0.687	0.651	0.689		3.11
26) C	2-Nitrophenol	0.063	0.075	0.106	0.120	0.141	0.146	0.147	0.114		29.95
27)	2,4-Dimethylph...	0.325	0.322	0.326	0.320	0.332	0.315	0.299	0.320		3.38
28)	bis(2-Chloroet...	0.444	0.434	0.435	0.409	0.424	0.404	0.375	0.418		5.63
29) C	2,4-Dichloroph...	0.237	0.254	0.267	0.267	0.280	0.271	0.254	0.262		5.43
30)	1,2,4-Trichlor...	0.316	0.305	0.307	0.295	0.303	0.287	0.267	0.297		5.40
31)	Naphthalene	1.099	1.048	1.037	0.966	0.987	0.937	0.878	0.993		7.52
32)	Benzoic acid		0.080	0.111	0.144	0.159	0.171	0.172	0.139		26.27
33)	4-Chloroaniline	0.420	0.420	0.414	0.398	0.409	0.381	0.353	0.399		6.12
34) C	Hexachlorobuta...	0.185	0.176	0.183	0.177	0.182	0.176	0.163	0.178		4.10
35)	Caprolactam		0.073	0.078	0.083	0.087	0.086	0.083	0.082		6.36
36) C	4-Chloro-3-met...	0.285	0.290	0.296	0.293	0.307	0.293	0.277	0.291		3.26
37)	2-Methylnaphth...	0.644	0.626	0.615	0.580	0.597	0.562	0.524	0.592		6.91
38)	1-Methylnaphth...	0.678	0.660	0.631	0.603	0.615	0.578	0.537	0.615		7.81

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF071525.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.616	0.600	0.602	0.578	0.601	0.561	0.522	0.583	5.56	
41) P	Hexachlorocycl...	0.254	0.306	0.323	0.356	0.356	0.344	0.323	12.16		
42) S	2,4,6-Tribromo...	0.136	0.157	0.181	0.188	0.204	0.194	0.188	0.178	13.23	
43) C	2,4,6-Trichlor...	0.300	0.312	0.377	0.376	0.400	0.393	0.377	0.362	10.94	
44)	2,4,5-Trichlor...	0.336	0.367	0.390	0.393	0.418	0.377	0.378	0.380	6.62	
45) S	2-Fluorobiphenyl	1.777	1.645	1.602	1.456	1.460	1.355	1.238	1.505	12.18	
46)	1,1'-Biphenyl	1.750	1.687	1.671	1.567	1.610	1.496	1.384	1.595	7.84	
47)	2-Chloronaphth...	1.271	1.203	1.206	1.159	1.195	1.120	1.048	1.172	6.10	
48)	2-Nitroaniline	0.180	0.224	0.296	0.328	0.372	0.363	0.355	0.303	24.51	
49)	Acenaphthylene	2.060	2.011	2.034	1.959	1.989	1.867	1.746	1.952	5.65	
50)	Dimethylphthalate	1.310	1.259	1.292	1.257	1.302	1.219	1.169	1.258	4.01	
51)	2,6-Dinitrotol...	0.133	0.176	0.222	0.234	0.262	0.262	0.247	0.219	22.07	
52) C	Acenaphthene	1.273	1.204	1.173	1.127	1.168	1.088	1.032	1.152	6.85	
53)	3-Nitroaniline	0.180	0.227	0.283	0.288	0.318	0.313	0.295	0.272	18.56	
54) P	2,4-Dinitrophenol	0.038	0.056	0.069	0.086	0.090	0.097	0.073	31.18		
55)	Dibenzofuran	1.895	1.795	1.785	1.697	1.729	1.604	1.516	1.717	7.37	
56) P	4-Nitrophenol	0.164	0.201	0.224	0.239	0.237	0.232	0.216	13.37		
57)	2,4-Dinitrotol...	0.147	0.197	0.253	0.295	0.326	0.319	0.320	0.265	26.25	
58)	Fluorene	1.478	1.387	1.328	1.246	1.284	1.186	1.105	1.288	9.67	
59)	2,3,4,6-Tetrac...	0.253	0.274	0.302	0.312	0.339	0.323	0.309	0.302	9.71	
60)	Diethylphthalate	1.221	1.217	1.257	1.225	1.260	1.187	1.130	1.214	3.68	
61)	4-Chlorophenyl...	0.666	0.642	0.631	0.614	0.625	0.586	0.555	0.617	5.92	
62)	4-Nitroaniline	0.169	0.208	0.237	0.249	0.270	0.263	0.259	0.236	15.25	
63)	Azobenzene	1.425	1.399	1.375	1.328	1.381	1.282	1.208	1.343	5.65	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...	0.035	0.053	0.064	0.079	0.084	0.086	0.067	30.46		
66) c	n-Nitrosodiphe...	0.735	0.709	0.736	0.682	0.717	0.685	0.654	0.703	4.30	
67)	4-Bromophenyl-...	0.231	0.219	0.234	0.222	0.239	0.228	0.219	0.227	3.38	
68)	Hexachlorobenzene	0.249	0.237	0.247	0.235	0.250	0.239	0.231	0.241	3.12	
69)	Atrazine	0.158	0.171	0.184	0.181	0.195	0.185	0.184	0.180	6.72	
70) C	Pentachlorophenol	0.098	0.125	0.132	0.149	0.145	0.145	0.132	14.47		
71)	Phenanthrene	1.190	1.136	1.111	1.042	1.089	1.021	0.968	1.079	6.95	
72)	Anthracene	1.158	1.126	1.126	1.057	1.106	1.042	0.983	1.085	5.62	
73)	Carbazole	1.038	1.026	1.008	0.938	0.998	0.927	0.873	0.972	6.27	
74)	Di-n-butylphth...	0.746	0.846	0.913	0.921	0.983	0.942	0.897	0.893	8.62	
75) C	Fluoranthene	1.044	1.053	1.009	0.960	0.998	0.938	0.869	0.981	6.59	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine	0.664	0.786	0.834	0.674	0.616	0.514	0.681	17.02		
78)	Pyrene	2.018	1.956	1.955	1.842	1.918	1.716	1.585	1.856	8.33	
79) S	Terphenyl-d14	1.563	1.471	1.437	1.343	1.385	1.246	1.133	1.368	10.53	
80)	Butylbenzylphth...	0.179	0.233	0.310	0.343	0.385	0.395	0.384	0.318	26.22	
81)	Benzo(a)anthra...	1.296	1.304	1.406	1.347	1.403	1.326	1.247	1.333	4.34	
82)	3,3'-Dichlorob...	0.314	0.406	0.385	0.428	0.411	0.377	0.387	10.35		
83)	Chrysene	1.286	1.234	1.211	1.201	1.270	1.228	1.162	1.227	3.41	
84)	Bis(2-ethylhex...	0.273	0.343	0.479	0.515	0.578	0.591	0.583	0.480	26.25	
85) c	Di-n-octyl pht...	0.474	0.811	0.922	1.073	1.115	1.167	0.927	27.83		

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF071525.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.345	1.411	1.526	1.521	1.620	1.519	1.470	1.488	5.98
88)	Benzo(b)fluora...	1.063	1.191	1.171	1.082	1.297	1.185	1.087	1.154	7.17
89)	Benzo(k)fluora...	1.175	1.066	1.111	1.148	1.088	1.060	1.056	1.101	4.23
90) C	Benzo(a)pyrene	0.989	1.019	1.108	1.124	1.196	1.124	1.078	1.091	6.41
91)	Dibenzo(a,h)an...	1.114	1.162	1.266	1.240	1.330	1.222	1.178	1.216	5.90
92)	Benzo(g,h,i)pe...	1.148	1.187	1.263	1.258	1.349	1.260	1.214	1.240	5.20

(#) = Out of Range

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
 Method File : 8270-BM070925.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 08 18:32:25 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BM050377.D 5 =BM050378.D 10 =BM050379.D 20 =BM050380.D 40 =BM050381.D 50 =BM050382.D 60 =BM050383.D 80 =BM050384.D

	Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.504	0.466	0.482	0.460	0.501	0.474	0.457	0.478	3.99	
3)	Pyridine	1.249	1.197	1.256	1.228	1.336	1.279	1.236	1.254	3.52	
4)	n-Nitrosodimet...	0.562	0.595	0.572	0.620	0.595	0.565	0.585		3.88	
5) S	2-Fluorophenol	1.126	1.079	1.149	1.152	1.258	1.202	1.168	1.162	4.89	
6)	Aniline	1.784	1.702	1.844	1.878	2.042	1.987	1.921	1.880	6.20	
7) S	Phenol-d6	1.367	1.327	1.441	1.466	1.607	1.547	1.498	1.465	6.67	
8)	2-Chlorophenol	1.135	1.105	1.219	1.217	1.341	1.301	1.258	1.225	6.89	
9)	Benzaldehyde	0.944	0.985	0.897	0.837	0.694		0.871		13.03	
10) C	Phenol	1.460	1.412	1.516	1.508	1.666	1.598	1.535	1.528	5.51	
11)	bis(2-Chloroet...	1.202	1.134	1.223	1.196	1.324	1.268	1.218	1.223	4.88	
12)	1,3-Dichlorobe...	1.504	1.425	1.496	1.448	1.580	1.511	1.465	1.490	3.40	
13) C	1,4-Dichlorobe...	1.553	1.459	1.536	1.465	1.610	1.538	1.487	1.521	3.57	
14)	1,2-Dichlorobe...	1.481	1.395	1.454	1.410	1.541	1.485	1.428	1.456	3.48	
15)	Benzyl Alcohol	0.907	0.995	1.022	1.125	1.090	1.052	1.032		7.45	
16)	2,2'-oxybis(1-...	1.835	1.741	1.818	1.752	1.907	1.821	1.741	1.802	3.41	
17)	2-Methylphenol	0.922	0.898	0.980	0.990	1.089	1.048	1.008	0.991	6.71	
18)	Hexachloroethane	0.518	0.501	0.532	0.520	0.574	0.556	0.543	0.535	4.66	
19) P	n-Nitroso-di-n...	0.790	0.805	0.808	0.888	0.911	1.005	0.968	0.924	0.887	9.01
20)	3+4-Methylphenols	1.170	1.296	1.325	1.458	1.393	1.338	1.330		7.29	
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.493	0.479	0.501	0.491	0.534	0.513	0.489	0.500	3.69	
23) S	Nitrobenzene-d5	0.362	0.355	0.390	0.392	0.429	0.415	0.400	0.392	6.77	
24)	Nitrobenzene	0.333	0.329	0.353	0.345	0.377	0.362	0.349	0.350	4.73	
25)	Isophorone	0.571	0.574	0.634	0.640	0.702	0.679	0.652	0.636	7.73	
26) C	2-Nitrophenol	0.107	0.112	0.132	0.147	0.167	0.167	0.167	0.143	18.26	
27)	2,4-Dimethylph...	0.295	0.275	0.301	0.299	0.329	0.317	0.308	0.303	5.63	
28)	bis(2-Chloroet...	0.407	0.395	0.422	0.417	0.454	0.436	0.420	0.422	4.61	
29) C	2,4-Dichloroph...	0.277	0.279	0.311	0.312	0.344	0.334	0.325	0.312	8.26	
30)	1,2,4-Trichlor...	0.370	0.349	0.369	0.362	0.397	0.385	0.378	0.373	4.21	
31)	Naphthalene	1.033	0.975	1.024	0.988	1.071	1.029	0.992	1.016	3.26	
32)	Benzoic acid	0.100	0.139	0.164	0.195	0.194	0.199	0.165		23.89	
33)	4-Chloroaniline	0.398	0.397	0.427	0.429	0.465	0.453	0.437	0.429	6.00	
34) C	Hexachlorobuta...	0.222	0.209	0.222	0.221	0.244	0.239	0.235	0.228	5.37	
35)	Caprolactam	0.064	0.081	0.087	0.096	0.093	0.090	0.085		13.87	
36) C	4-Chloro-3-met...	0.258	0.256	0.288	0.295	0.326	0.313	0.302	0.291	9.10	
37)	2-Methylnaphth...	0.611	0.595	0.638	0.635	0.693	0.671	0.647	0.641	5.19	
38)	1-Methylnaphth...	0.651	0.629	0.674	0.673	0.736	0.707	0.681	0.679	5.17	

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
 Method File : 8270-BM070925.M

39) I	Acenaphthene-d10	-----ISTD-----								
40)	1,2,4,5-Tetrac...	0.590	0.569	0.611	0.622	0.700	0.683	0.677	0.636	7.93
41) P	Hexachlorocycl...	0.320	0.361	0.396	0.450	0.454	0.459	0.407		14.08
42) S	2,4,6-Tribromo...	0.182	0.195	0.230	0.248	0.285	0.280	0.280	0.243	17.40
43) C	2,4,6-Trichlor...	0.330	0.335	0.377	0.394	0.443	0.434	0.427	0.391	11.81
44)	2,4,5-Trichlor...	0.359	0.376	0.415	0.430	0.484	0.470	0.460	0.428	11.13
45) S	2-Fluorobiphenyl	1.572	1.538	1.611	1.592	1.753	1.697	1.575	1.620	4.74
46)	1,1'-Biphenyl	1.464	1.419	1.479	1.441	1.586	1.527	1.471	1.484	3.79
47)	2-Chloronaphth...	1.164	1.126	1.187	1.150	1.267	1.217	1.171	1.183	3.93
48)	2-Nitroaniline	0.197	0.211	0.253	0.278	0.312	0.302	0.295	0.264	17.17
49)	Acenaphthylene	1.645	1.658	1.795	1.775	1.956	1.875	1.811	1.788	6.21
50)	Dimethylphthalate	1.323	1.285	1.368	1.351	1.506	1.432	1.373	1.377	5.29
51)	2,6-Dinitrotol...	0.196	0.224	0.263	0.273	0.308	0.297	0.289	0.264	15.41
52) C	Acenaphthene	1.086	1.050	1.126	1.111	1.231	1.196	1.158	1.137	5.51
53)	3-Nitroaniline	0.204	0.231	0.278	0.294	0.330	0.319	0.311	0.281	16.81
54) P	2,4-Dinitrophenol	0.074	0.097	0.117	0.140	0.147	0.150	0.121		25.40
55)	Dibenzofuran	1.747	1.674	1.756	1.709	1.880	1.784	1.711	1.751	3.84
56) P	4-Nitrophenol	0.182	0.225	0.245	0.276	0.265	0.258	0.242		14.15
57)	2,4-Dinitrotol...	0.234	0.277	0.340	0.371	0.422	0.407	0.401	0.350	20.31
58)	Fluorene	1.348	1.328	1.428	1.419	1.582	1.522	1.470	1.443	6.29
59)	2,3,4,6-Tetrac...	0.299	0.306	0.355	0.365	0.405	0.397	0.392	0.360	11.92
60)	Diethylphthalate	1.217	1.210	1.319	1.309	1.454	1.379	1.320	1.315	6.51
61)	4-Chlorophenyl...	0.687	0.673	0.733	0.747	0.839	0.811	0.790	0.754	8.28
62)	4-Nitroaniline	0.192	0.228	0.284	0.311	0.350	0.331	0.320	0.288	20.10
63)	Azobenzene	1.085	1.116	1.234	1.208	1.335	1.272	1.204	1.208	7.15
64) I	Phenanthrene-d10	-----ISTD-----								
65)	4,6-Dinitro-2-...	0.057	0.077	0.094	0.109	0.113	0.117	0.095		24.90
66) c	n-Nitrosodiphe...	0.572	0.577	0.620	0.623	0.678	0.658	0.654	0.626	6.49
67)	4-Bromophenyl-...	0.196	0.194	0.211	0.218	0.242	0.240	0.241	0.220	9.61
68)	Hexachlorobenzene	0.239	0.233	0.253	0.256	0.283	0.279	0.278	0.260	7.78
69)	Atrazine	0.166	0.175	0.199	0.211	0.232	0.228	0.227	0.206	12.95
70) C	Pentachlorophenol	0.122	0.145	0.161	0.181	0.180	0.183	0.162		15.23
71)	Phenanthrene	1.119	1.070	1.119	1.101	1.206	1.163	1.143	1.132	3.91
72)	Anthracene	1.028	1.017	1.109	1.117	1.233	1.204	1.179	1.127	7.44
73)	Carbazole	0.939	0.944	1.009	1.010	1.108	1.073	1.052	1.019	6.23
74)	Di-n-butylphth...	0.954	0.961	1.091	1.140	1.252	1.216	1.196	1.116	10.76
75) C	Fluoranthene	1.088	1.079	1.182	1.233	1.366	1.339	1.327	1.230	9.67
76) I	Chrysene-d12	-----ISTD-----								
77)	Benzidine	0.370	0.521	0.553	0.601	0.552	0.464	0.510		16.08
78)	Pyrene	1.198	1.177	1.260	1.260	1.388	1.366	1.316	1.281	6.26
79) S	Terphenyl-d14	1.095	1.110	1.225	1.258	1.298	1.094	0.878	1.137	12.43
80)	Butylbenzylpht...	0.317	0.334	0.401	0.441	0.492	0.491	0.479	0.422	17.39
81)	Benzo(a)anthra...	1.204	1.202	1.282	1.282	1.428	1.379	1.345	1.303	6.57
82)	3,3'-Dichlorob...	0.351	0.406	0.429	0.498	0.487	0.465	0.439		12.66
83)	Chrysene	1.162	1.153	1.200	1.209	1.327	1.303	1.249	1.229	5.45
84)	Bis(2-ethylhex...	0.499	0.546	0.658	0.705	0.780	0.763	0.742	0.670	16.33
85) c	Di-n-octyl pht...	0.742	0.931	1.058	1.193	1.194	1.183	1.050		17.44

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
Method File : 8270-BM070925.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.208	1.238	1.374	1.415	1.603	1.564	1.551	1.422	11.17
88)	Benzo(b)fluora...	1.090	1.075	1.177	1.230	1.368	1.344	1.327	1.230	9.83
89)	Benzo(k)fluora...	1.105	1.138	1.247	1.263	1.444	1.387	1.350	1.276	9.86
90) C	Benzo(a)pyrene	0.962	0.985	1.099	1.156	1.314	1.281	1.264	1.152	12.41
91)	Dibenzo(a,h)an...	1.022	1.035	1.159	1.193	1.352	1.320	1.301	1.198	11.23
92)	Benzo(g,h,i)pe...	0.992	1.001	1.097	1.121	1.263	1.232	1.216	1.132	9.73

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>PARS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2592</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>07/16/2025 19:39</u>
Lab File ID:	<u>BF143117.D</u>	Init. Calib. Date(s):	<u>07/15/2025 07/15/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>13:04 16:35</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.596	1.575		-1.3	
2-Fluorophenol	1.266	1.249		-1.3	
Phenol-d6	1.591	1.563		-1.8	
1,4-Dichlorobenzene	1.446	1.418		-1.9	20.0
2-Methylphenol	1.098	1.094		-0.4	
3+4-Methylphenols	1.346	1.368		1.6	
Nitrobenzene-d5	0.357	0.402		12.6	
Hexachloroethane	0.471	0.486		3.2	
Nitrobenzene	0.344	0.369		7.3	
Hexachlorobutadiene	0.178	0.178		0.0	20.0
2,4,6-Trichlorophenol	0.362	0.386		6.6	20.0
2-Fluorobiphenyl	1.505	1.421		-5.6	
2,4,5-Trichlorophenol	0.380	0.387		1.8	
2,4-Dinitrotoluene	0.265	0.307		15.8	
2,4,6-Tribromophenol	0.178	0.191		7.3	
Hexachlorobenzene	0.241	0.241		0.0	
Pentachlorophenol	0.132	0.144		9.1	20.0
Terphenyl-d14	1.368	1.302		-4.8	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>PARS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2592</u>
Instrument ID:	<u>BNA_M</u>	Calibration Date/Time:	<u>07/17/2025 13:42</u>
Lab File ID:	<u>BM050474.D</u>	Init. Calib. Date(s):	<u>07/08/2025 07/08/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>12:39 17:22</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.254	1.241		-1.0	
2-Fluorophenol	1.162	1.165		0.3	
Phenol-d6	1.465	1.479		1.0	
1,4-Dichlorobenzene	1.521	1.487		-2.2	20.0
2-Methylphenol	0.991	0.990		-0.1	
3+4-Methylphenols	1.330	1.341		0.8	
Nitrobenzene-d5	0.392	0.393		0.3	
Hexachloroethane	0.535	0.522		-2.4	
Nitrobenzene	0.350	0.343		-2.0	
Hexachlorobutadiene	0.228	0.231		1.3	20.0
2,4,6-Trichlorophenol	0.391	0.429		9.7	20.0
2-Fluorobiphenyl	1.620	1.584		-2.2	
2,4,5-Trichlorophenol	0.428	0.466		8.9	
2,4-Dinitrotoluene	0.350	0.407		16.3	
2,4,6-Tribromophenol	0.243	0.284		16.9	
Hexachlorobenzene	0.260	0.264		1.5	
Pentachlorophenol	0.162	0.180		11.1	20.0
Terphenyl-d14	1.137	1.120		-1.5	

All other compounds must meet a minimum RRF of 0.010.