

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2592

Analytical Method: 8270E

Client: PARSONS Engineering of New York, Inc.

DataFile: BM050476.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168885BS	Pyridine	50	39.1	ug/L	78				29	97	
	1,4-Dichlorobenzene	50	43.0	ug/L	86				76	103	
	2-Methylphenol	50	47.4	ug/L	95				69	109	
	3+4-Methylphenols	50	47.8	ug/L	96				67	106	
	Hexachloroethane	50	43.1	ug/L	86				76	118	
	Nitrobenzene	50	44.3	ug/L	89				58	106	
	Hexachlorobutadiene	50	44.8	ug/L	90				69	101	
	2,4,6-Trichlorophenol	50	50.9	ug/L	102				61	110	
	2,4,5-Trichlorophenol	50	51.7	ug/L	103				70	106	
	2,4-Dinitrotoluene	50	49.2	ug/L	98				60	115	
	Hexachlorobenzene	50	47.8	ug/L	96				73	106	
	Pentachlorophenol	100	110	ug/L	110				47	114	