

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>PARS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2592</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>07/16/2025</u> <u>09:26</u>
Lab File ID:	<u>VX047016.D</u>	Init. Calib. Date(s):	<u>07/02/2025</u> <u>07/02/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:11</u> <u>14:31</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.569	0.550		-3.34	20
1,1-Dichloroethene	0.542	0.535		-1.29	20
2-Butanone	0.274	0.363		32.48	20
Carbon Tetrachloride	0.484	0.468		-3.31	20
Chloroform	1.087	1.131		4.05	20
Benzene	1.385	1.335		-3.61	20
1,2-Dichloroethane	0.470	0.485		3.19	20
Trichloroethene	0.361	0.334		-7.48	20
Tetrachloroethene	0.333	0.298		-10.51	20
Chlorobenzene	1.081	1.054	0.3	-2.5	20
1,2-Dichloroethane-d4	0.661	0.692		4.69	20
Dibromofluoromethane	0.339	0.335		-1.18	20
Toluene-d8	1.197	1.120		-6.43	20
4-Bromofluorobenzene	0.455	0.448		-1.54	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.