

Surrogate Summary

SDG No.: Q2592

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2592-02	WC-SOIL-20250711	1,2-Dichloroethane-d4	50	53.3	107		74	125
		Dibromofluoromethane	50	47.8	96		75	124
		Toluene-d8	50	51.0	102		86	113
		4-Bromofluorobenzene	50	55.8	112		77	121
VX0716WBL01	VX0716WBL01	1,2-Dichloroethane-d4	50	53.6	107		74	125
		Dibromofluoromethane	50	49.5	99		75	124
		Toluene-d8	50	48.1	96		86	113
		4-Bromofluorobenzene	50	52.5	105		77	121
VX0716WBS01	VX0716WBS01	1,2-Dichloroethane-d4	50	52.4	105		74	125
		Dibromofluoromethane	50	48.8	98		75	124
		Toluene-d8	50	48.1	96		86	113
		4-Bromofluorobenzene	50	52.0	104		77	121
VX0716WBSD0	VX0716WBSD01	1,2-Dichloroethane-d4	50	55.4	111		74	125
		Dibromofluoromethane	50	50.2	100		75	124
		Toluene-d8	50	48.0	96		86	113
		4-Bromofluorobenzene	50	52.0	104		77	121